

**UNCLASSIFIED**

---

---

**AD 404 011**

*Reproduced  
by the*

**DEFENSE DOCUMENTATION CENTER**

**FOR**

**SCIENTIFIC AND TECHNICAL INFORMATION**

**CAMERON STATION, ALEXANDRIA, VIRGINIA**



---

---

**UNCLASSIFIED**

NOTICE: When government or other drawings, specifications or other data are used for any purpose other than in connection with a definitely related government procurement operation, the U. S. Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use or sell any patented invention that may in any way be related thereto.

6334

USNRDL-TR-629  
14 March 1963

CONSTRUCTION OF A MODIFIED HENDERSON APPARATUS

by  
J. F. Pribnow  
M. S. Silverman

AD. 404011

**404 011**

DDC  
MAY 17 1963  
TISIA A

AS AD NO. 404011

**U.S. NAVAL RADIOLOGICAL  
DEFENSE LABORATORY**

**SAN FRANCISCO 24, CALIFORNIA**

**MICROBIOLOGY AND IMMUNOLOGY BRANCH**  
**M. S. Silverman, Head**

**BIOLOGICAL AND MEDICAL SCIENCES DIVISION**  
**E. L. Alpen, Head**

---

**ADMINISTRATIVE INFORMATION**

This work was accomplished under the Bureau of Medicine and Surgery Task MR005.08-5200, Subtask 2, Technical Objective AW-6, as described in the U. S. Naval Radiological Defense Laboratory Annual Report to the Bureau of Medicine and Surgery (OPNAV FORM 3910-1) of 31 December 1962, and is listed in the U. S. Naval Radiological Defense Laboratory Technical Program Summary for Fiscal Years 1963-1965 of 1 November 1962 under Program A3, Problem 2, entitled "Nuclear Warfare Aspects of Whole Body Ionizing Radiation." This study was supported through funds provided by the Bureau of Medicine and Surgery, and the Defense Atomic Support Agency under NWER Program A4c, Subtask 03.027.

---

*Eugene P. Cooper*  
Eugene P. Cooper  
Scientific Director

*E. B. Roth*  
E. B. Roth, CAPT USN  
Commanding Officer and Director

#### ABSTRACT

An apparatus for exposing mice to an aerosol of microorganisms has been developed. It is a modification of the Henderson apparatus, adapted for exposure of smaller animals. The design allows for placement of the complete apparatus in a bacteriological hood. Although of relatively simple design and operation, it has proven highly dependable.

## SUMMARY

### The Problem

To develop an apparatus for exposing mice to an aerosol of microorganisms. The apparatus should be of relative simplicity of design, and small enough to be placed in a bacteriological hood.

### The Findings

A modified Henderson apparatus was designed for exposing mice in groups of up to 40 animals. Because of its basic design, the apparatus could be placed in a bacteriological hood, and has proven both reliable and serviceable.

## INTRODUCTION

The apparatus originally described by Henderson (1) is used to generate an aerosol of microorganisms. The microorganisms, mixed with air, flow past the experimental animals to sample collecting stations. This apparatus was too large to be placed in a bacteriological hood, of elaborate design, and not adaptable for use with animals as small as mice.

The modification described in this report was developed to expose mice, which had received continuous low level gamma radiation, to varying doses of the microorganism, Listeria monocytogenes. The apparatus is relatively small, easy to operate, and has proven very serviceable.

## PHYSICAL DESCRIPTION

The components of the apparatus are shown in Fig. 1. They consisted of a pressure pump, a Wells atomizer, mixing chamber, animal exposure chambers, air samplers, and a vacuum pump.

The Wells atomizer (2), an internal spray type, was made of glass, and utilized a peripheral fluid jet nozzle.

The mixing chamber consisted of an air mixing section and an additional aerosol mixing section. The air mixing section was a




Fig. 1 Modified Henderson apparatus for respiratory infection of mice.



brass tube  $2 \frac{3}{4}$  inches in diameter and  $6 \frac{3}{4}$  inches long. Two atmospheric inlets were located at the top and bottom of the tube  $\frac{1}{2}$  inch from the intake end. The aerosol mixing chamber was a copper tube  $2 \frac{1}{4}$  inches in diameter and 24 inches long. A water manometer was connected to the top of the section. The two sections were connected by  $90^\circ$  angle copper tubing.

Chambers for animal exposure consisted of air tight cast aluminum rectangular boxes 11 inches long,  $2 \frac{1}{2}$  inches deep and  $3 \frac{1}{8}$  inches wide (inside measurement). To insure an air tight fit, a rubber gasket was permanently affixed to the outer perimeter of the cover. The aerosol inlet and outlet located at either end were protected from the animals by use of baffles. Placing of the animals in the exposure chamber was facilitated by construction of boxes from  $\frac{1}{4}$  inch mesh wire screening hardware cloth. These boxes, which fit inside the exposure chamber, were hinged at one end, thus providing a small opening for inserting or removing mice. The chambers could hold up to ten mice each, and the apparatus accommodated four such chambers.

The air sampler consisted of metal impingers supplied through the courtesy of Mr. W. R. Leif of the U. S. Naval Biological Laboratory, Oakland, California. The sampler was made from a piece of 5 inch stainless steel tubing (outside diameter  $\frac{1}{4}$  inch). On the bottom end was soldered a  $\frac{1}{32}$  inch plate with a critical

orifice made by use of a #66 drill. The impinger flow rate was  $6.00 \pm 0.048$  liters per minute. The sampler collection bottles were four ounce French Square bottles.

The air pressure pump, Model 1032 A, and the vacuum pump, Model 1033 S, were obtained from the Arthur B. Thomas Co., P. O. Box 779, Philadelphia 5, Pa.

#### OPERATION

The standard operating procedure for this apparatus was as follows:

- a. The Wells atomizer was filled with 50 ml brain heart infusion broth culture of the test organism.
- b. Ten mice were placed in each exposure chamber.
- c. Air sampler bottles were filled with 10 ml of 1% tryptose broth plus 3 drops of Dow Corning Antifoam B. The impinger orifice was positioned 1 inch from the bottle bottom.
- d. To air wash the animals, the vacuum pump was activated for about 45 seconds before starting the pressure pump. The manometer indicated a slight pressure differential. After the desired run time, the pressure pump was stopped about 45 seconds before the vacuum pump.
- e. Plate counts of serial dilution of the broth in the sampler bottles were made. The number of organisms each

mouse received was obtained by the following calculation:

$$\frac{\text{Sampler count} \times \text{ml of collecting fluid}}{\text{impinger flow rate} \times \text{exposure time}} = \frac{\text{organisms}}{\text{liter}}$$

$$\frac{\text{Organisms}}{\text{liter}} \times 1.25 \times 10^{-3} \times \text{weight of mouse} \times \text{exposure time} = \frac{\text{organisms}}{\text{mouse}}$$

$1.25 \times 10^{-3}$  = average value for respiratory volume of mouse expressed in liter per minute per gram body weight (4).

#### SAFETY PRECAUTIONS

Because of the pathogenic nature of the test organism, operation of the apparatus was carried out in a bacteriological hood. The hood was vented to the outside atmosphere through a series of bacteriological glass fiber filters. Air drawn through the same filter system was removed to maintain a negative pressure in the laboratory room. Operating personnel wore a standard Navy B W gas mask and protective clothing, as it was necessary to open the hood to exchange mice and samples.

#### CONCLUSIONS

The modified Henderson apparatus described has proven highly satisfactory for exposing up to 40 mice to an aerosol of microorganisms. Its compact design allowed it to be placed in a bacteriological hood, thus providing protection to the operating personnel. Operation over a period of two year has demonstrated that this is a very serviceable and dependable design.

#### REFERENCES

1. Henderson, D. W. 1952 An apparatus for the study of air borne infection. J. Hygiene, 50: 53-68.
2. Wells, W. F. 1940 An apparatus for the study of experimental air borne disease. Science, 91: 172-174.
3. Guyton, A. C. 1947 Measurement of the respiratory volumes of laboratory animals. Am. J. Physiol., 150: 70-77.

Biology and Medicine

INITIAL DISTRIBUTION

Copies

NAVY

3 Chief, Bureau of Ships (Code 335)  
1 Chief, Bureau of Ships (Code 320)  
2 Chief, Bureau of Medicine and Surgery  
1 Chief of Naval Operations (Op-07T)  
1 Chief of Naval Research (Code 104)  
3 Director, Naval Research Laboratory (Code 2021)  
1 Office of Naval Research (Code 422)  
1 Office of Naval Research (Code 441)  
10 Office of Naval Research, FPO, New York  
3 Naval Medical Research Institute  
1 OIG, Radiation Exposure Evaluation Laboratory  
1 Director, Aviation Medical Acceleration Laboratory  
1 U.S. Naval Postgraduate School, Monterey  
1 Commander, Naval Ordnance Laboratory, Silver Spring  
1 Naval Missile Center (Code 5700)  
1 U.S. Naval Hospital, San Diego  
1 CO, Naval Medical Research Unit No. 2  
1 CO, Naval Medical Field Research Laboratory, Camp Lejeune

ARMY

1 Chief of Research and Development (Atomic Division)  
1 Chief of Research and Development (Life Science Division)  
1 Deputy Chief of Staff for Military Operations (CBR)  
1 Chief of Engineers (ENGMCD-DE)  
1 Chief of Engineers (ENGCW)  
1 CG, Army Materiel Command (AMCRD-DE-NE)  
1 CG, USA CBR Agency  
3 CO, BW Laboratories  
1 CO, Fort McClellan, Alabama  
1 Commandant, Chemical Corps Schools (Library)  
1 CG, CBR Combat Developments Agency  
1 CO, Chemical Research and Development Laboratories  
1 Commander, Chemical Corps Nuclear Defense Laboratory  
1 Hq., Army Environmental Hygiene Agency  
1 CG, Aberdeen Proving Ground

1 CO, Army Medical Research Laboratory  
1 Army Medical Research and Nutrition Laboratory (MEDEN-AD)  
1 CO, Army Medical Service Combat Development Agency  
2 Medical Field Service School, Fort Sam Houston (Stimson Lib.)  
1 Brooks Army Medical Center (Dept. Prev. Med.)  
1 Director, Surgical Research Unit, Fort Sam Houston  
1 Director, Walter Reed Army Medical Center  
1 Hq., Army Nuclear Medicine Research Detach., Europe  
1 CG, Combat Developments Command (CDCMR-V)  
1 CG, Quartermaster Res. and Eng. Command  
1 Hq. Dugway Proving Ground  
3 The Surgeon General (MEDNE)  
1 Office of the Surgeon General (Combat Dev.)  
1 CG, Engineer Res. and Dev. Laboratory  
1 Director, Office of Special Weapons Development  
1 CG, Munitions Command  
1 CO, Frankford Arsenal  
1 CG, Army Missile Command

#### AIR FORCE

1 Assistant Chief of Staff, Intelligence (AFCIN-3B)  
6 CG, Aeronautical Systems Division (ASAPRD-NS)  
1 CO, Radiological Health Laboratory Division  
1 Director, USAF Project RAND  
1 Commandant, School of Aerospace Medicine, Brooks AFB  
1 CO, School of Aviation Medicine, Gunter AFB  
1 6571st Aeromedical Research Lab., Holloman AFB  
1 Radiobiological Laboratory  
1 Office of the Surgeon (SUP3.1), Strategic Air Command  
1 Office of the Surgeon General  
1 CG, Special Weapons Center, Kirtland AFB  
1 Director, Air University Library, Maxwell AFB  
2 Commander, Technical Training Wing, 3415th TTG  
1 Hq., Second Air Force, Barksdale AFB  
1 Commander, Electronic Systems Division (CRZT)

#### OTHER DOD ACTIVITIES

3 Chief, Defense Atomic Support Agency (Library)  
1 Commander, FC/DASA, Sandia Base (FCDV)  
1 Commander, FC/DASA, Sandia Base (FCTG5, Library)  
1 Commander, FC/DASA, Sandia Base (FCWT)  
2 Office of Civil Defense, Washington  
2 Civil Defense Unit, Army Library  
1 Armed Forces Institute of Pathology

20 Armed Services Technical Information Agency  
1 Director, Armed Forces Radiobiology Research Institute

AEC ACTIVITIES AND OTHERS

1 Research Analysis Corporation  
1 Life Science Officer, AEC, Washington  
1 Director, Division of Biology and Medicine  
1 NASA, Ames Research Center, Moffett Field  
1 Naval Attache, Stockholm (for Commodore Troell)  
1 Aerojet General, Azusa  
5 Argonne Cancer Research Hospital  
10 Argonne National Laboratory  
2 Atomic Bomb Casualty Commission  
1 AEC Scientific Representative, France  
1 AEC Scientific Representative, Japan  
3 Atomic Energy Commission, Washington  
2 Atomic Energy of Canada, Limited  
3 Atomics International  
2 Battelle Memorial Institute  
1 Borden Chemical Company  
3 Brookhaven National Laboratory  
1 Chicago Patent Group  
1 Colorado State University  
1 Columbia University (Rossi)  
1 Committee on the Effects of Atomic Radiation  
3 Defence Research Member  
2 duPont Company, Aiken  
1 duPont Company, Wilmington  
1 Edgerton, Germeshausen and Grier, Inc., Goleta  
1 Edgerton, Germeshausen and Grier, Inc., Las Vegas  
2 General Dynamics, Fort Worth  
2 General Electric Company, Cincinnati  
8 General Electric Company, Richland  
1 General Electric Company, St. Petersburg  
1 General Scientific Corporation  
1 Hughes Aircraft Company, Culver City  
1 Iowa State University  
1 Journal of Nuclear Medicine  
1 Knolls Atomic Power Laboratory  
2 Los Alamos Scientific Laboratory (Library)  
1 Lovelace Foundation  
1 Martin-Marietta Corporation  
1 Massachusetts Institute of Technology  
1 Mound Laboratory  
1 National Academy of Sciences

2 NASA, Scientific and Technical Information Facility  
 1 National Bureau of Standards (Taylor)  
 1 National Cancer Institute  
 1 National Lead Company of Ohio  
 1 National Library of Medicine  
 1 New Jersey State Department of Health  
 1 New York Operations Office  
 1 New York University (Eisenbud)  
 1 Office of Assistant General Counsel for Patents  
 2 Phillips Petroleum Company  
 4 Pratt and Whitney Aircraft Division  
 2 Public Health Service, Washington  
 1 Public Health Service, Las Vegas  
 1 Public Health Service, Montgomery  
 1 Sandia Corporation, Albuquerque  
 1 Union Carbide Nuclear Company (ORGDP)  
 5 Union Carbide Nuclear Company (ORNL)  
 1 Union Carbide Nuclear Company (Paducah Plant)  
 1 United Nuclear Corporation (NDA)  
 1 U. S. Geological Survey, Denver  
 1 U. S. Geological Survey, Menlo Park  
 1 U. S. Geological Survey, Naval Gun Factory  
 1 U. S. Geological Survey, Washington  
 1 U. S. Weather Bureau, Washington  
 1 University of California, Davis  
 3 University of California Lawrence Radiation Lab., Berkeley  
 2 University of California Lawrence Radiation Lab., Livermore  
 1 University of California, Los Angeles  
 1 University of California, San Francisco  
 1 University of Chicago Radiation Laboratory  
 1 University of Hawaii  
 1 University of Puerto Rico  
 1 University of Rochester (Atomic Energy Project)  
 1 University of Tennessee (UTA)  
 1 University of Utah  
 1 University of Washington (Donaldson)  
 1 Wayne State University  
 1 Westinghouse Electric Corporation (Rahilly)  
 1 Westinghouse Electric Corporation (NASA)  
 1 Western Reserve University (Friedell)  
 25 Technical Information Extension, Oak Ridge

USNRDL

41 USNRDL, Technical Information Division

DISTRIBUTION DATE: 30 April 1963



<p>Naval Radiological Defense Laboratory USNRDL-TR-629</p> <p>CONSTRUCTION OF A MODIFIED HENDERSON APPARATUS by J. F. Pribnow and M. S. Silverman 14 March 1963 13 p. illus. 3 refs.</p> <p>UNCLASSIFIED</p> <p>An apparatus for exposing mice to an aerosol of microorganisms has been developed. It is a modification of the Henderson apparatus, adapted for exposure of smaller animals. The design allows for placement of the complete</p> <p>(over)</p>	<p>Naval Radiological Defense Laboratory USNRDL-TR-629</p> <p>CONSTRUCTION OF A MODIFIED HENDERSON APPARATUS by J. F. Pribnow and M. S. Silverman 14 March 1963 13 p. illus. 3 refs.</p> <p>UNCLASSIFIED</p> <p>An apparatus for exposing mice to an aerosol of microorganisms has been developed. It is a modification of the Henderson apparatus, adapted for exposure of smaller animals. The design allows for placement of the complete</p> <p>(over)</p>
<ol style="list-style-type: none"> <li>1. Aerosols.</li> <li>2. Exposure.</li> <li>3. Microorganisms.</li> <li>4. Test equipment.</li> <li>5. Mice.</li> </ol> <ol style="list-style-type: none"> <li>I. Pribnow, J. F.</li> <li>II. Silverman, M. S.</li> <li>III. Title.</li> <li>IV. MR005.08-5200.</li> </ol> <p><u>UNCLASSIFIED</u></p>	<ol style="list-style-type: none"> <li>1. Aerosols.</li> <li>2. Exposure.</li> <li>3. Microorganisms.</li> <li>4. Test equipment.</li> <li>5. Mice.</li> </ol> <ol style="list-style-type: none"> <li>I. Pribnow, J. F.</li> <li>II. Silverman, M. S.</li> <li>III. Title.</li> <li>IV. MR005.08-5200.</li> </ol> <p><u>UNCLASSIFIED</u></p>
<p>apparatus in a bacteriological hood. Although of relatively simple design and operation, it has proven highly dependable.</p> <p><u>UNCLASSIFIED</u></p>	<p>apparatus in a bacteriological hood. Although of relatively simple design and operation, it has proven highly dependable.</p> <p><u>UNCLASSIFIED</u></p>