



Submarine Base, Groton, Conn.

MEMORANDUM REPORT NO. 68-11

LIMITATIONS ON THE USE OF CIRCUMAURAL EARPHONES

by

Benson, R. W., Charan, K. K., Day, J. W., Harris, J. D.,
Niemoeller, A. F., Rudmose, W., Shaw, E. A. G., and Weissler, P.

Bureau of Medicine and Surgery, Navy Department
Research Work Unit MF022.01.04-9003.1

Approved and Released by:

Gerald J. Duffner, CAPT, MC, USN
COMMANDING OFFICER
Naval Submarine Medical Center

29 May 1968



This document has been approved for public release and sale; its distribution is unlimited.

THE UNIVERSITY OF CHICAGO
LIBRARY

1950

UNIVERSITY OF CHICAGO LIBRARY

U.S. NAVAL SUBMARINE MEDICAL CENTER
NAVAL SUBMARINE BASE NEW LONDON
GROTON, CONNECTICUT 06340

MEMORANDUM REPORT NO. 68-11

LIMITATIONS ON THE USE OF CIRCUMAUURAL EARPHONES

MF 022.01.04-9003.1

29 May 1968

Investigators:

Benson, R. W., Charan, K. K., Day, J. W., Harris, J. D.
Niemoeller, A. F., Rudmose, W., Shaw, E.A.G., and Weissler, P.

Submitted by:

J. Donald Harris
J. Donald Harris, Ph.D.
Head, Auditory Research Branch

Reviewed and Approved by:

Charles F. Gell
Charles F. Gell, M.D., D.Sc.
Scientific Director

Approved and Released by:

Gerald J. Duffner
Gerald J. Duffner
Captain, MC USN
Commanding Officer

SUMMARY PAGE

basis for this calibration. For this reason, the committee urges research on circumaural earphones, particularly on their construction, use, and calibration.

To summarize, it is the author's opinion that there are not sufficient data available at the present time to enable one to use circumaural earphones whenever sound-pressure levels of pure tones at the eardrum must be accurately defined, without thoroughly investigating *first* the calibration of the system consisting of the particular earphones and ears (both real and artificial) under consideration and *second* the reliability of this calibration. At this time the routine use of circumaural earphones in clinical and industrial audiometry cannot be justified, but should be restricted to laboratory practices where investigation of calibration procedures in relation to the specific earphones can be carried out.

CHARAN, K. K., COX, J. R., AND NIEMOELLER, A. F. (1965). "An Evaluation of New Couplers for Circumaural Earphones," *J. Acoust. Soc. Am.* *38* 945-955.

FORSYTH, S. E. (1966). "The Significance of Variances in the Real-Ear Sensitivity of a Circumaural Earphone," *J. Sound Vibration* *4* (1), 45-50.

SHAW, E. A. G. (1966). "Ear-Canal Pressure Generated by Circumaural and Supra-Aural Earphones," *J. Acoust. Soc. Am.* *39*, 471-479.

SHAW, E. A. G., AND THIESSEN, G. J. (1962). "Acoustics of Circumaural Earphones on Auditory Threshold," *J. Acoust. Soc. Am.* *34*, 1233-1246.

TILLMAN, T. W., AND GISH, K. D. (1964). "Comments on the Effects of Circumaural Earphones on Auditory Threshold," *J. Acoust. Soc. Am.* *36*, 969-970.

UNCLASSIFIED

Security Classification

DOCUMENT CONTROL DATA - R & D

(Security classification of title, body of abstract and indexing annotation must be entered when the overall report is classified)

1. ORIGINATING ACTIVITY (Corporate author) U.S. NAVAL SUBMARINE MEDICAL CENTER, Submarine Medical Research Laboratory		2a. REPORT SECURITY CLASSIFICATION UNCLASSIFIED	
		2b. GROUP	
3. REPORT TITLE LIMITATIONS ON THE USE OF CIRCUMAURAL EARPHONES			
4. DESCRIPTIVE NOTES (Type of report and Inclusive dates) Interim Report			
5. AUTHOR(S) (First name, middle initial, last name) R. W. Benson, K. K. Charan, J. W. Day, J. D. Harris, A. F. Niemoeller, W. Rudmose, E.A.G. Shaw, and P. Weissler			
6. REPORT DATE 29 May 1968		7a. TOTAL NO. OF PAGES 2	7b. NO. OF REFS Five
8a. CONTRACT OR GRANT NO.		9a. ORIGINATOR'S REPORT NUMBER(S) Memorandum Report No. 68-11	
b. PROJECT NO. MF 022.01.04-9003.1			
c.		9b. OTHER REPORT NO(S) (Any other numbers that may be assigned this report)	
d.			
10. DISTRIBUTION STATEMENT This document has been approved for public release and sale; its distribution is unlimited.			
11. SUPPLEMENTARY NOTES		12. SPONSORING MILITARY ACTIVITY U.S. Naval Submarine Medical Center Box 600, Naval Submarine Base New London Groton, Connecticut 06340	
13. ABSTRACT A literature survey and interpretation by a panel of experts of the NAS-NRC Committee on Hearing, Bioacoustics and Biomechanics shows that the now commonly-used large earmuffs, designed to protect hearing, cannot indiscriminately replace the small earphone cushion currently the standard for audiometry. If they are so used, hearing losses at 3000- 8000 cycles/second should be rechecked with the standard cushion.			

UNCLASSIFIED

Security Classification

14. KEY WORDS	LINK A		LINK B		LINK C	
	ROLE	WT	ROLE	WT	ROLE	WT
Protective equipment (for hearing in noise)						
Audiometer Equipment						
Earphone Calibration						
Circumaural Earphone Cushion						