

UNCLASSIFIED

AD NUMBER
AD848652
NEW LIMITATION CHANGE
TO Approved for public release, distribution unlimited
FROM Distribution authorized to U.S. Gov't. agencies and their contractors; Critical Technology; DEC 1968. Other requests shall be referred to Department of the Army, Fort Detrick, Attn: Technical Release Branch/TID, Frederick, MD 21701.
AUTHORITY
BDRL, D/A ltr, 29 Sep 1071

THIS PAGE IS UNCLASSIFIED

AD

TECHNICAL MANUSCRIPT 494

AD848652

ESTIMATION OF PRECIPITATING ANTIBODY
USING A TURBIDIMETRIC TECHNIQUE

William F. Vincent
Earl W. Harris
Sidney Yaverbaum

DECEMBER 1968

STATEMENT #2 UNCLASSIFIED

This document is subject to special export controls and each transmittal to foreign governments or foreign nationals may be made only with prior approval of Dept. of Army, Fort Detrick, ATTN: Technical Release Branch/TID, Frederick, Maryland 21701

DEPARTMENT OF THE ARMY
Fort Detrick
Frederick, Maryland

D D C
RECEIVED
MAR 7 1969
C

**Best
Available
Copy**

DISPOSITION INSTRUCTIONS		
DDCI	WRITE SECTION	<input type="checkbox"/>
DDC	DIFF SECTION	<input checked="" type="checkbox"/>
DDC	OTHER	<input type="checkbox"/>
DISTRIBUTION/AVAILABILITY CODES		
DIST.	AVAIL.	SPECIAL
2		

Reproduction of this publication in whole or in part is prohibited except with permission of the Commanding Officer, Fort Detrick, ATTN: Technical Releases Branch, Technical Information Division, Fort Detrick, Frederick, Maryland, 21701. However, DDC is authorized to reproduce the publication for United States Government purposes.

DDC AVAILABILITY NOTICES

Qualified requesters may obtain copies of this publication from DDC.

Foreign announcement and dissemination of this publication by DDC is not authorized.

Release or announcement to the public is not authorized.

DISPOSITION INSTRUCTIONS

Destroy this publication when it is no longer needed. Do not return it to the originator.

The findings in this publication are not to be construed as an official Department of the Army position, unless so designated by other authorized documents.

DEPARTMENT OF THE ARMY
Fort Detrick
Frederick, Maryland 21701

TECHNICAL MANUSCRIPT 494

ESTIMATION OF PRECIPITATING ANTIBODY
USING A TURBIDIMETRIC TECHNIQUE

William F. Vincent

Earl W. Harris

Sidney Yaverbaum

Physical Defense Division
COMMODITY DEVELOPMENT & ENGINEERING LABORATORIES

Project 1B662706A071

December 1968

BLANK PAGE

In conducting the research described in this report, the investigators adhered to the "Guide for Laboratory Animal Facilities and Care," as promulgated by the Committee on the Guide for Laboratory Animal Facilities and Care of the Institute of Laboratory Animal Resources, National Academy of Sciences-National Research Council.

ABSTRACT

The development of turbidity resulting from the precipitation between bovine serum albumin (BSA) and anti-BSA serum was studied in the spectrophotometer. The equivalence ratio and the titer of precipitating antibody could be determined by this method. Data obtained with the turbidimetric method correlated well with those obtained using the classical tube precipitin determination. Turbidimetric measurements were also applied successfully to the measurement of antibody titer in two other systems. Antibody titer can be determined with this method in about an hour as compared with the 3 or more days required by the tube test. In addition, this technique requires substantially less reagent than other methods.

I. INTRODUCTION

The combination of antibody with a soluble antigen is often accompanied by visible turbidity. The extent and rate of turbidity development are usually functions of the relative concentrations of antigen and antibody. Martin² and Boyden and DeFalco¹ were among the first to apply turbidimetric measurements to the study of precipitation. The latter investigators speculated that such measurements would be ideally suited to extensive serologic studies and would eliminate the use of tedious quantitative determinations. The use of turbidimetric measurements to study precipitation has resulted in the detection of small deviations not observable with other techniques. Gitlin and Edelhoch,⁴ Junge, Junge, and Krebs,⁶ and Hawkins⁵ employed turbidimetric measurements for the study of antigen-antibody interactions in precipitation. Little attention, however, has been devoted to the employment of such techniques for the routine determination of precipitating antibody.

This report describes investigations that employ a spectrophotometric technique to measure the rate of turbidity development in the precipitin reaction and apply these data to the determination of antibody titer.

II. MATERIALS AND METHODS

A. ANTIGENS

Crystalline bovine serum albumin (BSA) and bovine γ -globulin (BGG) were purchased from Pentex Corporation. Periodate-oxidized complexes of these proteins with adenosine-5'-monophosphate (AMP) were prepared according to the method of Erlanger and Beiser.³ For both immunization and testing, the antigens were suspended in 0.01 M phosphate buffer (pH 6.8) prepared in 0.15 M NaCl. This buffered saline was employed for all dilutions and suspensions.

B. ANTISERA

Antisera against BSA, BGG, and AMP-BSA were prepared in New Zealand white rabbits by intramuscular immunization with 4 to 6 weekly injections of 10 mg each of antigen emulsified in Freund's complete adjuvant.

C. PRECIPITIN REACTIONS

Equal volumes (1 ml) of serial dilutions of antigen and a constant concentration of antiserum were mixed and incubated at 37 C for 2 hours. After refrigeration at 4 C for 48 to 72 hours, the precipitates were collected by centrifugation, washed twice with 2-ml portions of cold buffered saline, and dissolved in 4 ml of 0.1 M NaOH. The protein concentration of the dissolved precipitates was determined using the Folin-Ciocalteu reagent.⁷

With various modifications, usually in determination of protein or nitrogen, the quantitative tube precipitin test just described is the most commonly employed method for determining precipitating antibody titer.

Direct turbidimetric measurements were performed by rapidly mixing equal volumes (0.2 ml) of antigen and antibody and transferring the mixture to a cuvette with a 1-mm path length. The development of turbidity was read at 15-second intervals at 350 m μ on a Beckman DK-2A spectrophotometer. A mixture of buffered saline and antisera in equal volumes was used in the reference cuvette to correct for absorbance by serum components such as hemoglobin. The temperature of the cell compartment remained relatively constant at about 27 C throughout the experiments.

III. RESULTS

The precipitin curve shown in Figure 1 was obtained by measuring the protein concentration of dissolved BSA:anti-BSA precipitates using the quantitative tube method. The rate of turbidity development using the same dilutions of reactants was measured in the spectrophotometer with results shown in Figure 2. The highest rate of turbidity development was observed with an antigen-antibody ratio corresponding to the equivalence ratio as determined by the tube test.

After a relatively short lag period, the rate of turbidity development was essentially constant for about 1 minute after mixing. The rate in the linear region was measured and was expressed as the Δ OD/min. The rate observed with equivalent amounts of antigen and antibody was designated as the E Δ OD/min. The equivalence ratio, quantity of precipitated protein, and the rate of turbidity development were determined for a number of antigen and antibody dilutions using both the tube precipitin technique and the turbidimetric method, and the results were compared. When the E Δ OD/min was plotted as a logarithmic function against antiserum dilutions, a linear relationship was obtained with the highest concentrations of antibody (Figure 3). A plot of log E Δ OD/min against precipitated protein, as shown in Figure 4, revealed a linear relationship. In contrast, however, with the plot obtained with the antiserum dilutions, linearity in this case was lost with the higher dilutions of antibody.




FIGURE 1. Quantitative Precipitation Between BSA and Anti-BSA Serum Measured by the Tube Precipitin Technique. The figures by the points correspond to the curves in Figure 2.




FIGURE 2. Development of Turbidity with Various Antigen to Antibody Ratios. The figures by each curve correspond to the points in Figure 1.




FIGURE 3. The E Δ OD/Min Values Obtained Using BSA as the Antigen with Various Dilutions of Anti-BSA Serum.




FIGURE 4. Comparison of the Values Obtained with the Quantitative Tube Precipitin Test (Protein Precipitated) with Those Obtained with the Turbidimetric Technique (Log E Δ OD/Min) in the BSA/Anti-BSA System.

The correlations observed in the BSA system were also observed in a BGG:anti-BGG system. Similar results were also obtained in the immunologic precipitation between anti-AMP-BSA serum and AMP-BGG as the testing antigen. In this case, an hapten-specific antibody was being measured.

IV. DISCUSSION

Hawkins⁵ noted that, in the equivalence and antigen-excess zones, there is a considerable lag before the development of turbidity. He noted, as Eagle had earlier,³ a significant increase in the velocity of the reaction when more concentrated reactants were employed, presumably due to a decrease in the average inter-particle space and the subsequent increase in the number of total collisions. In the experiments reported here, a very short lag was observed because high concentrations of serum were employed. In most cases, the lag period under these conditions was short enough to be disregarded in calculations.

With the higher concentrations of antibody, the rate of turbidity in the equivalence zone appeared to follow almost a first-order kinetics reaction rather than the second- or higher-order reactions observed in zones of antigen or antibody excess. With more dilute serum samples, the rate of turbidity at equivalence began to deviate from the apparent first-order reaction observed with higher concentrations. As a result, there was a deviation from linearity when dilute antisera were employed in titer determinations. The opposite relationship was observed when the rates were plotted against protein precipitated for the same antiserum, that is, there was a deviation from linearity with the more concentrated serum samples.

The experiments reported here indicate that the measurement of the rate of turbidity development can be used effectively to determine the antibody titer of serum when high concentrations of reactants are used. The test can be performed on a large number of serum samples in a very short time because there is no necessity for collection and measurement of precipitates. Turbidimetric measurement as a means of titrating antiserum employs far less reagent than the classical tube tests. The values obtained with multiple samples of the same dilution of antiserum or antigen were usually within one standard deviation. This is an important factor to be considered when the quantity of antiserum available for testing is limited.

Excellent correlations were obtained with the three serologic systems described in the current study. The routine application of this procedure for precipitating systems should be seriously considered.

LITERATURE CITED

1. Boyden, A.; DeFalco, R. 1943. Report on the use of the photon-reflectometer in serological comparisons. *Physiol. Zool.* 16:229-241.
2. Eagle, H. 1932. Specific agglutination and precipitation: II. Velocity of the reactions. *J. Immunol.* 23:153-186.
3. Erlanger, B.F.; Beiser, S.M. 1964. Antibodies specific for ribonucleosides and ribonucleotides and their reaction with DNA. *Proc. Nat. Acad. Sci.* 52:68-74.
4. Gitlin, D.; Edelhoch, H. 1951. A study of the reaction between human serum albumin and its homologous equine antibody through the medium of light scattering. *J. Immunol.* 66:67-77.
5. Hawkins, J.D. 1964. Some studies on the precipitin reaction using a turbidimetric method. *Immunology* 7:229-238.
6. Junge, J.McB.; Junge, C.O., Jr.; Krebs, E.G. 1955. A study of antigen-antibody reaction rates by light transmission measurements. *Arch. Biochem. Biophys.* 55:338-355.
7. Lowry, O.H.; Rosebrough, N.J.; Farr, A.L.; Randall, R.J. 1951. Protein measurement with the Folin phenol reagent. *J. Biol. Chem.* 193:265-275.
8. Martin, D.S. 1942. Use of the photonreflectometer in precipitin titrations of a multiple-antigen multiple-antibody mixture by the Dean-Webb method. *J. Bacteriol.* 43:95-96.

Unclassified
Security Classification

17

DOCUMENT CONTROL DATA - R & D		
<small>(Security classification of title, body of abstract and indexing classification must be entered when the control report is classified)</small>		
1. ORIGINATING ACTIVITY (Corporate author)	2a. REPORT SECURITY CLASSIFICATION	
Department of the Army Fort Detrick, Frederick, Maryland, 21701	Unclassified	
2b. GROUP		
3. REPORT TITLE		
ESTIMATION OF PRECIPITATING ANTIBODY USING A TURBIDIMETRIC TECHNIQUE		
4. DESCRIPTIVE NOTES (Type of report and inclusive dates)		
5. AUTHOR(S) (Print name, middle initial, last name)		
William F. Vincent Earl W. Harris Sidney (NMI) Yaverbaum		
6. REPORT DATE	7a. TOTAL NO. OF PAGES	7b. NO. OF REFS
December 1968	17	8
8a. CONTRACT OR GRANT NO.	8b. ORIGINATOR'S REPORT NUMBER(S)	
9. PROJECT NO. 1B662706A071	Technical Manuscript 494	
10. DISTRIBUTION STATEMENT	10a. OTHER REPORT NUM (Any other numbers that may be assigned this report)	
Qualified requesters may obtain copies of this publication from DDC. Foreign announcement and dissemination of this publication by DDC is not authorized. Release or announcement to the public is not authorized.		
11. SUPPLEMENTARY NOTES	12. SPONSORING MILITARY ACTIVITY	
	Department of the Army Fort Detrick, Frederick, Maryland, 21701	
13. ABSTRACT		
The development of turbidity resulting from the precipitation between bovine serum albumin (BSA) and anti-BSA serum was studied in the spectrophotometer. The equivalence ratio and the titer of precipitating antibody could be determined by this method. Data obtained with the turbidimetric method correlated well with those obtained using the classical tube precipitin determination. Turbidimetric measurements were also applied successfully to the measurement of antibody titer in two other systems. Antibody titer can be determined with this method in about an hour as compared with the 3 or more days required by the tube test. In addition, this technique requires substantially less reagent than other methods. () ←		
14. Key Words		
Turbidimetric estimation Bovine serum albumin Antibody equivalence ratio Antibody titer determination		

DD FORM 1473

REPLACES DD FORM 1473, 1 JAN 64, WHICH IS OBSOLETE FOR Army USE.

Unclassified

Security Classification