

UNCLASSIFIED

AD NUMBER
AD869301
NEW LIMITATION CHANGE
TO Approved for public release, distribution unlimited
FROM Distribution authorized to U.S. Gov't. agencies and their contractors; Administrative/Operational Use; MAY 1970. Other requests shall be referred to Department of the Army, Fort Detrick, Attn: Technical Release Branch/TID, Frederick, MD 21701.
AUTHORITY
BDRL D/A ltr, 29 Sep 1971

THIS PAGE IS UNCLASSIFIED

AD

AD 869301

TECHNICAL MANUSCRIPT 604

DIRECT ASSESSMENT OF VIRAL AEROSOLS  
ON CELL CULTURES

William A. Hankins  
Henry J. Hearn

MAY 1970

STATEMENT #2 UNCLASSIFIED

This document is subject to special export controls and each transmittal to foreign governments or foreign nationals may be made only with prior approval of Dept. of Army, Fort Detrick, ATTN: Technical Release Branch/TID, Frederick, Maryland 21701

DEPARTMENT OF THE ARMY

Fort Detrick  
Frederick, Maryland

DDC  
RECEIVED  
MAY 22 1970  
RELEASED

C 10

49823

ACCESS TO THE	
WHITE SECTION	<input type="checkbox"/>
BLUE SECTION	<input checked="" type="checkbox"/>
DISSEMINATED	<input type="checkbox"/>
DISSEMINATION AVAILABILITY NOTES	
DATE	AVAIL. TO THE PUBLIC
2	

Reproduction of this publication in whole or in part is prohibited except with permission of the Commanding Officer, Fort Detrick, ATTN: Technical Releases Branch, Technical Information Division, Fort Detrick, Frederick, Maryland, 21701. However, DDC is authorized to reproduce the publication for United States Government purposes.

DDC AVAILABILITY NOTICES

Qualified requesters may obtain copies of this publication from DDC.

Foreign announcement and dissemination of this publication by DDC is not authorized.

Release or announcement to the public is not authorized.

DISPOSITION INSTRUCTIONS

Destroy this publication when it is no longer needed. Do not return it to the originator.

The findings in this publication are not to be construed as an official Department of the Army position, unless so designated by other authorized documents.

DEPARTMENT OF THE ARMY  
Fort Detrick  
Frederick, Maryland 21701

TECHNICAL MANUSCRIPT 604

DIRECT ASSESSMENT OF VIRAL AEROSOLS ON CELL CULTURES

William A. Hankins

Henry J. Hearn

Experimental Aerobiology Division  
AEROBIOLOGY & EVALUATION LABORATORIES

Project 1B562602AD03

May 1970

## DIRECT ASSESSMENT OF VIRAL AEROSOLS ON CELL CULTURES\*

ABSTRACT

A technique is described for the direct exposure of cell cultures to airborne virus, enabling quantitation of the virus in concentrations as low as one plaque-forming unit per liter of air.

Methods of sampling airborne microorganisms, particularly viruses, usually require their impingement in a suitable collection fluid, although techniques may include collection on agar, on a filter, or by electrostatic precipitation.<sup>1</sup> The collected viral agent must then be quantitated by conventional assay procedures, i.e., by inoculation into laboratory animals, embryonated eggs, or cell cultures. This report describes a technique in which known volumes of viral aerosols were drawn directly into flasks containing cell culture monolayers that were then used to quantitate the virus by allowing plaques to develop. This technique permits the detection of as low as one plaque-forming unit (PFU) of airborne virus per liter of air.

Venezuelan equine encephalomyelitis (VEE) virus, Trinidad strain grown in embryonated eggs was selected for study. Viral aerosols were generated in a cylindrical, 1,200-liter chamber similar to that described by Wolfe,<sup>2</sup> using either an FK-8 gun<sup>3</sup> or a Collison device.<sup>4</sup> Monolayers of primary chicken embryo (CE) cells or an established line of guinea pig lung cells were prepared in 250-ml cell culture flasks,\*\* using Eagle's basal medium supplemented with 10% fetal calf serum, 100 units of penicillin, and 100  $\mu$ g of streptomycin per ml. The CE cells were subsequently shown to be more uniformly susceptible to the virus and were chosen for most of the experimental work.


Cell cultures drained of growth media were exposed to the viral aerosols by attaching the mouth of the flask to the chamber. Before attaching the flask to the chamber, a hole was burned through the opposite end of the flask with a hot metal rod; a small sterile skirted stopper was inserted in this hole. A 15-gauge needle attached to a vacuum source with rubber tubing was used to penetrate the rubber stopper and to draw the aerosol into the flask. The rate of flow was controlled by a critical orifice inserted into the rubber tubing. When flasks were connected in series, this was accomplished by substituting metal caps containing 15-gauge needles and rubber gaskets for the plastic caps (Fig. 1).

---

\* This report should not be used as a literature citation in material to be published in the open literature. Readers interested in referencing the information contained herein should contact the senior author to ascertain when and where it may appear in citable form.

\*\* Falcon Plastics, Los Angeles, California.

**PRECEDING PAGE BLANK**



**FIGURE 1. Cell Culture Flasks Used for Direct Assessment  
of Viral Aerosols.**

After exposure to aerosols, cell monolayers were overlaid with Eagle's basal medium supplemented with 5% fetal calf serum, 1% agar, 100 units of penicillin, and 100  $\mu$ g of streptomycin per ml. The length of time after exposure before the overlay was applied did not appear to be critical, but should be completed within an hour. Infected cells were incubated for 24 hours at 37 C; a second overlay containing, in addition, 0.01% neutral red was applied; plaques were counted in another 24 hours.

Plaque counts were not diminished by an air wash of the flasks immediately after exposure to the aerosol, suggesting a rapid and irreversible attachment of the virus to the cells. Treatment of the exposed flask with specific antiserum reduced or eliminated plaque formation.

The titers shown in Table 1 are average values of at least two tests and represent values corrected for a uniform total volume sampled. When virus diluted 1/300 was disseminated, the concentration of aerosolized virus as determined by the direct flask (DF) technique was approximately  $2.6 \times 10^3$  PFU per liter. Dilution of the virus in tenfold increments prior to aerosolization resulted in corresponding decreases in recoveries when the sampling rate was 1 liter per min. It is of interest that when the sampling rate was increased to 10 liters per min, the recovery titers decreased somewhat, suggesting that with this faster sampling rate some virus might have escaped from the flask. Titer values obtained by the impinger technique (Table 1) were comparable to those obtained by the DF technique in the same experiment. Aerosolized virus originating from suspension concentrations of 1/30,000 or greater, however, were not detected by conventional impinger techniques.

The results of experiments designed to test for virus escape from flasks exposed to aerosols of different concentrations are shown in Table 2. In these tests, the aerosols were drawn through flasks connected in series rather than into only one flask. Aerosol concentrations were reduced by a series of air dilutions (purses) of the chamber, and, after each purge, the virus titer of the aerosol was determined. Data indicate that about 10% of the virus bypassed one flask and was deposited in the next flask at all concentrations. Also included in Table 2 are data showing that the assessments of virus per liter of aerosol by the DF technique were approximately the same as obtained by assaying impinger samples by the plaque technique.

In summary, results of these tests indicate that the direct exposure of cell culture monolayers to viral aerosols (DF technique) provides a simple and highly sensitive technique for assessing infectious airborne viral particles. The DF technique permitted the recovery of virus in very low aerosol concentrations that could not be detected by conventional impinger techniques.

TABLE 1. SAMPLING RATES AND RECOVERIES OF DILUTED  
AEROSOLIZED VEE VIRUS USING DIRECT FLASK  
AND IMPINGER TECHNIQUES

Virus Dilution <sup>a/</sup>	Liters per Min <sup>b/</sup>	Time, min	Total Sample Volume, <sup>c/</sup> liter	PFU per Liter	
				Direct Flask	Impinger <sup>d/</sup>
1/300	1	0.1	0.1	2600	2600
1/3000	1	1.0	1.0	160	310
	1	0.1	0.1	420	NT <sup>e/</sup>
	10	0.1	1.0	70	NT
1/30,000	1	10.0	10.0	11	NR <sup>f/</sup>
	10	1.0	10.0	2	NR
1/300,000	1	10.0	10.0	1.2	NR

- a. In spray device prior to aerosolization.
- b. Liters per minute limited by critical orifice used to draw aerosol into flask.
- c. Total volume of air that passed through flask.
- d. Virus in impinger fluid assayed by plaque test.
- e. Not tested.
- f. No virus recovered by conventional impinger technique.



TABLE 2. SAMPLING OF VEE VIRAL AEROSOLS USING FLASKS<sup>a/</sup>  
IN SERIES AND IMPINGER TECHNIQUES

Sampling <sup>b/</sup>	Flask No.					Concn per liter, <sup>f/</sup> plaques	Concn per Liter, <sup>g/</sup> plaques
	1	2	3	4	5		
Initial	c <sup>d/</sup>	C	>750 <sup>e/</sup>	162	21	1.83 x 10 <sup>6</sup>	3.9 x 10 <sup>6</sup>
After Purge 1 <sup>c/</sup>	C	>800 <sup>e/</sup>	140	22	-	1.62 x 10 <sup>5</sup>	1.95 x 10 <sup>5</sup>
After Purge 2	>800 <sup>e/</sup>	95	3	-	-	9.80 x 10 <sup>3</sup>	2.38 x 10 <sup>4</sup>
After Purge 3	123	16	-	-	-	1.39 x 10 <sup>3</sup>	4.80 x 10 <sup>3</sup>

- a. As used in direct flask technique (see text and Fig. 1).  
b. 100 ml of aerosol sampled in 0.1 minute.  
c. Virus concentration reduced (purged) by air dilution of chamber.  
d. Clear; total cell destruction.  
e. TNTC, only estimate of count was possible.  
f. Assay values based upon countable plaques.  
g. Assay values based upon plaque counts obtained from impinger fluids.

LITERATURE CITED

1. Wolf, H.W.; Skaliy, P.; Hall, L.B.; Harris, M.M.; Decker, H.M.; Buchanan, L.M.; Dahlgren, C.M. 1959. Sampling microbiological aerosols, (Public Health Monograph 60). U.S. Government Printing Office, Washington, D.C. 53 p.
2. Wolfe, E.K. 1961. Quantitative characterization of aerosols. *Bacteriol. Rev.* 25:194-202.
3. Jemski, J.V.; Phillips, G.B. 1965. Aerosol challenge of animals, p. 286. In W.I. Gay (ed.) *Methods of animal experimentation*, Vol. I. Academic Press, Inc., New York.
4. Collison, W.E. 1935. *Inhalation therapy technique*, p. 77. Heineman, London.

DOCUMENT CONTROL DATA - R & D		
<i>(Security classification of title, body of abstract and indexing annotation must be entered when the overall report is classified)</i>		
1. ORIGINATING ACTIVITY (Corporate author) Department of the Army Fort Detrick, Frederick, Maryland, 21701		2a. REPORT SECURITY CLASSIFICATION Unclassified
		2b. GROUP
3. REPORT TITLE DIRECT ASSESSMENT OF VIRAL AEROSOLS ON CELL CULTURES		
4. DESCRIPTIVE NOTES (Type of report and inclusion dates)		
5. AUTHOR(S) (First name, middle initial, last name) William A. Hanks Henry J. Hearn		
6. REPORT DATE May 1970	7a. TOTAL NO. OF PAGES 11	7b. NO. OF REFS 5
8a. CONTRACT OR GRANT NO.	8b. ORIGINATOR'S REPORT NUMBER(S) Technical Manuscript 604	
a. PROJECT NO 1B562602AD03	8c. OTHER REPORT NO(S): (Any other numbers that may be assigned to the report)	
c. Task-Work Unit 01		
e. N/A	CMS 6678 AMXFD-AE-T 49823	
10. DISTRIBUTION STATEMENT Qualified requesters may obtain copies of this publication from DDC. Foreign announcement and dissemination of this publication by DDC is not authorized. <u>Release or announcement to the public is not authorized.</u>		
11. SUPPLEMENTARY NOTES Experimental Aerobiology Division	12. SPONSORING MILITARY ACTIVITY Department of the Army Fort Detrick, Frederick, Md., 21701	
13. ABSTRACT A technique is described for the direct exposure of cell cultures to airborne virus, enabling quantitation of the virus in concentrations as low as one plaque-forming unit per liter of air.		
14. Key Words Viral aerosols Plaque assay Cell culture		