

AD-A058 594

ARMY ARMAMENT RESEARCH AND DEVELOPMENT COMMAND DOVER--ETC F/G 13/2
CORRELATION OF OXYGEN DEMAND AND TOTAL ORGANIC CARBON TESTS ON --ETC(U)
JUN 78 M ROTH, J M MURPHY


UNCLASSIFIED

ARLCO-TR-77031-PT-1


SBIE-AD-E400 169

NL

| OF |
AD
A058 594



END
DATE
FILMED
11-78
DDC



MICROCOPY RESOLUTION TEST CHART
NATIONAL BUREAU OF STANDARDS-1963-A

AD A058594

12 LEVEL

AD

AD-E400 169

TECHNICAL REPORT ARLCD-TR-77031

CORRELATION OF OXYGEN DEMAND AND TOTAL ORGANIC CARBON TESTS ON WASTEWATERS FROM AMMUNITION PLANTS

PART I. COD AND TOC OF PURE EXPLOSIVES

MILTON ROTH
JOSEPH M. MURPHY, JR.

DDC
RECEIVED
SEP 15 1978

JUNE 1978

[Handwritten signature]

B

AD No. _____
DDC FILE COPY



US ARMY ARMAMENT RESEARCH AND DEVELOPMENT COMMAND
LARGE CALIBER
WEAPON SYSTEMS LABORATORY
DOVER, NEW JERSEY

APPROVED FOR PUBLIC RELEASE; DISTRIBUTION UNLIMITED.

78 07 26 017

The view, opinions, and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy or decision, unless so designated by other documentation.

Destroy this report when no longer needed. Do not return it to the originator.

Unclassified

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

REPORT DOCUMENTATION PAGE		READ INSTRUCTIONS BEFORE COMPLETING FORM
1. REPORT NUMBER Technical Report	2. GOVT ACCESSION NO. ARLCD-TR-77031 - PT-1	3. RECIPIENT'S CATALOG NUMBER
4. TITLE (and Subtitle) Correlation of Oxygen Demand and Total Organic Carbon Tests on Wastewaters from Ammunition Plants, Part I. COD and TOC of Pure Explosives,		5. TYPE OF REPORT & PERIOD COVERED Final report
6. AUTHOR(s) Milton Roth - ARRADCOM, Project Leader Joseph M. Murphy, Jr - ARRADCOM Project Engineer		7. PERFORMING ORG. REPORT NUMBER Environmental Qual R&D Funds
9. PERFORMING ORGANIZATION NAME AND ADDRESS US Army Armament Research & Development Command Dover, NJ 07801		8. CONTRACT OR GRANT NUMBER(s)
11. CONTROLLING OFFICE NAME AND ADDRESS US ARRADCOM ATTN: DRDAR-TSS Dover, NJ 07801		10. PROGRAM ELEMENT, PROJECT, TASK AREA & WORK UNIT NUMBERS
14. MONITORING AGENCY NAME & ADDRESS (if different from Controlling Office) STB-MTD-LCWSL US Army Armament Research & Development Command Dover, NJ 07801		12. REPORT DATE JUNE 1978
16. DISTRIBUTION STATEMENT (of this Report) Approved for public release; distribution unlimited.		13. NUMBER OF PAGES 46
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report) SBIE		15. SECURITY CLASS. (of this report) Unclassified
18. SUPPLEMENTARY NOTES AD-E400-169		15a. DECLASSIFICATION/DOWNGRADING SCHEDULE
19. KEY WORDS (Continue on reverse side if necessary and identify by block number) Correlation Chemical Oxygen Demand (COD) Explosive Manufacturing Sensitivity Total Organic Carbon (TOC) Precision Statistical Analysis Monitoring Total Organic Nitrogen		
20. ABSTRACT (Continue on reverse side if necessary and identify by block number) See reverse side		

DDC
RECEIVED
SEP 15 1978
RECEIVED
B

78 07 26 014/B
410-163

Unclassified

SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)

ABSTRACT

New Federal regulations for Effluent Limitations, Guidelines, and Standards for Explosives Manufacturing, Point Source Category, has led to the evaluation of the specified oxygen demand tests in order to determine their applicability to munition unique wastewaters. In part one of this study, results of chemical oxygen demand (COD) and total organic carbon (TOC) tests performed on solutions of pure energetic materials are reported and discussed.

The experimental data was analyzed statistically to determine the correlation between the COD and TOC tests when measuring known solutions of energetic materials; the relationship between actual and theoretical values was also determined.

Experimental results show that response to the COD test varies significantly between compounds and is not an accurate indicator of organic loading of munition wastewaters despite its high level of precision. TOC is found to be more accurate and more sensitive.

Unclassified

SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)

ACKNOWLEDGMENTS

Mr. Nathaniel S. Gelber, Analytical Chemist, U. S. Armament Research and Development Command, Dover, New Jersey provided invaluable assistance by preparing all the standard solutions and performing all the chemical oxygen demand tests. His comments on the details of the testing and his evaluation of the results were most helpful.

This study was funded under the Environmental Quality R&D program administered by the DARCOM Lead Laboratory, Aberdeen Proving Ground, MD.

ACCESSION for	
NTIS	White Section <input checked="" type="checkbox"/>
DDC	Buff Section <input type="checkbox"/>
UNANNOUNCED	<input type="checkbox"/>
JUSTIFICATION _____	
BY _____	
DISTRIBUTION/AVAILABILITY CODES	
Dist. AVAIL. and/or SPECIAL	
A	

TABLE OF CONTENTS

	Page No.
Introduction	1
Experimental	2
Experimental Design	2
Procedure	2
Discussion of Results	3
Comparison of COD and TOC for Nitrotoluenes and Toluene at High Concentrations	3
Comparison of COD and TOC for Nitrotoluenes and Toluene at Low Concentrations	3
Comparison of COD and TOC for Nitrobenzenes and Benzene at High Concentration	4
Comparison of COD and TOC for Nitrobenzenes and Benzene at Low Concentration	4
Comparison of COD and TOC Tests for Nitramines, Guanidine Nitrate, and Urea at High Concentration	4
Comparison of COD and TOC Tests for Nitramines, Guanidine Nitrate, and Urea at Low Concentration	5
Comparison of COD and TOC for Nitrate Esters, Alcohols, and Acids at High Concentration	5
Comparison of COD and TOC for Nitrate Esters, Alcohols, and Acids at Low Concentration	5
Experimental Data	5
Conclusions	6
Recommendations	7
References	7
Distribution List	35

Tables

1	Point sources from manufacturing and loading of explosives	9
2	EPA effluent limitations, guidelines and standards	10
3	COD and TOC test results for toluene and nitrotoluenes	11
4	COD and TOC test results for benzene and nitrobenzenes	16
5	COD and TOC test results for urea, nitramines, and guanidine nitrate	18
6	COD and TOC test results for nitrate esters	21
7	COD and TOC test results for alcohols and acids	22
8	Statistical parameters for aromatic compounds	24
9	<i>Statistical parameters for nonaromatic compounds</i>	25
10	Comparison of completeness of recovery by COD and TOC tests	26

Figures

1	Comparison of COD and TOC for nitrotoluenes and toluene at high concentration	27
2	Comparison of COD and TOC for nitrotoluenes and toluene at low concentration	28
3	Comparison of COD and TOC for nitrobenzenes and benzene at high concentration	29
4	Comparison of COD and TOC for nitrobenzenes and benzene at low concentration	30
5	Comparison of COD and TOC tests for nitramines, guanidine nitrate, and urea at high concentration	31
6	Comparison of COD and TOC tests for nitramines, guanidine nitrate, and urea at low concentration	32

7	Comparison of COD and TOC for nitrate esters, alcohols and acids at high concentration	33
8	Comparison of COD and TOC for nitrate esters, alcohols and acids at low concentration	34

INTRODUCTION

Effluent limitations and guidelines for existing sources, to be achieved by the application of the best practicable control technology currently available (BPCTCA), and the best available technology economically achievable (BATEA), were promulgated by the Environmental Protection Agency (EPA) in the Federal Register of 9 March 1976 for Explosive Manufacturing Point Source category. These regulations have led to an evaluation of the test methods for determining organic loading and the resulting oxygen demand of military-unique wastewaters. The interim final rule for effluent guidelines, for explosive manufacturing point source guidelines sets standards in terms of the following parameters: biochemical oxygen demand (BOD_5), chemical oxygen demand (COD), and total suspended solids (TSS).

The energetic materials found in wastewaters of different types of explosive manufacturing are listed in table 1. In the first part of this study, known concentrations were tested for COD and total organic carbon (TOC). The values obtained were statistically analyzed to determine which method is more appropriate for monitoring waste streams containing energetic materials and also to determine if any relationship exists between the methods.

The discharge standards set for explosive manufacturing point source category are listed in table 2 (ref. 1). These standards, in terms of kilograms (lb) BOD or kilograms (lb) COD allowed per 1000 kg (2200 lb) of product produced are established for BPCTCA and for BATEA. When the proposed discharge standards are projected over maximum production rates for nitroglycerin and trinitrotoluene and the water usage associated with their production, (ref. 2) the resulting discharge concentrations, in milligrams per liter (mg/l), of COD in the waste streams are extremely low (table 2). These standards present a problem of how to measure these parameters at such low levels. The COD analysis has limited accuracy below 10 mg/l of COD (ref. 3). In order to evaluate the EPA environmental water quality standards for munition wastewaters, the applicability of the conventional oxygen demand parameters to low concentration levels of COD of military-unique compounds was studied. In this study, consideration was given to other tests capable of detecting these munition-unique pollutants. Thus, the TOC test is compared to the COD for the same compounds at the same concentration.

An evaluation of this type will not only show the precision and accuracy of each method, but also will indicate which is the most appropriate for monitoring explosive manufacturing point sources.

EXPERIMENTAL

Experimental Design

Since one objective of this study was to compare different test methods to evaluate their suitability for estimating compliance with the EPA standards, a factorial design was used with the intention of applying a variety of statistical analyses. In this design, the following factors were considered:

1. Method (COD and TOC)
2. Concentrations (high and low levels)

Four replications were conducted on each combination of factors. The replications were randomized to eliminate the effect of time on the results. Analyst effect was eliminated by having one analyst perform all the COD tests and a second analyst performed all the TOC tests. The test solutions were prepared as described in the procedure paragraph below and aliquots were taken from the same solution for all tests. In general, the COD tests were completed before beginning the TOC tests.

Prior to conducting the analysis of variance of the data, the variances were checked for homogeneity by the Bartlett test (ref. 5). The result showed that the variances were not homogeneous; therefore, the analysis of variance, which is based on the assumption of homogeneity, was not performed. The statistical tests that were performed are discussed later in this report.


Procedure

Standard solutions that are typically found in the wastewaters of army ammunition plants were prepared. Isomers of the compounds were also included in the experiment in order to examine the effect of the structure of a compound on the test response. Each standard solution was prepared by dissolving weighed amounts (to the nearest 0.1 mg) of the compounds in distilled water. The mixture was stirred with a magnetic stirrer until solution was complete. Several compounds required more than two days of constant stirring to effect solution.

The solutions were cooled to ambient temperature before diluting to the mark with water. Aliquots were transferred from the high concentration level solutions and diluted appropriately to prepare the low concentration level.

The chemical oxygen demand analyses were performed according to the standard method (ref. 6). The total organic carbon analyses were also performed according to the standard method (ref. 7) using a Beckman 915-A Total Organic Carbon Analyzer, a Hamilton spring-loaded syringe, and a sample size of 20 micro-liters.

Theoretical COD values were used as a basis for calculating the percentage of recovery for the different compounds. The theoretical COD values were calculated on the assumption of complete oxidation of all elements in the energetic materials as follows:



Theoretical TOC values were used as the basis for calculating the percent of recovery. The percentage values were calculated on the assumption of complete combustion of the carbon in the compound to carbon dioxide.

DISCUSSION OF RESULTS

Comparison of COD and TOC for Nitrotoluenes and Toluene at High Concentrations

As shown in figure 1, at the high concentration level the TOC concentration shows a recovery of approximately one hundred percent. Toluene yielded low results, probably because of evaporation during the sample preparation (ref. 8).

The COD test at the high concentration level displayed results that varied with each compound. Low results were displayed by m-nitrotoluene; 2, 3 dinitrotoluene; 3, 4 dinitrotoluene; and toluene. Note that the nitrotoluenes yielding low results all had a nitro group in the meta position. All the other results were $\pm 5\%$ from 100% recovery.

Comparison of COD and TOC for Nitrotoluenes and Toluene at Low Concentrations

At low concentrations (fig. 2), the mononitrotoluenes all yielded recoveries that range from 95% - 105% in both TOC and COD. These results indicate that the mononitrotoluenes can be analyzed with good recovery by either test method. However, the more highly nitrated toluenes were significantly lower by the COD test than by the TOC.

An analysis of a solution containing 6.41 ppm TNT (equivalent to 2.22 mg/L TOC) is likely to produce erratic results since the detection limit, under ordinary procedures, is limited to 1 mg/L (ppm) of TOC (ref. 9).

Comparison of COD and TOC for Nitrobenzenes and Benzene at High Concentration

At the high level (fig. 3), TOC and COD deviate significantly from full recovery for most of the compounds tested. The COD results of this group are consistently low, with the dinitrobenzenes registering very minimal COD values. Nitrobenzene and benzene also yielded low results. This could be the result of the refractoriness of aromatics which typically are not oxidized appreciably in COD analysis.

Comparison of COD and TOC for Nitrobenzenes and Benzene at Low Concentration

At the low level concentration (fig. 4), good recoveries for TOC are exhibited by the nitrobenzenes. Benzene again yielded a low result, probably due to losses from volatilization of the compound.

The COD results follow the same pattern as at the high level, except that the nitrobenzene recovery is now complete. This indicates that either the concentration of dichromate or the digestion time at the high level is inadequate.

Comparison of COD and TOC Tests for Nitramines, Guanidine Nitrate, and Urea at High Concentration

All the compounds tested by the TOC method yielded low results except for tetryl and guanidine nitrate (fig. 5). The COD test also displayed consistently low results for this group of compounds. Nitroguanidine, guanidine nitrate, and urea did not respond at all to the COD. Although it has been recognized that certain aromatics give low results in the COD test, this appears to be the first time it has been found that the urea structure resists dichromate oxidation.

Tetryl and RDX displayed low COD results which indicates that nitramines also respond incompletely to the COD test. Tetryl is higher than the rest, probably because it contains three nitro groups in addition to a nitramine.

Comparison of COD and TOC Tests for Nitramines, Guanidine Nitrate, and Urea at Low Concentration

TOC recoveries were satisfactory for HMX and RDX, but high results were obtained on nitroguanidine and guanidine nitrate, and low results were obtained on tetryl and urea (fig. 6). The COD results followed the same pattern found at the high concentration level, indicating that the effect of structure is independent of concentration for this group of compounds.

Comparison of COD and TOC for Nitrate Esters, Alcohols, and Acids at High Concentration

Both TOC and COD (fig. 7) appear to be very appropriate for monitoring these compounds. Glycerin, pentaerythritol, and acetic acid are more typical of the type of compounds found in industrial rather than explosives wastewaters and are at least part of the basis for assuming that COD is a representative measure of organic loading in a stream. However, the data from explosives and urea type compounds indicates that the COD test has still-to-be-defined limitations.

Comparison of COD and TOC for Nitrate Esters, Alcohols, and Acids at Low Concentration

TOC results (fig. 8) are erratic, probably because the measurements were made at the lower level of sensitivity of the TOC analyzer. COD results, with the exception of acetic acid, yielded good recoveries with these more conventional materials. Acetic acid yielded a low recovery, probably because of losses due to volatilization during the dichromate oxidation.

Experimental Data

The experimental data is reported in tables 3 through 7. The tables include the standard deviations and the percent recoveries of COD and TOC test results. In order to complete the study it was necessary to determine the correlation between the test methods to see if the methods could be interchanged. Correlation coefficients were calculated using the recovery values obtained from the COD and TOC test methods of the combined high and low levels of concentration for individual compounds. Significant correlations at the 95% confidence level were obtained by several compounds in each table, but no overall correlation was found. Results from the statistical analyses are in tables 8 and 9.

Generally, in both the aromatic and nonaromatic groups, variances were always higher on results from total organic carbon analyses as opposed to chemical oxygen demand analyses. The larger variances might be an indication of the greater sensitivity of the TOC test method and the need for replication in order to obtain a reliable estimate of the true value.

As a final test of the overall completeness of recovery of the COD and TOC test methods, t-tests were performed to compare the data from each test with the theoretical recovery value of one hundred percent. The statistical evaluation (see table 10) showed that there was a significant difference from 100% recovery at the 90% level of confidence with 23 degrees of freedom for the COD analyses at both high and low concentration levels. With the TOC analyses, however, there is no significant difference from 100% recovery at either concentration level. This indicates that the TOC test is more accurate and appears to be a more appropriate measure of the organic loading of a wastewater.

CONCLUSIONS

1. The results obtained by the COD test do not give complete recovery with many compounds. Although this fact is recognized for aromatic compounds and is confirmed in this study, the study also revealed that compounds with urea structure, nitramine groups and certain nitro groups yield low results. However, the precision of the COD method surpasses that of the TOC.
2. The results obtained by the TOC test do give complete recovery as indicated by the t-test. The principle of the TOC test makes it independent of the structure of a compound. However, the precision of the TOC test, particularly at low concentrations, would require numerous replications to obtain a reliable estimate of the true mean of a sample.
3. From the point of view of enforcement of standards, the COD test appears inappropriate for the explosives industry. The TOC test would be more meaningful from the point of view of accuracy, but further study of this test is needed to improve its precision.
4. Although it has been shown that TOC and COD tests are correlated for many industrial wastes (ref. 10), their correlation is not a general rule because of the widely different principles upon which the tests are based. The data presented in this study indicates that correlations occur by chance rather than by a unifying principle. Thus, the existing standard based on the COD test for the explosive industry, cannot be converted to an equivalent standard based on the TOC test.

RECOMMENDATIONS

1. Further study should be made of the COD test to determine the effect of molecular structure on oxidizability. The efficiency of other oxidants, as compared to dichromate, should also be studied.
2. Consideration should be given to replacing the COD test with TOC in the EPA Limitations, Guidelines and Standards for the Explosives Industry.
3. Further study of the general indicators of organic loading in wastewaters and receiving waters is needed. For the explosives industry, a combination of TOC and total organic nitrogen should be considered to obtain a more pertinent measure of water quality.
4. Actual samples of wastewaters from explosive plants should be tested to confirm results obtained from the theoretical studies.

REFERENCES

1. Federal Register, Vol 41, No. 47 - Tuesday, March 9, 1976, pp. 10180 - 10188
2. D. H. Maybury, J. L. Evans, and A. Carotti, "Propellant Plant Pollution Abatement Improvement of Water Utilization at Radford Army Ammunition Plant," Picatinny Arsenal Report 4562
3. "Standard Methods for the Examination of Water and Wastewater," 14th Edition, American Public Health Association, Washington, DC, 1976, p. 551
4. J. P. Glennon, CPT, MSC, L. Reuter, LTC, MSC, "Environmental Quality Standards for Munitions - Unique Pollutants," U. S. Army Bioengineering Research and Development Laboratory, Fort Detrick, Frederick, Maryland. Presented at Seventeenth DOD Explosives Safety Seminar, Denver, Colorado, September 14, 1976
5. C. A. Bennett and N. L. Franklin, "Statistical Analysis in Chemistry and the Chemical Industry," Wiley and Sons, Inc., NY, 1954, p. 197
6. "Standard Methods for the Examination of Water and Wastewater," 14th Edition, American Public Health Association, Washington, DC, 1976, pp. 550 - 555

7. Ibid, pp. 532-534
8. C. E. Van Hall, D. Barth, and V. A. Stenger, *Analytical Chemistry*, Vol 37, No. 6, May 1965, pp. 769-773
9. "Standard Methods for the Examination of Water and Wastewater," 14th Edition, American Public Health Association, Washington, DC, 1976, p. 533
10. Environmental Protection Agency, "Handbook for Monitoring Industrial Wastewater," August, 1973, pp. 5-10

Table 1. Point sources from manufacturing and loading of explosives

TNT Production

Condensate Water: 2,3-dinitrotoluene (DNT), 2,4-DNT, 2,5-DNT, 2,6-DNT,
3,4-DNT, 3,5-DNT, mononitrotoluene, 1,3-dinitrobenzene

TNT Load, Assembly and Pack

Pink Water: 2,4,6-trinitrotoluene (TNT), 2,6-DNT, 4,6-DNT

RDX/HMX

Wastewaters: Cyclotrimethylene trinitramine (RDX), cyclotetramethylene
tetranitramine (HMX)

Nitrate Ester Wastewaters

Trinitroglycerol (NG), Triethyleneglycoldinitrate (TEGDN)

Table 2. EPA effluent limitations, guidelines and standards*

EFFLUENT CHARACTERISTIC	LIMITATIONS, kg/kg OR LBS/1000 LBS PRODUCT			
	FOR ANY 1 DAY, MAX	AV OF 30 CONSEC DAYS, MAX	BATEA	BATEA
	BPCTCA	BATEA	BPCTCA	BATEA
COD	-	0.85 ^a	-	0.55 ^b
EXISTING SOURCES	7.77		2.59	
NEW SOURCES	3.6		2.3	
BOD5	-	0.11	-	0.067
EXISTING SOURCES	0.72		0.24	
NEW SOURCES	0.35		0.23	
COD @ MAX PROD. RATES:				
		NG, mg/l	$\frac{a}{26}$	$\frac{b}{17}$
		TNT, mg/l	5.4	3.5

* 9 MAR 1976

Table 3. COD and TOC test results for toluene and nitrotoluenes

Compound	Concentration mg/l		COD, mg/l		Concentration mg/l	TOC, mg/l		Recovery, %		
	Added	Found	Added	Found		Added	Found	COD	TOC	
Toluene	169.5	408.0	529.9	169.5	169.5	54.0	154.5	76.9	31.8	
		385.0		33.8		72.6		19.9		
		391.0		17.2		73.7		10.1		
		362.0		33.8		68.3		20.0		
Toluene	17.0	387.0	53.0	\bar{x}	17.0	34.7	\bar{x}	73.0	20.4	
		19.0		s		15.06		s	3.60	8.88
		31.1		53.0		7.9		15.5	58.6	51.0
		33.0		30.6		7.0		7.9	62.3	45.2
o-NT	204.2	28.7	428.8	200.6	200.6	8.0	123.0	54.2	51.7	
		29.4		1.9		7.9		55.5	51.0	
		30.6		s		7.7		57.7	49.8	
		1.9		s		0.47		3.6	3.03	
o-NT	20.4	418.4	42.9	120.0	20.1	120.0	\bar{x}	97.5	97.5	
		418.4		s		122.0		97.5	99.2	
		418.4		42.9		123.6		97.5	100.0	
		419.4		0.5		118.0		97.8	95.9	
o-NT	20.4	418.7	42.9	120.9	20.1	120.9	\bar{x}	97.6	98.2	
		0.5		s		2.43		0.15	1.83	
		42.1		42.9		11.5		12.3	98.1	93.4
		39.7		42.1		13.5		12.3	92.5	109.7
o-NT	20.4	40.1	42.9	13.0	20.1	13.0	\bar{x}	93.4	105.6	
		41.7		s		14.8		97.2	120.3	
		40.9		42.9		13.2		12.3	95.3	107.3
		1.2		s		1.36		2.76	11.1	

Table 3. (Continued)

Compound	Concentration mg/l		COD, mg/l		Concentration mg/l	TOC, mg/l		Recovery, %	
	Added	Found	Added	Found		Added	Found	COD	TOC
m-NT	192.9	336.9	405.1	193.8	118.8	134.0	83.2	112.7	
		342.9				131.6	84.6	110.7	
		347.0				136.0	85.7	114.4	
		341.0				136.0	84.2	114.5	
m-NT		342.0	\bar{x}		\bar{x}	134.4	84.4	113.1	
		4.2	s		s	2.09	1.03	1.79	
	19.3	37.4	40.5	19.4	11.9	11.0	92.3	92.5	
		38.5				11.5	94.9	96.8	
		39.6				11.0	97.7	92.5	
p-NT		38.4				10.8	94.7	90.9	
		38.5	\bar{x}		\bar{x}	11.1	95.0	93.2	
		0.9	s		s	3.0	2.21	2.53	
	31.6	62.7	66.4	31.6	19.3	19.0	94.4	98.7	
		63.1				19.0	95.0	98.7	
p-NT		63.3				19.5	95.3	101.3	
		64.7				19.5	97.4	101.3	
		63.5	\bar{x}		\bar{x}	19.3	95.6	100.0	
		0.9	s		s	0.28	1.30	1.50	
	6.10	12.6	12.8	6.34	3.89	4.1	98.4	105.3	
p-NT		10.5				3.6	82.0	92.5	
		12.1				3.7	94.5	95.1	
		11.4				3.0	89.0	77.1	
		11.7	\bar{x}		\bar{x}	3.6	91.0	92.5	
		0.91	s		s	11.7	7.11	11.7	

Table 3. (Continued)

Compound	Concentration mg/l		COD, mg/l		Concentration mg/l	TOC, mg/l		Recovery, %	
	Added	Found	Added	Found		Added	Found	COD	TOC
2,3 DWT	79.8	62.8	50.5	23.4	24.5	78.7	103.8		
		63.2			24.8	79.2	105.7		
		62.7			24.8	78.6	105.7		
		57.8			25.00	72.4	106.8		
2,3 DWT	\bar{x}	61.6		\bar{x}	24.8	77.2	105.8		
	s	2.6		s	0.20	3.23	1.25		
	10.1	6.7	6.38	2.95	2.0	66.3	67.8		
		5.4			1.7	53.5	57.6		
2,4 DWT		6.2			1.9	61.4	64.4		
		7.0			1.7	69.3	57.6		
	\bar{x}	6.3		\bar{x}	1.83	62.6	62.0		
	s	0.7		s	0.15	6.90	5.1		
2,4 DWT	157.9	153.0	99.9	46.3	49.0	96.8	105.8		
		154.0			49.0	97.5	105.8		
		156.0			48.8	98.7	105.3		
		157.0			48.0	99.4	103.6		
2,4 DWT	\bar{x}	155.0		\bar{x}	48.7	98.1	105.2		
	s	1.83		s	0.48	1.17	1.04		
	10.6	10.2	6.7	3.10	2.8	96.2	90.3		
		10.1			3.0	95.3	96.8		
2,4 DWT		8.9			2.7	83.9	87.0		
		8.7			2.7	82.0	87.0		
	\bar{x}	9.5		\bar{x}	2.8	89.4	90.3		
	s	0.8		s	0.14	7.44	4.62		

Table 3. (Continued)

Compound	Concentration mg/l	COD, mg/l		Concentration mg/l	TOC, mg/l		Recovery, %		
		Added	Found		Added	Found	COD	TOC	
2,5 DNT	50.5	79.8	75.5	50.5	23.4	25.0	94.6	106.8	
			77.6				24.5	97.2	104.7
			75.2				24.5	94.2	104.7
			76.4				24.5	95.7	104.7
2,5 DNT	6.45	\bar{x}	76.2	6.45	\bar{x}	24.6	95.5	105.3	
		s	1.1		s	0.25	1.34	1.05	
		10.2	8.8		2.98	2.5	86.2	83.8	
			9.7			3.0	95.1	100.7	
2,6 DNT	50.8	\bar{x}	9.1	50.8	\bar{x}	3.13	89.0	105.0	
		s	0.5		s	0.63	5.05	21.1	
		80.3	77.8		23.5	24.7	96.9	105.1	
			79.2			23.8	98.6	101.2	
2,6 DNT	6.48	\bar{x}	78.6	6.48	\bar{x}	24.8	97.9	105.4	
		s	0.8		s	0.82	0.98	3.52	
		10.2	6.1		3.0	3.0	59.8	100.0	
			8.2			3.7	80.4	123.3	
2,6 DNT	6.48	\bar{x}	9.9	6.48	\bar{x}	3.2	97.0	106.7	
		s	9.9		s	3.0	97.0	100.0	
		8.5	8.5		3.23	3.0	83.0	107.5	
		1.8	1.8		0.33	0.33	17.7	11.0	

Table 3. (Continued)

Compound	Concentration mg/l		COD, mg/l		Concentration mg/l	TOC, mg/l		Recovery, %	
	Added	Found	Added	Found		Added	Found	COD	TOC
3,4 DWT	53.8	56.6	85.1	26.8	53.9	25.0	66.5	107.4	
		55.9		26.0			65.6	104.2	
3,4 DWT	6.23	54.9	9.8	25.5	6.23	2.88	64.5	102.2	
		53.2		26.0			62.5	104.2	
3,4 DWT	6.23	55.2	9.8	26.1	6.23	2.88	64.8	104.5	
		1.5		0.54			1.72	2.15	
3,4 DWT	6.23	5.9	9.8	2.1	6.23	2.88	60.2	72.9	
		4.3		2.0			43.8	69.4	
3,4 DWT	6.23	5.1	9.8	2.0	6.23	2.88	52.0	69.4	
		5.9		3.5			60.2	121.5	
2,4,6 TWT	100.4	5.3	127.3	2.4	100.4	37.2	54.1	83.3	
		0.77		0.74			7.85	25.5	
2,4,6 TWT	100.4	124.1	127.3	37.0	100.4	37.2	97.5	99.5	
		123.1		34.5			96.7	92.7	
2,4,6 TWT	6.41	124.1	8.13	35.0	6.41	2.22	97.5	94.1	
		120.8		37.0			94.8	99.5	
2,4,6 TWT	6.41	123.0	8.13	35.9	6.41	2.22	96.6	96.5	
		1.6		1.31			1.27	3.57	
2,4,6 TWT	6.41	6.6	8.13	3.1	6.41	2.22	81.2	139.6	
		6.3		4.5			77.5	202.0	
2,4,6 TWT	6.41	7.8	8.13	2.0	6.41	2.22	95.9	90.1	
		7.5		4.0			92.2	180.1	
2,4,6 TWT	6.41	7.1	8.13	3.4	6.41	2.22	86.7	153.0	
		0.71		1.10			8.75	49.2	

Table 4. COD and TOC test results for benzene and nitrobenzenes

Compound	Concentration mg/l		COD, mg/l		Concentration mg/l	TOC, mg/l		Recovery, %	
	Added	Found	Added	Found		Added	Found	COD	TOC
Benzene	147.1	307.0	452.0	148.0	136.2	72.5	67.9	53.2	53.2
		288.0		68.0		63.7			
Benzene	14.7	267.0	45.2	14.8	13.6	84.0	59.0	61.7	61.7
		270.0		75.0		59.7			
Benzene	14.7	283.0	45.2	14.8	13.6	74.9	62.6	55.0	55.0
		18.5		6.74		4.11			
Benzene	14.7	17.6	45.2	14.8	13.6	4.5	38.9	33.0	33.0
		17.0		4.5		37.6			
Benzene	14.7	16.3	45.2	14.8	13.6	5.0	36.1	36.8	36.8
		10.6		2.1		23.5			
Nitrobenzenes	150.2	15.4	214.5	150.2	64.4	4.03	34.0	29.6	29.6
		3.2		1.30		7.11			
m-DN Benzene	150.2	18.0	214.5	150.2	64.4	72.0	8.39	111.8	111.8
		16.1		58.0		7.50			
m-DN Benzene	150.2	12.2	214.5	150.2	64.4	58.0	5.69	90.1	90.1
		13.3		58.0		6.20			
m-DN Benzene	150.2	14.9	214.5	150.2	64.4	61.5	6.95	95.5	95.5
		2.64		7.0		1.23			
m-DN Benzene	150.2	1.9	214.5	150.2	64.4	7.0	8.86	108.7	108.7
		0.8		6.9		3.7			
m-DN Benzene	150.2	1.6	214.5	150.2	64.4	7.0	7.4	108.7	108.7
		0.4		6.0		1.8			
m-DN Benzene	150.2	1.2	214.5	150.2	64.4	6.73	5.44	104.4	104.4
		0.70		0.49		3.26			

Table 4. (Continued)

Compound	Concentration mg/l		COD, mg/l		Concentration mg/l	TOC, mg/l		Recovery, %	
	Added	Found	Added	Found		Added	Found	COD	TOC
p-DN Benzene	51.3	6.6	73.2	6.6	51.3	22.0	45.5	9.0	207.0
		7.5		7.5			21.6	10.2	98.3
		7.9		7.9			19.0	10.7	86.4
		6.3		6.3			24.5	8.6	111.5
p-DN Benzene		7.1		7.1		\bar{X}	27.7	9.61	125.8
		0.8		0.8		s	12.1	1.02	55.1
	15.4	1.2	22.0	1.2	15.4	6.60	6.8	5.4	103.0
		2.0		2.0			6.0	9.09	90.9
Nitrobenzene		2.7		2.7			7.3	12.3	110.6
		2.4		2.4			6.9	10.9	104.5
		2.1		2.1		\bar{X}	6.75	9.42	102.3
		0.7		0.7		s	0.55	2.99	8.25
Nitrobenzene	156.7	198.0	305.4	198.0	156.7	91.7	83.0	64.8	90.5
		180.0		180.0			80.0	58.9	87.2
		216.0		216.0			80.0	70.7	87.2
		201.0		201.0			99.0	65.8	107.9
Nitrobenzene		199.0		199.0		\bar{X}	85.5	65.1	93.2
		14.8		14.8		s	9.11	4.84	9.92
	15.7	29.9	30.5	29.9	15.7	9.17	9.0	97.9	98.1
		31.8		31.8			8.4	104.1	91.6
Nitrobenzene		29.8		29.8			8.5	97.5	92.7
		30.5		30.5			8.8	99.9	96.0
Nitrobenzene		30.5		30.5		\bar{X}	8.68	99.9	94.6
		0.9		0.9		s	0.28	3.02	2.99

Table 5. COD and TOC test results for urea, nitramines, and guanidine nitrate

Compound	Concentration mg/l		COD, mg/l		Concentration mg/l	TOC, mg/l		Recovery, %	
	Added	Found	Added	Found		Added	Found	COD	TOC
Urea	206.8	0.00	440.7	0.00	207.6	41.5	32.0	0	77.1
		-1.58		-1.58			30.0	-0.36	72.2
		2.35		2.35			33.0	0.53	79.5
		0.79		0.79			30.0	0.18	72.2
Urea		0.39	\bar{x}	0.39		\bar{x}	31.3	0.09	75.3
		1.62	s	1.62		s	1.5	0.37	3.6
	20.7	-0.38	44.1	-0.38	20.8	4.16	2.2	0.86	52.8
		-0.38		-0.38			2.0	0.86	48.0
NQ		-0.78		-0.78			2.5	-1.76	60.1
		-1.58		-1.58			3.5	-3.58	84.1
		0.57	\bar{x}	0.57		\bar{x}	2.55	0.90	61.2
			s			s	0.67	2.16	16.0
NQ	502.6	1.17	927.3	1.17	502.2	57.7	45.0	0.13	78.0
		1.55		1.55			44.5	0.17	77.1
		1.96		1.96			48.5	0.21	84.0
		1.18		1.18			49.0	-0.13	84.8
NQ		0.88	\bar{x}	0.88		\bar{x}	46.8	0.09	81.0
		1.41	s	1.41		s	2.33	0.04	4.0
	6.02	-0.24	11.1	-0.24	6.03	0.69	0.5	-2.16	72.5
		-0.40		-0.40			0.4	-3.6	57.9
NQ		-0.12		-0.12			1.5	-1.08	217.3
		-0.82		-0.82			1.0	7.38	144.9
		0.02	\bar{x}	0.02		\bar{x}	0.85	3.55	123.0
		0.55	s	0.55		s	0.51	2.75	73.4

Table 5. (Continued)

Compound	Concentration mg/l	COD, mg/l		Concentration mg/l	TOC, mg/l		Recovery, %		
		Added	Found		Added	Found	COD	TOC	
RDX	30.4	39.4	13.2	31.0	5.05	3.6	33.5	71.3	
			12.8				3.6	32.5	71.3
			11.8				3.6	29.9	71.3
			13.0				3.5	32.9	69.3
RDX	6.20	\bar{x}	12.7	6.20	\bar{x}	3.6	32.2	70.8	
		s	0.6		s	0.05	1.60	1.0	
		8.04	2.6		1.01	1.0	32.3	99.0	
			1.9			1.0	23.6	99.0	
			1.9			1.5	23.6	148.5	
			1.2			1.0	32.3	99.0	
Tetryl	51.1	\bar{x}	1.9	51.1	\bar{x}	1.13	28.0	111.4	
		s	0.6		s	.25	5.02	24.8	
		59.8	47.1			13.8	78.7	92.2	
			49.0			14.0	81.9	93.5	
			49.3			16.0	82.4	106.5	
			48.5			14.0	81.1	93.5	
Tetryl	6.14	\bar{x}	48.5	6.14	\bar{x}	14.5	81.1	96.5	
		s	1.0		s	1.03	1.64	6.74	
		7.18	4.1		1.80	1.0	57.1	55.5	
			4.8			1.7	66.9	94.4	
			4.6			1.0	64.1	55.5	
			6.3			1.7	87.7	94.4	
	4.95		1.35	68.9	75.0				
	0.9		0.40	13.2	22.5				

Table 5. (Continued)

Compound	Concentration mg/l		COD, mg/l		Concentration mg/l		TOC, mg/l		Recovery, %	
	Added	Found	Added	Found	Added	Found	Added	Found	COD	TOC
HMX	5.8	7.0	7.58	7.0	5.8	0.94	0.5	92.3	53.2	
		6.2		6.2			1.0	81.8	106.3	
		6.8		6.8			1.0	89.7	106.3	
		7.5		7.5			1.3	98.9	138.3	
Guanidine Nitrate		6.88		6.88		\bar{x}	0.95	90.7	101.2	
		0.54		0.54		s	0.33	7.07	35.3	
	300.9	0.8	473.0	0.8	301.0	29.6	28.8	0.17	97.3	
		-0.8		-0.8			30.5	-0.17	103.0	
		-1.2		-1.2			30.0	-0.25	101.3	
		0.0		0.0			29.8	0	100.7	
Guanidine Nitrate		-0.3		-0.3		\bar{x}	29.8	-0.06	100.6	
		0.9		0.9		s	0.71	0.19	2.39	
	30.12	0.4	47.4	0.4	30.1	3.47	5.2	0.84	149.8	
		-0.4		-0.4			5.0	-0.84	144.0	
		0.4		0.4			5.1	0.84	146.9	
		0		0			5.0	0	144.0	
	0.1		0.1		\bar{x}	5.08	0.21	146.18		
	0.4		0.4		s	0.10	0.80	2.78		

Table 6. COD and TOC test results for nitrate esters

Compound	Concentration mg/l		COD, mg/l		Concentration mg/l	TOC, mg/l		Recovery, %	
	Added	Found	Added	Found		Added	Found	COD	TOC
NG	500.0	242.8	246.6	84.0	500.0	79.5	84.0	98.5	105.6
		242.9		80.0			80.0	98.5	100.6
		242.9		82.0			82.0	98.5	103.1
		244.6		71.0			71.0	99.2	89.1
NG		243.3	\bar{x}	79.3		\bar{x}	79.3	98.7	99.7
		0.9	s	5.73		s	5.73	0.35	7.29
	5.93	4.1	2.92	1.0	5.93	0.94	1.0	140.4	106.0
		2.2		1.2			1.2	75.3	127.3
TEGDW	479.2	473.0	478.9	135.0	479.0	143.2	135.0	100.5	94.3
		475.4		141.0			141.0	99.2	94.3
		476.2		180.0			180.0	98.8	98.5
		12.6		148.0			148.0	99.4	125.7
TEGDW	47.96	48.2	47.93	14.3	47.9	14.3	14.3	99.4	94.4
		45.3		13.5			13.5	94.5	94.4
		46.5		14.1			14.1	97.0	98.6
		45.7		18.0			18.0	95.3	125.8
TEGDW		46.4	\bar{x}	14.8		\bar{x}	14.8	96.6	103.3
		1.3	s	2.17		s	2.17	2.17	15.14

Table 7. COD and TOC test results for alcohols and acids

Compound	Concentration mg/l		COD, mg/l		Concentration mg/l	TOC, mg/l		Recovery, %	
	Added	Found	Added	Found		Added	Found	COD	TOC
Glycerine	395.9	477.1	481.4	156.7	398.0	145.3	99.1	92.7	
		476.4				150.0	99.0	95.7	
		478.5				145.0	99.4	92.5	
		476.1				150.0	98.9	95.7	
Glycerine		477.0	\bar{x}	\bar{x}		147.5	99.1	94.2	
		1.1	s	s		2.80	0.22	1.80	
	39.6	46.7	48.2	15.6	39.8	14.5	96.9	92.9	
		48.5				14.5	100.6	92.9	
Pentaerythritol		47.2	\bar{x}	\bar{x}		15.0	97.9	96.2	
		0.9	s	s		15.0	95.9	96.2	
	301.7	417.0	425.5	133.21	302.0	132.0	98.0	99.0	
		417.0				134.2	98.0	100.7	
Pentaerythritol		416.0	\bar{x}	\bar{x}		129.0	97.8	96.8	
		418.0	s	s		134.0	98.2	100.6	
	6.14	417.0	8.66	2.71	6.14	132.3	98.0	99.3	
		0.82				0.67	0.18	1.82	
Pentaerythritol		7.4				2.2	85.5	81.2	
		10.2				2.5	117.7	92.3	
	6.14	7.5	8.66	2.71	6.14	3.1	86.6	114.4	
		8.5				3.7	98.2	136.5	
Pentaerythritol		8.4	\bar{x}	\bar{x}		2.88	97.0	106.1	
		1.3	s	s		0.67	15.0	24.5	

Table 7. (Continued)

<u>Compound</u>	<u>Concentration</u> <u>mg/l</u>		<u>COD, mg/l</u>		<u>Concentration</u> <u>mg/l</u>	<u>TOC, mg/l</u>		<u>Recovery, %</u>	
			<u>Added</u>	<u>Found</u>		<u>Added</u>	<u>Found</u>	<u>COD</u>	<u>TOC</u>
Acetic Acid	313.2	313.2	333.9	328.0	125.3	120.0	98.2	95.8	
				326.0		123.6	97.6	98.6	
				325.0		120.0	97.3	95.7	
				327.0		120.0	97.9	95.7	
		\bar{x}	327.0		\bar{x}	120.9	97.8	96.5	
		s	1.0		s	1.80	0.39	1.44	
Acetic Acid	6.69	6.69	7.13	5.1	2.68	2.1	71.5	78.4	
				4.3		1.6	60.3	59.7	
				4.3		1.0	60.3	37.3	
				3.9		1.0	54.6	37.3	
		\bar{x}	4.4		\bar{x}	1.43	62.0	53.2	
		s	0.5		s	0.53	7.08	19.9	

Table 8. Statistical parameters for aromatic compounds

<u>Compound</u>	<u>Variance</u>		<u>Correlation Coefficient</u>
	<u>COD</u>	<u>TOC</u>	
Toluene	77.1	282.2	-0.85 *
o-Nitrotoluene	4.75	78.1	-0.44
m-Nitrotoluene	33.9	116.6	-0.94 *
p-Nitrotoluene	28.3	75.2	0.64
2,3 Dinitrotoluene	85.7	557.0	0.85 *
2,4 Dinitrotoluene	46.2	72.6	0.83 *
2,5 Dinitrotoluene	21.9	191.5	0.33
2,6 Dinitrotoluene	193.2	58.5	-0.071
3,4 Dinitrotoluene	60.5	409.7	0.73 *
2,4,6 Trinitrotoluene	61.6	1956.3	-0.80 *
Benzene	282.2	234.1	0.93 *
m-Dinitrobenzene	5.86	97.4	0.38
p-Dinitrobenzene	4.37	1490.0	-0.11
Nitrobenzene	359.9	46.6	0.11
Tetryl	116.9	367.1	0.82 *

* Values significant at 95% confidence level

Table 9. Statistical parameters for nonaromatic compounds

<u>Compound</u>	<u>Variance</u>		<u>Correlation Coefficient</u>
	<u>COD</u>	<u>TOC</u>	
Urea	2.34	171.6	-0.48
Nitroguanidine	10.4	2819.6	0.35
Guanidine Nitrate	0.31	599.8	0.30
RDX	17.2	734.4	-0.72 *
HMX	50.0	1246.1	0.24
Nitroglycerin	334.9	488.4	-0.145
Triethylene Glycol Dinitrate	4.67	196.0	0.21
Glycerin	2.25	2.99	-0.45
Pentaerythritol	96.0	272.3	0.055
Acetic Acid	392.0	707.6	0.95 *

* Values significant at 95% confidence level

Table 10. Comparison of completeness of recovery by COD and TOC tests

<u>t - TEST VALUES</u>			
COD _{H1}	COD _{Lo}	TOC _{H1}	TOC _{Lo}
4.66*	5.76*	1.55	0.84

* VALUES ARE SIGNIFICANTLY DIFFERENT FROM 100% RECOVERY AT 90% LEVEL OF CONFIDENCE FOR 23 DEGREES OF FREEDOM.




Figure 1. Comparison of COD and TOC for nitrotoluenes and toluene at high concentration




Figure 2. Comparison of COD and TOC for nitrotoluenes and toluene at low concentration




Figure 3. Comparison of COD and TOC for nitrobenzenes and benzene at high concentration




Figure 4. Comparison of COD and TOC for nitrobenzenes and benzene at low concentration




Figure 5. Comparison of COD and TOC tests for nitramines, guanidine nitrate, and urea at high concentration




Figure 6. Comparison of COD and TOC tests for nitramines, guanidine nitrate, and urea at low concentration




Figure 7. Comparison of COD and TOC for nitrate esters, alcohols and acids at high concentration



Figure 8. Comparison of COD and TOC for nitrate esters, alcohols and acids at low concentration

DISTRIBUTION LIST

Commander
US Army Armament Research
and Development Command
ATTN: DRDAR-CG
DRDAR-LCM
DRDAR-LCM-SA (6)
DRDAR-LCP-F
DRDAR-TSS (5)
Dover, NJ 07801

Commander
US Army Materiel Development
and Readiness Command
ATTN: DRCDE
DRCDE-E
DRCIS
DRCPA-E
DRCRP-I
DRCDE
DRCDL
DRCSG-S
5001 Eisenhower Avenue
Alexandria, VA 22333

Commander
USDRC Installation & Services Agency
ATTN: DRCIS-RI-IU
DRCIS-RI-IC
Rock Island, IL 61299

Commander
US Army Armament Materiel
Readiness Command
ATTN: DRSAR-IMB-C (2)
DRSAR-RD
DRSAR-ISE (2)
DRSAR-SC
DRSAR-EN
DRSAR-PPW
DRSAR-ASF (2)
Rock Island, IL 61201

**Project Manager for Munition Production
Base Modernization and Expansion
US Army Materiel Development
and Readiness Command
ATTN: DRCPM-PBM-EC
DRCPM-PBM-T-EV
Dover, NJ 07801**

**Department of the Army
Office, Chief of Research, Development
and Acquisition
ATTN: DAMA-CSM-P
Washington, DC 20310**

**Commander
US Army Procurement Equipment Agency
ATTN: DRX-PE-MT
Rock Island, IL 61299**

**Department of the Army
ATTN: DAEN-ZCE
Washington, DC 20310**

**Commander
Chemical System Laboratory
ARRADCOM
ATTN: DRDAR-CLM-T
Aberdeen Proving Ground, MD 21010**

**Defense Contract Administration Services (2)
1610 S. Federal Building
100 Liberty Avenue
Pittsburgh, PA 15222**

**Defense Documentation Center (12)
Cameron Station
Alexandria, VA 22314**

Commander
US Army Construction Engineering
Research Laboratory
ATTN: CERL-ER
Champaign, IL 61820

Office, Chief of Engineers
ATTN: DAEN-MCZ-E
Washington, DC 20314

US Army Engineer District, New York
ATTN: Construction District
26 Federal Plaza
New York, NY 10007

US Army Engineer District, Baltimore
ATTN: Construction Division
P.O. Box 1715
Baltimore, MD 21203

US Army Engineer District, Norfolk
ATTN: Construction Division
803 Front Street
Norfolk, VA 23510

US Army Engineer District, Mobile
ATTN: Construction Division
P.O. Box 2288
Mobile, AL 36628

US Army Engineer District, Fort Worth
ATTN: Construction Division
P.O. Box 17300
Fort Worth, TX 76102

US Army Engineer District, Omaha
ATTN: Construction Division
6014 USPO and Courthouse
215 North 17th Street
Omaha, NE 68102

US Army Engineer District, Kansas City
ATTN: Construction Division (2)
700 Federal Building
Kansas City, MO 64106

US Army Engineer District, Sacramento
ATTN: Construction Division
650 Capitol Mall
Sacramento, CA 95814

US Army Engineer District, Huntsville
ATTN: Construction Division
P.O. Box 1600 West Station
Huntsville, AL 35807

Commander
US Army Environmental Hygiene Agency
ATTN: HSE-E (2)
Aberdeen Proving Ground, MD 21010

Commander
Badger Army Ammunition Plant
ATTN: SARBA-CE
Baraboo, WI 53913

Commander
Cornhusker Army Ammunition Plant
ATTN: SARCO-E
Grand Island, NB 68801

Commander
Holston Army Ammunition Plant
ATTN: SARHO-E
Kingsport, TN 37662

Commander
Indiana Army Ammunition Plant
ATTN: SARIN-OR
Charlestown, IN 47111

Commander
Naval Weapons Support Center
ATTN: Code 5042, Mr. C. W. Gilliam
Crane, IN 47522

Commander
Iowa Army Ammunition Plant
ATTN: SARIO-A
Middletown, IA 52638

Commander
Joliet Army Ammunition Plant
ATTN: SARJO-SS-E
Joliet, IL 60436

Commander
Kansas Army Ammunition Plant
ATTN: SARKA-CE
Parsons, KS 67537

Commander
Lone Star Army Ammunition Plant
ATTN: SARLS-IE
Texarkana, TX 57701

Commander
Longhorn Army Ammunition Plant
ATTN: SARLO-O
Marshall, TX 75670

Commander
Louisiana Army Ammunition Plant
ATTN: SARLA-S
Shreveport, LA 71102

Commander
Milan Army Ammunition Plant
ATTN: SARMI-S
Milan, TN 38358

Commander
Newport Army Ammunition Plant
ATTN: SARNE-S
Newport, IN 47966

Commander
Pine Bluff Arsenal
ATTN: SARPB-ETA
Pine Bluff, AR 71601

Commander
Radford Army Ammunition Plant
ATTN: SARRA-IE
Radford, VA 24141

Commander
Ravenna Army Ammunition Plant
Ravenna, OH 44266

Commander
Sunflower Army Ammunition Plant
ATTN: SARSU-O
Lawrence, KS 66044

Commander
Volunteer Army Ammunition Plant
ATTN: SARVO-T
Chattanooga, TN 34701

Army Logistics Management Center
Environmental Management
ATTN: Mr. Otto Nauman (2)
Fort Lee, VA 23801

Project Manager for Chemical Demilitarization
and Installation Restoration
ATTN: DRCPM-DRR, Mr. Harry Sholk
Aberdeen Proving Ground, MD 21010

Department of the Army
ATTN: DAEN-FEU
Washington, DC 20314

Dr. John A. Brown, Chairman
P.O. Box 145
Berkeley Heights, NJ 07922

Dr. Helmut Wolf, Consultant
120 Skyline Drive
Fayetteville, AR 72701

Dr. Fred Smetana, Consultant
5452 Parkwood Drive
Raleigh, NC 27612

Dr. Zachary Shetman, Consultant
109 N. Broadway
White Plains, NY 10603

Commander
US Army Armament Materiel and
Readiness Command
ATTN: DRSAR-LEP-L
Rock Island, IL 61299

Director
US Army TRADOC Systems
Analysis Activity
ATTN: ATAA-SL (Tech Lib)
White Sands Missile Range, NM 88002

Weapon System Concept Team/CSL
ATTN: DRDAR-ACW
Aberdeen Proving Ground, MD 21010

Technical Library
ATTN: DRDAR-CLJ-L
Aberdeen Proving Ground, MD 21010

Technical Library
ATTN: DRDAR-TSB-S
Aberdeen Proving Ground, MD 21005

Technical Library
ATTN: DRDAR-LCB-TL
Benet Weapons Laboratory
Watervliet, NY 12189

ED
78