

AD-A089 993

STANFORD UNIV CA EDWARD L GINZTON LAB OF PHYSICS F/6 20/3  
REMOTE ATMOSPHERIC MEASUREMENTS OF CH<sub>4</sub> USING A LINBO3 TUNABLE S--ETC(U)  
MAR 80 R L BYER, M ENDEMANN EPA-R-605750-01


UNCLASSIFIED

AFESC/ESL-TR-80-11 NL

1 1  
A  
20 0 1 4



END  
DATE  
FILMED  
10 80  
DTIC



MICROCOPY RESOLUTION TEST CHART  
NATIONAL BUREAU OF STANDARDS, 1963-A

AD A089993

**LEVEL**

ESL-TR-80-11

# REMOTE ATMOSPHERIC MEASUREMENTS OF CH<sub>4</sub> USING A LiNbO<sub>3</sub> TUNABLE SOURCE

ROBERT L. BYER  
MARTIN ENDEMANN  
EDWARD L. GINZTON LABORATORY OF PHYSICS  
STANFORD UNIVERSITY  
STANFORD, CA 94305

MARCH 1980  
FINAL REPORT  
OCTOBER 1978 - JUNE 1979

DTIC  
ELECTE  
S OCT 7 1980 D  
A

APPROVED FOR PUBLIC RELEASE; DISTRIBUTION UNLIMITED



ENGINEERING AND SERVICES LABORATORY  
AIR FORCE ENGINEERING AND SERVICES CENTER  
TYNDALL AIR FORCE BASE, FLORIDA 32403

80 10 3 132

**NOTICE**

Please do not request copies of this report from  
HQ AFESC/RD (Engineering and Services Laboratory).  
Additional copies may be purchased from:

National Technical Information Service  
5285 Port Royal Road  
Springfield, Virginia 22161

Federal Government agencies and their contractors  
registered with Defense Technical Information Center  
should direct requests for copies of this report to:

Defense Technical Information Center  
Cameron Station  
Alexandria, Virginia 22314





## TABLE OF CONTENTS

Section	Title	Page
I	INTRODUCTION	1
II	DESCRIPTION OF THE MONITORING SYSTEM	3
III	MEASUREMENTS	7
IV	CONCLUSIONS	19
	REFERENCES	20

## LIST OF FIGURES

Figure	Title	Page
1	Schematic of the Nd:YAG laser pumped LiNbO <sub>3</sub> tunable source transmitter.	4
2	Schematic of the electronic control and data processing system based on the PDP11 mini-computer.	6
3	A segment of an atmospheric transmission scan over a 1.55-kilometer path in the spectral range from 5500-5800 wave numbers.	8
4	A detailed segment of the scan shown in Figure 3.	9
5	A segment of an atmospheric absorption scan near the CH <sub>4</sub> fundamental absorption band at 3.4 microns.	10
6	An in-laboratory CH <sub>4</sub> absorption spectrum of the identical spectral region shown in Figure 5.	11
7	A section of an 18-hour measurement of CH <sub>4</sub> on 31 May 1979 over a 1.55-kilometer path.	12
8	A part of the CH <sub>4</sub> measurement showing CH <sub>4</sub> fluctuations over the path being monitored.	14
9	CH <sub>4</sub> ppm versus time over a 12-hour period.	15
10	CH <sub>4</sub> ppm versus time for a second measurement run.	16
11	A section of the H <sub>2</sub> O atmospheric absorption spectrum of use for H <sub>2</sub> O and atmospheric temperature measurements.	17

## SECTION I

### INTRODUCTION

In 1973 Byer and Garbuny (Reference 1) theoretically analyzed remote atmospheric pollutant monitoring by absorption. Two methods considered were long path absorption measurements using topographical targets and depth resolved absorption measurements using Mie scattering in the atmosphere as a distributed reflector.

Progress in remote monitoring since that first theoretical study has been delayed due to the slow development of tunable laser sources. Depth resolved absorption measurements referred to by the acronym DIAL-LIDAR have now been demonstrated in the visible, ultraviolet and infrared spectral ranges. The progress in remote measurements in tunable source LIDAR has been reviewed, (References 2 and 3) and this method is presently accepted as the primary measurement approach for long range atmospheric measurements using laser sources.

An early research effort at Stanford was to invent and develop a high power tunable source for remote monitoring applications. Under National Science Foundation (NSF) support a tunable infrared source based on a Nd:YAG-pumped  $\text{LiNbO}_3$  Optical Parametric Oscillator (OPO) was devised and demonstrated. The early device demonstrated a tuning range from 1.4 to 4.0 microns with a linewidth of 4 wavenumbers at an output energy of 10 millijoules per pulse at 10 Hertz.

The wide infrared tuning range and all solid-state construction of the tunable source showed promise for use in infrared remote monitoring. The solid state construction had a potential for reliable operation, and the wide infrared tuning range included molecular absorption bands of a wide variety of pollutant molecules. An additional benefit of the infrared source was eye-safe operation under either daylight or nighttime operations.

The OPO tunable source was used for remote measurements at Stanford in 1977-1978. The Stanford remote monitoring system, which includes a 16 inch diameter receiving telescope, minicomputer and Nd:YAG laser source was designed and constructed under NSF and Electric Power Research Institute (EPRI) support. The early measurements of  $\text{SO}_2$  and  $\text{CH}_4$  by Baumgartner and Byer (References 4 and 5) were conducted as a demonstration of the measurement potential of the system.

Following the early measurements, an extensive program of research on the OPO source and an optical parametric amplifier (OPA) for increased energy output was carried out. This program led to significant improvements in the tunable source energy, linewidth control, and automatic computer control (References 6

and 7). It also led to a design which demonstrated long term reliable operation under full automatic computer control. The improved OPO/OPA source was used as a transmitter in the present system.

Single-ended remote atmospheric measurements require considerable transmitted energy. For example, measurements using topographical targets typically require 0.01 to 0.1 joules per pulse transmitted energy. For DIAL-LIDAR measurements the required energy increases to near 0.1 to 1.0 joules due to the weak return signal. Future remote monitoring progress requires increased tunable laser capability.

Independent research has been carried out at Stanford to investigate new approaches to high pulse energy tunable infrared sources. The results show good potential for transmitted sources based on Nd:Glass and on pressure tunable CO<sub>2</sub> followed by Raman shifting. These future high energy sources make the potential for depth resolved infrared LIDAR appear particularly good.

The measurement program described in this report illustrates the advantages of the LIDAR approach for remote monitoring. These advantages include rapid measurement time, path averaged measurements, multiple pollutant capability, and excellent measurement precision. The measurement capability will improve as new higher energy sources become available.

## SECTION II

### DESCRIPTION OF THE MONITORING SYSTEM

The current monitoring system is a derivative of the one reported earlier by Baumgartner and Byer (References 4 and 5). Although the basic set-up remained the same, many essential changes have been made to increase the reliability of the system.

Figure 1 shows a schematic of the present set-up. The pump source for the Optical Parametric Oscillator (OPO) and the Optical Parametric Amplifier (OPA) is an unstable resonator Nd:YAG laser. However, to prevent damage to the  $\text{LiNbO}_3$  crystals in the OPO and OPA, two important changes were made. First, the cavity length of the Nd:YAG laser was doubled to 1.50 meters to increase the pulse length to 18 nanoseconds and so to decrease the peak optical intensity. This avoids the danger of crystal tracking due to self-focusing effects. The second change was the introduction of a far-field converter, which transforms the doughnut-shaped near-field laser output beam into an Airy-disk profile, which ensures better beam homogeneity. This, in turn, lowers the threshold of the OPO and also avoids hot spots in the beam profile and thus reduces the chance of crystal surface damage.

Extensive studies by Brosnan and Byer lead to improvements of the OPO itself (Reference 6). The introduction of a 10:1 prism beam expander before the grating decreases the output linewidth to less than one wavenumber over the entire tuning range. This linewidth is adequate to resolve most lines of atmospheric pollutants. If necessary, it is possible to reduce the linewidth to 0.1 wavenumbers by introducing an intracavity tilted etalon. To increase the available tunable output energy, an OPA, which was studied by Baumgartner and Byer (Reference 7), was added after the OPO source. The OPA now generates 20 millijoules output energy. However, optimization of beam overlap should increase the output energy to previously obtained values of 50 millijoules.

About 1 percent of the tunable output is reflected into a photo-acoustic cell to continuously monitor the cross-section of the observed species. The photo-acoustic cell is followed by an InSb detector, which is used as a reference detector to ratio out energy fluctuations of the OPA, and by an absorption cell to measure the absolute absorption cross-section of a pollutant at a given wavelength.

The main portion of the beam is expanded to about 3 centimeters diameter and transmitted to the telescope on the roof of the laboratory. Detection of the backscattered signal is done with a 40 centimeter,  $f=3$  Newtonian telescope with a liquid-nitrogen-cooled InSb detector at the focal plane.




Figure 1. Schematic of the Nd:YAG laser pumped LiNbO<sub>3</sub> tunable source transmitter. The source operates at 10 pulses per second with 20 millijoules output energy tunable under computer control over a 1.4-to 4.0-micron range at 1.0 wavenumber linewidth. An etalon can be used to narrow the linewidth to less than 0.1 wavenumber.

Figure 2 shows a schematic of the improved data processing and control electronics. It is centered around the PDP/11E10 minicomputer, which uses an extensive interactive program to tune the OPO, take data, process the data, and finally display it. The signals from the cooled InSb detector and the reference detector are integrated over one shot in a gated integrator and read out by the computer which ratios and averages them. In one mode the computer tunes the OPO on and off a resonance line, uses the signal from the photo-acoustic cell to correct for slow changes of the absorption cross-section, and immediately prints the measured concentration or stores the data for laboratory processing. Another operating mode provides automatic atmospheric or in-laboratory transmittance scans over wide spectral regions. The control program greatly simplifies the handling of the system.

Important improvements were made to the remote monitoring system. The tunable source was modified by adding a prism beam expander to the OPO resonator cavity. This led to a reduction in the grating-only linewidth from 4 wavenumbers to 0.9 wavenumbers. The OPO/OPA system was enclosed in a dust-free housing and operated at a lower energy level to ensure long-term damage-free operation. Reliable operation from January 1979 to June 1979 verified that the all solid-state tunable source can be trouble-free for extended operation times.

Recently, a tilted etalon for linewidth control to 0.1 wavenumbers was installed. The galvanometer tilt control proved to have inadequate resettability and is being replaced by a stepper motor device. However, the etalon did operate over an extended 40 wavenumber scan range under computer control. When the new mechanical system is completed, atmospheric measurements can be conducted at the improved resolution.

The second area of improvement was the system control software. The software package now allows control of the tunable source, monitoring of an absorption cross-section by means of the photo-acoustic cell, normalization of the transmitted pulse energy, data collection, and processing of the returned signal. Software was also developed to display the recorded spectrum in real time or to calculate and plot the measured pollutant concentration.

Figure 2 shows a schematic of the detection system and indicates the signal processing steps. The received signal  $S(\nu_1)$  and reference signal  $R(\nu_1)$  and the rms deviation of the normalized signal are calculated along with the quantity  $A = \log [S(\nu_1) / S(\nu_2)]$  which is proportional to the product  $N\sigma L$  where  $N$  is the pollutant density,  $\sigma$  is the cross-section, and  $L$  is the path. Typically 20 shots on resonance and 20 shots off resonance are processed and stored. Thus, one measurement takes 4 seconds. Further signal processing and averaging is done later with the stored values  $A$  and the rms variation of  $A$ .




Figure 2. Schematic of the electronic control and data processing system based on the PDP11 minicomputer.

### SECTION III

#### MEASUREMENTS

Using the tunable OPO/OPA source, single ended atmospheric absorption measurements were made. The typical path length was 1.54 kilometers, using the Hoover Tower structure as the topographical target. Spectral scans over a wavelength range from 1.45 to 4.0 microns have been made.

Figure 3 shows a section of a spectral scan in the 5500 to 5800 wavenumbers range. The 300 wavenumber scan is a part of a scan that extends to 2500 wavenumbers. The region shown in Figure 3 covers the overtone absorption band of H<sub>2</sub>O. The absorption peaks are not noise but are individual absorption lines of water vapor. To illustrate the resolution of the source, Figure 4 shows a section of water vapor absorption over a 40 wavenumber interval. In this scan the water vapor lines have been labeled. The scan has a resolution near 1 wavenumber and, therefore, does not resolve the water vapor lines which have half-widths near 0.03 to 0.1 wavenumbers. However, water vapor density can be determined with the 1 wavenumber resolution if an effective absorption cross-section is used.

Figure 5 shows a segment of the atmospheric absorption spectrum near the methane 3.4-micron fundamental band. Both CH<sub>4</sub> and H<sub>2</sub>O absorption lines are present. To clearly identify the CH<sub>4</sub> lines and to aid in selecting a transition free from water vapor, an absorption spectrum of CH<sub>4</sub> was taken in the laboratory. A section of the scan centered on the P-branch is shown in Figure 6. Figures 5 and 6 can be overlaid to show that the P10 methane transition is of proper strength and nearly water vapor free for atmospheric measurement over the 1.5 kilometer path.

The ability to select interference-free absorption bands and to verify the location of spectral lines by scanning the spectrum is a major advantage of the tunable source.

The P10 transition was selected for CH<sub>4</sub> measurements. The computer was programmed to tune onto the P10 line and off the line every 20 pulses. The data was normalized, ratioed, and processed to show ppm CH<sub>4</sub>. An effective cross-section for CH<sub>4</sub> (P10) was utilized based on the resolution of the OPO/OPA source. To correct for any long-term drift, the cross-section was continuously measured during a run by a photo-acoustic detector cell. The value of the cross-section was used to calculate the methane concentration in near real time.

Figure 7 shows a segment of an 18-hour CH<sub>4</sub> measurement run. Here the raw measurements are shown for 20 shots/line averaging




Figure 3. A segment of an atmospheric transmission scan over a 1.55-kilometer path in the spectral range from 5500 to 5800 wavenumbers.




Figure 4. A detailed segment of the scan shown in Figure 3. Here a region of H<sub>2</sub>O absorption is shown in which H<sub>2</sub>O lines have been identified and labeled.




Figure 5. A segment of an atmospheric absorption scan near the  $\text{CH}_4$  fundamental absorption band at 3.4 microns.




Figure 6. An in-laboratory CH<sub>4</sub> absorption spectrum of the identical spectral region shown in Figure 5. Figure 6 can be overlaid onto Figure 5 to show the CH<sub>4</sub> absorption peaks and to help select water vapor-free CH<sub>4</sub> absorption lines.




Figure 7. A section of an 18-hour measurement of CH<sub>4</sub> on 31 May 1979 over a 1.55 kilometer path. The 20-shot average yields a 4-second measurement time and a 0.6-ppm rms error.

or 4 seconds per point. The rms fluctuations are calculated to be 0.6 parts per million. To improve the precision of the measurement longer averaging times can be used. Figure 8 shows a section of the same CH<sub>4</sub> measurement run with an 80-second averaging time. Here the rms deviation is reduced to 0.03 ppm which is less than the CH<sub>4</sub> fluctuation level in the atmosphere.

During these measurements there was concern that the measured CH<sub>4</sub> levels were too high since the global average CH<sub>4</sub> level is near 1.5 ppm. Therefore, hourly CH<sub>4</sub> measurements were obtained from the Redwood City Bay Area Pollution Control District point monitoring station.

Figure 9 shows the measurements at Stanford over a 10-hour period and the hourly averaged measurements from the monitoring station which is located 8 kilometers north of Stanford. The agreement is excellent and is within the variations of CH<sub>4</sub> in the atmosphere. On the night of 30 May 1979 the weather was clear with a very light wind. The anomalously high CH<sub>4</sub> levels observed in the south bay area may be due to the large biomass production rate from marshes and city dumps located along the bay itself.

To check further on the correlation between our measurements and the station measurements, a second CH<sub>4</sub> run was made on 6 and 7 June 1979. Figure 10 shows the results and comparison. In this case the wind was blowing down the bay from Redwood City toward Stanford. The wind velocity and distance gave a 20 to 30-minute delay time from Redwood City to Stanford. This delay time is in good agreement with the peak CH<sub>4</sub> measurement at 22.00 at Redwood City which was observed at Stanford at 22.30. Later in the evening the wind ceased, making the measurements independent of the Redwood station.

It should be noted that the LIDAR system was making CH<sub>4</sub> measurements every 50 seconds. The Redwood City flame ionization detector measurements are averaged over 1-hour intervals for recording. Even though the measurement systems are separated physically, the agreement in measured CH<sub>4</sub> levels is excellent as should be expected in a region free of large sources of CH<sub>4</sub> in the area between Stanford and Redwood City.

During the final phase of this program the H<sub>2</sub>O spectrum was investigated to determine optimal wavelengths for density and temperature measurements. Figure 11 shows a section of the H<sub>2</sub>O absorption spectrum near 5650 wavenumbers. The labeled peaks have been selected for atmospheric temperature measurements due to their close proximity in wavenumber and their widely spaced lower levels. Here E<sub>0</sub> denotes the energy above the H<sub>2</sub>O ground state in wavenumbers.

These two H<sub>2</sub>O lines should allow temperature measurements to better than 1°C for a 1 percent precision in absorption




Figure 8. A part of the CH<sub>4</sub> measurement showing CH<sub>4</sub> fluctuations over the path being monitored. The statistical rms noise of the measurement is 0.03 ppm which is less than the CH<sub>4</sub> fluctuation level.




Figure 9. CH<sub>4</sub> ppm versus time over a 12-hour period. The Stanford LIDAR data is plotted as crosses. The point-monitoring hourly-averaged CH<sub>4</sub> levels are shown as open circles.




Figure 10. CH<sub>4</sub> ppm versus time for a second measurement run. The Stanford LIDAR data is plotted as crosses. Here a prevailing wind yielded a 30-minute delay between the atmosphere at Redwood City point monitoring station and its arrival near Stanford campus.




Figure 11. A section of the H<sub>2</sub>O atmospheric absorption spectrum of use for H<sub>2</sub>O and atmospheric temperature measurements.

measurement. The CH<sub>4</sub> measurements show that such a measurement precision is possible for an averaging time of approximately 5 minutes. Preliminary measurements have shown that the peak ratios of the two H<sub>2</sub>O lines do vary with atmospheric temperature as expected. Improvements are being made to the linewidth control of the OPO before attempting long-term atmospheric temperature measurements.

The primary emphasis of this program was to demonstrate long-term reliable remote air pollution measurement capability. The decision was made to concentrate on CH<sub>4</sub> rather than to attempt measurements on a wider variety of molecules. However, it should be noted that within the OPO 1.4- to 4.0-micron tuning range are absorption bands of a number of hydrocarbons, including ethane, propane, ethylene, and formaldehyde. Other molecules of potential interest include HCl, HF, N<sub>2</sub>O, NO<sub>2</sub>, H<sub>2</sub>S, NH<sub>3</sub>, CO, and CO<sub>2</sub>.

A measurement capability has been demonstrated using an OPO/OPA source. It is hoped to extend these results to other molecules and to atmospheric meteorological parameter measurements in the future.

## SECTION IV

### CONCLUSIONS

The reliable use of an OPO-based remote air pollution system has been demonstrated. Fully automatic operation of this system is possible. Future measurements will demonstrate the capability to measure many pollutants with the same system, thus showing another advantage of the tunability. The advantages of the continuous tunability for cross-section optimization and interference avoidance have been clearly demonstrated. The possibility of improving the measurement accuracy by appropriate signal averaging has also been demonstrated. Finally, long-term operation under full computer control of the all solid-state laser source and  $\text{LiNbO}_3$  OPO/OPA tunable source has demonstrated the potential for future use in remote atmospheric measurements.

## REFERENCES

1. R. L. Byer and M. Garbuny, "Pollutant Detection by Absorption Using Mie Scattering and Topographical Targets as Retroreflectors", *Applied Optics*, 12, p.1496 (1973).
2. R. L. Byer, "Review: Remote Air Pollution Measurements", *Optical and Quantum Electronics*, 7, p.147 (1975).
3. J. Murray and R. L. Byer, "Remote Monitoring Techniques", in Handbook of Air Pollution Analysis, ed. by R. Perry and R. J. Young, Chapman and Hall, London 1977.
4. R. A. Baumgartner and R. L. Byer, "Remote SO<sub>2</sub> Measurements at 4  $\mu$ m with a Continuously Tunable Source", *Optics Letters*, 2, p.163 (1978).
5. R. A. Baumgartner and R. L. Byer, "Continuously Tunable ir LIDAR with Applications to Remote Measurements of SO<sub>2</sub> and CH<sub>4</sub>", *Applied Optics*, 17, p.3555 (1978).
6. S. J. Brosnan and R. L. Byer, "Optical Parametric Oscillator Threshold and Linewidth Studies", *IEEE Journ. Quant. Elect.* QE-15, p.415 (1970).
7. R. A. Baumgartner and R. L. Byer, "Optical Parametric Amplification", *IEEE Journ. Quant. Elect.* QE-15, p 432 (1979).

INITIAL DISTRIBUTION

DTIC/DDA	2
HQ USAF/LEEV	1
OSAF/MIQ	1
OSAF/OI	1
OASD/(I&L)EES	1
AFIT/Library	1
EPA/ORD	1
USA Chief, R&D/EQ	1
OEHL/CC	4
USAFSAM/EDE	2
AFOSR/CC	1
AFOSR/NC	1
AUL/LSE 71-249	1
HQ USAFA/Library	1
HQ AFESC/TST	1
AFATL/DLODL	1
OUSDR&E	1
AFWL/SUL	1
USAFSAM/VNL	1
AFOSR/NP	1
EPA/ESRL	5
HQ AFESC/RDV	1
HQ AFESC/RDVC	5
Stanford University	2
NASA/Langley R.C.	1
MIT Lincoln Laboratory	1
USAEHA	1
US Army Med Bioengrg R&D Lab	2
1 Med Svc Wg/SGB	1
Chem Abstracts Service	1
Environmental Research Lab	1