



MR 80-8

Instrumentation at CERC's Field Research Facility, Duck, North Carolina

by H. Carl Miller


MISCELLANEOUS REPORT NO. 80-8 OCTOBER 1980

C

AD A 0917

FILE COPY

- 41-65





Approved for public release; distribution unlimited.

U.S. ARMY, CORPS OF ENGINEERS COASTAL ENGINEERING RESEARCH CENTER

> Kingman Building Fort Belvoir, Va. 22060

Reprint or republication of any of this material shall give appropriate credit to the U.S. Army Coastal Engineering Research Center.

Limited free distribution within the United States of single copies of this publication has been made by this Center. Additional copies are available from:

> National Technical Information Service ATTN: Operations Division 5285 Port Royal Road Springfield, Virginia 22161

Contents of this report are not to be used for advertising, publication, or promotional purposes. Citation of trade names does not constitute an official endorsement or approval of the use of such commercial products.

The findings in this report are not to be construed as an official Department of the Army position unless so designated by other authorized documents.

CERC-REPORT DOCUMENTATION PAGE	READ INSTRUCTIONS BEFORE COMPLETING FORM
REPORT NUBER 2. GOVT ACCESSIN MR-80-8 AD -AO	ON NO. 3. RECIPIENT'S CATALOG NUMBER
TITLE (and Subtilie) INSTRUMENTATION AT CERC'S FIELD RESEARCH	Miscellaneous Repert.
FACILITY, DUCK, NORTH CAROLINA	PERFORMING ORG. AEPORT HUMBER
H. Carl Miller	8. CONTRACT OR GRANT NUMBER(*)
PERFORMING ORGANIZATION NAME AND ADDRESS Department of the Army Coastal Engineering Research Center (CERRE-CO Kingman Building, Fort Belvoir, Virginia 220	
CONTROLLING OFFICE NAME AND ADDRESS Department of the Army	October 480
Coastal Engineering Research Center Kingman Building, Fort Belvoir, Virginia 220	
MONITORING AGENCY NAME & ADDRESS(II different from Controlling O	(fice) 15. SECURITY CLASS. (of this report) UNCLASSIFIED
The	15e. DECLASSIFICATION/DOWNGRADING SCHEDULE
Approved for public release; distribution unl	imited.
DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if differ	
DISTRIBUTION STATEMENT (of the abetract entered in Block 20, If different entered in Block 20, If different and the second secon	rent from Report)
DISTRIBUTION STATEMENT (of the abetract entered in Block 20, If different entered in Block 20, If different and the second secon	rent from Report)
 DISTRIBUTION STATEMENT (of the abstract entered in Block 20, it differ SUPPLEMENTARY NOTES KEY WORDS (Continue on reverse side if necessary and identify by block n Coastal Engineering Research Center Duck, North Carolina Field Research Facility 	rent from Report) number) Meteorological instruments Oceanographic instruments
 DISTRIBUTION STATEMENT (of the abstract entered in Block 20, it differ SUPPLEMENTARY NOTES KEY WORDS (Continue on reverse side if necessary and identify by block n Coastal Engineering Research Center Duck, North Carolina 	number) Meteorological instruments Oceanographic instruments momber) eorological instrumentation used coastal Engineering Research it Duck, North Carolina; the of the instrument data is also

4.20

PREFACE

This report provides a basic description of the oceanographic and meteorological instrumentation used for the collection of environmental data at the Coastal Engineering Research Center's (CERC) Field Research Facility (FRF) at Duck, North Carolina; it also provides the information necessary for proper interpretation of the instrument data. The report is intended as a reference for subsequent reports which will contain summaries of the data collected from the instruments. The work was carried out under the CERC's waves and coastal flooding program.

The report was prepared by H. Carl Miller, Oceanographer, under the supervision of Dr. C.L. Vincent, Chief, Coastal Oceanography Branch and C. Mason, Chief, Field Research Facility Group, Research Division.

The author acknowledges the helpful review comments from M.G. Mattie of CERC, R. Masterson, National Ocean Survey (NOS), and J. O'Brien, National Weather Service (NWS), National Oceanic and Atmospheric Administration (NOAA).

Comments on this publication are invited.

A STREET, SALES, STREET, STREE

Approved for publication in accordance with Public Law 166, 79th Congress, approved 31 July 1945, as supplemented by Public Law 172, 88th Congress, approved 7 November 1963.

TED E. BISHOP Colonel, Corps of Engineers Commander and Director

Accession For TIS CRA&1 DITO 11 101 i Grann-Just 11 • . cr 811 1.44 ۰₁. 1111 - 642. :. Avail chales : puot ni 127

CONTENTS

	Page	
	CONVERSION FACTORS, U.S. CUSTOMARY TO METRIC (SI) 6	
	I INTRODUCTION	
I	INSTRUMENTATION AT THE FRF 7 1. Oceanographic Instruments 7 2. Meteorological Instrumentation 23	
11	SAMPLING INFORMATION AND SUMMARY DESCRIPTIONS. 31 1. Digital Wave Data 31 2. Current Meter Data 31 3. Conductivity and Temperature Data 31 4. Monthly Tide Summaries. 31 5. Meteorological Observations 37	
1	v DATA REQUESTS	
	LITERATURE CITED	
APPE	CNDIX INSTRUMENT INSTALLATION CHART	
	FIGURES	
1	FRF location map	
	Instrument locations at FRF	
3	Baylor wave staff gage and support system	
4	Water depth envelope along the FRF pier	
5	Waverider buoys	
6	Waverider moorings	
7	Pressure wave gage	
8	T-bracket used for current meter and pressure gage installations 15	
9	Current meter installation	
10	Hydrolab Model TC-2 conductivity probe	
11	Hydrolab Model T-4 temperature probe	
	Hydrolab temperature and conductivity meters	
	NOS control station at end of pier	
	Leupold-Stevens tide gage	
15	Bristol pressure tide gage	

4

والمحالية والمحالية والمحالية

CONTENTS

FIGURES--Continued

	Page
16	Fischer-Porter tide gage
17	Metercraft pressure tide gage
18	Stilling well
19	A model F420C anemometer used to measure wind speed and direction 23
20	Wind speed and direction gages
21	Microbarograph
22	NWS barometer
23	Hygrothermograph
24	Wood shelter which houses the hygrothermograph
25	Recording rain gage
26	Six-inch plastic rain gage
27	Sling psychrometer
28	Mechanical pyranograph
29	Example of the CERC monthly wave data summary for significant wave heights and periods
30	Example of the CERC monthly wave data summary for distribution of
	significant wave height versus period
31	Example of graph showing significant wave height and period versus time
32	Example of spectrum plot
33	Example of NOS tide summary
34	Example of meteorological observations recorded on CERC Form 78-236 38
35	Example of daily meteorological observation summaries

i T

5

ं के लिये

CONVERSION FACTORS, U.S. CUSTOMARY TO METRIC (SI) UNITS OF MEASUREMENT

Multiply	by	To obtain
inches	25.4	millimeters
	2.54	centimeters
square inches	6.452	square centimeters
cubic inches	16.39	cubic centimeters
feet	30.48	centimeters
	0.3048	meters
square feet	0.0929	square meters
cubic feet	0.0283	cubic meters
yards	0.9144	meters
square yards	0.836	square meters
cubic yards	0.7646	cubic meters
miles	1.6093	kilometers
square miles	259.0	hectares
knots	1.852	kilometers per hour
acres	0.4047	hectares
foot-pounds	1.3558	newton meters
millibars	1.0197×10^{-3}	kilograms per square centimeter
ounces	28.35	grams
pounds	453.6	grams
Poundo	0.4536	kilograms
ton, long	1.0160	metric tons
ton, short	0.9072	metric tons
degrees (angle)	0.01745	radians
Fahrenheit degrees	5/9	Celsius degrees or Kelvins ¹

U.S. customary units of measurement used in this report can be converted to metric (SI) units as follows:

¹To obtain Celsius (C) temperature readings from Fahrenheit (F) readings, use formula: C = (5/9) (F -32).

To obtain Kelvin (K) readings, use formula: K = (5/9) (F - 32) + 273.15.

INSTRUMENTATION AT CERC'S FIELD RESEARCH FACILITY, DUCK, NORTH CAROLINA

by H. Carl Miller

I. INTRODUCTION

The Coastal Engineering Research Center's (CERC) Field Research Facility (FRF) is located on the Outer Banks, North Carolina, near the village of Duck (Fig. 1). The FRF provides a means for obtaining high-quality field data, particularly during storms, in support of the U.S. Army Corps of Engineers coastal research missions. The FRF consists of an 1,840 foot-long concrete research pier supported on 3-foot-diameter steel piles. The pier deck is 20 feet wide, 25 feet above mean sea level (MSL), and extends from behind the dunes to approximately the 25-foot depth contour. In addition, the main building contains offices, a technical library, an instrument repair shop, and a data acquisition room.

A Basic Environmental Measurements (BEM) program has been established to collect, analyze, and disseminate data on selected oceanographic and meteorological conditions at the FRF. Weekly bottom profiles along both sides of the pier and periodic bathymetric surveys (from behind the dune to 1 or 2 miles from shore) along ranges as far as 2.5 miles north and south of the pier are also performed as part of the BEM.

This report describes the instrumentation maintained at the FRF for the BEM study and the types of data that are obtained as part of that program. It is intended as a reference report to provide users with information necessary for an understanding of how and where the data are collected. Subsequent reports will summarize the data collected by the instruments described herein.

Section II describes the instruments and their locations. Section III describes the data sampling procedure and the format of each type of data summary; Section IV discusses how to request data from CERC. An Appendix contains installation summaries for each instrument described in this report.

II. INSTRUMENTATION AT THE FRF

CERC maintains most of the oceanographic and meteorologic instrumentation in the vicinity of the FRF. However, tide gages have been installed by the National Ocean Survey (NOS), National Oceanic and Atmospheric Administration (NOAA) to monitor water levels in the ocean along the pier and in the sound adjacent to the FRF. The NOS Tides Branch at Rockville, Maryland, prepares monthly water level summaries from selected gages. In addition, the National Weather Service (NWS), NOAA, maintains wind measuring anemometers at the facility.

1. Oceanographic Instruments.

and the second state of the second of the second state of the seco

The location of each oceanographic instrument at the FRF (described below) is shown in Figure 2; station numbers in the figure correspond to the distance (in feet) from a reference base line to the installation. The shore end of the pier is at station number 1+20.




Figure 1. FRF location map.

to and

10-2

٠


.

8

in aits

N 12 Mar

,



.

r.

.

í

ji .

;*

,

a. <u>Staff Wave Gages</u>. Three staff gages are used to collect wave data as part of the BEM. Two gages are located at stations 6+20 and 19+00; the third gage is located at Jennette's fishing pier (about 25 miles south of the FRF), Nags Head, North Carolina.

The wave staffs are parallel inductive cable types manufactured by the Baylor Company, Houston, Texas (Fig. 3). Each wave staff consists of two wire ropes 0.5 inch in diameter held parallel under tension 9 inches apart. Associated with each staff is a transducer element which yields an electrically linear (direct current) output that is proportional to the amount of cable above a short circuit on the staff caused by the conducting characteristics of seawater. The gage is designed for an accuracy and resolution of 1 and 0.1 percent full scale, respectively. (For additional details on the wave staffs, refer to Baylor Company, 1970.) The gage installations at stations 6+20 and 19+00 use a hanger bracket support at the top and an anchor system on the bottom which are centered under the pier deck midway between adjacent pilings (Fig. 3). The Nags Head staff has rigid top and bottom supports connected directly to a wooden piling on the north side of the pier.

The water depth at each gage fluctuates as a function of both the water level and bottom elevation. Figure 4 shows the maximum and minimum bottom elevations observed along the FRF pier between 27 July 1977 and 4 January 1979.

b. <u>Wave Measuring Buoys</u>. Two Datawell Waverider buoy wave sensors are deployed at the FRF (Fig. 5). One sensor is located near the seaward end of the pier in about 23 feet of water; the other sensor is 1.3 miles due east of the end of the pier in about 60 feet of water (Fig. 2). The Waverider is a buoy which follows the movement of the water surface with internal electronics that measure the vertical acceleration of the buoy. The signal is integrated twice to produce a surface displacement signal which is telemetered to a shore receiving station.

The mooring used to fix the buoy in position is an integral part of the wave measuring system, designed to allow the buoy to move freely with the sea surface but strong enough to withstand the large forces caused by storm waves. Figure 6 shows the nearshore and offshore Waverider mooring systems at the FRF. The mooring components and their dimensions are a function of the water depth where the buoys are deployed. The offshore mooring system (deployed in about 60 feet of water) consists of a black rubber stretch cord, a buoyant polypropylene white rope, and a nylon-covered, stainless-steel rope covered with an air compressor hose cut into 8-inch lengths (foreground of photo in Fig. 6). The air compressor hose protects the rope from abrasion as it is dragged back and forth on the ocean bottom. The nearshore mooring system consists only of the rubber cord for deployment in depths of 25 feet or less (background of photo in Fig. 6). Additional details on the buoy operation and mooring configuration are in Van Breugel, Verhagen, and Gerritzen (1972).

c. <u>Pressure Wave Gage</u>. A single pressure gage is installed under the seaward end of the FRF pier. The gage sensor consists of a transducer (manufactured by I.C. Transducers, Inc., San Jose, California), coupled to a Bellofram gage protector and enclosed in a PVC housing (Fig. 7). With this arrangement, the pressure is sensed by the diaphragm in the gage protector and transmitted by an oil-filled cavity to the transducer. (Additional details on the transducer may be obtained from the manufacturer.) The gage is located at station 19+40 in approximately 27 feet of water (Fig. 2). The pressure sensor is attached to a special T-shaped bracket anchored to the bottom (Fig. 8).






Figure 3. Baylor wave staff gage and support system.








Figure 5. Waverider buoys.


With an Allin


.... Ò

ł

Figure 6. Waverider moorings.

ALL AND ALL AND A





and the second second

15

1 (**1**97)


10 . **19** - 124

d. Current Meters. Two Marsh-McBirney (MMI) 551 electromagnetic current meters are deployed at the FRF, one at station 19+40 under the seaward end of the pier and the other at approximately the same distance from shore but 500 feet north of the pier (Fig. 2). The meter under the pier is attached to the T-bracket shown in Figure 8; the other meter installation is shown in Figure 9. The MMI current meters are solid-state instruments with no moving parts. The meter's operation is based on Faraday's principle of electromagnetic induction, i.e., a conductor moving through a magnetic field will produce a voltage proportional to the speed of the conductor. The probe for the meter consists of a 4-inch-diameter sphere which contains an electromagnet and two pairs of electrodes. The two orthogonal electrode pairs sense the vector component voltages induced when the water moves through the magnetic field produced by the electromagnet. The instrument range is ± 10 feet per second with an accuracy of ± 2 percent or ± 0.07 foot per second and a 0.2-second time constant; zero drift is less than 0.07 foot per second. Additional details on the MMI 551 meters are in Marsh-McBirney, Inc. (1978).


e. Conductivity and Temperature Meters. Two Hydrolab Model TC-2 conductivity probes (Fig. 10) and two Hydrolab Model T-4 temperature probes (Fig. 11) are located under the seaward end of the pier at station 1960. One conductivity and one temperature probe is 10 feet below MSL; the other pair is 22 feet below MSL. The TC-2 determines the total concentration of ions in solution using the four-electrode method of conductivity measurement. The method consists of separating the electrodes that supply the current in the conductivity cell from the electrodes that measure the voltage produced in the solution by passage of the current through the cell. This technique eliminates errors due to cable resistance and electrode polarization, and diminishes the effects of electrode fouling (Garner, 1972). Measurements of conductivity also vary with temperature; the TC-2 instruments are electronically corrected to 25° Celsius for temperature variations. Water temperature is measured by the TC-2 using a precision thermistor temperature sensor contained in the conductivity probe. The conductivity output data signal is stated to be accurate to within 2.5 percent of the reading and the temperature readings to within $\pm 0.5^{\circ}$ Celsius. Mounted a few inches above the TC-2 probes are T-4 Marine Hydrographic thermometers which have a range from -2° to 49° Celsius. The conductivity and temperature values are obtained by reading Hydrolab meters located in a stainlesssteel box on the pier deck (Fig. 12). (Additional details on the conductivity and temperature probes may be obtained from the Hydrolab Corporation, Austin, Texas.)

f. <u>Tide Gages</u>. Three NOS tide stations consisting of four gages are installed at the FRF. The NOS control station (Fig. 13) at the seaward end of the research pier (Fig. 2) consists of a Leupold-Stevens gage (Fig. 14; manufactured by Leupold and Stevens, Inc., Beaverton, Oregon) and a Bristol pressure gage (Fig 15; manufactured by the Bristol Company, Waterbury, Connecticut). The second tide station (located at station 7+20) has a Fischer-Porter tide gage (Fig. 16; manufactured by Fischer and Porter Company, Warminster, Pennsylvania). The third station (located about 1,000 feet west of the shore in the Currituck Sound) has a Metercraft pressure tide gage (Fig. 17; manufactured by Metercraft Corporation, Perry Hall, Maryland).

Both the Leupold-Stevens and Fischer-Porter analog-to-digital recorders (ADR) are float-activated, negator-spring, counterpoised instruments that mechanically convert the vertical motion of a float into a coded, punched



• ,




بعائد المراجة

and states

200

- 20



÷

Figure 10. Hydrolab Model TC-2 conductivity probe.



Figure 11. Hydrolab Model T-4 temperature probe.




Figure 12. Hydrolab temperature and conductivity meters.





Figure 13. NOS control station at end of pier.



;

İ

٢

91




Figure 14. Leupold-Stevens tide gage.




Figure 16. Fischer-Porter tide gage.




Figure 17. Metercraft pressure tide gage.

paper-tape record. The below-deck installations at stations 19+60 and 7+20 consist of 12-inch-diameter stilling wells with a 1-inch orifice and 0.5-inch orifice, respectively, and an 8.5-inch float (Fig. 18). The stilling well acts as a high-frequency filter to dampen the effects of waves, thereby giving accuracy to the measurements of water level fluctuations produced by the tide. The float, which responds to any variation in the water column within the stilling well, is coupled to a float-wire takeup drum and input shaft assembly via a stainless-steel wire. The angular position of the input shaft is recorded on paper tape every 6 minutes in a standard binary-decimal code.




Figure 18. Stilling well.

The Bristol and Metercraft pressure gages measure the hydrostatic pressure created by a column of water above a submerged orifice. The orifice and the recording instrument are linked by a nitrogen-filled, plastic supply line. Any change in the hydrostatic pressure, such as that caused by the rise and fall of the tide, is transmitted to the recorder where the pressure variations are recorded on a strip chart as changes in water depth.

To monitor tide gage operation and provide datum information, electric tape gages have been installed at stations 19+60 and 7+20 in lieu of a tide staff. The operating principle of the tape gage is based on the electrical conductivity of seawater. The gage consists of a stainless-steel tape on a takeup reel, 6inch-diameter stilling well, voltmeter, and battery. The weighted tape, graduated to hundredths of a foot, completes an electrical circuit upon contact with the seawater inside the metal stilling well. The distance to the water level at that instant is measured below a reading mark on the tape. The reading mark has a known elevation relative to tidal bench marks and the National Geodetic Vertical Control Network. The electric tape gage readings are compared with the automatically recorded tide record. Additional details on each gage may be obtained from the respective manufacturer.

2. Meteorological Instrumentation.

To monitor the meteorological conditions at the FRF, various equipment has been installed by the NWS and CERC. Figure 2 shows the location of the meteorological instruments. Although CERC is responsible for collecting the data, NWS provides guidance on installations and maintains the anemometer. In addition to any permanent records obtained from the devices with recording capabilities, each instrument is read daily (at approximately 0700 hours). Additional details on the instruments described below may be obtained from the respective manufacturer or from NWS.

a. <u>Anemometer</u>. A model F420C anemometer is used to measure wind speed and direction (Fig. 19). A cup rotor and a spread-tail wind vane were installed on a 21-foot-high tower about 65 feet from a temporary office trailer. The anemometer is coupled to wind speed and direction gages inside the trailer to monitor the onshore wind conditions (Fig. 20). The gages are manufactured by the Electric Speed Indicator Company, Cleveland, Ohio.




Figure 19. A model F420C anemometer used to measure wind speed and direction.




Figure 20. Wind speed and direction gages.

The accuracy of the speed transmitter and indicator assemblies is ± 1 knot from 2- to 100-knot winds and ± 2 knots for 100- to 200-knot winds at 20° to 25° Celsius. The wind-direction transmitter and indicator assemblies are accurate to $\pm 5^{\circ}$ at an airspeed of 5 knots or greater for any horizontal direction at 20° to 25° Celsius.

The wind-direction transmitter assembly includes a direction transmitter body and a spread-tail wind vane. The transmitter body consists of a 12-volt (direct current) synchro transmitter coupled to the wind-vane shaft. The moment of inertia of the vane is approximately 583,000 gram-centimeters squared. The synchro transmitting element gives a minimum resistance to the electrical current when the vane is directed to due north and increases the resistance as the vane turns away from north. Two potentiometers are used to determine which direction the vane has turned with respect to north.

The windspeed transmitter includes a windspeed transmitter body and a cup rotor. A direct-current magneto is coupled to the cup rotor and provides a voltage-versus revolution-per-minute ratio of 0.004007 with an electrical load of 428.6 ohms. The terminal resistance of the magneto is 40 ohms ±1 ohm; the moment of inertia of the cup rotor and magneto armature is 70,700 gram-centimeters squared (maximum). National Weather Service (1979) provides additional details on the wind measuring system. In late 1979, an analogue recorder was installed to provide continuous records of the wind speed and direction.

b. <u>Microbarograph</u>. This instrument, manufactured by the Belfort Instrument Company, Baltimore, Maryland, is located inside the building and is used to measure the atmospheric pressure and pressure tendency (Fig. 21). The device is an aneroid sensor which responds to pressure changes on the order of 0.005 inches of mercury. The pressure tendency can be determined from the chart record; i.e., the total change in pressure over a 3-hour period and the direction of change (increasing, decreasing, or some combination).



Figure 21. Microbarograph.

c. <u>Barometer</u>. A NWS barometer is also used for monitoring atmospheric pressure (Fig. 22). This device is an aneroid barometer with a 2-inch-diameter bimetallic wafer which expands or contracts with changes in pressure. The pressure (in inches) is read directly off the gage. Additional details on the barometer can be provided by the Basic Observations Branch, NWS.

d. <u>Hygrothermograph</u>. The hygrothermograph (Fig. 23) is located in an outdoor, ventilated instrument shelter (Fig. 24). This instrument, also manufactured by the Belfort Instrument Company, is used to measure and record both air temperature and relative humidity. The temperature sensor consists of a bimetal assembly made from two dissimilar metals which react to temperature changes at different rates. The recording styles (or pens) are driven by the expansion or contraction of the assembly. The relative humidity is sensed by



Figure 22. NWS barometer.




Figure 23. Hygrothermograph.

William Contraction



Figure 24. Wood shelter which houses the hygrothermograph.

a banjo-spread human hair element which expands when the humidity increases and contracts when it decreases, causing the second recording pen to deflect approiately. In addition to the daily observations, a 7-day continuous chart record is produced.

e. <u>Maximum and Minimum Thermometers</u>. The NWS thermometers (housed in the instrument shelter) are used to determine the daily extreme temperatures (Fig. 24). A constriction in the capillary of the maximum thermometer prevents the mercury from flowing back into the bulb. The mercury is forced past the constriction as the temperature rises but does not return through it when the temperature falls, causing a thread of mercury to remain in the tube at the highest temperature reached. The thermometers are held in a special mounting device called a Townsend Support (see Fig. 23) which allows a reset of the thermometers without removal. The high temperature thermometer is reset by spinning it in the Townsend Support, thereby causing enough centrifugal force to force the mercury back through the constriction.

The minimum thermometer has a small, dark glass index in the thermometer bore which is pulled toward the bulb by the surface tension of the mercury meniscus as the temperature drops and remains stationary when the temperature rises. The thermometer is mounted horizontally except when the instrument is reset. Resetting is facilitated by the Townsend Support and consists of raising the bulb of the thermometer, thus causing the index to slide downward (toward higher temperatures) in the tube. The thermometers are read and reset once a day. Additional details on the thermometers can be provided by the Basic Observations Branch, NWS. f. <u>Twelve-Inch Weighing Rain Gage</u>. This gage (manufactured by the Belfort Instrument Company, Baltimore, Maryland) is used to measure and record the daily amount of precipitation (Fig. 25), and is located near the instrument shelter (Fig. 2). The gage sensor consists of a l-foot-high collection bucket (8 inches in diameter), a weighing mechanism connected to the recorder pen, and a chart drive for continuously recording the amount of precipitation. The weighing device is a spring scale connected through a lever mechanism to a pen that moves across the chart. The rain capacity is expressed in inches but is measured in terms of weight; i.e., 1 inch is 29.02 ounces at 62.56° Celsius. The manufacturer's specifications indicate that the instrument accuracy is 0 to 6 inches ± 0.5 percent (± 0.03 inch of precipitation) and 6 to 12 inches ± 1.0 percent (± 0.06 inch of precipitation). Daily observations of the total precipitation are recorded and the chart records are retained for future reference.



Figure. 25. Recording rain gage.

g. <u>Six-Inch Plastic Rain Gage</u>. The 6-inch rain gage (a True Check Rain Gage manufactured by the Edwards Manufacturing Company, Alberta Tea, Minnesota) (Fig. 26) is located about 25 feet from the 12-inch weighing rain gage (Fig. 2). The gage is simply a catch with a calibrated scale and is used as a check for the 12-inch weighing rain gage. Daily totals are recorded and the gage is manually (emptied) reset.



Figure 26. Six-inch plastic rain gage.

h. <u>Sling Psychrometer</u>. A sling psychrometer is used to measure relative humidity and to determine the dew point (Fig. 27). The psychrometer has two thermometers mounted in a frame which can be rotated rapidly. A moistened muslin wick is attached to the bulb of one of the thermometers and the device is whirled to ventilate both thermometers. The temperature is read from both the "wet" and the "dry" bulb thermometers and a set of tables is used to obtain the relative humidity and dew point values from the different temperatures. The dew point is recorded daily along with the other meteorological observations. Additional details on the sling psychrometer can be provided by the Basic Observations Branch, NWS.

i. <u>Mechanical Pyranograph</u>. This device, located on top of the weather instrument shelter and made by the Weather Measure Corporation, Sacramento, California, provides a continuous record of the intensity of the sun and sky radiation (Fig. 28). Two black and two white bimetallic strips are coupled together with the black strips attached through a linkage to the recording pen. As the radiation heats the black strips, the pen deflects. The white strips provide compensation for ambient temperature changes. The area under the curve on the chart record is a measure of the total radiation. The chart records are collected and stored without analysis.



Figure 27. Sling psychrometer.



Figure 28. Mechanical pyranograph.

III. SAMPLING INFORMATION AND SUMMARY DESCRIPTIONS

1. Digital Wave Data.

Routinely, four digital records per day at times near 0100, 0700, 1300, and 1900 e.s.t. are obtained from the CERC wave gages. Each record is 20 minutes long with a sample frequency of four data points per second. Thompson (1977) describes the method used to analyze the digital data records. Monthly summaries of significant wave height and peak spectral period are available for dissemination.

For the pressure-wave gage, the surface pressure record is attenuated with depth. In order for the pressure record sensed by the gage to reflect surface conditions, linear wave theory is used to compensate for the hydrodynamic attenuation; Esteva and Harris (1970) provide a discussion of the compensation technique used by CERC. These data are sampled at 4 hertz for a 20-minute period as are the other wave sensors.

The FRF wave data are summarized in the standard CERC monthly wave data summary format which consists of (a) a listing of the significant wave height and peak spectral wave period by day and time (Fig. 29), (b) a joint distribution of significant wave height versus peak spectral period (Fig. 30), and (c) a graph of the significant wave height and peak spectral period versus time (Fig. 31). When requested, the spectral energy distribution as a function of wave frequency can be printed or plotted (Fig. 32).

2. Current Meter Data.

The current meter data are obtained as x and y components of velocity four times a day for 20 minutes at a frequency of four data points per second. The statistical package is designed to treat the data exactly like wave height information. Consequently, for each current component the mean of the data record, the standard deviation, and the spectral energy distribution are computed. The only summarization is in the form of Figure 29 where a listing of the mean current, standard deviation, and peak spectral period as a function of the day and time is generated.

3. Conductivity and Temperature Data.

The conductivity and temperature gages are read once daily and the data are recorded as notes on the CERC Littoral Environmental Observation (LEO) forms for the pier end LEO site. The LEO visually observed data are summarized by computer after the data have been punched on cards. An analog recorder (installed in April 1980) is used for continuously recording the water conductivity and temperature.

4. Monthly Tide Summaries.

The paper tapes and strip-chart records are sent to the NOS, Rockville, Maryland, where monthly tabulations of hourly tide heights and high and low water level summaries are prepared. Copies of the tabulated summaries are forwarded to CERC. Tidal datums at the FRF are computed by the NOS ortice. An example of the NOS hourly and high and low water summaries for 1 month of data is shown in Figure 33. 1

ميريد من من جي الاجراب في

	625 FRF	DUC#	Ş,	JAN 1978	516~11	SIGNIFICANT WAVE MEIGHTS AND PERILUDS	TELGHTS AND	PERTUDS	16- 625			
DAV/TIME		0	0100		0100	_		1300			1900	
	H(C#)	H(FT)	T(SEC)	(x))¥	H(FT)	T(SEC)	1 (CH)	H(F1)	T (SEC)	#(C#)	H(FT)	7(56C)
-	170.69	5.60	8.79		5.16	6.79	162.15	5.12	10.84	141.42	4 - 4	
~	52.10	3.00	10.84	76.01	2.52	10.64	95.10	3.12	10.64	80.47	34.2	
•	67.96	2.20	12.26	62.18	2.00	10.64	70.01	2.52				
9	e0.9e	2.10	5.01	95.10	3.12	5.03	•			40.04		
r	42.67	1.40	5.29	30.48	1.00	10.84						
. 1				1						69°57	1.44	6.34
~ (52.63	1.72	6.35	40.23	1.52	5.61	47.55	1.56	6.85	84.12	2.74	
60 (95.10	3.12	8°13	108.51	3.56	6. 03	124.36	4.0 6	9.71	151.18	10.1	12.26
•	178.00	5.84	A.79	220.08	7.24	12.20	114.60	3.70	10.04	92.26	5.04	14.12
2	10.58	2.12	12.20	76.05	2.56	12.20				41.55	1.56	12.26
-	20°67	2.64	4.75	151.19	49.3	6.3 ⁹				94.28	2.24	24.2
2	, (() , ()	74.	14.12	40.23	1.32	14.12	37.80	1.24	14.12	48.77	1.00	12.26
5	00°07	1.14	12.20	80 . u 7	2.04	3.95	121.92	4.00	8.03	190.29		6.71
	30.00	0~-0	12.26	199.95	6. 56	12.20	152.40	5.00	12.20	143.07	4.72	10.64
<u> </u>	72.61	3.12	10.84	129.24	4.24	10.64	151.18	96 7	10.84	117.04	3.04	12.20
<u>.</u>		20.2	12.26	75.59	5.48	10.84	17.71	2.32	12.20	71.93	2.36	12.26
1	59.55	1.02	10.84	50.00	2.12	12.20	91.44	3.00	4.75	210.42	50.0	0.70
	147.51	6.3B	9.71	135.33	77 . 7	10.84	110.95	3.64	10.84	97.54	5.20	9.71
	153.62	1 t	u.75	106.51	5.50	5.01				185.32	.0.0	5.01
2	264.57	8.58	6 ° 4 5	207.20	6. 60	e.70	170.69	5.60	9.71	129.24	1.24	10.04
2	132.99	9	9.71	74.57	2.44	9.71	74.37	2.44	10.84	01.00	2.23	14.12
22	73.15	2.40	12.20	119.48	3.92	12.26	121.92	4.00	12.20	90.22		
23	75.54	2.48	12.26	76.05	65.5	12.20	93.88	3.00	12.26	74.47	2.44	
3	65.ñu	2.10	10.64	65.8 4	2.10	10.84				52.43	1.72	9.71
2	53.04	1.70	12.0	57.50	1.68	9.71	85.34	2.80	9.71	110.95	2.00	6.79
4	153.62	5.06	9.71	200.04	0.10	9.71				84°70	\$ 0 ¥	10.94
		2.12	10.84	90.09	2.40	9.71	60.61	1.64	9.71	39.01	1.28	9.71
	3 - 3 S		10.54	40.23	1.32	16.05	00°07	1.64	7.39	62.19	2.0u	10.64
~		2.2	10.5				67.06	2.20	5.61	07.09	2.24	10.5
			5.29	78.03	2.50	5.97	40.22	2.90	5.47	76.03	4.50	2.47
15		~	6.38	00°40	2 .00	5.01	74.37	2.44	5.61	67.0 6	2.20	10.0
COL AVG	101.52	3.13	9.30	100.10	3.28	9.73	97.10	3.19	9.37	94.73	11.5	9.64
AVERAGE	AVERAGE SIGNIFICANT MAV	17 nave	METGHT .	3.23 FEET	AVENAGE	WAVE PERIUD	# 9.52 StC	U	112 URSERVATIONS	0 N S		
	70		a 00		00000		•					

Example of the CERC monthly wave data summary for significant wave heights and periods. Figure 29.

A Main

ł


MAVE CLIMATE FUM · 625 FAF DUCK, NC Distribution of Significant meight vs periou (in Ubservations per 1000 UBS) For 112 Observations of mon-calm conditiuns- o unservations of calm conditions mave not been included SIGNIFICANT REIGHT C01434


معفظهم ويستحد


(SECS)						110101 10101 101010		HE 16H7									
		2.1	1.0	1-4	5	5. b	1.4	7		4-10	10-11	4-9 4-10 10-11 11-12 12-13 13 4 10-14 10-40	11-21	•	TUTAL	ACC.	110 20 21
•. • •••	•		1)			•		:		•			00.00
1.0 - 1.9																	0.00
2.0 - 2.9																	00.0
3.0 - 3.9			•												•	o	9 2.50
4.0 - 4.9			٩٢	•		•									36	20	3.17
5.0 - 5.9		9 1	4 8				σ								103	202	205 2.75
0.0 = 0.0		27	•		•				•						29	208	3.21
7.0 - 7.9		•													0	277	1.50
8.0 - 6.9				27	•	27	. 4 1								ÛR	357	357 4.94
9.0 - 9.9		¢.0	ې ۲	•	27	61	52								101	518	3.61
10.0 -10.9		27	12	54	\$	σ									205	123	3.20
11.0 -11.9																725	0.00
12.0 -12.4		51	116	27	18	¢	o	•	•						223	986	3.30
13.0 -13.9																940	0.00
0.11- 0.11		5	•	•											45	199	2.10
15.0 -15.9																140	06.0
16.0 -16.9		•													œ	1000	1.50
17.0 -17.9																1000	00.0
10.0 -18.9																1000	00.0
19.0 -19.9																1000	0.00
20.0 -20.9																1000	0.00
Z1.0 + TOTAL		. 4.7	191			;	2	c									00.0
ACC. TOTAL		. 9		257			0 0 0	10		0001	0001	0.00					
COL. 4V6.	J.00 10.21	0.61		9.74	10.04		9.07	9.07 12.50	05.4	0.00	00.0	0.00 0.00	0.00	00.0	9.48	••••	
AVERAGE SIGWIFICANT AAVE MEIGMT = Variance of Significant Aave meight = Siandard deviation of Sig. Meight = Siandard devis i Links, Calm	IGWIFICANT AAVE MEIGMT = 3,23 FT AVEAGE FAVE PEDIOD = 9,52 SEC OF SIGNIFICANT AAVE MEIGMT = 2,71 FT SU VANIAYCE NÊ AAKE PENIOD = 9,12 SEC SU DEVIATION OF SIG, MEIGMT = 1,000 FT SIANDAHO NEVIATIOY OF PENIOD = 2,65 SEC (FUM TME ABOVE 3 LINES, CALM IS COUNTED AS 0,0 FEET FOF MAVE MEIGMT AND IS GMITTED FOR MAVE PENIOD)	AVE 76 NT 7 A1 NT 7 A1		5 н 1 н 1 н 1 н 1 н 1 н 1 н 1 н	3.23 F1 2.71 F1 S4 1.04 F1 15 COUNTED A	rt Ft Su VTED ≜	4 > 0	AVERGE FAVE PERIOD = 9,52 SEC Vaniate of mate peniod = 8,12 SEC SQ Statioand reviation of Peniod # 2,65 Sec Feet for mate meight and is cmitted for may	5 15 15 15 15 15 15 15 15 15 15 15 15 15	2 E A 100 4 E A 100 4 I 10 4 100 4 I 10 4 100	H 100 0	52 SEC 107 #	8.12 2.65 17760	566 566 567 560 560 560 560 560 560 560 560 560 560	ער ה ה ה	[00]	

Example of the CERC monthly wave data summary for distribution of significant wave height versus period. Figure 30.

RECOMD OGTAINED "ITH & BAVLOW STAFF WAVE GAGE LOCATED AT STATION 19400 (FRF pier). Data OBTAINED AT AVALYSIS OF A IN24-SECOVU STUITAL MECOND.









.

EAS FFEET) SEP 1974 TH 75N SEP 1974 TH 75N SEP 1974 TH 75N SEP 1974 TH 75N THE GHT THE GHT THE GHT THE GHT THE SECTION THE SE	DW WATER'S [FEET] SEP 1974 71 17 7.3 2.5 10 19 4104 TWE LGM H104 19 40 11 22501 20 20.5 14.4 2228 21 20.1 10.4 2228 22 11 20.1 10.4 2228 23 12.1 10.4 2228 24 0.4 12 22.1 25 11.2 20.0 17.1 20.0 19.2 20.0 10.2 20.0 10.0 0.10 0.10 0.10 0.0 10.0 0.0 10.	IGH AND LOW WATERS (FEET) SEP 1974 T) LEN DAY HIGH LGW HIGH LEN DAY LT 22.22 22.21 LEN 20.9 21.9 21.9 22.25 LEN 20.9 21.1 22.21 21.00 LEN 20.1 20.1 22.21 21.00 LEN 20.1 20.1 21.1 22.21 LEN 20.1 21.1 21.00 21.00 LEN 20.1 21.1 21.00 21.00 LEN 20.1 21.1 21.00 21.00 LEN 21.7 21.0 21.00 21.00 LEN 21.1 21.1 21.00 21.00 LEN 21.2 21.1 21.1 22.21 LEN 21.7 21.0 21.00 21.00 LEN 21.7 21.0 21.00<	TJ0855 MiGH AND LOW MATCHS FFE 13 SEP 1974 T1 HEIG-T DAY MIGH THE LCW HEIM 21,32 18,45 11 7,9 22,13 18,45 18,45 22,73 22,13 18,45 21 10,4 22,13 18,45 21 20,5 22,13 18,45 21 20,1 22,13 18,45 21 20,1 22,13 18,45 21 21,2 22,13 18,45 21 20,1 22,13 18,45 21 21,3 22,13 18,45 21 21,3 22,13 18,45 21 21,3 22,13 18,45 21,3 21,3 22,13 18,45 21,3 21,3 22,13 18,45 21,3 21,3 22,13 18,45 21,3 21,3 22,13 18,45 21,3 21,3 22,13 18,45 21,3 21,3 22,13 18,45 21,3 21,3 22,13	J'SIGE' SEP 1974 TIME SEP 1974 TIME HEIGH LEN DAY HIGH TIME HEIGH HIGH HIGH HIGH HIGH HIGH HIGH HI	TIDE-, ÞCLALY HEIGHTS (FEET) 8651370 duck frf ng (dutside)	•	1 2 3 4 5 6 7 8 9 10		0 18,49 19.21 19,40 19.78 20.30 20,68 21,25 21,21 21,54 22,09	15.01		:	3 19,39 19.73 19.44 19.62 10.87 10.87 16.88 19.21 20.31 21.07	4 20,27 20.56 20,05 19,44 ,5.16 18,95 18,62 18,79 19,50 20,44	5 20,99 21,23 20,96 20,15 15,77 19,41 18,77 18,71 19,11 19,78	6 21,29 21.95 21,70 21.1P 20:67 20.03 19,34 18.93 19.07 19.43	7 21,24 22,28 22,27 22,64 21,40 21,03 20,01 19,64 19,27 19,39	21,17 22,11 22,30 22,34 ,2,18 21,88 20,84 20,36 19,72 19,59	9 20,40 21.62 21,98 22.34 22.32 22.22 21.78 21.33 20.59 20.1b	10 19.63 20.88 21.37 21.77 22.C5 22.19 22.08 22.16 21.66 20.83	11 19,14 20,00 20,61 21,22 21,55 21,48 21,92 22,38 22,32 21,78	12 10.62 19.33 19.72 20.35 20.65 21.22 21.69 22.31 22.49 22.3C	13 16.62 19.03 19.16 19.53 15.67 20.35 20.79 21.96 22.54 22.54	19,08 19,14 15.24 19,58 20,06 21,30 22	19,39 19.26 18:54 19.02 19.21 20.53 21.45 22	Z0.74 20.38 20.05 19.67 19.19 18.99 18.69 19.91 20.67 21	20,83 20,24 15.59 19.25 18,79 19,53 20,10 20	22.05 22.04 21.71 21.19 20.24 19.72 19.02 19.38 19.82 20	19 22,19 22,34 22,38 21,98 21,26 20,38 19,49 19,81 19,68 19,57	20 21,91 22,28 22,41 22,41 21,83 21,27 20,22 20,36 19,53 19,47	21 21,35 21.41 22,14 22,28 22,12 21,63 20,95 20,99 20,53 14,74	22 20,60 21.05 21.53 21.88 22.C3 21.96 21.38 21.68 21.23 20.32
$\mathbf{F} = \mathbf{T} \mathbf{N} \mathbf{N} \mathbf{N} \mathbf{N} \mathbf{N} \mathbf{N} \mathbf{N} N$	0.4 HICH TIME 5.6 197, 10 HICH TIME 5.6 197, 11 19, 20, 10, 10, 12 20, 10, 10, 10, 13 20, 10, 10, 20, 13 20, 10, 10, 20, 21 20, 10, 10, 20, 23 10, 10, 10, 20, 24 0, 10, 10, 20, 25 10, 10, 20, 20, 24 0, 10, 20, 20, 25 10, 10, 20, 20, 26 10, 20, 20, 20, 25 10, 20, 20, 20, 25 10, 20, 20, 20, 27 10, 10, 20, 20, 25 10, 20, 20, 20, 26, 10, 20, 20, 20, 28, 10, 20, 20, 20, 26, 10, 20, 20, 20, 26, 10, 20, 20, <td< td=""><td>16. A.D. LOW WATERS (FEET) 5.F 197. 6.7 Day HICH LCH H 6.7 Day HICH LCH H 18.37 17 7.3 1.2 22.2 18.97 17 19.3 1.2 22.2 18.97 17 7.3 1.2 22.2 18.95 18 7.9 1.2 22.2 18.95 19 20.3 1.1 22.2 18.95 20 20.4 22.2 22.2 18.72 21 20.1 10.4 22.2 18.72 21 20.1 17.1 22.2 18.72 21 20.1 17.1 22.2 18.72 21 20.1 17.1 22.2 18.72 21 21.1 19.4 22.2 18.72 21.2 10.4 22.2 22.2 18.72 23.1 10.7 23.2 22.2 18.72 23.1 23.2 23.2 22.2 18.72 23.1 23.2 23.2 22.2 18.72 23.1 23.2 23.2 22.2 18.72 23.2 23.2 23.2 22</td><td>TILES HIGH AND LOW MATCHA THEFT SEP 1973 TILES HIGH AND LOW MIGH THE THE THE THE THE THE THE THE THE TH</td><td>MC (OUTSIDE) SEP 197 MC (MC (MT (MT (MT (MT (MT (MT (MT (MT (MT (MT</td><td>TA 75W</td><td></td><td></td><td>17.87 18.21</td><td>18.27</td><td>18,27 18,37</td><td>18,50</td><td>19.61</td><td></td><td>18.72</td><td>19,20</td><td></td><td>19.70</td><td>20.05</td><td>20.00</td><td></td><td>19455</td><td>19.20</td><td>I</td><td>14</td><td></td><td>G 0,15</td><td>11</td><td>2</td><td>19</td><td></td><td></td><td></td></td<>	16. A.D. LOW WATERS (FEET) 5.F 197. 6.7 Day HICH LCH H 6.7 Day HICH LCH H 18.37 17 7.3 1.2 22.2 18.97 17 19.3 1.2 22.2 18.97 17 7.3 1.2 22.2 18.95 18 7.9 1.2 22.2 18.95 19 20.3 1.1 22.2 18.95 20 20.4 22.2 22.2 18.72 21 20.1 10.4 22.2 18.72 21 20.1 17.1 22.2 18.72 21 20.1 17.1 22.2 18.72 21 20.1 17.1 22.2 18.72 21 21.1 19.4 22.2 18.72 21.2 10.4 22.2 22.2 18.72 23.1 10.7 23.2 22.2 18.72 23.1 23.2 23.2 22.2 18.72 23.1 23.2 23.2 22.2 18.72 23.1 23.2 23.2 22.2 18.72 23.2 23.2 23.2 22	TILES HIGH AND LOW MATCHA THEFT SEP 1973 TILES HIGH AND LOW MIGH THE THE THE THE THE THE THE THE THE TH	MC (OUTSIDE) SEP 197 MC (MC (MT	TA 75W			17.87 18.21	18.27	18,27 18,37	18,50	19.61		18.72	19,20		19.70	20.05	20.00		19455	19.20	I	14		G 0,15	11	2	19			
$ \begin{array}{cccccccccccccccccccccccccccccccccccc$	0.4 Hich The 1 1 1 1 1 1 1 1 2 2 2 2 2 2 2 1 2 2 2 1 3 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	16.1 0AY HITERS IFEE 6.7 0AY HITERS IFEE 18.37 17 19 1 18.95 13 17 1 18.95 13 17 1 18.95 13 17 1 18.95 13 17 1 18.95 13 13 1 18.95 13 20 1 18.95 21 20 1 18.95 21 20 1 18.95 21 20 1 18.95 21 20 1 18.95 23 12 1 18.95 23 12 1 18.95 25 12 1 18.95 26 23 1 18.95 26 1 1 18.95 26 1 1 18.95 1 1 1 18.95 1 1 1 18.95 1 1 1 18.95 1 1 1 18.95 1 1 1 18.95 1 1 1 18.95	TJUESS MIGE AND LOW MATERS THE IOUTSIDE: HEIG-T DAY HIGH THE H:GH LCh DAY HIGH HIGH 22129 19622 18 20.5 22139 19622 18 20.5 22139 19622 18 20.5 22139 19622 18 20.5 22139 19622 18 20.7 22145 19602 20.5 22145 18 65 20 10.1 22151 18 65 20 10.1 22151 18 65 20 10.1 22155 19612 20 10.1 22155 19612 20 10.1 22155 19612 20 10.1 22159 18 65 20 10.1 22151 18 25 20 10.1 22159 18 10.2 22159 18 10.1 22159 18 10.2 22159 18 10.2 2010 10.1 2010 10.1 2010 10.1 2010 10.1 2010 10.1 2010 10.1 2010 10.1 2010 20.1 2010 10.1 2010 20.1 2010 10.1 2010 20.1 2010 10.1 2010 20.1 2010 10.1 2010 20.1 2010 20.1	NC (OUTSIDE) NC		3	r	~			~ ~	4 ~	~	4 *** *	~ ~	~ `									20							•
		C C <td>T10255 H10F ADD L HE16-7 F59 16:72 22:23 19:05 22:29 19:09 22:29 19:09 22:29 19:09 22:29 19:09 22:29 19:09 22:29 19:09 22:29 19:09 22:29 19:09 22:25 19:00 22:25 19:00 20:00 2</td> <td>NC (0UTS10E) NC 222,39 NC<!--</td--><td>ERS LFEET</td><td>INE</td><td></td><td>.</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td>. •</td><td>0 F-</td><td></td><td>~ ~</td><td>r 01</td><td></td><td></td><td>4</td><td></td><td>36,6</td><td></td><td>00"00</td><td></td><td></td><td></td><td>DATUM 268</td><td></td></td>	T10255 H10F ADD L HE16-7 F59 16:72 22:23 19:05 22:29 19:09 22:29 19:09 22:29 19:09 22:29 19:09 22:29 19:09 22:29 19:09 22:29 19:09 22:29 19:09 22:25 19:00 22:25 19:00 20:00 2	NC (0UTS10E) NC 222,39 NC </td <td>ERS LFEET</td> <td>INE</td> <td></td> <td>.</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td>. •</td> <td>0 F-</td> <td></td> <td>~ ~</td> <td>r 01</td> <td></td> <td></td> <td>4</td> <td></td> <td>36,6</td> <td></td> <td>00"00</td> <td></td> <td></td> <td></td> <td>DATUM 268</td> <td></td>	ERS LFEET	INE		.								. •	0 F-		~ ~	r 01			4		36,6		00"00				DATUM 268	

A CARA

- 0

Example of NOS tide summary.--Continued Figure 33.

20,72 21,59 23,33 22,50 22,52 21,20 20,39 19,45 20,20 19,70 18,70 19,69 20,56 20,66 21,63 21,89 21,99 22,51 22,18 20,00 21.07 23,11 22.69 22.45 22.29 21.04 20.53 20.67 19.57 18.65 20.06 20.20 20.15 20.99 21.63 21.82 22.49 22.39 16.69 19.49 21.42 21.71 22.30 22.57 22.47 22.49 22.87 21.18 20.02 20.80 20.28 19.69 20.1C 20.22 20.63 21.46 21.60 18,41 18,92 20,26 20,55 21.39 22.07 22,52 22,54 23,40 22,05 20,73 21,51 20,64 19,83 20,04 15.89 20,02 20,78 20,89 19,34 19.25 19.09 18.7C 15.15 20.06 21.07 21.79 23.24 22.36 21.77 22.84 21.94 20.81 20.74 2C.C5 19.63 19.73 19.46 20.23 19.89 19.47 18.39 18.33 18.95 19.89 20.98 22.45 22.02 21.72 23.22 22.39 21.35 21.17 20.50 19.95 19.71 19.24 21.18 21.05 20.12 18.67 18.19 18.33 18.81 19.82 21.51 21.42 21.35 23.05 22.48 21.83 21.62 20.47 20.15 19.39 21.67 22.32 20:91 19.37 18.31 18.30 18.35 18.95 20.39 20.66 20.76 22.80 22.33 22.00 22.25 21.63 21.12 20.73 19.94 22,19 23,28 22,06 20,48 19.30 18,59 18,40 18,78 19,64 19.81 20,13 22,29 22,08 21,97 22,48 22,11 21,81 21,67 20,73 21.71 23,99 23,47 22.62 21.60 20.77 19.68 19.66 19.47 19.10 18,96 21.08 21.06 21.29 22.16 22.18 22.52 22.52 22.12 20,96 23,75 23,42 23,02 22,61 21,84 20,87 20,53 19,81 19,22 18,90 20,69 20,52 20,74 21,64 .1.84 22,09 22,56 22,50 20.88 20.70 21.05 21.82 22.34 23.01 22.14 21.08 20.07 21.02 20.16 19.69 20.05 15.54 20.28 20.80 21.13 0 Z0,26 19.32 Z0113 19.76 18.29 18.38 19.33 Z0,13 Z1,87 Z1.57 Z1.36 Z1.54 Z2.27 Z1.56 Z0.61 ZC.62 Z0.66 19.63 19.42 20,90 20,16 20,74 19.2, 16,19 17,97 18,52 19,01 20,67 20,87 20,91 21,13 22,30 21,67 21,36 21,15 20,50 20,18 19,55 21,35 21,05 21,74 19,57 18,71 18,11 18,47 20,12 20,05 20,23 21,06 22,11 21,77 21,83 21,17 21,19 21,00 20,21 21.47 21.60 22.65 21.16 15.73 18.88 18.72 18.35 19.77 19.37 19.54 20.57 21.73 21.60 21.75 22.13 21.73 21.74 20.90 21,19 21,79 23,24 22,1 20,59 19,87 19,27 18,67 19,83 19,11 18,85 20,07 21,12 21,55 21,55 22,12 22,02 22,20 21,04 19.21 20.25 22,30 22,53 22,70 22,70 22,47 21,69 21,56 20,34 19,36 20,22 20,13 19,82 20,44 20,86 21,28 22,03 22,22 18.73 18.95 19.45 19.51 26.25 21.22 21.99 22.36 23.50 22.48 21.5; 22.33 21.23 20.29 20.30 15.41 19.71 20.10 20.06 22.12 23.85 23.06 21.87 20.40 19.48 18.93 18.84 19.45 19.27 19.37 21.68 21.59 21 80 22.39 22.23 22.23 22.17 21.01 23.02 22,84 22,76 22.85 22,61 21,90 21,69 20,56 19.62 19.65 20.63 20.10 20,14 20.91 21.25 21,58 22,32 22,39 19.04 21.82 21.75 22.16 22.55 22.64 22.37 22.61 21.46 20.33 19.57 20.66 19.94 19.71 20.34 20.48 20.90 21.56 21.84 18,33 20.32 19.62 19.83 20.63 21.61 21.88 22.77 22.22 21.61 20.85 21.49 20.48 19.87 20.07 15.68 19.92 20.02 20.41 10.71 19.99 10.93 10.78 15.61 20.50 21.25 22.49 22.14 21.72 21.47 21.98 20.96 20.32 20.23 19.77 19.76 19.45 19.25 20.01 19.43 2 HOUR 2 :: 2 1 5 12 91 5 2 20 5 53 6 22

8

52

28

ñ

\$2

5

54

53

22

22

20

6.1

81

5

2

2

1

2

754

Ĩ

SEP 1978

NATIONAL CCEAN SURVEY (NOAA) TIDES, HCLRLY HEIGHTS (FEET)

۱

DA / CF FONTH

BAS1370 DUCK FRF NC (DUTSICE)

5. Meteorological Observations.

The meteorological instruments are read daily and the data are recorded on CERC Form 78-236 (Fig. 34). Daily monitoring of the instruments facilitates rapid identification of problems with the equipment as well as obtaining climatological information. An example of the daily meteorological observation summaries for 1 month of data from the FRF is shown in Figure 35.

IV. DATA REQUESTS

The Coastal Engineering Information and Analysis Center (CEIAC) is responsible for storing and disseminating most data collected at the FRF under the BEM. All data requests should be in writing and addressed to: Coastal Engineering Research Center, Corps of Engineers, Kingman Building, Fort Belvoir, Virginia 22060. Tidal data other than summaries should be obtained directly from the Tides Branch, National Ocean Survey, Rockville, Maryland 20850. A complete explanation of the exact data desired for specific dates or times will expedite filling any request. The request should also explain how the data will be used to help determine if other relevant data are available. For information regarding the availability of data, contact CEIAC at (202) 325-7386. Costs for collecting, copying, and mailing will be borne by the requestor.

		WEA	THE	r	OBSERVATIONS AT THE	/ATIO	NS A	T THF	202		I _										
	_	DUC	DUCK, NI	Ċ	1		: >	1			<u> </u>								****	1 0 1	
2 - Z	~~ [®] ~	E A I	R R R R R R R	*****	N PRES	5 PRES TEND	HICH	LOW DRY TEWP BULB	LB EULB	20 21 21	REL	WIND WIND	N N N N N N N N N N N N N N N N N N N	WIND WIND DIR SPO	340	VAR VAR	ANT SH PREC RI	SKY Rad	REA	EMARKS	003
121					** ** **	20.20	38 30 20	2		1-2	-	3	-			24 S7 38 39	8	2 2 2	47 44 44 70 21	2 12 1 2 2 2 2 2 2 2 2 2 1 2 1 2 1 2 1	78.70
ŝ	1	_	_		00	131	4	63111	10	1		=1	∳ ∳			E	09		75	1-	E
Ø:	30215	-					シンナ	122		a li	\$1		 		 						F
1	<u>ہ</u> ۔ چ						ین ج	561				5				Ē	4				E
м)!	8 8	_			~		4	633	1	_		3	 				8		100		
34 B.	57.03				_	1010	135	1000	<u>ि</u> जि	·		1310	-	ō	0		0 0		01		
U	0	_	_			31	1175370	7055	05150151/151	5120		314		511	4		00		1 076		E
Oi	010	_			O		3336	1002	1/12/21	0201				11	 റ		0	 	40 	4	-
11	1000		_		0	340	190	71/35	71/ 12/21/21/21/21/21/20	0150		1n		1	হ	E	0		5	4	E
\mathbf{n}	9	_			6	010	763	712 215	したのい	0119	_	10		<u>m</u>	5		0		6		F
-1	605		_		9	332	5	6921	121/12/	51/19		600		<u>е</u> С	 ເ		<u>0</u>		o o	3	Ē
000	50/3				0	433	074	640	12/12/2	010	, 	101	 	6	ñ		0		6	6	Ē
10	5/12	-			0	000	775	ମ୍ବାର	3016	い		Ś		160			0 0		6	4	
5 a Y	FIC:0		_		O	5	10	713	020	56		10	 	=	5+12	37	00			4	E
12	601	_			97	<u>a</u> 2	J/ 1085 7135	73515	16	())		215116		<u>m</u>	でキわ	600	0	<u> </u>	-	79	
< N	Ξ		_		0!	i 0	000	319115	3215	000		N		으	σ	a 8	00	 	- - 29	<u></u>	
0	Ø				u	5	01/3	010	27.0	2115		ŝ		m	040	m	00			3	
2	<u>o</u>		-		<u>U</u>	3	7185	21/10/0	1151	5		\sim		5	08+190	210	2	 		0	<u> </u>
0	8	_			(0)	1	ELL	631:15	3151	213		In	_	10		_	00			17	
01730	O				3	,a	<u>०</u> २ २	2116	1014	010		~		3	2		00	 		3	
SO.					0	5	3696	6015	71.15	4		1 C K		341	3		8	 		8	E
<u>O</u> ,	000		_		9-9-	310	37.4	665	2051	021/1020		10	 	4	1 1 1	31	00	 		4	;
835			_		8	30	m	733	いいで	20		10		$\overline{\sigma}$	8	т П	0			0	-
000	415				0	\hat{c}	493	(6	1.,	210	i 	CK.					00	 		4	2
0	R		_		00	730	20	é B L	<u>لم</u>							_	-8				
30	Ξ		_		\$		0	Ő,	τ,	5110		1					ম জ			6	
50	5		_		60	R R R	373	550	5	9		~					Ş			3	
14.11	21/18			+			\sim	$\frac{c}{v_0}$		1		୍ୟ	 				00				E
<u></u>	0		_			<u>-</u> 22 0	2		ت) مد	+ 7						_	 ا	↓ ↓			E
<u>.</u>			_		r S	3	17	<u>त</u> फ	<u>م</u>	<u>1</u>	<u>فت</u>	5586					80			<u>1</u>	_
Ø.	2	_			63	580	78 57	722	5	ن: <u>\</u>						_	ର ହ			19	
-			-	-	-	-												• •			-

Figure 34. Example of meteorological observations recorded on CERC Form 78-236.

Daily meteorological observations at the Field Research Facility, Duck, N.C.

4474EF Cobilions First wire Fontats Constrained First Constrained Functions First Foundation Functions First Fontaile Have Strong and Fortable Fort First Fortable Fort First Fortable First Fortabl RETHOTES ON MEATHER DESERVATIONS

Figure 35. Example of daily meteorological observation summaries.

LITERATURE CITED

BAYLOR COMPANY, "Handbook HB-13528," Houston, Tex., Mar. 1970.

A

- ESTEVA, E., and HARRIS, D.L., "Comparison of Pressure and Staff Wave Gage Records," *Proceedings of the 12th Coastal Engineering Conference*, American Society of Civil Engineers, Ch. 7, 1970, pp. 101-115 (also Reprint 2-71, U.S. Army, Corps of Engineers, Coastal Engineering Research Center, Washington, D.C., NTIS 732 637).
- GARNER, W., Jr., "Four-Electrode Conductivity Measurements," Hydrolab Technical Note 1, Austin, Tex., Dec. 1972.
- MARSH-McBIRNEY, INCORPORATED, "Model 551 Electromagnetic Water Current Meter," TWX No. 710-828-0083, Gaithersburg, Md., 1978.
- THOMPSON, E.F., "Wave Climate at Selected Locations Along U.S. Coasts," TR 77-1, U.S. Army, Corps of Engineers, Coastal Engineering Research Center, Fort Belvoir, Va., Jan 1977.
- NATIONAL WEATHER SERVICE, "Specification No. F420C-SP001, Wind Measuring System," Silver Spring, Md., Sept. 1970.
- VAN BREUGEL, J.G.A., VERHAGEN, C.M., and GERRITZEN, P.L., "Operation and Service Manual for Waverider," Haarlem, Netherlands, 1972.

		INST	UMENT INSTALLATION CHA	RT		
Sensor No.	Sensor type	Data type	Location	Medium depth(MSL) at sensor	Sensor range	Initial installation
				(ft)	(ft)	(date)
615	Baylor staff continuous wire	Wave	Station 620 FRF pier	7.5	28	23 Nov. 1977
625	Baylor staff continuous wire	Wave	Station 1900 FRF pier	27	31	11 Nov. 1977
112	Baylor staff continuous wire	Wave	Jennette's Fishing Pier, Nags Head, N.C.	17	25	Nov. 1977
610	Wavêrider buoy	Wave	350 ft NE. of seaward end of FRF pier	23	32	25 July 1978
620	Waverider buoy	Wave	1.3 mi. E. of seaward end of FRF pier	60	32	25 July 1978
611	Pressure gage	Wave	Station 1940 FRF pier	27	25	23 Nov. 1977
619, 629 (channels X and Y respec- tively)	Electromagnetic current meter	Mean and wave- induced bottom currents	Station 1940 FRF pier	27	±10 ft/s	19 Nov. 1977
639, 649 (channels X and Y respec-	Electromagnetic current meter		500 ft N.ot FiP pier at approximately Station 1940			
tively)		currents		21	±10 ft/s	19 Nov. 1977
612	Temperature probe	Water tempera- ture	9.6 ft below MSL at station 1960, FRF pier	27		20 Nov. 1977
613	Conductivity probe	Conduc- tivity of sea- water	9.6 ft below MSL at station 1'60, FRF pier	27		20 Nov. 1977
602	Temperature probe	Water tempora- ture	22 ft below N.L at station 1960, FRF pier	27		20 Nov. 1977
603	Conductivity probe	Conduc- tivity of sea- water	22 ft below MSL at station 1960, FRF pier	27		20 Nov. 1977
865-1370	Leupold- Stevens float-	Water level infor-	FRF pier end at station 1960			
	activated tlde gage	mation		27		14 Jan. 1979
865-1371	Fischer- Porter float- activated gage	Water level infor- mation	FRF pier at station 720	11		17 Dec. 1977
	Bubbler (pressure) tide gage	Water level infor- mation	1,000 ft W.of shore in Currituck Sound at FRF	5		28 Oct. 1977
	Fischer-Porter float-activated	Water	1,000 ft U.of shore in Currituck Sound	2		
	tide gage	infor- mation	at FRF	5		28 Oct. 1977

APPENDIX INSTRUMENT INSTALLATION CHART

والمحافظ والمحافظ والمحافظ والمحافظ والمحافظ والمحافظ والمحافظ والمحافظ والمحافظ والمحافظ والمحافظ والمحافظ

ľ

with a

12

41

.

1. 3 Total 1.


r y

APPENDIX

.

.

Sensor No.	Sensor type	Data type	Location	Medlum depth(MSL) at sensor	Sensor Fange	Initial installation
				at sensor (ft)	(ft)	(date)
865-1376	Metereraft	Water	1,000 ft W. of shor	e in		
	pressure tide	level infor-	Curvituck Sound at FRF			
	gage	mation		4		16 1.3. 1076
				•		15 July 1978
	Anemometer	Wind	190 ft behind the			
		speed	dune			
		and direc-				
		tion				25 Feb. 1978
	Microbarograph	Contin~	Inside office			
	nicrosarograph	uously	trailer			
		records				
		the at-				
		mospheric				
		pressure				4 Mar. 1978
	Barometer	At mos-	Inside office			
		pheric	trailer			
		pressure				4 Mar. 1978
	Hygro-	Contin-	Located in the inst	ru-		
	thermograph	uously	ment shack, 300 ft	from		
		records	dune			
		the air				
		tempera- ture and				
		relative				
		humidity				4 Mar. 1978
	Maximum ther-	Maximum	Instrument shelter			
	mometer	air tem-				
		perature				4 Mar. 1978
	Minimum ther-	Minimum	Instrument shelter			
	mometer	air tem-				
		perature				4 Mar. 1978
	12-inch weigh-	Contin-	285 ft behind the du	ine		
	ing rain gage	uously				
		records				
		the				
		amount of precipi-				
		tation				4 Mar. 197 8
	6-inch rain	Precipi-	270 ft behind the du	ane		
	gage	tation	· ·			4 Mar. 1978
	Sling	Dew	Used at instrument			
	psychrometer	point	shelter			11 Dec. 1978
	Mechanical	Solar	Mounted on top of			
	pyranograph	radia-	instrumer; shelter			
	• • • • • • • • •	tion				18 Jan. 1979



a with the second state of the

10.00