

AD-A129 339

SUMMARY OF WORK ON 'COOLED ION FREQUENCY STANDARD'
FISCAL YEAR 1983(U) NATIONAL BUREAU OF STANDARDS
BOULDER CO TIME AND FREQUENCY DIV. D J WINELAND ET AL.
UNCLASSIFIED 25 JUN 83 N00014-83-F-0003

1/1

F/G 7/4

NL


END

DATE

FILMED

7 83

DTIC



MICROCOPY RESOLUTION TEST CHART
NATIONAL BUREAU OF STANDARDS-1963-A

REPORT DOCUMENTATION PAGE

READ INSTRUCTIONS
BEFORE COMPLETING FORM

1. REPORT NUMBER None	2. GOVT ACCESSION NO. AD-A129339	3. RECIPIENT'S CATALOG NUMBER Annual Oct. 82 - Sept. 83
TITLE (and Subtitle) Summary of Work on "Cooled Ion Frequency Standard" ONR Contract No. N00014-83-F-0003		5. TYPE OF REPORT & PERIOD COVERED
7. AUTHOR(s) D. J. Wineland		6. PERFORMING ORG. REPORT NUMBER
9. PERFORMING ORGANIZATION NAME AND ADDRESS National Bureau of Standards Time and Frequency Division 524 Boulder, CO 80303		8. CONTRACT OR GRANT NUMBER(s) ONR Contract No. N00014-83-F-0003
11. CONTROLLING OFFICE NAME AND ADDRESS Office of Naval Research Physics Program Office Arlington, VA 22217		10. PROGRAM ELEMENT, PROJECT, TASK AREA & WORK UNIT NUMBERS 1731319.WIAE NR 407-004
14. MONITORING AGENCY NAME & ADDRESS (if different from Controlling Office)		12. REPORT DATE 6-25-83
		13. NUMBER OF PAGES
		15. SECURITY CLASS. (of this report)
		15a. DECLASSIFICATION/DOWNGRADING SCHEDULE
16. DISTRIBUTION STATEMENT (of this Report) Approved for public release; distribution unlimited		
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report) Same		
18. SUPPLEMENTARY NOTES		
19. KEY WORDS (Continue on reverse side if necessary and identify by block number) Atomic spectroscopy; Doppler narrowing; Doppler shifts; Frequency standards; High resolution spectroscopy; Ion storage; laser spectroscopy; Penning trap; Radiation pressure.		
20. ABSTRACT (Continue on reverse side if necessary and identify by block number) The purpose of this work is to develop techniques to overcome the fundamental limits of present frequency standards--the second and residual first-order Doppler shifts. To this end we study suitable frequency reference transitions in ions which are stored on electromagnetic traps and cooled by radiation pressure to < 1K.		

DTIC
ELECTE
JUN 15 1983
S A D

AD A129339

DTIC FILE COPY

(FY 83)

ONR Contract No. N00014-83-F-0003

Co-Principal Investigators

Wayne M. Itano
Frequency & Time Standards Group
524.10
National Bureau of Standards
Boulder, Colorado 80303
FTS: 320-5632
(303) 497-5632

1. ☐ None For
 2. ☐ Basic
 3. ☐ Full
 4. ☐ Advanced
 5. ☐ Classification
 6. ☐ None
 7. ☐ Basic
 8. ☐ Full
 9. ☐ Advanced
 10. ☐ Classification
 11. ☐ None
 12. ☐ Basic
 13. ☐ Full
 14. ☐ Advanced
 15. ☐ Classification
 16. ☐ None
 17. ☐ Basic
 18. ☐ Full
 19. ☐ Advanced
 20. ☐ Classification
 21. ☐ None
 22. ☐ Basic
 23. ☐ Full
 24. ☐ Advanced
 25. ☐ Classification
 26. ☐ None
 27. ☐ Basic
 28. ☐ Full
 29. ☐ Advanced
 30. ☐ Classification
 31. ☐ None
 32. ☐ Basic
 33. ☐ Full
 34. ☐ Advanced
 35. ☐ Classification
 36. ☐ None
 37. ☐ Basic
 38. ☐ Full
 39. ☐ Advanced
 40. ☐ Classification
 41. ☐ None
 42. ☐ Basic
 43. ☐ Full
 44. ☐ Advanced
 45. ☐ Classification
 46. ☐ None
 47. ☐ Basic
 48. ☐ Full
 49. ☐ Advanced
 50. ☐ Classification
 51. ☐ None
 52. ☐ Basic
 53. ☐ Full
 54. ☐ Advanced
 55. ☐ Classification
 56. ☐ None
 57. ☐ Basic
 58. ☐ Full
 59. ☐ Advanced
 60. ☐ Classification
 61. ☐ None
 62. ☐ Basic
 63. ☐ Full
 64. ☐ Advanced
 65. ☐ Classification
 66. ☐ None
 67. ☐ Basic
 68. ☐ Full
 69. ☐ Advanced
 70. ☐ Classification
 71. ☐ None
 72. ☐ Basic
 73. ☐ Full
 74. ☐ Advanced
 75. ☐ Classification
 76. ☐ None
 77. ☐ Basic
 78. ☐ Full
 79. ☐ Advanced
 80. ☐ Classification
 81. ☐ None
 82. ☐ Basic
 83. ☐ Full
 84. ☐ Advanced
 85. ☐ Classification
 86. ☐ None
 87. ☐ Basic
 88. ☐ Full
 89. ☐ Advanced
 90. ☐ Classification
 91. ☐ None
 92. ☐ Basic
 93. ☐ Full
 94. ☐ Advanced
 95. ☐ Classification
 96. ☐ None
 97. ☐ Basic
 98. ☐ Full
 99. ☐ Advanced
 100. ☐ Classification
 101. ☐ None
 102. ☐ Basic
 103. ☐ Full
 104. ☐ Advanced
 105. ☐ Classification
 106. ☐ None
 107. ☐ Basic
 108. ☐ Full
 109. ☐ Advanced
 110. ☐ Classification
 111. ☐ None
 112. ☐ Basic
 113. ☐ Full
 114. ☐ Advanced
 115. ☐ Classification
 116. ☐ None
 117. ☐ Basic
 118. ☐ Full
 119. ☐ Advanced
 120. ☐ Classification
 121. ☐ None
 122. ☐ Basic
 123. ☐ Full
 124. ☐ Advanced
 125. ☐ Classification
 126. ☐ None
 127. ☐ Basic
 128. ☐ Full
 129. ☐ Advanced
 130. ☐ Classification
 131. ☐ None
 132. ☐ Basic
 133. ☐ Full
 134. ☐ Advanced
 135. ☐ Classification
 136. ☐ None
 137. ☐ Basic
 138. ☐ Full
 139. ☐ Advanced
 140. ☐ Classification
 141. ☐ None
 142. ☐ Basic
 143. ☐ Full
 144. ☐ Advanced
 145. ☐ Classification
 146. ☐ None
 147. ☐ Basic
 148. ☐ Full
 149. ☐ Advanced
 150. ☐ Classification
 151. ☐ None
 152. ☐ Basic
 153. ☐ Full
 154. ☐ Advanced
 155. ☐ Classification
 156. ☐ None
 157. ☐ Basic
 158. ☐ Full
 159. ☐ Advanced
 160. ☐ Classification
 161. ☐ None
 162. ☐ Basic
 163. ☐ Full
 164. ☐ Advanced
 165. ☐ Classification
 166. ☐ None
 167. ☐ Basic
 168. ☐ Full
 169. ☐ Advanced
 170. ☐ Classification
 171. ☐ None
 172. ☐ Basic
 173. ☐ Full
 174. ☐ Advanced
 175. ☐ Classification
 176. ☐ None
 177. ☐ Basic
 178. ☐ Full
 179. ☐ Advanced
 180. ☐ Classification
 181. ☐ None
 182. ☐ Basic
 183. ☐ Full
 184. ☐ Advanced
 185. ☐ Classification
 186. ☐ None
 187. ☐ Basic
 188. ☐ Full
 189. ☐ Advanced
 190. ☐ Classification
 191. ☐ None

Reproduction in whole or in part is permitted for any purpose of the United States Government

Contract Description

→ The purpose of this work is to develop techniques to overcome the fundamental limits of present frequency standards--the second and residual first-order Doppler shifts. To this end, ~~we study~~ suitable frequency reference transitions in ions which are stored in electromagnetic traps and cooled by radiation pressure to ~~1K.~~ *we are at 1K. ←*
see below

Scientific Problem

The scientific problems are (1) to greatly suppress second order and residual first order Doppler shifts in atomic frequency standards in a fundamental way--by substantially reducing the kinetic energy of ions stored in electromagnetic traps, (2) to study suitable reference transitions in ions that can be used as frequency standards, and (3) to study the problems generic to all stored ion frequency standards. The goal is to achieve at least a factor of 100 improvement in accuracy over the present best device, the Cesium beam frequency standard, which has an accuracy of approximately one part in 10^{13} .

Scientific and Technical Approach

We will continue studies of laser cooling. We can now obtain ion temperatures near .1K in ion clouds--temperatures near 10^{-3} K should be possible. We will continue work to develop better traps for spectroscopic studies. We are completing experiments on Be^+ and will soon make improvements in trap design. We are developing a separate experiment for $^{201}\text{Hg}^+$ ions. This experiment has the goal of realizing a prototype microwave frequency standard with 10^{-15}

accuracy. We continue experiments on cloud dynamics. It appears that the largest systematic frequency shift in a frequency standard based on ions stored in a Penning trap will be due to a second order Doppler effect caused by the rotation of the ion cloud. Therefore a detailed study of this rotation is required. A superconducting magnet experiment is now being set up for this purpose.

PROGRESS DURING LAST CONTRACT PERIOD

A. $^{201}\text{Hg}^+$ studies

1. 194 nm generation. $^{201}\text{Hg}^+$ has been chosen as a prototype frequency standard because it appears to have the best potential performance of any ion that might conceivably be used in a microwave frequency standard. One of its chief drawbacks is that the coherent 194 nm radiation required for laser cooling/optical pumping is difficult to produce. Therefore a significant effort has gone into the development of this source. Tunable, narrowband cw radiation at 194 has now been generated by frequency doubling the output of a 514 nm argon ion laser and mixing this output (up to 80 mW at 257 nm) with the output of a 792 nm dye laser in a potassium pentaborate (KB5) crystal. Last year at this time, the basic generation scheme had been realized, but at low power ($\ll 1\mu\text{W}$). During this year the system has been improved to generate approximately 3 μW in a bandwidth $< 2\text{ MHz}$ which should be enough for initial cooling experiments on Hg^+ . This source of narrowband tunable radia-

tion < 200 nm is unique in the world and has resulted in two very significant publications.

2. Hg⁺ trap. A Penning trap apparatus for $^{198}\text{Hg}^+$ has been completed and initial testing performed. (Electron storage and electron cyclotron resonance for magnetic field calibration). In the near future, laser cooling experiments will be initiated.

B. $^9\text{Be}^+$ experiments

1. Cloud studies In order to optimize conditions for the $^{201}\text{Hg}^+$ experiments, we study the general ion storage problems using $^9\text{Be}^+$ ions where the required radiation ($\lambda=313$ nm) for optical pumping and laser cooling experiments is easier to produce. We have undertaken a careful study of ion cloud dynamics since it is expected that the largest systematic effect in a microwave frequency standard based on ions stored in a Penning trap will be due to a second order Doppler effect due to cloud rotation. To this end we have studied the effects of impurity ions on the ion "cloud" and observe a marked effect on the spreading of the cloud when the cooling laser is off. When the impurity ions such as H_3^+ , He^+ , BeH^+ , NH_3^+ , H_3O^+ and N_2H^+ are eliminated from the trap by strong resonance excitation at their cyclotron frequencies this spreading rate decreases by 1 to 2 orders of magnitude. This is very important because the Q's of "frequency standard" transitions in ions will be limited only by inter-

rogation time. Since the cooling laser must be off during this time, in order to avoid light shifts, then ion spreading must be kept to a minimum in order to obtain high Q.

We have developed a sensitive way to measure cloud size and density. The size of the cloud was determined by using a second probe laser. The density was determined by measuring the space charge shifted $\vec{E} \times \vec{B}$ cloud rotation frequency via the change in Doppler shift across the cloud using the probe laser. By using an optical depopulation technique, the steady state behavior of the ion cloud is undisturbed by the probe laser. This subject will be discussed in a future publication.

2. Observation of "liquid plasma" state. As an outgrowth of the experiments to determine cloud temperature and density, we have reached the condition where the ratio, Γ , of electrostatic energy between nearest neighbor ions and kinetic energy is approximately 10. The liquid state usually occurs for $\Gamma \approx 2$. We are now devising a way using laser pulse techniques to study liquid phenomenon and have already realized a way to detect plasma condensation into a "Wigner crystal" ($\Gamma \approx 150$).
3. High precision mass spectroscopy. As an outgrowth of the above cloud studies we have developed a sensitive way to measure ion cyclotron resonance. Comparison of cyclotron frequencies between two ion species in the same magnetic field then gives direct mass comparisons.

We irradiate a small cloud of ions (~ 20 in a cloud of diam. $\sim 100 \mu\text{m}$) with a laser beam of approximately the same diameter and observe the fluorescence scattering. If the ions are driven by rf electric fields at their cyclotron frequency, then the ion orbits increase in size which causes a decrease in overlap between the laser and ion cloud which causes a decrease in fluorescence. Using this method, we have made an indirect measurement of the proton/electron mass ratio to 0.3 ppm, limited by a theoretical calculation of $g_J(^9\text{Be}^+)$. (The spin flip frequency of $^9\text{Be}^+$ is used to calibrate the field.) The potential accuracy of mass comparisons using this method approaches one part in 10^{13} -- this is orders of magnitude higher than that envisioned by any other experiment.

4. Fine and hyperfine structure of $2p^2P$ state in Be^+ . Using optical/optical double resonance schemes, we have measured these transition frequencies by comparing to I_2 . This is interesting in Be^+ because these frequencies can be calculated fairly precisely due to the relatively simple structure of Be^+ .
5. Model frequency standard. We are in the process of operating the $^9\text{Be}^+$ apparatus as a frequency standard. We are using the $(M_1, M_J) = (-\frac{3}{2}, \frac{1}{2}) \rightarrow (-\frac{1}{2}, \frac{1}{2})$ transition at ~ 300 MHz for this purpose. We have extended relaxation times so we should expect linewidths of approximately 0.025 Hz and we are now trying to improve signal to noise by using adiabatic fast passage level

manipulation. Accuracies approaching that of the Cesium beam standard are expected. Significant improvement is expected in Hg^+ .

C. Theoretical studies

1. Trap design. Computer designed traps should greatly ease construction and improve performance by exactly nulling fourth and sixth order terms in the electric potential. The simple trap pictured here should have performance better than any of the complicated hyperbolic traps constructed to date. Two journal publications will be prepared on this work.

2. Magnetic field dependence of (s,l) electron configurations. For the future, optical frequency standards might be based on weakly allowed intercombination transitions in group III A ions. Unfortunately at zero field, these transition frequencies are linearly dependent on magnetic field. To counteract this one could work at "field independent" magnetic fields as we do in our microwave/rf experiments. Using the modified Breit Wills theory for hfs interactions, we have calculated these field independent points in III A ions. In particular, In^+ looks like a very promising candidate for future work.

D. New apparatus

1. Superconducting magnet system. We now have a 6 T magnet system installed and running. An NMR diagnostic system is now being constructed. First ion trap experiments will be performed on Mg^+ or Hg^+ ions.
2. Real time imaging system. A real time (response time 1 ms) imaging system (resolution 10-15 μm) has been designed and the image intensifier tube has been ordered. This will aid cloud dynamical studies.
3. Hg^+ Penning trap apparatus. Described above.

E. Ion storage publications in preparation or published since May 80.

1. "Generation of continuous-wave 194 nm radiation by sum frequen-

- cy mixing in an external ring cavity". H. Hemmati, J. C. Bergquist, and Wayne M. Itano, Optics Letters 8, 73 (1983).
2. "Laser Fluorescence Mass Spectroscopy". D. J. Wineland, J. J. Bollinger, and Wayne M. Itano, Phys. Rev. Lett. 50, 628 (1983).
 3. "High Power Second Harmonic Generation of 257 nm Radiation in an External Ring Cavity". J. C. Bergquist, H. Hemmati and Wayne M. Itano, Optics Commun. 43, 437 (1982).
 4. "Time and Frequency Standards Based on Charged Particle Trapping." Wayne M. Itano, D. J. Wineland, H. Hemmati, J. C. Bergquist, H. Hemmati and J. J. Bollinger, I.E.E.E. Trans. Nuc. Sci, Vol. N.S. 30, No. 2, Apr. 1983.
 5. "High Resolution Spectroscopy of Stored Ions". D. J. Wineland, Wayne M. Itano and R. S. Van Dyck Jr., to be published as a review in Advances in Atomic and Molecular Physics, Vol. 19, ed. by Bederson & Bates. (Academic Press, Aug. 1983).
 6. "Sum-Frequency Generation of Continuous-Wave 243 nm Radiation". H. Hemmati and J. C. Bergquist, submitted to Optics Commun.
 7. "Microwave-optical double resonance spectroscopy of stored atomic ions". Wayne M. Itano, D. J. Wineland, H. Hemmati, J. C. Bergquist, and J. J. Bollinger, Abstracts for the eighth International Conference on Atomic Physics.

8. "Spectroscopy of Hg^+ Ions by Continuous-Wave Coherent Radiation at 194 nm". H. Hemmati, J. C. Bergquist, and Wayne M. Itano, Bull. Am. Phys. Soc. 28, May 1983.
9. "Temperature, Density, and Size Measurement of Laser-Cooled Ion Clouds in a Penning Trap". J. J. Bollinger and D. J. Wineland, Bull. Am. Phys. Soc. 28, May 1983.
10. "Ground-state hfs - Zeeman Spectrum of $^9\text{Be}^+$ ". Wayne M. Itano, D. J. Wineland, and J. J. Bollinger, Bull. Am. Phys. Soc. 28, May 1983.
11. "Laser Fluorescence Detection of Ion Cyclotron Resonance". D. J. Wineland, J. J. Bollinger, and Wayne M. Itano, Bull. Am. Phys. Soc. 28, May 1983.
12. "Properties of an Electromagnetic Trap for Charged Particles". E. C. Beaty, Bull. Am. Phys. Soc. 28, May 1983.
13. "Measurements of Fine and Hyperfine Structures and Optical Transition Frequencies of the $2p^2P$ States of $^9\text{Be}^+$ ". J. S. Wells, J. J. Bollinger, and D. J. Wineland, Bull. Am. Phys. Soc. 28, May 1983.
14. "Magnetic Field Dependence of (s,l) Electron Configurations". D. J. Wineland and Wayne M. Itano, Bull. Am. Phys. Soc. 28, May 1983.

F. Invited talks on ion storage since May 1982

1. PTB, West Germany Aug., 1982 (Itano).
2. Optical Society of America - Annual Meeting. Tucson, AZ Oct., 1982 (Wineland).
3. Seventh Conference on the Application of Accelerators in Research and Industry. Denton, TX Nov., 1982 (Itano).
4. International Union of Radio Scientists. Boulder, CO, Jan. '83 (Wineland).
5. Colloquium at Harvard Univ. Cambridge MA Jan. '83 (Wineland).
6. Colloquium at Colorado State Univ. Fort Collins, CO Dec. '82 (Bergquist).
7. NBS Research Science Seminar. Gaithersburg, MD Jan. '83 (Bergquist).
8. Colloquium at Colorado School of Mines. Golden, CO Jan. '83 (Bergquist).
9. Sixth International Conf. on Laser Spectroscopy. Interlocken, Switzerland June '83 (Bollinger and Hemmati).

10. Gordon Conference on Atomic Physic. New London, NH July, 1983
(Itano).

11. Conference of the Society of Photo-Optical Instrumentation Engineers
(S.P.I.E.) San Diego, CA Aug. 1983 (Wineland).

Miscellaneous

A. All FY83 funds will be spent by end of contract year.

B. Other contract support of principal investigators:

David J. Wineland and Wayne M. Itano

AFOSR \$140K "Precision Frequency Metrology Using Laser Cooled Ions."

KEY PERSONNEL (FY83)

Co-Principal Investigators	D. J. Wineland	(40%)
	Wayne M. Itano	(40%)
Senior Staff Scientists	J. C. Bergquist	(40%)
	C. Manney	(50%)
	E. Beaty (started	
	June 1/82)	(50%)
Research Associate	J. Wells	(25%)
	H. Hemmati	(40%)
NRC Postdoctoral	J. Bollinger	(50%)
Research Associate		
Guest Research Associate	R. Blatt	(20%)