


AD-A162 189


GPIB INTERFACE AND TIMING GENERATOR FOR THE
HEWLETT-PACKARD 2401C INTEGRA (U) AEROSPACE CORP EL
SEGUNDO CA ELECTRONICS RESEARCH LAB M F BOTTJER ET AL
30 SEP 85 TR-0084A(5925-00)-1 SD-TR-85-67 F/G 14/2

1/1

UNCLASSIFIED

NL





MICROCOPY RESOLUTION TEST CHART
NATIONAL BUREAU OF STANDARDS-1963-A

12

AD-A162 189

**GPIB Interface and Timing Generator
for the Hewlett-Packard 2401C
Integrating Digital Voltmeter**

**M. F. BOTTJER, W. A. JOHNSON,
T. T. MORI and D. Y. WATANABE**
Electronics Research Laboratory
The Aerospace Corporation
El Segundo, CA 90245

30 September 1985

**DTIC
ELECTE**
DEC 10 1985
S D
A

APPROVED FOR PUBLIC RELEASE;
DISTRIBUTION UNLIMITED

DTIC FILE COPY

Prepared for
SPACE DIVISION
AIR FORCE SYSTEMS COMMAND
Los Angeles Air Force Station
P.O. Box 92960, Worldway Postal Center
Los Angeles, CA 90009-2960

85 12 -9 068


This report was submitted by The Aerospace Corporation, El Segundo, CA 90245, under Contract No. F04701-83-C-0084 with the Space Division, P.O. Box 92960, Worldway Postal Center, Los Angeles, CA 90009-2960. It was reviewed and approved for The Aerospace Corporation by M. J. Daugherty, Director, Electronics Research Laboratory. 1st Lt Alfredo Carbonel, Jr., SD, was the Air Force project officer.

This report has been reviewed by the Public Affairs Office (PAS) and is releasable to the National Technical Information Service (NTIS). At NTIS, it will be available to the general public, including foreign nationals.

This technical report has been reviewed and is approved for publication. Publication of this report does not constitute Air Force approval of the report's findings or conclusions. It is published only for the exchange and stimulation of ideas.



ALFREDO CARBONEL, Jr., 1Lt, USAF
Prgm Mgr, COMSAT Device Development



JOSEPH HESS, GM-15
Director, AFSTC West Coast Office
AFSTC/WCO OL-AB

REPORT DOCUMENTATION PAGE		READ INSTRUCTIONS BEFORE COMPLETING FORM	
1. REPORT NUMBER SD-TR-85-67	2. GOVT ACCESSION NO. AD-A162 189	3. RECIPIENT'S CATALOG NUMBER	
4. TITLE (and Subtitle) GPIB INTERFACE AND TIMING GENERATOR FOR THE HEWLETT-PACKARD 2401C INTEGRATING DIGITAL VOLTMETER		5. TYPE OF REPORT & PERIOD COVERED	
		6. PERFORMING ORG. REPORT NUMBER TR-0084A(5925-08)-1	
7. AUTHOR(s) Martin F. Bottjer, Walter A. Johnson Tsutomu T. Mori, and Donald Y. Watanabe		8. CONTRACT OR GRANT NUMBER(s)	F04701-83-C-0084
		9. PERFORMING ORGANIZATION NAME AND ADDRESS The Aerospace Corporation El Segundo, Calif. 90245	10. PROGRAM ELEMENT, PROJECT, TASK AREA & WORK UNIT NUMBERS
11. CONTROLLING OFFICE NAME AND ADDRESS Space Division Los Angeles Air Force Station Los Angeles, Calif. 90009-2960		12. REPORT DATE 30 September 1985	
		13. NUMBER OF PAGES 25	
14. MONITORING AGENCY NAME & ADDRESS (if different from Controlling Office)		15. SECURITY CLASS. (of this report)	Unclassified
		15a. DECLASSIFICATION/DOWNGRADING SCHEDULE	
16. DISTRIBUTION STATEMENT (of this Report) Approved for public release; distribution unlimited.			
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report)			
18. SUPPLEMENTARY NOTES			
19. KEY WORDS (Continue on reverse side if necessary and identify by block number) Low-Frequency Drift Measurements Programmable Integrator			
20. ABSTRACT (Continue on reverse side if necessary and identify by block number) GPIB interface logic and a timing generator for the Hewlett-Packard 2401C integrating digital voltmeter were designed, constructed, and checked out. This logic enables the control of the 2401C by a Hewlett Packard 9825T calculator. This hardware is described herein. Also described is software for a typical usage for the voltmeter, calculator, and interface system with a printer and plotter as outputting devices. <i>key words:</i>			

CONTENTS

I.	INTRODUCTION.....	7
II.	EXAMPLE OF A TYPICAL USAGE.....	15
III.	INSTRUCTIONS FOR USING DATA ACQUISITION AND REDUCTION PROGRAM.....	29
IV.	2401C INTEGRATING DIGITAL VOLTMETER CONTROL SETTINGS AND CONNECTIONS TO THE INTERFACE BUS.....	33
APPENDIX A.	REJECTION OF 60 Hz AND HARMONICALLY RELATED COMPONENTS.....	A-1
APPENDIX B.	QUANTIZATION LEVEL OF THE 2401C.....	B-1
APPENDIX C.	EFFECT ON THE CALCULATION OF σ OF THE TIME- BANDWIDTH PRODUCT OF THE INPUT VOLTAGE.....	C-1



FIGURES

1.	Functional Block Diagram.....	8
2.	GPIB Interface Logic for 2401C (IDVM).....	9
3.	Address and Mode Select GPIB Interface Circuit.....	10
4.	GPIB Interface Drivers, Card 2.....	11
5.	Data Acquisition Block Diagram.....	16
6.	Data Acquisition Flow Chart Using 9825T Calculator and 2401C IDVM.....	20
7.	Plot of Table 3 Data Set.....	22

TABLES

1.	Interconnections for 2401C and GPIB Interface Logic.....	12
2.	Data Acquisition and Reduction Program.....	17
3.	Manual Mode Output.....	21
4.	Automatic Mode Output.....	23

I. INTRODUCTION

The GPIB interface and timing generator for the Hewlett Packard (or DYMEC) 2401C allow calculator control of the integration time of the 2401C and also enable the calculator to read the 2401C output.

The block diagram (Fig. 1) shows the functional elements of the "GPIB/TIMING GENERATOR INTERFACE." The GPIB (IEE 488) protocol is handled by a 74LS488 integrated circuit along with a few support integrated circuits. A 10 MHz crystal clock required for operation of the 74LS488 is divided by 10^5 to produce a 10 ms clock for the timing generator. This clock is used for counting down a four-digit presetable counter. The timing generator provides the start/stop information for the 2401C. The timing range is 0.01 to 99.99 sec in steps of 10 ms. We should note that the time between integration times (i.e., the "dead time") is a minimum of about 17 ms for the circuit implementation described in this report. The problem of integration time selection to achieve rejection of 60 Hz and harmonic components is discussed in Appendix A. The quantization level of the 2401C and how it is set are discussed in Appendix B.

Output data from the 2401C is formatted as eight binary-coded decimal (BCD) words of six digits plus sign and exponent. These data are parallel loaded into four 8 bit shift registers. The data are then outputted sequentially over the GPIB as eight data words. The circuit details of the unit are given in Figs. 2 through 4. In Table 1 are listed the interconnections for the 2401C and GPIB interface logic.





Fig. 1. Functional Block Diagram



- 1 15 02 21
- 2 15 04 22, 3
- 3 15 74 24, 5
- 4 15 139 26
- 5 15 166 27, 8, 9, 10
- 6 15 192 211, 12, 13, 14, 15
- 7 15 221 216, 17
- 8 15 240 218
- 9 15 377 219
- 10 15 390 220, 21
- 11 96 15 488 222

NOTE: H (HIGH)
 L (LOW)
 1K RESISTOR TO +5V

Fig. 2. GPIB Interface Logic for 2401C (IDVM)




Fig. 3. Address and Mode Select GPIB Interface Circuit




Fig. 4. GPIB Interface Drivers, Card 2.

Table 1. Interconnections for 2401C and GPIB Interface Logic

2401C Connector J2 Pin No.	BCD Weight and Coding of Outputs from 2401C ^a	3M Connector 3429-4005 26 Pins (2 each)	CA3081 Base Card 2	CA3081 Collector Card 2	On-Board Terminal	Shift Register 74LS166 Card 1
1	10 ⁻ⁿ	1	Z4-11	Z4-12	1	Z10-2
2	10 ⁻ⁿ	2	Z4-6	Z4-4	2	Z9-2
3	10 ⁰	1	Z1-16	Z1-1	3	Z10-3
4	10 ⁰	2	Z1-3	Z1-2	4	Z9-3
5	10 ¹	1	Z1-6	Z1-4	5	Z10-4
6	10 ¹	2	Z1-10	Z1-9	6	Z9-4
7	10 ²	1	Z2-3	Z2-2	7	Z10-5
8	10 ²	2	Z2-13	Z2-14	8	Z9-5
9	10 ³	1	Z2-10	Z2-9	9	Z10-10
10	10 ³	2	Z2-8	Z2-7	10	Z9-10
11	10 ⁴	1	Z3-13	Z3-14	11	Z10-11
12	10 ⁴	2	Z3-11	Z3-12	12	Z9-11
13	10 ⁵	1	Z3-8	Z3-7	13	Z10-12
14	10 ⁵	2	Z4-16	Z4-1	14	Z9-12
15	Function 1	8	Z5-16	Z5-1	15	Z10-14
16	Function 2	21	Z5-3	Z5-2	16	Z9-14
17					Record Pin 17	
18					Reset	18
19					+ 5	19
20					Start	37 Z5-8
21					Ground	20
22						
23						

Table 1. Interconnections for 2401C and GPIB Interface Logic (Continued)

2401C Connector J2 Pin No.	BCD Weight and Coding of Outputs from 2401C ^a	3M Connector 3429-4005 26 Pins (2 each)	CA3081 Base Card 2	CA3081 Collector Card 2	On-Board Terminal	Shift Register 74LS166 Card 1
24						
25						
26	10 ⁻ⁿ 4	1	Z4-10	Z4-9	21	Z8-2
27	10 ⁻ⁿ 8	14	Z4-8	Z4-7	22	Z7-2
28	10 ⁰ 4	2	Z1-13	Z1-14	23	Z8-3
29	10 ⁰ 8	15	Z1-11	Z1-12	24	Z7-3
30	10 ¹ 4	3	Z1-8	Z1-7	25	Z8-4
31	10 ¹ 8	16	Z2-16	Z2-1	26	Z7-4
32	10 ² 4	4	Z2-11	Z2-1	27	Z8-5
33	10 ² 8	17	Z2-6	Z2-4	28	Z7-5
34	10 ³ 4	5	Z3-16	Z3-1	29	Z8-10
35	10 ³ 8	18	Z3-3	Z3-2	30	Z7-10
36	10 ⁴ 4	6	Z3-6	Z3-4	31	Z8-11
37	10 ⁴ 8	19	Z3-10	Z3-9	32	Z7-11
38	10 ⁵ 4	7	Z4-3	Z4-2	33	Z8-12
39	10 ⁵ 8	20	Z4-13	Z4-14	34	Z7-12
40	Function 4	8	Z5-13	Z5-14	35	Z8-14
41	Function 8	21	Z5-11	Z5-12	36	Z7-14
50	Ground	25				

^aListing is for a 1248 code (option M21). Other options use a 122'4 code.

II. EXAMPLE OF A TYPICAL USAGE

A typical data acquisition test setup is shown in Fig. 5. Following Fig. 5 is Table 2, which is a listing of a data acquisition and reduction program used with the test setup. Following the program are instructions for using the program (Section III), the 2401C control settings and connections (Section IV), and lastly a program flow chart, depicted as Fig. 6.

To illustrate the use of the program, the noise output of a LM194 op-amp set for a voltage gain of 50 and with a 2000 ohm termination was used as an input to the 2401C. Table 3 and Fig. 7 give the results for a manual mode run for 50, 1-sec integration data points. The output for a typical automatic mode sequence is shown in Table 4. The sequence is for 50 integration times of 1 and 2 sec, three blocks of each, and two iterations.




Fig. 5. Data Acquisition Block Diagram

Table 2. Data Acquisition and Reduction Program

```

0: "IDVM Ver 2.0 (MFB 831012 [12/16/83])":
1: "For use with 2401C interface Don-1":
2: "added avg std dev & avg dev\T 2/22/84":
3: dim D[4],D$(8),P$(10),L$(80),T[22],X$(80),T$(20)
4: dev "M",723,"P",701;fmt 8,c,z
5:
6: ent "1 2 2' 4 code (Y)?" ,P$
7: if cap(P$)="Y";sfg 8
8: ent "AUTO MODE (Y)?" ,P$
9: if cap(P$)="Y";sfg 2
10:
11: "Bgn":c11 'Time';wrt "P","IDVM Ver 2.0 "&T$
12: 1)T)T[1];ent "Label",L$
13: if flg13;X$)L$
14: L$)X$;wrt "P",L$
15: ent "NUMBER OF SAMPLES ?",N;int(N))N
16: if N<=1;jmp -1
17: if flg2;goto "Setup"
18: ent "SAMPLE INTEGRATION TIME [Sec]?" ,T[1]
19: if T[1]>99.99 or T[1]<.01;jmp -1
20:
21: "Loop":if flg2;gsb "GEN"
22: 0)P)0)X)Y;fxd 0;sfg 1
23: if flg8;conv 54,52,55,53,60,54,61,55,62,56,63,57
24: 10000+int(.5+100T[T]))D
25: str(D))D$;D$(3))D$
26: for I=1 to 4;48+161+val(D$(I,I))D[I];next I
27:
28: for I=1 to N;dsp I
29: wtb "M",D[4],D[3],D[2],D[1]
30: red "M.8",D$
31: val(D$(2,7))V
32: V/tn^val(D$(8,8))V
33: if D$(1,1)="2";-V)V
34: V)rI;next I
35:
36: r1/T[T])H)L
37: for I=1 to N;rI/T(T))V)rI
38: if V>H;V)H
39: if V<L;V)L
40: I+X)X;V+Y)Y
41: IV+P)P;II+Q)Q
42: next I
43: (NP-XY)/(NQ-XX))A
44: (YQ-XP)/(NQ-XX))B
45: Y/N)M;0)P)Y; for I=1 to N

```

Table 2. Data Acquisition and Reduction Program (Continued)

```

46: rI-AI-B)V;V+Y)Y;VV+P)P
47: next I
48: (NP-YY)/N(N-1))V;\V)S
49: conv;if flg2;gto "LIST"
50:
51: "Out":beep;ent "PLOT/LIST/NEXT (P/L/N) ?",P$
52: if flg13;gto +4
53: if cap(P$[1])="P";gto "PLOT"
54: if cap(P$[1])="L";gto "LIST"
55: if cap(P$[1])="N";gto "Bgn"
56: prt "No. of Pts.="&str(N)
57: fxd 6;prt "Mean=",M
58: prt "Std Dv=",S
59: prt "Slope=",A
60: prt "Y Int=",B
61: gto "Bgn"
62:
63: "PLOT":pclr;sc1 0,N,M-1.5(M-L),M+1.5(H-M)
64: pen# 3;fxd 0;xax M,1,0,N,5
65: fxd int(.5-log((H-L)/10))E;yax 0,tn^(-E),L,H,2
66: pen# 2;plt 1,rI,-2
67: for I=1 to N;plt I,rI;next I
68: sc1 0,10,0,10
69: plt 2,10,1
70: pen# 1;lb1 L$
71: pen# ;gto "Out"
72:
73: "LIST":wrt "P"
74: cll 'Time';wrt "P",T$
75: if W=1;gto "L"
76: fmt 1,f12.7,z
77: fmt 2,f12.7
78: for I=1 to 1+int(N/6)
79: for J=1 to 6
80: 6<I-1>+J)K
81: if K>N;gto "L"
82: if K=N or J=6;wrt "P.2",rK;gto +2
83: wrt "P.1",rk
84: next J
85: next I
86:
87: "L":fmt 1,c18,f11.8;fxd 0;wrt "P"
88: wrt "P"," No. of Points="&str(N)
89: wrt "P.1","Integ. Time=",T[T]
90: wrt "P.1","Mean=",M
91: wrt "P.1","Std. Deviation=",S

```

Table 2. Data Acquisition and Reduction Program (Continued)

```

92: wrt "P.1","Slope=",A
93: wrt "P.1","Y Intercept=",B
94: wrt "P";if flg2;gto "Loop"
95: gto "Out"
96:
97: "Setup":1)I;0)G)S)T)Z;fxd 0
98: for J=1 to 22;0)T[J];next J
99: str(I))P$;dsp "SAMPLE INTEGRATION TIME: T["&P$[2]&"]=?"
100: ent "",T[I]
101: if T[I]>99.99;jmp -2
102: if I<19 and not flg13;I+1)I;gto -3
103: ent "NUMBER OF BLOCKS OF SAMPLES?", T[20]
104: ent "NUMBER OF ITERATIONS?",T[21]
105: ent "DATA POINTS LISTED (N)?",P$
106: cap(P$)="N")W
107" gto "Loop"
108:
109: "GEN":G+1)G;if G>T[20];1)G
110: SS+Z)Z
111: if G#1;ret
112: if Z=0;gto +4
113: wrt "P.1","Avg std dev=",\ (Z/T[20]))Z
114: wrt "P.1","std dev \ T=",Z\T[T]
115: 0)S)Z
116: T+1)T;if T[T]=0;1)T;T[22]+1)T[22]
117: if T[22]<T[21];ret
118: gto "Bgn"
119:
120: "Time":wrt 8,"R"
121: red 8,T$
122: T$[7])P$
123: "/" )T$[3,3])T$[6,6];"84")T$[7,8]
124: P$&" "&T$[1,8])T$;ret
*4134

```




Fig. 6. Data Acquisition Flow Chart Using 9825T Calculator and 2401C IDVM

Table 3. Manual Mode Output

IDVM Ver 2.0 13:20:30 04/18/84
Noise meas. LM194 G=50 input term. 2k

13:23:48 04/18/84

-0.0284150	-0.0284150	-0.0284470	-0.0284260	-0.0284360	-0.0284100
-0.0283810	-0.0283860	-0.0283820	-0.0283860	-0.0283780	-0.0283610
-0.0283660	-0.0283520	-0.0283580	-0.0283810	-0.0284030	-0.0284130
-0.0284140	-0.0284250	-0.0284210	-0.0284090	-0.0284070	-0.0284170
-0.0284570	-0.0284530	-0.0284330	-0.0284420	-0.0284650	-0.0284760
-0.0284910	-0.0285020	-0.0284950	-0.0284710	-0.0284280	-0.0284080
-0.0284280	-0.0284430	-0.0284400	-0.0284630	-0.0284390	-0.0284450
-0.0284530	-0.0284460	-0.0284500	-0.0284650	-0.0284700	-0.0284680
-0.0284830	-0.0285100				

No. of Points= 50
Integ. Time= 1.0000000
Mean=-0.02843026
Std. Deviation= 0.00002863
Slope=-0.00000180
Y Intercept=-0.02838436




Fig. 7. Plot of Table 3 Data Set

Table 4. Automatic Mode Output

IDVM Ver 2.0 13:53:45 04/18/84
 noise meas. LM194 G=50 input term 2k

13:57:20 04/18/84

-0.0277580	-0.0277910	-0.0277790	-0.0277930	-0.0277530	-0.0277630
-0.0277700	-0.0277560	-0.0277540	-0.0277320	-0.0277090	-0.0277300
-0.0277310	-0.0277130	-0.0277310	-0.0277440	-0.0277570	-0.0277380
-0.0277420	-0.0277300	-0.0277700	-0.0277710	-0.0277900	-0.0277880
-0.0277980	-0.0277760	-0.0277630	-0.0277380	-0.0277490	-0.0277790
-0.0277990	-0.0277870	-0.0277590	-0.0277620	-0.0277690	-0.0277810
-0.0278140	-0.0277970	-0.0277790	-0.0278040	-0.0278020	-0.0277850
-0.0277900	-0.0278000	-0.0277960	-0.0278030	-0.0278110	-0.0278170
-0.0278140	-0.0277920				

No. of Points= 50
 Integ. Time= 1.00000000
 Mean=-0.02777114
 Std. Deviation= 0.00002185
 Slope=-0.00000120
 Y Intercept=-0.02774059

13:58:14 04/18/84

-0.0278050	-0.0277820	-0.0277710	-0.0277650	-0.0277630	-0.0278000
-0.0278040	-0.0278070	-0.0278060	-0.0277970	-0.0278200	-0.0278140
-0.0277790	-0.0277660	-0.0277910	-0.0278040	-0.0278250	-0.0278130
-0.0278110	-0.0277970	-0.0278230	-0.0278610	-0.0278770	-0.0278630
-0.0278800	-0.0278740	-0.0278620	-0.0278470	-0.0278400	-0.0278310
-0.0277890	-0.0277720	-0.0277830	-0.0277840	-0.0277580	-0.0277640
-0.0277580	-0.0277830	-0.0277980	-0.0277930	-0.0277960	-0.0277760
-0.0277650	-0.0277650	-0.0277450	-0.0277650	-0.0277670	-0.0277440
-0.0277370	-0.0277360				

No. of Points= 50
 Integ. Time= 1.00000000
 Mean=-0.02779712
 Std. Deviation= 0.00003508
 Slope= 0.00000083
 Y Intercept=-0.02781835

Table 4. Automatic Mode Output (Continued)

13:59:09 04/18/84

-0.0277200	-0.0277210	-0.0277020	-0.0276900	-0.0277180	-0.0277120
-0.0276900	-0.0276960	-0.0277090	-0.0277010	-0.0277020	-0.0277110
-0.0277120	-0.0276920	-0.0277000	-0.0277120	-0.0277260	-0.0277300
-0.0277330	-0.0277040	-0.0276890	-0.0276960	-0.0276880	-0.0277300
-0.0277170	-0.0277150	-0.0277250	-0.0277430	-0.0277500	-0.0277600
-0.0277420	-0.0277550	-0.0277430	-0.0277460	-0.0277440	-0.0277250
-0.0277270	-0.0277610	-0.0277570	-0.0277630	-0.0277430	-0.0277570
-0.0277250	-0.0277220	-0.0277370	-0.0277170	-0.0277060	-0.0276870
-0.0277150	-0.0277060				

No. of Points= 50
 Integ. Time= 1.00000000
 Mean=-0.02772144
 Std. Deviation= 0.00001960
 Slope=-0.00000066
 Y Intercept=-0.02770460

Avg std dev= 0.00002641
 std dev \ T= 0.00002641

14:00:53 04/18/84

-0.0276815	-0.0277020	-0.0277145	-0.0276870	-0.0277030	-0.0276930
-0.0276990	-0.0276830	-0.0276545	-0.0276370	-0.0276160	-0.0276735
-0.0276430	-0.0276385	-0.0276150	-0.0276045	-0.0275730	-0.0276010
-0.0275920	-0.0276230	-0.0276135	-0.0276135	-0.0276115	-0.0276320
-0.0276100	-0.0276210	-0.0276120	-0.0276130	-0.0276215	-0.0276045
-0.0276260	-0.0276410	-0.0276455	-0.0276620	-0.0276590	-0.0276485
-0.0276880	-0.0276970	-0.0276835	-0.0277220	-0.0277535	-0.0277580
-0.0277705	-0.0277730	-0.0277495	-0.0277220	-0.0277340	-0.0277630
-0.0277400	-0.0277080				

No. of Points= 50
 Integ. Time= 2.00000000
 Mean=-0.02766661
 Std. Deviation= 0.00004873
 Slope=-0.00000161
 Y Intercept=-0.02762565

Table 4. Automatic Mode Output (Continued)

14:02:38 04/18/84

-0.0276540	-0.0276790	-0.0276845	-0.0277025	-0.0276745	-0.0276795
-0.0276665	-0.0276665	-0.0276500	-0.0276640	-0.0276545	-0.0276675
-0.0276890	-0.0276810	-0.0276770	-0.0276520	-0.0276515	-0.0276280
-0.0275985	-0.0275895	-0.0275960	-0.0275860	-0.0275975	-0.0275810
-0.0275980	-0.0275675	-0.0275620	-0.0275580	-0.0275745	-0.0275660
-0.0275270	-0.0275575	-0.0275545	-0.0275450	-0.0275360	-0.0275210
-0.0275295	-0.0275390	-0.0275390	-0.0275325	-0.0274965	-0.0275115
-0.0275445	-0.0275365	-0.0275230	-0.0275365	-0.0275415	-0.0275375
-0.0275310	-0.0275665				

No. of Points= 50
 Integ. Time= 2.00000000
 Mean=-0.02759404
 Std. Deviation= 0.00002406
 Slope= 0.00000384
 Y Intercept=-0.02769185

14:04:23 04/18/84

-0.0275810	-0.0275540	-0.0275465	-0.0275230	-0.0275070	-0.0275305
-0.0275130	-0.0274780	-0.0274775	-0.0274865	-0.0274815	-0.0274710
-0.0274920	-0.0274850	-0.0274625	-0.0274700	-0.0274830	-0.0274905
-0.0275150	-0.0275495	-0.0275445	-0.0275660	-0.0275410	-0.0275485
-0.0275595	-0.0274965	-0.0274955	-0.0274875	-0.0275055	-0.0274895
-0.0274840	-0.0275015	-0.0275100	-0.0275250	-0.0274680	-0.0274555
-0.0274810	-0.0275005	-0.0275055	-0.0274785	-0.0274630	-0.0274630
-0.0274765	-0.0274760	-0.0274695	-0.0274580	-0.0274555	-0.0274340
-0.0274095	-0.0274005				

No. of Points= 50
 Integ. Time= 2.00000000
 Mean=-0.02749492
 Std. Deviation= 0.00003091
 Slope= 0.00000150
 Y Intercept=-0.02753304

Avg std dev= 0.00003610
 std dev \ T= 0.00005105

Table 4. Automatic Mode Output (Continued)

14:05:18 04/18/84

-0.0274530	-0.0274590	-0.0274640	-0.0274520	-0.0274280	-0.0274400
-0.0274540	-0.0274820	-0.0275000	-0.0274840	-0.0274950	-0.0275160
-0.0275150	-0.0275020	-0.0275200	-0.0275150	-0.0275190	-0.0274940
-0.0274880	-0.0275250	-0.0275490	-0.0275550	-0.0275530	-0.0275790
-0.0275800	-0.0275440	-0.0275290	-0.0274790	-0.0274900	-0.0275000
-0.0275200	-0.0275090	-0.0274950	-0.0274790	-0.0274930	-0.0275020
-0.0275060	-0.0275190	-0.0275300	-0.0275230	-0.0275360	-0.0275510
-0.0275450	-0.0275480	-0.0275490	-0.0275370	-0.0275360	-0.0275450
-0.0275460	-0.0275410				

No. of Points= 50

Integ. Time= 1.00000000

Mean=-0.02751146

Std. Deviation= 0.00002718

Slope=-0.00000154

Y Intercept=-0.02747222

14:06:12 04/18/84

-0.0275210	-0.0274870	-0.0274780	-0.0274990	-0.0275030	-0.0275050
-0.0275190	-0.0275320	-0.0275260	-0.0275340	-0.0275340	-0.0275550
-0.0275620	-0.0275430	-0.0275350	-0.0275530	-0.0275600	-0.0275520
-0.0275680	-0.0275850	-0.0275810	-0.0275500	-0.0275390	-0.0275240
-0.0275250	-0.0275500	-0.0275270	-0.0275470	-0.0275430	-0.0275630
-0.0275910	-0.0275960	-0.0275820	-0.0275760	-0.0275770	-0.0275750
-0.0275740	-0.0275640	-0.0275530	-0.0275260	-0.0275590	-0.0275880
-0.0275970	-0.0275770	-0.0275760	-0.0275580	-0.0275370	-0.0275450
-0.0275510	-0.0275420				

No. of Points= 50

Integ. Time= 1.00000000

Mean=-0.02754888

Std. Deviation= 0.00002237

Slope=-0.00000114

Y Intercept=-0.02751981

Table 4. Automatic Mode Output (Continued)

14:07:07 04/18/84

-0.0275590	-0.0275550	-0.0275580	-0.0275690	-0.0276030	-0.0276100
-0.0276040	-0.0276230	-0.0276040	-0.0276050	-0.0276090	-0.0275770
-0.0275770	-0.0275780	-0.0275570	-0.0275290	-0.0275200	-0.0275380
-0.0275520	-0.0275530	-0.0275660	-0.0275620	-0.0275820	-0.0275840
-0.0275880	-0.0275870	-0.0275850	-0.0276010	-0.0275900	-0.0275830
-0.0275550	-0.0275630	-0.0275400	-0.0275170	-0.0275130	-0.0275140
-0.0275380	-0.0275460	-0.0275180	-0.0275240	-0.0275470	-0.0275720
-0.0275980	-0.0276220	-0.0276540	-0.0276670	-0.0276330	-0.0276220
-0.0276250	-0.0276090				

No. of Points= 50
 Integ. Time= 1.00000000
 Mean=-0.02757570
 Std. Deviation= 0.00003667
 Slope=-0.00000030
 Y Intercept=-0.02756801

Avg std dev= 0.00002935
 std dev \ T= 0.00002935

14:08:52 04/18/84

-0.0275745	-0.0275860	-0.0275935	-0.0275765	-0.0275550	-0.0274985
-0.0275025	-0.0275390	-0.0275710	-0.0275700	-0.0275755	-0.0275945
-0.0276145	-0.0275840	-0.0275950	-0.0275860	-0.0275715	-0.0275800
-0.0275800	-0.0275460	-0.0275770	-0.0276015	-0.0276040	-0.0275810
-0.0275720	-0.0275985	-0.0276100	-0.0275825	-0.0275670	-0.0275790
-0.0275920	-0.0275705	-0.0275815	-0.0276005	-0.0276025	-0.0275815
-0.0275815	-0.0276095	-0.0276030	-0.0275945	-0.0275760	-0.0275620
-0.0275780	-0.0275735	-0.0275745	-0.0275615	-0.0275680	-0.0275665
-0.0275790	-0.0275860				

No. of Points= 50
 Integ. Time= 2.00000000
 Mean=-0.02757816
 Std. Deviation= 0.00002184
 Slope=-0.00000034
 Y Intercept=-0.02756950

Table 4. Automatic Mode Output (Continued)

14:10:36 04/18/84

-0.0275715	-0.0276125	-0.0276510	-0.0276415	-0.0276370	-0.0276460
-0.0276250	-0.0276145	-0.0275795	-0.0275225	-0.0275260	-0.0274825
-0.0274520	-0.0274545	-0.0274370	-0.0274660	-0.0274800	-0.0274910
-0.0274755	-0.0274850	-0.0274750	-0.0275183	-0.0275450	-0.0275660
-0.0275885	-0.0275910	-0.0275940	-0.0275690	-0.0275315	-0.0275320
-0.0275485	-0.0275655	-0.0275825	-0.0275610	-0.0275870	-0.0275615
-0.0275670	-0.0275275	-0.0275460	-0.0275820	-0.0275810	-0.0275600
-0.0275550	-0.0275685	-0.0275730	-0.0275955	-0.0276140	-0.0276255
-0.0276305	-0.0276545				

No. of Points= 50
 Integ. Time= 2.00000000
 Mean=-0.02755904
 Std. Deviation= 0.00005602
 Slope=-0.00000062
 Y Intercept=-0.02754323

14:12:21 04/18/84

-0.0276275	-0.0275875	-0.0276180	-0.0276250	-0.0276135	-0.0276160
-0.0275975	-0.0275870	-0.0275900	-0.0276170	-0.0276540	-0.0276600
-0.0276455	-0.0276545	-0.0276495	-0.0276310	-0.0275970	-0.0275945
-0.0275970	-0.0275915	-0.0276265	-0.0276165	-0.0276335	-0.0275905
-0.0275915	-0.0275650	-0.0275590	-0.0275830	-0.0275435	-0.0275500
-0.0275470	-0.0275610	-0.0275770	-0.0275565	-0.0275625	-0.0275700
-0.0275425	-0.0275590	-0.0275630	-0.0275425	-0.0275250	-0.0275445
-0.0275665	-0.0275405	-0.0275455	-0.0275470	-0.0275395	-0.0275735
-0.0275470	-0.0275210				

No. of Points= 50
 Integ. Time= 2.00000000
 Mean=-0.02758487
 Std. Deviation= 0.00002284
 Slope= 0.00000204
 Y Intercept=-0.02763691

Avg std dev= 0.00003714
 std dev \ T= 0.00005252
 IDVM Ver 2.0 14:12:24 04/18/84

III. INSTRUCTIONS FOR USING DATA ACQUISITION AND REDUCTION PROGRAM

The instructions for using the data acquisition and reduction program are described herein.

I. Load Tape

Press "LOAD", enter tape file number, press "EXECUTE".

II. Run Program

Press "RUN".

III. Operation of Program

A. 122'4 code(Y)? (see Note 1)

Enter "Y" and press "CONTINUE" for 122'4 code

or

Enter "N" and press "CONTINUE" or just "CONTINUE" for 1248 code.

B. Auto Mode (Y)?

Enter "Y" and press "CONTINUE" for automatic mode (see Note 2).

Enter "N" and press "CONTINUE" or just "CONTINUE" for manual mode (see Note 3).

1. Auto Mode

a. Label

Enter the title for which this test is to be conducted (up to 80 character positions) and press "CONTINUE".

b. Number of samples?

Enter number of samples desired and press "CONTINUE".

c. Sample integration time: T[1]=? (see Note 4)

Enter first integration time and press "CONTINUE".

d. Sample integration time: T[2]=?

Enter second integration time and press "CONTINUE".

e. Repeat as desired for a maximum of 19 integration times. Pressing "CONTINUE" instead of entering another integration time moves the program to the next step.

- f. Number of blocks of samples? (see Note 5)
Enter number of blocks of samples and press "CONTINUE".
- g. Number of iterations? (see Note 6)
Enter number of iterations and press "CONTINUE".
- h. Data points listed (N)?
Enter "N" and press "CONTINUE" (see Note 7)
or
Enter "Y" and press "CONTINUE" (see Note 8)
- i. Once "CONTINUE" is pressed the computer begins to collect data, automatically prints the number of samples, returns to step 1.a. ("Label") and waits for the next test sequence (see Note 9).

2. Manual Mode

- a. Label
Enter the title (up to 80 character positions), and press "CONTINUE".
- b. Number of samples?
Enter number of samples and press "CONTINUE".
- c. Sample integration time (sec)?
Enter integration time (≥ 0.01 to ≤ 99.99 sec) and press "CONTINUE". Data are now acquired.
- d. Plot/list/or next (P/L/N)?
Enter "P" and press "CONTINUE" for a plot of the data acquired in 2.c. vs. the number of the sample.
or
Enter "L" and press "CONTINUE" for a list of the data acquired in 2.c. (see Note 8)
or
Enter "N" and press "CONTINUE". This step returns to 2.a. for the next sample cycle
or
If no letters are used, the following information will be listed on 9825T calculator tape:
Number of points
Means
Standard deviation of means
Slope of best least squares fit straight line
Y-intercept of best least squares fit straight line
- e. (See Note 8).

NOTES:

1. The 2401C with option M21 has a 1248 code. Otherwise the 2401C has a 122'4 code.
2. In the automatic mode a wide variety of data taking sequences can be specified.
3. The manual mode is used for listing and/or plotting one block of data.
4. A maximum of 19 different integration times can be selected between > 0.01 to < 99.99 sec.
5. The number of blocks indicates the number of times data are repeated for each integration time. For example if two blocks are selected for integration times of 1, 2, and 5 sec (integration times are selected in l.c.), each of which has 50 samples (number of samples are selected in l.b.). The computer will accumulate data as follows

50 samples, 1 second integration time for each sample
50 samples, 1 second integration time for each sample
50 samples, 2 second integration time for each sample
50 samples, 2 second integration time for each sample
50 samples, 5 second integration time for each sample
50 samples, 5 second integration time for each sample

The total is six blocks of data, and each block is stored, processed, and outputted separately.

6. The number of iterations indicates the number of times the data sequence described in Note 5 is repeated. For example, if the three iterations were asked for the sequence described in Note 5 would be repeated 3 times for a total of 18 blocks of data. Each block would be stored, processed, and outputted separately.
7. The printer will list:
 - Number of points
 - Integration time
 - Mean of data
 - Standard deviation of the data about a linear least squares fit to the data
 - Slope of the linear least squares fit
 - Y intercept of the linear least squares fit
 - The standard deviation is computed as

$$\sigma = \left\{ \frac{1}{N-1} \sum_{i=1}^N [y_i - (mx_i + b)]^2 \right\}^{1/2}$$

where

N = number of data points

x_i, y_i = data set

m = slope of the linear least squares fit

b = Y intercept of the linear least squares fit

In Appendix C is discussed the effect of the time-bandwidth product of the input voltage on the calculation of σ .

8. The printer will list each measured integrated voltage (i.e., each mean) and also the information listed in Note 7 above.
9. At this time or any other time if one desires to get into the manual mode or from a manual mode into an auto mode or simply to get back to the beginning of the program because of an error, press "STOP" and press "RUN". The computer will advance to the beginning of the program which is step III.A., which asks the question "122'4 code (Y)?".

IV. 2401C INTEGRATING DIGITAL VOLTMETER CONTROL SETTINGS
AND CONNECTIONS TO THE INTERFACE BUS

- I. Front Panel of 2401C
 - A. Range: Depending on the amplitude of the input voltage, select one of five positions: 0.1, 1.0, 10, 100, or 1000 V.
 - B. Function: Set to "VOLT" position.
 - C. Sample Period: Set to "EXT SEL" position.
 - D. Sample Rate: Setting not important.
 - E. Attenuation: Setting not important.
- II. Back Panel of 2401C
 - A. STORE DURING COUNT/DISPLAY DURING COUNT: Either position.
 - B. INT/EXT: Either position.
 - C. Program control (J1 connector): A coax is connected from the "START" BNC fitting on the interface box to J1 with the center conductor going to J1-b and the shield to J1-a and J1-z. A special coax is furnished with the interface box to make these connections.
 - D. Counter reset: Connect a coax from counter reset to the "RESET" input on the interface box.
 - E. BCD output (J2 connector): Connect to the interface box "BCD in" with 50 pin ribbon cable supplied with the interface box.
 - F. FREQ input: No connection.
 - G. 100 KHz STD, INPUT/OUTPUT: No connection.
 - H. 100. KHz STD, INT/EXT: Either position.
- III. Interface Box to Calculator
 - A. Connect HP-IB from calculator to the interface box GPIB input.

APPENDIX A

REJECTION OF 60 Hz AND HARMONICALLY RELATED COMPONENTS

The addition of unwanted 60 Hz (i.e., pickup voltages) and harmonically related components to the wanted signal is a common laboratory problem. Depending on the integration time chosen the 2401C will provide a varying degree of rejection to the unwanted signal. Given an input unwanted voltage of amplitude V_u at frequency f_u , we can define a measure of rejection R as

$$R = \frac{\sigma_1}{\sigma_o(T)} \quad (A-1)$$

where

- σ_1 = standard deviation of the unwanted voltage = $V_u/\sqrt{2}$
- $\sigma_o(T)$ = standard deviation of the output mean voltage for an integration time of T and caused by the unwanted voltage

The output mean voltage $V_o(T)$ for an integration time of T and caused by the unwanted voltage is given by

$$V_o(T) = \frac{V_u}{\Delta\theta} \int_{\theta}^{\theta+\Delta\theta} \sin 2\pi f_u t \, d(2\pi f_u t) = \frac{V_u}{\Delta\theta} [\cos \theta - \cos(\theta + \Delta\theta)] \quad (A-2)$$

where $\Delta\theta = (T/\frac{1}{f_u}) \times 2\pi$ radians

The variance of $V_o(T)$ is, by definition,

$$\text{Var } V_o(T) = E[V_o - E(V_o)]^2 \quad (A-3)$$

where $E[]$ denotes mathematical expectation. Substituting (A-2) into (A-3), and noting that $E(V_o) = 0$ we have

$$\begin{aligned} \text{Var } V_o(T) &= \left(\frac{V_u}{\Delta\theta}\right)^2 E [\cos^2 \theta + \cos^2(\theta + \Delta\theta) - 2 \cos \theta \cos(\theta + \Delta\theta)] \\ &= \left(\frac{V_u}{\Delta\theta}\right)^2 \left\{ \frac{1}{2} + \frac{1}{2} - 2E[\cos \theta \cos(\theta + \Delta\theta)] \right\} \end{aligned}$$

$$\begin{aligned}
&= \left(\frac{V_u}{\Delta\theta}\right)^2 \left[1 - 2 \times \frac{1}{2\pi} \int_0^{2\pi} (\cos^2\theta \cos \Delta\theta - \cos \theta \sin \theta \sin \Delta\theta) d\theta\right] \\
&= \left(\frac{V_u}{\Delta\theta}\right)^2 (1 - \cos \Delta\theta) \tag{A-4}
\end{aligned}$$

From which we get

$$\sigma_o(T) = [\text{Var } V_o(T)]^{1/2} = V_u \frac{(1 - \cos \Delta\theta)^{1/2}}{\Delta\theta} \tag{A-5}$$

and finally

$$R = \frac{\sigma_i}{\sigma_o(T)} = \frac{\Delta\theta}{\sqrt{2}(1 - \cos \Delta\theta)^{1/2}} \tag{A-6}$$

A plot of R vs. $\Delta\theta$ is given in Fig. A-1. Equation (A-6) was checked experimentally by putting a known 60 Hz voltage into the 2401C and then varying the sampling time. In Table A-1 are shown the results for a constant input of 10.3 mV rms for integration times of 10 through 190 ms every 10 ms. For each integration time 100 samples were taken, e.g., for an integration time of 40 ms, this means we would take 100 averages and each would be 40 ms long. As can be seen from the table the agreement between the calculated and experimental values of R are within about 5%, the experimental value being consistently higher, probably due to some systematic error, e.g., an error in the 10.3 mV rms measurement.




Fig. A-1. Rejection vs. $\Delta\theta$

Table A-1. Experimental Determination of R as a Function of Integration Time (T)

T (ms)	R (predicted)	σ_o (mV rms)	$10.3/\sigma_o$
10	1.98	5.004	2.06
20	6.41	1.557	6.62
30	9.62	1.017	10.13
40	7.93	1.214	8.48
50	-	0.0125	0.824×10^3
60	11.89	0.8081	12.75
70	22.45	0.4449	23.15
80	25.66	0.3881	26.54
90	17.84	0.5615	18.34
100	-	0.00301	3.42×10^3
110	21.80	0.4446	23.17
120	38.48	0.2592	39.74
130	41.69	0.2363	43.59
140	27.75	0.3544	29.06
150	-	0.00263	3.92×10^3
160	31.71	0.3114	33.08
170	54.52	0.1831	56.25
180	57.72	0.1706	60.38
190	37.66	0.2645	38.94

APPENDIX B

QUANTIZATION LEVEL OF THE 2401C

The 2401C consists of a very precise voltage to frequency converter (VFC) and a counter with a six digit display. The highest frequency out of the VFC is 310 kHz. In general, the number of counts, C, displayed on the counter is given by

$$C = \frac{10^5 V_1}{S} \times T$$

where

V_1 = input voltage

S = voltage scale setting, choice of 0.1, 1, 10, 100, or 1000 V

T = counting time, choice of 0.01, 0.1, or 1.0 sec or can be externally controlled

Since the maximum frequency out of the VFC is 310 kHz, this means that the maximum allowed value of V_1 is $3.1 \times S$. Values above that level will result in an overload indication. The maximum quantization level if the internal time base is used is then 1 part in 310,000 for a 1 sec count, 1 part in 31,000 for a 0.1 sec count, and 1 part in 3100 for a 0.01 sec count. If the count time is controlled externally, the maximum quantization is 1 part in 999,999. This of course requires that T be at least $1/0.31 = 3.2258$ sec.

APPENDIX C

EFFECT ON THE CALCULATION OF σ OF THE TIME-BANDWIDTH PRODUCT OF THE INPUT VOLTAGE

For an ideal low-pass input bandwidth of B Hz and for an integration time of T , the error in σ has been shown¹ to be about 5, 2.5, and 1% for $BT = 1, 2,$ and 4 , respectively. To eliminate problems with overload of the 2401C caused by large high-frequency inputs, e.g., as might be caused by sharp spike pickup, we commonly low pass the input to the 2401C. Therefore, in order to make accurate calculations of σ , we choose the low-pass bandwidth B such that $BT \gg 1$.

¹W. A. Johnson, "The Effects of a Finite Time-Bandwidth Product on the Variance of the Sample Mean of a Low-Pass Stationary Random Process," IEEE Transactions on Information Theory, Vol. IT-19, No. 1, January 1973, pp. 115-117. See Figure 2.

LABORATORY OPERATIONS

The Laboratory Operations of The Aerospace Corporation is conducting experimental and theoretical investigations necessary for the evaluation and application of scientific advances to new military space systems. Versatility and flexibility have been developed to a high degree by the laboratory personnel in dealing with the many problems encountered in the nation's rapidly developing space systems. Expertise in the latest scientific developments is vital to the accomplishment of tasks related to these problems. The laboratories that contribute to this research are:

Aerophysics Laboratory: Launch vehicle and reentry fluid mechanics, heat transfer and flight dynamics; chemical and electric propulsion, propellant chemistry, environmental hazards, trace detection; spacecraft structural mechanics, contamination, thermal and structural control; high temperature thermomechanics, gas kinetics and radiation; cw and pulsed laser development including chemical kinetics, spectroscopy, optical resonators, beam control, atmospheric propagation, laser effects and countermeasures.

Chemistry and Physics Laboratory: Atmospheric chemical reactions, atmospheric optics, light scattering, state-specific chemical reactions and radiation transport in rocket plumes, applied laser spectroscopy, laser chemistry, laser optoelectronics, solar cell physics, battery electrochemistry, space vacuum and radiation effects on materials, lubrication and surface phenomena, thermionic emission, photosensitive materials and detectors, atomic frequency standards, and environmental chemistry.

Computer Science Laboratory: Program verification, program translation, performance-sensitive system design, distributed architectures for spaceborne computers, fault-tolerant computer systems, artificial intelligence and microelectronics applications.

Electronics Research Laboratory: Microelectronics, GaAs low noise and power devices, semiconductor lasers, electromagnetic and optical propagation phenomena, quantum electronics, laser communications, lidar, and electro-optics; communication sciences, applied electronics, semiconductor crystal and device physics, radiometric imaging; millimeter wave, microwave technology, and RF systems research.

Materials Sciences Laboratory: Development of new materials: metal matrix composites, polymers, and new forms of carbon; nondestructive evaluation, component failure analysis and reliability; fracture mechanics and stress corrosion; analysis and evaluation of materials at cryogenic and elevated temperatures as well as in space and enemy-induced environments.

Space Sciences Laboratory: Magnetospheric, auroral and cosmic ray physics, wave-particle interactions, magnetospheric plasma waves; atmospheric and ionospheric physics, density and composition of the upper atmosphere, remote sensing using atmospheric radiation; solar physics, infrared astronomy, infrared signature analysis; effects of solar activity, magnetic storms and nuclear explosions on the earth's atmosphere, ionosphere and magnetosphere; effects of electromagnetic and particulate radiations on space systems; space instrumentation.

END

FILMED

1-86

DTIC