

# Unclassified

# Space Shuttle HCl Gas Detection

M. G. Scott and C. W. Pender, Jr. Sverdrup Technology, Inc.

May 1987

Final Report for Period April 15, 1984 - May 15, 1985

3

/47-PDC

SER A

<u>و</u> ب





Unclassified

### NOTICES

.

When U. S. Government drawings, specifications, or other data are used for any purpose other than a definitely related Government procurement operation, the Government thereby incurs no responsibility nor any obligation whatsoever, and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise, or in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Qualified users may obtain copies of this report from the Defense Technical Information Center.

References to named commercial products in this report are not to be considered in any sense as an endorsement of the product by the United States Air Force or the Government.

This report has been reviewed by the Office of Public Affairs (PA) and is releasable to the National Technical Information Service (NTIS). At NTIS, it will be available to the general public, including foreign nations.

#### APPROVAL STATEMENT

This report has been reviewed and approved.

Bradley le Biekn

BRADLEY A. BIEHN, Captain, USAF Facility Technology Division Directorate of Technology Deputy for Operations

Approved for publication:

FOR THE COMMANDER

anin L. garty

MARION L. LASTER Director of Technology Deputy for Operations

UNCLASSIFIED SECURITY CLASSIFICATION OF THIS PAGE

REPORT DOCUMENTATION PAGE					
1. REPORT SECURITY CLASSIFICATION		15. RESTRICTIVE MARKINGS			
28 SECURITY CLASSIFICATION AUTHORITY		3 DISTRIBUTION/AVAILABILITY OF REPORT			
26 DECLASSIFICATION/DOWNGRADING SCHEDULE		SEE REVERSE OF THIS PAGE			
4 PERFORMING ORGANIZATION REPORT NUM	BER(S)	5. MONITORING OR	GANIZATION RE	PORT NUMBER(S)	
AEDC-TR-85-52					
6. NAME OF PERFORMING ORGANIZATION Arnold Engineering Development Center	6b. OFFICE SYMBO⊾ <i>(If applicable)</i> DOT	75. NAME OF MONIT	ORING ORGANI	ZATION	
6c. ADDRESS (City, State and ZIP Code)	•	76 ADDRESS (City,	State and ZIP Cod	e)	
Air Force Systems Command Arnold Air Force Station, TN 37389-5000					
SA NAME OF FUNDING/SPONSORING ORGANIZATION AIr Force	85 OFFICE SYMBOL (If applicable)	9 PROCUREMENT I	NSTRUMENT ID	ENTIFICATION NU	IMBER
Engineering Services Center	RDVS				
Tundall Ain Canas Pasa - 27-mid	- 22/02	PROGRAM	PROJECT	TASK	WORK UNIT
Iynaall Alr Force Base, Florid	ia 32403	ELEMENT NO	NÖ	NO	ND
11 TITLE (Include Security Classification)	<u></u> .	921A67		ĺ	
SEE REVERSE OF THIS PAGE					ł
Scott, M. G. and Pender, C. W.	, Jr., Sverdrup	Technology, J	Inc., AEDC	Group	
136 TYPE OF REPORT 136 TIME C Final FROM 4/1	оvered 5/84 <sub>то</sub> 5/15/85	14. DATE OF REPORT May 1987	ητ (Yr Mo., Day)	15. PAGE CO	DUNT
16. SUPPLEMENTARY NOTATION					
Available in Defense Technic	al Information	Center (DTIC).			
17 COSATI CODES	18 SUBJECT TERMS IC	ntinue on reverse if re	cessory and identi	/y by block number, colid_ny	, conellant
FIELD GROUP SUB GR	HUI CONCENTRAL   Fourier transfi	orm data aci	nutcie nuisition	rocket	t boosters
	spectrometer	combust	ion byprodu	cts	
19 ABSTRACT (Continue on reverse if necessary and	i identify by block number				
Hydrogen chloride (HCl) gas is	one of the byp	roducts products	ced by the	Space Shuttl	e
technique to measure concentra	itions of daseou	s HCl for four	r davs foll	owing Shuttl	e
launches 41D and 51A at Kenned	ly Space Center	(KSC). The p	eak concent	rations	
measured near ground level abo	out 400 m from t	he launch str	ucture were	found to be	9
less than 10 ppm averaged over	ra 500-m patna	rter both lau	ncnes.		
20. DISTRIBUTION/AVAILABILITY OF ABSTRA	т	21 ABSTRACT SEC	JAITY CLASSIFI	CATION	
UNCLASSIFIED/UNLIMITED		UNCLASSIFI	ED		
22. NAME OF RESPONSIBLE INDIVIDUAL		226 TELEPHONE N	UMBER	22c OFFICE SYM	BOL
W. O. Cole		(615)454-781	3	DOS	
DD FORM 1473, 83 APR	EDITION OF 1 JAN 73	SOBSOLETE	[ [N		
			SECURI	TY CLASSIFICATI	ON OF THIS PAGE

## UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE

## 3. DISTRIBUTION/AVAILABILITY OF REPORT

Approved for public release; distribution is unlimited.

11. TITLE

-

Space Shuttle HC1 Gas Detection

#### PREFACE

The work reported herein was conducted by the Arnold Engineering Development Center (AEDC), Air Force Systems Command (AFSC), at the request of Air Force Engineering Services Center (AFESC/RDVS), Tyndall Air Force Base, Florida, and the Atmospheric Effects Branch, Marshall Space Flight Center, NASA H-73999B. The reported measurements were conducted and results obtained by Sverdrup Technology, Inc., AEDC Group, operating contractor for the propulsion test facilities at the AEDC, AFSC, Arnold Air Force Station, Tennessee, under Project Number DB07EW. The Air Force Project Managers were Captain Frank Tanji and Captain Bradley Biehn, AEDC/DOTR. The data analysis was completed on May 15, 1985, and the manuscript was submitted for publication on July 1, 1985.

## CONTENTS

## Page

1.0	INTRODUCTION	5
2.0	TECHNIQUE	6
3.0	APPARATUS	7
4.0	CALIBRATIONS AND ANALYSIS	8
5.0	RESULTS	10
6.0	SUMMARY	11
7.0	RECOMMENDATIONS	11
	REFERENCES	12

## **ILLUSTRATIONS**

## Figure

.

1.	Kennedy Space Center 39A Launch Pad	13
2.	Sample Spectrum Showing HC1 Fundamental Absorption Band	
	a. Covering 3,200 to 2,200 cm <sup>-1</sup>	14
	b. Showing Only the P Branch and How I <sub>o</sub> and I Are Determined	15
3.	P(1) Absorption Line	
	a. Calculated	16
	b. Convolved with ILS	16
4.	51A Launch Spectra	
	a. Typical Spectrum Obtained After the 51A Launch	17
	b. Background Spectrum Obtained Before the 51A Launch	18
5.	FTS Schematic	
	a. Calibration Setup	19
	b. Test Setup	19
6.	Source Configuration	
	a. 41D	20
	b. 5)A	20
7.	Calibration Curves from HC1 Cells	
	a. P(1) Line	21
	b. P(2) Line	21
	c. P(3) Line	22
	d. P(4) Line	22

3

## Page

## Figure

	e. P(5) Line	23
	f. P(6) Line	23
8.	Summary of 41D HC1 Concentrations Measured Using Several	
	P-Branch Absorption Lines	24
9.	Summary of 51A HC1 Concentrations Measured Using Several	
	P-Branch Absorption Lines	25
10.	51A HC1 Concentration Using P(5) Line for	
	10 hr Following Launch	26

Page

.

#### **1.0 INTRODUCTION**

Large amounts of gaseous hydrogen chloride (HCl) are one of the combustion by-products generated during the launch of the Space Transportation System (Shuttle) by the Shuttle's solid-propellant rocket boosters (SRB's). Significant quantities of HCl are present in the plume cloud that lingers near the launch structure. The HCl gas remains entrained in the atmosphere near the launch facility, is dispersed by wind, and is deposited on surfaces, such as grass or ground. (See Ref. 1 for deposition mechanisms.) The HCl gas that remains in the vicinity of the launch structure, from either initial release or secondary ground release, is important because it corrodes metals and electronic equipment and because it is a health hazard. The problem is currently restricted to the Kennedy Space Center (KSC) launch area but will also be present at Vandenberg when the Shuttle Western Test Range launch facility is activated. The problem may be exacerbated at Vandenberg because of differences in the launch mount, sound-suppression water systems, and the proximity of the launch facilities.

The objective of this project was to develop and apply a nonobtrusive absorption technique employing a Fourier Transform Spectrometer (FTS) to monitor the concentration of gaseous HCl present near the KSC launch pad for several days after a Shuttle launch. Because of safety considerations and the requirement to collect data immediately after a launch, the measurement technique was automated to permit untended operation. The technique was applied for launches 41D and 51A, Pad 39A (Fig. 1), which occurred on August 30, 1984 and November 8, 1984, respectively, at KSC.

To accomplish the objective, a technique based on the absorption of infrared (IR) radiation by the diatomic HCl molecules was developed. The characteristic HCl lines are attributable to absorption of radiation associated with vibrational and rotational transitions within the HCl molecule. A simple model (Ref. 2) to describe this absorption process is a molecule in which the individual atoms, held together by chemical bonds, are in vibratory motion along these bonds, while the entire molecule is rotating. The HCl is in a state of vibratory motion brought about by the alternate stretching and contracting of the chemical bond as the hydrogen and chlorine atoms move away and toward each other, respectively. This vibratory motion is superimposed on a rotation of the molecule about an axis perpendicular to the chemical bond. When IR radiation of the proper frequency (i.e. energy) impinges on the molecule and is absorbed, the vibration and/or rotation states are changed. These changes must satisfy certain selection rules that give rise to discrete absorption lines. These lines are labeled according to their frequency. The lines with frequencies greater than that of the band center are said to be in the R branch, whereas those with frequencies less than that of the band center are in the P branch (Fig. 2). The individual lines making up the P and R branches are identified as P(1), P(2), etc. as the frequency of the lines moves away from the band center.

#### 2.0 TECHNIQUE

A nonobtrusive IR absorption technique was used for monitoring gaseous HCl concentrations over long paths near the Shuttle launch pad (Fig. 1). The technique involved use of an IR spectrometer located approximately 500 m from an IR source. An IR spectrum was acquired at regular intervals (20 min for 41D, 10 min for 51A) for several days following the Shuttle launches. The spectra were then inspected to find the magnitude of absorption because of the presence of gaseous HCl. The concentration of HCl was then calculated using the absorption coefficient measured in the laboratory. The spectral line used to determine the amount of absorption is in the band centered at 2,886 cm<sup>-1</sup>. This technique is not dependent on absolute magnitudes of the spectra; therefore, the intensity of the source is not relevant to the measurement, and only relative intensities will be presented in this report.

Naturally occurring HCl contains two isotopes of chlorine. They are <sup>35</sup>Cl and <sup>37</sup>Cl, in the ratio of 3 to 1, respectively. The IR absorption lines of H<sup>35</sup>Cl and H<sup>37</sup>Cl are separate, but very close. A spectrum of these isotopic species is shown in Fig. 2b.

The absorption attributable to HCl gas is well understood and documented. The HCl absorption lines are described by (Ref. 3)

$$I(\lambda) = I_{o}(\lambda) \exp \left\{-SP\gamma_{o}L/[(\lambda - \lambda_{o})^{2} + \gamma_{o}^{2}]\pi\right\}$$

where

I<sub>o</sub> ( $\lambda$ ) is the intensity of the incident radiation I ( $\lambda$ ) is the intensity of the transmitted radiation S is the line strength of the absorption line (cm<sup>-2</sup> atm<sup>-1</sup>) P is the partial pressure of the absorbing species (atm)  $\gamma_0$  is the pressure-broadened half-width at half-maximum of the HCl line (cm<sup>-1</sup>)  $\lambda_0$  is the line center of the absorption line (cm<sup>-1</sup>) L is path length

Figure 3 shows the P(1) lines calculated using parameters contained in Ref. 4. The figure shows the theoretical absorption assuming 6 ppm HCl over a 500-m path. Because of limited resolution and the apodization function, the Fourier transform spectrometer (FTS) that produced the actual spectra influences the spectra shape. The convolution of the instrument line shape (ILS), using 0.5-cm<sup>-1</sup> resolution and a triangular apodization function, produces the P(1) line shown in Fig. 3b. As can be seen by comparing Figs. 3a and b, the ILS shortens

and broadens the spectral lines. (Note: The scales of these plots are not the same.) The effect is detrimental in that the resultant spectrum is not an accurate portrayal of the true line shape. However, it is still possible to use the resultant spectra since the effect can be calibrated. The calibration will be discussed later in this report.

A complicating factor was the presence of additional absorption lines in the spectra obtained in the field. The effect of neighboring absorption lines is to give the impression of a greater amount of gaseous HCl. The spectral region of interest included absorption lines attributable to other molecules present in the atmosphere such as H<sub>2</sub>O and, Hydrogen Deuterium Oxide (HDO) (Ref. 5). Therefore, the spectra were obtained at 0.5-cm<sup>-1</sup> resolution to separate the HCl lines from the lines of interference. In Fig. 4a the HCl doublets can be easily discerned when compared to a background spectrum with no HCl present, Fig. 4b. Figure 4a is a portion of a typical spectrum acquired during the 51A launch. This figure shows only the lines within the P branch. The R branch is unusable because of domination by water vapor bands at 2.7 and 3.2  $\mu$ m. The P(5) line was used to determine the HCl concentrations because it had the least interference.

## **3.0 APPARATUS**

The Shuttle launches took place at KSC Pad 39A. Figure 1 shows the major features of the launch pad and the AEDC equipment location. The test apparatus consisted of an FTS and a collimated light source with associated control and data acquisition instrumentation. The sample path length was 500 m. The distance to the launch structure was approximately 350 m. The equipment was located so that the plume deflected by the flame trench passed through the center of the sample path.

The FTS used was a Block Model RS197 field-rated instrument (Refs. 6 and 7). It was configured to have the maximum sensitivity possible in the spectral region of interest, 3,000 to 2,700 cm<sup>-1</sup>. This was accomplished by using a germanium beamsplitter on a potassium bromide substrate and an indium antimonide detector cooled to 77 K. The detector dewar had a maximum hold time of 6 hr, so an automatic fill system was developed to replenish the liquid nitrogen (LN<sub>2</sub>) from a large auxiliary reservoir. The FTS was rigidly mounted to maintain optical alignment with the source during the large vibrations associated with the Shuttle launch. The instrument was enclosed in a Plexiglas<sup>®</sup> container that was kept at a slight positive pressure with a gaseous nitrogen (N<sub>2</sub>) purge to reduce the possibility of HCl corrosion (Fig. 5b). The IR radiation entered the FTS through a hole in the Plexiglas container that matched the entrance aperture of the FTS. An O-ring seal was made between the FTS and the container. In this way, the FTS was in the same configuration in the field test as in the calibration. The only difference in the optical paths was that the quartz sample cell was fitted with IR transmisive quartz windows on each end.

The source units used in launches 41D and 51A were different. Modifications were made to the equipment for launch 51A because of insight gained during the 41D launch. The original configuration (See Fig. 6a) used in the 41D launch consisted of a Barnes Model 112017 blackbody with collimator and an Optronics Model 100U, 1,000-w tungsten halogen lamp with a 25.4-cm Cassegrain telescope used as a collimator. The use of multiple sources was required to provide adequate intensity since the FTS was located 500 m distant and also as insurance against optical misalignment from the shock of the booster ignition. In the second configuration (See Fig. 6b), for the 51A launch, three, 30.48-cm spherical mirrors with 1,000-w tungsten halogen lamps at their focal points were used to provide an intensity increase of 400 percent. In each case the sources were securely mounted and covered with a hood that provided protection from inclement weather and SRB plume debris. The sources were mounted so that the optical path between the FTS and sources was approximately 1.2 m above the ground (Fig. 6b).

The digital data acquisition system collected a set of 124 interferograms (the raw data) during each sampling period of approximately 2 min. Each set was averaged and stored on magnetic tape as one data point. A spectrum is obtained by taking the Fourier transform of the interferogram (Ref. 6). For launch 41D, data points were collected every 20 min. This time interval was reduced to 10 min for 51A after improvements were made to the software that reduced the processing time from 17 to 8 min. The data were not converted to the spectral domain until later in order to permit the collection of additional data during the postlaunch period.

## 4.0 CALIBRATION AND ANALYSIS

The absorption technique involved the comparison of spectral data acquired in the field with similar data acquired in the laboratory. As explained in Section 2.0, calibration data generated in the laboratory were used in lieu of published HCl absorption coefficients to negate the influence of the spectrometer instrument line shape.

In the laboratory (Fig. 5a), a previously evacuated quartz sample cell (2.54 diam by 25.4 cm) was filled with mixtures of HCl and dry N<sub>2</sub> with known concentrations simulating the absorption expected following the launch. The concentration of HCl in the sample cell is stated in terms of partial pressures (x) with units of psia, and the concentration (y) in the field is expressed as parts per million (ppm). The relationship is x/14.7 = y/1,000,000. Specific sample cell concentrations of HCl were achieved using a partial pressure technique in which

the cell was filled in several stages. Each stage consisted of partially evacuating the cell followed by bringing it back up to 14.7 psia with dry  $N_2$ . This process was continued until the desired partial pressure of HCl was reached. To simulate the field conditions, the concentrations used in the laboratory were greater by an amount dictated by Beer's law. A simple form of Beer's law,

$$\ln(I/I_0) = -ABC$$

where

I	= transmitted radiation
Io	= incident radiation
A	= absorptivity (constant at a given wavelength)
B	= path length
С	= concentration of the absorbing medium

states that the ratio of  $I/I_0$  will remain constant if C is varied inversely to B. For example,  $I/I_0$  is equivalent for a gas at either 1 psia over a 1-m path or for 10 psia over 0.1 m. Therefore, in the laboratory the concentrations needed to be greater by a factor of 1,968 (the ratio of the range in the field, 500 m, to the cell length, 25.4 cm).

A set of nine spectra corresponding to nine calibration pressures were obtained. The x axis of these curves is in terms of path-averaged concentrations over 500 m. Calibration curves were obtained by plotting  $\ln(I_o/I)$  versus x for six P lines. The plots are shown in Figs. 7a through f. The partial pressures of HCl, in psia, used in the test cell were 0.018, 0.036, 0.053, 0.080, 0.120, 0.181, 0.272, 0.544, and 1.633. The conversion relationship to the equivalent field concentration is given by y (ppm) = 34.6 x(psia).

The analysis of the KSC spectra consisted of measuring the difference between the baseline and the peak depth of the six P absorption lines. An important feature of this technique is that atmospheric haze does not invalidate the measurement. Atmospheric scattering attributable to haze does not affect the procedure since both I and I<sub>o</sub> (Fig. 2b) are influenced by the same amount, leaving the ratio  $I/I_o$  constant. This feature negates the necessity of performing an instrument or source calibration.

The error associated with the resultant measurements of the 41D launch is estimated to be  $\pm 19$  percent (root sum of the squares). This error is caused by an error of  $\pm 10$  percent because of pressure transducer uncertainty,  $\pm 10$ -percent gas-handling technique,  $\pm 10$  percent

in FTS measurement uncertainty, and  $\pm 8$  percent in data handling (digitization and processing). The error associated with the resultant measurements of the 51A launch is estimated to be  $\pm 15$  percent (root sum of the squares). This reduction in error is caused by the use of the alternate source unit with a higher intensity during the second measurement. The FTS measurement uncertainty was reduced to  $\pm 3$  percent, and the data handling uncertainty was reduced to  $\pm 2$  percent.

## 5.0 RESULTS

The application of the IR absorption technique following the 41D launch showed the presence of average concentrations of gaseous HCl as great as 4 ppm ( $\pm$ 19 percent). Figure 4a is a representative spectrum taken postlaunch. The HCl lines are clearly apparent among other atmospheric absorption lines. Figure 8 gives the concentration of gaseous HCl as yielded by examination of five H<sup>35</sup>Cl absorption lines. The spread in the results is attributed to the presence of absorption lines of other molecules. Note that in Fig. 8 the concentration determined using the P(5) line is less than concentrations determined using the other lines. In the case of most of the HCl absorption lines, their frequencies are nearly the same as other atmospheric constituents, such as HDO and H<sub>2</sub>O (Ref. 5). Therefore, the resultant absorption is greater than that caused only by the HCl. Since it is difficult to determine how much of the absorption is caused by the other molecules, the most isolated line, the P(5) line, was used to determine the HCl concentration.

The concentration peaked approximately 100 min postlaunch, remained high for nearly 1 hr, and then began to diminish. There unfortunately exists a gap in the data because of a period in which instrumentation was realigned following the loss of one of the sources during the high vibration experienced during the launch. The instrumentation was operated for three days following the launch, and HCl was detected only in the 6-hr interval following the launch.

Data obtained from the 51A launch are better in two ways. Data were acquired more frequently and with an improvement in the signal-to-noise ratio by nearly a factor of four. During the 51A launch, higher HCl concentrations were measured than were detected during the 41D launch. Figure 9 is a summary of six concentration curves from the P-branch lines. The curves have a significant scatter caused by absorption attributed to other molecules present in the sample path. The most isolated line again is the P(5) line. Figure 10 shows the best estimate of the HCl concentration. The HCl concentration peaked at about 9 ppm  $\pm$  15 percent approximately 1 hr postlaunch, fluctuated for about an hour, and then decayed. In the two days following the launch, HCl was detected in minute quantities after sunrise until early afternoon. Long-term, low levels of HCl were visible in the 51A launch either because of greater revolatilization attributable to atmospheric and ground conditions or better signalto-noise ratio. Atmospheric conditions during the measurement periods of both launches are available (Refs. 8 and 9).

### 6.0 SUMMARY

The objectives of the project were completely fulfilled. An unobtrusive absorption technique to monitor concentrations of gaseous hydrogen chloride was developed and then applied on two space shuttle launches at the Kennedy Space Center. The technique developed exploited the intrinsic property of gaseous HCl to absorb IR radiation in specific, narrow, spectral lines. This IR absorption is measured by a Fourier Transform Spectrometer that features accuracy and precision without a calibrated source. The minimum detectable concentration of the instrument is 0.10 ppm with an error of 15 percent (obtained by taking the square root of the sum of the squares of all possible errors or uncertainties in the measurement system including calibration). The application portion of this project enabled refinement of field use of the measurement technique. Increased source radiance and larger source collimating optics improved the signal-to-noise ratio by a factor of four. Improvement in data processing software decreased uncertainty by 4 percent.

Measured quantities of gaseous HCl following launches of both 41D and 51A displayed similar trends. The concentration after each launch first increased, then decreased, then increased to a maximum approximately 100 sec after the launch. The concentration decreased gradually to less than 1 ppm approximately 10 hr after each launch. The peak concentration measured was 4 and 9 ppm for 41D and 51A, respectively. For two days after the 51A launch only, detectable levels were recorded for a period of approximately 6 hr following sunrise. Atmospheric and ground conditions were different for each launch (Refs. 8 and 9). Additional considerations in interpreting the concentration levels are that the data are both path and time averaged, and that the ground scar from the plume makes up approximately 60 percent of the 500-m sample path length. These considerations imply that the concentration along the path is not uniform. Additionally only gaseous HCl will be detected by this technique.

#### 7.0 RECOMMENDATIONS

Improvements to the instrument developed on this project can be made in two areas. The signal can be increased by enlarging the collection optics on the FTS, and the noise can be reduced by increasing the data averaging period. Some improvement in calibration accuracy could be realized by using gas-mixture ratios prepared by the National Bureau of Standards.

The technique developed by this project could use computed tomography with multiple paths to generate a two-dimensional map of HCl concentration levels to determine areas of highest concentration and change over time. This technique would enable determination of safe areas around solid-propellant rocket motor launch complexes where people, wildlife, or machinery may be adversely impacted. Additional analysis (not within the scope of this project) should be conducted to determine the impact of atmospheric and ground conditions, aerosol entrapment of HCl, and flame trench cooling water runoff and holding pond contributions to gaseous HCl concentration levels.

#### REFERENCES

- Anderson, B. J. and Keller, V. W. "Space Shuttle Exhaust Cloud Properties." NASA TP-2258, December 1983.
- Szymanski, H. A. Theory and Practice of Infrared Spectroscopy. Plenum Press, New York, 1964.
- 3. Thorne, N. P. Spectraphysics. Chapman and Hall, London, 1974.
- Lin, C. L. et al. "Line Parameters of HCl Obtained by Simultaneous Analysis of Spectra." *Optical Spectroscopic Radiation Transfer*. Vol. 20. Pergamon Press Ltd., Great Britain, 1978.
- 5. Wolf, W. L., editor. Handbook of Military Infrared Technology. US Government Printing Office, Washington, D. C., 1965.
- 6. Bell, E. E. Introduction to Fourier Transform Spectrometry. Academic Press, New York, 1972.
- 7. Martin, A. E. Infrared Interferometric Spectrometers. Elsevier Scientific Publishing Company, New York, 1980.
- Johnson, D. L. et al. "Atmospheric Environment for Space Shuttle (STS-41D) Launch." NASA-TM-86484, October 1984.
- Johnson, D. L. et al. "Atmospheric Environment for Space Shuttle (STS-51A) Launch." NASA-TM-86497, December 1984.





13



## HC1 Absorption Sepctrum

a. Covering 3,200 to 2,200 cm<sup>-1</sup> Figure 2. Sample spectrum showing HCl fundamental absorption band.



b. Showing only the P branch and how  $I_0$  and I are determined Figure 2. Concluded.

15



cm<sup>-1</sup>

b. Convolved with the ILS Figure 3. P(1) absorption line.



a. Typical spectrum obtained after the 51A launch. Figure 4. 51A launch spectra.





18



a. Calibration setup



b. Test setup Figure 5. FTS schematic.







b. 51A Figure 6. Source configuration.







b. P(2) line Figure 7. Calibration curves for HCl cells.



c. P(3) line



d. P(4) line Figure 7. Continued.

![](_page_25_Figure_1.jpeg)

![](_page_25_Figure_2.jpeg)

![](_page_25_Figure_3.jpeg)

f. P(6) line Figure 7. Concluded.

![](_page_26_Figure_1.jpeg)

Figure 8. Summary of 41D HCl concentrations measured using several P-branch absorption lines.

![](_page_27_Figure_1.jpeg)

Time Following Launch in Minutes

Figure 9. Summary of 51A HCI concentrations measured using several P-branch absorption lines.

![](_page_28_Figure_1.jpeg)

Figure 10. 51A HCl concentration using P(5) line for 10 hr following launch.