


REPORT DOCUMENTATION PAGE

2

AD-A199 925

2b. RESTRICTIVE MARKINGS	
3. DISTRIBUTION / AVAILABILITY OF REPORT Approved for public release; distribution is unlimited	
2d. DECLASSIFICATION / DOWNGRADING SCHEDULE	
4. PERFORMING ORGANIZATION REPORT NUMBER(S) NMRI 88-9	
5. MONITORING ORGANIZATION REPORT NUMBER(S)	
6a. NAME OF PERFORMING ORGANIZATION Naval Medical Research	6b. OFFICE SYMBOL (if applicable)
7a. NAME OF MONITORING ORGANIZATION Naval Medical Command	
6c. ADDRESS (City, State, and ZIP Code) Bethesda, Maryland 20814-5055	
7b. ADDRESS (City, State, and ZIP Code) Department of the Navy Washington, D.C. 20372-5120	
8a. NAME OF FUNDING / SPONSORING ORGANIZATION Naval Medical Research and Development Command	8b. OFFICE SYMBOL (if applicable)
9. PROCUREMENT INSTRUMENT IDENTIFICATION NUMBER	
8c. ADDRESS (City, State, and ZIP Code) Bethesda, Maryland 20814-5055	
10. SOURCE OF FUNDING NUMBERS	
PROGRAM ELEMENT NO. 62233N	PROJECT NO. MM33C30
TASK NO. 006.1003	WORK UNIT ACCESSION NO. DN377031
11. TITLE (Include Security Classification) MICROCOMPUTER SOFTWARE AND INTERFACE FOR CONTROL OF A MICROSCOPE SCANNING STAGE	
12. PERSONAL AUTHOR(S) Benjamin R. Lee, Helmut Ludl, C. Patrick Reynolds	
13a. TYPE OF REPORT Report No. 8	13b. TIME COVERED FROM TO
14. DATE OF REPORT (Year, Month, Day) 1988	
15. PAGE COUNT 7	
16. SUPPLEMENTARY NOTATION Reprinted from: Analytical and Quantitative Cytology and Histology v. 10, n. 1, Feb. 1988 pp. 47-53	
17. COSATI CODES	
FIELD	GROUP SUB-GROUP
18. SUBJECT TERMS (Continue on reverse if necessary and identify by block number) Computers; Image Processing, Computer-Assisted; Software; Microcomputers; Microscopy; User-Computer Interface	
19. ABSTRACT (Continue on reverse if necessary and identify by block number)	
<p>Accession For</p> <p>NTIS GRA&amp;I <input checked="" type="checkbox"/></p> <p>DTIC TAB <input type="checkbox"/></p> <p>Unannounced <input type="checkbox"/></p> <p>Justification</p> <p>By _____</p> <p>Distribution/</p> <p>Availability Codes</p> <p>Dist Avail and/or Special</p> <p>A-1 20</p>	
20. DISTRIBUTION / AVAILABILITY OF ABSTRACT <input checked="" type="checkbox"/> UNCLASSIFIED/UNLIMITED <input type="checkbox"/> SAME AS RPT <input type="checkbox"/> DTIC USERS	
21. ABSTRACT SECURITY CLASSIFICATION Unclassified	
22a. NAME OF RESPONSIBLE INDIVIDUAL Phyllis Blum, Information Services Division	22b. TELEPHONE (include Area Code) 202-295-2188
22c. OFFICE SYMBOL ISD/ADMIN/NMRI	



# Microcomputer Software and Interface for Control of a Microscope Scanning Stage

Benjamin R. Lee  
Helmut Ludl  
C. Patrick Reynolds, M.D., Ph.D.

*Motorized scanning stages are valuable in microscopy systems that employ digital image analysis and for the development of semiautomatic computer-assisted microscope systems; the development of standard software "tools" to control such stages will facilitate their integration into a variety of computer-based systems. A set of Microsoft BASIC and Turbo PASCAL programs that interfaces a microprocessor-controlled stepper motor microscope stage (MDACE 1000) to an IBM PC or PC-AT or compatible microcomputer via a serial interface (RS-232) is described. These programs can be integrated into other software written in either BASIC or PASCAL, or used via a menu program that directs the routines to control*

*scanning patterns and to locate the microscope stage to a selected area of the slide. Coordinates of significant events on a slide can be stored on a disk file to allow future examination. The software and interface also provide control of a filter wheel in the microscope for use in multicolor fluorescence assays.*

Motorized microscope stage movement can significantly increase the power of cell analysis systems using digital image processing. Such stages are also valuable in microscopy without image analysis. Through automation, precise movement of the stage is achieved, thereby allowing uniform, complete and accurate scans of a slide.<sup>18-21</sup> Systems have been developed that use stepper motor stages to track the movement of living cells,<sup>8,22-37</sup> to trace and reconstruct nerve pathways,<sup>7,9,10,12,27,29,38</sup> to scan cervical preparations for dysplasia,<sup>15,26,28</sup> to scan blood smears to produce an automated differential blood count<sup>20,30,33,34</sup> and to perform other image analysis procedures.<sup>2,14,16</sup> Such systems were developed either as complete analyzers with dedicated computers or using various minicomputers. However, with constantly decreasing prices and increasing power, microcomputers are becoming attractive for developing image analysis systems.<sup>5,23,24,31</sup> Moreover, the low cost of microcomputers should allow their widespread use in manual microscopy for data collection. Cell analysis with such systems, whether performed by image analysis or by manual methods, would benefit from interfacing the microcomputer to a motorized scanning stage.

This paper describes a package of modular programs that interface a microcomputer to a commercially available microprocessor-controlled stepper motor stage. These programs have been developed in both PASCAL and BASIC, allowing their incorporation as software "tools" into microcomputer-based microscopy systems developed in either of these languages.

From the Transplantation Research Program Center, Naval Medical Research Institute, Bethesda, Maryland, and Ludl Electronic Products, Scarsdale, New York, U.S.A.

Mr. Lee, an Undergraduate Student at Cornell University, was Biological Aid, Naval Medical Research Institute.

Mr. Ludl owns Ludl Electronic Products.

Dr. Reynolds was Research Medical Officer, Naval Medical Research Institute, and is now Assistant Professor of Pediatrics, Jonsson Comprehensive Cancer Center, University of California at Los Angeles Medical Center, Los Angeles, California

Address reprint requests to: C. Patrick Reynolds, M.D., Ph.D., Department of Pediatrics, UCLA Medical Center, Los Angeles, CA 90024

This study was supported by Naval Medical Research and Development Command Work Unit MF58.527.007.0004. The opinions and assertions contained herein are the private views of the authors and are not to be construed as official or reflecting the views of the Navy Department or the Naval Service at large

Received for publication January 21, 1987

Accepted for publication June 4, 1987


0884-6812/88/1001-0047/\$02.00/0 © The International Academy of Cytology  
Analytical and Quantitative Cytology and Histology

**Materials and Methods**


The programs to control the stepper motor stage were developed on an IBM PC-AT with a standard serial interface. The stage controller consists of a 6-MHz 8085 microprocessor with 24K RAM and 24K ROM in addition to a proportional X,Y joystick (MDACE 1000, Ludl Electronics, Scarsdale, New York). The MDACE 1000 provides a serial RS-232 interface that connects an IBM PC or PC-AT microcomputer to the MDACE controller, allowing the microcomputer to control variable scan patterns and to interrogate stage coordinates. The MDACE controller also controls a filter wheel (positioned between the microscope and a video camera), which allows switching of six band-pass filters. The stage interfaced to the controller contains stepper motors and 1.0-mm lead screws in the steppers with drivers for the X and Y axes. The various components of the MDACE 1000 consist of the stage controller, filter wheel controller, main power supply, programming module and central processing unit (Figure 1).

Routines to control the stage were first written in Microsoft BASIC (Microsoft Corporation, Bellevue, Washington) and then translated to Turbo PASCAL (Borland International, Scotts Valley, California). Serial input/output (I/O) procedures from the Turbo Asynch Tools Package (Blaise Computing, Berkeley, California) were used to control the serial interface. A DCE RS-232 cable was used to interface the controller to the computer (Figure 2).

The MDACE 1000 allows programmable incre-



**Figure 1**  
Main components of the MDACE 1000 stage controller. Instruction codes from the computer are sent to the stage controller microprocessor using a serial RS-232 interface.



**Figure 2**  
Wiring diagram for the RS-232 asynchronous communication cable used to interface the computer to the stage controller.

ments from one step to 8,000,000 steps, a scanning distance of  $\pm 8,000,000$  steps, a scanning speed from 0.001 mm to a maximum of 16.00 mm per second (50,000 steps per second) and a communication rate from 110 to 9600 baud. The use of 1.0-mm lead screws allows stepping increments as small as 0.1  $\mu\text{m}$ .

**Results**

As originally designed, the MDACE 1000 provides programmable control of the stage stepper motors via a keypad on the controller. Scan patterns can be programmed using the keypad and stored in the non-volatile random access memory of the MDACE 1000. To improve the user interface and versatility of the system and to allow interaction with computer image analysis systems, we developed software to control the stage from a microcomputer.

The communications syntax follows standard RS-232 asynchronous serial communications protocol. In sending data to the stage controller, such as target positions or stepper motor speed, the process described in Figure 3 is followed. The stage controller acknowledges each instruction by returning an "echo" of the command. However, reading information from the stage controller, such as the current position of the motors, requires that an inquiry code is sent (Figure 4). An "echo" of the instruction is not

Fig  
Flo  
dat.

ret  
inf  
th.  
ad  
Fig  
ser  
ad.  
lin.




Figure 3  
 Flow chart of the RS-232 communications protocol used to send data to the stage controller.

returned; rather, the returning bytes are the desired information. Table I lists the major instruction codes that are used to control the stage.

An example of the program in Turbo PASCAL that adjusts the stage's stepper motor speed is shown in Figure 5. The program opens the serial port and then sends the appropriate ASCII code to the stage, which adjusts the stepper motor rate using the protocol outlined in Figure 3. Similar programs control other fea-


Table I ASCII and Hexadecimal Equivalent Instruction Codes Used to Control the Stage Via the Serial Communications Port

Code description	HEX code	ASCII code
READ DISPLAY	\$H72	114
ENABLE CPU DISPLAY	\$H44	68
DISABLE CPU DISPLAY	\$H45	69
FILTER WHEEL ADVANCE	\$H46	70
FILTER WHEEL REVERSE	\$H4E	78
TEST BUSY (MOTOR)	\$H0	00
LOAD BASE	\$H41	65
READ BASE	\$H61	97
LOAD SPEED	\$H53	83
RUN MOTOR	\$H47	71

tures of the stage by using the appropriate ASCII codes from Table I. Although identical routines for control of the stage were written in both Microsoft BASIC and Turbo PASCAL, the PASCAL syntax allows defining tasks as "procedures." The individual procedures can then be easily utilized according to the programmer's needs. Table II lists the individual PASCAL procedures developed to control the stage; Table III lists the functions provided by the Turbo Asynch Tools used in serial I/O communications. By combining the appropriate procedures and functions, programs for scanning various types of specimens can be easily and rapidly developed.

Another feature this software provides is the ability to store coordinates of any event identified during scanning. Marking a significant event is achieved by storing to a disk file the number of motor steps moved along the X and Y axis from a predetermined "base" or zero point. Returning to the event is accomplished simply by incrementing to the same number of motor steps away from a predetermined base. This method allows the software to store an unlimited number of events. The 1- $\mu$ m resolution of the stage allows accurate relocation of a significant object.

Two methods of defining scanning patterns can be used with this software. One method consists of setting the size of the scan increment, establishing the number of increments, and looping through the movement procedure the appropriate number of times. The second method reads X-Y coordinates from a file, and then sends the stage to the desired location. A necessary capability for both methods is to set a reference point, or coordinate origin of the slide, so that the stage is aligned and can then return to marked locations of significant events. Manual



**Figure 4**  
Flow chart showing how the computer interrogates the stage controller for data, such as coordinates of the current slide location.

control of the stage motors is provided through a joystick controller, which is useful for setting this reference point.

To provide a "user-friendly" interface with the stage operator, the various routines for control of the stage can be integrated under control of a menu-driven program. The menu from an example control program is shown in Figure 6. The program is designed to allow variable scan patterns for particular

**Table II** List of Turbo PASCAL "Procedures" (Subroutines) Developed for Control of the Microscope Scanning Stage

Procedures	Description
Convert	Converts data into integers
stage_drive	I/O using Turbo Asynch Tools
Open_Com1	Opens the serial port calling Asynch functions
stage_write	I/O using Turbo PASCAL standard I/O functions
Speed_Rate	Sets the stepper speed of the motor
Position_Absolute	Sets the coordinate origin (0,0)
Move_Stage	Sends codes to the respective motors
InquireCoordinates	Reads the position of the stepper motor
ReadDisplay	Reads the keypad display
Exist	Checks the disk directory for an existing file
Stepper	Sets a target position and moves the stage there
StageX	Sends a target position to the x-axis motor
Test_Busy	Tests to see if a motor is still moving
StageY	Sends a target position to the y-axis motor
SetUp	Sets up the scan patterns
Scan_Slide	Implements the scan pattern and saves coordinates to a disk file
Filter	Sends codes to the filter wheel
Speed	Changes the stepper rate of the motors
Readback	Recalls saved coordinates of significant locations
Menu	Displays a menu of the possible choices

slides with options to (1) move defined distances along the X and Y axes, (2) adjust the stepper motor speed, (3) set the coordinate origin (reference point = 0,0) as the current stage location, (4) display coordinates of the current location, (5) change the scan increment size, (6) rotate the filter wheel and (7) perform user-defined scan patterns of a slide.

#### Discussion

The microcomputer software developed for a commercially available motorized microscope stage controller provides microcomputer control of various scan patterns and motor speeds and storage of coordinates allowing relocation of areas of interest. Via a

**Table III** Modular Functions for Serial I/O Provided by the Turbo Asynch Tools

Turbo Asynch functions	Description
ToHex	Converts a decimal value to hexadecimal
Open A1	Initializes the serial port
Close A1	Closes the serial port
WrtChA1	Writes a character to the output buffer of the serial port
RdChA1	Reads a character from the input buffer of the serial port

```
procedure stage_drive(order:integer);
begin { stage_write }
  Errorcode:= WrtchA1(COM1,chr(order));
  Errorcode:= RdChA1(COM1,ch,InqSize,Portstatus);
end { stage_write };

procedure Open_Com1;
begin; { Open_Com1 }
  Errorcode:= OpenA1(COM1,100,100,0,0,addr(iobuffer));
  if (Errorcode = Port_Open_Already) then
    write('Com1 is already opened.')
  else if (Errorcode = OK) then
    write('Com1 successfully opened.')
  else
    write('Cannot open Com1. Error Code = ',Errorcode:3);
  write(InqSize,' characters remain in the input buffer.');
```

```
end; { Open_com1 }

procedure stage_write(order:integer);
begin { stage_write }
  write(aux, chr(order));
  read(aux, ch);
  writeln('code echo ', ch);
end { stage_write };

procedure Speed_Rate(device:integer); { write speed to the stage }
begin; { Speed_Rate }
  Stage_drive(58); { first : }
  Stage_drive(device); { dev = 0 }
  Stage_drive(83); { inst = $H53 }
  Stage_drive(2); { data = 2 }
  Convert(50000.0);
  Stage_drive(idat_1);
  Stage_drive(idat_2);
  Stage_drive(58); { last : }
end; { Speed_Rate }
```

Figure 5  
I/O procedures. Instructions are sent via serial RS-232 asynchronous format. Shown are sample procedures for reading and writing to the serial port and setting the stepping rate of the motor. Parameters follow normal PASCAL syntax.

serial (RS-232) interface, the microcomputer is able to record the number of motor steps that the stage moves, thus defining an X-Y coordinate axis. The ability to interrogate and store coordinates of observed "events" to a disk file allows recall of the observed object for reobservation at a later time.<sup>18,21,37</sup> In addition to control of the stage, the MDACE controller provides control of a filter wheel positioned between the microscope and an intensified (ISIT) video camera. The software developed also allows control of the filter wheel by the microcomputer. We have found the filter wheel useful for changing band-pass filters in multicolor fluorescence assays.

Computer-controlled microscope stages provide rapid, accurate slide movement, thus reducing operator fatigue in scanning large numbers of specimens. In addition, such a system allows semiautomatic collection of various data during microscopy, which can be coordinated with locations on the slide.<sup>4,9,17,18,21,25</sup> Examples of techniques in which such "computer-aided microscopy" can be useful include grain counting in autoradiography,<sup>17,25,32</sup> reading various immunologic assays<sup>11,19</sup> and neuroanatomic mapping.<sup>1,9,10,12,38</sup> The programs described here will also be useful for more automated systems that use digital image analysis for detection and quantitation of various types of cells.<sup>1,3,6,20,30,35</sup>

An advantage of developing programs for microcomputers is that a large number of programmers are working in an identical (or at least compatible) environment. This has led several vendors to develop powerful sets of programming "tools" that provide callable subroutines for various tasks, thus speeding up program development. We have made use of such tools in developing this software (Blaise Computing Turbo Asynch). The PASCAL software we have developed can be modified and compiled with Microsoft PASCAL (Microsoft Corporation, Bellevue, Washington) or PASCAL-2 (Oregon Software, Portland, Oregon), both of which allow linking to compiled FORTRAN or C programs, thus extending the value of this software as "tools" for programmers in other languages.

One of the obstacles to developing effective automated microscopy systems has been the necessity of obtaining hardware (often custom designed) that will accommodate software developed by other investiga-

#### MICROCOMPUTER INTERFACE FOR CONTROL OF A MICROSCOPE SCANNING STAGE

```
[X] Stage Movement - x-axis variable stage movement
[Y] Stage Movement - y-axis variable stage movement
[S]peed Control - Change the scanning rate

Position [A]bsolute - reset current stage position as (0,0)
[D]isplay the keyboard CPU coordinates
Inquire [C]oordinates - read the coordinates of the motor stepper

[F]ilter Wheel - step forward one filter
Filter [R]everse - step reverse one filter
Display [C]urrent filter
[M]ove - combined X and Y stage movement
Redisplay this men[u]

[Q]uit, and return to data processing

Enter the letter of selection: X,Y,P,A,D,C,F,R,C,M,U,Q
```

Figure 6  
An example of a menu for the microcomputer-controlled microscope scanning stage programs.

Table IV Comparison of Several Commercially Available Microscope Scanning Stages

Stage (and vendor)	Resolution ( $\mu\text{m}$ )	Scan area (mm)	Maximum speed (mm/s)	Approximate cost
Marzhauser Mechanical Stage EK32 (Bunton Instruments, Rockville, MD 20850)	0.1	75 $\times$ 50	20	\$5,000 (stage) \$5,000 (controller)
Boeckeler Motorized X-Y System (Boeckeler Instruments, Tucson, AZ 85705)	5.0	150 $\times$ 150	50	\$10,875*
	1.0	50 $\times$ 50	25	\$9,345*
Stepper Motor Positioning Stages (Stoelting, Chicago, IL 60623)	1.0	88 $\times$ 88	6.7	\$8,895
MDACE 1000 XY Stage Controller (OPELCO, Washington, DC 20041-17127)	0.1	75 $\times$ 75	16	\$5,948 (stage) \$4,374 (controller)
	0.2	75 $\times$ 75	32	\$5,500

\*User must provide the stage from an existing microscope; price includes modification of existing stage and object code for IBM PC control software

tors, or of writing the software *de novo* for the hardware available. Table IV summarizes the characteristics of several commercially available stage-controller systems. Although written for the MDACE 1000, the software described here should be adaptable to any of the scanning stages listed in Table IV, allowing the user some flexibility in the design and assembly of a system.

The rapid increase in graphics power available for low-cost microcomputers indicates that they will find widespread use in the field of digital image microscopy. Several vendors already offer frame grabbers for personal computers, mainly for microcomputers using MS-DOS (IBM PC and compatible).<sup>13,31,39</sup> The availability of such hardware for a large number of compatible computers suggests that standard software "tools" will be developed that will speed up program development for image processing. Several vendors are already releasing such software, and some is even being implemented on graphics chips.<sup>36</sup> As the number of such software tools increases, development of microcomputer digital image microscopy systems will become increasingly easy. The software presented here provides a useful set of tools for microscope stage control, not only for integration with image processing systems, but also for use in semimanual (computer-assisted) microscopy.

## References

- Arndt-Jovin DJ, Robert-Nicoud M, Kaufman SJ, Jovin TM: Fluorescence digital imaging microscopy in cell biology. *Science* 230:247-256, 1985
- Auffermann W, Repges R, Bocking A: Rapid diagnostic DNA cytometry with an automatic microscope and a TV image-analysis system. *Analyt Quant Cytol* 6:179-188, 1984
- Bahr GF: Frontiers of quantitative cytochemistry: A review of recent developments and potentials. *Analyt Quant Cytol* 1:1-19, 1979
- Bahr GF, Boccia JA: Marking the origin of coordinates in high-resolution light-microscopy scanning. *J Microsc* 132:137-141, 1983
- Barrows GH, Siskin JE, Allegra JC, Grascch SD: Measurement of fluorescence using digital integration of video images. *J Histochem Cytochem* 32:741-746, 1984
- Bartels PH, Abmayr W, Bibbo M, Burger G, Soost HJ, Taylor J, Wied GL: Computer recognition of ectocervical cells: Image features. *Analyt Quant Cytol* 3:157-163, 1981
- Berbel PJ, Villanueva JJ, Regidor J, Lopez-Garcia C: A method for the study of the spatial distribution of the neuronal dendritic tree using a universal stage. *J Neurosci Methods* 4:141-152, 1981
- Berns GS, Berns MW: Computer-based tracking of living cells. *Exp Cell Res* 142:103-109, 1982
- Capowski JJ, Cruce WLR: How to configure a computer-aided neuron reconstruction and graphics display system. *Comput Biomed Res* 12:569-587, 1979
- Capowski JJ, Rethelyi M: Neuron reconstruction using a Quantimet image analyzing computer system. *Acta Morphol Acad Sci Hung* 30:241-249, 1982
- Darling W, Justin O, Pike BL, Jennings G, Shortman K: An indexing stage for microscopic scanning of microtitre tray wells. *J Immunol Methods* 52:39-42, 1982
- Davis BJ: The electronic pantograph: Amplifier couples microscope stage to X-Y plotter. *Brain Res Bull* 15:533-536, 1985
- Dawson B: PC transformed into image processor. *Digital Design* 16:63-73, 1986
- Donovan RM, Goldstein E: A charge-coupled device-based image cytophotometry system for quantitative histochemistry and cytochemistry. *J Histochem Cytochem* 33:551-556, 1985
- Dytch HE, Bartels PH, Bibbo M, Pishotta FT, Wied GL: The rejection of noncellular artifacts in Papanicolaou-stained slide specimens by an automated high-resolution system. *Analyt Quant Cytol* 5:241-249, 1983
- Gamel JW, McLean JW: Computerized histopathologic assessment of malignant potential: III. Refinements of measurement and data analysis. *Analyt Quant Cytol* 6:37-44, 1984
- Giaretti WA, Gais P, Jütting U, Rodenacker K, Dörmer P: Correlation between chromatin morphology as derived by digital image analysis and autoradiographic labeling pattern. *Analyt Quant Cytol* 5:79-89, 1983
- Glaser EM, McMullen NT: Semiautomatic microscopy as a

- tool for cytologists. *Analyt Quant Cytol Histol* 8:116-127, 1986
19. Goldrosen MH, Giller GA, Warshaw WB: A program for the control of an image analysis system used in *in vitro* assays of cell-mediated immunity. *Comput Programs Biomed* 14:121-126, 1982
  20. Gunzer U, Harms H, Haucke M, Aus HM, ter Meulen V: Computer-aided image analysis for the differentiation of mononuclear cells in peripheral blood smears from leukemic patients. *Analyt Quant Cytol* 3:26-32, 1981
  21. Hunter NR, Taylor GR, Swank PR, Winkler DG, Thompson JL, Montalvo J, Greenberg SD, McFadyen GM, Estrada R: Precision cell location and relocation techniques: An application for cell image analysis. *Analyt Quant Cytol* 6:139-143, 1984
  22. Inoue S: Applications of video to light microscopy. *In Video Microscopy*. Edited by S Inoue. New York, Plenum Press, 1986, pp 393-422
  23. Isseroff A, Lancet D: An inexpensive microcomputer-based image-analysis system: Novel applications to quantitative autoradiography. *J Neurosci Methods* 12:265-275, 1985
  24. Jarvis LR: A microcomputer system for video image analysis and diagnostic microdensitometry. *Analyt Quant Cytol Histol* 8:201-209, 1986
  25. Kemper HL, Kallman RF: SACCAS: A semiautomated counting machine with an accommodating threshold selector. *Analyt Quant Cytol* 5:138-142, 1983
  26. Lin YK, Fu KS: Segmentation of Papanicolaou smear images. *Analyt Quant Cytol* 3:201-206, 1981
  27. Mize RR: A microcomputer plotter for use with light and electron microscopes. *In The Microcomputer in Cell and Neurobiology Research*. Edited by RR Mize. New York, Elsevier, 1985, pp 112-133
  28. Noguchi Y, Tenjin Y, Sugishita T: Cancer-cell detection system based on multispectral images. *Analyt Quant Cytol* 5:143-151, 1983
  29. Pearlstein RA, Kirschner L, Simon SJ, Machell S, White WF, Sidman RL: A multimodal system for reconstruction and quantification of neurologic structures. *Analyt Quant Cytol Histol* 8:108-115, 1986
  30. Preston K Jr: Automation of the analysis of cell images. *Analyt Quant Cytol* 2:1-14, 1979
  31. Puls JH, Bibbo M, Dytch HE, Bartels PH, Wied GI: MicroTICAS: The design of an inexpensive video-based microphotometer/computer system for DNA ploidy studies. *Analyt Quant Cytol Histol* 8:1-7, 1986
  32. Ramm P, Kulich JH: Principles of computer-assisted imaging in autoradiographic densitometry. *In The Microcomputer in Cell and Neurobiology Research*. Edited by RR Mize. New York, Elsevier, 1985, pp 312-334
  33. Rosvoll RV, Mengason AP, Smith L, Patel HJ, Maynard J, Connor F: Visual and automated differential leukocyte counts: A comparison of three instruments. *Am J Clin Pathol* 71:695-703, 1979
  34. Schoentag RA, Pedersen JT: Evaluation of an automated blood smear analyzer. *Am J Clin Pathol* 71:685-694, 1979
  35. Shapiro B, Lemkin P, Lipkin L: The application of artificial intelligence techniques to biologic cell identification. *J Histochem Cytochem* 22:741-750, 1974
  36. Stock R, Robertson B: New chips unleash super graphics. *Computer Graphics World* 9:24-32, 1986
  37. Thurston G, Jaggi B, Palcic B: Cell motility measurements with an automated microscope system. *Exp Cell Res* 165:380-390, 1986
  38. Williams FG, Elde R: A microcomputer-aided system for the graphic reproduction of neurohistochemical maps. *Comput Programs Biomed* 15:93-102, 1982
  39. Wilson A: Imaging rises to meet user demands. *Digital Design* 16:23, 1986