

AAMRL-SR-90-516

AD-A254 332



VISUAL EVOKED RESPONSE (VER) DETECTION OF LOSS OF
PERIPHERAL VISION

WILLIAM B. ALBERY, Ph.D.

Harry G. Armstrong Aerospace Medical Research Laboratory

RICHARD T. GILL, Ph.D.

University of Idaho
Moscow, Idaho

DTIC
ELECTE
AUG 11 1992
S A D

DECEMBER 1990

FINAL REPORT FOR THE PERIOD OCTOBER 1984 THROUGH AUGUST 1987

Approved for public release; distribution is unlimited

HARRY G. ARMSTRONG AEROSPACE MEDICAL RESEARCH LABORATORY
HUMAN SYSTEMS DIVISION
AIR FORCE SYSTEMS COMMAND
WRIGHT-PATTERSON AIR FORCE BASE, OHIO 45433-6573

92 8 10 115

92-22633

NOTICES

When US Government drawings, specifications, or other data are used for any purpose other than a definitely related Government procurement operation, the Government thereby incurs no responsibility nor any obligation whatsoever, and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise, as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use or sell any patented invention that may in any way be related thereto.

Please do not request copies of this report from the Armstrong Laboratory. Additional copies may be purchased from:

National Technical Information Service
5285 Port Royal Road
Springfield VA 22314


TECHNICAL REVIEW AND APPROVAL

AAMRL-SR-90-516

This report has been reviewed by the Office of Public Affairs (PA) and is releasable to the National Technical Information Service (NTIS). At NTIS, it will be available to the general public, including foreign nations.

This technical report has been reviewed and is approved for publication.

FOR THE COMMANDER



JAMES W. BRINKLEY, SES
Director, Biodynamics and Bioengineering Division
Armstrong Laboratory

DTIC QUALITY INSPECTED 5

Accession For	
NTIS	CRA&I <input checked="checked" type="checkbox"/>
DTIC	TAB <input type="checkbox"/>
Unannounced	<input type="checkbox"/>
Justification	
By	
Distribution /	
Availability Codes	
Dist	Avail and/or Special
A-1	

REPORT DOCUMENTATION PAGE

Form Approved
OMB No. 0704-0188

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.

1. AGENCY USE ONLY (Leave blank)		2. REPORT DATE March 1991	3. REPORT TYPE AND DATES COVERED Special Report Oct 84 - Aug 87	
4. TITLE AND SUBTITLE Visual Evoked Response (VER) Detection of Loss of Peripheral Vision			5. FUNDING NUMBERS 72312508	
6. AUTHOR(S) William B. Albery				
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Harry G. Armstrong Aerospace Medical Research Laboratory Wright-Patterson AFB, OH 45433-6573			8. PERFORMING ORGANIZATION REPORT NUMBER AAMRL-SR-90-516	
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES)			10. SPONSORING/MONITORING AGENCY REPORT NUMBER	
11. SUPPLEMENTARY NOTES				
12a. DISTRIBUTION/AVAILABILITY STATEMENT Approved for Public Release; Distribution is Unlimited			12b. DISTRIBUTION CODE	
13. ABSTRACT (Maximum 200 words) A novel concept of objectively determining when a centrifuge subject and subsequently, a pilot, loses peripheral vision due to the physiological effects of sustained acceleration was attempted. The approach was to modulate lights in the peripheral vision of the subject as the subject stared at a centrally fixed indicator. It was demonstrated by using a special, lock-in amplifier, that a reliable visual evoked response (VER) could be elicited in the peripheral vision of a seated subject at 1 G. The next step, determining if the technique worked on subjects under sustained acceleration was never accomplished. Reliable, non-invasive means of monitoring centrifuge subjects came available during the later stages of this research which overcame the need to demonstrate this technique on the centrifuge.				
14. SUBJECT TERMS (U) Electroencephalograms, (U) EEG, (U) Peripheral Light Loss, (U) Acceleration tolerance, (U) Visual Evoked Response			15. NUMBER OF PAGES	
			16. PRICE CODE	
17. SECURITY CLASSIFICATION OF REPORT Unclassified	18. SECURITY CLASSIFICATION OF THIS PAGE Unclassified	19. SECURITY CLASSIFICATION OF ABSTRACT Unclassified	20. LIMITATION OF ABSTRACT UL	

Objective: The objective of this research was to assess the feasibility of using visually induced steady state electroencephalograms (EEG) to estimate acceleration induced peripheral light loss (PLL), rather than relying on subjects' subjective determinations of PLL. The goal was to use this technique as an objective measure of PLL during acceleration tolerance experiments.

Approach: Experiments were conducted in a normal 1-G environment to determine the optimal conditions (i.e., the precise form of the stimulus and the analysis technique) for obtaining the strongest EEG response to both foveal and peripherally localized stimuli. The visual evoked response (VER) was evaluated for both foveal and peripheral stimuli. A semi-circular light bar was constructed (Figure 1) and data were collected on both foveal and peripheral stimulation (lights A, B, C, D in Figure 1). A special amplifier was designed and developed which allowed a narrow band of EEG frequencies to be processed.

Results: Both on-axis and off-axis VERs were observed and recorded at 1 G. Only two subjects were evaluated for VERs stimulated by peripheral lights (Figure 1). The technique was never tested on the centrifuge. VER analysis took at least 3 seconds to complete which made the technique unacceptable for early detection of PLL during centrifuge experimentation.

Discussion: Although the technique of developing a VER by using lights modulated in the periphery of the test subject was demonstrated on at least two subjects, this technique was never evaluated on the centrifuge. Other objective means for monitoring the physiological status of the centrifuge subject became available, such as the ultrasonic flow transducer for temporal artery blood flow. In addition, researchers at USAFSAM, Brooks AFB, TX, completely instrumented subjects for recording EEG during a G-LOC study and found that the frequency of the brain waves shifts from high (alpha) to low (delta) when the centrifuge subject loses consciousness. The technique described here could potentially eliminate the subjectivity involved in estimating the sudden loss of peripheral vision (tunnel vision), but problems with detection and quick analysis of the VER preclude its use at this time. Other reliable non-invasive physiological monitoring techniques such as the transcranial Doppler device, pulse oximetry, and temporal artery flow velocity have been found to be acceptable techniques by centrifuge/acceleration researchers.

Conclusion: The concept of stimulating a visual evoked response by lights in one's periphery was novel in that it represented a potential objective means of determining when a human was losing peripheral vision as a result of the physiological effects of sustained acceleration. It was assumed that if one lost peripheral vision into a cone of 60 degrees, for example, that the VER from lights modulated outside the cone, say at 70 degrees, could no longer be evoked. The peripheral vision-evoked technique was not tested on the centrifuge because other non-

invasive physiological monitoring devices became available that eliminated the time delay problem.




Figure 1: Top View of Conceptual Design For Using EEG Analysis To Determine Human Field-of-View