

AD-A270 731



AD _____

MIPR NO: 91MM1505

TITLE: A COMPARISON OF POLYGRAPHIC AND ACTIGRAPHIC MONITORING OF SLEEP USING A 5-CHANNEL PROGRAMMABLE-SENSITIVITY ACTIGRAPH

**PRINCIPAL INVESTIGATOR: Timothy F. Elsmore
Paul Naitoh**

**CONTRACTING ORGANIZATION: Naval Health Research Center
P. O. Box 85122
San Diego, California 92186-5122**

REPORT DATE: November 30, 1992

TYPE OF REPORT: Final Report

**PREPARED FOR: U.S. Army Medical Research and
Development Command, Fort Detrick
Frederick, Maryland 21702-5012**

**DISTRIBUTION STATEMENT: Approved for public release;
distribution unlimited**

The views, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy or decision unless so designated by other documentation.

**SDTIC
ELECTE
OCT 18 1993
S E D**

REPORT DOCUMENTATION PAGE			Form Approved OMB No. 0704-0188	
Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503				
1. AGENCY USE ONLY (Leave blank)		2. REPORT DATE 30 Nov 92	3. REPORT TYPE AND DATES COVERED Final Report (11/15/90-11/14/92)	
4. TITLE AND SUBTITLE A Comparison of Polygraphic and Actigraphic Monitoring of Sleep Using a 5-Channel Programmable-Sensitivity Actigraph			5. FUNDING NUMBERS MIPR No. 91MM1505	
6. AUTHOR(S) Timothy F. Elsmore Paul Naitoh				
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Naval Health Research Center P.O. Box 85122 San Diego, California 92186-5122			8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES) U.S. Army Medical Research & Development Command Fort Detrick Frederick, Maryland 21702-5012			10. SPONSORING / MONITORING AGENCY REPORT NUMBER	
11. SUPPLEMENTARY NOTES Report covers A2 for the period November 15, 1991 - November 14, 1992.				
12a. DISTRIBUTION / AVAILABILITY STATEMENT Approved for public release; distribution unlimited			12b. DISTRIBUTION CODE	
13. ABSTRACT (Maximum 200 words) The purpose of this project was to investigate the utility of a new device, the AMA-32 actigraph, for monitoring the general activity of human subjects both during sleep and wake. In the first phase of the project, a software program for the analysis of AMA-32 data files was developed. In the second phase, the effects of frequency passband and threshold settings of the AMA-32 amplifier were examined in a 96-hour study. It was concluded that discrimination of sleep from wake was best with the 2-3 Hz passband, which emulates earlier-generation actigraphs, and that existing algorithms for automatic sleep scoring of actigraphs are not applicable to actigraph data collected with non-traditional analog amplifier characteristics. Additional studies are needed to further explore the utility of the additional features offered by the AMA-32.				
14. SUBJECT TERMS Actigraph, Activity, Sleep, Movement, Sleep scoring, Sleep Cycles, Performance, Military, RAD V			15. NUMBER OF PAGES	
			16. PRICE CODE	
17. SECURITY CLASSIFICATION OF REPORT Unclassified	18. SECURITY CLASSIFICATION OF THIS PAGE Unclassified	19. SECURITY CLASSIFICATION OF ABSTRACT Unclassified	20. LIMITATION OF ABSTRACT Unlimited	

FOREWORD

Opinions, interpretations, conclusions and recommendations are those of the author and are not necessarily endorsed by the U.S. Army.

Where copyrighted material is quoted, permission has been obtained to use such material.

Where material from documents designated for limited distribution is quoted, permission has been obtained to use the material.

Citations of commercial organizations and trade names in this report do not constitute an official Department of the Army endorsement or approval of the products or services of these organizations.

In conducting research using animals, the investigator(s) adhered to the "Guide for the Care and Use of Laboratory Animals," prepared by the Committee on Care and Use of Laboratory Animals of the Institute of Laboratory Animal Resources, National Research Council (NIH Publication No. 86-23, Revised 1985).

For the protection of human subjects, the investigator(s) have adhered to policies of applicable Federal Law 45CFR46.

In conducting research utilizing recombinant DNA technology, the investigator(s) adhered to current guidelines promulgated by the National Institutes of Health.

Accession For	
NTIS CRA&I	☒
DTIC TAB	☐
Unannounced	☐
Justification	
By	
Distribution /	
Availability Codes	
Dist	Avail and/or Special
A-1	

Anthony F. Elmore 29 Jan '93
PI Signature Date

391642

93-24198



FUNDING NO: 91MM1505

Bibliography:

Elsmore, T. F. Actigraphs and their applications, DOD Human Factors Engineering Technical Advisory Group meeting, Huntsville, AL, November 1992.

Personnel paid from this contract:

None

A comparison of polygraphic and actigraphic monitoring of sleep¹

Timothy F. Elsmore² and Paul Naitoh³

¹ Final report of work supported by MIPR #91MM1505 from the US Army Medical Research and Development Command to the Naval Health Research Center, San Diego, CA.

² Walter Reed Army Institute of Research, Washington, DC, 20307-5100, USA. This work was done while serving as a visiting Scientist at the Naval Health Research Center, San Diego CA.

³ Naval Health Research Center, San Diego, CA.

TABLE OF CONTENTS

Introduction	3
Software Development	5
Systematic Exploration of AMA-32 Parameters	6
Methods	6
Results	7
General effects of passband and thresho'd variations	7
Waking activity	10
Sleep-wake discrimination	19
Automatic discrimination of sleep from wake	22
Discussion and Conclusions	24
Acknowledgements	27
References	28
Appendices	
1. AMA-32 amplifier settings	31
2. GAAP Software details	32
3. Actigraph raw data	38
4. Frequency distributions by day	45
5. Nighttime activity and EEG somnograms	52

INTRODUCTION

Activity measuring devices are being used in a wide variety of clinical and field settings for assessment of both daytime activity and sleep. Clinical applications include the diagnosis and treatment of a variety of disorders including mood disorders, schizophrenia, hyperactivity in children, attention deficit disorders, sleep disorders, substance abuse, eating disorders, and chronic disease (Brown, Smolensky, D'Alonzo, and Redmond, 1990; Tryon, 1991). These devices are also being used in military research where the emphasis has been on assessment of sleep in field operations (e.g. Brooks, Shergold, Angus, Heslegrave, and Redmond, 1988; Pleban, Valentine, Penetar, Redmond, and Belenky, 1990). The purpose of the present study was to evaluate the utility of a newly-designed, highly programmable type of actigraph (the AMA-32), which permits specification of several analog amplifier parameters.

In order to explore the properties of the AMA-32, it was first necessary to develop a computer program for the analysis of actigraph data. The *General Activity Analysis Program* provides rapid menu-driven access to actigraph data, quantitatively summarizes activity records, scores them for sleep, and outputs graphical and numeric data to the computer screen, printer, or data files.

In initial laboratory studies, effects of altering several amplifier parameters were investigated. Subjects lived in the laboratory for several days, undergoing a simulated time-zone shift. In addition to wearing actigraphs, brain activity was monitored during sleep, and the records were scored for sleep stages using standard techniques. As a result of these studies it was concluded that the 2-3 Hz passband, High Threshold setting of the AMA-32 provides the best sleep/wake discrimination. Additional studies will be necessary to further evaluate the AMA-32 for other purposes.

Wrist-worn actigraphs

A variety of different devices have been used for measurement of activity, the most versatile being the AMA-16 actigraph produced by Precision Control Designs Inc., described by Redmond and Hegge (1985). The device measures approximately 6.4 x 8.9 x 1.9 cm and was attached to the wrist by a velcro wrist band, using a solid-state accelerometer to detect motion. After processing by appropriate analog circuitry, the data are stored in a 16 Kbyte memory. The AMA-16 is programmed by inserting it into an interface unit which in turn is attached to an IBM PC-compatible microcomputer. This device has been used extensively in laboratory and field research (Brooks, Shergold, Angus, Heslegrave, and Redmond, 1988; Mullaney, Kripke, and Messin 1980; Konigsberg and Cushing, 1986; Tryon, 1991), and yields useful information on the impact of environmental stressors on patterns of sleep and wake. However, it has some drawbacks. It is large enough to cause difficulties in some environments. For example, some pilots refuse to wear the device since it interferes with the normal wearing of a tight-fitting flight suit (French, 1992), its memory capacity is only 16 Kbytes limiting the amount of data that may be stored, and its recording parameters are fixed. That is, the frequency passband, threshold levels, and gain of the circuitry may not be changed.

A new generation of actigraphs (Precision Control Designs Inc., Model AMA-32) addresses many of the drawbacks of the AMA-16. This model is much smaller, measuring only 4.45 x 3.48 x 1.27 cm., about the size of a diver's watch. The memory capacity has been doubled to 32 Kbytes, making longer studies possible, and it is highly programmable, permitting user specification of frequency passband, threshold, and amplification factors. This versatility theoretically should permit fine-tuning of measurements for specific purposes (e.g. daytime vs nighttime activity). Details of the available AMA-32 actigraph settings are shown in Appendix 1. However, there are

severe shortcomings both in information on the significance of the available AMA-32 settings as well as in methods for analysis of AMA-32 data. The purpose of the present project was to address these shortcomings in order to maximize the utility of the AMA-32 in laboratory, clinical, and field research settings.

SOFTWARE DEVELOPMENT

The lack of user-friendly software has been a significant impediment to the widespread application of actigraph technology. In fact, at the beginning of the present project there was *no* user-friendly software for the derivation of research-grade data from actigraph data files. Programs do exist for clinical applications of actigraphy, but these programs lack the versatility necessary for the types of exploratory analyses desired for research purposes. Thus, the first step in this project was the development of a software program for the derivation of research-grade data from actigraph data files. The General Activity Analysis Program (GAAP) permits rapid analysis of data files generated with either AMA-16 or AMA-32 actigraphs. The program is described in detail in Appendix

2. The principal features of GAAP are:

1. Graphical display of activity at two levels of resolution
2. Daily or hourly summaries of sleep and activity
3. Daily or hourly distributions of activity data
4. Multiple sleep-scoring algorithms
5. Sleep scoring editing capability
6. Screen, Printer, or File data output
7. Epoch-by-epoch data file output
8. Graphics file output for hard-copy plots of activity
9. Mouse-driven user interface

GAAP enables individuals with no programming expertise to quickly access and analyze activity data, generate figures, and produce output appropriate for analysis by conventional statistical and spreadsheet programs. This program is available upon request from the first author or may be downloaded from the OMPAT computer Bulletin Board (phone 1-800-542-7844).

SYSTEMATIC EXPLORATION OF AMA-32 PARAMETERS

While a complete exploration of the possibilities presented by the AMA-32 was not possible in this project, to begin the process of defining the utility of this device, a number of subjects were run as part of a larger experiment.

Method

Subjects and Procedure

Young adult male civilian volunteer subjects were recruited for participation in a study of the effects of bright light on adaptation to a simulated time-zone shift. Data from seven subjects will be reported. As part of the study, EEG recordings were made during four eight-hour sleep periods. During the study, each subject continuously wore two actigraphs on his non-dominant wrist for up to 96 consecutive hours. The actigraphs were each set to the same frequency passband with a high gain, differing only in threshold, and were operated in zero-crossing mode (see Appendix 2 for more details on actigraph operating modes). Amplifier settings used are shown in Table 1 (see Appendix 1 for more details on the available amplifier settings).

Table 1 -- Amplifier Settings

<u>Subject</u>	<u>Settings</u>	<u>Passband (Hz)</u>
BL201	5, 6	.1 - 3
BL203	5, 6	.1 - 3
BL205	9, 10	.1 - 9
BL206	9, 10	.1 - 9
BL207	17, 18	2 - 3
BL208	17, 18	2 - 3
BL210	19, 20	2 - 9

Results

Although there will be some overlap, descriptions of results will be divided into five sections, 1) general effects of passband and threshold variations; 2) activity during wake; 3) activity during sleep; 4) sleep/wake discrimination; 5) automatic discrimination of sleep from wake. Because of the exploratory nature of this project and the large amount of data collected, representative selected data will be presented in the body of the paper and more detailed data will be provided in appendices.

General effects of passband and threshold variations

As indicated by Redmond and Hegge (1987), the effects of alterations in the analog sensitivity of the circuitry of the actigraph are quite complex and not easily characterized. However, the data collected in this study permit some generalizations. Figure 1 illustrates complete actigraph records from two subjects. The left column is from Subject BL203, whose two actigraphs were set for a passband of 0.1 - 3 Hz, and the right is from Subject BL208 whose actigraphs were set for a 2 - 3 Hz passband. The top row is for the actigraphs with a low threshold setting, and the bottom row is for the high-threshold actigraphs. In these figures, each line represents activity for a full day, from 1200 to 1200. All records were plotted on the same scale.




Figure 1. Complete actigraph records for subject 203 (passband, 0.1-3 Hz) and 207 (passband, 2-3 Hz). The top and bottom rows represent low and high threshold actigraphs, respectively.

There was a clear difference in overall activity levels between the two passbands, with considerably more activity shown for 2 - 3 Hz than for 0.1 - 3 Hz. The other two subjects with these amplifier settings showed the same effect to approximately the same degree. This finding is somewhat counterintuitive, since 0.1 - 3 Hz is a broader passband than 2 - 3 Hz and thus should accumulate more movements. However, this simplistic notion does not take the complex nature of the signal generated by the motion transducer in the actigraph into account. Recall that the actigraphs were operated in zero-crossing mode, in which the beginning of a movement is defined as occurring when the voltage from the transducer exceeds the threshold. The low end of the 0.1 - 3 passband is sensitive to very slow movements (i.e. 1 movement / 10 seconds), thus when a

slow movement begins (i.e. the voltage exceeds the threshold), faster movements will not register, since the threshold voltage has already been exceeded. Hence, fewer total counts may be registered when the passband includes extremely low frequencies.

A second phenomenon that is apparent from Figure 1 is that, during sleep, more counts are registered with the 0.1 - 3 Hz passband than with the 2 - 3 Hz passband. On day 1, the sleep period was from 2200 - 0600 hours, and on subsequent days, it was from 0800 - 1600 hours. In some cases it is quite difficult to visually distinguish sleep from wake in the raw activity records with the 0.1 - 3 Hz passband, whereas the sleep periods are quite clearly visible with the 2 - 3 Hz passband. Complete activity records for all subjects are attached to this report as Appendix 3. inspection of the records for subject BL210 (Appendix 3-7) suggested that his actigraphs were not functioning reliably, hence his data must be regarded as highly preliminary.

Figure 1 also illustrates an effect of threshold, which, for these passbands, resulted in less activity for the higher threshold than the low threshold. However, this was not always the case. The probable reason for this is illustrated in Figure 2 which shows a hypothetical activity waveform, and three different thresholds. If the activity waveform were a perfect sinusoid, varying the threshold would have a binary effect, that is, either a movement would register or not, and lower thresholds would result in more movements being counted. However, movements do not produce perfectly sinusoidal waveforms. In this figure, choosing thresholds A or C would result in the registration of two movements. However, choosing threshold B would result in 6 movements being counted. When the additional factors of amplifier gain and passband are considered, the relationships become even more complex.




Figure 2. Hypothetical waveforms of actigraph transducer voltage output illustrating effects of varying thresholds on recorded movement frequency.

Waking activity

Assessment of daytime activity has been largely confined to summaries of hourly or total daily activity (Tryon, 1991). These measurements have been useful in distinguishing between various groups, such as normal and depressed patients (eg. Godfrey and Knight, 1984), hyperactive and normal children (eg. Porrino, Rappoport, Behar, Sceery, Ismund. and Bunney, 1983), and drugged vs control patients (eg. Mattman, Loepfe, Scheitlin, Schmidlin, Gerne, Strauch, Lehman, and Borbely, 1982). With more detailed analysis, it should be possible to detect treatment effects with greater precision.

Table 2 presents summary descriptive statistics for individual subject activity during three "days", with day 1 representing the 24-hour period beginning at 0800 following the first night of sleep in the laboratory. No sleep was permitted during this period. Days 2 and 3 were 16-hour periods beginning at 1600 following the first and second daytime sleep periods. Subject BL203's high-threshold actigraph failed during the third day. Table 2 shows that in most cases the day-to-day reliability of the actigraphs was quite high, with daily variations in mean activity generally within 15% of the three-day mean. Agreement between subjects for comparable actigraphs (e.g. BL201, low threshold vs. BL203 low threshold) was similarly high, with no more than a 10% difference between the three-day mean score for each subject and the mean of the two three-day scores for both subjects. This table also shows a reasonable degree of agreement between the means and medians of movements per epoch, indicating that the movements per epoch distributions are reasonably uniform.

Table 2
Movements per epoch for continuous waking activity

Passband	Subject	Threshold	Day	Mean	SD	Median	Q1	Q3
0.1-3 Hz	BL201	L	1	26.74	9.71	25	20	32
			2	24.52	10.50	24	18	30
			3	25.28	11.67	23	18	32
		H	1	19.56	8.51	19	14	24
			2	19.37	10.00	19	13	26
			3	21.51	9.22	21	16	27
	BL203	L	1	23.67	11.12	21	16	29
			2	19.10	9.56	17	13	23
			3	19.96	9.37	19	14	24
		H	1	18.43	9.52	18	12	24
			2	18.09	9.75	18	12	23
			3	..	..	.	.	.
0.1-9 Hz	BL205	L	1	37.32	17.38	33	25	46
			2	29.09	15.44	27	19	37
			3	41.57	19.41	40	30	54
		H	1	41.72	18.41	42	30	53
			2	37.20	17.41	37	26	48
			3	30.05	15.30	31	22	39
	BL206	L	1	38.00	17.00	35	25	49
			2	45.53	19.69	44	31	58
			3	43.26	18.54	43	32	56
		H	1	39.30	17.93	36	27	51
			2	34.85	15.78	33	24	45
			3	36.60	18.06	35	25	48
2-3 Hz	BL207	L	1	56.47	19.03	61	48	88
			2	54.38	21.77	62	47	68
			3	47.41	25.87	58	26	66
		H	1	36.25	22.00	36	17	53
			2	36.57	23.25	41	16	57
			3	36.96	22.75	41	17	56
	BL208	L	1	54.20	18.80	60	46	67
			2	50.96	21.13	58	42	56
			3	50.48	21.16	57	37	56
		H	1	34.45	20.65	35	17	53
			2	34.48	21.86	36	15	54
			3	31.77	21.43	31	13	52
2-9 Hz	BL210	L	1	68.87	21.82	75	59	84
			2	80.70	17.66	83	76	90
			3	82.03	9.91	82	77	88
		H	1	51.19	20.66	54	36	68
			2	52.99	23.01	59	37	71
			3	57.20	21.05	64	42	74

This uniformity of distributions is also evident in Figure 3 which in which the medians and inter-quartile ranges (IQR) of movements per epoch are displayed. In most cases, the IQR bars extend approximately equally above and below the median. The most marked asymmetries occur when the medians are highest, i.e. for the 2-3 Hz, low threshold actigraphs. This figure also illustrates that for the 2-3 Hz, high threshold actigraphs, the IQR's are quite large, indicating that the distributions are relatively flat. This point will be further illustrated in the sleep/wake section.




Figure 3. Median movements per epoch. The heavy vertical lines on each bar span the inter-quartile range.

In addition to the summary statistics presented thus far, consideration of actigraph records as time series provides another set of approaches to data analysis that may prove useful in addressing issues related to daytime activity patterns.

Autocorrelation -- Day 1




Figure 4. Autocorrelation functions for six subjects for 24 hours. Each row represents two subjects with the same passband setting.

Examples of two time series analysis techniques will be provided. One approach to time series analysis is autocorrelation which assesses the degree of correlation between increasingly distant epochs. Figure 5 shows autocorrelation functions for both high- and low-threshold actigraphs for day 1. Correlations are plotted for lags 1 to 20, where lag 1 represents immediately adjacent epochs, and lag 20 correlates epochs that were 20 epochs (10 minutes) apart. These functions all have the same general shape, with the correlation between epochs decreasing as the distance between the epochs increases. The functions are all quite smooth and orderly, indicating that there were no short-term periodicities (i.e. period less than 20 epochs) in the data. There was also good agreement between each subject's actigraphs, though for both BL201 and BL205 the autocorrelation functions for the high threshold actigraph was consistently higher than that for the low threshold actigraph.

Figure 5 shows the results of submitting the data from day 1 to a Fourier analysis (Fast Fourier Transform or FFT). In all cases, the power or magnitude is greatest at the higher frequencies (periods less than 20 minutes). Figures 4 and 5 are presented here merely to illustrate the type of data that can result from time series analyses. It remains for more extensive studies to determine the degree to which such analyses may be useful in assessment of the effects of other variables of interest on activity.

Activity during sleep periods

Summary statistics for each subject for three eight-hour sleep periods are presented in Table 3. The first sleep period was from 2200 on day 1 to 0600 the following day. Subsequent sleep periods began at 0800. For simplicity, sleep periods will be referred to as "nights". Thus, night 2 began following 26 hours of time awake. Since the low threshold actigraph for subject BL210 registered an extremely large increase in

Fourier Analysis -- Day 1




Figure 5. Periodograms for 24 hours of waking activity.

movements during the last two nights, it is assumed that device was malfunctioning, thus, those data will be discounted.

In most cases, the stability of activity across nights was less than that seen above during wake, with several examples of greater than 50% night-to-night differences. There was also little consistency in the pattern of changes across subjects. The degree to which this variability reflected the effects of the experimental protocol is not clear from the present data. It is strikingly clear from Table 3, however, that the overall activity level was least for the 2-3 Hz passband, particularly so for the high threshold actigraphs. The median number of movements recorded by these actigraphs was consistently zero during sleep periods, thus the movement distributions are highly asymmetrical. On the other hand, the other passbands all continued to have fairly uniform distributions. Figure 6 shows these effects graphically by plotting the medians and interquartile ranges of movements per epoch from night 1.

Table 3
Movements per epoch during sleep periods

<u>Passband</u>	<u>Subject</u>	<u>Threshold</u>	<u>Sleep Period</u>	<u>Mean</u>	<u>SD</u>	<u>Median</u>	<u>Q1</u>	<u>Q3</u>
0.1-3 Hz	BL201	L	1	13.26	9.74	9	6	17
			2	16.53	10.30	13	7	24
			3	16.95	7.72	15	11	22
		H	1	10.27	7.82	11	3	18
			2	11.04	7.21	11	6	15
			3	7.38	4.56	6	5	9
	BL203	L	1	25.74	13.80	24	15	33
			2	23.88	11.82	22	15	30
			3	16.18	10.65	13	9	20
		H	1	5.69	5.84	4	0	8
			2	6.82	6.54	6	2	10
			3	7.65	7.31	6	1	13
0.1-9 Hz	BL205	L	1	27.62	18.01	24	16	35
			2	8.77	13.81	3	1	8
			3	54.55	23.22	52	38	67
		H	1	16.61	16.03	12	2	28
			2	11.07	10.88	9	3	16
			3	8.61	10.96	5	1	12
	BL206	L	1	30.62	14.37	29	19	48
			2	51.95	17.99	53	41	64
			3	49.22	18.86	53	33	64
		H	1	14.83	10.34	14	6	22
			2	20.66	11.58	20	14	28
			3	18.45	10.47	20	10	26
2-3 Hz	BL207	L	1	11.20	13.00	3	0	19
			2	7.37	9.98	4	0	11
			3	12.09	14.82	6	1	19
		H	1	3.17	9.71	0	0	0
			2	1.49	6.13	0	0	0
			3	4.92	13.08	0	0	1
	BL208	L	1	6.31	12.84	1	0	8
			2	4.11	11.60	0	0	2
			3	5.40	10.19	0	0	6
		H	1	2.13	7.33	0	0	0
			2	2.21	8.90	0	0	0
			3	1.29	4.85	0	0	0
2-9 Hz	BL210	L	1	5.07	9.32	0	0	8
			2	72.02	17.14	74	61	84
			3	92.18	13.59	91	82	102
		H	1	30.16	20.81	25	13	37
			2	12.12	12.57	10	3	17
			3	23.01	17.13	20	11	32




Figure 6. Median and inter-quartile ranges of movements per epoch from night 1.

Sleep-wake discrimination

The raw data displays illustrated in Tables 1 and 2, Figure 2, and Appendix 3 provide a broad overview of activity across several days. A more focused look at the data is illustrated in Figure 7 which dramatically illustrates the effect of amplifier passband on movement frequency distributions during one 24-hour period (Day 1). These figures plot the number of epochs containing a given number of movements as a joint function of number of movements and time of day, with movement frequency going from left to right, and time of day going from front to back. The top graph is data from subject BL201 with a passband of 0.1 - 3 Hz, and the bottom is from BL207 with a 2 - 3 Hz passband. In the top frame, the movement frequency distribution during waking periods





Figure 7. Movement frequency distributions for two actigraphs on day 1. See text for an explanation.

was fairly uniform, with a peak in the neighborhood of 20 movements per epoch. During the sleep period (2200-0600), the distributions shifted to the left, occasionally peaking at 0 movements during sleep. The bottom graph shows a much clearer difference between sleep and waking distributions, with the waking distributions showing no clear peak, and the sleep distributions unambiguously peaking at zero. Appendix 4 presents similar plots for each subject for the three 24-hour periods of the study beginning at 1800 on day 1.

To further clarify the differences between passbands and thresholds in sleep/wake discrimination, Table 4 shows the average sleep/wake ratios for nights 1 - 3. The values were obtained by dividing mean movements per epoch from a given night by mean movements per epoch for the following day.

Table 4
Sleep/Wake Ratios

<u>Passband (Hz)</u>	<u>Subject</u>	<u>Low Threshold</u>	<u>High Threshold</u>
0.1 - 3	BL201	0.61	0.47
	BL203	1.05	0.37
0.1 - 9	BL205	0.84	0.33
	BL206	1.04	0.49
2 - 3	BL207	0.19	0.09
	BL208	0.10	0.06

Two effects are suggested by the data in this table. First, regardless of the passband, in *all* cases the ratios are lower for the high threshold than for the low threshold. Secondly, as was suggested by Figure 7, the ratios are *much* lower for 2 - 3 Hz than for the other passbands, suggesting that the 2 - 3 Hz passband is the band of choice for distinguishing sleep from wake.

Automatic discrimination of sleep from wake

There have been a variety of efforts to develop computer programs to automatically discriminate sleep from wake based on actigraph data (Cole and Kripke, 1988; Cole, Kripke, Gruen, Mullaney, and Gillin, 1992; Kripke, Webster, Mullaney, Messin, and Fleck, 1983; Mullaney, Kripke, and Messin, 1980; Pleban, Valentine, Penetar, Redmond, and Belenky, 1990; Sadeh, Alster, Urbach and Lavie, 1989; Webster, Kripke, Messin, Mullaney, and Wyborney, 1982). All of these techniques involve derivation of complex weighted sliding averages to determine whether or not a given epoch is to be scored as sleep or wake. These techniques perform reasonably well, typically achieving 80%-90% agreement with polygraphically-scored sleep. Actigraphic sleep scoring systems typically fail, that is incorrectly score epochs as sleep, during periods of low activity during the day (reading or watching TV), and during sleep onset (depression in movement prior to onset of EEG-defined sleep) (Elsmore, 1992; Tryon, 1991).

The data in this paper show that activity measurements are heavily dependent on the amplifier settings (i.e. passband, threshold, and presumably gain, though that was not varied in this study) of the actigraph. Thus, algorithms for automatic determination of sleep from activity data will similarly be dependent on amplifier settings. The studies studied cited in the preceding paragraph used actigraphs whose analog characteristics approximate setting 18 of the AMA-32 actigraph, thus, in the present study the extant activity/sleep algorithms can only be expected to apply to the data obtained with that setting, i.e. the high threshold actigraphs for subjects BL207 and BL208. Table 5 summarizes the results of applying three different sleep-scoring algorithms to these data from the first three sleep periods. All three algorithms were developed for data collected in 1-minute epochs. Since in this study, the data were collected in 30-second epochs,

successive epochs were combined prior to applying the algorithms. Standard EEG records were collected during all sleep periods, and the records were scored for sleep stages according to the criteria set forth in Rechtschaffen and Kales (1968). For purposes of comparison with the actigraph-based scores, all epochs scored as sleep stages 1-4 and REM were considered to be sleep.

Table 5
Comparison of EEG and actigraph-based sleep measures

<u>Subject</u>	<u>Night</u>	Epochs scored as sleep				Percent agreement with EEG-based scores		
		<u>EEG</u>	<u>A1</u>	<u>A2</u>	<u>A3</u>	<u>A1</u>	<u>A2</u>	<u>A3</u>
BL207	1	850	908	920	682	93.5	92.2	75.4
	2	806	940	942	746	82.7	82.5	71.0
	3	904	870	888	650	86.2	87.2	65.0
BL208	1	845	930	930	696	89.6	89.2	72.6
	2	786	926	926	744	79.5	79.5	69.5
	3	750	960	956	694	78.1	78.1	63.3

A1 = Cole & Kripke
A2 = Sadeh et al.
A3 = Walter Reed

These data confirm the reports in the literature (Cole and Kripke, 1988; Cole, Kripke, Gruen, Mullaney, and Gillin, 1992; and Sadeh, Alster, Urbach and Lavie, 1989) that actigraphs can produce sleep data that is around 80 - 90% in agreement with EEG-based sleep measures when subjects are known to be in bed. The Walter Reed algorithm (Pleban et al., 1990) was much less accurate, yielding only around 65 - 75% accuracy. The probable reason for this is that this algorithm considers only the absolute activity level for the past five minutes, while the other two algorithms are much more context-sensitive, allowing for the normal movements that occur during sleep. However, this

feature also makes these algorithms more likely to produce false positives when the subject is known to be awake.

Table 6 shows the percent of epochs scored as sleep by all three algorithms when the subjects were known to be awake. Day 1 data represents the 24-hour period from 0800 following the first sleep period to 0800 on the next day. Days 2 and 3 represent the 16-hour periods commencing at 1600 on subsequent days. This table shows that the algorithms that perform best when the subject is known to be in bed and trying to sleep, perform worst when the subject is known to be awake. In this case, the Walter Reed algorithm is clearly superior.

Table 6
Percent of epochs incorrectly scored as sleep

<u>Subject</u>	<u>Day</u>	<u>A1</u>	<u>A2</u>	<u>A3</u>
BL207	1	12.2	26.0	3.5
	2	18.1	27.3	8.8
	3	16.2	26.5	8.2
BL208	1	9.7	26.4	1.8
	2	14.2	25.2	5.2
	3	17.5	32.3	3.3

Discussion and Conclusions

The initial goal of this project was to thoroughly evaluate the utility of the new features of the AMA-32 actigraph in both daytime and sleep recording. The first thing that became apparent was the lack of adequate software for analysis of actigraph data for research purposes. Hence the first year of the project was devoted to developing a general-purpose activity data analysis program that operates on IBM-compatible PC's. GAAP is the result of that effort, and copies will be provided by the first author upon

request. This program will enable widespread use of actigraph technology in both the military and civilian research and clinical communities, which, until now, were required to either enlist the services of an actigraphy specialist, write custom software to analyze actigraph data files, or purchase expensive commercial software designed for clinical applications.

The AMA-32 is a highly programmable device. The amplifier can be set to 5 different passbands, with high or low gain and high or low threshold. In addition the device can operate in zero-crossing mode or time-above threshold mode, 40 possibilities in all. The present study surveyed only six of these possibilities, three passbands with either high or low thresholds, with high gain setting in all cases. The passbands chosen represented low to medium (0.1 - 3 Hz), medium (2 - 3 Hz), and low to high (0.1 - 9 Hz) frequencies. Only the zero-crossing mode was used.

The clearest result of this study was that of the passbands and thresholds used, for discrimination of sleep from waking activity the medium (2 - 3 Hz) passband with high threshold (setting 18) was clearly superior to the other five combinations tested, even with the limited numbers of subjects in the study. Interestingly, this set of parameters most closely mimics that of the earlier, non-programable AMA-16 actigraph. The credit for this clearly goes to Daniel Redmond and Fredrick Hegge of the Walter Reed Army Institute of Research who did the developmental engineering for the AMA-16 (Redmond and Hegge, 1987).

The effects of threshold variations were also fairly clear, with higher thresholds usually producing fewer counts, and showing more sensitivity to sleep/wake conditions [see discussion of Figure 2 for explanation of the opposite possibility]. There may be circumstances, however, when lower thresholds might be preferable. For example, if one

is interested in movement during sleep, a high threshold may produce a floor effect, limiting the usefulness of the data.

In the present study, daytime activities were highly circumscribed. That is, the subjects were confined to a laboratory with little opportunity for physical activity. If one is interested in changes in daytime activity, non-traditional actigraph parameters may yield more useful data. In the absence of relevant research this issue remains unresolved.

Given the great differences produced by varying the actigraph parameters, it became obvious that attempts to apply automatic sleep scoring algorithms to data obtained with non-traditional actigraph parameters would not be meaningful. It is possible that new algorithms for processing these data could yield superior estimates of sleep parameters than the existing ones. This also remains a topic for future research.

It must be noted that the sleep-scoring algorithms that were surveyed are far from perfect. Those that performed well when the subjects were known to be in bed, produced many false positives (i.e. scored waking epochs as sleep) when the subjects were known to be awake, and vice versa. This is clearly an area in which a new approach is needed, particularly when accurate partitioning of sleep from wake is required. Two possible candidates are neural networks that are trained to discriminate sleep from wake, and sliding pattern recognition techniques.

Actigraphy is becoming a useful tool in the assessment of human behavior for clinical and research purposes. The AMA-32 is a device with great, though unrealized potential. A great deal of work remains to be done to fully understand how this device can best be exploited. GAAP provides the researcher and clinician with easy access to activity data, and the data presented here represent a beginning.

ACKNOWLEDGEMENTS

The valuable guidance and council of Dr. Fred Hegge during this project is gratefully acknowledged. Dr. Tamsin Kelly was the principal investigator on the Bright Light study. Without her gracious cooperation these data could not have been collected. The technical assistance of Steve Gomez, Matthew Sinclair, Scott Engelland, and Karen Schlangen is also gratefully acknowledged. Special thanks to Lorene Irwin who scored the EEG sleep records.

REFERENCES

- Brooks, C. J., Shergold, D. J., Angus, R. G., Heslegrave, and Redmond, D. P. (1988). Actigraph measurements of work/sleep patterns during a navy operation. *Journal of the Royal Navy Medical Service*, 74, 153-164.
- Brown, A. C., Smolensky, M. H., D'Alonzo, G. E., and Redmond, D. P. (1990). Actigraphy: a means of assessing circadian patterns in human activity. *Chronobiology International*, 7, 125-133.
- Cole, R. J. and Kripke, D. F. (1988). Progress in automatic sleep/wake scoring by wrist actigraph. *Sleep Research*, 17, 331.
- Cole, R. J., Kripke, D. F., Gruen, W., Mullaney, D. J., and Gillin, J. C. (1992). Automatic sleep/wake identification from wrist activity, *Sleep*, 15, 461-469.
- Elsmore, T. F. (1992). Quarterly report, MIPR No. 91MM1505, Naval Health Research Center, 10 Feb 92.
- French, J. (1992). Personal communication, observations during operation Desert Shield.
- Godfrey, H. P. D., and Knight, R. G. (1984). The validity of actometer and speech activity measures in the assessment of depressed patients. *British Journal of Psychiatry*, 145, 159-163.

- Konigsberg, E. and Cushing, T. (1986). Chemical Defense User Safety System. Report of USAMRDC Contract DAMD1785C5257.
- Kripke, D. F., Webster, J. B., Mullaney, D. J., Messin, S., and Fleck, P. (1983). Measuring sleep by wrist actigraph. Final report of USAMRDC Contract DAMD1778C8040.
- Mattman, P., Loepfe, M., Scheitlin, T., Schmidlin, D., Gerne, M., Strauch, I., Lehmann, D, and Borbely, A. A. (1982). Day-time effects and motor activity after three benzodiazepene hypnotics. *Arzneimittel-Forschung/Drug Research*, 32, 461-465.
- Mullaney, D. J., Kripke, D. F., and Messin, S. (1980). Wrist-actigraphic estimation of sleep time. *Sleep*, 3, 83-92.
- Pleban, R. J., Valentine, P. J., Penetar, D. M., Redmond, D. P., and Belenky, G. L. (1990). Characterization of sleep and body composition changes during ranger training. *Military Psychology*, 2, 145-156.
- Porrino, L. J., Rappoport, J. L., Behar, D., Sceery, W., Ismund, D. R., and Bunney, Jr., W.E. (1983), A naturalistic assessment of the motor activity of hyperactive boys I. Comparison with normal controls. *Archives of General Psychiatry*, 40, 681-687.
- Rechtshaffen, A. and Kales, A. (1968). A manual of standardized terminology, techniques and scoring systems for sleep stages of human subjects. Washington, D. C., U. S. Government Printing Office.

- Redmond, D. P. and Hegge, F. W. (1985). Observations on the design and specification of a wrist-worn human activity monitoring system. *Behavior Research Methods, Instrumentation, and Computers*, 17, 659-669.
- Redmond, D. P. and Hegge, F. W. (1987). The design of human activity monitors. In L. E. Scheving, F. Halberg, and C. F. Ehret (Eds.) *Chronobiotechnology and Chronobiological Engineering*, Boston: Martinus Nijhoff, pp 202-215.
- Sadeh, A., Alster, J., Urbach, D., and Lavie, P. (1989). Actigraphically based automatic sleep-wake scoring: Validity and clinical applications. *Journal of Ambulatory Monitoring*, 2, 209-216.
- Tryon, Warren W. (1991). *Activity Measurement in Psychology and Medicine*, New York: Plenum Press.
- Webster, J. B., Kripke, D. F., Messin, S., Mullaney, and Wyborney, G. (1982). An activity-based sleep monitor system for ambulatory use. *Sleep*, 5, 389-399.

Appendix 1

Appendix 1: AMA-32 amplifier settings

Option	FR	G	T	Option	FR	G	T	Option	FR	G	T
1	.1-1	H	L	8	.1-3	L	H	13	2-3	L	L
2	.1-1	H	H	9	.1-9	H	L	14	2-3	L	H
3	.1-1	L	L	10	.1-9	H	H	19	2-9	H	L
4	.1-1	L	H	11	.1-9	L	L	20	2-9	H	H
5	.1-3	H	L	12	.1-9	L	H	15	2-9	L	L
6	.1-3	H	H	17	2-3	L	H	16	2-9	L	H
7	.1-3	L	L	18	2-3	H	H				

FR: Frequency response in HZ

G: Gain, L = 5; H = 26

T: Threshold, L = 6mV; H = 24mV

Appendix 2

General Activity Analysis Program (Version 0.95): User's Guide

GENERAL DESCRIPTION

The General Activity Analysis Program (GAAP) was designed for the analysis of data files generated with the AMA-32 actigraph manufactured by Precision Control Designs Inc. (PCD) of Ft. Walton Beach FL. It will also work with files generated by the AMA-16 actigraph, also manufactured by PCD. It provides a means of quickly viewing and summarizing activity data, application of sleep scoring algorithms to the data, generating epoch-by-epoch ASCII files, and generating hard copy tabular ASCII and graphical (HPGL) output files. The program runs on IBM-compatible PC's and requires a VGA monitor and a mouse.

GAAP is (almost) entirely menu-driven, with menu selections made by pointing and clicking the mouse. This file is available for help on-line. However, it was hoped that the program menu choices would be intuitively obvious, enabling the knowledgeable user to quickly master the fundamentals and get on with the real work. If this is not the case, this Users Guide should help clear up some of the gaps in GAAP.

GAAP MENU STRUCTURE

I. MAIN MENU

- A. Input New File
- B. Process Data
 - 1. Sleep Scoring
 - a. Cole & Kripke
 - b. Sadeh et al.
 - c. Walter Reed
 - 2. Set Markers
 - 3. Display/Edit
 - a. Wide View \$
 - b. Zoom/Edit
 - 4. Quit
- C. Output
 - 1. Display/Edit (same as above)
 - 2. Sleep Periods !
 - a. Sleep Summary by Day
 - b. Sleep Summary by Hour
 - c. Distributions by Day
 - d. Distributions by Hour
 - 3. Wake Periods !

- a. Summary by Day
- b. Summary by Hour
- c. Distributions by Day
- d. Distributions by Hour
- 4. HPGL Output File \$
- 5. Epoch-by-epoch File
- 6. Quit
- D. Help
- E. Quit

II. SCALE MENU

- A. Auto Scale
- B. Force Scale

III. OUTPUT DESTINATION MENU

- A. Screen
- B. Printer
- C. File

Note:

- \$ Automatically calls Scale Menu
- ! Automatically calls Output Destination Menu

MENU DETAILS

I. MAIN MENU

- A. Input New File. You will be prompted to set the path for data files, initially set to the default directory. A list of all "DAT" files is then displayed. Highlight the desired one, and double-click to select.
- B. Process Data. Perhaps a misnomer, but ...
 - 1. Sleep Scoring. Implements three published algorithms for scoring sleep from activity data. All scoring is done using 60-second epochs since that is the way the algorithms were written. To see the results you need to use the display/edit option, the output sleep options, or the epoch-by-epoch file option.
 - a. Cole, R. J. and Kripke, D. F. (1988). Progress in automatic sleep/wake scoring by wrist actigraph. *Sleep Research*, 17, 331.
 - b. Sadeh, A., Alster, J., Urbach, D., and Lavie, P. (1989). Actigraphically based automatic sleep-wake scoring: Validity and clinical applications. *Journal of Ambulatory Monitoring*, 2, 209-216.
 - c. Pleban, R. J., Valentine, P. J., Penetar, D. M., Redmond, D. P., and Belenky, G. L. (1990). Characterization of sleep and body composition changes during ranger training. *Military Psychology*, 2, 145-156.

Appendix 2

2. Set Markers. The "Markers" are times the program needs to do its business. The markers and defaults are as follows:

Marker	Value	Use
1. Day start time:	1200	Time at which 24-hour plots begin
2. Analysis start (hour):	1800	2-5 are not used
3. Analysis start (day):	1	
4. Analysis end (hour):	1200	
5. Analysis end (day):	6	
6. Sleep period start:	2200	Begin sleep summary
7. Sleep period end:	0600	End sleep summary
8. Wake period start:	0600	Begin wake summary
9. Wake period end:	2200	End wake summary

3. Display / Edit. Screen displays of activity data.

- Wide View. Displays up to 10 full days of data, one day per line. If a sleep scoring algorithm has been run, the epochs scored as sleep will be displayed in red immediately below the line for each day.
- Zoom/Edit. Displays 12 hours of data, one hour per line. File can be paged through by clicking on "Forward" and "Back" buttons. Using the mouse, minute-by-minute activity scores are displayed. Results of the sleep-scoring algorithm are also displayed below each epoch (red = sleep). The sleep scores can be edited using the mouse. Pressing the left button sets the sleep score for the current epoch to sleep, and the right button sets it to wake. If no sleep scoring had been done to the present file, all epochs will be set to sleep.

4. Quit. Returns to main menu

C. Output

- Display / Edit. Same as 3 above
- Sleep periods. This provides summaries of activity only during the period of time bracketed by markers 6 and 7. The OUTPUT DESTINATION routine will automatically be called. If FILE output is chosen, the default filename will have the same name as the input data file, but the extension "SLP". See III. below for more discussion on file names.
 - Sleep summary by day. Minutes of sleep and wake, number, mean duration, and longest sleep episode, total activity counts, mean, standard deviation, first quartile, median, and third quartile of counts per epoch.
 - Sleep summary by hour. Same as a. but hourly.
 - Distributions by day. Displays a distribution (14 bins + overflow bin) for each day of counts per epoch. Bin size is selectable, and the user is queried regarding bin size each time this routine is run. The default is 10. Thus, bin 1 has the number of epochs containing 0-9 counts, bin 2, 10-19, etc.

d. Distributions by hour. Same as c. but hourly.

3. Wake periods. Summaries of activity only during the period of time bracketed by markers 8 and 9. As above, the OUTPUT DESTINATION routine will automatically be called. If FILE output is chosen, the default filename will have the same name as the input data file, but the extension "WAK".

- a. Summary by day. Total activity, mean, standard deviation, first quartile, median, and third quartile of counts per epoch.
- b. Summary by hour. Same as a. but hourly.
- c. Same as 2c above.
- d. Same as 2d above.

4. HPGL output file. Produces a file of HPGL (Hewlett Packard Graphics Language) commands which will generate a hard copy of the "wide view" display discussed above (I.B.3.a.). This file can be interpreted by any number of graphics programs, or printed by utilities such as PRINTGL, which is a shareware program. The file can also be directly printed on a HP pen plotter. These files will have names like "name.plt" where name is the same as the "DAT" file being processed. (i.e. if you are working with file "ACTIVITY.DAT", the plot file will be "ACTIVITY.PLT". These files tend to become VERY long, easily reaching 450k bytes in length. Thus they should either be deleted after use or archived, then deleted to prevent cluttering up your hard disk. This routine also automatically calls the SCALE MENU.

5. Epoch-by-Epoch file. Creates an ASCII file with one record per epoch. Fields in the record are:

- epoch number
- time of day
- number of counts
- sleep score (1=wake, 0=sleep)

NOTE: if no sleep scoring has been done, this field is eliminated.

The output filename will be the same as the input file name, but have the extension "ASC". Again, these files can become very long, so good disk hygiene is warranted.

6. Quit. Return to main menu.

D. Help. Reads this file on-line.

E. Quit. Exit program.

II. SCALE MENU. This menu allows you to set the scale for 24- hour displays or plots. The scale value is the value which will be plotted as 80% of the distance between successive days on the multi-day plot.

A. Auto Scale. The scale value is set to the maximum value in the input file.

B. Force Scale. Allows input of a scaling factor. If several plots are to be compared, this is the preferred method, since it will allow them all to be plotted with the same scale. The input value is retained between plots.

III. OUTPUT DESTINATION MENU. Called by all routines discussed above in I.C.2-4 above.

A. Screen.

B. Printer. No form feed is sent. Thus, multiple calls can print data on the same sheet of paper.

C. File. If the same file name is used in successive calls, the new data will be APPENDED to the old, permitting you to build up a data file with results from multiple subjects, multiple runs from the same subject, different marker parameters, etc.

GAAP Restrictions

1. "DAT" FILES ONLY. PCD provides the program "PCD.EXE" which can save data in either of two formats, "PCD" or "DAT". The PCD format is the more compact of the two, since it stores the data in a "packed" fashion, very much the way it is actually stored internally in the actigraph memory during data collection. The DAT format expands the data such that the activity from each epoch is stored as two bytes (16 bits), making decoding of the data much simpler. If your data are saved as PCD files, you will have to read it into the PCD program and write it back out as DAT files.

2. SLEEP SCORING FOR 30 AND 60 SECOND EPOCHS ONLY. A truly general program would not have this restriction, but due to the fact that the extant sleep algorithms require a one-minute epoch, the decision was made to keep this restriction. For sleep analysis with 30-second epochs, data from successive 2-epoch blocks is used.

3. VGA ONLY. Sorry, if your system doesn't have at least VGA, you're out of luck.

4. MOUSE REQUIRED. I hate keyboards. Actually GAAP requires keyboard input in a few places. It was unavoidable.

An Actigraph Primer

The AMA-32 actigraph is a wristwatch-sized device which records movement. The AMA-16 is somewhat larger. The active element in the device is called an "accelerometer" which generates a voltage when it is moved through space. The actigraph contains electronic circuitry for sensing this analog voltage and converting it to digital pulses which are in turn stored in an on-board memory. In the AMA-32 (but not the AMA-16), the parameters of this analog-to-digital conversion (frequency passband, amplifier gain, and voltage threshold) are programmable. Details are available in the documentation for PCD.EXE. For reliable operation, the AMA-32 battery should be replaced after about two weeks of data collection. During data collection, activity is accumulated in "epochs" of time. That is, the number of movements occurring in each epoch is counted and stored.

Interaction with the AMA-32 is accomplished with an Actigraph Interface Unit (AIU) connected to a serial port on an IBM-compatible PC. The menu-driven PCD.EXE program has routines for initializing actigraphs, monitoring their operation, and downloading data from the actigraph into the PC. Initialization parameters include a "packing option" (memory bits per epoch), the data collection epoch, data collection start time and date, current time and date, amplifier setting, and data identifier. The duration of a data collection run depends on the amount of memory in a given actigraph (16 or 32K) and

Appendix 2


the data packing option. In general, since more activity occurs in longer epochs, more memory space is required to store the data, thus requiring a larger packing option (more bits per epoch). The most commonly used epoch duration is 60 seconds. With a 16K memory and 60 sec epochs, approximately 11 days of data can be collected before the actigraph memory is filled.

The AMA-32 can actually be operated in two different modes, Zero Crossing (ZC) and Time Above Threshold (TAT). Virtually all of the work that has been done has used the ZC mode. In this mode, two counts are generated for each movement, one when the voltage output from the accelerometer exceeds the threshold, and one when it falls back below the threshold. While usually counts begin and end in the same epoch and only occasionally, will a movement will begin in one epoch and end in the next. This results in an inappropriately high number of epochs with an even number of counts. To correct for this artifact, if ZC mode is in effect, the data from each epoch are corrected while the file is being read, resulting in only one count per movement. The formula for this correction is:


$$\text{NEW_VALUE} = \text{INTEGER} ((\text{OLD_VALUE} + 1) / 2)$$

In TAT mode, movement duration is recorded by accumulating the time in seconds that the voltage remains above threshold.

Appendix 3




File: BL201-5.DAT, Epoch: 30, Amplifier: 05, Scale: 100
Starting Time: Jul 13, 1992, 14:00:00, SN:0053




File: BL201-6.DAT, Epoch: 30, Amplifier: 06, Scale: 100
Starting Time: Jul 13, 1992, 14:00:00, SN:0057

Appendix 3




File: BL203-5.DAT, Epoch: 30, Amplifier: 05, Scale: 100
Starting Time: Jul 13, 1992, 14:00:00, SN: 61




File: BL203-6.DAT, Epoch: 30, Amplifier: 06, Scale: 100
Starting Time: Jul 13, 1992, 14:00:00, SN: 0062

Appendix 3




File: BL205-9.DAT, Epoch: 30, Amplifier: 09, Scale: 100
Starting Time: Jul 20, 1992, 15:30:00, SN:0056




File: BL205-10.DAT, Epoch: 30, Amplifier: 10, Scale: 100
Starting Time: Jul 20, 1992, 14:00:00, SN:0060

Appendix 3




File: BL206-9.DAT, Epoch: 30, Amplifier: 09, Scale: 100
Starting Time: Jul 20, 1992, 14:00:00, SN: 61




File: BL206-10.DAT, Epoch: 30, Amplifier: 10, Scale: 100
Starting Time: Jul 20, 1992, 14:00:00, SN:0062


Appendix 3




File: BL207-17.DAT, Epoch: 30, Amplifier: 17, Scale: 100
Starting Time: Aug 10, 1992, 12:00:00, SN:0053



File: BL207-18.DAT, Epoch: 30, Amplifier: 18, Scale: 100
Starting Time: Aug 10, 1992, 12:00:00, SN:0054




File: BL208-17.DAT, Epoch: 30, Amplifier: 17, Scale: 100
Starting Time: Aug 10, 1992, 12:00:00, SN:0055




File: BL208-18.DAT, Epoch: 30, Amplifier: 18, Scale: 100
Starting Time: Aug 10, 1992, 12:00:00, SN:0056







Appendix 3









File: BL210-19.DAT, Epoch: 30, Amplifier: 19, Scale: 100
Starting Time: Aug 16, 1992, 18:00:00, SN:0053









File: BL210-20.DAT, Epoch: 30, Amplifier: 20, Scale: 100
Starting Time: Aug 16, 1992, 18:00:00, SN:0054









Appendix 4















Appendix 4









Appendix 4












Subject BL201, 7/13/92




Subject BL203, 7/13/92




Subject BL205, 7/20/92




Subject BL206, 7/20/92



Subject BL210, 8/17/92



Subject BL207, 8/20/92



Subject BL208, 8/20/92

