

AD _____

MIPR NO: 93MM3579


TITLE: NAVAL BLOOD RESEARCH LABORATORY ANALYSIS OF REFLUPS
BAGSETS

PRINCIPAL INVESTIGATOR(S): C. R. VALERI
AL EDWARDS

CONTRACTING ORGANIZATION: NAVAL BLOOD RESEARCH LABORATORY
BOSTON, MA 02118

REPORT DATE: DECEMBER 10, 1993

TYPE OF REPORT: FINAL



PREPARED FOR: Commander
U.S. Army Medical Research and Materiel Command
Fort Detrick, Maryland 21702-5012

DISTRIBUTION STATEMENT: Approved for public release;
distribution unlimited

The views, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy or decision unless so designated by other documentation.

19950731 020

REPORT DOCUMENTATION PAGE

Form Approved
OMB No. 0704-0188

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.

| | | |
|----------------------------------|-----------------------------|----------------------------------|
| 1. AGENCY USE ONLY (Leave blank) | 2. REPORT DATE 10 Dec 93 | 3. REPORT TYPE AND DATES COVERED |
|----------------------------------|-----------------------------|----------------------------------|

| | |
|--------------------------------------------------------------------------------------|--------------------------------|
| 4. TITLE AND SUBTITLE Naval Blood Research Laboratory Analysis of Reflups Bagsets | 5. FUNDING NUMBERS 93MM3579 |
|--------------------------------------------------------------------------------------|--------------------------------|

| | |
|--------------------------------------------|--|
| 6. AUTHOR(S) C. R. Valeri Al Edwards | |
|--------------------------------------------|--|

| | |
|----------------------------------------------------------------------------------------------------------------------|------------------------------------------|
| 7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Naval Blood Research Laboratory Boston, Massachusetts 02118 | 8. PERFORMING ORGANIZATION REPORT NUMBER |
|----------------------------------------------------------------------------------------------------------------------|------------------------------------------|

| | |
|-------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------|
| 9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) U.S. Army Medical Research and Materiel Command Fort Detrick, Maryland 21702-5012 | 10. SPONSORING/MONITORING AGENCY REPORT NUMBER |
|-------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------|

| |
|-------------------------|
| 11. SUPPLEMENTARY NOTES |
|-------------------------|

| | |
|---------------------------------------------------------------------------------------------------|------------------------|
| 12a. DISTRIBUTION / AVAILABILITY STATEMENT Approved for public release, distribution unlimited | 12b. DISTRIBUTION CODE |
|---------------------------------------------------------------------------------------------------|------------------------|

| | | | | | | | | | | | | | | | | | | | | | |
|----------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------|--|------------|-------------------------------------|----------|--------------------------|-------------|--------------------------|---------------------|--|----------|--|----------------|--|--------------------|--|------|-------------------------|-----|--|
| 13. ABSTRACT (Maximum 200 words) | <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td colspan="2" style="padding: 2px;">Accession For</td> </tr> <tr> <td style="padding: 2px;">NTIS CRA&I</td> <td style="padding: 2px; text-align: center;"><input checked="" type="checkbox"/></td> </tr> <tr> <td style="padding: 2px;">DTIC TAB</td> <td style="padding: 2px; text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td style="padding: 2px;">Unannounced</td> <td style="padding: 2px; text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td colspan="2" style="padding: 2px;">Justification</td> </tr> <tr> <td colspan="2" style="padding: 2px;">By</td> </tr> <tr> <td colspan="2" style="padding: 2px;">Distribution /</td> </tr> <tr> <td colspan="2" style="padding: 2px; text-align: center;">Availability Codes</td> </tr> <tr> <td style="padding: 2px; text-align: center;">Dist</td> <td style="padding: 2px; text-align: center;">Avail and/or Special</td> </tr> <tr> <td style="padding: 2px; text-align: center;">A-1</td> <td style="padding: 2px;"></td> </tr> </table> | Accession For | | NTIS CRA&I | <input checked="" type="checkbox"/> | DTIC TAB | <input type="checkbox"/> | Unannounced | <input type="checkbox"/> | Justification | | By | | Distribution / | | Availability Codes | | Dist | Avail and/or Special | A-1 | |
| Accession For | | | | | | | | | | | | | | | | | | | | | |
| NTIS CRA&I | <input checked="" type="checkbox"/> | | | | | | | | | | | | | | | | | | | | |
| DTIC TAB | <input type="checkbox"/> | | | | | | | | | | | | | | | | | | | | |
| Unannounced | <input type="checkbox"/> | | | | | | | | | | | | | | | | | | | | |
| Justification | | | | | | | | | | | | | | | | | | | | | |
| By | | | | | | | | | | | | | | | | | | | | | |
| Distribution / | | | | | | | | | | | | | | | | | | | | | |
| Availability Codes | | | | | | | | | | | | | | | | | | | | | |
| Dist | Avail and/or Special | | | | | | | | | | | | | | | | | | | | |
| A-1 | | | | | | | | | | | | | | | | | | | | | |

| | |
|-------------------|---------------------------|
| 14. SUBJECT TERMS | 15. NUMBER OF PAGES 79 |
| | 16. PRICE CODE |

| | | | |
|-------------------------------------------------------|----------------------------------------------------------|---------------------------------------------------------|-----------------------------------------|
| 17. SECURITY CLASSIFICATION OF REPORT Unclassified | 18. SECURITY CLASSIFICATION OF THIS PAGE Unclassified | 19. SECURITY CLASSIFICATION OF ABSTRACT Unclassified | 20. LIMITATION OF ABSTRACT Unlimited |
|-------------------------------------------------------|----------------------------------------------------------|---------------------------------------------------------|-----------------------------------------|

GENERAL INSTRUCTIONS FOR COMPLETING SF 298

The Report Documentation Page (RDP) is used in announcing and cataloging reports. It is important that this information be consistent with the rest of the report, particularly the cover and title page. Instructions for filling in each block of the form follow. It is important to *stay within the lines* to meet optical scanning requirements.

Block 1. Agency Use Only (Leave blank).

Block 2. Report Date. Full publication date including day, month, and year, if available (e.g. 1 Jan 88). Must cite at least the year.

Block 3. Type of Report and Dates Covered. State whether report is interim, final, etc. If applicable, enter inclusive report dates (e.g. 10 Jun 87 - 30 Jun 88).

Block 4. Title and Subtitle. A title is taken from the part of the report that provides the most meaningful and complete information. When a report is prepared in more than one volume, repeat the primary title, add volume number, and include subtitle for the specific volume. On classified documents enter the title classification in parentheses.

Block 5. Funding Numbers. To include contract and grant numbers; may include program element number(s), project number(s), task number(s), and work unit number(s). Use the following labels:

| | |
|----------------------|------------------------------|
| C - Contract | PR - Project |
| G - Grant | TA - Task |
| PE - Program Element | WU - Work Unit Accession No. |

Block 6. Author(s). Name(s) of person(s) responsible for writing the report, performing the research, or credited with the content of the report. If editor or compiler, this should follow the name(s).

Block 7. Performing Organization Name(s) and Address(es). Self-explanatory.

Block 8. Performing Organization Report Number. Enter the unique alphanumeric report number(s) assigned by the organization performing the report.

Block 9. Sponsoring/Monitoring Agency Name(s) and Address(es). Self-explanatory.

Block 10. Sponsoring/Monitoring Agency Report Number. (if known)

Block 11. Supplementary Notes. Enter information not included elsewhere such as: Prepared in cooperation with...; Trans. of...; To be published in.... When a report is revised, include a statement whether the new report supersedes or supplements the older report.

Block 12a. Distribution/Availability Statement. Denotes public availability or limitations. Cite any availability to the public. Enter additional limitations or special markings in all capitals (e.g. NOFORN, REL, ITAR).

DOD - See DoDD 5230.24, "Distribution Statements on Technical Documents."

DOE - See authorities.

NASA - See Handbook NHB 2200.2.

NTIS - Leave blank.

Block 12b. Distribution Code.

DOD - Leave blank.

DOE - Enter DOE distribution categories from the Standard Distribution for Unclassified Scientific and Technical Reports.

NASA - Leave blank.

NTIS - Leave blank.

Block 13. Abstract. Include a brief (*Maximum 200 words*) factual summary of the most significant information contained in the report.

Block 14. Subject Terms. Keywords or phrases identifying major subjects in the report.

Block 15. Number of Pages. Enter the total number of pages.

Block 16. Price Code. Enter appropriate price code (*NTIS only*).

Blocks 17. - 19. Security Classifications. Self-explanatory. Enter U.S. Security Classification in accordance with U.S. Security Regulations (i.e., UNCLASSIFIED). If form contains classified information, stamp classification on the top and bottom of the page.

Block 20. Limitation of Abstract. This block must be completed to assign a limitation to the abstract. Enter either UL (unlimited) or SAR (same as report). An entry in this block is necessary if the abstract is to be limited. If blank, the abstract is assumed to be unlimited.

TABLE OF CONTENTS:

| | Page # |
|---------------------------------------------------------------------------------|--------|
| Objective & Summary ----- | 4 |
| Analytical Parameters & Methods ----- | 5 |
| Definition of a Product Failure ----- | 8 |
| Results ----- | 10 |
| <u>Tables</u> | |
| Table 1 - Fill Volume Summarized Data ----- | 11 |
| Table 2 - Temperature Indicator & Reading Summarized Data ----- | 11 |
| Table 3 - Bag Leaks & Breakage Summarized Data ----- | 12 |
| Table 4 - Particulate Testing Summarized Data ----- | 13 |
| Table 5 - Sodium Concentration of REFLUPS Solutions Summarized data ----- | 13 |
| Table 6 - Potassium Concentration of REFLUPS Solutions Summarized data ----- | 14 |
| Table 7 - pH Values of REFLUPS Solutions Summarized data ----- | 14 |
| Table 8 - Lactate Concentration of REFLUPS Solutions Summarized data ----- | 15 |
| Table 9 - Chloride Concentration of REFLUPS Solutions Summarized data ----- | 15 |
| Table 10 - Osmolality of REFLUPS Solutions Summarized data ----- | 16 |
| Table 11 - Glucose Concentration of REFLUPS Solutions Summarized data ----- | 16 |
| Table 12 - Calcium Concentration of REFLUPS Solutions Summarized data ----- | 17 |
| Table 13 - Endotoxin Testing Summarized Data ----- | 17 |
| Table 14 - Sterility Testing Summarized Data ----- | 18 |

Appendixes

Appendix 1.1

Individual data for ARMY Manifolds Lactated Ringers 500ml

Appendix 1.2

Individual data for ARMY Manifolds Lactated Ringers 1000ml

Appendix 1.3

Individual data for ARMY Manifolds 0.9% NaCl

Appendix 1.4

Individual data for ARMY Manifolds 0.9% NaCl 0.2% Dextrose

Appendix 2.1

Fill Volume data for ARMY Manifolds Lactated Ringers 500ml

Appendix 2.2

Fill Volume data for ARMY Manifolds Lactated Ringers 1000ml

Appendix 2.3

Fill Volume data for ARMY Manifolds 0.9% NaCl

Appendix 2.4

Fill Volume data for ARMY Manifolds 0.9% NaCl 0.2% Dextrose

TEST AND EVALUATION OF THE EDM REFLUPS SYSTEM USED IN THE
 PREPARATION OF RESUSCITATIVE FLUIDS AT US ARMY TEST SITE - FORT
 SAM HOUSTON, SAN-ANTONIO, TX.

OBJECTIVE : This study was done to evaluate the end products
 produced by the EDM REFLUPS system at the US ARMY test site at
 Fort Sam Houston, San Antonio, TX.

The study period for the production of the fluids was September
 1993 to October 1993 and evaluation of the fluids is now ongoing
 at the Naval Blood Research Laboratory.

SUMMARY: NBRL received 155 manifolds for Testing.

The distribution type was as follows;

| | |
|-----------------------------------------------------------|-----------------------|
| 0.9% Sodium Chloride 0.2% Dextrose Red Cell Wash (RCW) | = 29 manifolds |
| 0.9% Sodium Chloride (NaCl) | = 43 manifolds |
| Lactated Ringers 0.5 liter (LR5) | = 48 manifolds |
| Lactated Ringers 1 liter (LR1) | = 41 manifolds |
| <u>TOTAL</u> | <u>=161 Manifolds</u> |

Fill Volume has been measured on all bags received.
 Solutions prepared with the EDM Model REFLUPS system at
 US ARMY test site at Fort Sam Houston, San Antonio,
 TX and tested to date were shown to be Sterile and Endotoxin
 free with a chemical composition equivalent to commercial
 solutions.

Defects were observed in 17 of 2621 bags (see page 12).

(1) Analytical Parameters and Methods

0.9% NaCl

| Analytical Parameter | USP Method | Equivalent Method | Method chosen by NBRL | Rationale |
|----------------------|--------------------------------|--------------------------------------------------------|-----------------------|--------------------------------------------------------------------|
| Sodium | * <191> Flame Photometer | --- | USP | |
| pH | * <791> pH Test | ---- | USP | |
| Volume | * <1> Weight | -- -- | USP | |
| Sterility | * <71> Membrane Filtration | ---- | USP | |
| Particulate Matter | * <788> Membrane Filtration | Electronic liquid borne Particle Counter System - HIAC | HIAC | Less subjective more accurate method. USP equivalency established. |
| Pyrogen | * <151> Pyrogen test | LAL -Pyrotell Method | LAL | A more Quantitative & sensitive test USP equivalency established |
| Chloride | Titration | | USP | Using a Buchler Chloridometer |
| Osmolality | | Freezing point Depression | not required USP | |

*Refers to Relevant Paragraph/page of U.S. Pharmacopeia (USP) XXII.

Lactated Ringers (1L and 1/2L)

| Analytical Parameter | USP Method | Equivalent Method | Method chosen by NBRL | Rationale |
|----------------------|-----------------------------------|-----------------------------------------------------------------|-----------------------|--------------------------------------------------------------------|
| Sodium & Potassium | * <191> Flame Photometer | --- | | USP |
| Lactate * | Spectrophotometric Assay - Manual | Automated Spectrophotometric COBAS FARA Chemistry System | COBAS | A more Quantitative test USP Equivalency established |
| Total Calcium | *pg 1228 Manual Colormetric | COBAS FARA CHEMISTRY SYSTEM Spectrophotometer | COBAS | USP Equivalency Established |
| pH | * <791> pH Test | ---- | | USP |
| Volume | * <1> Weight | ---- | | USP |
| Sterility | * <71> Membrane Filtration | ---- | | USP |
| Particulate Matter | * <788> Membrane Filtration | Electronic liquid borne Particle Counter System - HIAC | HIAC | Less subjective more accurate method. USP equivalency established. |
| Pyrogen | * <151> Pyrogen test | LAL -Pyrotell Method | LAL | More Quantitative & sensitive test. USP equivalency established |
| Osmolality | | Freezing point Depression | Not required by USP | |
| Chloride | Titration | | USP | Using a Buchler Chloridometer |

*Refers to Relevant Paragraph/page of U.S. Pharmacopeia (USP) XXII.

0.2% Dextrose and 0.9% NaCl

| Analytical Parameter | USP Method | Equivalent Method | Method chosen by NBRL | Rationale |
|----------------------|-------------------------------|---------------------------------------------------------------|-------------------------|-----------------------------------------------------------------|
| Sodium & Potassium | *<191> Flame Photometer | --- | USP | |
| pH | *<791> pH Test | ---- | USP | |
| Volume | *<1> Weight | ---- | USP | |
| Sterility | *<71> Membrane Filtration | ---- | USP | |
| Particulate Matter | *<788> Membrane Filtration | Electronic liquid borne Particle Counter System - HIAC | HIAC | Less subjective & more accurate method. USP equivalent |
| Osmolality | | Freezing point Depression | not a USP required test | |
| Pyrogen | *<151> Pyrogen test | LAL -Pyrotell Method | LAL | More Quantitative & sensitive test. USP equivalency established |
| Chloride | Titration | | USP | Using a Buchler Chloridometer |
| Glucose | Manual Spectrophotometric | Automated Spectrophotometric Using the ABBOTT VP Super System | VP Super System | USPEquivalency Established |

*Refers to Relevant Paragraph/page of U.S. Pharmacopeia (USP) XXII.

USP states that "Compliance, may be determined, by the use of alternative methods, chosen for advantages in accuracy, sensitivity, precision & selectivity"

2) DEFINITION OF A PRODUCT FAILURE:

STERILITY- Sterility was evaluated using the Millipore Steritest. The Aerobic media used was a Tryp Soy Broth & the Anaerobic media used was a Fluid Thio Glycollate Broth. Sterility tests were done on, one bag in the 18 one liter manifold, one bag in the 18 half liter manifold & one bag from the 6 three liter manifold. If the first bag in the manifold of 18 bags tests positive then repeat tests will be done on 3 additional bags from this manifold. If the first bag in the manifold of 6 bags tests positive, then repeat test was done on 1 other bag in the manifold. A positive finding in the additional sample renders the manifold a product failure.

(If surplus bags are available, further testing is carried out to provide an insight into any positive results.)

NOTE: All Sterility testing was carried out in a certified clean room to reduce the possibility of exogenously contaminating the solutions as they were being tested.

ENDOTOXIN - Endotoxin (pyrogens) was measured using LAL-500 (Associates of Cape Cod, MA) and the kinetic Turbidmetric assay. Samples for endotoxin testing were obtained from one bag of the 18 one liter manifold, one bag of the 18 half liter manifold and from one bag of the 6 three liter manifold are tested for endotoxin. If the sample from the first bag is positive , 3 other bags from the same manifold containing 18 bags or one other bag from the same manifold containing 6 bags, are retested for endotoxin. If the second bag is also positive, the manifold is considered a product failure.

CHEMICAL COMPOSITION - Analysis for chemical composition was made in one bag in the 18 one liter manifold , one bag in the 18 half liter manifold and one bag from the 6 three liter manifold. If the sample from the first bag is outside the acceptable range¹, 3 other bags from the same manifold containing 18 bags or one other bag from the same manifold containing 6 bags, are retested for chemical composition. If the second bag is also outside the acceptable range, the manifold is considered a product failure.

Note: The osmolality specifications for the REFLUPS solutions were provided to NBRL by STERIMATICS Corporation and were as follows;

NaCl 273-310 mOSM

Lactated Ringers 241-267 mOSM

Red Cell Wash 281-311 mOSM

PARTICULATE MATTER - Particulate Matter was analyzed using a particle size analyzer, calibrated using NIST Certified particle-size standards (Duke Scientific, CA) . Samples are analyzed four consecutive times, using a volume of 5ml. Data from the first sample is discarded and the final three counts are averaged. If there are more than 50 particles per ml that are 10 microns or larger or 5 particles that are 25 microns or larger, the solution is considered a product failure.

1 "ACCEPTABLE RANGE" IS DETERMINED BY USP GUIDELINES I.E. $\pm 5\%$ OF LABELED VALUE. IF LABELLED VALUE IS EXPRESSED AS A RANGE THEN A VALUE IS ACCEPTABLE ONLY IF IT IS WITHIN THE SPECIFIED RANGE (E.G. PH, CALCIUM, OSMOLALITY, VOLUME). IF THE FINAL RESULT IS NOT WITHIN $\pm 5\%$ OF LABELED VALUE, THEN A DETERMINATION OF PRODUCT FAILURE IS BASED ON THE EQUIVALENCY OF THE REFLUPS SOLUTIONS TO COMMERCIAL SOLUTIONS, MEASURED BY NBRL.

FILL VOLUME : The volume of the Reflups solutions was determined by weighing the overwrapped bag, deducting the weight of the overwrap & bag and then dividing this weight by the density of the solution. The densities used were those measured by NBRL and are as follows;

0.9% Sodium Chloride = 1.0036g/ml

Lactated Ringers = 1.0033g/ml

0.9% Sodium Chloride 0.2% dextrose = 1.0045g/ml

Fill volume specifications are those provided to NBRL by STERIMATICS Corporation for stored product , as follows:

500 - 560 ml for 1/2 liter manifolds

1000 - 1090 ml for 1 liter manifolds

3000 - 3180 ml for 3 liter manifolds

ANY BAG OUTSIDE THE ABOVE RANGE WAS CONSIDERED A PRODUCT FAILURE.

RESULTS:

Summary Results obtained on solutions prepared using the EDM REFLUPS are reported in Tables 1 to 14. Individual values for each parameter are reported in Appendix 1. Volume measurements for each REFLUPS bag are reported in Appendix 2.

Solutions prepared with the EDM Model REFLUPS system at the US Army Test site at FORT SAM HOUSTON, TX., tested to date, were shown to be Sterile and Endotoxin free with a chemical composition equivalent to commercial solutions.

Defects were observed in 17 of 2621 bags (see page 12).

TABLE 1**VOLUME OF ARMY REFLUPS SOLUTIONS**

Volume measurements were carried out on all bags.

| <u>SOLUTION TYPE</u> | # OF BAGS <u>WEIGHED</u> | <u>Mean</u> ml | <u>SD</u> | # OF FAILURES | | <u>% FAILURES</u> |
|----------------------|-----------------------------|-------------------|-----------|------------------|-----------------|-------------------|
| | | | | <u>underfill</u> | <u>overflow</u> | |
| NaCl - (1 Liter) | 773 | 1051.3 | 5.3 | 7 | 1 | 1.0% |
| LR (0.5 Liter) | 810 | 534.7 | 2.3 | 2 | 1 | 0.4% |
| RCW (3 Liter) | 174 | 3083.3 | 33.2 | 3 | 1 | 2.3% |
| LR (1 Liter) | 864 | 1050.5 | 12.4 | 8 | 0 | 0.9% |
| Total | 2621 | | | | | 0.9% |

INDIVIDUAL DATA IS REPORTED IN APPENDIX 2.

(Note: Appendix 1 reports the volume measurements on only those manifolds that have been tested for Endotoxin, Sterility, Particulate Matter or Chemistries.)

TABLE 2

INCIDENCE OF ARMY MANIFOLDS EXPOSED TO TEMPERATURES HIGHER THAN THOSE SPECIFIED IN THE SHIPPING PROTOCOL.

| <u>SOLUTION TYPE</u> | <u># OF MANIFOLDS RECEIVED</u> | <u># OF MANIFOLDS EXPOSED TO HIGH TEMPERATURES</u> |
|----------------------|--------------------------------|----------------------------------------------------|
| NaCl - (1 Liter) | 29 | 2 |
| LR1 - (1 Liter) | 43 | 3 |
| LR (0.5 Liter) | 48 | 1 |
| RCW (3 Liter) | 41 | 0 |
| | 161 | 6 |

Total incidence = 3.7%

INCIDENCE & TYPE OF DEFECTIVE BAGS OF ARMY REFLUPS SOLUTIONS

INCIDENCE & TYPE OF DEFECTIVE BAGS OF ARMY REFLUPS SOLUTIONS

NOTE: Each REFLUPS bag has 2 ports , a delivery port which allows the bag to be entered and fluid to be delivered and a filling port which is used to fill the bag during production.

Three types of defects were observed with the ARMY REFLUPS bags.

Type of Defect

A = Delivery port loose and falls out upon sampling The nature of this defect allows it to be detected only when bags are sampled for testing.

B = Puncture in bag (pin-hole or split in bag)

C = Filling port loose & leaking fluid

Defects were discovered at Log In , where the bag is weighed & given an NBRL Label, and upon sampling for testing.

Table 3 shows the incidence of defective bags & type of defects that were observed at NBRL.

Note: As testing is still ongoing at the Naval Blood Research Laboratory the incidence of observed defective bags may increase.

TABLE 3

| Solution Type | Run # | Defect Type | Stage Discovered |
|---------------|--------|-------------|------------------|
| NaCl- 1 liter | 284002 | C | Log In |
| NaCl- 1 liter | 285048 | C | Log In |
| NaCl- 1 liter | 285049 | A | Log In |
| NaCl- 1 liter | 280017 | C | Log In |
| NaCl- 1 liter | 279050 | C | Log In |
| NaCl- 1 liter | 279047 | C | Log In |
| NaCl- 1 liter | 279046 | C | Log In |
| NaCl- 1 liter | 279046 | C | Log In |
| NaCl- 1 liter | 273011 | C | Log In |
| NaCl- 1 liter | 274007 | C | Log In |
| LR 1Liter | 285008 | C | Log In |
| LR 1Liter | 280030 | C | Log In |
| LR 1Liter | 280030 | C | Sampling |
| LR 1Liter | 283021 | C | Log In |
| LR 1Liter | 275046 | C | Log In |
| RCW 3 Liter | 279029 | C | Log In |
| LR 0.5 Liter | 282027 | B | Sampling |

17 bags out of 2621 had defects (i.e 0.65%)

TABLE 4

PARTICULATE TESTING OF ARMY REFLUPS SOLUTIONS

| <u>SOLUTION TYPE</u> | <u># OF BAGS TESTED</u> | <u># OF FAILURES</u> | <u>% FAILURES</u> |
|----------------------|-------------------------|----------------------|-------------------|
| NaCl - (1 Liter) | 12 | 0 | 0% |
| LR (0.5 Liter) | 11 | 0 | 0% |
| LR (1 Liter) | 24 | 0 | 0% |
| RCW (3 Liter) | 12 | 0 | 0% |

All solutions tested were shown to be within the USP Standard for particulate matter. As a control, commercial solutions were tested with the REFLUPS solutions & at all times they were comparable.

INDIVIDUAL DATA IS REPORTED IN APPENDIX 1

TABLE 5

SODIUM CONCENTRATION OF ARMY REFLUPS SOLUTIONS

| <u>SOLUTION TYPE</u> | <u># OF BAGS TESTED</u> | <u>Mean mEq/l</u> | <u>SD</u> | <u># OF FAILURES</u> | <u>%FAILURES</u> |
|----------------------|-------------------------|-------------------|-----------|----------------------|------------------|
| NaCl - (1 Liter) | 12 | 153.8 | 1.1 | 0 | 0% |
| LR - (0.5 Liter) | 12 | 130.7 | 1.8 | 0 | 0% |
| LR - (1 Liter) | 24 | 130.1 | 1.5 | 0 | 0% |
| RCW (3 Liter) | 12 | 152.7 | 1.7 | 0 | 0% |

All solutions were shown to have acceptable sodium & again they were comparable to commercial solutions of a similar nature.

INDIVIDUAL DATA IS REPORTED IN APPENDIX 1

TABLE 6

POTASSIUM CONCENTRATION OF ARMY REFLUPS SOLUTIONS

| <u>SOLUTION TYPE</u> | <u># OF BAGS TESTED</u> | <u>Mean mEq/l</u> | <u>SD</u> | <u># OF FAILURES</u> | <u>% FAILURES</u> |
|----------------------|-------------------------|-------------------|-----------|----------------------|-------------------|
| LR (0.5 Liter) | 12 | 4.04 | 0.04 | 0 | 0% |
| LR (1 Liter) | 24 | 4.05 | 0.10 | 0 | 0% |

All solutions were shown to have acceptable potassium & again they were comparable to commercial solutions of a similar nature.

INDIVIDUAL DATA IS REPORTED IN APPENDIX 1

TABLE 7

pH VALUES OF ARMY REFLUPS SOLUTIONS

| <u>SOLUTION TYPE</u> | <u># OF BAGS TESTED</u> | <u>Mean</u> | <u>SD</u> | <u># OF FAILURES</u> | <u>% FAILURES</u> |
|----------------------|-------------------------|-------------|-----------|----------------------|-------------------|
| NaCl - (1 Liter) | 12 | 5.8 | 0.1 | 0 | 0% |
| LR (0.5 Liter) | 10 | 7.0 | 0.2 | 0 | 0% |
| LR (1 Liter) | 24 | 6.8 | 0.1 | 0 | 0% |
| RCW (3 Liter) | 12 | 5.1 | 0.2 | 0 | 0% |

All solutions were shown to have an acceptable pH level & comparable to commercial solutions.

INDIVIDUAL DATA IS REPORTED IN APPENDIX 1

TABLE 8

LACTATE CONCENTRATION OF ARMY REFLUPS SOLUTIONS

| <u>SOLUTION TYPE</u> | <u># OF BAGS TESTED</u> | <u>Mean</u> mg/dl | <u>SD</u> | <u># OF FAILURES</u> | <u>% FAILURES</u> |
|----------------------|-----------------------------|----------------------|-----------|----------------------|-------------------|
| LR (0.5 Liter) | 12 | 252.5 | 3.0 | 0 | 0% |
| LR (1 Liter) | 21 | 253.2 | 4.5 | 0 | 0% |

This measurement was only made on Lactated Ringers solutions

All solutions had acceptable levels of Lactate & were comparable to commercial solutions of a similar nature.

INDIVIDUAL DATA IS REPORTED IN APPENDIX 1

TABLE 9

CHLORIDE CONCENTRATION OF ARMY REFLUPS SOLUTIONS

| <u>SOLUTION TYPE</u> | <u># OF BAGS TESTED</u> | <u>Mean</u> mEq/l | <u>SD</u> | <u># OF FAILURES</u> | <u>% FAILURES</u> |
|----------------------|-----------------------------|----------------------|-----------|----------------------|-------------------|
| NaCl - (1 Liter) | 12 | 152 | 1.5 | 0 | 0% |
| LR (0.5 Liter) | 12 | 108 | 1.1 | 0 | 0% |
| LR (1 Liter) | 24 | 108 | 1.8 | 0 | 0% |
| RCW (3 Liter) | 12 | 149 | 1.5 | 0 | 0% |

All solutions tested were shown to have an acceptable chloride level & were comparable to commercial solutions.

INDIVIDUAL DATA IS REPORTED IN APPENDIX 1

TABLE 10**OSMOLALITY OF ARMY REFLUPS SOLUTIONS**

| <u>SOLUTION TYPE</u> | <u># OF BAGS TESTED</u> | <u>Mean</u> mOsmol/l | <u>SD</u> | <u># OF FAILURES</u> | <u>% FAILURES</u> |
|----------------------|-------------------------|-------------------------|-----------|----------------------|-------------------|
| NaCl - (1 Liter) | 12 | 284 | 2.5 | 0 | 0% |
| LR (0.5 Liter) | 12 | 253 | 2.5 | 0 | 0% |
| LR (1 Liter) | 24 | 254 | 3.5 | 0 | 0% |
| RCW (3 Liter) | 12 | 292 | 1.9 | 0 | 0% |

All solutions tested were shown to have an acceptable osmolality value, comparable to commercial solutions.

INDIVIDUAL DATA IS REPORTED IN APPENDIX 1

TABLE 11**GLUCOSE CONCENTRATION OF ARMY REFLUPS SOLUTIONS**

| <u>SOLUTION TYPE</u> | <u># OF BAGS TESTED</u> | <u>Mean</u> mg/dl | <u>SD</u> | <u># OF FAILURES</u> | <u>% FAILURES</u> |
|----------------------|-------------------------|----------------------|-----------|----------------------|-------------------|
| RCW (3 Liter) | 12 | 173 | 2 | 0 | 0% |

This measurement was only made on 0.9% NaCl 0.2% Dextrose solutions. The data shows that the glucose values are not within 5% of labelled value but the values are comparable to those for commercial solutions. The established acceptable range for Glucose is 170mg/dl to 180mg/d.

INDIVIDUAL DATA IS REPORTED IN APPENDIX 1

TABLE 12

CALCIUM CONCENTRATION OF ARMY REFLUPS SOLUTIONS

| <u>SOLUTION TYPE</u> | <u># OF BAGS TESTED</u> | <u>Mean</u> mg/dl | <u>SD</u> | <u># OF FAILURES</u> | <u>% FAILURES</u> |
|----------------------|-------------------------|----------------------|-----------|----------------------|-------------------|
| LR (0.5 Liter) | 12 | 5.3 | 0.1 | 0 | 0% |
| LR (1 Liter) | 24 | 5.3 | 0.1 | 0 | 0% |

This measurement was only made on Lactated Ringers solutions (1 liter & 1/2 liter). All solutions had acceptable levels of Calcium & were comparable to commercial solutions of a similar nature.

TABLE 13

ENDOTOXIN EVALUATION OF ARMY REFLUPS SOLUTIONS

| <u>SOLUTION TYPE</u> | <u># OF BAGS TESTED</u> | <u># OF FAILURES</u> | <u>% FAILURES</u> |
|----------------------|-------------------------|----------------------|-------------------|
| NaCl - (1 Liter) | 11 | 0 | 0% |
| LR (0.5 Liter) | 0 | 0 | 0% |
| LR (1 Liter) | 16 | 0 | 0% |
| RCW (3 Liter) | 12 | 0 | 0% |

All solutions tested were shown to be pyrogen free.

The data presented reflects the final endotoxin results, details of potency, spikes & recoveries and validation data have been archived.

INDIVIDUAL DATA IS REPORTED IN APPENDIX 1

TABLE 14

STERILITY TESTING OF ARMY REFLUPS SOLUTIONS

| <u>SOLUTION TYPE</u> | <u># OF BAGS TESTED</u> | <u># OF FAILURES</u> | <u>% FAILURES</u> |
|----------------------|-------------------------|----------------------|-------------------|
| NaCl - (1 Liter) | 0 | 0 | 0% |
| LR (0.5 Liter) | 12 | 0 | 0% |
| LR (1 Liter) | 16 | 0 | 0% |
| RCW (3 Liter) | 6 | 0 | 0% |

NOTE: All Sterility testing is carried out in a certified clean room to reduce the possibility of exogenously contaminating the solutions as they are being tested.

The clean room is located at;
Haemonetics Corporation,
Braintree, MA.

All solutions tested were shown to be sterile.
INDIVIDUAL DATA IS REPORTED IN APPENDIX 1

| LACTATED RINGERS 500ml | | | | | | | | | | | | | | |
|------------------------|---------------|----------------|-----------|-----------|------------------|-------------|------|------------------|------------------|-------------------|-----------------|--------------------|-----------------------|-------------------------|
| Solution | Run # | Sterility Test | | Endotoxin | Glucose mg/dl | Particulate | pH | Calcium mg/dl | Lactate mg/dl | Chloride mEq/L | Sodium mEq/L | Potassium mEq/L | Osmolality mOsmo/L | Average Volume ml |
| | | Aerobic | Anaerobic | | | | | | | | | | | |
| N/A = NOT APPLICABLE | | | | | | | | | | | | | | |
| N/T = NOT TESTED | | | | | | | | | | | | | | |
| LR5 | 279021 | Pass | Pass | N/T | N/A | Pass | 7.11 | 5.30 | 250 | 107 | 130.10 | 4.05 | 252 | 529.9 |
| LR5 | 274035 | Pass | Pass | N/T | N/A | Pass | 7.03 | 5.50 | 259 | 110 | 132.80 | 4.11 | 256 | 535.4 |
| LR5 | 274035 | Pass | Pass | N/T | N/A | Pass | 6.58 | 5.20 | 251 | 109 | 132.40 | 4.09 | 255 | 535.4 |
| LR5 | 275031 | Pass | Pass | N/T | N/A | Pass | 7.23 | 5.40 | 252 | 106 | 132.20 | 4.06 | 250 | 534.8 |
| LR5 | 276039 | Pass | Pass | N/T | N/A | Pass | 7.08 | 5.30 | 252 | 108 | 130.40 | 4.01 | 250 | 535.3 |
| LR5 | 279019 | Pass | Pass | N/T | N/A | Pass | 7.11 | 5.40 | 254 | 107 | 131.60 | 4.04 | 253 | 534.9 |
| LR5 | 277040 | Pass | Pass | N/T | N/A | Pass | 6.96 | 5.30 | 250 | 108 | 129.40 | 3.97 | 250 | 534.5 |
| LR5 | 269044 | Pass | Pass | N/T | N/A | Pass | 6.93 | 5.40 | 253 | 109 | 131.20 | 4.03 | 252 | 535.6 |
| LR5 | 276033 | Pass | Pass | N/T | N/A | Pass | N/T | 5.30 | 254 | 108 | 131.60 | 4.04 | 252 | 535.0 |
| LR5 | 272045 | Pass | Pass | N/T | N/A | Pass | N/T | 5.30 | 255 | 107 | 131.80 | 4.03 | 251 | 534.2 |
| LR5 | 275025 | Pass | Pass | N/T | N/A | Pass | 7.03 | 5.30 | 253 | 108 | 128.20 | 4.02 | 253 | 534.1 |
| LR5 | 275032 | Pass | Pass | N/T | N/A | N/T | 7.21 | 5.30 | 247 | 108 | 127.00 | 3.97 | 258 | 535.6 |
| | Count | 12 | 12 | 0 | | 11 | 10 | 12 | 12 | 12 | 12 | 12 | 12 | 12 |
| | Average | | | | | | 7.03 | 5.33 | 252.5 | 107.9 | 130.7 | 4.04 | 253 | 535 |
| | Std Deviation | | | | | | 0.2 | 0.1 | 3.0 | 1.1 | 1.8 | 0.04 | 2.5 | 1.6 |

| LACTATED RINGERS 1000ml | | | | | | | | | | | | | | |
|-------------------------|--------|----------------|-----------|----------------------|---------------|-------------|------|---------------|------------------|----------------|--------------|-----------------|-------------------|--------------------|
| Solution | Run # | Sterility Test | | N/A = NOT APPLICABLE | | | | | N/T = NOT TESTED | | | | Average Volume ml | |
| | | Acrobic | Anaerobic | Endotoxin | Glucose mg/dl | Particulate | PH | Calcium mg/dl | Lactate mg/dl | Chloride mEq/L | Sodium mEq/L | Potassium mEq/L | | Osmolality mOsmo/L |
| LRI | 270038 | Pass | Pass | Pass | N/A | Pass | 6.89 | 5.4 | 255 | 108 | 128.9 | 4.02 | 256 | 1025.3 |
| LRI | 276013 | Pass | Pass | Pass | N/A | Pass | 7.00 | 5.5 | 259 | 110 | 131.6 | 4.09 | 263 | 1033.6 |
| LRI | 269028 | Pass | Pass | Pass | N/A | Pass | 6.87 | 5.2 | 251 | 107 | 127.3 | 3.96 | 260 | 1047.0 |
| LRI | 273035 | N/T | N/T | Pass | N/A | Pass | 6.63 | 5.3 | 247 | 106 | 126.4 | 3.94 | 249 | 1056.3 |
| LRI | 274045 | Pass | Pass | Pass | N/A | Pass | 6.87 | 5.3 | 260 | 106 | 126.4 | 3.94 | 250 | 1052.6 |
| LRI | 271039 | N/T | N/T | Pass | N/A | Pass | 6.86 | 5.4 | 254 | 108 | 128.2 | 3.99 | 254 | 1054.3 |
| LRI | 278037 | Pass | Pass | Pass | N/A | Pass | 6.93 | 5.4 | 263 | 108 | 129.8 | 4.03 | 256 | 1053.3 |
| LRI | 276008 | Pass | Pass | Pass | N/A | Pass | 6.94 | 5.4 | N/T | 108 | 129.8 | 4.03 | 255 | 1052.6 |
| LRI | 275042 | Pass | Pass | Pass | N/A | Pass | 6.90 | 5.4 | N/T | 108 | 128.9 | 4.00 | 254 | 1051.7 |
| LRI | 271033 | Pass | Pass | Pass | N/A | Pass | 6.31 | 5.5 | N/T | 109 | 130.1 | 4.00 | 257 | 1056.1 |
| LRI | 275050 | Pass | Pass | Pass | N/A | Pass | 6.69 | 5.4 | 255 | 105 | 127.2 | 3.96 | 249 | 1052.2 |
| LRI | 272028 | Pass | Pass | Pass | N/A | Pass | 6.78 | 5.4 | 260 | 110 | 130.5 | 4.10 | 253 | 1054.1 |
| LRI | 272031 | Pass | Pass | Pass | N/A | Pass | 6.66 | 5.4 | 252 | N/T | 131.5 | 4.10 | 255 | 1054.6 |
| LRI | 279039 | Pass | Pass | N/T | N/A | Pass | 6.90 | 5.3 | 250 | 106 | 129.1 | 4.01 | 251 | 1055.2 |
| LRI | 274045 | Pass | Pass | N/T | N/A | Pass | 6.70 | 5.4 | 249 | 108 | 129.7 | 4.04 | 252 | 1052.6 |
| LRI | 279042 | N/T | N/T | Pass | N/A | Pass | 6.70 | 5.3 | 248 | 107 | 129.0 | 4.01 | 254 | 1054.4 |
| LRI | 274041 | Pass | Pass | Pass | N/A | Pass | 6.86 | 5.4 | 249 | 108 | 129.6 | 4.02 | 254 | 1047.0 |
| LRI | 277033 | Pass | Pass | N/T | N/A | Pass | 6.51 | 5.4 | 252 | 108 | 129.1 | 4.02 | 252 | 1046.3 |
| LRI | 275041 | N/T | N/T | N/T | N/A | Pass | 6.94 | 5.2 | 256 | 110 | 132.4 | 4.10 | 256 | 1055.5 |
| LRI | 275046 | N/T | N/T | N/T | N/A | Pass | 6.93 | 5.2 | 253 | 109 | 131.4 | 4.07 | 252 | 1043.6 |
| LRI | 279040 | N/T | N/T | N/T | N/A | Pass | 6.73 | 5.1 | 248 | 107 | 128.8 | 4.00 | 249 | 1042.5 |
| LRI | 277037 | N/T | N/T | N/T | N/A | Pass | 6.84 | 5.1 | 255 | 114 | 132.2 | 4.14 | 255 | 1055.8 |
| LRI | 272028 | N/T | N/T | N/T | N/A | Pass | 6.88 | 5.1 | 249 | 108 | 128.5 | 4.00 | 250 | 1054.1 |
| LRI | 273029 | Pass | Pass | Pass | N/A | Pass | 6.74 | 5.3 | 253 | 107 | 128.4 | 4.00 | 257 | 1057.0 |
| Count | 16 | 16 | 16 | 16 | | 24 | 24 | 24 | 21 | 23 | 24 | 24 | 24 | 24 |
| Average | | | | | | | 6.79 | 5.33 | 253.24 | 108.04 | 129.37 | 4.02 | 253.9 | 1051.2 |
| Std Deviation | | | | | | | 0.2 | 0.1 | 4.5 | 1.8 | 1.7 | 0.1 | 3.5 | 6.8 |

| 0.9% SODIUM CHLORIDE 1000ml | | | | | | | | | | | | | | |
|-----------------------------|----------------------|----------------|-----------|-----------|----------------------|-------------|-------------|------------------|------------------|-------------------|-----------------|--------------------|-----------------------|-------------------------|
| Solution | Run # | Sterility Test | | Endotoxin | Glucose mg/dl | Particulate | PH | Calcium mg/dl | Lactate mg/dl | Chloride mEq/L | Sodium mEq/L | Potassium mEq/L | Osmolality mOsmo/L | Average Volume ml |
| | | Aerobic | Anaerobic | | | | | | | | | | | |
| | | | | | N/A = NOT APPLICABLE | | | | | | | | | |
| | | | | | N/T = NOT TESTED | | | | | | | | | |
| NaCl | 274006 | N/T | N/T | Pass | N/A | Pass | 6.12 | N/A | N/A | 152 | 153.40 | N/A | 284 | 1051.6 |
| NaCl | 278007 | N/T | N/T | Pass | N/A | Pass | 5.86 | N/A | N/A | 151 | 153.60 | N/A | 283 | 1052.1 |
| NaCl | 270033 | N/T | N/T | Pass | N/A | Pass | 5.90 | N/A | N/A | 154 | 156.20 | N/A | 288 | 1069.1 |
| NaCl | 275002 | N/T | N/T | N/T | N/A | Pass | 5.57 | N/A | N/A | 153 | 154.80 | N/A | 287 | 1051.9 |
| NaCl | 276018 | N/T | N/T | Pass | N/A | Pass | 5.80 | N/A | N/A | 154 | 154.90 | N/A | 282 | 1056.7 |
| NaCl | 277023 | N/T | N/T | Pass | N/A | Pass | 5.83 | N/A | N/A | 150 | 153.20 | N/A | 282 | 1053.9 |
| NaCl | 271005 | N/T | N/T | Pass | N/A | Pass | 5.80 | N/A | N/A | 151 | 154.00 | N/A | 282 | 1053.7 |
| NaCl | 273007 | N/T | N/T | Pass | N/A | Pass | 5.57 | N/A | N/A | 149 | 151.70 | N/A | 279 | 1054.2 |
| NaCl | 269020 | N/T | N/T | Pass | N/A | Pass | 5.84 | N/A | N/A | 151 | 153.90 | N/A | 284 | 1055.9 |
| NaCl | 275007 | N/T | N/T | Pass | N/A | Pass | 5.79 | N/A | N/A | 151 | 154.00 | N/A | 284 | 1056.9 |
| NaCl | 270035 | N/T | N/T | Pass | N/A | Pass | 5.83 | N/A | N/A | 153 | 153.40 | N/A | 286 | 1044.0 |
| NaCl | 270022 | N/T | N/T | Pass | N/A | Pass | 5.79 | N/A | N/A | 152 | 152.90 | N/A | 284 | 1051.3 |
| | Count | 0 | 0 | 11 | | 12 | 12 | | | 12 | 12 | | 12 | 12 |
| | Average | | | | | | 5.81 | | | 151.8 | 153.8 | | 284 | 1054 |
| | Std Deviation | | | | | | 0.1 | | | 1.5 | 1.1 | | 2.5 | 5.8 |

| RED CELL WASH 3000ml | | Sterility Test | | N/A = NOT APPLICABLE | | N/T = NOT TESTED | | | | | | | | | | | | | |
|----------------------|---------------|----------------|-----------|----------------------|------------------|------------------|------|------------------|------------------|-------------------|-----------------|--------------------|-----------------------|-------------------------|--|--|--|--|--|
| Solution | Run # | Aerobic | Anaerobic | Endotoxin | Glucose mg/dl | Particulate | PH | Calcium mg/dl | Lactate mg/dl | Chloride mEq/L | Sodium mEq/L | Potassium mEq/L | Osmolality mOsmo/L | Average Volume ml | | | | | |
| RBC | 277014 | Pass | Pass | Pass | 175 | Pass | 5.36 | N/A | N/A | 149 | 153.2 | N/A | 290 | 3079.6 | | | | | |
| RBC | 272002 | Pass | Pass | Pass | 174 | Pass | 5.06 | N/A | N/A | 150 | 153.0 | N/A | 291 | 3184.6 | | | | | |
| RBC | 273001 | Pass | Pass | Pass | 175 | Pass | 5.11 | N/A | N/A | 150 | 153.5 | N/A | 294 | 3087.3 | | | | | |
| RBC | 271010 | Pass | Pass | Pass | 173 | Pass | 5.02 | N/A | N/A | 149 | 154.8 | N/A | 290 | 3089.6 | | | | | |
| RBC | 276026 | Pass | Pass | Pass | 173 | Pass | 5.14 | N/A | N/A | 150 | 154.6 | N/A | 293 | 3096.7 | | | | | |
| RBC | 272002 | Pass | Pass | Pass | 174 | Pass | 5.49 | N/A | N/A | 151 | 151.0 | N/A | 294 | 3184.6 | | | | | |
| RBC | 279028 | N/T | N/T | Pass | 173 | Pass | 5.27 | N/A | N/A | 150 | 154.4 | N/A | 291 | 3082.2 | | | | | |
| RBC | 275024 | N/T | N/T | Pass | 172 | Pass | 5.10 | N/A | N/A | 145 | 151.5 | N/A | 294 | 3085.5 | | | | | |
| RBC | 279029 | N/T | N/T | Pass | 171 | Pass | 5.06 | N/A | N/A | 149 | 151.8 | N/A | 290 | 3092.0 | | | | | |
| RBC | 276024 | N/T | N/T | Pass | 172 | Pass | 5.09 | N/A | N/A | 150 | 150.5 | N/A | 295 | 3075.6 | | | | | |
| RBC | 273003 | N/T | N/T | Pass | 170 | Pass | 5.02 | N/A | N/A | 149 | 150.3 | N/A | 291 | 3059.3 | | | | | |
| RBC | 271009 | N/T | N/T | Pass | 171 | Pass | 4.88 | N/A | N/A | 149 | 152.8 | N/A | 294 | 3093.6 | | | | | |
| | Count | 6 | 6 | 12 | 12 | 12 | 12 | | | 12 | 12 | | 12 | 12 | | | | | |
| | Average | | | | 173 | | 5.13 | | | 149.3 | 152.6 | | 292 | 3101 | | | | | |
| | Std Deviation | | | | 2 | | 0.2 | | | 1.5 | 1.6 | | 1.9 | 40.3 | | | | | |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-269041-01 | 582.4 | 533.8 | Pass |
| LR500 | 0000-269041-02 | 584.5 | 535.9 | Pass |
| LR500 | 0000-269041-03 | 584.0 | 535.4 | Pass |
| LR500 | 0000-269041-04 | 583.1 | 534.5 | Pass |
| LR500 | 0000-269041-05 | 586.0 | 537.4 | Pass |
| LR500 | 0000-269041-06 | 584.2 | 535.6 | Pass |
| LR500 | 0000-269041-07 | 581.9 | 533.3 | Pass |
| LR500 | 0000-269041-08 | 585.2 | 536.6 | Pass |
| LR500 | 0000-269041-09 | 582.4 | 533.8 | Pass |
| LR500 | 0000-269041-10 | 583.3 | 534.7 | Pass |
| LR500 | 0000-269041-11 | 585.6 | 537.0 | Pass |
| LR500 | 0000-269041-12 | 582.1 | 533.5 | Pass |
| LR500 | 0000-269041-13 | 583.5 | 534.9 | Pass |
| LR500 | 0000-269041-14 | 585.0 | 536.4 | Pass |
| LR500 | 0000-269041-15 | 580.8 | 532.2 | Pass |
| LR500 | 0000-269041-16 | 577.3 | 528.8 | Pass |
| LR500 | 0000-269041-17 | 582.6 | 534.0 | Pass |
| LR500 | 0000-269041-18 | 583.6 | 535.0 | Pass |
| LR500 | 0000-269044-01 | 584.6 | 536.0 | Pass |
| LR500 | 0000-269044-02 | 585.2 | 536.6 | Pass |
| LR500 | 0000-269044-03 | 584.9 | 536.3 | Pass |
| LR500 | 0000-269044-04 | 584.8 | 536.2 | Pass |
| LR500 | 0000-269044-05 | 584.1 | 535.5 | Pass |
| LR500 | 0000-269044-06 | 584.3 | 535.7 | Pass |
| LR500 | 0000-269044-07 | 584.8 | 536.2 | Pass |
| LR500 | 0000-269044-08 | 584.6 | 536.0 | Pass |
| LR500 | 0000-269044-09 | 585.5 | 536.9 | Pass |
| LR500 | 0000-269044-10 | 583.7 | 535.1 | Pass |
| LR500 | 0000-269044-11 | 582.9 | 534.3 | Pass |
| LR500 | 0000-269044-12 | 580.0 | 531.4 | Pass |
| LR500 | 0000-269044-13 | 583.0 | 534.4 | Pass |
| LR500 | 0000-269044-14 | 584.0 | 535.4 | Pass |
| LR500 | 0000-269044-15 | 584.2 | 535.6 | Pass |
| LR500 | 0000-269044-16 | 584.5 | 535.9 | Pass |
| LR500 | 0000-269044-17 | 584.4 | 535.8 | Pass |
| LR500 | 0000-269044-18 | 585.1 | 536.5 | Pass |
| LR500 | 0000-270005-01 | 582.6 | 534.0 | Pass |
| LR500 | 0000-270005-02 | 583.9 | 535.3 | Pass |
| LR500 | 0000-270005-03 | 580.9 | 532.3 | Pass |
| LR500 | 0000-270005-04 | 584.5 | 535.9 | Pass |
| LR500 | 0000-270005-05 | 584.0 | 535.4 | Pass |
| LR500 | 0000-270005-06 | 583.4 | 534.8 | Pass |
| LR500 | 0000-270005-07 | 583.4 | 534.8 | Pass |
| LR500 | 0000-270005-08 | 583.4 | 534.8 | Pass |
| LR500 | 0000-270005-09 | 584.8 | 536.2 | Pass |
| LR500 | 0000-270005-10 | 584.2 | 535.6 | Pass |

Reflups Volume Army_2 Bags

| | | | | |
|----------------------------|----------------|-----------------------|---------------|--------|
| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-270005-11 | 583.8 | 535.2 | Pass |
| LR500 | 0000-270005-12 | 581.9 | 533.3 | Pass |
| LR500 | 0000-270005-13 | 585.3 | 536.7 | Pass |
| LR500 | 0000-270005-14 | 584.8 | 536.2 | Pass |
| LR500 | 0000-270005-15 | 583.8 | 535.2 | Pass |
| LR500 | 0000-270005-16 | 585.4 | 536.8 | Pass |
| LR500 | 0000-270005-17 | 583.8 | 535.2 | Pass |
| LR500 | 0000-270005-18 | 581.9 | 533.3 | Pass |
| LR500 | 0000-270006-01 | 585.8 | 537.2 | Pass |
| LR500 | 0000-270006-02 | 583.3 | 534.7 | Pass |
| LR500 | 0000-270006-03 | 583.5 | 534.9 | Pass |
| LR500 | 0000-270006-04 | 584.1 | 535.5 | Pass |
| LR500 | 0000-270006-05 | 583.4 | 534.8 | Pass |
| LR500 | 0000-270006-06 | 582.9 | 534.3 | Pass |
| LR500 | 0000-270006-07 | 585.3 | 536.7 | Pass |
| LR500 | 0000-270006-08 | 583.9 | 535.3 | Pass |
| LR500 | 0000-270006-09 | 586.5 | 537.9 | Pass |
| LR500 | 0000-270006-10 | 576.5 | 528.0 | Pass |
| LR500 | 0000-270006-11 | 582.3 | 533.7 | Pass |
| LR500 | 0000-270006-12 | 585.6 | 537.0 | Pass |
| LR500 | 0000-270006-13 | 584.6 | 536.0 | Pass |
| LR500 | 0000-270006-14 | 580.9 | 532.3 | Pass |
| LR500 | 0000-270006-15 | 581.8 | 533.2 | Pass |
| LR500 | 0000-270006-16 | 586.7 | 538.1 | Pass |
| LR500 | 0000-270006-17 | 581.8 | 533.2 | Pass |
| LR500 | 0000-270006-18 | 584.4 | 535.8 | Pass |
| LR500 | 0000-271015-01 | 583.6 | 535.0 | Pass |
| LR500 | 0000-271015-02 | 583.3 | 534.7 | Pass |
| LR500 | 0000-271015-03 | 582.6 | 534.0 | Pass |
| LR500 | 0000-271015-04 | 585.4 | 536.8 | Pass |
| LR500 | 0000-271015-05 | 582.9 | 534.3 | Pass |
| LR500 | 0000-271015-06 | 583.8 | 535.2 | Pass |
| LR500 | 0000-271015-07 | 579.0 | 530.4 | Pass |
| LR500 | 0000-271015-08 | 584.9 | 536.3 | Pass |
| LR500 | 0000-271015-09 | 582.0 | 533.4 | Pass |
| LR500 | 0000-271015-10 | 584.7 | 536.1 | Pass |
| LR500 | 0000-271015-11 | 584.5 | 535.9 | Pass |
| LR500 | 0000-271015-12 | 584.4 | 535.8 | Pass |
| LR500 | 0000-271015-13 | 581.5 | 532.9 | Pass |
| LR500 | 0000-271015-14 | 584.5 | 535.9 | Pass |
| LR500 | 0000-271015-15 | 584.1 | 535.5 | Pass |
| LR500 | 0000-271015-16 | 582.4 | 533.8 | Pass |
| LR500 | 0000-271015-17 | 576.7 | 528.2 | Pass |
| LR500 | 0000-271015-18 | 584.0 | 535.4 | Pass |
| LR500 | 0000-272045-01 | 583.0 | 534.4 | Pass |
| LR500 | 0000-272045-02 | 584.2 | 535.6 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-272045-03 | 582.4 | 533.8 | Pass |
| LR500 | 0000-272045-04 | 582.9 | 534.3 | Pass |
| LR500 | 0000-272045-05 | 584.9 | 536.3 | Pass |
| LR500 | 0000-272045-06 | 582.4 | 533.8 | Pass |
| LR500 | 0000-272045-07 | 583.4 | 534.8 | Pass |
| LR500 | 0000-272045-08 | 582.4 | 533.8 | Pass |
| LR500 | 0000-272045-09 | 584.1 | 535.5 | Pass |
| LR500 | 0000-272045-10 | 583.0 | 534.4 | Pass |
| LR500 | 0000-272045-11 | 583.1 | 534.5 | Pass |
| LR500 | 0000-272045-12 | 576.5 | 528.0 | Pass |
| LR500 | 0000-272045-13 | 583.0 | 534.4 | Pass |
| LR500 | 0000-272045-14 | 583.7 | 535.1 | Pass |
| LR500 | 0000-272045-15 | 583.1 | 534.5 | Pass |
| LR500 | 0000-272045-16 | 584.1 | 535.5 | Pass |
| LR500 | 0000-272045-17 | 581.4 | 532.8 | Pass |
| LR500 | 0000-272045-18 | 582.4 | 533.8 | Pass |
| LR500 | 0000-273051-01 | 580.9 | 532.3 | Pass |
| LR500 | 0000-273051-02 | 581.0 | 532.4 | Pass |
| LR500 | 0000-273051-03 | 581.1 | 532.5 | Pass |
| LR500 | 0000-273051-04 | 581.2 | 532.6 | Pass |
| LR500 | 0000-273051-05 | 580.6 | 532.0 | Pass |
| LR500 | 0000-273051-06 | 582.9 | 534.3 | Pass |
| LR500 | 0000-273051-07 | 580.0 | 531.4 | Pass |
| LR500 | 0000-273051-08 | 582.2 | 533.6 | Pass |
| LR500 | 0000-273051-09 | 581.5 | 532.9 | Pass |
| LR500 | 0000-273051-10 | 575.2 | 526.7 | Pass |
| LR500 | 0000-273051-11 | 581.5 | 532.9 | Pass |
| LR500 | 0000-273051-12 | 581.0 | 532.4 | Pass |
| LR500 | 0000-273051-13 | 581.4 | 532.8 | Pass |
| LR500 | 0000-273051-14 | 580.1 | 531.5 | Pass |
| LR500 | 0000-273051-15 | 581.6 | 533.0 | Pass |
| LR500 | 0000-273051-16 | 581.2 | 532.6 | Pass |
| LR500 | 0000-273051-17 | 580.2 | 531.6 | Pass |
| LR500 | 0000-273051-18 | 581.3 | 532.7 | Pass |
| LR500 | 0000-274028-01 | 584.3 | 535.7 | Pass |
| LR500 | 0000-274028-02 | 586.4 | 537.8 | Pass |
| LR500 | 0000-274028-03 | 584.3 | 535.7 | Pass |
| LR500 | 0000-274028-04 | 585.1 | 536.5 | Pass |
| LR500 | 0000-274028-05 | 583.6 | 535.0 | Pass |
| LR500 | 0000-274028-06 | 584.0 | 535.4 | Pass |
| LR500 | 0000-274028-07 | 583.2 | 534.6 | Pass |
| LR500 | 0000-274028-08 | 583.2 | 534.6 | Pass |
| LR500 | 0000-274028-09 | 582.5 | 533.9 | Pass |
| LR500 | 0000-274028-10 | 583.8 | 535.2 | Pass |
| LR500 | 0000-274028-11 | 583.4 | 534.8 | Pass |
| LR500 | 0000-274028-12 | 584.3 | 535.7 | Pass |

Reflups Volume Army_2 Bags

| | | | | |
|----------------------------|----------------|--------------|---------------|--------|
| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
| Product | Bag Number | Gross Weight | Volume | Status |
| | | Grams | ml | |
| LR500 | 0000-274028-13 | 584.5 | 535.9 | Pass |
| LR500 | 0000-274028-14 | 582.9 | 534.3 | Pass |
| LR500 | 0000-274028-15 | 584.1 | 535.5 | Pass |
| LR500 | 0000-274028-16 | 584.6 | 536.0 | Pass |
| LR500 | 0000-274028-17 | 584.7 | 536.1 | Pass |
| LR500 | 0000-274028-18 | 579.8 | 531.2 | Pass |
| LR500 | 0000-274031-01 | 581.3 | 532.7 | Pass |
| LR500 | 0000-274031-02 | 580.1 | 531.5 | Pass |
| LR500 | 0000-274031-03 | 582.7 | 534.1 | Pass |
| LR500 | 0000-274031-04 | 582.2 | 533.6 | Pass |
| LR500 | 0000-274031-05 | 580.2 | 531.6 | Pass |
| LR500 | 0000-274031-06 | 582.0 | 533.4 | Pass |
| LR500 | 0000-274031-07 | 582.1 | 533.5 | Pass |
| LR500 | 0000-274031-08 | 581.2 | 532.6 | Pass |
| LR500 | 0000-274031-09 | 581.0 | 532.4 | Pass |
| LR500 | 0000-274031-10 | 576.1 | 527.6 | Pass |
| LR500 | 0000-274031-11 | 580.3 | 531.7 | Pass |
| LR500 | 0000-274031-12 | 816.9 | 767.6 | FAIL |
| LR500 | 0000-274031-13 | 583.2 | 534.6 | Pass |
| LR500 | 0000-274031-14 | 582.5 | 533.9 | Pass |
| LR500 | 0000-274031-15 | 580.6 | 532.0 | Pass |
| LR500 | 0000-274031-16 | 583.1 | 534.5 | Pass |
| LR500 | 0000-274031-17 | 582.9 | 534.3 | Pass |
| LR500 | 0000-274031-18 | 583.1 | 534.5 | Pass |
| LR500 | 0000-274035-01 | 584.2 | 535.6 | Pass |
| LR500 | 0000-274035-01 | 584.3 | 535.7 | Pass |
| LR500 | 0000-274035-02 | 584.4 | 535.8 | Pass |
| LR500 | 0000-274035-02 | 586.4 | 537.8 | Pass |
| LR500 | 0000-274035-03 | 584.1 | 535.5 | Pass |
| LR500 | 0000-274035-03 | 584.9 | 536.3 | Pass |
| LR500 | 0000-274035-04 | 584.7 | 536.1 | Pass |
| LR500 | 0000-274035-04 | 577.0 | 528.5 | Pass |
| LR500 | 0000-274035-05 | 579.2 | 530.6 | Pass |
| LR500 | 0000-274035-05 | 583.3 | 534.7 | Pass |
| LR500 | 0000-274035-06 | 583.4 | 534.8 | Pass |
| LR500 | 0000-274035-06 | 583.9 | 535.3 | Pass |
| LR500 | 0000-274035-07 | 583.3 | 534.7 | Pass |
| LR500 | 0000-274035-07 | 544.5 | 496.1 | FAIL |
| LR500 | 0000-274035-08 | 584.7 | 536.1 | Pass |
| LR500 | 0000-274035-08 | 582.4 | 533.8 | Pass |
| LR500 | 0000-274035-09 | 583.3 | 534.7 | Pass |
| LR500 | 0000-274035-09 | 585.7 | 537.1 | Pass |
| LR500 | 0000-274035-10 | 585.0 | 536.4 | Pass |
| LR500 | 0000-274035-10 | 583.7 | 535.1 | Pass |
| LR500 | 0000-274035-11 | 584.0 | 535.4 | Pass |
| LR500 | 0000-274035-11 | 585.2 | 536.6 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-274035-12 | 584.8 | 536.2 | Pass |
| LR500 | 0000-274035-12 | 587.7 | 539.1 | Pass |
| LR500 | 0000-274035-13 | 583.7 | 535.1 | Pass |
| LR500 | 0000-274035-13 | 584.6 | 536.0 | Pass |
| LR500 | 0000-274035-14 | 583.7 | 535.1 | Pass |
| LR500 | 0000-274035-14 | 582.3 | 533.7 | Pass |
| LR500 | 0000-274035-15 | 584.6 | 536.0 | Pass |
| LR500 | 0000-274035-15 | 582.5 | 533.9 | Pass |
| LR500 | 0000-274035-16 | 583.6 | 535.0 | Pass |
| LR500 | 0000-274035-16 | 581.9 | 533.3 | Pass |
| LR500 | 0000-274035-17 | 583.0 | 534.4 | Pass |
| LR500 | 0000-274035-17 | 585.1 | 536.5 | Pass |
| LR500 | 0000-274035-18 | 583.1 | 534.5 | Pass |
| LR500 | 0000-274035-18 | 583.3 | 534.7 | Pass |
| LR500 | 0000-275025-01 | 582.7 | 534.1 | Pass |
| LR500 | 0000-275025-02 | 583.8 | 535.2 | Pass |
| LR500 | 0000-275025-03 | 584.0 | 535.4 | Pass |
| LR500 | 0000-275025-04 | 583.7 | 535.1 | Pass |
| LR500 | 0000-275025-05 | 584.2 | 535.6 | Pass |
| LR500 | 0000-275025-06 | 577.4 | 528.9 | Pass |
| LR500 | 0000-275025-07 | 583.2 | 534.6 | Pass |
| LR500 | 0000-275025-08 | 583.1 | 534.5 | Pass |
| LR500 | 0000-275025-09 | 582.8 | 534.2 | Pass |
| LR500 | 0000-275025-10 | 583.6 | 535.0 | Pass |
| LR500 | 0000-275025-11 | 583.8 | 535.2 | Pass |
| LR500 | 0000-275025-12 | 583.3 | 534.7 | Pass |
| LR500 | 0000-275025-13 | 574.6 | 526.1 | Pass |
| LR500 | 0000-275025-14 | 582.7 | 534.1 | Pass |
| LR500 | 0000-275025-15 | 585.2 | 536.6 | Pass |
| LR500 | 0000-275025-16 | 583.3 | 534.7 | Pass |
| LR500 | 0000-275025-17 | 583.0 | 534.4 | Pass |
| LR500 | 0000-275025-18 | 583.0 | 534.4 | Pass |
| LR500 | 0000-275031-01 | 583.4 | 534.8 | Pass |
| LR500 | 0000-275031-02 | 582.8 | 534.2 | Pass |
| LR500 | 0000-275031-03 | 583.5 | 534.9 | Pass |
| LR500 | 0000-275031-04 | 583.6 | 535.0 | Pass |
| LR500 | 0000-275031-05 | 577.8 | 529.3 | Pass |
| LR500 | 0000-275031-06 | 583.6 | 535.0 | Pass |
| LR500 | 0000-275031-07 | 583.4 | 534.8 | Pass |
| LR500 | 0000-275031-08 | 585.2 | 536.6 | Pass |
| LR500 | 0000-275031-09 | 583.7 | 535.1 | Pass |
| LR500 | 0000-275031-10 | 584.4 | 535.8 | Pass |
| LR500 | 0000-275031-11 | 584.1 | 535.5 | Pass |
| LR500 | 0000-275031-12 | 583.3 | 534.7 | Pass |
| LR500 | 0000-275031-13 | 583.9 | 535.3 | Pass |
| LR500 | 0000-275031-14 | 582.7 | 534.1 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-275031-15 | 583.0 | 534.4 | Pass |
| LR500 | 0000-275031-16 | 584.6 | 536.0 | Pass |
| LR500 | 0000-275031-17 | 583.3 | 534.7 | Pass |
| LR500 | 0000-275031-18 | 584.5 | 535.9 | Pass |
| LR500 | 0000-275032-01 | 584.2 | 535.6 | Pass |
| LR500 | 0000-275032-02 | 580.6 | 532.0 | Pass |
| LR500 | 0000-275032-03 | 584.1 | 535.5 | Pass |
| LR500 | 0000-275032-04 | 584.5 | 535.9 | Pass |
| LR500 | 0000-275032-05 | 584.3 | 535.7 | Pass |
| LR500 | 0000-275032-06 | 584.3 | 535.7 | Pass |
| LR500 | 0000-275032-07 | 584.0 | 535.4 | Pass |
| LR500 | 0000-275032-08 | 584.4 | 535.8 | Pass |
| LR500 | 0000-275032-09 | 583.8 | 535.2 | Pass |
| LR500 | 0000-275032-10 | 584.8 | 536.2 | Pass |
| LR500 | 0000-275032-11 | 585.0 | 536.4 | Pass |
| LR500 | 0000-275032-12 | 584.6 | 536.0 | Pass |
| LR500 | 0000-275032-13 | 584.6 | 536.0 | Pass |
| LR500 | 0000-275032-14 | 583.9 | 535.3 | Pass |
| LR500 | 0000-275032-15 | 585.3 | 536.7 | Pass |
| LR500 | 0000-275032-16 | 584.1 | 535.5 | Pass |
| LR500 | 0000-275032-17 | 583.7 | 535.1 | Pass |
| LR500 | 0000-275032-18 | 584.3 | 535.7 | Pass |
| LR500 | 0000-276033-01 | 584.3 | 535.7 | Pass |
| LR500 | 0000-276033-02 | 584.9 | 536.3 | Pass |
| LR500 | 0000-276033-03 | 583.1 | 534.5 | Pass |
| LR500 | 0000-276033-04 | 583.1 | 534.5 | Pass |
| LR500 | 0000-276033-05 | 583.5 | 534.9 | Pass |
| LR500 | 0000-276033-06 | 584.0 | 535.4 | Pass |
| LR500 | 0000-276033-07 | 578.4 | 529.9 | Pass |
| LR500 | 0000-276033-08 | 584.5 | 535.9 | Pass |
| LR500 | 0000-276033-09 | 583.2 | 534.6 | Pass |
| LR500 | 0000-276033-10 | 584.2 | 535.6 | Pass |
| LR500 | 0000-276033-11 | 584.2 | 535.6 | Pass |
| LR500 | 0000-276033-12 | 583.3 | 534.7 | Pass |
| LR500 | 0000-276033-13 | 584.6 | 536.0 | Pass |
| LR500 | 0000-276033-14 | 584.1 | 535.5 | Pass |
| LR500 | 0000-276033-15 | 583.6 | 535.0 | Pass |
| LR500 | 0000-276033-16 | 582.7 | 534.1 | Pass |
| LR500 | 0000-276033-17 | 584.1 | 535.5 | Pass |
| LR500 | 0000-276033-18 | 583.9 | 535.3 | Pass |
| LR500 | 0000-276039-01 | 583.6 | 535.0 | Pass |
| LR500 | 0000-276039-02 | 584.3 | 535.7 | Pass |
| LR500 | 0000-276039-03 | 583.7 | 535.1 | Pass |
| LR500 | 0000-276039-04 | 584.3 | 535.7 | Pass |
| LR500 | 0000-276039-05 | 583.3 | 534.7 | Pass |
| LR500 | 0000-276039-06 | 584.2 | 535.6 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-276039-07 | 584.7 | 536.1 | Pass |
| LR500 | 0000-276039-08 | 583.8 | 535.2 | Pass |
| LR500 | 0000-276039-09 | 584.7 | 536.1 | Pass |
| LR500 | 0000-276039-10 | 583.0 | 534.4 | Pass |
| LR500 | 0000-276039-11 | 584.4 | 535.8 | Pass |
| LR500 | 0000-276039-12 | 584.4 | 535.8 | Pass |
| LR500 | 0000-276039-13 | 585.1 | 536.5 | Pass |
| LR500 | 0000-276039-14 | 584.5 | 535.9 | Pass |
| LR500 | 0000-276039-15 | 583.2 | 534.6 | Pass |
| LR500 | 0000-276039-16 | 583.4 | 534.8 | Pass |
| LR500 | 0000-276039-17 | 581.4 | 532.8 | Pass |
| LR500 | 0000-276039-18 | 583.6 | 535.0 | Pass |
| LR500 | 0000-277040-01 | 584.1 | 535.5 | Pass |
| LR500 | 0000-277040-02 | 584.7 | 536.1 | Pass |
| LR500 | 0000-277040-03 | 584.6 | 536.0 | Pass |
| LR500 | 0000-277040-04 | 581.5 | 532.9 | Pass |
| LR500 | 0000-277040-05 | 575.3 | 526.8 | Pass |
| LR500 | 0000-277040-06 | 583.5 | 534.9 | Pass |
| LR500 | 0000-277040-07 | 585.1 | 536.5 | Pass |
| LR500 | 0000-277040-08 | 581.7 | 533.1 | Pass |
| LR500 | 0000-277040-09 | 584.0 | 535.4 | Pass |
| LR500 | 0000-277040-10 | 583.8 | 535.2 | Pass |
| LR500 | 0000-277040-11 | 583.9 | 535.3 | Pass |
| LR500 | 0000-277040-12 | 585.8 | 537.2 | Pass |
| LR500 | 0000-277040-13 | 583.4 | 534.8 | Pass |
| LR500 | 0000-277040-14 | 577.7 | 529.2 | Pass |
| LR500 | 0000-277040-15 | 584.6 | 536.0 | Pass |
| LR500 | 0000-277040-16 | 581.6 | 533.0 | Pass |
| LR500 | 0000-277040-17 | 581.9 | 533.3 | Pass |
| LR500 | 0000-277040-18 | 584.5 | 535.9 | Pass |
| LR500 | 0000-278044-01 | 579.7 | 531.1 | Pass |
| LR500 | 0000-278044-02 | 580.8 | 532.2 | Pass |
| LR500 | 0000-278044-03 | 581.4 | 532.8 | Pass |
| LR500 | 0000-278044-04 | 574.9 | 526.4 | Pass |
| LR500 | 0000-278044-05 | 580.9 | 532.3 | Pass |
| LR500 | 0000-278044-06 | 580.1 | 531.5 | Pass |
| LR500 | 0000-278044-07 | 581.1 | 532.5 | Pass |
| LR500 | 0000-278044-08 | 580.7 | 532.1 | Pass |
| LR500 | 0000-278044-09 | 580.6 | 532.0 | Pass |
| LR500 | 0000-278044-10 | 580.0 | 531.4 | Pass |
| LR500 | 0000-278044-11 | 580.9 | 532.3 | Pass |
| LR500 | 0000-278044-12 | 580.5 | 531.9 | Pass |
| LR500 | 0000-278044-13 | 580.2 | 531.6 | Pass |
| LR500 | 0000-278044-14 | 580.2 | 531.6 | Pass |
| LR500 | 0000-278044-15 | 580.7 | 532.1 | Pass |
| LR500 | 0000-278044-16 | 581.2 | 532.6 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-278044-17 | 581.4 | 532.8 | Pass |
| LR500 | 0000-278044-18 | 580.8 | 532.2 | Pass |
| LR500 | 0000-278049-01 | 582.2 | 533.6 | Pass |
| LR500 | 0000-278049-02 | 584.6 | 536.0 | Pass |
| LR500 | 0000-278049-03 | 577.2 | 528.7 | Pass |
| LR500 | 0000-278049-04 | 583.7 | 535.1 | Pass |
| LR500 | 0000-278049-05 | 586.1 | 537.5 | Pass |
| LR500 | 0000-278049-06 | 582.9 | 534.3 | Pass |
| LR500 | 0000-278049-07 | 582.7 | 534.1 | Pass |
| LR500 | 0000-278049-08 | 583.0 | 534.4 | Pass |
| LR500 | 0000-278049-09 | 581.8 | 533.2 | Pass |
| LR500 | 0000-278049-10 | 583.1 | 534.5 | Pass |
| LR500 | 0000-278049-11 | 581.5 | 532.9 | Pass |
| LR500 | 0000-278049-12 | 583.7 | 535.1 | Pass |
| LR500 | 0000-278049-13 | 583.3 | 534.7 | Pass |
| LR500 | 0000-278049-14 | 584.3 | 535.7 | Pass |
| LR500 | 0000-278049-15 | 582.9 | 534.3 | Pass |
| LR500 | 0000-278049-16 | 583.4 | 534.8 | Pass |
| LR500 | 0000-278049-17 | 582.5 | 533.9 | Pass |
| LR500 | 0000-278049-18 | 582.4 | 533.8 | Pass |
| LR500 | 0000-278052-01 | 580.4 | 531.8 | Pass |
| LR500 | 0000-278052-02 | 581.2 | 532.6 | Pass |
| LR500 | 0000-278052-03 | 575.5 | 527.0 | Pass |
| LR500 | 0000-278052-04 | 583.0 | 534.4 | Pass |
| LR500 | 0000-278052-05 | 580.4 | 531.8 | Pass |
| LR500 | 0000-278052-06 | 581.8 | 533.2 | Pass |
| LR500 | 0000-278052-07 | 581.4 | 532.8 | Pass |
| LR500 | 0000-278052-08 | 581.6 | 533.0 | Pass |
| LR500 | 0000-278052-09 | 582.0 | 533.4 | Pass |
| LR500 | 0000-278052-10 | 580.8 | 532.2 | Pass |
| LR500 | 0000-278052-11 | 579.8 | 531.2 | Pass |
| LR500 | 0000-278052-12 | 583.5 | 534.9 | Pass |
| LR500 | 0000-278052-13 | 581.0 | 532.4 | Pass |
| LR500 | 0000-278052-14 | 582.1 | 533.5 | Pass |
| LR500 | 0000-278052-15 | 582.4 | 533.8 | Pass |
| LR500 | 0000-278052-16 | 580.6 | 532.0 | Pass |
| LR500 | 0000-278052-17 | 582.4 | 533.8 | Pass |
| LR500 | 0000-278052-18 | 582.6 | 534.0 | Pass |
| LR500 | 0000-279019-01 | 585.4 | 536.8 | Pass |
| LR500 | 0000-279019-02 | 586.0 | 537.4 | Pass |
| LR500 | 0000-279019-03 | 583.8 | 535.2 | Pass |
| LR500 | 0000-279019-04 | 584.7 | 536.1 | Pass |
| LR500 | 0000-279019-05 | 582.8 | 534.2 | Pass |
| LR500 | 0000-279019-06 | 583.3 | 534.7 | Pass |
| LR500 | 0000-279019-07 | 583.7 | 535.1 | Pass |
| LR500 | 0000-279019-08 | 581.3 | 532.7 | Pass |

Reflups Volume Army_2 Bags

12/7/93

| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|--------------|---------------|--------|
| Product | Bag Number | Gross Weight | Volume | Status |
| | | Grams | ml | |
| LR500 | 0000-279019-09 | 582.9 | 534.3 | Pass |
| LR500 | 0000-279019-10 | 583.0 | 534.4 | Pass |
| LR500 | 0000-279019-11 | 584.0 | 535.4 | Pass |
| LR500 | 0000-279019-12 | 584.2 | 535.6 | Pass |
| LR500 | 0000-279019-13 | 578.9 | 530.3 | Pass |
| LR500 | 0000-279019-14 | 583.9 | 535.3 | Pass |
| LR500 | 0000-279019-15 | 583.4 | 534.8 | Pass |
| LR500 | 0000-279019-16 | 583.7 | 535.1 | Pass |
| LR500 | 0000-279019-17 | 582.9 | 534.3 | Pass |
| LR500 | 0000-279019-18 | 583.6 | 535.0 | Pass |
| LR500 | 0000-279021-01 | 584.2 | 535.6 | Pass |
| LR500 | 0000-279021-02 | 583.6 | 535.0 | Pass |
| LR500 | 0000-279021-03 | 583.9 | 535.3 | Pass |
| LR500 | 0000-279021-04 | 583.5 | 534.9 | Pass |
| LR500 | 0000-279021-05 | 582.4 | 533.8 | Pass |
| LR500 | 0000-279021-06 | 583.2 | 534.6 | Pass |
| LR500 | 0000-279021-07 | 584.8 | 536.2 | Pass |
| LR500 | 0000-279021-08 | 583.4 | 534.8 | Pass |
| LR500 | 0000-279021-09 | 582.9 | 534.3 | Pass |
| LR500 | 0000-279021-10 | 583.2 | 534.6 | Pass |
| LR500 | 0000-279021-11 | 499.1 | 450.8 | FAIL |
| LR500 | 0000-279021-12 | 584.2 | 535.6 | Pass |
| LR500 | 0000-279021-13 | 577.2 | 528.7 | Pass |
| LR500 | 0000-279021-14 | 582.4 | 533.8 | Pass |
| LR500 | 0000-279021-15 | 585.3 | 536.7 | Pass |
| LR500 | 0000-279021-16 | 584.1 | 535.5 | Pass |
| LR500 | 0000-279021-17 | 582.4 | 533.8 | Pass |
| LR500 | 0000-279021-18 | 582.7 | 534.1 | Pass |
| LR500 | 0000-279024-01 | 584.7 | 536.1 | Pass |
| LR500 | 0000-279024-02 | 584.0 | 535.4 | Pass |
| LR500 | 0000-279024-03 | 584.1 | 535.5 | Pass |
| LR500 | 0000-279024-04 | 584.6 | 536.0 | Pass |
| LR500 | 0000-279024-05 | 584.2 | 535.6 | Pass |
| LR500 | 0000-279024-06 | 583.2 | 534.6 | Pass |
| LR500 | 0000-279024-07 | 584.0 | 535.4 | Pass |
| LR500 | 0000-279024-08 | 583.6 | 535.0 | Pass |
| LR500 | 0000-279024-09 | 584.8 | 536.2 | Pass |
| LR500 | 0000-279024-10 | 583.7 | 535.1 | Pass |
| LR500 | 0000-279024-11 | 583.5 | 534.9 | Pass |
| LR500 | 0000-279024-12 | 584.4 | 535.8 | Pass |
| LR500 | 0000-279024-13 | 584.2 | 535.6 | Pass |
| LR500 | 0000-279024-14 | 584.1 | 535.5 | Pass |
| LR500 | 0000-279024-15 | 583.2 | 534.6 | Pass |
| LR500 | 0000-279024-16 | 584.3 | 535.7 | Pass |
| LR500 | 0000-279024-17 | 575.3 | 526.8 | Pass |
| LR500 | 0000-279024-18 | 583.3 | 534.7 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | 46.8 | Density(mg/l) | 1.0033 | |
|----------------------------|----------------|-----------------------|--------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-280040-01 | 583.2 | 534.6 | Pass |
| LR500 | 0000-280040-02 | 583.2 | 534.6 | Pass |
| LR500 | 0000-280040-03 | 584.0 | 535.4 | Pass |
| LR500 | 0000-280040-04 | 583.5 | 534.9 | Pass |
| LR500 | 0000-280040-05 | 583.4 | 534.8 | Pass |
| LR500 | 0000-280040-06 | 578.9 | 530.3 | Pass |
| LR500 | 0000-280040-07 | 583.5 | 534.9 | Pass |
| LR500 | 0000-280040-08 | 584.7 | 536.1 | Pass |
| LR500 | 0000-280040-09 | 582.5 | 533.9 | Pass |
| LR500 | 0000-280040-10 | 584.4 | 535.8 | Pass |
| LR500 | 0000-280040-11 | 584.2 | 535.6 | Pass |
| LR500 | 0000-280040-12 | 584.8 | 536.2 | Pass |
| LR500 | 0000-280040-13 | 583.4 | 534.8 | Pass |
| LR500 | 0000-280040-14 | 583.3 | 534.7 | Pass |
| LR500 | 0000-280040-15 | 581.7 | 533.1 | Pass |
| LR500 | 0000-280040-16 | 583.3 | 534.7 | Pass |
| LR500 | 0000-280040-17 | 583.1 | 534.5 | Pass |
| LR500 | 0000-280040-18 | 584.7 | 536.1 | Pass |
| LR500 | 0000-281039-01 | 583.8 | 535.2 | Pass |
| LR500 | 0000-281039-02 | 583.9 | 535.3 | Pass |
| LR500 | 0000-281039-03 | 584.3 | 535.7 | Pass |
| LR500 | 0000-281039-04 | 584.1 | 535.5 | Pass |
| LR500 | 0000-281039-05 | 583.4 | 534.8 | Pass |
| LR500 | 0000-281039-06 | 583.2 | 534.6 | Pass |
| LR500 | 0000-281039-07 | 583.1 | 534.5 | Pass |
| LR500 | 0000-281039-08 | 583.8 | 535.2 | Pass |
| LR500 | 0000-281039-09 | 584.1 | 535.5 | Pass |
| LR500 | 0000-281039-10 | 578.0 | 529.5 | Pass |
| LR500 | 0000-281039-11 | 584.1 | 535.5 | Pass |
| LR500 | 0000-281039-12 | 583.3 | 534.7 | Pass |
| LR500 | 0000-281039-13 | 583.9 | 535.3 | Pass |
| LR500 | 0000-281039-14 | 582.9 | 534.3 | Pass |
| LR500 | 0000-281039-15 | 583.7 | 535.1 | Pass |
| LR500 | 0000-281039-16 | 586.0 | 537.4 | Pass |
| LR500 | 0000-281039-17 | 582.5 | 533.9 | Pass |
| LR500 | 0000-281039-18 | 583.5 | 534.9 | Pass |
| LR500 | 0000-281043-01 | 583.6 | 535.0 | Pass |
| LR500 | 0000-281043-02 | 584.5 | 535.9 | Pass |
| LR500 | 0000-281043-03 | 583.0 | 534.4 | Pass |
| LR500 | 0000-281043-04 | 583.6 | 535.0 | Pass |
| LR500 | 0000-281043-05 | 583.5 | 534.9 | Pass |
| LR500 | 0000-281043-06 | 584.8 | 536.2 | Pass |
| LR500 | 0000-281043-07 | 584.4 | 535.8 | Pass |
| LR500 | 0000-281043-08 | 578.9 | 530.3 | Pass |
| LR500 | 0000-281043-09 | 584.4 | 535.8 | Pass |
| LR500 | 0000-281043-10 | 585.5 | 536.9 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|--------------|---------------|--------|
| Product | Bag Number | Gross Weight | Volume | Status |
| | | Grams | ml | |
| LR500 | 0000-281043-11 | 584.6 | 536.0 | Pass |
| LR500 | 0000-281043-12 | 584.1 | 535.5 | Pass |
| LR500 | 0000-281043-13 | 585.3 | 536.7 | Pass |
| LR500 | 0000-281043-14 | 583.7 | 535.1 | Pass |
| LR500 | 0000-281043-15 | 583.5 | 534.9 | Pass |
| LR500 | 0000-281043-16 | 583.2 | 534.6 | Pass |
| LR500 | 0000-281043-17 | 583.4 | 534.8 | Pass |
| LR500 | 0000-281043-18 | 583.3 | 534.7 | Pass |
| LR500 | 0000-281045-01 | 571.7 | 523.2 | Pass |
| LR500 | 0000-281045-02 | 582.1 | 533.5 | Pass |
| LR500 | 0000-281045-03 | 585.9 | 537.3 | Pass |
| LR500 | 0000-281045-04 | 581.1 | 532.5 | Pass |
| LR500 | 0000-281045-05 | 584.5 | 535.9 | Pass |
| LR500 | 0000-281045-06 | 582.8 | 534.2 | Pass |
| LR500 | 0000-281045-07 | 583.4 | 534.8 | Pass |
| LR500 | 0000-281045-08 | 585.2 | 536.6 | Pass |
| LR500 | 0000-281045-09 | 584.6 | 536.0 | Pass |
| LR500 | 0000-281045-10 | 582.8 | 534.2 | Pass |
| LR500 | 0000-281045-11 | 584.8 | 536.2 | Pass |
| LR500 | 0000-281045-12 | 585.0 | 536.4 | Pass |
| LR500 | 0000-281045-13 | 583.6 | 535.0 | Pass |
| LR500 | 0000-281045-14 | 584.0 | 535.4 | Pass |
| LR500 | 0000-281045-15 | 583.9 | 535.3 | Pass |
| LR500 | 0000-281045-16 | 584.6 | 536.0 | Pass |
| LR500 | 0000-281045-17 | 583.2 | 534.6 | Pass |
| LR500 | 0000-281045-18 | 583.4 | 534.8 | Pass |
| LR500 | 0000-281050-01 | 584.1 | 535.5 | Pass |
| LR500 | 0000-281050-01 | 578.8 | 530.3 | Pass |
| LR500 | 0000-281050-02 | 584.0 | 535.4 | Pass |
| LR500 | 0000-281050-02 | 585.4 | 536.8 | Pass |
| LR500 | 0000-281050-03 | 583.1 | 534.5 | Pass |
| LR500 | 0000-281050-03 | 584.9 | 536.3 | Pass |
| LR500 | 0000-281050-04 | 583.8 | 535.2 | Pass |
| LR500 | 0000-281050-04 | 586.0 | 537.4 | Pass |
| LR500 | 0000-281050-05 | 584.5 | 535.9 | Pass |
| LR500 | 0000-281050-05 | 584.5 | 535.9 | Pass |
| LR500 | 0000-281050-06 | 585.1 | 536.5 | Pass |
| LR500 | 0000-281050-06 | 584.0 | 535.4 | Pass |
| LR500 | 0000-281050-07 | 583.6 | 535.0 | Pass |
| LR500 | 0000-281050-07 | 584.1 | 535.5 | Pass |
| LR500 | 0000-281050-08 | 583.6 | 535.0 | Pass |
| LR500 | 0000-281050-08 | 583.9 | 535.3 | Pass |
| LR500 | 0000-281050-09 | 585.1 | 536.5 | Pass |
| LR500 | 0000-281050-09 | 584.7 | 536.1 | Pass |
| LR500 | 0000-281050-10 | 584.1 | 535.5 | Pass |
| LR500 | 0000-281050-10 | 585.1 | 536.5 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-281050-11 | 584.1 | 535.5 | Pass |
| LR500 | 0000-281050-11 | 584.1 | 535.5 | Pass |
| LR500 | 0000-281050-12 | 583.7 | 535.1 | Pass |
| LR500 | 0000-281050-12 | 584.6 | 536.0 | Pass |
| LR500 | 0000-281050-13 | 584.7 | 536.1 | Pass |
| LR500 | 0000-281050-13 | 585.8 | 537.2 | Pass |
| LR500 | 0000-281050-14 | 584.2 | 535.6 | Pass |
| LR500 | 0000-281050-14 | 583.6 | 535.0 | Pass |
| LR500 | 0000-281050-15 | 583.9 | 535.3 | Pass |
| LR500 | 0000-281050-15 | 584.8 | 536.2 | Pass |
| LR500 | 0000-281050-16 | 585.0 | 536.4 | Pass |
| LR500 | 0000-281050-16 | 585.0 | 536.4 | Pass |
| LR500 | 0000-281050-17 | 584.5 | 535.9 | Pass |
| LR500 | 0000-281050-17 | 584.4 | 535.8 | Pass |
| LR500 | 0000-281050-18 | 578.4 | 529.9 | Pass |
| LR500 | 0000-281050-18 | 585.0 | 536.4 | Pass |
| LR500 | 0000-282022-01 | 578.7 | 530.2 | Pass |
| LR500 | 0000-282022-02 | 584.5 | 535.9 | Pass |
| LR500 | 0000-282022-03 | 585.3 | 536.7 | Pass |
| LR500 | 0000-282022-04 | 584.7 | 536.1 | Pass |
| LR500 | 0000-282022-05 | 586.3 | 537.7 | Pass |
| LR500 | 0000-282022-06 | 583.1 | 534.5 | Pass |
| LR500 | 0000-282022-07 | 585.4 | 536.8 | Pass |
| LR500 | 0000-282022-08 | 585.3 | 536.7 | Pass |
| LR500 | 0000-282022-09 | 584.3 | 535.7 | Pass |
| LR500 | 0000-282022-10 | 584.7 | 536.1 | Pass |
| LR500 | 0000-282022-11 | 583.9 | 535.3 | Pass |
| LR500 | 0000-282022-12 | 585.1 | 536.5 | Pass |
| LR500 | 0000-282022-13 | 586.0 | 537.4 | Pass |
| LR500 | 0000-282022-14 | 584.6 | 536.0 | Pass |
| LR500 | 0000-282022-15 | 584.2 | 535.6 | Pass |
| LR500 | 0000-282022-16 | 584.7 | 536.1 | Pass |
| LR500 | 0000-282022-17 | 584.7 | 536.1 | Pass |
| LR500 | 0000-282022-18 | 583.1 | 534.5 | Pass |
| LR500 | 0000-282027-01 | 584.3 | 535.7 | Pass |
| LR500 | 0000-282027-02 | 581.1 | 532.5 | Pass |
| LR500 | 0000-282027-03 | 584.0 | 535.4 | Pass |
| LR500 | 0000-282027-04 | 586.0 | 537.4 | Pass |
| LR500 | 0000-282027-05 | 582.9 | 534.3 | Pass |
| LR500 | 0000-282027-06 | 581.8 | 533.2 | Pass |
| LR500 | 0000-282027-07 | 584.4 | 535.8 | Pass |
| LR500 | 0000-282027-08 | 585.4 | 536.8 | Pass |
| LR500 | 0000-282027-09 | 584.1 | 535.5 | Pass |
| LR500 | 0000-282027-10 | 582.8 | 534.2 | Pass |
| LR500 | 0000-282027-11 | 583.8 | 535.2 | Pass |
| LR500 | 0000-282027-12 | 581.6 | 533.0 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-282027-13 | 584.5 | 535.9 | Pass |
| LR500 | 0000-282027-14 | 577.7 | 529.2 | Pass |
| LR500 | 0000-282027-15 | 582.7 | 534.1 | Pass |
| LR500 | 0000-282027-16 | 586.6 | 538.0 | Pass |
| LR500 | 0000-282027-17 | 582.1 | 533.5 | Pass |
| LR500 | 0000-282027-18 | 584.4 | 535.8 | Pass |
| LR500 | 0000-282029-01 | 583.0 | 534.4 | Pass |
| LR500 | 0000-282029-02 | 585.8 | 537.2 | Pass |
| LR500 | 0000-282029-03 | 584.6 | 536.0 | Pass |
| LR500 | 0000-282029-04 | 584.2 | 535.6 | Pass |
| LR500 | 0000-282029-05 | 583.1 | 534.5 | Pass |
| LR500 | 0000-282029-06 | 581.9 | 533.3 | Pass |
| LR500 | 0000-282029-07 | 582.3 | 533.7 | Pass |
| LR500 | 0000-282029-08 | 584.3 | 535.7 | Pass |
| LR500 | 0000-282029-09 | 584.2 | 535.6 | Pass |
| LR500 | 0000-282029-10 | 585.9 | 537.3 | Pass |
| LR500 | 0000-282029-11 | 577.4 | 528.9 | Pass |
| LR500 | 0000-282029-12 | 584.0 | 535.4 | Pass |
| LR500 | 0000-282029-13 | 583.0 | 534.4 | Pass |
| LR500 | 0000-282029-14 | 581.7 | 533.1 | Pass |
| LR500 | 0000-282029-16 | 582.3 | 533.7 | Pass |
| LR500 | 0000-282029-17 | 585.1 | 536.5 | Pass |
| LR500 | 0000-282029-18 | 585.2 | 536.6 | Pass |
| LR500 | 0000-283040-01 | 583.3 | 534.7 | Pass |
| LR500 | 0000-283040-02 | 583.6 | 535.0 | Pass |
| LR500 | 0000-283040-03 | 584.5 | 535.9 | Pass |
| LR500 | 0000-283040-04 | 582.8 | 534.2 | Pass |
| LR500 | 0000-283040-05 | 583.5 | 534.9 | Pass |
| LR500 | 0000-283040-06 | 583.2 | 534.6 | Pass |
| LR500 | 0000-283040-07 | 583.3 | 534.7 | Pass |
| LR500 | 0000-283040-08 | 582.8 | 534.2 | Pass |
| LR500 | 0000-283040-09 | 582.6 | 534.0 | Pass |
| LR500 | 0000-283040-10 | 582.4 | 533.8 | Pass |
| LR500 | 0000-283040-11 | 583.6 | 535.0 | Pass |
| LR500 | 0000-283040-12 | 583.1 | 534.5 | Pass |
| LR500 | 0000-283040-13 | 585.4 | 536.8 | Pass |
| LR500 | 0000-283040-14 | 583.5 | 534.9 | Pass |
| LR500 | 0000-283040-15 | 584.1 | 535.5 | Pass |
| LR500 | 0000-283040-16 | 584.5 | 535.9 | Pass |
| LR500 | 0000-283040-17 | 583.9 | 535.3 | Pass |
| LR500 | 0000-283040-18 | 579.7 | 531.1 | Pass |
| LR500 | 0000-283050-01 | 584.1 | 535.5 | Pass |
| LR500 | 0000-283050-02 | 583.5 | 534.9 | Pass |
| LR500 | 0000-283050-03 | 584.0 | 535.4 | Pass |
| LR500 | 0000-283050-04 | 583.9 | 535.3 | Pass |
| LR500 | 0000-283050-05 | 583.4 | 534.8 | Pass |

Reflups Volume Army_2 Bags

| | | | | |
|----------------------------|----------------|--------------|---------------|--------|
| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
| Product | Bag Number | Gross Weight | Volume | Status |
| | | Grams | ml | |
| LR500 | 0000-283050-06 | 584.0 | 535.4 | Pass |
| LR500 | 0000-283050-07 | 583.1 | 534.5 | Pass |
| LR500 | 0000-283050-08 | 584.1 | 535.5 | Pass |
| LR500 | 0000-283050-09 | 583.8 | 535.2 | Pass |
| LR500 | 0000-283050-10 | 580.2 | 531.6 | Pass |
| LR500 | 0000-283050-11 | 584.1 | 535.5 | Pass |
| LR500 | 0000-283050-12 | 584.3 | 535.7 | Pass |
| LR500 | 0000-283050-13 | 583.3 | 534.7 | Pass |
| LR500 | 0000-283050-14 | 584.6 | 536.0 | Pass |
| LR500 | 0000-283050-15 | 585.3 | 536.7 | Pass |
| LR500 | 0000-283050-16 | 583.4 | 534.8 | Pass |
| LR500 | 0000-283050-17 | 584.0 | 535.4 | Pass |
| LR500 | 0000-283050-18 | 583.2 | 534.6 | Pass |
| LR500 | 0000-284024-01 | 579.3 | 530.7 | Pass |
| LR500 | 0000-284024-02 | 578.7 | 530.2 | Pass |
| LR500 | 0000-284024-03 | 581.0 | 532.4 | Pass |
| LR500 | 0000-284024-04 | 579.2 | 530.6 | Pass |
| LR500 | 0000-284024-05 | 579.7 | 531.1 | Pass |
| LR500 | 0000-284024-06 | 578.2 | 529.7 | Pass |
| LR500 | 0000-284024-07 | 580.5 | 531.9 | Pass |
| LR500 | 0000-284024-08 | 580.4 | 531.8 | Pass |
| LR500 | 0000-284024-09 | 580.3 | 531.7 | Pass |
| LR500 | 0000-284024-10 | 578.5 | 530.0 | Pass |
| LR500 | 0000-284024-11 | 580.1 | 531.5 | Pass |
| LR500 | 0000-284024-12 | 577.7 | 529.2 | Pass |
| LR500 | 0000-284024-13 | 583.2 | 534.6 | Pass |
| LR500 | 0000-284024-14 | 578.4 | 529.9 | Pass |
| LR500 | 0000-284024-15 | 579.4 | 530.8 | Pass |
| LR500 | 0000-284024-16 | 580.5 | 531.9 | Pass |
| LR500 | 0000-284024-17 | 579.9 | 531.3 | Pass |
| LR500 | 0000-284024-18 | 572.8 | 524.3 | Pass |
| LR500 | 0000-284025-01 | 584.2 | 535.6 | Pass |
| LR500 | 0000-284025-02 | 582.3 | 533.7 | Pass |
| LR500 | 0000-284025-03 | 577.6 | 529.1 | Pass |
| LR500 | 0000-284025-04 | 583.9 | 535.3 | Pass |
| LR500 | 0000-284025-05 | 585.0 | 536.4 | Pass |
| LR500 | 0000-284025-06 | 584.5 | 535.9 | Pass |
| LR500 | 0000-284025-07 | 583.9 | 535.3 | Pass |
| LR500 | 0000-284025-08 | 584.0 | 535.4 | Pass |
| LR500 | 0000-284025-09 | 585.5 | 536.9 | Pass |
| LR500 | 0000-284025-10 | 584.8 | 536.2 | Pass |
| LR500 | 0000-284025-11 | 585.5 | 536.9 | Pass |
| LR500 | 0000-284025-12 | 582.9 | 534.3 | Pass |
| LR500 | 0000-284025-13 | 583.8 | 535.2 | Pass |
| LR500 | 0000-284025-14 | 587.3 | 538.7 | Pass |
| LR500 | 0000-284025-15 | 582.7 | 534.1 | Pass |

Reflups Volume Army_2 Bags

| | | | | |
|----------------------------|----------------|-----------------------|---------------|--------|
| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-284025-16 | 583.5 | 534.9 | Pass |
| LR500 | 0000-284025-17 | 583.7 | 535.1 | Pass |
| LR500 | 0000-284025-18 | 582.9 | 534.3 | Pass |
| LR500 | 0000-284033-01 | 577.2 | 528.7 | Pass |
| LR500 | 0000-284033-02 | 581.4 | 532.8 | Pass |
| LR500 | 0000-284033-03 | 583.8 | 535.2 | Pass |
| LR500 | 0000-284033-04 | 583.6 | 535.0 | Pass |
| LR500 | 0000-284033-05 | 584.3 | 535.7 | Pass |
| LR500 | 0000-284033-06 | 581.8 | 533.2 | Pass |
| LR500 | 0000-284033-07 | 583.5 | 534.9 | Pass |
| LR500 | 0000-284033-08 | 581.8 | 533.2 | Pass |
| LR500 | 0000-284033-09 | 583.6 | 535.0 | Pass |
| LR500 | 0000-284033-10 | 582.9 | 534.3 | Pass |
| LR500 | 0000-284033-11 | 583.6 | 535.0 | Pass |
| LR500 | 0000-284033-12 | 583.7 | 535.1 | Pass |
| LR500 | 0000-284033-13 | 581.1 | 532.5 | Pass |
| LR500 | 0000-284033-14 | 584.3 | 535.7 | Pass |
| LR500 | 0000-284033-15 | 583.9 | 535.3 | Pass |
| LR500 | 0000-284033-16 | 583.3 | 534.7 | Pass |
| LR500 | 0000-284033-17 | 582.6 | 534.0 | Pass |
| LR500 | 0000-284033-18 | 584.3 | 535.7 | Pass |
| LR500 | 0000-285034-01 | 579.5 | 530.9 | Pass |
| LR500 | 0000-285034-02 | 581.4 | 532.8 | Pass |
| LR500 | 0000-285034-03 | 579.3 | 530.7 | Pass |
| LR500 | 0000-285034-04 | 580.8 | 532.2 | Pass |
| LR500 | 0000-285034-05 | 579.8 | 531.2 | Pass |
| LR500 | 0000-285034-06 | 581.2 | 532.6 | Pass |
| LR500 | 0000-285034-07 | 573.3 | 524.8 | Pass |
| LR500 | 0000-285034-08 | 583.9 | 535.3 | Pass |
| LR500 | 0000-285034-09 | 581.4 | 532.8 | Pass |
| LR500 | 0000-285034-10 | 581.9 | 533.3 | Pass |
| LR500 | 0000-285034-11 | 581.4 | 532.8 | Pass |
| LR500 | 0000-285034-12 | 581.8 | 533.2 | Pass |
| LR500 | 0000-285034-13 | 583.7 | 535.1 | Pass |
| LR500 | 0000-285034-14 | 584.8 | 536.2 | Pass |
| LR500 | 0000-285034-15 | 583.1 | 534.5 | Pass |
| LR500 | 0000-285034-16 | 582.2 | 533.6 | Pass |
| LR500 | 0000-285034-17 | 584.4 | 535.8 | Pass |
| LR500 | 0000-285034-18 | 584.3 | 535.7 | Pass |
| LR500 | 0000-285040-01 | 583.6 | 535.0 | Pass |
| LR500 | 0000-285040-02 | 582.6 | 534.0 | Pass |
| LR500 | 0000-285040-03 | 583.9 | 535.3 | Pass |
| LR500 | 0000-285040-04 | 583.0 | 534.4 | Pass |
| LR500 | 0000-285040-05 | 578.8 | 530.3 | Pass |
| LR500 | 0000-285040-06 | 583.9 | 535.3 | Pass |
| LR500 | 0000-285040-07 | 582.3 | 533.7 | Pass |

Reflups Volume Army_2 Bags

12/7/93

| Bag and Overwrap Weight(g) | | 46.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR500 | 0000-285040-08 | 585.4 | 536.8 | Pass |
| LR500 | 0000-285040-09 | 583.8 | 535.2 | Pass |
| LR500 | 0000-285040-10 | 584.5 | 535.9 | Pass |
| LR500 | 0000-285040-11 | 583.2 | 534.6 | Pass |
| LR500 | 0000-285040-12 | 583.8 | 535.2 | Pass |
| LR500 | 0000-285040-13 | 585.0 | 536.4 | Pass |
| LR500 | 0000-285040-14 | 584.8 | 536.2 | Pass |
| LR500 | 0000-285040-15 | 584.0 | 535.4 | Pass |
| LR500 | 0000-285040-16 | 584.2 | 535.6 | Pass |
| LR500 | 0000-285040-17 | 582.5 | 533.9 | Pass |
| LR500 | 0000-285040-18 | 588.4 | 539.8 | Pass |
| LR500 | 0000-285042-01 | 584.2 | 535.6 | Pass |
| LR500 | 0000-285042-02 | 583.4 | 534.8 | Pass |
| LR500 | 0000-285042-03 | 581.2 | 532.6 | Pass |
| LR500 | 0000-285042-04 | 584.6 | 536.0 | Pass |
| LR500 | 0000-285042-05 | 583.3 | 534.7 | Pass |
| LR500 | 0000-285042-06 | 583.6 | 535.0 | Pass |
| LR500 | 0000-285042-07 | 585.0 | 536.4 | Pass |
| LR500 | 0000-285042-08 | 582.4 | 533.8 | Pass |
| LR500 | 0000-285042-09 | 584.3 | 535.7 | Pass |
| LR500 | 0000-285042-10 | 582.8 | 534.2 | Pass |
| LR500 | 0000-285042-11 | 578.2 | 529.7 | Pass |
| LR500 | 0000-285042-12 | 584.5 | 535.9 | Pass |
| LR500 | 0000-285042-13 | 584.9 | 536.3 | Pass |
| LR500 | 0000-285042-14 | 583.6 | 535.0 | Pass |
| LR500 | 0000-285042-15 | 586.7 | 538.1 | Pass |
| LR500 | 0000-285042-16 | 587.2 | 538.6 | Pass |
| LR500 | 0000-285042-17 | 584.9 | 536.3 | Pass |
| LR500 | 0000-285042-18 | 584.1 | 535.5 | Pass |
| Number | 810.0 | Failures | | 3 |

Reflups Volume Army_2 Bags

12/7/93

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-269028-01 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-269028-02 | 1125 | 1056.7 | Pass |
| LR1000 | 0000-269028-03 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-269028-04 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-269028-05 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-269028-06 | 1116.7 | 1048.4 | Pass |
| LR1000 | 0000-269028-07 | 1123.2 | 1054.9 | Pass |
| LR1000 | 0000-269028-08 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-269028-09 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-269028-10 | 1117.6 | 1049.3 | Pass |
| LR1000 | 0000-269028-11 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-269028-12 | 1015 | 947.1 | FAIL |
| LR1000 | 0000-269028-13 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-269028-14 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-269028-15 | 1120 | 1051.7 | Pass |
| LR1000 | 0000-269028-16 | 1116.9 | 1048.6 | Pass |
| LR1000 | 0000-269028-17 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-269028-18 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-270038-01 | 1114.1 | 1045.8 | Pass |
| LR1000 | 0000-270038-02 | 1124.7 | 1056.4 | Pass |
| LR1000 | 0000-270038-03 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-270038-04 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-270038-05 | 1123.6 | 1055.3 | Pass |
| LR1000 | 0000-270038-06 | 602.4 | 535.8 | FAIL |
| LR1000 | 0000-270038-07 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-270038-08 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-270038-09 | 1122.5 | 1054.2 | Pass |
| LR1000 | 0000-270038-10 | 1123.6 | 1055.3 | Pass |
| LR1000 | 0000-270038-11 | 1122.5 | 1054.2 | Pass |
| LR1000 | 0000-270038-13 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-270038-14 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-270038-15 | 1123.6 | 1055.3 | Pass |
| LR1000 | 0000-270038-16 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-270038-17 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-270038-18 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-271033-01 | 1117.8 | 1049.5 | Pass |
| LR1000 | 0000-271033-01 | 1120.5 | 1052.2 | Pass |
| LR1000 | 0000-271033-02 | 1126.9 | 1058.6 | Pass |
| LR1000 | 0000-271033-02 | 1129.5 | 1061.2 | Pass |
| LR1000 | 0000-271033-03 | 1120.6 | 1052.3 | Pass |
| LR1000 | 0000-271033-03 | 1126.3 | 1058.0 | Pass |
| LR1000 | 0000-271033-04 | 1125 | 1056.7 | Pass |
| LR1000 | 0000-271033-04 | 1129.8 | 1061.5 | Pass |
| LR1000 | 0000-271033-05 | 1123.8 | 1055.5 | Pass |
| LR1000 | 0000-271033-05 | 1126.6 | 1058.3 | Pass |
| LR1000 | 0000-271033-06 | 1124.3 | 1056.0 | Pass |

Reflups Volume Army_2 Bags

| | | | | |
|----------------------------|----------------|--------------|---------------|--------|
| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
| Product | Bag Number | Gross Weight | Volume | Status |
| | | Grams | ml | |
| LR1000 | 0000-271033-06 | 1129.7 | 1061.4 | Pass |
| LR1000 | 0000-271033-07 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-271033-07 | 1126.6 | 1058.3 | Pass |
| LR1000 | 0000-271033-08 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-271033-08 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-271033-09 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-271033-09 | 1123.7 | 1055.4 | Pass |
| LR1000 | 0000-271033-10 | 1124.4 | 1056.1 | Pass |
| LR1000 | 0000-271033-10 | 1127.3 | 1059.0 | Pass |
| LR1000 | 0000-271033-11 | 1124.4 | 1056.1 | Pass |
| LR1000 | 0000-271033-11 | 1127.3 | 1059.0 | Pass |
| LR1000 | 0000-271033-12 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-271033-12 | 1125.7 | 1057.4 | Pass |
| LR1000 | 0000-271033-13 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-271033-13 | 1124.4 | 1056.1 | Pass |
| LR1000 | 0000-271033-14 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-271033-15 | 1124.4 | 1056.1 | Pass |
| LR1000 | 0000-271033-15 | 1127.3 | 1059.0 | Pass |
| LR1000 | 0000-271033-16 | 1127.6 | 1059.3 | Pass |
| LR1000 | 0000-271033-17 | 1123.8 | 1055.5 | Pass |
| LR1000 | 0000-271033-18 | 1125.1 | 1056.8 | Pass |
| LR1000 | 0000-271039-01 | 1125.4 | 1057.1 | Pass |
| LR1000 | 0000-271039-02 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-271039-03 | 1124.2 | 1055.9 | Pass |
| LR1000 | 0000-271039-04 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-271039-05 | 1116 | 1047.7 | Pass |
| LR1000 | 0000-271039-06 | 1125.6 | 1057.3 | Pass |
| LR1000 | 0000-271039-07 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-271039-08 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-271039-09 | 1120.5 | 1052.2 | Pass |
| LR1000 | 0000-271039-10 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-271039-11 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-271039-12 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-271039-13 | 1123.7 | 1055.4 | Pass |
| LR1000 | 0000-271039-14 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-271039-15 | 1125.7 | 1057.4 | Pass |
| LR1000 | 0000-271039-16 | 1122.5 | 1054.2 | Pass |
| LR1000 | 0000-271039-17 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-271039-18 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-272028-01 | 1103.5 | 1035.3 | Pass |
| LR1000 | 0000-272028-01 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-272028-02 | 1123.6 | 1055.3 | Pass |
| LR1000 | 0000-272028-02 | 1122.5 | 1054.2 | Pass |
| LR1000 | 0000-272028-03 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-272028-03 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-272028-04 | 1117 | 1048.7 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-272028-04 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-272028-05 | 1123.8 | 1055.5 | Pass |
| LR1000 | 0000-272028-05 | 1126.5 | 1058.2 | Pass |
| LR1000 | 0000-272028-06 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-272028-06 | 1125.9 | 1057.6 | Pass |
| LR1000 | 0000-272028-07 | 1124.6 | 1056.3 | Pass |
| LR1000 | 0000-272028-07 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-272028-08 | 1123.7 | 1055.4 | Pass |
| LR1000 | 0000-272028-08 | 1124.1 | 1055.8 | Pass |
| LR1000 | 0000-272028-09 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-272028-09 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-272028-10 | 1124 | 1055.7 | Pass |
| LR1000 | 0000-272028-10 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-272028-11 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-272028-11 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-272028-12 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-272028-12 | 1119.6 | 1051.3 | Pass |
| LR1000 | 0000-272028-13 | 1122.5 | 1054.2 | Pass |
| LR1000 | 0000-272028-13 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-272028-14 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-272028-14 | 1119.1 | 1050.8 | Pass |
| LR1000 | 0000-272028-15 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-272028-15 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-272028-16 | 1121 | 1052.7 | Pass |
| LR1000 | 0000-272028-16 | 1119.8 | 1051.5 | Pass |
| LR1000 | 0000-272028-17 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-272028-17 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-272028-18 | 1123.2 | 1054.9 | Pass |
| LR1000 | 0000-272028-18 | 1124.4 | 1056.1 | Pass |
| LR1000 | 0000-272031-01 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-272031-02 | 1117.4 | 1049.1 | Pass |
| LR1000 | 0000-272031-03 | 1115.9 | 1047.6 | Pass |
| LR1000 | 0000-272031-04 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-272031-05 | 1125.3 | 1057.0 | Pass |
| LR1000 | 0000-272031-06 | 1127.4 | 1059.1 | Pass |
| LR1000 | 0000-272031-07 | 1124.9 | 1056.6 | Pass |
| LR1000 | 0000-272031-08 | 1134 | 1065.7 | Pass |
| LR1000 | 0000-272031-09 | 1125.4 | 1057.1 | Pass |
| LR1000 | 0000-272031-10 | 1121.2 | 1052.9 | Pass |
| LR1000 | 0000-272031-11 | 1118.8 | 1050.5 | Pass |
| LR1000 | 0000-272031-12 | 1117.6 | 1049.3 | Pass |
| LR1000 | 0000-272031-13 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-272031-14 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-272031-15 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-272031-16 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-272031-17 | 1123.4 | 1055.1 | Pass |

Reflups Volume Army_2 Bags

| | | | | |
|----------------------------|----------------|-----------------------|---------------|--------|
| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-272031-18 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-273029-01 | 1125.5 | 1057.2 | Pass |
| LR1000 | 0000-273029-02 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-273029-03 | 1127.4 | 1059.1 | Pass |
| LR1000 | 0000-273029-04 | 1125.2 | 1056.9 | Pass |
| LR1000 | 0000-273029-05 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-273029-06 | 1128.5 | 1060.2 | Pass |
| LR1000 | 0000-273029-07 | 1124.8 | 1056.5 | Pass |
| LR1000 | 0000-273029-08 | 1126.1 | 1057.8 | Pass |
| LR1000 | 0000-273029-09 | 1125.1 | 1056.8 | Pass |
| LR1000 | 0000-273029-10 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-273029-11 | 1127.7 | 1059.4 | Pass |
| LR1000 | 0000-273029-12 | 1126.3 | 1058.0 | Pass |
| LR1000 | 0000-273029-13 | 1126.3 | 1058.0 | Pass |
| LR1000 | 0000-273029-14 | 1125.6 | 1057.3 | Pass |
| LR1000 | 0000-273029-15 | 1127.6 | 1059.3 | Pass |
| LR1000 | 0000-273029-16 | 1125.7 | 1057.4 | Pass |
| LR1000 | 0000-273029-17 | 1118.7 | 1050.4 | Pass |
| LR1000 | 0000-273029-18 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-273035-16 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-273035-17 | 1119.4 | 1051.1 | Pass |
| LR1000 | 0000-273035-18 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-274041-01 | 1093.4 | 1025.2 | Pass |
| LR1000 | 0000-274041-02 | 1002.5 | 934.6 | FAIL |
| LR1000 | 0000-274041-03 | 1117.6 | 1049.3 | Pass |
| LR1000 | 0000-274041-04 | 1125.2 | 1056.9 | Pass |
| LR1000 | 0000-274041-05 | 1124.4 | 1056.1 | Pass |
| LR1000 | 0000-274041-06 | 1125.4 | 1057.1 | Pass |
| LR1000 | 0000-274041-07 | 1124.4 | 1056.1 | Pass |
| LR1000 | 0000-274041-08 | 1124.4 | 1056.1 | Pass |
| LR1000 | 0000-274041-09 | 1125.3 | 1057.0 | Pass |
| LR1000 | 0000-274041-10 | 1124.3 | 1056.0 | Pass |
| LR1000 | 0000-274041-11 | 1125.1 | 1056.8 | Pass |
| LR1000 | 0000-274041-12 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-274041-13 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-274041-14 | 1125.3 | 1057.0 | Pass |
| LR1000 | 0000-274041-15 | 1116.4 | 1048.1 | Pass |
| LR1000 | 0000-274041-16 | 1124.1 | 1055.8 | Pass |
| LR1000 | 0000-274041-17 | 1125.7 | 1057.4 | Pass |
| LR1000 | 0000-274041-18 | 1124.9 | 1056.6 | Pass |
| LR1000 | 0000-274045-01 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-274045-01 | 1043.8 | 975.8 | FAIL |
| LR1000 | 0000-274045-02 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-274045-02 | 1126.2 | 1057.9 | Pass |
| LR1000 | 0000-274045-03 | 1120.1 | 1051.8 | Pass |
| LR1000 | 0000-274045-03 | 1123.8 | 1055.5 | Pass |

Reflups Volume Army_2 Bags

12/7/93

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-274045-04 | 1119.3 | 1051.0 | Pass |
| LR1000 | 0000-274045-04 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-274045-05 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-274045-05 | 1124 | 1055.7 | Pass |
| LR1000 | 0000-274045-06 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-274045-06 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-274045-07 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-274045-07 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-274045-08 | 1121.2 | 1052.9 | Pass |
| LR1000 | 0000-274045-08 | 1125.6 | 1057.3 | Pass |
| LR1000 | 0000-274045-09 | 1123.2 | 1054.9 | Pass |
| LR1000 | 0000-274045-09 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-274045-10 | 1119.6 | 1051.3 | Pass |
| LR1000 | 0000-274045-10 | 1116.4 | 1048.1 | Pass |
| LR1000 | 0000-274045-11 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-274045-11 | 1125 | 1056.7 | Pass |
| LR1000 | 0000-274045-12 | 1120.3 | 1052.0 | Pass |
| LR1000 | 0000-274045-12 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-274045-13 | 1109.8 | 1041.6 | Pass |
| LR1000 | 0000-274045-13 | 1123.8 | 1055.5 | Pass |
| LR1000 | 0000-274045-14 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-274045-14 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-274045-15 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-274045-15 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-274045-16 | 1114.7 | 1046.4 | Pass |
| LR1000 | 0000-274045-16 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-274045-17 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-274045-17 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-274045-18 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-274045-18 | 1124.9 | 1056.6 | Pass |
| LR1000 | 0000-275041-01 | 1109.7 | 1041.5 | Pass |
| LR1000 | 0000-275041-02 | 1125.4 | 1057.1 | Pass |
| LR1000 | 0000-275041-03 | 1117.6 | 1049.3 | Pass |
| LR1000 | 0000-275041-04 | 1123.8 | 1055.5 | Pass |
| LR1000 | 0000-275041-05 | 1125.3 | 1057.0 | Pass |
| LR1000 | 0000-275041-06 | 1125.5 | 1057.2 | Pass |
| LR1000 | 0000-275041-07 | 1127 | 1058.7 | Pass |
| LR1000 | 0000-275041-08 | 1124.9 | 1056.6 | Pass |
| LR1000 | 0000-275041-09 | 1124.5 | 1056.2 | Pass |
| LR1000 | 0000-275041-10 | 1123.6 | 1055.3 | Pass |
| LR1000 | 0000-275041-11 | 1124.5 | 1056.2 | Pass |
| LR1000 | 0000-275041-12 | 1124.3 | 1056.0 | Pass |
| LR1000 | 0000-275041-13 | 1125.6 | 1057.3 | Pass |
| LR1000 | 0000-275041-14 | 1127.1 | 1058.8 | Pass |
| LR1000 | 0000-275041-15 | 1125 | 1056.7 | Pass |
| LR1000 | 0000-275041-16 | 1125.6 | 1057.3 | Pass |

Reflups Volume Army_2 Bags

12/7/93

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-275041-17 | 1125.2 | 1056.9 | Pass |
| LR1000 | 0000-275041-18 | 1123.7 | 1055.4 | Pass |
| LR1000 | 0000-275042-01 | 1120.1 | 1051.8 | Pass |
| LR1000 | 0000-275042-02 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-275042-03 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-275042-04 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-275042-05 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-275042-06 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-275042-07 | 1119.1 | 1050.8 | Pass |
| LR1000 | 0000-275042-08 | 1120 | 1051.7 | Pass |
| LR1000 | 0000-275042-09 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-275042-10 | 1109.1 | 1040.9 | Pass |
| LR1000 | 0000-275042-11 | 1120.6 | 1052.3 | Pass |
| LR1000 | 0000-275042-12 | 1121.2 | 1052.9 | Pass |
| LR1000 | 0000-275042-13 | 1115.5 | 1047.2 | Pass |
| LR1000 | 0000-275042-14 | 1121 | 1052.7 | Pass |
| LR1000 | 0000-275042-15 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-275042-16 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-275042-17 | 1121.1 | 1052.8 | Pass |
| LR1000 | 0000-275042-18 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-275046-01 | 1124.4 | 1056.1 | Pass |
| LR1000 | 0000-275046-02 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-275046-03 | 1098.2 | 1030.0 | Pass |
| LR1000 | 0000-275046-04 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-275046-05 | 1087.4 | 1019.2 | Pass |
| LR1000 | 0000-275046-06 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-275046-07 | 983.1 | 915.3 | FAIL |
| LR1000 | 0000-275046-08 | 1125.1 | 1056.8 | Pass |
| LR1000 | 0000-275046-09 | 1116.9 | 1048.6 | Pass |
| LR1000 | 0000-275046-10 | 1124.5 | 1056.2 | Pass |
| LR1000 | 0000-275046-11 | 1116.6 | 1048.3 | Pass |
| LR1000 | 0000-275046-12 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-275046-13 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-275046-14 | 1125.7 | 1057.4 | Pass |
| LR1000 | 0000-275046-15 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-275046-16 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-275046-17 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-275046-18 | 1123.2 | 1054.9 | Pass |
| LR1000 | 0000-275050-01 | 1123.6 | 1055.3 | Pass |
| LR1000 | 0000-275050-02 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-275050-03 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-275050-04 | 1121.1 | 1052.8 | Pass |
| LR1000 | 0000-275050-05 | 1119.6 | 1051.3 | Pass |
| LR1000 | 0000-275050-06 | 1120.6 | 1052.3 | Pass |
| LR1000 | 0000-275050-07 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-275050-08 | 1120.9 | 1052.6 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|--------------|---------------|--------|
| Product | Bag Number | Gross Weight | Volume | Status |
| | | Grams | ml | |
| LR1000 | 0000-275050-09 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-275050-10 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-275050-11 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-275050-12 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-275050-13 | 1110.3 | 1042.1 | Pass |
| LR1000 | 0000-275050-14 | 1121.1 | 1052.8 | Pass |
| LR1000 | 0000-275050-15 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-275050-16 | 1119.7 | 1051.4 | Pass |
| LR1000 | 0000-275050-17 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-275050-18 | 1117.5 | 1049.2 | Pass |
| LR1000 | 0000-276008-01 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-276008-02 | 1121.2 | 1052.9 | Pass |
| LR1000 | 0000-276008-03 | 1116.5 | 1048.2 | Pass |
| LR1000 | 0000-276008-05 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-276008-06 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-276008-07 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-276008-08 | 1120.1 | 1051.8 | Pass |
| LR1000 | 0000-276008-09 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-276008-10 | 1121.1 | 1052.8 | Pass |
| LR1000 | 0000-276008-11 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-276008-12 | 1121 | 1052.7 | Pass |
| LR1000 | 0000-276008-13 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-276008-14 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-276008-15 | 1120.4 | 1052.1 | Pass |
| LR1000 | 0000-276008-16 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-276008-17 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-276008-18 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-276013-01 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-276013-02 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-276013-03 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-276013-04 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-276013-05 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-276013-06 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-276013-07 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-276013-08 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-276013-09 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-276013-10 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-276013-11 | 1115.3 | 1047.0 | Pass |
| LR1000 | 0000-276013-12 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-276013-13 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-276013-14 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-276013-15 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-276013-16 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-276013-17 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-276013-18 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-277033-01 | 983.5 | 915.7 | FAIL |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-277033-02 | 1118.8 | 1050.5 | Pass |
| LR1000 | 0000-277033-03 | 1125 | 1056.7 | Pass |
| LR1000 | 0000-277033-04 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-277033-05 | 1126.1 | 1057.8 | Pass |
| LR1000 | 0000-277033-06 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-277033-07 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-277033-08 | 1124.8 | 1056.5 | Pass |
| LR1000 | 0000-277033-09 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-277033-10 | 1127.4 | 1059.1 | Pass |
| LR1000 | 0000-277033-11 | 1124.9 | 1056.6 | Pass |
| LR1000 | 0000-277033-12 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-277033-13 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-277033-14 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-277033-15 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-277033-16 | 1102.2 | 1034.0 | Pass |
| LR1000 | 0000-277033-17 | 1125.7 | 1057.4 | Pass |
| LR1000 | 0000-277033-18 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-277037-01 | 1124.2 | 1055.9 | Pass |
| LR1000 | 0000-277037-02 | 1125.2 | 1056.9 | Pass |
| LR1000 | 0000-277037-03 | 1125.7 | 1057.4 | Pass |
| LR1000 | 0000-277037-04 | 1125.1 | 1056.8 | Pass |
| LR1000 | 0000-277037-05 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-277037-06 | 1124.2 | 1055.9 | Pass |
| LR1000 | 0000-277037-07 | 1125.1 | 1056.8 | Pass |
| LR1000 | 0000-277037-08 | 1124 | 1055.7 | Pass |
| LR1000 | 0000-277037-09 | 1123 | 1054.7 | Pass |
| LR1000 | 0000-277037-10 | 1123.6 | 1055.3 | Pass |
| LR1000 | 0000-277037-11 | 1125.2 | 1056.9 | Pass |
| LR1000 | 0000-277037-12 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-277037-13 | 1125.3 | 1057.0 | Pass |
| LR1000 | 0000-277037-14 | 1123 | 1054.7 | Pass |
| LR1000 | 0000-277037-15 | 1117.1 | 1048.8 | Pass |
| LR1000 | 0000-277037-16 | 1124.5 | 1056.2 | Pass |
| LR1000 | 0000-277037-17 | 1126.7 | 1058.4 | Pass |
| LR1000 | 0000-277037-18 | 1124.1 | 1055.8 | Pass |
| LR1000 | 0000-278037-01 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-278037-02 | 1116.3 | 1048.0 | Pass |
| LR1000 | 0000-278037-03 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-278037-04 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-278037-05 | 1124.5 | 1056.2 | Pass |
| LR1000 | 0000-278037-06 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-278037-07 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-278037-08 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-278037-09 | 1123.8 | 1055.5 | Pass |
| LR1000 | 0000-278037-10 | 1124.3 | 1056.0 | Pass |
| LR1000 | 0000-278037-11 | 1123.8 | 1055.5 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-278037-12 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-278037-13 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-278037-14 | 1124.5 | 1056.2 | Pass |
| LR1000 | 0000-278037-15 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-278037-16 | 1099 | 1030.8 | Pass |
| LR1000 | 0000-278037-17 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-278037-18 | 1123.7 | 1055.4 | Pass |
| LR1000 | 0000-279034-01 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-279034-02 | 1114.9 | 1046.6 | Pass |
| LR1000 | 0000-279034-03 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-279034-04 | 1119.9 | 1051.6 | Pass |
| LR1000 | 0000-279034-05 | 1121.4 | 1053.1 | Pass |
| LR1000 | 0000-279034-06 | 1119.9 | 1051.6 | Pass |
| LR1000 | 0000-279034-07 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-279034-08 | 1121.4 | 1053.1 | Pass |
| LR1000 | 0000-279034-09 | 1119.3 | 1051.0 | Pass |
| LR1000 | 0000-279034-10 | 1120.2 | 1051.9 | Pass |
| LR1000 | 0000-279034-11 | 1120 | 1051.7 | Pass |
| LR1000 | 0000-279034-12 | 1120.1 | 1051.8 | Pass |
| LR1000 | 0000-279034-13 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-279034-14 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-279034-15 | 1121 | 1052.7 | Pass |
| LR1000 | 0000-279034-16 | 1121.4 | 1053.1 | Pass |
| LR1000 | 0000-279034-17 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-279034-18 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-279039-01 | 1121.2 | 1052.9 | Pass |
| LR1000 | 0000-279039-01 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-279039-02 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-279039-02 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-279039-03 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-279039-03 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-279039-04 | 1124.2 | 1055.9 | Pass |
| LR1000 | 0000-279039-04 | 1115.8 | 1047.5 | Pass |
| LR1000 | 0000-279039-05 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-279039-05 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-279039-06 | 1123.2 | 1054.9 | Pass |
| LR1000 | 0000-279039-06 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-279039-07 | 1124.2 | 1055.9 | Pass |
| LR1000 | 0000-279039-07 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-279039-08 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-279039-08 | 1121.3 | 1053.0 | Pass |
| LR1000 | 0000-279039-09 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-279039-09 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-279039-10 | 1124 | 1055.7 | Pass |
| LR1000 | 0000-279039-10 | 1124.5 | 1056.2 | Pass |
| LR1000 | 0000-279039-11 | 1123.1 | 1054.8 | Pass |

Reflups Volume Army_2 Bags

12/7/93

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-279039-11 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-279039-12 | 1123.6 | 1055.3 | Pass |
| LR1000 | 0000-279039-12 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-279039-13 | 1125.6 | 1057.3 | Pass |
| LR1000 | 0000-279039-13 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-279039-14 | 1123.8 | 1055.5 | Pass |
| LR1000 | 0000-279039-14 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-279039-15 | 1126.5 | 1058.2 | Pass |
| LR1000 | 0000-279039-15 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-279039-16 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-279039-16 | 1118.5 | 1050.2 | Pass |
| LR1000 | 0000-279039-17 | 1125.4 | 1057.1 | Pass |
| LR1000 | 0000-279039-17 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-279039-18 | 1125.2 | 1056.9 | Pass |
| LR1000 | 0000-279039-18 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-279040-01 | 1123.8 | 1055.5 | Pass |
| LR1000 | 0000-279040-02 | 1125.6 | 1057.3 | Pass |
| LR1000 | 0000-279040-03 | 1117.8 | 1049.5 | Pass |
| LR1000 | 0000-279040-04 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-279040-05 | 1123.7 | 1055.4 | Pass |
| LR1000 | 0000-279040-06 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-279040-07 | 899.9 | 832.4 | FAIL |
| LR1000 | 0000-279040-08 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-279040-09 | 1123 | 1054.7 | Pass |
| LR1000 | 0000-279040-10 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-279040-11 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-279040-12 | 1123 | 1054.7 | Pass |
| LR1000 | 0000-279040-13 | 1124.7 | 1056.4 | Pass |
| LR1000 | 0000-279040-14 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-279040-15 | 1123.6 | 1055.3 | Pass |
| LR1000 | 0000-279040-16 | 1123.8 | 1055.5 | Pass |
| LR1000 | 0000-279040-17 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-279040-18 | 1124.1 | 1055.8 | Pass |
| LR1000 | 0000-279042-01 | 1123.9 | 1055.6 | Pass |
| LR1000 | 0000-279042-02 | 1123.7 | 1055.4 | Pass |
| LR1000 | 0000-279042-03 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-279042-04 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-279042-05 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-279042-06 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-279042-07 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-279042-08 | 1123.6 | 1055.3 | Pass |
| LR1000 | 0000-279042-09 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-279042-10 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-279042-11 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-279042-12 | 1124.8 | 1056.5 | Pass |
| LR1000 | 0000-279042-13 | 1124.4 | 1056.1 | Pass |

Reflups Volume Army_2 Bags

12/7/93

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-279042-14 | 1123.7 | 1055.4 | Pass |
| LR1000 | 0000-279042-15 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-279042-16 | 1117.1 | 1048.8 | Pass |
| LR1000 | 0000-279042-17 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-279042-18 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-280021-01 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-280021-02 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-280021-03 | 1121.3 | 1053.0 | Pass |
| LR1000 | 0000-280021-04 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-280021-05 | 1121.1 | 1052.8 | Pass |
| LR1000 | 0000-280021-06 | 1078.2 | 1010.1 | Pass |
| LR1000 | 0000-280021-07 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-280021-08 | 1119.9 | 1051.6 | Pass |
| LR1000 | 0000-280021-09 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-280021-10 | 1120.6 | 1052.3 | Pass |
| LR1000 | 0000-280021-11 | 1121 | 1052.7 | Pass |
| LR1000 | 0000-280021-12 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-280021-13 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-280021-14 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-280021-15 | 1119.8 | 1051.5 | Pass |
| LR1000 | 0000-280021-16 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-280021-17 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-280021-18 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-280024-01 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-280024-02 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-280024-03 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-280024-04 | 1120.3 | 1052.0 | Pass |
| LR1000 | 0000-280024-05 | 1120.2 | 1051.9 | Pass |
| LR1000 | 0000-280024-06 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-280024-07 | 1121 | 1052.7 | Pass |
| LR1000 | 0000-280024-08 | 1119 | 1050.7 | Pass |
| LR1000 | 0000-280024-09 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-280024-10 | 1118.7 | 1050.4 | Pass |
| LR1000 | 0000-280024-11 | 1121.3 | 1053.0 | Pass |
| LR1000 | 0000-280024-12 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-280026-01 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-280026-02 | 1119.4 | 1051.1 | Pass |
| LR1000 | 0000-280026-03 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-280026-04 | 1122.5 | 1054.2 | Pass |
| LR1000 | 0000-280026-05 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-280026-06 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-280026-07 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-280026-08 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-280026-09 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-280026-10 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-280026-11 | 1123.7 | 1055.4 | Pass |

Reflups Volume Army_2 Bags

12/7/93

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-280026-12 | 1123.2 | 1054.9 | Pass |
| LR1000 | 0000-280030-01 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-280030-02 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-280030-03 | 1120.4 | 1052.1 | Pass |
| LR1000 | 0000-280030-04 | 1123.2 | 1054.9 | Pass |
| LR1000 | 0000-280030-05 | 1120 | 1051.7 | Pass |
| LR1000 | 0000-280030-06 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-280030-07 | 1120.6 | 1052.3 | Pass |
| LR1000 | 0000-280030-08 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-280030-09 | 1121 | 1052.7 | Pass |
| LR1000 | 0000-280030-10 | 1097.4 | 1029.2 | Pass |
| LR1000 | 0000-280030-11 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-280030-12 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-280030-13 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-280030-14 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-280030-15 | 1115.2 | 1046.9 | Pass |
| LR1000 | 0000-280030-16 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-280030-17 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-280030-18 | 1121.8 | Bad Bag | Pass |
| LR1000 | 0000-281017-01 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-281017-01 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-281017-02 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-281017-02 | 1122.5 | 1054.2 | Pass |
| LR1000 | 0000-281017-03 | 1121 | 1052.7 | Pass |
| LR1000 | 0000-281017-03 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-281017-04 | 1117.9 | 1049.6 | Pass |
| LR1000 | 0000-281017-04 | 1119.5 | 1051.2 | Pass |
| LR1000 | 0000-281017-05 | 1121.3 | 1053.0 | Pass |
| LR1000 | 0000-281017-05 | 1119.8 | 1051.5 | Pass |
| LR1000 | 0000-281017-06 | 1115.9 | 1047.6 | Pass |
| LR1000 | 0000-281017-06 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-281017-07 | 1109 | 1040.8 | Pass |
| LR1000 | 0000-281017-07 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-281017-08 | 1104.1 | 1035.9 | Pass |
| LR1000 | 0000-281017-08 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-281017-09 | 1104.8 | 1036.6 | Pass |
| LR1000 | 0000-281017-09 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-281017-10 | 1097.1 | 1028.9 | Pass |
| LR1000 | 0000-281017-10 | 1120.5 | 1052.2 | Pass |
| LR1000 | 0000-281017-11 | 1105.4 | 1037.2 | Pass |
| LR1000 | 0000-281017-11 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-281017-12 | 1102.5 | 1034.3 | Pass |
| LR1000 | 0000-281017-12 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-281017-13 | 1110.6 | 1042.4 | Pass |
| LR1000 | 0000-281017-14 | 1103.3 | 1035.1 | Pass |
| LR1000 | 0000-281017-15 | 1102.8 | 1034.6 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-281017-16 | 1107.3 | 1039.1 | Pass |
| LR1000 | 0000-281017-17 | 1095.1 | 1026.9 | Pass |
| LR1000 | 0000-281017-18 | 1108.2 | 1040.0 | Pass |
| LR1000 | 0000-281019-01 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-281019-02 | 1115.4 | 1047.1 | Pass |
| LR1000 | 0000-281019-03 | 1121.2 | 1052.9 | Pass |
| LR1000 | 0000-281019-04 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-281019-05 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-281019-06 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-281019-07 | 1119.7 | 1051.4 | Pass |
| LR1000 | 0000-281019-08 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-281019-09 | 1121.3 | 1053.0 | Pass |
| LR1000 | 0000-281019-10 | 1121.2 | 1052.9 | Pass |
| LR1000 | 0000-281019-11 | 1120.5 | 1052.2 | Pass |
| LR1000 | 0000-281019-12 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-281019-13 | 1120.3 | 1052.0 | Pass |
| LR1000 | 0000-281019-14 | 1122.5 | 1054.2 | Pass |
| LR1000 | 0000-281019-15 | 1045.3 | 977.3 | FAIL |
| LR1000 | 0000-281019-16 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-281019-17 | 1120.2 | 1051.9 | Pass |
| LR1000 | 0000-281019-18 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-281021-01 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-281021-02 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-281021-03 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-281021-04 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-281021-05 | 1116.8 | 1048.5 | Pass |
| LR1000 | 0000-281021-06 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-281021-07 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-281021-08 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-281021-09 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-281021-10 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-281021-11 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-281021-12 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-281021-13 | 1094.1 | 1025.9 | Pass |
| LR1000 | 0000-281021-14 | 1124.7 | 1056.4 | Pass |
| LR1000 | 0000-281021-15 | 1102.8 | 1034.6 | Pass |
| LR1000 | 0000-281021-16 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-281021-17 | 1121.2 | 1052.9 | Pass |
| LR1000 | 0000-281021-18 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-282034-01 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-282034-02 | 1120.4 | 1052.1 | Pass |
| LR1000 | 0000-282034-03 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-282034-04 | 1121.3 | 1053.0 | Pass |
| LR1000 | 0000-282034-05 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-282034-06 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-282034-07 | 1121.4 | 1053.1 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-282034-08 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-282034-09 | 1120 | 1051.7 | Pass |
| LR1000 | 0000-282034-10 | 1117.9 | 1049.6 | Pass |
| LR1000 | 0000-282034-11 | 1120.1 | 1051.8 | Pass |
| LR1000 | 0000-282034-12 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-282034-13 | 1118.7 | 1050.4 | Pass |
| LR1000 | 0000-282034-14 | 1119.9 | 1051.6 | Pass |
| LR1000 | 0000-282034-15 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-282034-16 | 1121.4 | 1053.1 | Pass |
| LR1000 | 0000-282034-17 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-282034-18 | 1120.2 | 1051.9 | Pass |
| LR1000 | 0000-282044-01 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-282044-02 | 1120.6 | 1052.3 | Pass |
| LR1000 | 0000-282044-03 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-282044-04 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-282044-05 | 1125.7 | 1057.4 | Pass |
| LR1000 | 0000-282044-06 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-282044-07 | 1124.5 | 1056.2 | Pass |
| LR1000 | 0000-282044-08 | 1124.5 | 1056.2 | Pass |
| LR1000 | 0000-282044-09 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-282044-10 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-282044-11 | 1116.5 | 1048.2 | Pass |
| LR1000 | 0000-282044-12 | 1120.4 | 1052.1 | Pass |
| LR1000 | 0000-282044-13 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-282044-14 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-282044-15 | 1124.1 | 1055.8 | Pass |
| LR1000 | 0000-282044-16 | 1124 | 1055.7 | Pass |
| LR1000 | 0000-282044-17 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-282044-18 | 1124.7 | 1056.4 | Pass |
| LR1000 | 0000-282045-01 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-282045-02 | 1124.8 | 1056.5 | Pass |
| LR1000 | 0000-282045-03 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-282045-04 | 1119.5 | 1051.2 | Pass |
| LR1000 | 0000-282045-05 | 1124.2 | 1055.9 | Pass |
| LR1000 | 0000-282045-06 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-282045-07 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-282045-08 | 1121.1 | 1052.8 | Pass |
| LR1000 | 0000-282045-09 | 1123.2 | 1054.9 | Pass |
| LR1000 | 0000-282045-10 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-282045-11 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-282045-12 | 1123 | 1054.7 | Pass |
| LR1000 | 0000-282045-13 | 1122.5 | 1054.2 | Pass |
| LR1000 | 0000-282045-14 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-282045-15 | 1116 | 1047.7 | Pass |
| LR1000 | 0000-282045-16 | 1125.6 | 1057.3 | Pass |
| LR1000 | 0000-282045-17 | 1121 | 1052.7 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-282045-18 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-283010-01 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-283010-02 | 1119.4 | 1051.1 | Pass |
| LR1000 | 0000-283010-03 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-283010-04 | 1120 | 1051.7 | Pass |
| LR1000 | 0000-283010-05 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-283010-06 | 1119.5 | 1051.2 | Pass |
| LR1000 | 0000-283010-07 | 1119.6 | 1051.3 | Pass |
| LR1000 | 0000-283010-08 | 1119.2 | 1050.9 | Pass |
| LR1000 | 0000-283010-09 | 1113.8 | 1045.5 | Pass |
| LR1000 | 0000-283010-10 | 1118.3 | 1050.0 | Pass |
| LR1000 | 0000-283010-11 | 1120.4 | 1052.1 | Pass |
| LR1000 | 0000-283010-12 | 1119.4 | 1051.1 | Pass |
| LR1000 | 0000-283010-13 | 1120 | 1051.7 | Pass |
| LR1000 | 0000-283010-14 | 1121.2 | 1052.9 | Pass |
| LR1000 | 0000-283010-15 | 1119.6 | 1051.3 | Pass |
| LR1000 | 0000-283010-16 | 1099.2 | 1031.0 | Pass |
| LR1000 | 0000-283010-17 | 1118.8 | 1050.5 | Pass |
| LR1000 | 0000-283010-18 | 1119.1 | 1050.8 | Pass |
| LR1000 | 0000-283013-01 | 1118 | 1049.7 | Pass |
| LR1000 | 0000-283013-02 | 1119.2 | 1050.9 | Pass |
| LR1000 | 0000-283013-03 | 1118.5 | 1050.2 | Pass |
| LR1000 | 0000-283013-04 | 1119.4 | 1051.1 | Pass |
| LR1000 | 0000-283013-05 | 1118.5 | 1050.2 | Pass |
| LR1000 | 0000-283013-06 | 1117.6 | 1049.3 | Pass |
| LR1000 | 0000-283013-07 | 1121.3 | 1053.0 | Pass |
| LR1000 | 0000-283013-08 | 1120.5 | 1052.2 | Pass |
| LR1000 | 0000-283013-09 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-283013-10 | 1118.5 | 1050.2 | Pass |
| LR1000 | 0000-283013-11 | 1117.1 | 1048.8 | Pass |
| LR1000 | 0000-283013-12 | 1118.5 | 1050.2 | Pass |
| LR1000 | 0000-283015-01 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-283015-02 | 1121.4 | 1053.1 | Pass |
| LR1000 | 0000-283015-03 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-283015-04 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-283015-05 | 1119.8 | 1051.5 | Pass |
| LR1000 | 0000-283015-06 | 1120.6 | 1052.3 | Pass |
| LR1000 | 0000-283015-07 | 1114.2 | 1045.9 | Pass |
| LR1000 | 0000-283015-08 | 1121.1 | 1052.8 | Pass |
| LR1000 | 0000-283015-09 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-283015-10 | 1120.3 | 1052.0 | Pass |
| LR1000 | 0000-283015-11 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-283015-12 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-283015-13 | 1121.4 | 1053.1 | Pass |
| LR1000 | 0000-283015-14 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-283015-15 | 1122 | 1053.7 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-283015-16 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-283015-17 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-283015-18 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-283021-01 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-283021-02 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-283021-03 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-283021-04 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-283021-05 | 1114.8 | 1046.5 | Pass |
| LR1000 | 0000-283021-06 | 1120.3 | 1052.0 | Pass |
| LR1000 | 0000-283021-07 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-283021-08 | 1120.7 | 1052.4 | Pass |
| LR1000 | 0000-283021-09 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-283021-10 | 1121.4 | 1053.1 | Pass |
| LR1000 | 0000-283021-11 | 1120.4 | 1052.1 | Pass |
| LR1000 | 0000-283021-12 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-283021-13 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-283021-14 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-283021-15 | 1120.6 | 1052.3 | Pass |
| LR1000 | 0000-283021-16 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-283021-17 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-283021-18 | 1120.5 | Bad Bag | Pass |
| LR1000 | 0000-284035-01 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-284035-02 | 1124 | 1055.7 | Pass |
| LR1000 | 0000-284035-03 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-284035-04 | 1121 | 1052.7 | Pass |
| LR1000 | 0000-284035-05 | 1121.1 | 1052.8 | Pass |
| LR1000 | 0000-284035-06 | 1119.9 | 1051.6 | Pass |
| LR1000 | 0000-284035-07 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-284035-08 | 1124.6 | 1056.3 | Pass |
| LR1000 | 0000-284035-09 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-284035-10 | 1122.9 | 1054.6 | Pass |
| LR1000 | 0000-284035-11 | 1120.1 | 1051.8 | Pass |
| LR1000 | 0000-284035-12 | 1119.6 | 1051.3 | Pass |
| LR1000 | 0000-284037-01 | 1121.1 | 1052.8 | Pass |
| LR1000 | 0000-284037-02 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-284037-03 | 1119.8 | 1051.5 | Pass |
| LR1000 | 0000-284037-04 | 1119.5 | 1051.2 | Pass |
| LR1000 | 0000-284037-05 | 1120.4 | 1052.1 | Pass |
| LR1000 | 0000-284037-06 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-284037-07 | 1113.3 | 1045.1 | Pass |
| LR1000 | 0000-284037-08 | 1120.4 | 1052.1 | Pass |
| LR1000 | 0000-284037-09 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-284037-10 | 1119.9 | 1051.6 | Pass |
| LR1000 | 0000-284037-11 | 1120.6 | 1052.3 | Pass |
| LR1000 | 0000-284037-12 | 1120 | 1051.7 | Pass |
| LR1000 | 0000-284042-01 | 1122.9 | 1054.6 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-284042-01 | 1119.7 | 1051.4 | Pass |
| LR1000 | 0000-284042-02 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-284042-02 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-284042-03 | 1116.1 | 1047.8 | Pass |
| LR1000 | 0000-284042-03 | 1119.6 | 1051.3 | Pass |
| LR1000 | 0000-284042-04 | 1123 | 1054.7 | Pass |
| LR1000 | 0000-284042-04 | 1120.9 | 1052.6 | Pass |
| LR1000 | 0000-284042-05 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-284042-05 | 1121.1 | 1052.8 | Pass |
| LR1000 | 0000-284042-06 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-284042-06 | 1120 | 1051.7 | Pass |
| LR1000 | 0000-284042-07 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-284042-07 | 1120.5 | 1052.2 | Pass |
| LR1000 | 0000-284042-08 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-284042-08 | 1115.9 | 1047.6 | Pass |
| LR1000 | 0000-284042-09 | 1122.5 | 1054.2 | Pass |
| LR1000 | 0000-284042-09 | 1120.6 | 1052.3 | Pass |
| LR1000 | 0000-284042-10 | 1125.2 | 1056.9 | Pass |
| LR1000 | 0000-284042-10 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-284042-11 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-284042-11 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-284042-12 | 1125.1 | 1056.8 | Pass |
| LR1000 | 0000-284042-12 | 1120.3 | 1052.0 | Pass |
| LR1000 | 0000-284042-13 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-284042-13 | 1118.8 | 1050.5 | Pass |
| LR1000 | 0000-284042-14 | 1123.1 | 1054.8 | Pass |
| LR1000 | 0000-284042-14 | 1117.8 | 1049.5 | Pass |
| LR1000 | 0000-284042-15 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-284042-15 | 1119.6 | 1051.3 | Pass |
| LR1000 | 0000-284042-16 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-284042-16 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-284042-17 | 1125.5 | 1057.2 | Pass |
| LR1000 | 0000-284042-17 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-284042-18 | 1122.6 | 1054.3 | Pass |
| LR1000 | 0000-284042-18 | 1119.2 | 1050.9 | Pass |
| LR1000 | 0000-285005-01 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-285005-02 | 1125.4 | 1057.1 | Pass |
| LR1000 | 0000-285005-03 | 1115.8 | 1047.5 | Pass |
| LR1000 | 0000-285005-04 | 1118.2 | 1049.9 | Pass |
| LR1000 | 0000-285005-05 | 1125.6 | 1057.3 | Pass |
| LR1000 | 0000-285005-06 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-285005-07 | 1127 | 1058.7 | Pass |
| LR1000 | 0000-285005-08 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-285005-09 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-285005-10 | 1113.6 | 1045.4 | Pass |
| LR1000 | 0000-285005-11 | 1120.3 | 1052.0 | Pass |

Reflups Volume Army_2 Bags

12/7/93

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0033 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| LR1000 | 0000-285005-12 | 1120.4 | 1052.1 | Pass |
| LR1000 | 0000-285005-13 | 1100.2 | 1032.0 | Pass |
| LR1000 | 0000-285005-14 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-285005-15 | 1116.2 | 1047.9 | Pass |
| LR1000 | 0000-285005-16 | 1121.9 | 1053.6 | Pass |
| LR1000 | 0000-285005-17 | 1130.8 | 1062.5 | Pass |
| LR1000 | 0000-285005-18 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-285008-01 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-285008-02 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-285008-03 | 1121.5 | 1053.2 | Pass |
| LR1000 | 0000-285008-04 | 1123 | 1054.7 | Pass |
| LR1000 | 0000-285008-05 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-285008-06 | 1122.1 | 1053.8 | Pass |
| LR1000 | 0000-285008-07 | 1122.8 | 1054.5 | Pass |
| LR1000 | 0000-285008-08 | 1122 | 1053.7 | Pass |
| LR1000 | 0000-285008-09 | 1122.4 | 1054.1 | Pass |
| LR1000 | 0000-285008-10 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-285008-11 | 1124.3 | 1056.0 | Pass |
| LR1000 | 0000-285008-12 | 1123.5 | 1055.2 | Pass |
| LR1000 | 0000-285008-13 | 1123.4 | 1055.1 | Pass |
| LR1000 | 0000-285008-14 | 1121.3 | 1053.0 | Pass |
| LR1000 | 0000-285008-15 | 1121.6 | 1053.3 | Pass |
| LR1000 | 0000-285008-16 | 1122.7 | 1054.4 | Pass |
| LR1000 | 0000-285008-17 | 1116.3 | 1048.0 | Pass |
| LR1000 | 0000-285008-18 | 1123.4 | Bad Bag | Pass |
| LR1000 | 0000-285055-01 | 1120.8 | 1052.5 | Pass |
| LR1000 | 0000-285055-02 | 1121.4 | 1053.1 | Pass |
| LR1000 | 0000-285055-03 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-285055-04 | 1121 | 1052.7 | Pass |
| LR1000 | 0000-285055-05 | 1122.3 | 1054.0 | Pass |
| LR1000 | 0000-285055-06 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-285055-07 | 1121.8 | 1053.5 | Pass |
| LR1000 | 0000-285055-08 | 1116.4 | 1048.1 | Pass |
| LR1000 | 0000-285055-09 | 1124.7 | 1056.4 | Pass |
| LR1000 | 0000-285055-10 | 1121.4 | 1053.1 | Pass |
| LR1000 | 0000-285055-11 | 1121.2 | 1052.9 | Pass |
| LR1000 | 0000-285055-12 | 1123.3 | 1055.0 | Pass |
| LR1000 | 0000-285055-13 | 1123 | 1054.7 | Pass |
| LR1000 | 0000-285055-14 | 1122.5 | 1054.2 | Pass |
| LR1000 | 0000-285055-15 | 1124.7 | 1056.4 | Pass |
| LR1000 | 0000-285055-16 | 1122.2 | 1053.9 | Pass |
| LR1000 | 0000-285055-17 | 1121.7 | 1053.4 | Pass |
| LR1000 | 0000-285055-18 | 1120.2 | 1051.9 | Pass |
| Number | 864 | | Failures | 8 |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-269020-01 | 1125.7 | 1057.1 | Pass |
| NACL | 0000-269020-02 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-269020-03 | 1125.5 | 1056.9 | Pass |
| NACL | 0000-269020-04 | 1125.2 | 1056.6 | Pass |
| NACL | 0000-269020-05 | 1124.3 | 1055.7 | Pass |
| NACL | 0000-269020-06 | 1127.3 | 1058.7 | Pass |
| NACL | 0000-269020-07 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-269020-08 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-269020-09 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-269020-10 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-269020-11 | 1118.0 | 1049.4 | Pass |
| NACL | 0000-269020-12 | 1124.3 | 1055.7 | Pass |
| NACL | 0000-269020-13 | 1127.0 | 1058.4 | Pass |
| NACL | 0000-269020-14 | 1125.7 | 1057.1 | Pass |
| NACL | 0000-269020-15 | 1124.7 | 1056.1 | Pass |
| NACL | 0000-269020-16 | 1125.2 | 1056.6 | Pass |
| NACL | 0000-269020-17 | 1124.4 | 1055.8 | Pass |
| NACL | 0000-269020-18 | 1125.8 | 1057.2 | Pass |
| NACL | 0000-270031-01 | 1120.7 | 1052.1 | Pass |
| NACL | 0000-270031-02 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-270031-03 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-270031-04 | 1117.5 | 1048.9 | Pass |
| NACL | 0000-270031-05 | 1121.1 | 1052.5 | Pass |
| NACL | 0000-270031-06 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-270031-07 | 1118.9 | 1050.3 | Pass |
| NACL | 0000-270031-08 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-270031-09 | 1120.6 | 1052.0 | Pass |
| NACL | 0000-270031-10 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-270031-11 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-270031-12 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-270031-13 | 1120.5 | 1051.9 | Pass |
| NACL | 0000-270031-14 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-270031-15 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-270031-16 | 1094.2 | 1025.7 | Pass |
| NACL | 0000-270031-17 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-270031-18 | 1121.2 | 1052.6 | Pass |
| NACL | 0000-270033-01 | 1121.6 | 1053.0 | Pass |
| NACL | 0000-270033-02 | 1124.1 | 1055.5 | Pass |
| NACL | 0000-270033-03 | 1124.3 | 1055.7 | Pass |
| NACL | 0000-270033-04 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-270033-05 | 1124.9 | 1056.3 | Pass |
| NACL | 0000-270033-06 | 1119.8 | 1051.2 | Pass |
| NACL | 0000-270033-07 | 1126.6 | 1058.0 | Pass |
| NACL | 0000-270033-08 | 1388.6 | 1319.1 | FAIL |
| NACL | 0000-270033-09 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-270033-10 | 1119.4 | 1050.8 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-270033-11 | 1124.0 | 1055.4 | Pass |
| NACL | 0000-270033-12 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-270033-13 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-270033-14 | 1120.6 | 1052.0 | Pass |
| NACL | 0000-270033-15 | 1124.1 | 1055.5 | Pass |
| NACL | 0000-270033-16 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-270033-17 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-270033-18 | 1125.3 | 1056.7 | Pass |
| NACL | 0000-271005-01 | 1126.3 | 1057.7 | Pass |
| NACL | 0000-271005-02 | 1124.1 | 1055.5 | Pass |
| NACL | 0000-271005-03 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-271005-04 | 1124.7 | 1056.1 | Pass |
| NACL | 0000-271005-05 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-271005-06 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-271005-07 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-271005-08 | 1125.0 | 1056.4 | Pass |
| NACL | 0000-271005-09 | 1119.0 | 1050.4 | Pass |
| NACL | 0000-271005-10 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-271005-11 | 1120.1 | 1051.5 | Pass |
| NACL | 0000-271005-12 | 1119.7 | 1051.1 | Pass |
| NACL | 0000-271005-13 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-271005-14 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-271005-15 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-271005-16 | 1119.1 | 1050.5 | Pass |
| NACL | 0000-271005-17 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-271005-18 | 1121.3 | 1052.7 | Pass |
| NACL | 0000-272020-01 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-272020-02 | 1113.1 | 1044.5 | Pass |
| NACL | 0000-272020-03 | 1130.8 | 1062.2 | Pass |
| NACL | 0000-272020-04 | 1131.2 | 1062.6 | Pass |
| NACL | 0000-272020-05 | 1125.8 | 1057.2 | Pass |
| NACL | 0000-272020-06 | 1131.4 | 1062.8 | Pass |
| NACL | 0000-272020-07 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-272020-08 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-272020-09 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-272020-10 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-272020-11 | 1120.9 | 1052.3 | Pass |
| NACL | 0000-272020-12 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-272020-13 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-272020-14 | 1121.6 | 1053.0 | Pass |
| NACL | 0000-272020-15 | 1125.5 | 1056.9 | Pass |
| NACL | 0000-272020-16 | 1120.7 | 1052.1 | Pass |
| NACL | 0000-272020-17 | 1124.2 | 1055.6 | Pass |
| NACL | 0000-272020-18 | 1124.4 | 1055.8 | Pass |
| NACL | 0000-272022-01 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-272022-02 | 1124.0 | 1055.4 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-272022-03 | 1124.1 | 1055.5 | Pass |
| NACL | 0000-272022-04 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-272022-05 | 1120.9 | 1052.3 | Pass |
| NACL | 0000-272022-06 | 1124.1 | 1055.5 | Pass |
| NACL | 0000-272022-07 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-272022-08 | 1119.9 | 1051.3 | Pass |
| NACL | 0000-272022-09 | 1126.8 | 1058.2 | Pass |
| NACL | 0000-272022-10 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-272022-11 | 1091.9 | 1023.4 | Pass |
| NACL | 0000-272022-12 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-272022-13 | 1124.7 | 1056.1 | Pass |
| NACL | 0000-272022-14 | 1122.8 | 1054.2 | Pass |
| NACL | 0000-272022-15 | 1116.9 | 1048.3 | Pass |
| NACL | 0000-272022-16 | 1123.3 | 1054.7 | Pass |
| NACL | 0000-272022-17 | 1124.0 | 1055.4 | Pass |
| NACL | 0000-272022-18 | 1101.1 | 1032.6 | Pass |
| NACL | 0000-273007-01 | 1119.8 | 1051.2 | Pass |
| NACL | 0000-273007-02 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-273007-03 | 1116.9 | 1048.3 | Pass |
| NACL | 0000-273007-04 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-273007-05 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-273007-06 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-273007-07 | 1123.4 | 1054.8 | Pass |
| NACL | 0000-273007-08 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-273007-09 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-273007-10 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-273007-11 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-273007-12 | 1124.5 | 1055.9 | Pass |
| NACL | 0000-273007-13 | 1124.8 | 1056.2 | Pass |
| NACL | 0000-273007-14 | 1122.8 | 1054.2 | Pass |
| NACL | 0000-273007-15 | 1123.3 | 1054.7 | Pass |
| NACL | 0000-273007-16 | 1123.4 | 1054.8 | Pass |
| NACL | 0000-273007-17 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-273007-18 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-273011-01 | 1120.4 | 1051.8 | Pass |
| NACL | 0000-273011-02 | 1124.6 | 1056.0 | Pass |
| NACL | 0000-273011-03 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-273011-04 | 1124.7 | 1056.1 | Pass |
| NACL | 0000-273011-05 | 1123.1 | 1054.5 | Pass |
| NACL | 0000-273011-06 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-273011-07 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-273011-08 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-273011-09 | 1124.4 | 1055.8 | Pass |
| NACL | 0000-273011-10 | 1063.1 | 994.7 | FAIL |
| NACL | 0000-273011-11 | 1117.1 | 1048.5 | Pass |
| NACL | 0000-273011-12 | 1120.7 | 1052.1 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-273011-13 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-273011-14 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-273011-15 | 1126.5 | 1057.9 | Pass |
| NACL | 0000-273011-16 | 1124.1 | 1055.5 | Pass |
| NACL | 0000-273011-17 | 1106.4 | 1037.9 | Pass |
| NACL | 0000-273011-18 | 1116.8 | Bad Bag | Pass |
| NACL | 0000-274006-01 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-274006-02 | 1125.9 | 1057.3 | Pass |
| NACL | 0000-274006-03 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-274006-04 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-274006-05 | 1092.2 | 1023.7 | Pass |
| NACL | 0000-274006-06 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-274006-07 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-274006-08 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-274006-09 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-274006-10 | 1114.1 | 1045.5 | Pass |
| NACL | 0000-274006-11 | 1121.2 | 1052.6 | Pass |
| NACL | 0000-274006-12 | 1120.0 | 1051.4 | Pass |
| NACL | 0000-274006-13 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-274006-14 | 1119.7 | 1051.1 | Pass |
| NACL | 0000-274006-15 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-274006-16 | 1120.9 | 1052.3 | Pass |
| NACL | 0000-274006-17 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-274006-18 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-274007-01 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-274007-02 | 1123.4 | 1054.8 | Pass |
| NACL | 0000-274007-03 | 1116.3 | 1047.7 | Pass |
| NACL | 0000-274007-04 | 1120.7 | 1052.1 | Pass |
| NACL | 0000-274007-05 | 1121.2 | 1052.6 | Pass |
| NACL | 0000-274007-06 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-274007-07 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-274007-08 | 1124.7 | 1056.1 | Pass |
| NACL | 0000-274007-09 | 1119.6 | 1051.0 | Pass |
| NACL | 0000-274007-10 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-274007-11 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-274007-12 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-274007-13 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-274007-14 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-274007-15 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-274007-16 | 1120.9 | 1052.3 | Pass |
| NACL | 0000-274007-17 | 1124.9 | 1056.3 | Pass |
| NACL | 0000-274007-18 | 1122.8 | Bad Bag | Pass |
| NACL | 0000-275002-01 | 1122.8 | 1054.2 | Pass |
| NACL | 0000-275002-02 | 1097.2 | 1028.7 | Pass |
| NACL | 0000-275002-03 | 1101.4 | 1032.9 | Pass |
| NACL | 0000-275002-04 | 1122.6 | 1054.0 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-275002-05 | 1121.4 | 1052.8 | Pass |
| NACL | 0000-275002-06 | 1123.4 | 1054.8 | Pass |
| NACL | 0000-275002-07 | 1140.0 | 1071.3 | Pass |
| NACL | 0000-275002-08 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-275002-09 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-275002-10 | 1120.0 | 1051.4 | Pass |
| NACL | 0000-275002-11 | 1122.8 | 1054.2 | Pass |
| NACL | 0000-275002-12 | 1120.5 | 1051.9 | Pass |
| NACL | 0000-275002-13 | 1125.4 | 1056.8 | Pass |
| NACL | 0000-275002-14 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-275002-15 | 1123.1 | 1054.5 | Pass |
| NACL | 0000-275002-16 | 1122.8 | 1054.2 | Pass |
| NACL | 0000-275002-17 | 1115.3 | 1046.7 | Pass |
| NACL | 0000-275002-18 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-275007-01 | 1125.5 | 1056.9 | Pass |
| NACL | 0000-275007-02 | 1125.1 | 1056.5 | Pass |
| NACL | 0000-275007-03 | 1127.7 | 1059.1 | Pass |
| NACL | 0000-275007-04 | 1125.3 | 1056.7 | Pass |
| NACL | 0000-275007-05 | 1125.1 | 1056.5 | Pass |
| NACL | 0000-275007-06 | 1127.6 | 1059.0 | Pass |
| NACL | 0000-275007-07 | 1124.9 | 1056.3 | Pass |
| NACL | 0000-275007-08 | 1125.5 | 1056.9 | Pass |
| NACL | 0000-275007-09 | 1125.7 | 1057.1 | Pass |
| NACL | 0000-275007-10 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-275007-11 | 1126.2 | 1057.6 | Pass |
| NACL | 0000-275007-12 | 1125.3 | 1056.7 | Pass |
| NACL | 0000-275007-13 | 1126.2 | 1057.6 | Pass |
| NACL | 0000-275007-14 | 1126.3 | 1057.7 | Pass |
| NACL | 0000-275007-15 | 1125.4 | 1056.8 | Pass |
| NACL | 0000-275007-16 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-275007-17 | 1124.6 | 1056.0 | Pass |
| NACL | 0000-275007-18 | 1126.1 | 1057.5 | Pass |
| NACL | 0000-276018-01 | 1125.3 | 1056.7 | Pass |
| NACL | 0000-276018-02 | 1124.8 | 1056.2 | Pass |
| NACL | 0000-276018-03 | 1128.3 | 1059.7 | Pass |
| NACL | 0000-276018-04 | 1125.1 | 1056.5 | Pass |
| NACL | 0000-276018-05 | 1124.8 | 1056.2 | Pass |
| NACL | 0000-276018-06 | 1129.1 | 1060.5 | Pass |
| NACL | 0000-276018-07 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-276018-08 | 1125.0 | 1056.4 | Pass |
| NACL | 0000-276018-09 | 1125.5 | 1056.9 | Pass |
| NACL | 0000-276018-10 | 1126.7 | 1058.1 | Pass |
| NACL | 0000-276018-11 | 1124.4 | 1055.8 | Pass |
| NACL | 0000-276018-12 | 1120.1 | 1051.5 | Pass |
| NACL | 0000-276018-13 | 1125.1 | 1056.5 | Pass |
| NACL | 0000-276018-14 | 1125.3 | 1056.7 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-276018-15 | 1126.4 | 1057.8 | Pass |
| NACL | 0000-276018-16 | 1126.1 | 1057.5 | Pass |
| NACL | 0000-276018-17 | 1125.9 | 1057.3 | Pass |
| NACL | 0000-276018-18 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-277021-01 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-277021-02 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-277021-03 | 1126.5 | 1057.9 | Pass |
| NACL | 0000-277021-04 | 1120.6 | 1052.0 | Pass |
| NACL | 0000-277021-05 | 1115.7 | 1047.1 | Pass |
| NACL | 0000-277021-06 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-277021-07 | 1123.1 | 1054.5 | Pass |
| NACL | 0000-277021-08 | 1068.8 | 1000.4 | Pass |
| NACL | 0000-277021-09 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-277021-10 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-277021-11 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-277021-12 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-277021-13 | 1139.6 | 1070.9 | Pass |
| NACL | 0000-277021-14 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-277021-15 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-277021-16 | 1120.1 | 1051.5 | Pass |
| NACL | 0000-277021-17 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-277021-18 | 1120.2 | 1051.6 | Pass |
| NACL | 0000-277023-01 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-277023-02 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-277023-03 | 1116.1 | 1047.5 | Pass |
| NACL | 0000-277023-04 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-277023-05 | 1125.4 | 1056.8 | Pass |
| NACL | 0000-277023-06 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-277023-07 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-277023-08 | 1120.3 | 1051.7 | Pass |
| NACL | 0000-277023-09 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-277023-10 | 1121.1 | 1052.5 | Pass |
| NACL | 0000-277023-11 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-277023-12 | 1120.9 | 1052.3 | Pass |
| NACL | 0000-277023-13 | 1124.1 | 1055.5 | Pass |
| NACL | 0000-277023-14 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-277023-15 | 1125.2 | 1056.6 | Pass |
| NACL | 0000-277023-16 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-277023-17 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-277023-18 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-278006-01 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-278006-02 | 1121.0 | 1052.4 | Pass |
| NACL | 0000-278006-03 | 1114.7 | 1046.1 | Pass |
| NACL | 0000-278006-04 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-278006-05 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-278006-06 | 1121.3 | 1052.7 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-278006-07 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-278006-08 | 1121.3 | 1052.7 | Pass |
| NACL | 0000-278006-09 | 1119.2 | 1050.6 | Pass |
| NACL | 0000-278006-10 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-278006-11 | 1121.0 | 1052.4 | Pass |
| NACL | 0000-278006-12 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-278006-13 | 1123.3 | 1054.7 | Pass |
| NACL | 0000-278006-14 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-278006-15 | 1115.6 | 1047.0 | Pass |
| NACL | 0000-278006-16 | 1124.6 | 1056.0 | Pass |
| NACL | 0000-278006-17 | 1123.3 | 1054.7 | Pass |
| NACL | 0000-278006-18 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-278007-01 | 1121.4 | 1052.8 | Pass |
| NACL | 0000-278007-02 | 1120.8 | 1052.2 | Pass |
| NACL | 0000-278007-03 | 1121.4 | 1052.8 | Pass |
| NACL | 0000-278007-04 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-278007-05 | 1121.7 | 1053.1 | Pass |
| NACL | 0000-278007-06 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-278007-07 | 1120.4 | 1051.8 | Pass |
| NACL | 0000-278007-08 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-278007-09 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-278007-10 | 1114.5 | 1045.9 | Pass |
| NACL | 0000-278007-11 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-278007-12 | 1120.9 | 1052.3 | Pass |
| NACL | 0000-278007-13 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-278007-14 | 1120.8 | 1052.2 | Pass |
| NACL | 0000-278007-15 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-278007-16 | 1112.0 | 1043.4 | Pass |
| NACL | 0000-278007-17 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-278007-18 | 1120.5 | 1051.9 | Pass |
| NACL | 0000-278009-01 | 1116.0 | 1047.4 | Pass |
| NACL | 0000-278009-02 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-278009-03 | 1124.0 | 1055.4 | Pass |
| NACL | 0000-278009-04 | 1124.4 | 1055.8 | Pass |
| NACL | 0000-278009-05 | 1109.0 | 1040.5 | Pass |
| NACL | 0000-278009-06 | 1124.3 | 1055.7 | Pass |
| NACL | 0000-278009-07 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-278009-08 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-278009-09 | 1123.3 | 1054.7 | Pass |
| NACL | 0000-278009-10 | 1104.7 | 1036.2 | Pass |
| NACL | 0000-278009-11 | 1123.3 | 1054.7 | Pass |
| NACL | 0000-278009-12 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-278009-13 | 1123.4 | 1054.8 | Pass |
| NACL | 0000-278009-14 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-278009-15 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-278009-16 | 1123.6 | 1055.0 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-278009-17 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-278009-18 | 1124.2 | 1055.6 | Pass |
| NACL | 0000-279046-01 | 1121.2 | 1052.6 | Pass |
| NACL | 0000-279046-02 | 1121.6 | 1053.0 | Pass |
| NACL | 0000-279046-03 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-279046-04 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-279046-05 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-279046-06 | 1121.1 | 1052.5 | Pass |
| NACL | 0000-279046-07 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-279046-08 | 1120.9 | 1052.3 | Pass |
| NACL | 0000-279046-09 | 1119.9 | 1051.3 | Pass |
| NACL | 0000-279046-10 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-279046-11 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-279046-12 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-279046-13 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-279046-14 | 1120.5 | 1051.9 | Pass |
| NACL | 0000-279046-15 | 1120.5 | 1051.9 | Pass |
| NACL | 0000-279046-16 | 1115.9 | 1047.3 | Pass |
| NACL | 0000-279046-17 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-279046-18 | 1120.9 | 1052.3 | Pass |
| NACL | 0000-279047-01 | 1121.1 | 1052.5 | Pass |
| NACL | 0000-279047-02 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-279047-03 | 1123.1 | 1054.5 | Pass |
| NACL | 0000-279047-04 | 1121.1 | 1052.5 | Pass |
| NACL | 0000-279047-05 | 1114.4 | 1045.8 | Pass |
| NACL | 0000-279047-06 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-279047-07 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-279047-08 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-279047-09 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-279047-10 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-279047-11 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-279047-12 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-279047-13 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-279047-14 | 1111.1 | 1042.5 | Pass |
| NACL | 0000-279047-15 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-279047-16 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-279047-17 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-279047-18 | 1123.4 | 1054.8 | Pass |
| NACL | 0000-279050-01 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-279050-02 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-279050-03 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-279050-04 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-279050-05 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-279050-06 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-279050-07 | 1121.6 | 1053.0 | Pass |
| NACL | 0000-279050-08 | 1120.0 | 1051.4 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-279050-09 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-279050-10 | 1121.1 | 1052.5 | Pass |
| NACL | 0000-279050-11 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-279050-12 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-279050-13 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-279050-14 | 1124.5 | 1055.9 | Pass |
| NACL | 0000-279050-15 | 1113.9 | 1045.3 | Pass |
| NACL | 0000-279050-16 | 1114.4 | 1045.8 | Pass |
| NACL | 0000-279050-17 | 1120.9 | 1052.3 | Pass |
| NACL | 0000-279050-18 | 1121.4 | Bad Bag | Pass |
| NACL | 0000-279052-01 | 1121.4 | 1052.8 | Pass |
| NACL | 0000-279052-02 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-279052-03 | 1120.5 | 1051.9 | Pass |
| NACL | 0000-279052-04 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-279052-05 | 1116.8 | 1048.2 | Pass |
| NACL | 0000-279052-06 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-279052-07 | 1126.3 | 1057.7 | Pass |
| NACL | 0000-279052-08 | 1125.2 | 1056.6 | Pass |
| NACL | 0000-279052-09 | 1124.5 | 1055.9 | Pass |
| NACL | 0000-279052-10 | 1124.8 | 1056.2 | Pass |
| NACL | 0000-279052-11 | 1125.7 | 1057.1 | Pass |
| NACL | 0000-279052-12 | 1125.1 | 1056.5 | Pass |
| NACL | 0000-279052-13 | 1123.2 | 1054.6 | Pass |
| NACL | 0000-279052-14 | 1125.0 | 1056.4 | Pass |
| NACL | 0000-279052-15 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-279052-16 | 1124.4 | 1055.8 | Pass |
| NACL | 0000-279052-17 | 1124.5 | 1055.9 | Pass |
| NACL | 0000-279052-18 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-280017-01 | 1116.6 | 1048.0 | Pass |
| NACL | 0000-280017-02 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-280017-03 | 912.1 | 844.3 | FAIL |
| NACL | 0000-280017-04 | 1123.4 | 1054.8 | Pass |
| NACL | 0000-280017-05 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-280017-06 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-280017-07 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-280017-08 | 1124.8 | 1056.2 | Pass |
| NACL | 0000-280017-09 | 1124.2 | 1055.6 | Pass |
| NACL | 0000-280017-10 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-280017-11 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-280017-12 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-280017-13 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-280017-14 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-280017-15 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-280017-16 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-280017-17 | 1124.3 | 1055.7 | Pass |
| NACL | 0000-280017-18 | 1121.3 | Bad Bag | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-281005-01 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-281005-02 | 1121.4 | 1052.8 | Pass |
| NACL | 0000-281005-03 | 1121.4 | 1052.8 | Pass |
| NACL | 0000-281005-04 | 1121.6 | 1053.0 | Pass |
| NACL | 0000-281005-05 | 1114.7 | 1046.1 | Pass |
| NACL | 0000-281005-06 | 1123.1 | 1054.5 | Pass |
| NACL | 0000-281005-07 | 1120.3 | 1051.7 | Pass |
| NACL | 0000-281005-08 | 1121.0 | 1052.4 | Pass |
| NACL | 0000-281005-09 | 1121.2 | 1052.6 | Pass |
| NACL | 0000-281005-10 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-281005-11 | 1121.2 | 1052.6 | Pass |
| NACL | 0000-281005-12 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-281005-13 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-281005-14 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-281005-15 | 1119.7 | 1051.1 | Pass |
| NACL | 0000-281005-16 | 1115.5 | 1046.9 | Pass |
| NACL | 0000-281005-17 | 1121.1 | 1052.5 | Pass |
| NACL | 0000-281005-18 | 1128.0 | 1059.4 | Pass |
| NACL | 0000-281006-01 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-281006-02 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-281006-03 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-281006-04 | 1123.1 | 1054.5 | Pass |
| NACL | 0000-281006-05 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-281006-06 | 1123.3 | 1054.7 | Pass |
| NACL | 0000-281006-07 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-281006-08 | 1123.2 | 1054.6 | Pass |
| NACL | 0000-281006-09 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-281006-10 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-281006-11 | 1117.2 | 1048.6 | Pass |
| NACL | 0000-281006-12 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-281006-13 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-281006-14 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-281006-15 | 1123.1 | 1054.5 | Pass |
| NACL | 0000-281006-16 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-281006-17 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-281006-18 | 1052.9 | 984.6 | FAIL |
| NACL | 0000-281007-01 | 1112.1 | 1043.5 | Pass |
| NACL | 0000-281007-02 | 1083.2 | 1014.7 | Pass |
| NACL | 0000-281007-03 | 1121.0 | 1052.4 | Pass |
| NACL | 0000-281007-04 | 1123.4 | 1054.8 | Pass |
| NACL | 0000-281007-05 | 1034.0 | 965.7 | FAIL |
| NACL | 0000-281007-06 | 1116.4 | 1047.8 | Pass |
| NACL | 0000-281007-07 | 1123.2 | 1054.6 | Pass |
| NACL | 0000-281007-08 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-281007-09 | 1063.9 | 995.5 | FAIL |
| NACL | 0000-281007-10 | 1121.3 | 1052.7 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-281007-11 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-281007-12 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-281007-13 | 1124.6 | 1056.0 | Pass |
| NACL | 0000-281007-14 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-281007-15 | 1117.8 | 1049.2 | Pass |
| NACL | 0000-281007-16 | 1122.8 | 1054.2 | Pass |
| NACL | 0000-281007-17 | 1124.5 | 1055.9 | Pass |
| NACL | 0000-281007-18 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-281009-01 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-281009-02 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-281009-03 | 1121.7 | 1053.1 | Pass |
| NACL | 0000-281009-04 | 1121.7 | 1053.1 | Pass |
| NACL | 0000-281009-05 | 1117.3 | 1048.7 | Pass |
| NACL | 0000-281009-06 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-281009-07 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-281009-08 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-281009-09 | 1116.4 | 1047.8 | Pass |
| NACL | 0000-281009-10 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-281009-11 | 1121.6 | 1053.0 | Pass |
| NACL | 0000-281009-12 | 1120.5 | 1051.9 | Pass |
| NACL | 0000-281009-13 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-281009-14 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-281009-15 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-281009-16 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-281009-17 | 1121.3 | 1052.7 | Pass |
| NACL | 0000-281009-18 | 1121.3 | 1052.7 | Pass |
| NACL | 0000-281011-01 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-281011-02 | 1121.3 | 1052.7 | Pass |
| NACL | 0000-281011-03 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-281011-04 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-281011-05 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-281011-06 | 1114.2 | 1045.6 | Pass |
| NACL | 0000-281011-07 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-281011-08 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-281011-09 | 1124.1 | 1055.5 | Pass |
| NACL | 0000-281011-10 | 1120.7 | 1052.1 | Pass |
| NACL | 0000-281011-11 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-281011-12 | 1121.3 | 1052.7 | Pass |
| NACL | 0000-281011-13 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-281011-14 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-281011-15 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-281011-16 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-281011-17 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-281011-18 | 1123.4 | 1054.8 | Pass |
| NACL | 0000-282004-01 | 1124.7 | 1056.1 | Pass |
| NACL | 0000-282004-02 | 1122.9 | 1054.3 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-282004-03 | 1120.2 | 1051.6 | Pass |
| NACL | 0000-282004-04 | 1123.3 | 1054.7 | Pass |
| NACL | 0000-282004-05 | 1116.8 | 1048.2 | Pass |
| NACL | 0000-282004-06 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-282004-07 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-282004-08 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-282004-09 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-282004-10 | 1124.9 | 1056.3 | Pass |
| NACL | 0000-282004-11 | 1125.4 | 1056.8 | Pass |
| NACL | 0000-282004-12 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-282004-13 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-282004-14 | 1125.2 | 1056.6 | Pass |
| NACL | 0000-282004-15 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-282004-16 | 1121.6 | 1053.0 | Pass |
| NACL | 0000-282004-17 | 1123.2 | 1054.6 | Pass |
| NACL | 0000-282004-18 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-282005-01 | 1122.8 | 1054.2 | Pass |
| NACL | 0000-282005-02 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-282005-03 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-282005-04 | 1121.6 | 1053.0 | Pass |
| NACL | 0000-282005-05 | 1125.0 | 1056.4 | Pass |
| NACL | 0000-282005-06 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-282005-07 | 1116.8 | 1048.2 | Pass |
| NACL | 0000-282005-08 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-282005-09 | 1125.2 | 1056.6 | Pass |
| NACL | 0000-282005-10 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-282005-11 | 1124.0 | 1055.4 | Pass |
| NACL | 0000-282005-12 | 1125.7 | 1057.1 | Pass |
| NACL | 0000-282005-13 | 1124.1 | 1055.5 | Pass |
| NACL | 0000-282005-14 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-282005-15 | 1124.0 | 1055.4 | Pass |
| NACL | 0000-282005-16 | 1122.8 | 1054.2 | Pass |
| NACL | 0000-282005-17 | 1121.4 | 1052.8 | Pass |
| NACL | 0000-282005-18 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-282007-01 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-282007-02 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-282007-03 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-282007-04 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-282007-05 | 1121.3 | 1052.7 | Pass |
| NACL | 0000-282007-06 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-282007-07 | 1115.9 | 1047.3 | Pass |
| NACL | 0000-282007-08 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-282007-09 | 1121.7 | 1053.1 | Pass |
| NACL | 0000-282007-10 | 1124.1 | 1055.5 | Pass |
| NACL | 0000-282007-11 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-282007-12 | 1122.5 | 1053.9 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-282007-13 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-282007-14 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-282007-15 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-282007-16 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-282007-17 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-282007-18 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-282008-01 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-282008-02 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-282008-03 | 1121.0 | 1052.4 | Pass |
| NACL | 0000-282008-04 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-282008-05 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-282008-06 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-282008-07 | 1121.7 | 1053.1 | Pass |
| NACL | 0000-282008-08 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-282008-09 | 1116.6 | 1048.0 | Pass |
| NACL | 0000-282008-10 | 1116.1 | 1047.5 | Pass |
| NACL | 0000-282008-11 | 1121.3 | 1052.7 | Pass |
| NACL | 0000-282008-12 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-282008-13 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-282008-14 | 1124.1 | 1055.5 | Pass |
| NACL | 0000-282008-15 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-282008-16 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-282008-17 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-282008-18 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-283027-01 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-283027-02 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-283027-03 | 1114.4 | 1045.8 | Pass |
| NACL | 0000-283027-04 | 1121.8 | 1053.2 | Pass |
| NACL | 0000-283027-05 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-283027-06 | 1122.8 | 1054.2 | Pass |
| NACL | 0000-283027-07 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-283027-08 | 1122.8 | 1054.2 | Pass |
| NACL | 0000-283027-09 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-283027-10 | 1121.3 | 1052.7 | Pass |
| NACL | 0000-283027-11 | 1125.1 | 1056.5 | Pass |
| NACL | 0000-283027-12 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-283027-13 | 1121.6 | 1053.0 | Pass |
| NACL | 0000-283027-14 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-283027-15 | 1103.4 | 1034.9 | Pass |
| NACL | 0000-283027-16 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-283027-17 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-283027-18 | 1125.8 | 1057.2 | Pass |
| NACL | 0000-283030-01 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-283030-02 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-283030-03 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-283030-04 | 1122.5 | 1053.9 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-283030-05 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-283030-06 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-283030-07 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-283030-08 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-283030-09 | 1123.3 | 1054.7 | Pass |
| NACL | 0000-283030-10 | 1117.6 | 1049.0 | Pass |
| NACL | 0000-283030-11 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-283030-12 | 1123.3 | 1054.7 | Pass |
| NACL | 0000-283030-13 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-283030-14 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-283030-15 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-283030-16 | 1123.7 | 1055.1 | Pass |
| NACL | 0000-283030-17 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-283030-18 | 1112.9 | 1044.3 | Pass |
| NACL | 0000-284002-01 | 1117.0 | 1048.4 | Pass |
| NACL | 0000-284002-02 | 1117.2 | 1048.6 | Pass |
| NACL | 0000-284002-03 | 1111.8 | 1043.2 | Pass |
| NACL | 0000-284002-04 | 1113.8 | 1045.2 | Pass |
| NACL | 0000-284002-05 | 1113.2 | 1044.6 | Pass |
| NACL | 0000-284002-06 | 1110.7 | 1042.1 | Pass |
| NACL | 0000-284002-07 | 1114.1 | 1045.5 | Pass |
| NACL | 0000-284002-08 | 1116.0 | 1047.4 | Pass |
| NACL | 0000-284002-09 | 1111.9 | 1043.3 | Pass |
| NACL | 0000-284002-10 | 1106.4 | 1037.9 | Pass |
| NACL | 0000-284002-11 | 1102.0 | 1033.5 | Pass |
| NACL | 0000-284002-12 | 1101.1 | 1032.6 | Pass |
| NACL | 0000-284002-13 | 1117.1 | 1048.5 | Pass |
| NACL | 0000-284002-14 | 1110.0 | 1041.5 | Pass |
| NACL | 0000-284002-15 | 1117.1 | 1048.5 | Pass |
| NACL | 0000-284002-17 | 1118.0 | 1049.4 | Pass |
| NACL | 0000-284002-18 | 636.1 | Bad Bag | Pass |
| NACL | 0000-285046-01 | 719.5 | 652.4 | FAIL |
| NACL | 0000-285046-02 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-285046-03 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-285046-04 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-285046-05 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-285046-06 | 1121.7 | 1053.1 | Pass |
| NACL | 0000-285046-07 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-285046-08 | 1120.3 | 1051.7 | Pass |
| NACL | 0000-285046-09 | 1123.4 | 1054.8 | Pass |
| NACL | 0000-285046-10 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-285046-11 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-285046-12 | 1121.2 | 1052.6 | Pass |
| NACL | 0000-285046-13 | 1120.7 | 1052.1 | Pass |
| NACL | 0000-285046-14 | 1115.2 | 1046.6 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-285046-15 | 1119.9 | 1051.3 | Pass |
| NACL | 0000-285046-16 | 1128.3 | 1059.7 | Pass |
| NACL | 0000-285046-17 | 1126.8 | 1058.2 | Pass |
| NACL | 0000-285046-18 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-285048-01 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-285048-01 | 1123.8 | 1055.2 | Pass |
| NACL | 0000-285048-02 | 1120.2 | 1051.6 | Pass |
| NACL | 0000-285048-02 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-285048-03 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-285048-03 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-285048-04 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-285048-04 | 1119.6 | 1051.0 | Pass |
| NACL | 0000-285048-05 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-285048-05 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-285048-06 | 1120.3 | 1051.7 | Pass |
| NACL | 0000-285048-06 | 1124.9 | 1056.3 | Pass |
| NACL | 0000-285048-07 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-285048-07 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-285048-08 | 1116.2 | 1047.6 | Pass |
| NACL | 0000-285048-08 | 1125.1 | 1056.5 | Pass |
| NACL | 0000-285048-09 | 1120.6 | 1052.0 | Pass |
| NACL | 0000-285048-09 | 1117.4 | 1048.8 | Pass |
| NACL | 0000-285048-10 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-285048-10 | 1127.0 | 1058.4 | Pass |
| NACL | 0000-285048-11 | 1122.2 | 1053.6 | Pass |
| NACL | 0000-285048-11 | 1122.8 | 1054.2 | Pass |
| NACL | 0000-285048-12 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-285048-12 | 1124.0 | 1055.4 | Pass |
| NACL | 0000-285048-13 | 1124.8 | 1056.2 | Pass |
| NACL | 0000-285048-13 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-285048-14 | 1122.4 | 1053.8 | Pass |
| NACL | 0000-285048-14 | 1124.7 | 1056.1 | Pass |
| NACL | 0000-285048-15 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-285048-15 | 1123.1 | 1054.5 | Pass |
| NACL | 0000-285048-16 | 1121.9 | 1053.3 | Pass |
| NACL | 0000-285048-16 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-285048-17 | 1123.6 | 1055.0 | Pass |
| NACL | 0000-285048-17 | 1123.3 | 1054.7 | Pass |
| NACL | 0000-285048-18 | 1121.1 | Bad Bag | Pass |
| NACL | 0000-285048-18 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-285049-01 | 1124.2 | 1055.6 | Pass |
| NACL | 0000-285049-02 | 1120.4 | 1051.8 | Pass |
| NACL | 0000-285049-03 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-285049-04 | 1119.4 | 1050.8 | Pass |
| NACL | 0000-285049-05 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-285049-06 | 1131.5 | 1062.9 | Pass |

Reflups Volume Army_2 Bags

12/7/93

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-285049-07 | 1121.5 | 1052.9 | Pass |
| NACL | 0000-285049-08 | 1120.9 | 1052.3 | Pass |
| NACL | 0000-285049-09 | 1088.1 | 1019.6 | Pass |
| NACL | 0000-285049-10 | 1087.9 | 1019.4 | Pass |
| NACL | 0000-285049-11 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-285049-12 | 1123.1 | 1054.5 | Pass |
| NACL | 0000-285049-13 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-285049-14 | 1123.9 | 1055.3 | Pass |
| NACL | 0000-285049-15 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-285049-16 | 1122.3 | 1053.7 | Pass |
| NACL | 0000-285049-17 | 1116.1 | 1047.5 | Pass |
| NACL | 0000-285049-18 | 1061.6 | Bad Bag | Pass |
| NACL | 0000-285050-01 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-285050-02 | 1121.7 | 1053.1 | Pass |
| NACL | 0000-285050-03 | 1120.6 | 1052.0 | Pass |
| NACL | 0000-285050-04 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-285050-05 | 1119.7 | 1051.1 | Pass |
| NACL | 0000-285050-06 | 1123.5 | 1054.9 | Pass |
| NACL | 0000-285050-07 | 1122.9 | 1054.3 | Pass |
| NACL | 0000-285050-08 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-285050-09 | 1122.1 | 1053.5 | Pass |
| NACL | 0000-285050-10 | 1121.4 | 1052.8 | Pass |
| NACL | 0000-285050-11 | 1123.1 | 1054.5 | Pass |
| NACL | 0000-285050-12 | 1124.5 | 1055.9 | Pass |
| NACL | 0000-285050-13 | 1124.4 | 1055.8 | Pass |
| NACL | 0000-285050-14 | 1116.3 | 1047.7 | Pass |
| NACL | 0000-285050-15 | 1123.0 | 1054.4 | Pass |
| NACL | 0000-285050-16 | 1125.1 | 1056.5 | Pass |
| NACL | 0000-285050-17 | 1122.5 | 1053.9 | Pass |
| NACL | 0000-285050-18 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-286011-01 | 1110.7 | 1042.1 | Pass |
| NACL | 0000-286011-02 | 1121.1 | 1052.5 | Pass |
| NACL | 0000-286011-03 | 1114.5 | 1045.9 | Pass |
| NACL | 0000-286011-04 | 1120.3 | 1051.7 | Pass |
| NACL | 0000-286011-05 | 1124.5 | 1055.9 | Pass |
| NACL | 0000-286011-06 | 1123.2 | 1054.6 | Pass |
| NACL | 0000-286011-07 | 1119.7 | 1051.1 | Pass |
| NACL | 0000-286011-08 | 1125.1 | 1056.5 | Pass |
| NACL | 0000-286011-09 | 1122.0 | 1053.4 | Pass |
| NACL | 0000-286011-10 | 1122.6 | 1054.0 | Pass |
| NACL | 0000-286011-11 | 1119.7 | 1051.1 | Pass |
| NACL | 0000-286011-12 | 1121.2 | 1052.6 | Pass |
| NACL | 0000-286011-13 | 1121.6 | 1053.0 | Pass |
| NACL | 0000-286011-14 | 1121.0 | 1052.4 | Pass |
| NACL | 0000-286011-15 | 1122.7 | 1054.1 | Pass |
| NACL | 0000-286011-16 | 1121.8 | 1053.2 | Pass |

Reflups Volume Army_2 Bags

12/7/93

| Bag and Overwrap Weight(g) | | 64.8 | Density(mg/l) | 1.0036 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| NACL | 0000-286011-17 | 1118.4 | 1049.8 | Pass |
| NACL | 0000-286011-18 | 1120.8 | 1052.2 | Pass |
| NACL | 0001-270035-01 | 1123.0 | 1054.4 | Pass |
| NACL | 0001-270035-02 | 1121.6 | 1053.0 | Pass |
| NACL | 0001-270035-03 | 1123.6 | 1055.0 | Pass |
| NACL | 0001-270035-04 | 1119.4 | 1050.8 | Pass |
| NACL | 0001-270035-05 | 1123.2 | 1054.6 | Pass |
| NACL | 0001-270035-06 | 938.8 | 870.9 | FAIL |
| NACL | 0001-270035-07 | 1124.0 | 1055.4 | Pass |
| NACL | 0001-270035-08 | 1122.8 | 1054.2 | Pass |
| NACL | 0001-270035-09 | 1122.9 | 1054.3 | Pass |
| NACL | 0001-270035-10 | 1123.1 | 1054.5 | Pass |
| NACL | 0001-270035-11 | 1122.8 | 1054.2 | Pass |
| NACL | 0001-270035-12 | 1124.5 | 1055.9 | Pass |
| NACL | 0001-270035-13 | 1123.2 | 1054.6 | Pass |
| NACL | 0001-270035-14 | 1123.1 | 1054.5 | Pass |
| NACL | 0001-270035-15 | 1123.5 | 1054.9 | Pass |
| NACL | 0001-270035-16 | 1122.3 | 1053.7 | Pass |
| NACL | 0001-270035-17 | 1122.6 | 1054.0 | Pass |
| NACL | 0001-270035-18 | 1122.1 | 1053.5 | Pass |
| NACL | 0001-278006-01 | 1122.3 | 1053.7 | Pass |
| NACL | 0001-278006-02 | 1122.7 | 1054.1 | Pass |
| NACL | 0001-278006-03 | 1122.5 | 1053.9 | Pass |
| NACL | 0001-278006-04 | 1122.9 | 1054.3 | Pass |
| NACL | 0001-278006-05 | 1122.9 | 1054.3 | Pass |
| NACL | 0001-278006-06 | 1124.1 | 1055.5 | Pass |
| NACL | 0001-278006-07 | 1122.1 | 1053.5 | Pass |
| NACL | 0001-278006-08 | 1123.3 | 1054.7 | Pass |
| NACL | 0001-278006-09 | 1123.0 | 1054.4 | Pass |
| NACL | 0001-278006-10 | 1123.1 | 1054.5 | Pass |
| NACL | 0001-278006-11 | 1124.0 | 1055.4 | Pass |
| NACL | 0001-278006-12 | 1123.9 | 1055.3 | Pass |
| NACL | 0001-278006-13 | 1122.8 | 1054.2 | Pass |
| NACL | 0001-278006-14 | 1121.4 | 1052.8 | Pass |
| NACL | 0001-278006-15 | 1122.0 | 1053.4 | Pass |
| NACL | 0001-278006-16 | 1123.6 | 1055.0 | Pass |
| NACL | 0001-278006-17 | 1122.8 | 1054.2 | Pass |
| NACL | 0001-278006-18 | 1120.4 | 1051.8 | Pass |
| Number | 773 | | Failures | 8 |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 103.8 | Density(mg/l) | 1.0045 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| RCW | 0000-271009-01 | 3195.4 | 3077.8 | Pass |
| RCW | 0000-271009-02 | 3213.3 | 3095.6 | Pass |
| RCW | 0000-271009-03 | 3217.1 | 3099.4 | Pass |
| RCW | 0000-271009-04 | 3216.8 | 3099.1 | Pass |
| RCW | 0000-271009-05 | 3212.4 | 3094.7 | Pass |
| RCW | 0000-271009-06 | 3213.1 | 3095.4 | Pass |
| RCW | 0000-271010-01 | 3202.3 | 3084.6 | Pass |
| RCW | 0000-271010-02 | 3221.0 | 3103.2 | Pass |
| RCW | 0000-271010-03 | 3184.9 | 3067.3 | Pass |
| RCW | 0000-271010-04 | 3207.9 | 3090.2 | Pass |
| RCW | 0000-271010-05 | 3214.3 | 3096.6 | Pass |
| RCW | 0000-271010-06 | 3213.4 | 3095.7 | Pass |
| RCW | 0000-271011-01 | 3209.7 | 3092.0 | Pass |
| RCW | 0000-271011-02 | 3208.8 | 3091.1 | Pass |
| RCW | 0000-271011-03 | 3200.6 | 3082.9 | Pass |
| RCW | 0000-271011-04 | 3160.0 | 3042.5 | Pass |
| RCW | 0000-271011-05 | 3209.8 | 3092.1 | Pass |
| RCW | 0000-271011-06 | 3209.4 | 3091.7 | Pass |
| RCW | 0000-272002-01 | 3192.6 | 3075.0 | Pass |
| RCW | 0000-272002-01 | 3211.4 | 3093.7 | Pass |
| RCW | 0000-272002-02 | 3215.4 | 3097.7 | Pass |
| RCW | 0000-272002-02 | 3221.6 | 3103.8 | Pass |
| RCW | 0000-272002-03 | 3213.3 | 3095.6 | Pass |
| RCW | 0000-272002-03 | 3213.0 | 3095.3 | Pass |
| RCW | 0000-272002-04 | 3214.3 | 3096.6 | Pass |
| RCW | 0000-272002-04 | 3208.1 | 3090.4 | Pass |
| RCW | 0000-272002-05 | 3214.1 | 3096.4 | Pass |
| RCW | 0000-272002-05 | 3756.9 | 3636.7 | FAIL |
| RCW | 0000-272002-06 | 3216.2 | 3098.5 | Pass |
| RCW | 0000-272002-06 | 3206.9 | 3089.2 | Pass |
| RCW | 0000-273001-01 | 3216.9 | 3099.2 | Pass |
| RCW | 0000-273001-02 | 3164.6 | 3047.1 | Pass |
| RCW | 0000-273001-03 | 3214.8 | 3097.1 | Pass |
| RCW | 0000-273001-04 | 3208.7 | 3091.0 | Pass |
| RCW | 0000-273001-05 | 3215.0 | 3097.3 | Pass |
| RCW | 0000-273001-06 | 3210.1 | 3092.4 | Pass |
| RCW | 0000-273003-01 | 2955.3 | 2838.7 | FAIL |
| RCW | 0000-273003-02 | 3209.8 | 3092.1 | Pass |
| RCW | 0000-273003-03 | 3186.0 | 3068.4 | Pass |
| RCW | 0000-273003-04 | 3286.1 | 3168.0 | Pass |
| RCW | 0000-273003-05 | 3211.0 | 3093.3 | Pass |
| RCW | 0000-273003-06 | 3213.2 | 3095.5 | Pass |
| RCW | 0000-274021-01 | 3212.6 | 3094.9 | Pass |
| RCW | 0000-274021-02 | 3212.7 | 3095.0 | Pass |
| RCW | 0000-274021-03 | 3214.1 | 3096.4 | Pass |
| RCW | 0000-274021-04 | 3209.4 | 3091.7 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 103.8 | Density(mg/l) | 1.0045 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| RCW | 0000-274021-05 | 3205.2 | 3087.5 | Pass |
| RCW | 0000-274021-06 | 3211.6 | 3093.9 | Pass |
| RCW | 0000-275024-01 | 3211.5 | 3093.8 | Pass |
| RCW | 0000-275024-02 | 3213.0 | 3095.3 | Pass |
| RCW | 0000-275024-03 | 3191.2 | 3073.6 | Pass |
| RCW | 0000-275024-04 | 3214.0 | 3096.3 | Pass |
| RCW | 0000-275024-05 | 3203.1 | 3085.4 | Pass |
| RCW | 0000-275024-06 | 3186.5 | 3068.9 | Pass |
| RCW | 0000-276024-01 | 3177.8 | 3060.2 | Pass |
| RCW | 0000-276024-02 | 3229.5 | 3111.7 | Pass |
| RCW | 0000-276024-03 | 3199.3 | 3081.6 | Pass |
| RCW | 0000-276024-04 | 3210.7 | 3093.0 | Pass |
| RCW | 0000-276024-05 | 3129.7 | 3012.3 | Pass |
| RCW | 0000-276024-06 | 3212.3 | 3094.6 | Pass |
| RCW | 0000-276026-01 | 3215.3 | 3097.6 | Pass |
| RCW | 0000-276026-02 | 3215.0 | 3097.3 | Pass |
| RCW | 0000-276026-03 | 3216.4 | 3098.7 | Pass |
| RCW | 0000-276026-04 | 3216.4 | 3098.7 | Pass |
| RCW | 0000-276026-05 | 3216.4 | 3098.7 | Pass |
| RCW | 0000-276026-06 | 3207.2 | 3089.5 | Pass |
| RCW | 0000-277014-01 | 3208.7 | 3091.0 | Pass |
| RCW | 0000-277014-02 | 3143.1 | 3025.7 | Pass |
| RCW | 0000-277014-03 | 3208.3 | 3090.6 | Pass |
| RCW | 0000-277014-04 | 3208.2 | 3090.5 | Pass |
| RCW | 0000-277014-05 | 3209.3 | 3091.6 | Pass |
| RCW | 0000-277014-06 | 3205.9 | 3088.2 | Pass |
| RCW | 0000-277016-01 | 3210.3 | 3092.6 | Pass |
| RCW | 0000-277016-02 | 3212.6 | 3094.9 | Pass |
| RCW | 0000-277016-03 | 3210.5 | 3092.8 | Pass |
| RCW | 0000-277016-04 | 3194.4 | 3076.8 | Pass |
| RCW | 0000-277016-05 | 3209.2 | 3091.5 | Pass |
| RCW | 0000-277016-06 | 3191.8 | 3074.2 | Pass |
| RCW | 0000-278001-01 | 3201.8 | 3084.1 | Pass |
| RCW | 0000-278001-01 | 3216.4 | 3098.7 | Pass |
| RCW | 0000-278001-02 | 3210.9 | 3093.2 | Pass |
| RCW | 0000-278001-02 | 3174.9 | 3057.3 | Pass |
| RCW | 0000-278001-03 | 3213.9 | 3096.2 | Pass |
| RCW | 0000-278001-03 | 3214.4 | 3096.7 | Pass |
| RCW | 0000-278001-04 | 3212.1 | 3094.4 | Pass |
| RCW | 0000-278001-04 | 3212.9 | 3095.2 | Pass |
| RCW | 0000-278001-05 | 3207.8 | 3090.1 | Pass |
| RCW | 0000-278001-05 | 3203.7 | 3086.0 | Pass |
| RCW | 0000-278001-06 | 3212.5 | 3094.8 | Pass |
| RCW | 0000-278001-06 | 3174.9 | 3057.3 | Pass |
| RCW | 0000-279027-01 | 3206.0 | 3088.3 | Pass |
| RCW | 0000-279027-02 | 3214.4 | 3096.7 | Pass |

Reflups Volume Army_2 Bags

| Bag and Overwrap Weight(g) | | 103.8 | Density(mg/l) | 1.0045 |
|----------------------------|----------------|-----------------------|---------------|--------|
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| RCW | 0000-279027-03 | 3203.4 | 3085.7 | Pass |
| RCW | 0000-279027-04 | 3196.8 | 3079.1 | Pass |
| RCW | 0000-279027-05 | 3212.3 | 3094.6 | Pass |
| RCW | 0000-279027-06 | 3211.8 | 3094.1 | Pass |
| RCW | 0000-279028-01 | 3209.9 | 3092.2 | Pass |
| RCW | 0000-279028-02 | 3213.8 | 3096.1 | Pass |
| RCW | 0000-279028-03 | 3214.1 | 3096.4 | Pass |
| RCW | 0000-279028-04 | 3207.2 | 3089.5 | Pass |
| RCW | 0000-279028-05 | 3199.3 | 3081.6 | Pass |
| RCW | 0000-279028-06 | 3154.9 | 3037.4 | Pass |
| RCW | 0000-279029-01 | 3213.2 | 3095.5 | Pass |
| RCW | 0000-279029-02 | 3213.5 | 3095.8 | Pass |
| RCW | 0000-279029-03 | 3199.0 | 3081.3 | Pass |
| RCW | 0000-279029-04 | 3213.5 | 3095.8 | Pass |
| RCW | 0000-279029-05 | 3209.4 | 3091.7 | Pass |
| RCW | 0000-279029-06 | 3203.9 | Bad Bag | Pass |
| RCW | 0000-280035-01 | 3212.1 | 3094.4 | Pass |
| RCW | 0000-280035-02 | 3215.1 | 3097.4 | Pass |
| RCW | 0000-280035-03 | 3219.9 | 3102.1 | Pass |
| RCW | 0000-280035-04 | 3211.8 | 3094.1 | Pass |
| RCW | 0000-280035-05 | 3215.8 | 3098.1 | Pass |
| RCW | 0000-280035-06 | 3216.1 | 3098.4 | Pass |
| RCW | 0000-280037-01 | 3201.9 | 3084.2 | Pass |
| RCW | 0000-280037-02 | 3213.1 | 3095.4 | Pass |
| RCW | 0000-280037-03 | 3217.4 | 3099.7 | Pass |
| RCW | 0000-280037-04 | 3215.0 | 3097.3 | Pass |
| RCW | 0000-280037-05 | 3216.0 | 3098.3 | Pass |
| RCW | 0000-280037-06 | 3217.8 | 3100.0 | Pass |
| RCW | 0000-281002-01 | 3216.7 | 3099.0 | Pass |
| RCW | 0000-281002-02 | 3214.2 | 3096.5 | Pass |
| RCW | 0000-281002-03 | 3208.7 | 3091.0 | Pass |
| RCW | 0000-281002-04 | 3218.5 | 3100.7 | Pass |
| RCW | 0000-281002-05 | 3218.0 | 3100.2 | Pass |
| RCW | 0000-281002-06 | 3214.4 | 3096.7 | Pass |
| RCW | 0000-282050-01 | 3185.2 | 3067.6 | Pass |
| RCW | 0000-282050-02 | 3216.2 | 3098.5 | Pass |
| RCW | 0000-282050-03 | 3214.3 | 3096.6 | Pass |
| RCW | 0000-282050-04 | 3217.9 | 3100.1 | Pass |
| RCW | 0000-282050-05 | 3211.7 | 3094.0 | Pass |
| RCW | 0000-282050-06 | 3216.6 | 3098.9 | Pass |
| RCW | 0000-283035-01 | 3219.2 | 3101.4 | Pass |
| RCW | 0000-283035-02 | 3210.8 | 3093.1 | Pass |
| RCW | 0000-283035-03 | 3219.7 | 3101.9 | Pass |
| RCW | 0000-283035-04 | 3215.4 | 3097.7 | Pass |
| RCW | 0000-283035-05 | 3218.2 | 3100.4 | Pass |
| RCW | 0000-283035-06 | 2593.7 | 2478.7 | FAIL |

Reflups Volume Army_2 Bags

12/7/93

| | | | | |
|----------------------------|----------------|-----------------------|---------------|--------|
| Bag and Overwrap Weight(g) | | 103.8 | Density(mg/l) | 1.0045 |
| Product | Bag Number | Gross Weight Grams | Volume ml | Status |
| RCW | 0000-283036-01 | 3124.3 | 3007.0 | Pass |
| RCW | 0000-283036-02 | 3215.7 | 3098.0 | Pass |
| RCW | 0000-283036-03 | 3170.9 | 3053.4 | Pass |
| RCW | 0000-283036-04 | 3219.7 | 3101.9 | Pass |
| RCW | 0000-283036-05 | 3209.1 | 3091.4 | Pass |
| RCW | 0000-283036-06 | 3207.9 | 3090.2 | Pass |
| RCW | 0000-284049-01 | 3213.1 | 3095.4 | Pass |
| RCW | 0000-284049-02 | 3222.4 | 3104.6 | Pass |
| RCW | 0000-284049-03 | 3192.3 | 3074.7 | Pass |
| RCW | 0000-284049-04 | 3209.7 | 3092.0 | Pass |
| RCW | 0000-284049-05 | 3205.4 | 3087.7 | Pass |
| RCW | 0000-284049-06 | 3207.3 | 3089.6 | Pass |
| RCW | 0000-284050-01 | 3212.0 | 3094.3 | Pass |
| RCW | 0000-284050-01 | 3155.2 | 3037.7 | Pass |
| RCW | 0000-284050-02 | 3212.4 | 3094.7 | Pass |
| RCW | 0000-284050-02 | 3216.9 | 3099.2 | Pass |
| RCW | 0000-284050-03 | 3212.2 | 3094.5 | Pass |
| RCW | 0000-284050-03 | 3200.1 | 3082.4 | Pass |
| RCW | 0000-284050-04 | 3206.0 | 3088.3 | Pass |
| RCW | 0000-284050-04 | 3204.9 | 3087.2 | Pass |
| RCW | 0000-284050-05 | 3213.2 | 3095.5 | Pass |
| RCW | 0000-284050-05 | 3197.1 | 3079.4 | Pass |
| RCW | 0000-284050-06 | 3214.2 | 3096.5 | Pass |
| RCW | 0000-284050-06 | 3209.5 | 3091.8 | Pass |
| RCW | 0000-284051-01 | 3152.1 | 3034.6 | Pass |
| RCW | 0000-284051-02 | 3196.2 | 3078.5 | Pass |
| RCW | 0000-284051-03 | 3208.4 | 3090.7 | Pass |
| RCW | 0000-284051-04 | 3203.8 | 3086.1 | Pass |
| RCW | 0000-284051-05 | 3207.5 | 3089.8 | Pass |
| RCW | 0000-284051-06 | 3216.8 | 3099.1 | Pass |
| RCW | 0000-285016-01 | 2607.5 | 2492.5 | FAIL |
| RCW | 0000-285016-02 | 3206.5 | 3088.8 | Pass |
| RCW | 0000-285016-03 | 3215.5 | 3097.8 | Pass |
| RCW | 0000-285016-04 | 3204.3 | 3086.6 | Pass |
| RCW | 0000-285016-05 | 3207.0 | 3089.3 | Pass |
| RCW | 0000-285016-06 | 3208.3 | 3090.6 | Pass |
| Number | 174 | Failures | 4 | |