Army Research Laboratory



A Review of Laser Photofragmentation/Fragment Detection Techniques for Chemical Analysis of the Gas Phase

by Rosario C. Sausa and Joseph B. Simeonsson

ARL-TR-2304

September 2000

Approved for public release; distribution is unlimited.

DTIC QUALITY INSCHOOLD 4



The findings in this report are not to be construed as an official Department of the Army position unless so designated by other authorized documents.

Citation of manufacturer's or trade names does not constitute an official endorsement or approval of the use thereof.

Destroy this report when it is no longer needed. Do not return it to the originator.

Army Research Laboratory

Aberdeen Proving Ground, MD 21005-5066

ARL-TR-2304

September 2000

A Review of Laser Photofragmentation/Fragment Detection Techniques for Chemical Analysis of the Gas Phase

Rosario C. Sausa Weapons and Materials Research Directorate, ARL

Joseph B. Simeonsson Department of Chemistry, University of Iowa

Approved for public release; distribution is unlimited.

Abstract

Laser photofragmentation/fragment detection techniques utilize one or multiple lasers for photolysis of the analyte species and spectroscopic detection of the characteristic photofragments. Fragment detection techniques include laser-induced fluorescence, photoionization, prompt emission, stimulated emission, and laser ionization recombination emission spectroscopies. Applications for chemical analysis of the gas phase have been performed by the detection of atomic and molecular fragments. Recent reports on photofragmentation/fragment detection approaches are presented, and the outlook for future applications is discussed.

Acknowledgments

We thank our numerous colleagues who provided us with reprints/preprints of their work. R. C. Sausa also thanks Dr. V. Swayanbunathan for helpful discussions. J. B. Simeonsson gratefully acknowledges support from the University of Iowa through the Old Gold Fellowship program.

INTENTIONALLY LEFT BLANK.

.

·

Table of Contents

	Acknowledgments	iii
	List of Figures	vii
	List of Tables	vii
1.	Introduction	1
2.	PF/FD Approach	1
3.	PF/FD Applications Via the Detection of Atomic Fragments	5
4.	PF/FD Applications Via the Detection of Molecular Fragments	7
4.1	NO/NO2/NOx	8
4.2	NH3	11
4.3	H_2SO_4 , SO_2	11
4.4	HO ₂	11
4.5	Nitrocompounds Including Energetic Materials	12
5.	Future Outlook for PF/FD Applications	13
6.	References	15
	Glossary	19
	Distribution List	21
	Report Documentation Page	23

.

INTENTIONALLY LEFT BLANK.

List of Figures

<u>Figure</u>		Page Page
1.	Schematic Diagram Illustrating Several PF/FD Processes, Including LIF, SE, LIRE, PE, PPI, and PI Fragment Detection.	3
2.	Schematic Diagram Showing the PF/FD Approach for Detecting NO ₂ Through NO Fragment Detection Via LIF and REMPI Processes	9

List of Tables

<u>Table</u>		Page
1.	PF/FD Chemical Analysis by Atomic Fragment Detection	14
2.	PF/FD Chemical Analysis by Molecular Fragment Detection	14

INTENTIONALLY LEFT BLANK.

1. Introduction

Laser-based spectrometric techniques are useful for chemical analysis because they offer the combination of high sensitivity and selectivity with real-time monitoring capabilities. There continues to be interest in the development of laser-based photofragmentation/fragment detection (PF/FD) techniques for gaseous analytes due to the fact that laser technology is rapidly maturing and is often capable of being implemented outside of the research laboratory. Also, there is a constant need for ever higher levels of analytical performance in terms of sensitivity, selectivity, and temporal response. The purpose of this article is to review the status of spectrochemical techniques based on PF/FD methods. The PF/FD approach is first discussed in terms of its analytical utility and the strategies that can be implemented for fragment detection. Early reports viewed as important to the development of current techniques are noted, as are previous reviews of this field of PF/FD research. Presented are recent applications of PF/FD techniques for chemical analysis that include the detection of analytes by characteristic atomic fragments or molecular fragments. Finally, the results of the review are summarized and areas of future applications are identified. A glossary of useful abbreviations and terms is also included at the end of the review.

2. PF/FD Approach

The PF/FD approach is most often utilized when the analyte molecule does not lend itself to direct spectroscopic detection. In general, atoms and small molecules (e.g., diatomic and triatomic species) can be detected directly by absorption, fluorescence, or photoionization spectroscopy techniques due to a favorable combination of usually strong optical transitions and sharp, well-resolved spectral features. However, it is often the case for larger molecules that the transitions are weaker and the spectral features are broad and poorly defined, in which case direct detection of the molecule by the above methods is not analytically practical. While direct detection may not be feasible, the photofragmentation products of polyatomic molecules are often monatomic and diatomic species that are readily detected. Since the fragments are characteristic of the chemical composition of the precursor (analyte), they also contribute to the selectivity of the method. It is

important to note that the PF process has a spectral dependence which can be exploited to increase the selectivity of PF/FD approaches. One important feature of the PF/FD method is its utility for the detection of classes of compounds. When molecules share a common functionality (e.g., -Cl, -Br, -I, $-NO_2$, $-SO_2/-SO_4$), the functionality may be targeted for fragmentation and detection. Optimization of the technique for the detection of a characteristic fragment allows a class of compounds to be detected by a single spectroscopic approach.

A schematic diagram showing the basis of the PF/FD method is presented in Figure 1. The analyte or precursor molecule, R-F, is initially excited by a single or multiphoton absorption process using laser radiation, hv_1 , resonant with an absorption feature of the analyte molecule. Following excitation, fragmentation occurs and yields the characteristic fragment, F, with internal energy, E(j,v,e), and its companion radical, R. The letters j, v, and e denote the fragment's internal energy due to rotational, vibrational, and electronic excitation. A second laser source, hv_2 , which may be time-delayed from the first, can then be used to probe the fragment distribution by laser-induced fluorescence (LIF), photoionization (PI), laser ionization recombination emission (LIRE), or stimulated emission (SE). In some cases, the photofragment population may be produced with sufficient electronic excitation so that it either emits spontaneously or absorbs additional photons, resulting in prompt photoionization (PPI). The former process is sometimes referred to as prompt emission (PE) or fragment fluorescence (FF) and should be distinguished from LIF following the photofragmentation process. A special case of the PF/FD method occurs when the condition $hv_1 = hv_2$ is satisfied. In this case, one laser can be used for both photofragmentation of the precursor molecules and excitation of the characteristic fragments.

Spectrochemical measurements using each of the above-mentioned detection techniques have been reported. The relative effectiveness of each approach depends on the analyte (precursor) and fragment species in question. Due to the high laser intensity required for the photofragmentation process, the photoproduction of interferences is a key consideration in selecting the optimum method of fragment detection. Although PF/LIF has seen the widest application and perhaps the highest degree of success, it should be pointed out that the combination of laser photofragmentation followed by PI detection of the fragment with mass spectrometric detection has perhaps the

2



Figure 1. Schematic Diagram Illustrating Several PF/FD Processes, Including LIF, SE, LIRE, PE, PPI, and PI Fragment Detection.

greatest potential in terms of analytical selectivity and sensitivity, as discussed by Antonov and Letokhov [1, 2] and Letokhov [3]. Perhaps the only limitation to PF/PI approaches is the practicality outside the research laboratory.

It is worth noting early reports and reviews, which have been significant in the development of current applications of PF/FD methods. Perhaps the earliest suggestion of PF/FD as an analytical methodology was by Hurst et al. in 1977 [4]. In concluding remarks to their paper, "A Demonstration of One-Atom Detection," the authors identified analytical applications for a resonance ionization spectroscopy approach that included the analysis of "molecular substances, e.g., the classification of a compound by dissociation and subsequent identification of resultant atoms (fragments)."

Since that time, there have been several demonstrations of the feasibility and merits of identifying molecular substances by PF/FD techniques. In 1978, Antonov et al. [5] reported on the detection of several compounds by photoionization and PF/PI techniques, including NO₂, benzaldehyde, benzophenone, nitrobenzene, and *para-* and *ortho-*nitrotoluenes. Antonov and Letokhov [1] later reviewed the use of multiphoton fragmentation and multiphoton ionization techniques for chemical detection. Included in the review was the proposal of a laser selective molecular detector, which incorporated lasers for optical excitation and photoionization multistep scheme where resonant absorption processes proceeding through vibrational and then electronic intermediate states would be most useful due to the high selectivity that could be achieved.

In 1979, Halpern et al. [6] reported on the study of an ArF laser PF/PE technique used for detecting NH_3 molecules by the emission of NH photofragments. The authors noted the potential for related PF/FD techniques in detecting species such as NO_2 and NOx, making it the first such suggestion for utilizing PF/FD techniques to monitor atmospheric analytes. Then in 1980, Rodgers et al. [7] published a paper entitled, "Photofragmentation-Laser Induced Fluorescence: A New Method for Detecting Atmospheric Gases," in which PF/LIF was presented as a general strategy for detecting a variety of important atmospheric species. Although no experimental results were

4

presented, the analytical capabilities were estimated using nominal values for important experimental parameters. Also, they indicated the feasibility of PF/LIF for measuring NO_2 , NO_3 , HNO_2 and laid the foundation for a series of reports on this approach for atmospheric measurements.

Laser multiphoton ionization and fragmentation methods have been reviewed previously by Syage and Wessel [8]. Although the review emphasized PF/FD applications to physical studies, the authors noted an increase in the number of analytical reports and the strong analytical potential of these approaches. Jinkins et al. [9] published a review of PF/FD fluorescence techniques that includes a discussion of the merits of fragment fluorescence spectrometry and a presentation of examples of PF/PE and PF/LIF studies of nonfluorescent species. Important features of the PF/PE and PF/LIF approaches were noted and included the following: (1) their applicability to fluorescent and nonfluorescent molecular analytes; (2) the complementary nature of the information obtained by PF/PE and PF/LIF (regarding neutral fragments), as compared to information by mass spectrometry (regarding ionic fragments); and (3) the "compound class" selectivity obtained by monitoring photofragmented functional groups.

Most recently, Simeonsson and Sausa have reviewed the analytical applications of PF/FD techniques through 1996 [10]. Included in that review are applications of the PF/FD method to the detection of gas phase species by PE, LIF, SE, LIRE, and PI fragment detection. This review is more limited, reporting primarily on applications of the PF/FD method that have appeared since the former review. It is intended to demonstrate the current capabilities of PF/FD techniques and to indicate where future applications are likely to occur in this growing field. For a more comprehensive review of earlier reports, see Simeonsson and Sausa [10].

3. PF/FD Applications Via the Detection of Atomic Fragments

The detection of trace metals in high temperature environments is important both for regulating toxic metal emissions and for controlling emissions of alkali species that are corrosive and result in the degradation of combustion systems. The measurement of these species represents a major area

of application of PF/FD techniques. The analytical utility of the PF/PE approach for detecting gas phase salts of Na and K was initially reported in 1986 by Oldenborg and Baughcum [11], who used an ArF excimer laser (with and without stimulated Raman shifting) as the ultraviolet laser source. Recently, this approach has received renewed attention from several groups. Chadwick et al. [12] have explored the utility of a PF/PE approach for selectively by measuring NaOH in the gas phase in the presence of other Na species. The approach is based on a two photon excited PF/PE where the NaOH molecule absorbs two 355-nm photons from a frequency tripled Nd:YAG laser. This excitation scheme is convenient since it takes advantage of the mature laser technology. Furthermore, the technique selectively produces prompt emission from Na fragments from NaOH, but not those from NaCl. The limit of detection (LOD) for this technique is estimated to be 0.5 parts per million (ppm), and while it is not as sensitive as the use of 193-nm laser radiation, it is selective for NaOH and demonstrates how spectral selectivity of the PF process can be used to distinguish chemical species that contain a common fragment.

Greger et al. [13] have reported on the PF/PE technique for in-situ alkali concentration measurements in a pressurized, fluidized bed reactor. An excimer laser operating at 193 nm photodissociates NaCl and KCl, producing excited Na(3 2 P) and K(4 2 P) atoms whose emissions are simultaneously detected in two separate channels at 589 and 768 nm. Studies on lignite and hard coal under several conditions were performed, and an LOD of <0.2 parts per billion (ppb) was reported for both elements. The dynamic range for the measurements exceeds three orders of magnitude.

Buckley et al. [14] have extended the PF/FD technique to the detection of hazardous metal species in postcombustion gases of a methane/air burner-stabilized flame and reactor [15]. Various metal chloride aerosols were injected in the burner or reactor and exited with 193-nm radiation. Prompt emission from Ba, Cr, Mn, Ni, Pb, and Tl atoms were monitored. LODs in the sub-ppm range have been reported for single- and 100-shot averaging, with negligible spectral interferences between the metal species or background interferences from the combustion products. The highest sensitivity was observed for Tl, with an LOD of 0.5 ppb for 100-shot averaging. The authors reported that the signal changed linearly with concentration for each metal compound injected in the

postflame gases, and in some cases changed with aerosol size and surface area. This latter variation makes the technique a strong function of temperature in the region of the postflame zone where condensation and nucleation are important. The authors show that LODs in the ppb level are readily obtained for the above elements at moderate laser power, despite considerable atmospheric quenching. They indicate that this method is likely to be capable of meeting proposed regulations for metal emissions of combustion devices.

In the PF/PE technique, the major competition for fluorescence is quenching. Depending on the gas composition and temperature, the effective lifetimes for typical metals transitions at atmospheric pressure are one to two orders of magnitude less than their radiative lifetimes. As a result, the fluorescence quantum efficiency Y, Y = A/(A + Q), where A is the spontaneous emission rate and Q is the total quenching rate, is in the range of 1–10% at 1 atm. Hartinger et al. [16] have measured the quenching of fluorescence from Na(3 ²P) and K (4 ²P) atoms by various collision partners at 973 and 1,273 K following photodissociation of NaCl and KCl at 193 nm. They report quenching crosssections for CO₂, O₂, and N₂ (10–60 Å²) and Ar (<1 Å²), and calculate an effective fluorescence quantum efficiency for Na (3 ²P) in the range of 5–30% at 1 atm using a radiative lifetime of 16 ns and an effective lifetime in the range of 0.8–0.5 ns.

The above works demonstrate that PF/FE is a viable diagnostic for real-time and in-situ monitoring of industrial combustion systems. It is sensitive, nonintrusive, and fast compared to standard sampling methods that require considerably longer measurement times and bear the risk of the sample reacting in the sampling lines. Furthermore, the PF/PE technique is readily implemented due to its simplicity; it only requires a single laser and simple detection optics.

4. PF/FD Applications Via the Detection of Molecular Fragments

PF/FD techniques have been used in a wide range of applications by their characteristic molecular photofragments. Major areas of application have included the measurement of reactive trace gases in the atmosphere, the measurement of halogenated organics in combustion

environments, and the detection of energetic materials (nitrocompounds). For an extensive review of earlier applications, see Simeonsson and Sausa [10].

4.1 NO/NO₂/NOx. Several groups have reported PF/FD techniques for nitrogen oxides. Peng et al. [17] have explored the utility of two PF-PI schemes for detection of NOx (NOx = NO + NO₂). At 226 nm, NO is photoionized by a 1 + 1 process via the A-X (0,0) transition. NO₂ is photolyzed efficiently at this wavelength, and the resulting NO fragments can be detected by the above 1 + 1 PI approach (see Figure 2). The single shot detection limit for NO is reported as 2–3 ppb at 226 nm. The sensitivity for NO₂ is lower and the LOD is estimated to be 10–15 ppb. Alternatively, a 2 + 1 approach at 384 nm can be used to detect NO, for which the LOD is also about 2–3 ppb. Signal averaging is estimated to improve the LOD for the approach to near 0.2 ppb.

Recently, Ledingham et al. [18] explored the use of picosecond (ps) and femtosecond (fs) laser pulses to bypass the dissociative state of NO₂ by fast up-pumping to the NO₂ ion state, thereby increasing the parent ion signal compared to the NO fragment signal. The multiphoton ionization and dissociation of NO₂ is performed at 248 and 496 nm using pulse widths ranging from 15 ns to 300 fs. At 248, the NO₂/NO ion ratio varied from ~1% using 15-ns pulses to about 60% using 300-fs pulses. By comparison, the ion ratios at 496 nm are similar for both 5-ps and 500-fs pulses, but very much reduced from those after photodissociation at 248 nm. The authors interpret their results using a simple rate equation model and conclude that the dissociation rate at 248 nm increases with increasing pulse width for the NO₂ -> NO + O(¹D), which opens at 40126 cm⁻¹ (249 nm). More recently, the same group studied the 375-nm multiphoton ionization and dissociation of NO₂ with 10-ns and 50-fs laser pulses [19]. In all spectra, the main peak observed is due to the ionization of the fragment NO, which is formed from the 375-nm dissociation of NO₂. The parent NO₂ ion peak is not observed with nanosecond pulses, but is observed with femtosecond pulses with a NO₂/NO ion ratio of ~10%. These results suggest that selective photoionization of NO and NO₂ species is possible using picosecond and femtosecond laser sources.

Sandholm et al. [20] have reported on the analytical capabilities of PF-LIF techniques for measuring atmospheric NO₂ and NOy (NOy = NO + NO₂ + all other sources of NO).



Figure 2. Schematic Diagram Showing the PF/FD Approach for Detecting NO₂ Through NO Fragment Detection Via LIF and REMPI Processes.

The techniques are based on the detection of the NO, which is either photofragmented or produced chemically in a catalytic converter. A recent report describes the latest developments in the detection of NO by LIF, which is based on excitation of NO via the A-X (0,0) transition at 226 nm

and D-A (0,0) near 1,100 nm. This results in fluorescence at the D-X (0,0) transition near 190 nm [20]. Due to the low fluorescence wavelength, the technique is signal limited, resulting in excellent signal-to-noise characteristics and LODs near 1 pptr for NO and other analytes. Instruments based on this detection scheme have been used for aircraft-based measurements of tropospheric NO, NO₂, and NOy over Alaska, Canada [21, 22], and the Pacific Ocean [23].

Simeonsson and Sausa [24] have reported the trace analysis of NO₂ in the presence of NO by PF/PI at visible wavelengths. NO and NO₂ are differentiated by using laser-visible radiation; NO is detected near 452 nm by (2 + 2) REMPI via its A-X (0,0) transitions, and NO₂ is detected by laser photofragmentation with subsequent fragment NO ionization by means of its A-X (0,0) and (1,1) transitions. Spectral differentiation is possible because the internal energy of the NO photofragment differs from that of "ambient" NO. The LODs for NO and NO₂ are in the 20–40 ppbv range at 449.2, 450.7, and 452.6 nm for a 10-s integration time, and compare favorably with the sensitivity achieved previously for NO₂ at 226 nm. The LOD of NO₂ at 517.5 nm is 75 ppbv.

Simeonsson et al. [25] are exploring techniques for detecting NO by LIF, and NO₂ and NOx by PF/LIF. Studies using excitation via the A-X (1,0) band near 215 nm, and fluorescence detection via the A-X (0,1) band at 237 nm demonstrate the high sensitivity of this approach. LODs for NO and NO₂ at atmospheric pressure are 0.03 ppb and 5 ppb. Studies are being performed to evaluate the utility of this approach for other nitrogen oxides and the potential for interferences from species such as O_2 , O_3 , SO_2 , and H_2O .

Swayambunathan and Sausa [26] have explored the use of a two-color PF/FD scheme for detecting NO₂. An Nd: YAG laser operating at 355 nm was used to both excite NO₂ and to pump a dye laser whose output is frequency doubled to produce 226-nm radiation. The NO fragment is detected by LIF using its A-X (0,0) transitions. It was found that the sensitivity approximately doubled from 200 to 100 ppb when using both 355- and 226-nm radiation, compared to using only 226-nm radiation. The increase in sensitivity is attributed to the generation of more NO from the dissociation of NO₂ at 355 nm.

4.2 NH₃. The detection of atmospheric NH₃ is an important application of PF/LIF. Studies by Schendel et al. [27] and Williams et al. [28] at Georgia Tech have shown in the laboratory and in the field that a PF/LIF approach based on the PF of NH₃ at 193 nm with subsequent LIF of the NH photofragments (excitation at 450 nm, detection at 325 nm) is highly effective for measurements in the open atmosphere [27, 28]. The spectral selectivity of this technique is such that isotopic tracer studies can be performed using a radio labeled ¹⁵NH₃. Intercomparison studies have also demonstrated the superiority of the PF/LIF approach over any other method in terms of accuracy, precision, and response time [28]. A version of this instrument was flown as part of the NSF-ACE I program, although the results have not been published [29].

4.3 H_2SO_4 , SO_2 . Sulfur compounds play a critical role in atmospheric chemistry and the measurement of sulfur species by PF techniques has been reported. Omenetto et al. [30] and Omenetto and Panne [31] have reported on a PF/PE technique for the detection of H_2SO_4 aerosol droplets and SO₂ using an ArF excimer laser operating at 193 nm. While irradiation of the H_2SO_4 aerosols produced broad band emission spectra, irradiation of SO₂ resulted in SO₂ emissions that were superimposed with emissions attributed to SO photofragments. The SO emissions are believed to result from SO photofragments that are subsequently excited by ArF laser radiation. An alternative PF/PD approach for measuring H_2SO_4 droplets based on their reaction with NaOH and emission-volatilized Na atoms is also being explored [32].

4.4 HO₂. The hydroperoxyl radical (HO₂) is important in both atmospheric and low temperature combustion environments. Hynes et al. [33] have developed a two-color PF/LIF technique for detecting HO₂. HO₂ is photolyzed at 212 nm by the fifth harmonic output of an Nd:YAG laser, generating the O and OH fragments. The OH is then detected by LIF by means of its A-X (1,0) band near 282 nm. The OH LIF signal is directly proportional to the HO₂ concentration because both the photolysis and probe pulses are generated simultaneously or slightly delayed by the same laser system. A detection limit of 2×10^{11} cm⁻³ has been reported in 300 Torr of N₂, indicating that the approach has the sensitivity and temporal resolution to be useful for kinetic studies of HO₂ under atmospheric conditions.

11

4.5 Nitrocompounds Including Energetic Materials. Various nitro-containing organic compounds, including energetic materials, have been measured using PF/FD. The multiphoton ionization and dissociation of nitrometane using nanosecond and femtosecond laser pulses at 375 and 750 nm was investigated by Kilic et al. [34]. The nanosecond time-of-flight spectra revealed only one peak at m/e = 30, corresponding to the NO photofragment. But the femtosecond spectra revealed a number of peaks, the prominent ones attributable to the parent ion, NO₂, NO, and CH₃. The spectra recorded using 90 fs of 750-nm radiation were similar to those recorded using 90-fs of 375-nm radiation, but with the parent peak considerably reduced.

Recently, the same group extended their nanosecond and picosecond laser studies to the ionization and photodissociation of nitrobenzene and o-, m-, and p-nitrotoluene isomers at 375 nm [35]. As in the nitromethane studies, these studies show that the parent molecule can be ionized with 90 fs pulses with great efficiency by bypassing its dissociative, molecular-excited states. This enhances the analytical utility for nitrocompound detection by minimizing parent photofragmentation and increasing parent ionization. Furthermore, the mass spectra of the nitrotoluene isomers show significant differences, enhancing their selectivity.

Wu et al. [36] report the detection of TNT by PF/LIF as part of an effort to develop a laser-based optical sensor for monitoring soil contaminated with energetic materials. A 226-nm laser is used to both photofragment the target compound and detect the characteristic NO fragment by LIF via its A-X (0,0) transitions. The concentration of the target compound is inferred from the intensity of the NO fluorescence. The effects of temperature and pressure on the PF/LIF signal were studied, and it was found that the signal significantly increased when the sample was heated above 343 K. Physical and chemical changes of the sample were noted at these temperatures. A detection limit of 4 ppm is reported for TNT in soil, showing that the PF/LIF holds great promise as a detection technique for energetic materials in soil.

Recently, Swayambunathan and Sausa [26] applied the PF/FD technique for the detection of RDX, PETN, and TNT. The technique employed a single laser operating at 226 nm and low pulse energies (100 μ J) for both photofragmenting the parent molecule and detecting the NO fragment by

12

REMPI or LIF. The REMPI technique appears to be more useful for ambient studies because the ion signal can be measured without heating the sample to high temperatures where decomposition can occur, as observed in the case of RDX and PETN. The REMPI LODs (S/N=3) are 70 ppm for TNT, 2 ppm for PETN, and 7 ppm for RDX.

5. Future Outlook for PF/FD Applications

As the above reports indicate, the majority of applications of PF/FD techniques have been in two areas: (1) the measurement of trace atmospheric species, especially nitrogen oxides and NH₃; and (2) the measurement of volatile alkali and heavy metal compounds in combustion exhaust gases. Other important areas of application include the measurement of chlorinated compounds in combustion exhaust gases and the development of PF/LIF and PF/PI approaches for measuring nitrocompounds (R-NO₂) with the goal of detecting energetic materials through in situ detection of their vapors. A summary of the PF/FD reports described in this review is given in Tables 1 and 2. A common theme for these efforts is the strong need for analytical approaches that allow in situ measurements and provide high sensitivity and selectivity on a real-time or near real-time basis. As more analytical chemistry measurements are pursued in situ, there will be even greater motivation for developing laser-based PF/FD approaches.

PF/FD sensors have already been demonstrated for measurements of NO₂/NOx/NOy, HNO₃, HONO, and NH₃, and it is expected that these techniques will continue to be developed and utilized for these same species. It is also expected that PF/FD approaches will be extended to other atmospheric species such as CH₃I, OI, CH₃Br, and dimethylsulfide (DMS). Furthermore, it is anticipated that PF/FD approaches will be useful for monitoring metal containing species and halogenated species in combustion environments and will continue to be explored for the detection of energetic materials and explosives. Continued advances in solid-state laser technology, combined with the fact that PF/FD approaches can be carried out remotely through optical fibers, indicates a strong potential for sensing and monitoring applications, especially in difficult or hostile environments. The high sensitivity and selectivity of PF/FD approaches is certain to lead to future growth in the number of applications of these powerful analytical techniques.

F	R-F	Method	λ _[PF] (nm)	λ _[FD] (nm)	Transition	LOD (ppb)	Ref.
Na	NaOH	PF/PE	193	589	$3p^{2}P^{\circ} - 3s^{2}S$	500	[12]
Na	NaCl	PF/PE	193	589	$3p^{2}P^{\circ} - 3s^{2}S$	<0.2	[13]
K	KC1	PF/PE	193	768	$4p^{2}P^{\circ} - 4s^{2}S$	<0.2	[13]
Ba	BaCl ₂	PF/PE	193	455.4	$6p^{1} {}^{2}P_{3/2} - 6s^{1} {}^{2}S_{1/2}$	30	[14, 15]
	2			493.4	$6p^{1} {}^{2}P_{1/2} - 6s^{2} {}^{2}S_{1/2}$		
				553.6	$6p^{1} P_{1} - 6s^{2} S_{0}$		
Cr	CrCl ₂	PF/PE	193	357.8	NI	300	[14, 15]
Mn	MnCl ₂	PF/PE	193	403.0	$3d^5 4s^1 4p^1 z {}^6P - 3d^5 4s^2 a {}^6S$	200	[14, 15]
Ni	NiCl ₂	PF/PE	193	300.3	NI	150	[14, 15]
Pb	PbCl ₂	PF/PE	193	405.8	${}^{3}P_{1}^{o} - {}^{3}P_{2}$	15	[14, 15]
T 1	TlCl ₃	PF/PE	193	377.5	${}^{2}S_{1/2} - {}^{2}P^{o}_{1/2}$	1.5	[14, 15]

Table 1. PF/FD Chemical Analysis by Atomic Fragment Detection

Table 2. PF/FD Chemical Analysis by Molecular Fragment Detection

F	R-F	Method	λ _[PF]	$\lambda_{[FD]}$	Transition	$\lambda_{[FL]}$ (nm)	LOD	Ref.
	NO	DE/DEMDI	226	226	$A - \mathbf{X}(0, 0)$		10-15 ppb	[17]
	INO_2	rr/KEIVIPI	450 7	450.7	A-X(0,0)		25 ppb	[24]
			452.6	452.6	A-X(0.0)		15 ppb	[24]
			517.5	517.5	A-X(0,3)		75 ppb	[24]
			226	226	A-X(0,0)		22 ppb	[37]
NO	NO ₂	PF/LIF	350-420	226/1100	A-X(0,0)/D-A(0,0)	190	6 ppt	[22]
			215	215	A-X (1,0)	237	5 ppb	[25]
NH	NH ₃	PF/LIF	193	450	A-X	325	4-10 ppt	[27]
so	SO ₂	PF/PE	193	913	NI		NI	[30–32]
	H ₂ SO ₄							
OH	OH ₂	PF/LIF	212	212	A-X(1,0)	308	$2 \times 10^{11} \text{ cm}^{-3}$	[33]
NO	TNT	PF/PI	226	226	A-X(0,0)		70 ppb	[26]
	RDX	PF/PI	226	226	A-X(0,0)		7 ppb	[26]
	PETN	PF/PI	226	226	A-X(0,0)		2 ppb	[26]
NO	TNT	PF/LIF	226	226	A-X(0,0)	226-260	37 ppb	[26]

Note: NI = not indicated.

6. References

- 1. Antonov, V. S., and V. S. Letokhov. Applied Physics B. Vol. 24, pp. 89-106, 1981.
- 2. Antonov, V. S., and V. S. Letokhov. Laser Analytical Spectrochemistry. Edited by V. S. Letokhov, Adam, Hilger, Bristol, 1985.
- 3. Letokhov, V. S. Laser Photoionization Spectroscopy. Orlando, FL: Academic Press, 1987.
- 4. Hurst, G. S., M. H. Nayfeh, and J. P. Young. Applied Physics Letters. Vol. 30, p. 229, 1977.
- 5. Antonov, V. S., I. N. Knyazev, V. S. Letokhov, V. M. Matiuk, V. G. Movshev, and V. K. Potapov. *Optics Letters*. Vol. 3, p. 37, 1978.
- 6. Halpern, J. B., W. M. Jackson, and V. McCrary. Applied Optics. Vol. 18, p. 590, 1979.
- 7. Rodgers, M. O., K. Asai, and D. D. Davis. Applied Optics. Vol. 19, p. 3597, 1980.
- 8. Syage, J. A., and J. E. Wessel. Applied Spectroscopy Reviews. Vol. 24, pp. 1-79, 1988.
- 9. Jinkins, J. G., J. Schendel, and E. L. Wehry. *Laser Techniques in Luminescence Spectroscopy*. ASTM STP 1066, Philadelphia, PA: American Society for Testing and Materials, edited by T. Vo-Dinh and D. Eastwood, 1990.
- 10. Simeonsson, J. B., and R. C. Sausa. Applied Spectroscopy Reviews. Vol. 31, pp. 1-72, 1996.
- 11. Oldenborg, R. C., and S. L. Baughcum. Analytical Chemistry. Vol. 58, p. 1430, 1986.
- 12. Chadwick, B. L., P. G. Griffin, and R. J. S. Morrison. Applied Spectroscopy. Vol. 51, pp. 990–993, 1997.
- 13. Greger, F., K. T. Hartinger, P. B. Monkhouse, J. Wolfrum, H. Bauman, and B. Bonn. 26th Symposium (International) on Combustion. The Combustion Institute, p. 3301, 1996.
- 14. Buckley, S. G., C. P. Koshland, R. F. Sawyer, and D. Lucas. 26th Symposium (International) on Combustion. The Combustion Institute, p. 2455, 1996.
- 15. Buckley, S. G., C. S. McEnally, R. Sawyer, C. P. Koshland, and D. Lucas. *Combustion Science and Technology*. Vol. 118, p. 169, 1996.
- 16. Hartinger, K. T., S. Nord, and P. B. Monkhouse. Applied Physics B. Vol. 64, p. 363, 1997.
- 17. Peng, W. X., K. W. D. Ledingham, A. Marshall, and R. P. Singhal. Analyst 120, Vol. 2537, 1995.

- 18. Ledingham, K. W. D., C. Kosmidis, S. Georgiou, S. Couris, and R. P. Singhal. Chemical Physics Letters. Vol. 247, p. 555, 1995.
- 19. Singhal, R. P., H. S. Kilic, K. W. D. Ledingham, C. Kosmidis, T. McCanny, A. J. Langley, and W. Shaikh. *Chemical Physical Letters*. Vol. 253, p. 81, 1996.
- 20. Sandholm, S., S. Smyth, R. Bai, and J. Bradshaw. Journal of Geophysical Research. Vol. 102, p. 28651, 1997.
- Sandholm, S., J. D. Bradshaw, G. Chen, H. B. Singh, R. W. Talbot, G. L. Gregory, D. R. Blake, G. W. Sachse, E. V. Browell, J. D. W. Barrick, M. A. Shipman, A. S. Bachmeier, and D. Owen. *Journal of Geophysical Research*. Vol. 97, p. 16481, 1992.
- Sandholm, S., J. Olson, J. Bradshaw, R. Talbot, H. Singh, G. Gregory, D. Blake, B. Anderson, G. Sachse, J. Barrick, J. Collins, K. Klemm, B. Lefer, O. Klemm, K. Gorzelska, D. Herlth, and D. O' Hara. *Journal of Geophysical Research*. Vol. 99, p. 1837, 1994.
- 23. Crawford, J., D. Davis, G. Chen, J. Bradshaw, S. Sandholm, G. Gregory, G. Sachse, B. Anderson, J. Collins, D. Blake, H. Singh, B. Heikes, R. Talbot, and J. Rodriguez. *Journal of Geophysical Research*. Vol. 101, p. 2053, 1996.
- 24. Simeonsson, J. B., and R. C. Sausa. Applied Spectroscopy. Vol. 50, p. 1277, 1996.
- 25. Simeonsson, J. B., R. N. Carter, and A. Peck. Unpublished data.
- 26. Swayambunathan, V., and R. C. Sausa. Unpublished data.
- 27. Schendel, J. S., R. E. Stickel, C. A. van Dijk, S. T. Sandholm, D. D. Davis, and J. D. Bradshaw. Applied Optics. Vol. 29, p. 4924, 1990.
- Williams, E. J., S. T. Sandholm, J. D. Bradshaw, J. S. Schendel, A. O. Langford, P. K. Quinn, P. J. LeBel, S. A. Vay, P. D. Roberts, R. B. Norton, B. A. Watkins, M. P. Buhr, D. D. Parrish, J. G. Calvert, and F. C. Fehsenfeld. *Journal of Geophysical Research*. Vol. 97, p. 11591, 1992.
- 29. Sandholm, S. Personal communication with J. B. Simeonsson. Georgia Institute of Technology, Atlanta, GA.
- 30. Omenetto, N., P. Cavalli, M. Hidalgo, and G. A. Petrucci. Annali di Chimica. Vol. 87, p. 241, 1997.
- 31. Omenetto, N., and U. Panne. Fresenius Journal of Analytical Chemistry. Vol. 355, p. 227, 1996.
- 32. Hidalgo-Nunez, M., P. Cavalli, G. A. Petrucci, and N. Omenetto. Laser Applications to Chemical and Environmental Analysis Symposium. Orlando, FL, March 1998.

- 33. Hynes, A. J., R. C. Richter, and C. J. Nien. Chemical Physics Letters. Vol. 258, p. 633, 1996.
- Kilic, H. S., K. W. D. Ledingham, C. Kosmidis, T. McCanny, R. P. Singhal, S. L. Wang, D. J. Smith, A. J. Langley, and W. Shaikh. *Journal of Physical Chemistry A*. Vol. 101, p. 817, 1997.
- 35. Kosmidis, C., K. W. D. Ledingham, H. S. Kilic, T. McCanny, R. P. Singhal, A. J. Langley, and W. Shaikh. *Journal of Physical Chemistry A*. Vol. 101, p. 2264, 1997.
- 36. Wu, D., J. P. Singh, F. Y. Yueh, and D. L. Monts. Applied Optics. Vol. 35, p. 3998, 1996.
- 37. Simeonsson, J. B., G. W. Lemire, and R. C. Sausa. Analytical Chemistry. Vol. 66, p. 2272, 1994.

INTENTIONALLY LEFT BLANK.

18

Glossary

FD	Fragment detection
LIF	Laser-induced fluorescence
LIRE	Laser ionization recombination emission
LOD	Limit of detection (3 s unless specified otherwise)
MS	Mass spectrometry
PE	Prompt emission
PF	Photofragmentation
PIS	Photoionization spectrometry
REMPI	Resonance-enhanced multiphoton ionization
SE	Stimulated emission
TPI	Two photon ionization

.

•

.

INTENTIONALLY LEFT BLANK.

.

.

NO. OF COPIES ORGANIZATION

- 2 DEFENSE TECHNICAL INFORMATION CENTER DTIC DDA 8725 JOHN J KINGMAN RD STE 0944 FT BELVOIR VA 22060-6218
- 1 HQDA DAMO FDT 400 ARMY PENTAGON WASHINGTON DC 20310-0460
- 1 OSD OUSD(A&T)/ODDDR&E(R) R J TREW THE PENTAGON WASHINGTON DC 20301-7100
- 1 DPTY CG FOR RDA US ARMY MATERIEL CMD AMCRDA 5001 EISENHOWER AVE ALEXANDRIA VA 22333-0001
- 1 INST FOR ADVNCD TCHNLGY THE UNIV OF TEXAS AT AUSTIN PO BOX 202797 AUSTIN TX 78720-2797
- 1 DARPA B KASPAR 3701 N FAIRFAX DR ARLINGTON VA 22203-1714
- 1 NAVAL SURFACE WARFARE CTR CODE B07 J PENNELLA 17320 DAHLGREN RD BLDG 1470 RM 1101 DAHLGREN VA 22448-5100
- 1 US MILITARY ACADEMY MATH SCI CTR OF EXCELLENCE MADN MATH MAJ HUBER THAYER HALL WEST POINT NY 10996-1786

NO. OF COPIES ORGANIZATION

1 DIRECTOR US ARMY RESEARCH LAB AMSRL D D R SMITH 2800 POWDER MILL RD

1 DIRECTOR US ARMY RESEARCH LAB AMSRL DD 2800 POWDER MILL RD ADELPHI MD 20783-1197

ADELPHI MD 20783-1197

- 1 DIRECTOR US ARMY RESEARCH LAB AMSRL CI AI R (RECORDS MGMT) 2800 POWDER MILL RD ADELPHI MD 20783-1145
- 3 DIRECTOR US ARMY RESEARCH LAB AMSRL CI LL 2800 POWDER MILL RD ADELPHI MD 20783-1145
- 1 DIRECTOR US ARMY RESEARCH LAB AMSRL CI AP 2800 POWDER MILL RD ADELPHI MD 20783-1197

ABERDEEN PROVING GROUND

4 DIR USARL AMSRL CI LP (BLDG 305)

NO. OF

COPIES ORGANIZATION

ABERDEEN PROVING GROUND

20 DIR USARL AMSRL WM BD **B E FORCH** W R ANDERSON S W BUNTE C F CHABALOWSKI A COHEN **R DANIEL** D DEVYNCK R A FIFER **B E HOMAN** A J KOTLAR K L MCNESBY M MCQUAID M S MILLER A W MIZIOLEK J B MORRIS R A PESCE-RODRIGUEZ B M RICE R C SAUSA M A SCHROEDER J A VANDERHOFF

REPORT DOC	Form Approved OMB No. 0704-0188						
Public reporting burden for this collection of informatio	n is estimated to average 1 hour per response, in	cluding the time for reviewing instru- Send comments regarding this burg	ctions, searching existing data sources, ien estimate or any other aspect of this				
gathering and maintaining the data needed, and complet collection of information, including suggestions for red	ucing this burden, to Washington Headquarters	Services, Directorate for Information	Operations and Reports, 1215 Jefferson Washington, DC 20503.				
1. AGENCY USE ONLY (Leave blank)	2. REPORT DATE	3. REPORT TYPE AND	DATES COVERED				
	September 2000	Final, Jan 98–Dec	99				
4. TITLE AND SUBTITLE			5. FUNDING NUMBERS				
A Review of Laser Photofrag	gmentation/Fragment Detect	ion Techniques for					
Chemical Analysis of the Gas Pha	ase		622618.AH80				
-							
6. AUTHOR(S)							
Rosario C. Sausa and Joseph B. S	Simeonsson*						
7. PERFORMING ORGANIZATION NAME	(S) AND ADDRESS(ES)		REPORT NUMBER				
ATTN. AMODI WM PD							
ATTN: AMSRL-WW-BD	21005 5066		ARL-TR-2304				
Aberdeen Proving Ground, MD	21005-5000						
9. SPONSORING/MONITORING AGENCY	(NAMES(S) AND ADDRESS(ES)		10.SPONSORING/MONITORING				
			AGENCY REPORT NUMBER				
11. SUPPLEMENTARY NOTES		500 IO 100 I					
*University of Iowa, Department	of Chemistry, Iowa City, IO	52242-1294					
122 DISTRIBUTION/AVAILABILITY STA	TEMENT		12b. DISTRIBUTION CODE				
Approved for public release; dist	ribution is unlimited						
13. ABSTRACT (Maximum 200 words)			1 C 1 their of the evolute				
Laser photofragmentation/fr	agment detection techniques	utilize one or multiple	e lasers for photolysis of the analyte				
species and spectroscopic dete	ction of the characteristic	photofragments. Frag	gment detection techniques include				
laser-induced fluorescence, phot	coionization, prompt emission	n, stimulated emission	, and laser ionization recombination				
emission spectroscopies. Appli	cations for chemical analysis	of the gas phase has	been performed by the detection of				
atomic and molecular fragments	 Recent reports on photofr 	agmentation/fragment	detection approaches are presented,				
and the outlook for future applic	ations is discussed.						
1							
14 SUBJECT TERMS			15. NUMBER OF PAGES				
laser photofragmentation/frag	induced 26						
fluorescence photoionization et	16. PRICE CODE						
fuorescence, photoromzation, en	monon operatoropy						
17. SECURITY CLASSIFICATION 18	CATION 20. LIMITATION OF ABSTRACT						
OF REPORT			FD TT				
UNCLASSIFIED	UNCLASSIFIED	UNCLASSIFI	Standard Form 208 (Rev. 2-80)				
NSN 7540-01-280-5500		73	Prescribed by ANSI Std 239-18 298-102				

INTENTIONALLY LEFT BLANK.

USER EVALUATION SHEET/CHANGE OF ADDRESS

This Laboratory undertakes a continuing effort to improve the quality of the reports it publishes. Your comments/answers to the items/questions below will aid us in our efforts.

1. ARL Report Number/Author <u>ARL-TR-2304 (Sausa)</u> Date of Report <u>September 2000</u>

2. Date Report Received ______

3. Does this report satisfy a need? (Comment on purpose, related project, or other area of interest for which the report will be used.)

4. Specifically, how is the report being used? (Information source, design data, procedure, source of ideas, etc.)

5. Has the information in this report led to any quantitative savings as far as man-hours or dollars saved, operating costs avoided, or efficiencies achieved, etc? If so, please elaborate.

6. General Comments. What do you think should be changed to improve future reports? (Indicate changes to organization, technical content, format, etc.)

Organization

Name

CURRENT ADDRESS

Street or P.O. Box No.

City, State, Zip Code

7. If indicating a Change of Address or Address Correction, please provide the Current or Correct address above and the Old or Incorrect address below.

E-mail Name

Organization

OLD ADDRESS

Name

Street or P.O. Box No.

City, State, Zip Code

(Remove this sheet, fold as indicated, tape closed, and mail.) (DO NOT STAPLE)