

NON-INVASIVE BLOOD FLOW MONITORING ON THE WRIST

M. Maier, L-G. Lindberg

Department of Biomedical Engineering, University of Linköping, Sweden

The Swedish Competence Center of Non-invasive Medical Measurements NIMED, University of Linköping, Sweden

Abstract- Blood flow changes was measured by PPG in vitro and in patients utilizing a new custom designed optical sensor. The results indicate that the sensor may monitor central related blood pressure variations and congestive heart failure. **Keywords** -Photoplethysmography, light reflection, radial artery, red blood cell orientation

I. INTRODUCTION

Measurements of light reflection in moving blood in vitro give information about the orientation and movement pattern of red blood cells in a circular conduit, being a simple model of a human vessel. The relation between light reflection and shear rate, both in steady flow and pulsatile flow conditions, reflects specific scattering characteristics correlated to the orientation of red blood cells.

The purpose of this study was to present and validate a technique and a sensor for monitoring of blood flow changes over the radial artery.

II. METHODOLOGY

A. Photoplethysmography

Photoplethysmography (PPG) is a common non-invasive method mainly used to measure blood oxygenation level, heartbeat and changes in blood flow [1]. The PPG signal also contains information about the circulatory dynamics [2] and respiratory rate and volume [3]. Principally the method utilizes light reflection and scattering in tissue. The skin or organ is illuminated by radiation in the range of 600 to 1300 nm. The reflected or transmitted light is recorded by a photodetector and variations in light intensity are caused by changes in flow and pressure pulsations in blood. The mechanism behind the origin of the PPG signal are not fully understood. Both red blood cell orientation and volume changes are important factors behind the generation of the PPG signal.

B. In vitro experiments

Experiments were performed on blood circulating in a silicone tubing system described earlier [4]. A waveform generator regulated a roller pump, which produced a simulated pressure waveform resembling the human pulsatile blood pressure. A flow-through hole ($\varnothing=2\text{mm}$) was drilled in a piece of acrylic glass. In reflection mode an optical glass fibre was placed adjacent to the illuminating fiber with a center to center distance of 2.5 mm. The light from the detecting fiber was guided to a silicon photodetector. A light emitting diode (LED, wavelength=880 nm) was used as the light source.

C. Human experiments

Blood flow was measured using PPG on the wrist on heart failure patients. Invasive blood pressure was measured in the contralateral radial artery. A custom designed optical sensor was attached over the radial artery with the optical module slightly curving into the tissue when the sensor was secured with a Velcro closing.

III. RESULTS

A. In vitro experiments

Figure 1 demonstrates relative changes in the AC signal of light reflection (RAC) versus relative changes in pulsatile blood pressure when measurements were performed on blood from 12 healthy blood donors in vitro. The correlation coefficient was 0.95 and the corresponding correlation coefficient when plotting the absolute values of reflection against pressure was 0.82.




Figure 1 Relative changes in RAC versus blood pressure variations.

B. Human experiments

Figure 2 shows a typical PPG recording on a healthy male subject at rest

Report Documentation Page

Report Date 25OCT2001	Report Type N/A	Dates Covered (from... to) -
Title and Subtitle Non-Invasive Blood Flow Monitoring on the Wrist	Contract Number	
	Grant Number	
	Program Element Number	
Author(s)	Project Number	
	Task Number	
	Work Unit Number	
Performing Organization Name(s) and Address(es) Department of Biomedical Engineering, University of Linköping, Sweden	Performing Organization Report Number	
Sponsoring/Monitoring Agency Name(s) and Address(es) US Army Research, Development & Standardization Group (UK) PSC 802 Box 15 FPO AE 09499-1500	Sponsor/Monitor's Acronym(s)	
	Sponsor/Monitor's Report Number(s)	
Distribution/Availability Statement Approved for public release, distribution unlimited		
Supplementary Notes Papers from the 23rd Annual International Conference of the IEEE ENgineering in Medicine and Biology Society, October 25-28, 2001, held in Istanbul, Turkey. See also ADM001351 for entire conference on cd-rom.		
Abstract		
Subject Terms		
Report Classification unclassified	Classification of this page unclassified	
Classification of Abstract unclassified	Limitation of Abstract UU	
Number of Pages 2		




Figure 2 A typical PPG recording on a healthy male subject at rest. The slow variation is due to breathing.

IV. DISCUSSION AND CONCLUSION

According to figure 1 there is a close linear relationship between changes in light reflection and changes in pulsatile blood pressure. This means that, despite substantial variations in the haemoglobin concentration (Hb=116-162), which certainly affects the blood viscosity and therefore the orientation capability of the red blood cells, there is a correlation between the pulsatile optical signal and the pulsatile pressure.

A non-invasive sensor for central related blood flow monitoring has been developed. Variations in signal amplitude may reflect different physiological events corresponding to blood pressure variations and congestive heart failure.

ACKNOWLEDGMENT

The authors wish to thank S. Walther, Department of Cardiology and P. Sveider and B. Ragnemalm, Department of Biomedical Engineering. This study was supported by The Swedish Competence Center of Non-invasive Medical Measurements NIMED.

REFERENCES

- [1] A. A. Kamal, J. B. Harness, G. Irving, and A. J. Mearns, "Skin photoplethysmography--a review," *Computer Methods & Programs in Biomedicine*, vol. 28, pp. 257-69, 1989.
- [2] W. B. Murray and P. A. Foster, "The peripheral pulse wave: information overlooked. [see comments]," *Journal of Clinical Monitoring*, vol. 12, pp. 365-77, 1996.
- [3] A. Johansson and P. A. Oberg, "Estimation of respiratory volumes from the photoplethysmographic signal. Part 2: A model study," *Medical & Biological Engineering & Computing*, vol. 37, pp. 48-53, 1999.
- [4] L.-G. Lindberg and P. A. Oberg, "Optical properties of blood in motion," *Optical Engineering*, vol. 32, pp. 253-257, 1993.