

AFRL-MN-EG-TP-2005-7408

REFLECTION PROPERTIES OF SCARABAEIDAE

Dennis H. Goldstein
Air Force Research Laboratory
Munitions Directorate
AFRL/MNGI
Eglin AFB, FL 32542-6810



AUGUST 2005

Conference paper presented at the 50th annual meeting of SPIE, Optics and Photonics, San Diego, California, 1-4 August 2005. This work has been submitted for publication in the Proceedings of SPIE Volume 5888. The author is a U.S. Government employee; therefore, the U.S. Government is owner of the work. SPIE will assert copyright for the proceedings volume within which this paper is published. Nevertheless, the United States has for itself and others acting on its behalf an unlimited, nonexclusive, irrevocable, paid-up, royalty-free worldwide license to use for its purposes.

DISTRIBUTION A: Approved for public release; distribution unlimited.

AIR FORCE RESEARCH LABORATORY, MUNITIONS DIRECTORATE

Air Force Materiel Command

■ United States Air Force

■ Eglin Air Force Base

REPORT DOCUMENTATION PAGE

Form Approved
OMB No. 0704-0188

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports (0704-0188), 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number. **PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ADDRESS.**

1. REPORT DATE (DD-MM-YYYY) 03-08-2005		2. REPORT TYPE Conference Paper		3. DATES COVERED (From - To)	
4. TITLE AND SUBTITLE Reflection properties of Scarabaeidae				5a. CONTRACT NUMBER	
				5b. GRANT NUMBER	
				5c. PROGRAM ELEMENT NUMBER 62602F	
6. AUTHOR(S) Dennis H. Goldstein				5d. PROJECT NUMBER 2068	
				5e. TASK NUMBER 30	
				5f. WORK UNIT NUMBER 45	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Air Force Research Laboratory, Munitions Directorate AFRL/MNGI Eglin AFB, FL 32542-6810				8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES) Air Force Research Laboratory Munitions Directorate AFRL/MNGI Eglin AFB, FL 32542-6810				10. SPONSOR/MONITOR'S ACRONYM(S) AFRL-MN-EG	
				11. SPONSOR/MONITOR'S REPORT NUMBER(S) AFRL-MN-EG-TP-2005-7408	
12. DISTRIBUTION / AVAILABILITY STATEMENT Approved for Public Release; Distribution is Unlimited.					
13. SUPPLEMENTARY NOTES Conference paper presented at the 50th annual meeting of SPIE, Optics and Photonics, San Diego, California, 1-4 August 2005. This work has been submitted for publication in the Proceedings of SPIE Volume 5888. The author is a U.S. Government employee; therefore, the U.S. Government is owner of the work. SPIE will assert copyright for the proceedings volume within which this paper is published. Nevertheless, the United States has for itself and others acting on its behalf an unlimited, nonexclusive, irrevocable, paid-up, royalty-free worldwide license to use for its purposes.					
14. ABSTRACT Beetles of the scarab family have been found to reflect circularly polarized light from incident unpolarized light. There are many known animals that use polarized light in some form and several that actually create it, but there are few examples of the creation of circularly polarized light in nature. Previous work is summarized, and new measurements of several scarab specimens are presented.					
15. SUBJECT TERMS Scarab beetles, Mueller matrix, reflectometer, polarized light, polarimeter.					
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT	18. NUMBER OF PAGES	19a. NAME OF RESPONSIBLE PERSON
a. REPORT	b. ABSTRACT	c. THIS PAGE			Dennis H. Goldstein
UNCLASSIFIED	UNCLASSIFIED	UNCLASSIFIED	SAR	10	19b. TELEPHONE NUMBER (include area code) 850-883-0885

INSTRUCTIONS FOR COMPLETING SF 298

1. REPORT DATE. Full publication date, including day, month, if available. Must cite at least the year and be Year 2000 compliant, e.g. 30-06-1998; xx-06-1998-, xx-xx-1998.

2. REPORT TYPE. State the type of report, such as final, technical, interim, memorandum, master's thesis, progress, quarterly, research, special, group study, etc.

3. DATES COVERED. Indicate the time during which the work was performed and the report was written, e.g., Jun 1997 - Jun 1998; 1-10 Jun 1996; May - Nov 1998; Nov 1998.

4. TITLE. Enter title and subtitle with volume number and part number, if applicable. On classified documents, enter the title classification in parentheses.

Ba. CONTRACT NUMBER. Enter all contract numbers as they appear in the report, e.g. F33615-86-C-5169.

5b. GRANT NUMBER. Enter all grant numbers as they appear in the report, e.g. AFOSR-82-1234.

5c. PROGRAM ELEMENT NUMBER. Enter all program element numbers as they appear in the report, e.g. 61101A.

5d. PROJECT NUMBER. Enter all project numbers as they appear in the report, e.g. 1F665702D1257; ILIR.

5e. TASK NUMBER. Enter all task numbers as they appear in the report, e.g. 05; RF0330201; T4112.

5f. WORK UNIT NUMBER. Enter all work unit numbers as they appear in the report, e.g. 001; AFAPL30480105.

6. AUTHOR(S). Enter name(s) of person(s) responsible for writing the report, performing the research, or credited with the content of the report. The form of entry is the last name, first name, middle initial, and additional qualifiers separated by commas, e.g. Smith, Richard, J, Jr.

7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES). Self-explanatory.

8. PERFORMING ORGANIZATION REPORT NUMBER.

Enter all unique alphanumeric report numbers assigned by the performing organization, e.g. BRL-1234; AFWL-TR-85-4017-Vol-21-PT-2.

9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES). Enter the name and address of the organization(s) financially responsible for and monitoring the work.

10. SPONSOR/MONITOR'S ACRONYM(S). Enter, if available, e.g. BRL, ARDEC, NADC.

11. SPONSOR/MONITOR'S REPORT NUMBER(S).

Enter report number as assigned by the sponsoring/monitoring agency, if available, e.g. BRL-TR-829; -21 5.

12. DISTRIBUTION/AVAILABILITY STATEMENT. Use agency-mandated availability statements to indicate the public availability or distribution limitations of the report. If additional limitations/ restrictions or special markings are indicated, follow agency authorization procedures, e.g. RD/FRD, PROPIN, ITAR, etc. Include copyright information.

13. SUPPLEMENTARY NOTES. Enter information not included elsewhere such as: prepared in cooperation with; translation of; report supersedes; old edition number, etc.

14. ABSTRACT. A brief (approximately 200 words) factual summary of the most significant information.

15. SUBJECT TERMS. Key words or phrases identifying major concepts in the report.

16. SECURITY CLASSIFICATION. Enter security classification in accordance with security classification regulations, e.g. U, C, S, etc. If this form contains classified information, stamp classification level on the top and bottom of this page.

17. LIMITATION OF ABSTRACT. This block must be completed to assign a distribution limitation to the abstract. Enter UU (Unclassified Unlimited) or SAR (Same as Report). An entry in this block is necessary if the abstract is to be limited.

Reflection properties of Scarabaeidae

Dennis H. Goldstein
Air Force Research Laboratory, Eglin AFB FL 32542

ABSTRACT

Beetles of the scarab family have been found to reflect circularly polarized light from incident unpolarized light. There are many known animals that use polarized light in some form and several that actually create it, but there are few examples of the creation of circularly polarized light in nature. Previous work is summarized, and new measurements of several scarab specimens are presented.

1. INTRODUCTION

The creation of polarized light is common in nature, but the generation of circularly polarized light from unpolarized light is quite rare. A. A. Michelson noted in 1911¹ that reflected light from the scarab beetle *Plusiotis resplendens*, a beetle that appears to have been fashioned out of brass or gold, is circularly polarized. Later authors have discussed this effect as well²⁻⁹ and it has been found that only scarabs possess the ability to generate circularly polarized light. Fig. 1a shows *Plusiotis resplendens* in the absence of polarizing optics, and Fig. 1b shows the animal with a circular polarizer in front of the camera. The effect for this creature is more subtle than other scarabs; it is difficult to discern the difference between these two images. A more impressive example is *Plusiotis gloriosa*, shown in Fig. 2. I have examined several of these scarabs for this study and have measured the spectral reflectance and Mueller matrix for the visible region in order to quantify this effect. In Section 2 I review the instrumentation used to make the spectropolarimetric measurements, and in Section 3 I discuss data reduction issues. Results are presented in Section 4, and conclusions are drawn in Section 5.



a. No polarizer b. Circular polarizer in front of camera

Figure 1: *Plusiotis resplendens*



a. No polarizer b. Circular polarizer in front of camera
Figure 2: *Plusiotis gloriosa*

2. DUAL ROTATING-RETARDER MUELLER MATRIX POLARIMETER AND MUELLER MATRIX REFLECTOMETER

An instrument that measures spectral polarization properties of materials in transmission and reflection has been previously designed, patented¹⁰, and described¹¹. The spectropolarimetric reflectometer is based around a commercial Fourier transform spectrometer. The spectropolarimeter generates radiation from the ultraviolet to the far infrared, and has a Visual Basic executive program that allows control over the spectrometer, control over motorized rotation stages, and spectropolarimetric data processing. The spectrometer is a Bio-Rad FTS-6000 with Win-IR Pro software. For purposes of the measurements described here, the spectrometer is used with one source, a xenon lamp, and one silicon detector. The Bio-Rad spectrometer serves as a radiation source for the polarimetric portion of the instrument and is operated in the conventional absorption spectroscopy mode. The radiation generated by the spectrometer is brought out through the spectrometer's external port. Fig. 3 shows the basic optical schematic of the instrument for monostatic reflectance measurements. The detector is fixed in the position shown. The sample is mounted a short distance past the beamsplitter. The optical system that collects light for the detector consists of an off-axis parabolic mirror. This mirror is fixed to look toward the beamsplitter and focus light onto the detector that is mounted perpendicularly to the light coming from the beamsplitter. The parabolic mirror, detector, and mounting devices are referred to as the detector assembly. The sample is mounted vertically. The reflectometer may be used without any modification of the polarization of the source radiation, i.e., no polarization elements, and in this mode, it is a spectral reflectometer. In order to obtain spectropolarimetric measurements, a dual rotating retarder Mueller matrix polarimeter, described by Azzam¹², is included in the system. This polarimeter consists of a polarization state generator before the sample and a polarization state analyzer after the sample. The polarization state generator consists of a linear polarizer followed by a quarter wave retarder. The polarization state analyzer consists of a quarter wave retarder followed by a linear polarizer in front of the detector assembly. Although I use retarders that are nominally quarter wave in the spectral region being measured, the exact retardance is not critical. When the retarders are rotated in a five to one ratio, all sixteen elements of the sample Mueller matrix are encoded onto twelve harmonics of the detected signal, which can then be Fourier analyzed to recover the Mueller matrix elements. Other previous implementations of this Mueller matrix polarimeter have been described elsewhere¹³.

3. DATA REDUCTION AND ERROR COMPENSATION

The data reduction algorithm for this polarimeter as originally presented by Azzam assumes ideal polarization elements and no orientation errors. The data reduction algorithms may be generalized to compensate for systematic errors that result when orientation misalignment and non-ideal retarders are used. If the polarization elements are rotationally misaligned, or the retarders do not have exactly one-quarter wave of retardance, the changes in Fourier amplitudes and phases result in errors in the sample Mueller matrix. Even small orientation and retardance errors ($<1^\circ$) can lead to large errors in the measured Mueller matrix ($> 10\%$ in some matrix elements). These errors become especially important when the retardance and alignment vary

significantly from their nominal values such as in multi-wavelength or spectral instruments. I have incorporated correction terms for orientation and retardance errors into the dual rotating retarder data reduction algorithm. These correction terms are collected during calibration of the instrument. Small angle approximation error correction equations are given in Goldstein and Chipman¹⁴, and this is generalized to larger angles in Chenault et al.¹⁵ These equations are quite lengthy and will not be presented here.




Figure 3: Monostatic spectropolarimetric reflectance measurement configuration

The beamsplitter indicated in Fig. 3 is a polka-dot beamsplitter, i.e. a plate to which silvered dots have been applied. This type of beamsplitter produces good achromatic performance. There is no off-the-shelf beamsplitter known to this author that does not introduce retardance. Beamsplitters that produce a 50/50 split with no retardance have been designed¹⁶, but they have not yet been fabricated. Because of this residual retardance in the beamsplitter, the measured Mueller matrix will contain a retardance that is not associated with the sample. This must be removed to obtain the true Mueller matrix of the sample.

The apparent measured Mueller matrix of the sample is the product of three component Mueller matrices. In order of light transmission, these are 1) the Mueller matrix of the beamsplitter in transmission, 2) the Mueller matrix of the sample, and 3) the Mueller matrix of the beamsplitter in reflection on the beam reflection from the sample. The measured Mueller matrix can then be expressed as

$$M_{measured} = M_{BSR} M_s M_{BST} \quad (1)$$

where the subscript BSR indicates the beamsplitter in reflection and the subscript BST indicates the beamsplitter in transmission. The true Mueller matrix of the sample can be found by solving the equation for M_s so that

$$M_s = M_{BSR}^{-1} M_{measured} M_{BST}^{-1} \quad (2)$$

M_{BST} can be measured directly by removing the sample and placing the polarization state analyzer and detector assembly in line with the original beam. M_{BSR} is found indirectly. The sample is replaced by a high-quality front surface aluminum mirror placed normally to the beam in the original instrument configuration as in Figure 3. The Mueller matrix for an ideal mirror is

$$M_M = \begin{bmatrix} 1 & 0 & 0 & 0 \\ 0 & 1 & 0 & 0 \\ 0 & 0 & -1 & 0 \\ 0 & 0 & 0 & -1 \end{bmatrix} \quad (3)$$

M_{BSR} is then

$$M_{BSR} = M_T M_{BST}^{-1} M_M^{-1} \quad (4)$$

where M_T is the measured Mueller matrix through the whole system with the mirror in the sample position. Measured retardances on the order of 15° have been found for commercial 50/50 beamsplitters designed to be nonpolarizing and used at 45°.

Having established the beamsplitter retardances, the polarimetric reflectometer is operated and saves the data that has been corrected for Mueller matrix polarimeter errors as described in Section 2. The data is then compensated for beamsplitter retardances in Mathcad or MATLAB programs that also plot the data. Tests have been performed with the mirror and other samples to insure that residual retardances all originate from the beamsplitter.

4. MEASUREMENT RESULTS

Results of measurements of the two scarabs in Figs. 1 and 2 are shown in Figs. 4-7. All data was collected at 16 cm^{-1} resolution. Fig. 4 shows the spectral reflectance of *Plusiotis gloriosa* while Fig. 6 shows the spectral reflectance for *Plusiotis resplendens*. The large spike in each of these plots at $0.6328 \text{ }\mu\text{m}$ is due to the Helium-Neon laser in the spectrometer. From Fig. 4 it is evident that most of the reflected light from *Plusiotis gloriosa* is in the green to yellow with two peaks at about $0.53 \text{ }\mu\text{m}$ and $0.58 \text{ }\mu\text{m}$. *Plusiotis resplendens* has a more uniform reflectance distribution with most of the energy at the yellow to red end of the spectrum.

The Mueller matrices for these two scarabs are shown in Figs. 5 and 7, and show objects that are unlike anything else I have measured. The Mueller matrix for *Plusiotis gloriosa* in Fig. 5 is almost a textbook example of a matrix for a homogeneous nonlinear (circular) polarizer, i.e.

$$\begin{bmatrix} 1 & 0 & 0 & 1 \\ 0 & 0 & 0 & 0 \\ 0 & 0 & 0 & 0 \\ 1 & 0 & 0 & 1 \end{bmatrix}$$




Figure 4: Spectral reflectance of *Plusiotis gloriosa*




Figure 5: Mueller matrix for *Plusiotis gloriosa*




Figure 6: Spectral reflectance of *Plusiotis resplendens*




Figure 7: Mueller matrix for *Plusiotis resplendens*

The behavior of the Mueller matrix for *Plusiotis resplendens* has more features. The matrix is of an object that generates circularly polarized light from unpolarized light but with considerable variation from one end of the visible spectrum to the other. At the same time, at the short wave end of the spectrum it has mirror-like qualities. The hand of the matrix component generating circularly polarized light from unpolarized light, M30 in Fig. 7, actually reverses twice in going from 0.4 to 0.7 μm .

5. CONCLUSIONS

I have presented reflectance spectra and Mueller matrices for two scarab beetles. This is the first explicit instance of this data in the literature known to me. Michelson¹ (1911) describes the fact that the hand of the circular polarization reverses from the blue end of the spectrum to the red, and Gaubert² (1924) agrees with this. In a much later paper (1969), Neville and Caveney⁵ disagree with this. The current results validate the results of Michelson and Gaubert with regard to this hand reversal, and in fact, it appears that there are two reversals in the visible wavelength region.

REFERENCES

1. A. A. Michelson, "On Metallic Colouring in Birds and Insects," *Phil. Mag.* **21**, 554-567 (1911).
2. P. Gaubert, "Sur la polarisation circulaire de la lumière réfléchie par les Insectes," *Comptes Rendus* **179**, 1148-1150 (1924).
3. J.P. Mathieu and N. Faraggi, "Etude de la lumière polarisée circulairement réfléchie par certains Coleopteres," *Comptes Rendus* **205**, 1378-1380 (1937).
4. C. Robinson, "The cholesteric phase in polypeptide solutions and biological structures," *Mol. Cryst. I*, 467-494 (1966).
5. A.C. Neville and S. Caveney, "Scarabaeid beetle exocuticle as an optical analogue of cholesteric liquid crystals," *Biol. Rev.* **44**, 531-562 (1969).
6. G. P. Können, *Polarized Light in Nature*, Cambridge University Press (1985).
7. G. W. Kattawar, "A Search for Circular Polarization in Nature," *Optics and Photonics News*, 42-43, September 1994.
8. M. Srinivasarao, "Nano-Optics in the Biological World: Beetles, Butterflies, Birds, and Moths," *Chem. Rev.* **99**, 1935-1961 (1999).
9. G. Horváth and D. Varjú, *Polarized Light in Animal Vision*, Springer (2004).
10. D. H. Goldstein and D. B. Chenault, "Spectropolarimetric reflectometer," *Opt. Eng.* **41**, pp, 1013-1020, 2002.
11. D. H. Goldstein, D. B. Chenault, and M. Owens, "Spectropolarimetric Reflectometer," U. S. Patent No. 6.618,145, 9 September 2003.
12. R.M.A. Azzam, "Photopolarimetric measurement of the Mueller matrix by Fourier analysis of a single detected signal," *Opt. Lett.* **2**, pp. 148-150, 1978.

13. D.H. Goldstein, "Mueller matrix dual-rotating retarder polarimeter," *Appl. Opt.* **31**, pp. 6676-6683, 1992.
14. D. H. Goldstein and R. A. Chipman, "An Error Analysis of a Mueller Matrix Polarimeter," *J. Opt. Soc. Am. A* **7**, pp. 693-700, 1990.
15. D. B. Chenault, J. L. Pezzaniti, and R. A. Chipman, "Mueller matrix algorithms," *Polarization Analysis and Measurement*, D. H. Goldstein and R. A. Chipman, eds., Proc. SPIE 1746, pp. 231-246, San Diego CA, 19-21 July 1992.
16. A. De and R. M. A. Azzam, "Optimal coated silicon membrane beam splitters for the division-of-amplitude photopolarimeter (DOAP)," presented at the OSA Annual Meeting, October 2, 2002.