## WRAIR PROTOCOLS FOR SOLDIER STATUS AND READINESS TO ORGANOPHOSPHATE EXPOSURE: UNPROCESSED WHOLE BLOOD CHOLINESTERASE AND PYRIDOSTIGMINE BROMIDE QUANTIFICATION

Gregory E. Garcia, Shawn R. Feaster, Deborah R. Moorad, Bhupendra P. Doctor, and Richard K. Gordon Walter Reed Army Institute of Research, Division of Biochemistry, Department of Biochemical Pharmacology, Silver Spring, MD 20910-7500,

Connie R. Clark, J. Richard Smith, and Brian J. Lukey U.S. Army Medical Research Institute of Chemical Defense,

Raven E. Reitstetter U.S. Army Center for Health Promotion and Preventive Medicine Aberdeen Proving Ground, MD 21010, and

## ABSTRACT

Exposure to nerve agents, OPs, pesticides, anesthetics, terrorists' chemical agents, and drugs of abuse such as cocaine, heroin, and some neuro-degenerative disease states selectively reduces AChEand/or BChE activity. Since urban terrorism is on the rise, Federal, State, and local authorities need a reliable, fast, inexpensive method for confirming such an assault to initiate appropriate containment, decontamination, and treatment measures. Thus, we developed a semi-automated medical diagnostic microplate procedure capable of screening unprocessed whole blood samples for the concentrations of AChE and BChE (patent pending, WRAIR) to overcome limitations of current methods. We applied our technique to whole blood in man, non-human primates, and rodents. In addition, pyridostigmine bromide (PB) treatment is the only prophylactic treatment for troops who expect to be exposed to OPs. Thus, we developed a sensitive HPLC technique to quantify PB in human blood and also in Rhesus and rodent blood and tissue. The technique uses solid phase extraction, lyophilization for concentration, and HPLC of the reconstituted samples followed by strong-cation exchange chromatography via isocratic elution. These techniques will provide the soldier, field medic, hospital, company, or

civilian organization with the ability to rapidly determine exposure to OPs and aid in treatment while the HPLC assay will aid in verifying troop compliance when required for PB prophylaxis against OPs.

Report Documentation Page				Form Approved OMB No. 0704-0188	
Public reporting burden for the collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to a penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number.					
1. REPORT DATE 01 JUL 2003	2. REPORT TYPE N/A		3. DATES COVERED		
4. TITLE AND SUBTITLE				5a. CONTRACT NUMBER	
WRAIR Protocols For Soldier Status And Readiness To Organophosphate Exposure: Unprocessed Whole Blood Cholinesterase				5b. GRANT NUMBER	
And Pyridostigmine Bromide Quantification				5c. PROGRAM ELEMENT NUMBER	
6. AUTHOR(S)				5d. PROJECT NUMBER	
				5e. TASK NUMBER	
				5f. WORK UNIT NUMBER	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Walter Reed Army Institute of Research, Division of Biochemistry, Department of Biochemical Pharmacology, Silver Spring, MD 20910-7500				8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES)				10. SPONSOR/MONITOR'S ACRONYM(S)	
				11. SPONSOR/MONITOR'S REPORT NUMBER(S)	
12. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release, distribution unlimited					
13. SUPPLEMENTARY NOTES See also ADM001523.					
14. ABSTRACT					
15. SUBJECT TERMS					
16. SECURITY CLASSIFICATION OF: 17. LIMITATION O				18. NUMBER	19a. NAME OF
a. REPORT unclassified	b. ABSTRACT unclassified	c. THIS PAGE unclassified	UU	4	NESPONSIBLE PERSON

Standard Form 298 (Rev. 8-98) Prescribed by ANSI Std Z39-18

## INTRODUCTION

The physiological role of acetylcholinesterase (AChE, EC 3.1.1.7) is the acceleration of hydrolysis of the neurotransmitter acetylcholine at nerve-nerve and neuromuscular junctions. This enzyme possesses one of the fastest turnover rates known. Inhibitors of AChE have been used medicinally in the treatment of glaucoma, myasthenia gravis, and recently for Alzheimer's disease with FDA approval of tacrine and Aricept. Other inhibitors have been used agriculturally and domestically as pesticides and insecticides. Unfortunately, some of the most potent inhibitors of cholinesterases have been developed as chemical warfare agents (CWA), e.g., GA (tabun), GB (sarin, the agent used by the terrorist group in Japan in 1995), GD (soman), and VX, These organophosphorus (OP) nerve agents are a serious threat to US military personnel. The sequel to AChE poisoning is a cholinergic crisis in man; the clinical effects are directly related to acetylcholine accumulation.

Exposure to nerve agents, OPs, pesticides, anesthetics, terrorists' chemical agents, drugs of abuse such as cocaine, heroin, and some neuro-degenerative disease states selectively reduces AChE and/or BChE activity. Thus, blood cholinesterase activity can be exploited as a tool for confirming exposure to such toxic agents. Furthermore, since urban terrorism is on the rise, Federal, State, and local authorities need a reliable, fast, inexpensive, and standard method for confirming such an assault in order to initiate appropriate containment, decontamination, and treatment measures. Finally, pyridostigmine bromide (PB) treatment is the only prophylactic treatment for troops who expect to be exposed to OPs. PB (30 mg tablet, twice daily recommended) results in about 30% reversible inhibition of normal levels of blood cholinesterase. Therefore, the screening method described here can also be utilized to determine the level of blood cholinesterase protected by PB and thus the protective status of troops prior to deployment.

Current clinical determination of cholinesterase levels of blood includes the Michel, microMichel, pH stat, Ellman, and microEllman methodologies. These methods, however, have long clinical turn around times, lack of standardization among labs, and normally determine only the serum or red blood cell cholinesterase concentration, but not both. In addition to the clinical methodologies, a field deployable unit is commercially available: the Test-Mate OPTM system. The kit requires several processing steps and the use of a selective BChE inhibitor for complete AChE and BChE screening.To overcome these limitations of current methods, we developed a semi-automated medical diagnostic microplate procedure capable of screening unprocessed whole blood samples for the concentrations of AChE and BChE at unprecedented precision (<1 %) and accuracy (>99%) (patent pending, WRAIR). In addition, the procedure is being ported to a high-throughput robotic system. We have successfully applied our technique to whole blood in the presence and absence of selective (huperzine-a and Iso-OMPA), and non-selective (pyridostigmine bromide) cholinesterase inhibitors. This is possible because blood contains two cholinesterases that possess different affinities for any given substrate, and a linear correlation exists between enzyme activity and concentration. Thus, if one determines the activity of any given blood sample with two different substrates, then it is feasible to calculate the precise concentrations of both proteins (i.e., two equations with two unknowns). Furthermore, monitoring the activity with three substrates (Figure 1) provides three fold degenerate data [i.e., three sets of two equations with two unknowns]. The rates of substrate hydrolysis represented by R1, R2, and R3 correspond to the turnover of substrate 1, substrate 2, and substrate 3, respectively. [AChE] and [BChE] refer to the actual concentrations of AChE and BChE contained in the sample. Finally, the coefficients in each equation (i.e., x<sub>1</sub>, x<sub>2</sub>, x<sub>3</sub>, and y<sub>1</sub>, y<sub>2</sub>, y<sub>3</sub>) represent the sensitivity coefficients. These coefficients represent the contribution that AChE and BChE independently contribute to the overall rate of substrate hydrolysis (R<sub>1</sub>, R<sub>2</sub>, R<sub>3</sub>). Simultaneously solving these three sets of degenerate equations provides three independent estimates for the concentrations of AChE and BChE. Therefore, determining the mean value and the standard deviation for these independently derived values provides an excellent estimate of the authentic concentrations of each protein.



Figure 1. Mathematical representation of the WRAIR sensitivity coefficient method.

We found that the results obtained from these experiments parallel those in the literature. In addition, the sensitivity coefficient method has been applied to several species including human, rhesus monkey, Hartley guinea pig, and Sprague Dawley rat. Thus, blood levels of cholinesterases after exposure to CWAs such as GD and pharmaceuticals such as huperzine-a and pyridostigmine have been determined in animals and man. We have also performed comparisons with the COBAS/FARA and TestMate  $OP^{TM}$  systems, which yielded identical trends. However, unlike the conventional clinical tests or the Test-Mate  $OP^{TM}$  unit, our method provides a full analysis of the patient's cholinesterase levels, does not rely on the addition of selective AChE or BChE inhibitors, uses a single non-invasive blood collection technique (currently a finger prick), is not labor intensive, and produces results in less than six minutes.

A field unit for blood cholinesterase (i.e., individual use) is in development; it is based on the new technology of a single use disposable silicon needle (US patent #5,801,057). This needle is minimally invasive and painless, removes less than a drop of blood, and is now being commercialized for diabetic blood glucose monitoring. Detection for measuring the rate of cholinesterase-catalyzed hydrolysis has been demonstrated with a single substrate redox-mediated electrochemical sensor utilizing butyrylthiocholine as a substrate, which is hydrolyzed by the enzyme to form thiocholine. The latter product is consumed by reacting with ferricyanide, and is then oxidized at the electrode to form ferrocyanide. Within 60 sec after initiation of the assay, a level of cholinesterase reduced more than 50% is readily observed by the electrochemical detector. The hollow silicon microneedle, comparable in cross-section to a human hair, is now being integrated with lab-on-a-chip microfluidics.

Lastly, to verify compliance of ordered PB consumption (when expecting OP exposure), we have developed a sensitive HPLC technique to quantify PB in human blood and also in Rhesus, and rodent blood and tissue. This is an alternate and confirmatory technique for measurement of PB levels directly to supplement the indirect assay by inhibition of ChEs. The PB HPLC technique is based on solid phase extraction, lyophilization for concentration, and HPLC of the reconstituted samples using strong-cation

exchange chromatography and isocratic elution. The linear dynamic range of sensitivity covers at least 500 to 0.00125  $\mu$ g of PB. This assay lends itself to high-throughput and direct PB quantification. The utility and robustness of the technology is shown by adoption at other DOD agencies involving measurements of PB in plasma samples and it is used in studies at the Uniform Services University of Health Sciences for low-dose combination effects of PB, atropine, and pyrimethamine in Gulf War syndrome investigations.

While the cholinesterase assay method using a handheld device is designed for the  $1^{st}$  echelon (or by first responders), the high-throughput assay is readily performed at  $3^{rd}/4^{th}$  echelons (and in civilian clinical laboratories). These techniques will provide the soldier, field medic, hospital, company, or civilian organization with the ability to rapidly determine exposure to OPs and aid in treatment. In addition, the sensitive HPLC assay for quantifying PB levels will aid in verifying troop compliance when required for prophylaxis against OPs.