

<b>REPORT DOCUMENTATION PAGE</b>			<i>Form Approved</i> <i>OMB No. 0704-0188</i>		
Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports (0704-0188), 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number. <b>PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ADDRESS.</b>					
<b>1. REPORT DATE (DD-MM-YYYY)</b> 25 Mar 2013		<b>2. REPORT TYPE</b> Consultative Letter		<b>3. DATES COVERED (From – To)</b> Oct 2012 – Feb 2013	
<b>4. TITLE AND SUBTITLE</b>  PRL S-030A Verification Survey at Former McClellan AFB, Sacramento, CA			<b>5a. CONTRACT NUMBER</b>		
			<b>5b. GRANT NUMBER</b>		
			<b>5c. PROGRAM ELEMENT NUMBER</b>		
<b>6. AUTHOR(S)</b> Maj Alan Hale			<b>5d. PROJECT NUMBER</b>		
			<b>5e. TASK NUMBER</b>		
			<b>5f. WORK UNIT NUMBER</b>		
<b>7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES)</b> USAF School of Aerospace Medicine Occupational and Environmental Health Dept/OEC 2510 Fifth St. Wright-Patterson AFB, OH 45433-7913			<b>8. PERFORMING ORGANIZATION REPORT NUMBER</b>  AFRL-SA-WP-CL-2013-0005		
<b>9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES)</b>			<b>10. SPONSORING/MONITOR'S ACRONYM(S)</b>		
			<b>11. SPONSOR/MONITOR'S REPORT NUMBER(S)</b>		
<b>12. DISTRIBUTION / AVAILABILITY STATEMENT</b>  Distribution A: Approved for public release; distribution is unlimited. Case Number: 88ABW-2013-1395, 25 Mar 2013					
<b>13. SUPPLEMENTARY NOTES</b>					
<b>14. ABSTRACT</b> At the request of the U.S. Air Force Radioisotope Committee Secretariat (RICS), the U.S. Air Force School of Aerospace Medicine Consultative Services Division (USAFSAM/OEC) completed an independent radiological assessment/verification survey from 23-25 Oct 12 at site PRL S-030A on former McClellan AFB, CA. This verification survey was conducted at the site after the contractor had completed the majority of the excavation of contaminated soils from the site. Radium-226 was the sole radionuclide of concern. Cabrera Services, Inc., under contract with Tetra Tech, conducted all radiological field work to include the Final Status Survey (FSS). This letter details the findings of this visit and is meant to assist the RICS when evaluating the contractor's FSS report of this site.					
<b>15. SUBJECT TERMS</b> USAF School of Aerospace Medicine (USAFSAM), former McClellan AFB, radium-226, verification survey, final status survey, independent radiological assessment					
<b>16. SECURITY CLASSIFICATION OF:</b>			<b>17. LIMITATION OF ABSTRACT</b>	<b>18. NUMBER OF PAGES</b>	<b>19a. NAME OF RESPONSIBLE PERSON</b>
<b>a. REPORT</b> U	<b>b. ABSTRACT</b> U	<b>c. THIS PAGE</b> U			Col Mark Smallwood
			SAR	27	<b>19b. TELEPHONE NUMBER (include area code)</b>



**DEPARTMENT OF THE AIR FORCE**  
**AIR FORCE RESEARCH LABORATORY**  
**WRIGHT-PATTERSON AFB OHIO**

25 March 2013

MEMORANDUM FOR AFMSA/SG3PB

ATTN: MAJ DANIEL SHAW  
USAF RADIOISOTOPE COMMITTEE SECRETARIAT  
AIR FORCE MEDICAL SUPPORT AGENCY  
7700 ARLINGTON BOULEVARD, SUITE 5158  
FALLS CHURCH, VA 22042-5158

FROM: USAFSAM/OEC  
2510 Fifth Street  
Wright-Patterson AFB, OH 45433

SUBJECT: Consultative Letter, AFRL-SA-WP-CL-2013-0005, PRL S-030A Verification  
Survey at Former McClellan AFB, Sacramento, CA

1. INTRODUCTION:

a. *Purpose:* At the request of the U.S. Air Force Radioisotope Committee Secretariat (RICS), the U.S. Air Force School of Aerospace Medicine Consultative Services Division (USAFSAM/OEC) completed an independent radiological assessment/verification survey from 23-25 Oct 2012 at site PRL S-030A on former McClellan AFB, CA. This verification survey was conducted at the site after the contractor had completed the majority of the excavation of contaminated soils from the site. Radium-226 (Ra-226) was the sole radionuclide of concern. Cabrera Services, Inc., under contract with Tetra Tech, conducted all radiological field work including the Final Status Survey (FSS). This letter details the findings of this visit and is meant to assist the RICS when evaluating the contractor's FSS report of this site.

b. *Survey Personnel:*

- (1) Chief, Radiation Health Consulting Branch, USAFSAM/OEC
- (2) Health Physics Technician, USAFSAM/OEC

c. *Personnel Contacted:*

- (1) Radiation Program Manger, AFCEC/CIBW
- (2) Radiation Safety Officer, AFCEC/CIBW

d. *Equipment:*

- (1) Ludlum-Model 2221, serial number 218606, calibrated on 23 Nov 11
- (2) Ludlum-Model 44-10, serial number PR276614, calibrated on 23 Nov 11
- (3) Ludlum-Model 2221, serial number 78153, calibrated on 25 Nov 11
- (4) Ludlum-Model 44-10, serial number PR276615, calibrated on 25 Nov 11

## 2. METHODOLOGY:

a. *Background Area:* The background area of 400 m<sup>2</sup> has been routinely used as a reference area since it was determined to be radiologically nonimpacted. The background area was characterized with the 2x2 sodium iodide detector using a gamma walkover technique. No soil samples were taken in the background area, as the background soil concentrations for former McClellan AFB were determined previously and were not relevant for this survey. After gamma walkover surveys at the survey sites, it was determined that the background area was not consistent with the survey areas, most likely due to the fact the background area still had vegetation in place. The background data are, therefore, irrelevant and are not presented.

b. *Survey Areas:* Cabrera Services, Inc. was responsible for conducting a total of two FSSs at two different sites. The week of the visit, PRL S-030A was anticipated to be available for USAFSAM verification. Therefore, after consultation with the RICS, it was chosen for this survey. During the week of the visit, PRL S-030A had not been completely remediated. The eastern portion of the site had residual contamination that the contractors were still surveying and remediating; therefore, the FSS of the site had not been completed. As a result, USAFSAM was only able to survey the area of PRL S-030A that had been remediated. The survey area underwent gamma walkover surveys and soil sampling. This verification survey required at least 10% of the survey area to be scanned by the gamma walkover technique and soil sampling totaling at least 10% of the number taken by the contractor. Further information on the surveys can be found in Attachments 1-2.

(1) *Gamma Walkover Survey in Survey Area:* The goals of the walkover survey were to detail site radiological conditions, identify potential spots of elevated residual Ra-226 concentrations, and identify locations for biased soil sampling. Since the gamma walkover data were to qualitatively assess the site in terms of mean reading plus amount of standard deviation (SD), minimum detectable concentration and count rates were not calculated. Table 1, in section 3, summarizes the results of the gamma walkover survey.

(2) *Soil Sampling in Survey Area:* Since the contractor planned to collect 64 soil samples at PRL S-030A, 8 biased soil samples were collected based on walkover data exhibiting elevated count rates. Each sample was taken from an area of about 8 square inches to a depth of 6 inches. The volume sampled was enough for laboratory analysis (approximately 0.25 gallon). Field soil sampling procedures were in place to prevent cross-contamination of samples. Table 2, in section 3, summarizes the soil sample results.

3. RESULTS: The results are summarized below along with some basic discussion. Additional substantiating data, including maps and survey data, can be found in Attachments 1-5.

**Table 1. Gamma Walkover Results Summary**

<b>Instrument</b>	<b>Mean Count Rate (cpm)</b>	<b>SD (cpm)</b>	<b>No. of Readings &gt; Mean + 3 SD</b>
Ludlum Model 2221/44-10 (SN: 218606/PR276614)	7313	931	14
Ludlum Model 2221/44-10 (SN: 78153/PR276615)	7282	828	84

Table 2. Soil Sampling Results Summary

Sample No.	Ra-226 Concentration (pCi/g)	Ra-226 Concentration Uncertainty (pCi/g)
CS00001	17.2	±0.248
CS00002	1.09	±0.0613
CS00003	1.30	±0.0525
CS00004	7.44	±0.149
CS00005	1.99	±0.0672
CS00006	1.83	±0.0588
CS00007	0.831	±0.0489
CS00008	1.51	±0.0571


#### 4. CONCLUSIONS:

a. All soil samples were taken in biased locations where elevated Ra-226 concentrations were expected to be found. The maximum Ra-226 concentration was  $17.2 \pm 0.248$  pCi/g. This area was identified during the survey as needing remediation; see Figure A2-3 in Attachment 2. Additionally two other samples, CS0004 and CS0005, were greater than the cleanup goal of 2.0 pCi/g when uncertainty is added.

b. USAFSAM acknowledges that the site had not undergone a comprehensive contractor FSS when the portion of the site was verified by USAFSAM. However, the contractor failed to identify the areas USAFSAM detected during its remediation support surveys.

c. Based upon the findings of this verification survey, USAFSAM recommends the RICS carefully review the contractor's FSS report for PRL S-030A before the site is approved for unrestricted free release of radiological controls.

5. If you have any questions or need further information, please contact Maj Alan Hale at 937-938-3320 (DSN 798-3320) or alan.hale@us.af.mil.



ALAN C. HALE, Maj, USAF, BSC  
Chief, Radiation Health Consulting Branch

#### 5 Attachments:

1. Instrumentation and Survey Methodology
2. Survey of PRL S-030A
3. Laboratory Soil Analysis
4. Instrument Calibration Sheets
5. Radiation Meter QC Log

## ATTACHMENT 1

### INSTRUMENTATION AND SURVEY METHODOLOGY

1. *Sodium Iodide Detector*: The sodium iodide detectors used were Ludlum 44-10 2x2-inch detectors coupled with a Ludlum 2221 ratemeter/scaler. For walkover surveys, the Ludlum 2221 was connected via cable to a Trimble GeoXT handheld GPS unit. The Trimble GeoXT logged the count rates with the corresponding GPS coordinates every second. The detector was held at 10 cm above the ground for soil readings and walkover surveys. During walkovers, the scan speed was approximately 0.5 meters per second. This is consistent with the contractor methodology used during gamma walkover surveys.

2. *Laboratory Analysis of Soil*: Soil samples were counted at the USAFSAM Radioanalytical Laboratory (OEAL) at WPAFB, OH. The soil was counted on a high purity germanium detector. The soils were also analyzed by OEAL using the appropriate in-growth method to determine Ra-226 levels in soil. Attachment 3 contains the laboratory reports.

3. All field instruments were function checked and field tested before and after use with a check source. All instruments were tested to ensure a 20% tolerance during field checks. All of the instruments passed the tolerance function check except the post-survey check of survey meter serial number 218606 with probe serial number PR276614 as documented in Attachment 5. A review of the data collected with this meter verified that the data were valid. The suspected reason for the instrument failing the post-survey check was a loose connection in the probe. While surveying, the probe was held vertically, which did not affect the connection; however, during the function check the probe was in a horizontal position, which affected the connection. All instruments are calibrated on an annual basis at WPAFB, OH. Attachment 4 contains all annual calibration sheets and Attachment 5 is the Radiation Meter QC Log.

## ATTACHMENT 2

### SURVEY OF PRL S-030A

1. *Location:* Figures A2-1, A2-2, and A2-3 are photographs of the PRL S-030A site during verification survey work. Note standing water in the excavation area (Figures A2-1 and A2-2), which limited the area available for the gamma walkover survey.



**Figure A2-1. Photograph of PRL S-030A, Gamma Walkover Survey**



**Figure A2-2. Photograph of PRL S-030A, Soil Sampling**



**Figure A2-3. Photograph of PRL S-030A, Hotspot Found Corresponding to Soil Sample CS0001**

2. *Survey Results:* Eight soil samples were taken at the PRL S-030A site. The gross Ra-226 soil concentration results are in Table A2-1, which is inclusive of the background Ra-226 concentration and can be compared directly to the cleanup goal of 2.0 pCi/g. The locations of the soil samples are in Figure A2-4. The PRL S-030A site area of approximately 5000 m<sup>2</sup> was characterized with a sodium iodide detector using a gamma walkover technique. Scan coverage was approximately 75%. Due to the shallow nature of the excavation, only the floor of the excavation was scanned and not the side walls. The resultant data of this walkover are listed in Figure A2-5. Since there was a disparity in the mean count rates of the background survey area and the survey unit due to soil difference, through analysis of the statistics the mean count rate of the survey unit itself and associated statistics were used as the “background” for this survey unit. The lowest readings of the survey are marked in green and correspond to less than 2 SDs above the mean background value. This green area is where soil concentrations are expected to be the lowest. The other colors represent areas of statistical significance where higher concentrations are expected to be found. Yellow and red colors correspond to greater than 2 and 3 SDs, respectively, above the mean count rate value. The color scheme demonstrates a scale of instrument data based upon SDs from background, where no regulatory values are implied.



Table A2-1. Soil Sample Results for PRL S-030A

Sample No.	GPS Coordinates (°N/°E)	Gross Ra-226 Concentration (pCi/g)
CS00001	38.650007/ -121.406722	17.2 ± 0.248
CS00002	38.650110/ -121.406746	1.09 ± 0.0613
CS00003	38.649986/ -121.406863	1.30 ± 0.0525
CS00004	38.649945/ -121.406921	7.44 ± 0.149
CS00005	38.649909/ -121.406917	1.99 ± 0.0672
CS00006	38.649484/ -121.406786	1.83 ± 0.0588
CS00007	38.649473/ -121.406710	0.831 ± 0.0489
CS00008	38.649502/ -121.407166	1.51 ± 0.0571




Figure A2-4. Soil Sample Locations for PRL S-030A



Figure A2-5. Gamma Walkover Data for PRL S-030A

ATTACHMENT 3

LABORATORY SOIL ANALYSIS

---

Sample Analysis Results Reported on 07-Jan-2013

---

USAFSAM/OEHHL ID 11200280  
Customer Address 01172  
77 MDOS/SGFB  
5942 DUDLEY BLVD

MCCLELLAN AFB CA 95652-1074

---

IDENTIFICATION

---

Base Sample # CS00001 Serial #  
Date Collected 10/25/2012 Received 10/29/2012 Completed 12/20/2012

---

Analytes	Activity +/- Uncertainty	Lc / MDA
RADIUM 226	1.72E+01 +/- 2.48E-01 pc/g	7.71E-02 / 1.57E-01 pc/g

---

COMMENTS  
RA226

---

RESULTS ACCURATE TO 2 SIGNIFICANT FIGURES.  
UNCERTAINTY AT 95% CONFIDENCY LEVEL.

---

If you have any questions concerning the information provided above, please contact the ESOH Service Center at 1-888-232-2353.

ROBERT D. SCHMIDTGOESSLING, Major, USAF, BSC  
Chief, Analytical Laboratory Branch

Sample Analysis Results Reported on 07-Jan-2013

USAFSAM/OHHL ID 11200281

Customer Address 0117Z

77 MDOS/SGPB  
5342 DUDLEY BLVD

MCCLELLAN AFB CA 95652-1074

IDENTIFICATION

Base Sample # CS00002 Serial #  
Date Collected 10/25/2012 Received 10/29/2012 Completed 12/20/2012

Analytes	Activity +/- Uncertainty	Lc / MDA
RADIUM 226	1.09E+00 +/- 6.13E-02 pcig	5.63E-02 / 1.15E-01 pcig

COMMENTS

RESULTS ACCURATE TO 2 SIGNIFICANT FIGURES.  
UNCERTAINTY AT 95% CONFIDENCY LEVEL.

If you have any questions concerning the information provided above, please contact the ESOH Service Center at 1-888-232-2383.

ROBERT D. SCHMIDTGOESSLING, Major, USAF, BSC  
Chief, Analytical Laboratory Branch

Sample Analysis Results Reported on 07-Jan-2013

USAFSAM/OBHHL ID 11200282

Customer Address 0117Z

77 MDOS/SGFB  
5342 DUDLEY BLVD

MCCLELLAN AFB CA 95652-1074

IDENTIFICATION

Base Sample # CS00003 Serial #  
Date Collected 10/25/2012 Received 10/29/2012 Completed 12/20/2012

Analytes	Activity +/- Uncertainty	Lc / MDA
RADIUM 226	130E+00 +/- 5.25E-02 pcig	4.43E-02 / 9.07E-02 pcig

COMMENTS

RESULTS ACCURATE TO 2 SIGNIFICANT FIGURES.  
UNCERTAINTY AT 95% CONFIDENCY LEVEL.

If you have any questions concerning the information provided above, please  
contact the ESOH Service Center at 1-888-232-2353.

ROBERT D. SCHMIDTGOESSLING, Major, USAF, ESO  
Chief, Analytical Laboratory Branch

Sample Analysis Results Reported on 07-Jan-2013

USAFSAM/OMHHL ID 11200283

Customer Address 0117Z

77 MDOS/SGFB  
5342 DUDLEY BLVD

MCCLELLAN AFB CA 95652-1074

IDENTIFICATION

Base Sample # CS00004 Serial #  
Date Collected 10/25/2012 Received 10/29/2012 Completed 12/20/2012

Analytes	Activity +/- Uncertainty	Lc / MDA
RADIUM 226	7.44E+00 +/- 1.49E-01 pci/g	6.11E-02 / 1.25E-01 pci/g

COMMENTS

RESULTS ACCURATE TO 2 SIGNIFICANT FIGURES.  
UNCERTAINTY AT 95% CONFIDENCY LEVEL.

If you have any questions concerning the information provided above, please  
contact the ESOH Service Center at 1-866-232-2353.

ROBERT D. SCHMIDTGOESSLING, Major, USAF, BSC  
Chief, Analytical Laboratory Branch

Sample Analysis Results Reported on 07-Jan-2013

USAFSAM/OMHHL ID 11200284  
Customer Address 01172  
77 MDOS/SGPB  
5342 DUDLEY BLVD

MCCLELLAN AFB CA 95652-1074

IDENTIFICATION

Base Sample # CS00005 Serial #  
Date Collected 10/25/2012 Received 10/29/2012 Completed 12/20/2012

Analytes	Activity +/- Uncertainty	Lc / MDA
RADIUM 226	1.99E+00 +/- 6.72E-02 pd/g	4.31E-02 / 8.87E-02 pd/g

COMMENTS

RESULTS ACCURATE TO 2 SIGNIFICANT FIGURES.  
UNCERTAINTY AT 95% CONFIDENCY LEVEL.

If you have any questions concerning the information provided above, please  
contact the ESOH Service Center at 1-888-232-2353.

ROBERT D. SCHMIDTGOESSLING, Major, USAF, BSC  
Chief, Analytical Laboratory Branch

Sample Analysis Results Reported on 07-Jan-2013

USAFSAM/OMHHL ID 11200285

Customer Address 01172

77 MDOS/SGPB  
5342 DUDLEY BLVD

MCCLELLAN AFB CA 95652-1074

IDENTIFICATION

Base Sample # CS00006 Serial #  
Date Collected 10/25/2012 Received 10/29/2012 Completed 12/20/2012

Analytes	Activity +/- Uncertainty	Lc / MDA
RADIUM 226	1.83E+00 +/- 5.88E-02 pc/g	4.46E-02 / 9.18E-02 pc/g

COMMENTS

RESULTS ACCURATE TO 2 SIGNIFICANT FIGURES.  
UNCERTAINTY AT 95% CONFIDENCY LEVEL.

If you have any questions concerning the information provided above, please contact the ESOH Service Center at 1-888-232-2383.

ROBERT D. SCHMIDTGOESSLING, Major, USAF, BSC  
Chief, Analytical Laboratory Branch



Sample Analysis Results Reported on 07-Jan-2013

USAFSAM/OMHHL ID 11200286

Customer Address 01172

77 MDOS/SGPB  
5342 DUDLEY BLVD

MCCLELLAN AFB CA 95652-1074

IDENTIFICATION

Base Sample # CS00007 Serial #  
Date Collected 10/25/2012 Received 10/29/2012 Completed 12/20/2012

Analytes	Activity +/- Uncertainty	Lc / MDA
RADIUM 226	8.31E-01 +/- 4.89E-02 pci/g	4.11E-02 / 8.47E-02 pci/g

COMMENTS

RESULTS ACCURATE TO 2 SIGNIFICANT FIGURES.  
UNCERTAINTY AT 95% CONFIDENCY LEVEL.

If you have any questions concerning the information provided above, please  
contact the ESOH Service Center at 1-888-232-2353.

ROBERT D. SCHMIDTGOESSLING, Major, USAF, BSC  
Chief, Analytical Laboratory Branch

Sample Analysis Results Reported on 07-Jan-2013

USAFSAM/OEHHL ID 11200287  
Customer Address 01172  
77 MDOS/SGPB  
5342 DUDLEY BLVD

MCCLELLAN AFB CA 95652-1074

IDENTIFICATION

Base Sample # CS00008 Serial #  
Date Collected 10/25/2012 Received 10/29/2012 Completed 12/20/2012

Analytes	Activity +/- Uncertainty	Lc / MDA
RADIUM 226	1.51E+00 +/- 5.71E-02 pci/g	3.86E-02 / 7.92E-02 pci/g

COMMENTS

RESULTS ACCURATE TO 2 SIGNIFICANT FIGURES.  
UNCERTAINTY AT 95% CONFIDENCY LEVEL.

If you have any questions concerning the information provided above, please contact the ESOH Service Center at 1-888-232-2353.

ROBERT D. SCHMIDTGOESSLING, Major, USAF, BSC  
Chief, Analytical Laboratory Branch

ATTACHMENT 4

INSTRUMENT CALIBRATION SHEETS

Page 1 of 4



DEPARTMENT OF THE AIR FORCE  
 USAF SCHOOL OF AEROSPACE MEDICINE (AFMC)  
 OCCUPATIONAL ENVIRONMENTAL HEALTH/RADIATION HEALTH (OEHH)  
 WRIGHT-PATTERSON AFB OHIO  
**CERTIFICATE OF CALIBRATION**

Mfg. Ludlum Model 2221 Serial # 218606 Index # 099333 Date 23 Nov 11  
 Mfg. Ludlum Model 44-10 Serial # PR276614 Index # 100861 Cal. Due Date: 23 Nov 12

TEST, MEASUREMENT AND DIAGNOSTIC EQUIPMENT

NIST Traceable Check Sources				Reference Instruments			
Isotope	Serial #	Cal. Date	DPM	Mfg.	Model	Serial #	Cal. Due Date
Cs-137	RP3067	1 Nov 04	2,454,000	Ludlum	500-1	102951	8 Feb 2012

Measurement Standards and test equipment used are traceable to the National Institute of Standards and Technology, to the extent allowed by the Institute's calibration facilities.

- Battery Ck.  Mechanical Ck.  Meter Zeroed  Reset Ck.  Alarm Ck.  
 Audio Ck.  Geotropism Ck.  F/S Resp. Ck.  Window Op.

As Found HV 991 VDC Temperature 72.7 °F Relative Humidity 55 %  
 Final Volt. Set 100 VDC Threshold (LLD) 10 mV Window (ULD) 20 mV Window width 10 mV  
 HV Readout (2 points) Reference: 500 V Reference: 1000 V  
 Inst. Readout: 500 V ± 2% Inst. Readout: 1000 V ± 2%

RANGE MULTIPLIER	REFERENCE CAL. POINT	"AS FOUND" READING	CORRECTED READING
x 1000	400 CPM	400,000 CPM	400,000 CPM
x 1000	100 CPM	100,000 CPM	100,000 CPM
x 100	400 CPM	40,000 CPM	40,000 CPM
x 100	100 CPM	10,000 CPM	10,000 CPM
x 10	400 CPM	4,000 CPM	4,000 CPM
x 10	100 CPM	1,000 CPM	1,000 CPM
x 1	400 CPM	400 CPM	400 CPM
x 1	100 CPM	100 CPM	100 CPM
Log Scale	200 CPM	200 CPM	200 CPM

DIGITAL SCALER READOUT

CAL. REF. POINT	AS FOUND READING	CORRECTED READING
40,000 CPM	39,886 CPM	39,886 CPM

\*UNCERTAINTY WITHIN ± 10% CORRECTION FACTOR WITHIN ± 20%

COMMENTS: Calibration Interval = 1 year Use "Window Out"  
 Cs-137 Eff: 6,500 CPM/μCi/m<sup>2</sup>@12"  
 Procedural Authority - ICP#22210000

Calibrated By: STU Hutchinson

Date: 23 Nov 2011

Reviewed By: Piper Miller

Date: 30 Nov 11



DEPARTMENT OF THE AIR FORCE  
USAF SCHOOL OF AEROSPACE MEDICINE (AFMC)  
OCCUPATIONAL ENVIRONMENTAL HEALTH/RADIATION HEALTH (OEHH)  
WRIGHT-PATTERSON AFB OHIO  
CERTIFICATE OF CALIBRATION

Meter Mfg. Ludlum Model 2221 Serial # 218606 Index # 099333 Date: 23 Nov 11  
Cal. Due Date: 23 Nov 12

TEST, MEASUREMENT AND DIAGNOSTIC EQUIPMENT

NIST Traceable Check Sources

Isotope	Serial #	Cert. Date	DPM
Cs-137	RP3067	1 Nov 04	2,454,000

Reference Instruments

Mfg.	Model	Serial #	Cal. Due Date
Ludlum	500-1	102951	8 Feb 2012

Measurement Standards and test equipment used are traceable to the National Institute of Standards and Technology, to the extent allowed by the Institute's calibration facilities.

NaI DETECTOR HIGH VOLTAGE OPTIMIZATION

Probe #1

Mfg. Ludlum  
Model 44-10  
Serial # PR276614  
Index # 100861  
Isotope: Cs-137

High Voltage	CPM
750	7834
800	9964
850	11523
900	12240
950	12771
1000	12851
1050	13589
1100	13500
1150	13697
1200	13684
1250	13908
1300	13966
1350	17910
Bkg@ 1100	4664

Final Volt. Set 1100 VDC

Efficiency 6500 CPM/ $\mu$ cm<sup>2</sup>@12"

Probe #2

Mfg. \_\_\_\_\_  
Model \_\_\_\_\_  
Serial # \_\_\_\_\_  
Index # \_\_\_\_\_  
Isotope: \_\_\_\_\_

High Voltage	CPM

Final Volt. Set \_\_\_\_\_ VDC

Efficiency \_\_\_\_\_ CPM/ $\mu$ cm<sup>2</sup>@12"

Probe #3

Mfg. \_\_\_\_\_  
Model \_\_\_\_\_  
Serial # \_\_\_\_\_  
Index # \_\_\_\_\_  
Isotope: \_\_\_\_\_

High Voltage	CPM

Final Volt. Set \_\_\_\_\_ VDC

Efficiency \_\_\_\_\_ CPM/ $\mu$ cm<sup>2</sup>@12"

COMMENTS: Calibration Interval = 1 year      Use "Window LUT"

Calibrated By: Stu Hutchinson

Date: 23 Nov 2011

Reviewed By: Karl Will

Date: 30 Nov 11

HotSpot FIDLER Text File Output  
HotSpot FIDLER Calibration Information

Report Date : Nov 23 2011 07:07 AM  
Calibration Date : 23 Nov, 2011  
Target Mix : Other Nuclide Check Source  
Radionuclide : Cs-137  
Detector Barcode Number : 100861  
Meter Barcode Number : 099333  
Detector Manufacturer : Ludlum  
Detector Model Number : 44-10  
Detector Serial Number : PR276614  
Meter Manufacturer : Ludlum  
Meter Model Number : 2221  
Meter Serial Number : 218606

Check Source I.D. : RP 3067  
Calibration Date : 23 Nov, 2011  
Calibrated by : Stu Hutchinson  
Check Source Activity (uCi) : 1.100E+00  
Check Source 17-keV Self : 1.000E+00

Sample Counting Time (minutes) : 1.000E+00  
Detector Height (cm) : 3.000E+01

Cs-137 window information:  
Background (cpm) : 4,669  
Areal Limit of Sensitivity (uCi/m2) : 4.9E-02  
Point Limit of Sensitivity (uCi) : 1.0E-01  
K-factor (m2) : 2.07

Counting Data (counts):  
0-cm: 8120  
20-cm: 7179  
40-cm: 6081  
60-cm: 5469  
80-cm: 5167  
100-cm: 5069

Instrument Type : other  
Window Option: Only 60 keV  
Units: Classic


This is an actual 2 x 2 calibration and the values are typical of most 2 x 2 configurations.

Detector Calibration Results

Cs-137 window information:  
Cs-137 Detector Efficiency (cpm/(uCi/m2)): 6.5E+03  
Cs-137 Detector Areal LOS (uCi/m2) : 4.9E-02  
Cs-137 Detector Point LOS (uCi) : 1.0E-01  
Cs-137 Detector Background Rate (cpm) : 4,669  
Cs-137 Detector Check Source Rate (cpm) : 3,451  
Cs-137 Detector K-Factor (m2) : 2.07  
Cs-137 Detector K-Factor sdev (%) : 7.5

Cs-137 Eff: 6,500 CPM/uCi/m<sup>2</sup> @ 12"

VDC	Cs-137 CPM	DU Slug	Ludlum 2221	SN 218606
			Ludlum 44-10	SN 276614
750	7834	7417		
800	9969	10456		
850	11523	13713		
900	12240	16030		
950	12771	16869		
1000	12851	17336		
1050	13589	17835		
1100	13500	17515		
1150	13697	17816		
1200	13684	17764		
1250	13908	17946		
1300	13966	18133		
1350	17910	18023		
1400		18159		
1450		19730		
1500				
1550				
1600				





**DEPARTMENT OF THE AIR FORCE  
USAF SCHOOL OF AEROSPACE MEDICINE (AFMC)  
OCCUPATIONAL ENVIRONMENTAL HEALTH/RADIATION HEALTH (OEHH)  
WRIGHT-PATTERSON AFB OHIO  
CERTIFICATE OF CALIBRATION**

Mfg. Ludlum Model 2221 Serial # 78153 Index # 04125 Date: 25 Nov 11  
 Mfg. Ludlum Model 44-10 Serial # PR276615 Index # 100862 Cal. Due Date: 25 Nov 12

**TEST, MEASUREMENT AND DIAGNOSTIC EQUIPMENT**

NIST Traceable Check Sources				Reference Instruments			
Isotope	Serial #	Cert. Date	DPM	Mfg.	Model	Serial #	Cal. Due Date
Cs-137	RP3067	1 Nov 04	2,454,000	Ludlum	500-1	102951	8 Feb 2012

Measurement Standards and test equipment used are traceable to the National Institute of Standards and Technology, to the extent allowed by the Institute's calibration facilities.

- Battery Ck.  
  Mechanical Ck.  
  Meter Zeroed  
  Reset Ck.  
  Alarm Ck.  
 Audio Ck.  
  Geotropism Ck.  
  FIS Resp. Ck.  
  Window Op.

As Found HV 703 VDC    Temperature 73.1 °F    Relative Humidity 54.6 %

Final Volt. Set 1250 VDC    Threshold (LLD) 10 mV    Window (ULD) 20 mV    Window width 10 mV

HV Readout (2 points)    Reference: 500 V    Reference: 1000 V  
 Inst. Readout: 499 V ± 2%    Inst. Readout: 1000 V ± 2%

RANGE MULTIPLIER	REFERENCE CAL. POINT	"AS FOUND" READING	CORRECTED READING
x 1000	400 CPM	400,000 CPM	400,000 CPM
x 1000	100 CPM	100,000 CPM	100,000 CPM
x 100	400 CPM	40,000 CPM	40,000 CPM
x 100	100 CPM	10,000 CPM	10,000 CPM
x 10	400 CPM	4,000 CPM	4,000 CPM
x 10	100 CPM	1,000 CPM	1,000 CPM
x 1	400 CPM	400 CPM	400 CPM
x 1	100 CPM	100 CPM	100 CPM
Log Scale	200 CPM	199 CPM	199 CPM

**DIGITAL SCALER READOUT**

CAL. REF. POINT	AS FOUND READING	CORRECTED READING
40,000 CPM	39,889 CPM	39,889 CPM

\*UNCERTAINTY WITHIN ± 10% CORRECTION FACTOR WITHIN ± 20%

COMMENTS: Calibration Interval = 1 year    Use "Window Dot"  
 Cs-137 Eff. 2,900 CPM/μCi/m<sup>2</sup> @ 12"  
 DU-238 response curve determined.    Procedural Authority - ICP#22210000

Calibrated By: Steve Hutchinson  
 Reviewed By: Piperville

Date: 25 Nov 2011  
 Date: 30 Nov 11



**DEPARTMENT OF THE AIR FORCE  
USAF SCHOOL OF AEROSPACE MEDICINE (AFMC)  
OCCUPATIONAL ENVIRONMENTAL HEALTH/RADIATION HEALTH (OEHH)  
WRIGHT-PATTERSON AFB OHIO  
CERTIFICATE OF CALIBRATION**

Meter Mfg. Ludlum Model 2221 Serial # 78153 Index # 04125 Cal. Due Date: 25 Nov 11  
Date: 25 Nov 11

**TEST, MEASUREMENT AND DIAGNOSTIC EQUIPMENT**

NIST Traceable Check Sources				Reference Instruments			
Isotope	Serial #	Cert. Date	DPM	Mfg.	Model	Serial #	Cal. Due Date
Cs-137	RP3067	1 Nov 04	2,454,000	Ludlum	500-1	102951	8 Feb 2012

Measurement Standards and test equipment used are traceable to the National Institute of Standards and Technology, to the extent allowed by the Institute's calibration facilities.

**NaI DETECTOR HIGH VOLTAGE OPTIMIZATION**

Probe #1  
Mfg. Ludlum  
Model 44-10  
Serial # PR276615  
Index # 100862  
Isotope: Cs-137

Probe #2  
Mfg. Ludlum  
Model 44-10  
Serial # PR276615  
Index # 100862  
Isotope: DU-slug

Probe #3  
Mfg. \_\_\_\_\_  
Model \_\_\_\_\_  
Serial # \_\_\_\_\_  
Index # \_\_\_\_\_  
Isotope: \_\_\_\_\_

High Voltage	CPM
900	11393
950	12953
1000	13581
1050	14078
1100	14215
1150	15057
1200	15173
1250	15017
1300	15267
1350	15501
1400	15212
1450	15503
Bkgd @ 1250	4532

High Voltage	CPM
900	10618
950	13590
1000	16229
1050	17132
1100	17581
1150	17881
1200	18157
1250	18419
1300	18455
1350	18452
1400	18499
HHS	

High Voltage	CPM

Final Volt. Set 1250 VDC      Final Volt. Set 1250 VDC      Final Volt. Set \_\_\_\_\_ VDC  
Efficiency 2960 CPM/μm<sup>2</sup>@12"      Efficiency NA CPM/μm<sup>2</sup>@12"      Efficiency \_\_\_\_\_ % 2π@1"

COMMENTS: Calibration Interval = 1 year      Use "Window Bot"  
DU-238 response curve determined.

Calibrated By: Stu Hutchinson      Date: 25 Nov 2011  
Reviewed By: Peter Mills      Date: 30 Nov 11



HotSpot FIDLER Calibration Information  
HotSpot FIDLER Text File Output

Report Date : Nov 23 2011 11:12 AM  
Calibration Date : 23 Nov, 2011  
Target Mix : Other Nuclide Check Source  
Radionuclide : Cs-137  
Detector Barcode Number : 100862  
Meter Barcode Number : 04125  
Detector Manufacturer : Ludlum  
Detector Model Number : 44-10  
Detector Serial Number : PR276615  
Meter Manufacturer : Ludlum  
Meter Model Number : 2221  
Meter Serial Number : 78153

Check Source I.D. : RP 3067  
Calibration Date : 23 Nov, 2011  
Calibrated by : Stu Hutchinson  
Check Source Activity (uCi) : 1.000E+00  
Check Source 17-keV Self : 1.000E+00

Sample Counting Time (minutes) : 1.000E+00  
Detector Height (cm) : 3.000E+01

Cs-137 Window Information:  
Background (cpm) : 4,532  
Areal Limit of Sensitivity (uCi/m2) : 1.1E-01  
Point Limit of Sensitivity (uCi) : 9.3E-02  
K-factor (m2) : 0.84

Counting Data (counts):  
0-cm: 8279  
20-cm: 7135  
40-cm: 6023  
60-cm: 5398  
80-cm: 5067  
100-cm: 4747

Instrument Type : other  
Window Option: Only 60 keV  
Units: Classic


This is an actual 2 x2 calibration and the values are typical of most 2 x2 configurations.

Detector Calibration Results

Cs-137 Window Information:  
Cs-137 Detector Efficiency (cpm/(uCi/m2)) : 2.9E+03  
Cs-137 Detector Areal LOS (uCi/m2) : 1.1E-01  
Cs-137 Detector Point LOS (uCi) : 9.3E-02  
Cs-137 Detector Background Rate (cpm) : 4,532  
Cs-137 Detector Check Source Rate (cpm) : 3,747  
Cs-137 Detector K-Factor (m2) : 0.84  
Cs-137 Detector K-Factor sdev (%) : 7.4

Cs-137 Eff: 2900 CPM/uCi/m<sup>2</sup> @ 12"

VDC	Cs-137 CPM	DU Slug	Ludlum 2221	SN78153			
			Ludlum 44-10	SN276615			
750		3301					
800		5205					
850		7923					
900	11393	10618					
950	12953	13590					
1000	13581	16229					
1050	14078	17132					
1100	14215	17581					
1150	15057	17881					
1200	15173	18157					
1250	15017	18419					
1300	15267	18455					
1350	15501	18452					
1400	15212	18499					
1450	15503						
1500							
1550							
1600							



ATTACHMENT 5

RADIATION METER QC LOG

Radiation Meter QC Log

Model	S/N	Date/Time	HV/cables/Bat check	Source Check Reading	Acceptable Range
2221 <del>2222</del>	78153 PR 276615	22oct12/1800	✓	21772 cpm	17418 - 26126 cpm
2221 222	218606 PR 276614	22oct12/1500	✓	20476 cpm	16381 - 24571 cpm
2221 222	78153 PR 276615	23oct12/1400	✓	20283 cpm	17418 - 26126 cpm
2221 222	218606 PR 276614	23oct12/1400	✓	19821 cpm	16381 - 24571 cpm
2221 222	78153 PR 276615	23oct12/1446	✓	19839 cpm	17418 - 26126 cpm
2221 222	218606 PR 276614	23oct12/1447	✓	19210 cpm	16381 - 24571 cpm
2221 222	78153 PR 276615	24oct12/1300	✓	22047 cpm	17418 - 26126 cpm
2221 222	218606 PR 276614	24oct12/1300	✓	19763 cpm	16381 - 24571 cpm
2221 222	78153 PR 276615	24oct12/1529	✓	21346 cpm	17418 - 26126 cpm
2221 222	218606 PR 276614	24oct12/1532	✓	656 cpm	16381 - 24571 cpm
2221 222	78153 PR 276615	25OCT12/1300	✓	19,572 cpm	17418 - 26,126 cpm
2221 222	78153 PR 276615	25OCT12/1433	✓	21,413 cpm	17418 - 26,126 cpm

★ The past check did not meet acceptable range. It affects the crystal connection to PMT or diodes are bad. Data collected with this meter will be closely evaluated to see if is usable.