REPORT DO	CUMENT	ATION PAG	E		Form Approved
Public reporting burden for th	nis collection of information	is estimated to average 1	hour per response, including t	he time for reviewing ins	OMB No. 0704-0188 tructions, searching existing data sources, gathering and
maintaining the data needed	, and completing and revie	wing this collection of infor	mation. Send comments rega	rding this burden estima	te or any other aspect of this collection of information, including ons and Reports (0704-0188), 1215 Jefferson Davis Highway, Suite
1204, Arlington, VA 22202-4	1302. Respondents should	be aware that notwithstan	ding any other provision of law	v, no person shall be sub	pject to any penalty for failing to comply with a collection of
1. REPORT DATE (L		2. REPO	OO NOT RETURN YOUR FOI RT TYPE	KM TO THE ABOVE AD	3. DATES COVERED (From – To)
31 Mar 2015	,	Consulta	tive Letter		July 2014 – December 2014
4. TITLE AND SUBT	TTLE	,			5a. CONTRACT NUMBER
	(CS) 024 Verificat	tion Survey at For	mer McClellan AFI	3, Sacramento,	5b. GRANT NUMBER
California					5c. PROGRAM ELEMENT NUMBER
6. AUTHOR(S)					5d. PROJECT NUMBER
Matthew W. Uelen					
					5e. TASK NUMBER
					5f. WORK UNIT NUMBER
7. PERFORMING OF			SS(ES)		8. PERFORMING ORGANIZATION REPORT NUMBER
USAF School of A					NUMBER
Occupational & En Consultative Service		th Dept			AFRL-SA-WP-CL-2015-0012
2510 Fifth St.	Les Division				2010 0012
Wright-Patterson A	FB. OH 45433-7	913			
Wilght Luncison I	22, 011 10 100 7	, 10			
9. SPONSORING / N	ONITORING AGE	NCY NAME(S) AND	ADDRESS(ES)		10. SPONSORING/MONITOR'S ACRONYM(S)
					11. SPONSOR/MONITOR'S REPORT NUMBER(S)
12. DISTRIBUTION	AVAILABILITY CT	ATCMENT			
12. DISTRIBUTION	AVAILABILITT ST	AIEWENI			
Distribution A: Ap	proved for public	release; distributi	ion is unlimited. Ca	se Number: 88.	ABW-2015-1641, 31 Mar 2015
1	1	,			,
13. SUPPLEMENTA	RY NOTES				
14. ABSTRACT	a IIS Air Force	Padioisotona Con	mittaa Sacratariat (DICS) the U.S.	Air Force School of Aerospace Medicine,
					performed an independent radiological
					, California, on 29-31 July 2014.
					nine if radiation levels complied with criteria
					unrestricted release. Radium-226 was the
					n, performed a Final Status Survey (FSS)
			This letter details to	he findings of th	is visit and will assist the RICS with
evaluating the cont	ractor's FSS repo	rt of this site.			
45 SUBJECT TERM	ıe				
15. SUBJECT TERM		e (USAFSAM) f	ormer McClellan Al	FR radium-226	verification survey, final status survey,
independent radiole			ormer inteciental A	, radium 220,	, remieuton survey, mui suitus survey,
16. SECURITY CLAS			17. LIMITATION	18. NUMBER	19a. NAME OF RESPONSIBLE PERSON
			OF ABSTRACT	OF PAGES	Col Krystal L. Murphy
a. REPORT	b. ABSTRACT	c. THIS PAGE	CAD	40	19b. TELEPHONE NUMBER (include area
U	U	U	SAR	40	code)



DEPARTMENT OF THE AIR FORCEUSAF SCHOOL OF AEROSPACE MEDICINE (AFMC) WRIGHT-PATTERSON AFB OH

31 March 2015

MEMORANDUM FOR AFMSA/SG3PB

ATTN: MAJ DANIEL SHAW
USAF RADIOISOTOPE COMMITTEE SECRETARIAT
AIR FORCE MEDICAL SUPPORT AGENCY
7700 ARLINGTON BOULEVARD, SUITE 5158
FALLS CHURCH VA 22042-5158

FROM: USAFSAM/OEC 2510 Fifth Street

Wright-Patterson AFB OH 45433-7913

SUBJECT: Consultative Letter, AFRL-SA-WP-CL-2015-0012, Consolidated Site (CS) 024 Verification Survey at Former McClellan AFB, Sacramento, California

1. INTRODUCTION:

a. *Purpose*: At the request of the U.S. Air Force Radioisotope Committee Secretariat (RICS), the U.S. Air Force School of Aerospace Medicine, Consultative Services Division (USAFSAM/OEC), Radiation Health Consulting Branch performed an independent radiological verification survey of the CS 024 hazardous waste site, located on former McClellan AFB, California, on 29-31 July 2014. The purpose of this verification survey was to provide an independent evaluation of the final radiological conditions at a decommissioned site. USAFSAM/OEC surveyed the CS 024 site to measure residual surface activity and determine if radiation levels complied with criteria agreed upon between representatives of the State of California and the U.S. Air Force for unrestricted release. Radium-226 (Ra-226) was the sole radionuclide of concern. Cabrera Services, Inc., under contract with URS Corporation, performed a Final Status Survey (FSS) and all radiological fieldwork previously in July 2014. This letter details the finding of this visit and is intended to assist the RICS with evaluating the contractor's FSS report of this site. Personnel utilized techniques and procedures taken from Nuclear Regulatory Guide 1575, *Multi-Agency Radiation Survey and Site Investigation Manual (MARSSIM)*, Rev 1.

b. *Background:* Site CS 024 is an inactive disposal burn pit that became operative in 1953. Between 1953 and 1969, base personnel used the site for burning and burial of demolition debris and scrap material. Workers frequently used the eastern half of the pit to burn Air Force generated wastes; workers used the western half to discard unburned waste material. Cleanup personnel have detected a wide variety of contaminants at the site to include volatile organic compounds, polychlorinated biphenyls, dioxins/furans, and metals. The spatial distribution of radioactive materials across the site, namely Ra-226, varies. This is due to the variety of wastes burned and stored on the site. Site CS 024 is located in the far southern border of the McClellan Park boundary, parallel to a rail line that runs along the southeastern edge of the site. At the time of this survey, the site had not yet undergone an FSS. Cabrera Services began surveying the sides of the trench after USAFSAM/OEC personnel completed the verification survey. Prior to this survey, Cabrera Services had previously completed cleanup operations on the trench floor. See Figure 1 below.



Figure 1. View of CS 024

c. Survey Personnel:

- (1) Maj. Marcus Grant, Health Physicist, USAFSAM/OEC
- (2) SSgt Michael Ames, Lead Health Physics Technician, USAFSAM/OEC

d. Personnel Contacted:

- (1) Radiation Safety Officer, AFCEC/CIBW
- (2) Base Realignment and Closure Program Management/Western Region, AFCEC/CIBW
- (3) Radiation Project Manager, Cabrera Services, Inc.

e. Equipment:

- (1) Ludlum Model 2221 with Scionix 76BRS76/2ME1-X 3x3 Sodium Iodide Detector, Meter SN 287511, Detector SN SAG420, Calibration date 16 Jun 2014
- (2) Ludlum Model 2221 with Scionix 76BRS76/2ME1-X 3x3 Sodium Iodide Detector, Meter SN 290803, Detector SN SAG427, Calibration date 3 Jun 2014

2. METHODOLOGY:

- a. *Overview:* The verification survey consisted of soil sample collection and gamma walkover measurements. The goal of the walkover survey was to identify locations for soil sampling. This verification survey required scanning at least 10% of the trench floor by the gamma walkover technique and soil sample collection totaling at least 10% of the number taken by Cabrera Services. Prior to the survey, Cabrera Services identified 40 soil sampling locations; therefore, at least 4 soil samples are necessary to meet the 10% minimum requirement. Survey personnel used the gamma walkover data to qualitatively assess the site in terms of mean count rate and standard deviation. Minimum detectable concentration and the associated count rates were not calculated.
- b. *Soil Sampling:* Each sample was taken from an area of approximately 8 square inches to a depth of 6 inches. The volume sampled was sufficient for laboratory analysis (approximately 1 quart). Field soil sampling procedures were in place to prevent cross contamination of samples. Personnel collected a background sample from uncontaminated soil in an adjacent area to the CS 024 site. The USAFSAM Radioanalytical Laboratory (OEA) at Wright-Patterson AFB, Ohio, analyzed soil samples by counting on a high purity germanium detector, using the appropriate ingrowth method to determine Ra-226 levels in soil. Prior to the 28-day in-growth period, OEA sifted the soil samples for any rocks and large debris and then dried to remove any moisture potentially affecting analysis. Attachment 1 contains laboratory analysis for the samples. An "x" within Attachment 2 indicates the locations of soil samples.
- c. Gamma Walkover Survey: Scionix Model 76BRS76/3M-E1-X 3x3-inch thallium-doped sodium iodide (NaI) detectors and Ludlum Model 2221 ratemeter/scalers were the primary radiation detection instruments utilized to perform the gamma walkover survey. Survey personnel connected the instruments to a Trimble GeoXT handheld GPS unit via RS-232 serial cable. The Trimble GeoXT logged observed count rates with corresponding GPS coordinates every second. Survey personnel held the detectors at a height of 10 cm above the ground. The scan speed was approximately 0.5 meters per second, with the scan lines spaced about 0.5 meters apart. All field instruments were function checked at the start and end of each workday using a gamma-emitting check source. Personnel tested each instrument to ensure a 10% tolerance during field checks. Qualified staff calibrate the instruments annually. Attachment 3 contains the annual calibration certificates for all instruments used during this survey. Attachment 4 contains the quality control logs used to document instrument performance with a check source both before and after use.

3. RESULTS:

- a. Gamma Walkover Survey: The scan coverage exceeded the minimum scan requirement of 10% of the trench floor. Some areas were inaccessible using the walkover technique, notably in areas that presented a climbing hazard. The team did not survey the sides of the trench. Personnel used the mean count rate as the background for the scanned area and did not survey a reference area. USAFSAM/OEC personnel calculated the mean and standard deviation of the dataset based on the instrument and the particular area surveyed. In some cases, USAFSAM/OEC also excluded data outside the background distribution (e.g., outlying data points, data generated due to software anomalies or equipment malfunctions). Personnel then compiled the data using commercial geospatial mapping software. The locations marked in green correspond to the lowest readings of the survey. Given the data collected, green dots represent all measured activity points below 2 standard deviations (SD) above the mean value [less than 26981 counts per minute (cpm)]. Yellow dots correspond to data points between 2 and 3 SD above the mean (between 26988 and 28649 cpm). Red dots represent all data points above 3 SD (greater than 28650 cpm). Survey personnel expect higher concentrations in areas marked in red. The color scheme demonstrates a scale of instrument data based upon SDs from background radiation levels, where no regulatory values are implied. A table of statistics compiled from the dataset can be found in Attachment 5.
- b. *Soil Sampling*: Survey personnel collected five soil samples from CS 024: four samples within the permitted site boundary and one sample outside of CS 024 in an area of soil to be used as backfill for the site. The survey team used this backfill sample to compare to background radiation levels. None of the samples collected exceeded the cleanup goal of 2.0 pCi/g. Table 1 below details both the gross and background subtracted soil results and the associated GPS coordinates for each sample. When comparing results to the cleanup level of 2.0 pCi/g, utilize the gross (background included) results listed below. The uncertainties listed are at the 95% confidence level (2 SD). Laboratory staff determines concentrations of Ra-226 in the soil by gamma spectroscopy analysis of Ra-226 and the associated bismuth-214 / lead-213 daughters in secular equilibrium. The Ra-226 concentrations in the table below are labeled as "Ra_D_214" within Attachment 1.

Table 1. Soil Sample Results for CS 024

Client Sample ID	GPS Coordinates (WGS1984, Decimal Degrees, N/E)	Gross Ra-226 Concentration (pCi/g) ¹	Net Ra-226 Concentration above Background (pCi/g)
GS-14-0010	38.642017258, -121.402916535	0.58 ± 0.04	N/A
GS-14-0011	38.641801046, -121.403217583	1.07 ± 0.05	0.31 ± 0.06
GS-14-0012	38.641697144, -121.403497604	0.76 ± 0.04	0.00 ± 0.06
GS-14-0013	38.641369635, -121.403706972	0.75 ± 0.04	N/A
GS-14-0014 ²	38.640995840, -121.403987551	0.76 ± 0.04	N/A

¹Ra-226 concentrations with associated uncertainties from soil sample results.

² Background sample for CS 024.

4. DISCUSSION:

- a. *Gamma Walkover Survey:* An area outside the permitted area was surveyed using the gamma walkover technique to verify that contamination did not migrate outside of the controlled area. The surface was primarily asphalt, whereas the bulk of the walkover data collected for the trench was dirt. Consequently, the background counts observed in this location are different from the trench, approximately 6100 cpm less on average.
- b. *Soil Sampling:* The highest recorded soil sample result for CS 024 was 1.07 pCi/g for gross Ra-226. Accordingly, gamma walkover scans also indicated an increase in gamma count rate near this sample. The highest recorded count rate in this area was 30,666 cpm. The mean count rate observed within the entire area of CS 024 (excluding data recorded within the asphalt area) was 23,616 cpm, with an SD of 1,836 cpm.

5. CONCLUSIONS AND RECOMMENDATIONS:

- a. Based on the results of the laboratory analysis, none of the areas surveyed exceeded the cleanup goal of 2.0 pCi/g. Laboratory analysis is the primary indicator of site conditions for regulatory decisions.
- b. USAFSAM/OEC detected no areas of elevated radiological activity above background that would require additional investigative actions using the gamma walkover technique.
- c. USAFSAM/OEC recommends the survey unit be considered for unrestricted free release of radiological controls, with the approval of the RICS.
- d. If you have any questions regarding this report, please contact the ESOH Service Center at (888) 232-3764 (DSN 798-3764) or ESOH.Service.Center@us.af.mil.

MATTHEW W. UELEN USAFSAM Health Physicist

be le leken

5 Attachments:

- 1. Laboratory Soil Analysis
- 2. Map of Gamma Walkover with Soil Sample Locations
- 3. Instrument Calibration Sheets
- 4. Radiation Meter Quality Control (QC) Log
- 5. Statistical Summary of Instrument Data

Attachment 1 Laboratory Soil Analysis





MEMORANDUM FOR: SSgt Michael Ames USAFSAM/OEC 2510 Fifth St. Bldg. 840

WRIGHT-PATTERSON AFB, OH 45433

FROM: USAFSAM OE Industrial Hygiene 2510 Fifth Street, Bldg 20840, Room W327 Wright Patterson Air Force Base, OH 45433-7913

REF: Order No.: S1408005 Dear SSgt Michael Ames:

Enclosed are the sample reports from 14 samples received on 8/5/2014.

Samples, not consumed in analysis, will be held according to the appropriate regulatory authority unless you specifically request otherwise. Should you choose to reproduce this report, we recommend you do so in its entirety so that the integrity of the data package is kept intact.

If you have questions, or if we may be of further assistance to you, please do not hesitate to contact us.

Sincerely,

ERIC L WEATHERHOLT, Capt, USAF Analytical Services Program Manager Tel: (937) 938-2523 (DSN Prefix: 798) https://hpws.afrl.af.mil/dhp/OE/ESOHSC/pages/index.cfm?id=742

Note: Sample analysis performed by: USAFSAM/OE Radioanalytical Division

This report is intended solely for the purpose of the person to whom it is addressed. If received in error, please notify the Program Manager listed above.

Page 1 of 17 Pages

USAFSAM OE Industrial Hygiene

CLIENT: USAFSAM/OEC

Project: CASE NARRATIVE

Lab Work Order: S1408005

There were no problems associated with the samples or analysis except where noted below. Unless otherwise noted, sample results are not blank corrected, and all quality control associated with the samples were within acceptable limits.

These results relate only to the items tested.

Customer requested Ra226 analysis via gamma spectrometry. Analysis report will show results for Ra_D_214. This result is equivalent to Ra226 plus the ingrowth of Bi214/Pb214 daughters using all higher energy, high yield energy lines. Ra_D_214 = Ra226 via gamma spec analysis with ingrowth.

Sample Preparation Comments:

SampID	TestCode	Comments
S1408005-01A	RAD_GAMMA_SOLID_	Ingrowth Time: 6-Aug-14 16:15
S1408005-02A	RAD_GAMMA_SOLID_	Ingrowth Time: 6-Aug-14 16:15
S1408005-03A	RAD_GAMMA_SOLID_	Ingrowth Time: 6-Aug-14 16:15
S1408005-04A	RAD_GAMMA_SOLID_	Ingrowth Time: 6-Aug-14 16:15
S1408005-05A	RAD_GAMMA_SOLID_	Ingrowth Time: 6-Aug-14 16:15
S1408005-06A	RAD_GAMMA_SOLID_	Ingrowth Time: 6-Aug-14 16:15
S1408005-07A	RAD GAMMA SOLID	Ingrowth Time: 6-Aug-14 16:15

Page 2 of 17 Pages



Certificate of Analysis

Company: USAFSAM/OEC

Address: 2510 Fifth St. Bldg 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140001

 Lab Sample ID:
 S1408005-01A

Matrix: Soil Client ID: 02060 C

 Collection Date:
 7/29/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

D	O116	A -41-44	TT	т.	MDA	TT34	A14	D-4-/Ti	D -4 -1-
Parameter	Qualifier	Activity	Uncertainty	Le	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec I	Full								
U-235	G,N	9.07E-02	2.82E-02	2.20E-02	4.48E-02	pCi/g	JRO	09/05/14 1241	7630
TL-208	G,N	3.88E-01	3.46E-02	2.13E-02	4.38E-02	pCi/g	JRO	09/05/14 1241	7630
TH-234	G,N	1.15E+00	8.12E-01	9.62E-01	1.95E+00	pCi/g	JRO	09/05/14 1241	7630
TH-228	G,N	1.68E+00	8.11E-01	1.21E+00	2.47E+00	pCi/g	JRO	09/05/14 1241	7630
PB-212	G,N	1.46E+00	5.98E-02	2.67E-02	5.46E-02	pCi/g	JRO	09/05/14 1241	7630
BI-212	G,N	1.35E+00	3.09E-01	2.20E-01	4.59E-01	pCi/g	JRO	09/05/14 1241	7630
AC-228	G,N	1.26E+00	7.51E-02	7.17E-02	1.49E-01	pCi/g	JRO	09/05/14 1241	7630
RA-224		9.58E-01	2.91E-01	2.67E-01	5.46E-01	pCi/g	JRO	09/05/14 1241	7630
Ra_D_214		8.64E-01	4.41E-02	3.68E-02	7.58E-02	pCi/g	JRO	09/05/14 1241	7630
K-40		1.38E+01	9.03E-01	2.57E-01	5.35E-01	pCi/g	JRO	09/05/14 1241	7630

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 3 of 17 Pages



Certificate of Analysis

Company: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140002

 Lab Sample ID:
 S1408005-02A

Matrix: Soil Client ID: 02060 C

 Collection Date:
 7/29/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Lc	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec I	'ull								
TL-208	G,N	2.32E-01	2.89E-02	1.93E-02	3.97E-02	pCi/g	JRO	09/05/14 1842	7630
TH-234	G,N	1.94E+00	7.29E-01	9.86E-01	2.00E+00	pCi/g	JRO	09/05/14 1842	7630
PB-212	G,N	8.80E-01	4.41E-02	2.80E-02	5.71E-02	pCi/g	JRO	09/05/14 1842	7630
BI-212	G,N	8.89E-01	2.82E-01	2.23E-01	4.63E-01	pCi/g	JRO	09/05/14 1842	7630
AC-228	G,N	7.29E-01	7.34E-02	6.90E-02	1.43E-01	pCi/g	JRO	09/05/14 1842	7630
RA-224		7.66E-01	2.55E-01	2.90E-01	5.92E-01	pCi/g	JRO	09/05/14 1842	7630
Ra_D_214		6.24E-01	3.47E-02	3.33E-02	6.87E-02	pCi/g	JRO	09/05/14 1842	7630
K-40		2.08E+01	1.17E+00	2.15E-01	4.49E-01	pCi/g	JRO	09/05/14 1842	7630

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 4 of 17 Pages



Certificate of Analysis

ompany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140003

 Lab Sample ID:
 S1408005-03A

Matrix: Soil Client ID: 02060 C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Lc	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec F	'ull								
U-235	G,N	4.38E-02	2.77E-02	2.26E-02	4.59E-02	pCi/g	JRO	09/06/14 044	7630
TL-208	G,N	2.35E-01	2.75E-02	1.68E-02	3.48E-02	pCi/g	JRO	09/06/14 044	7630
TH-234	G,N	8.92E-01	7.02E-01	8.91E-01	1.81E+00	pCi/g	JRO	09/06/14 044	7630
PB-212	G,N	8.10E-01	4.21E-02	2.82E-02	5.76E-02	pCi/g	JRO	09/06/14 044	7630
BI-212	G,N	7.30E-01	2.86E-01	2.17E-01	4.52E-01	pCi/g	JRO	09/06/14 044	7630
AC-228	G,N	7.01E-01	6.74E-02	6.85E-02	1.43E-01	pCi/g	JRO	09/06/14 044	7630
RA-224		5.10E-01	2.38E-01	2.81E-01	5.75E-01	pCi/g	JRO	09/06/14 044	7630
Ra_D_214		5.02E-01	3.32E-02	3.71E-02	7.61E-02	pCi/g	JRO	09/06/14 044	7630
K-40		1.94E+01	1.12E+00	2.04E-01	4.29E-01	pCi/q	JRO	09/06/14 044	7630

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 5 of 17 Pages



Certificate of Analysis

ompany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140004

 Lab Sample ID:
 S1408005-04A

Matrix: Soil Client ID: 02060 C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Lc	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec I	full								
U-235	G,N	7.08E-02	3.02E-02	2.43E-02	4.94E-02	pCi/g	JRO	09/06/14 645	7630
TL-208	G,N	2.85E-01	3.02E-02	1.87E-02	3.86E-02	pCi/g	JRO	09/06/14 645	7630
TH-234	G,N	1.20E+00	7.38E-01	8.78E-01	1.79E+00	pCi/g	JRO	09/06/14 645	7630
TH-228	G,N	2.91E+00	7.97E-01	1.09E+00	2.21E+00	pCi/g	JRO	09/06/14 645	7630
PB-212	G,N	9.93E-01	4.96E-02	2.76E-02	5.64E-02	pCi/g	JRO	09/06/14 645	7630
PA-234M	G,N	5.92E+00	4.23E+00	4.31E+00	9.15E+00	pCi/g	JRO	09/06/14 645	7630
BI-212	G,N	8.43E-01	4.66E-01	3.70E-01	7.59E-01	pCi/g	JRO	09/06/14 645	7630
AC-228	G,N	8.85E-01	6.84E-02	6.86E-02	1.43E-01	pCi/g	JRO	09/06/14 645	7630
EU- 15 4	E,N	1.12E-01	8.77E-02	1.11E-01	2.36E-01	pCi/g	JRO	09/06/14 645	7630
CD-109	D, F, N	5.78E-01	2.94E-01	3.70E-01	7.53E-01	pCi/g	JRO	09/06/14 645	7630
RA-224		6.38E-01	2.61E-01	2.68E-01	5.49E-01	pCi/g	JRO	09/06/14 645	7630
Ra_D_214		6.98E-01	3.62E-02	3.10E-02	6.42E-02	pCi/g	JRO	09/06/14 645	7630
K-40		1.39E+01	9.00E-01	2.51E-01	5.21E-01	pCi/g	JRO	09/06/14 645	7630

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 6 of 17 Pages



Certificate of Analysis

ompany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140005

 Lab Sample ID:
 S1408005-05A

Matrix: Soil Client ID: 02060C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Lc	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec I	Full								
U-235	G,N	1.01E-01	3.45E-02	2.74E-02	5.56E-02	pCi/g	JRO	09/06/14 1247	7630
TL-208	G,N	3.34E-01	4.62E-02	3.83E-02	7.78E-02	pCi/g	JRO	09/06/14 1247	7630
TH-234	G,N	2.36E+00	8.38E-01	9.97E-01	2.03E+00	pCi/g	JRO	09/06/14 1247	7630
TH-228	G,N	2.87E+00	8.53E-01	1.03E+00	2.10E+00	pCi/g	JRO	09/06/14 1247	7630
PB-212	G,N	1.17E+00	5.80E-02	2.84E-02	5.80E-02	pCi/g	JRO	09/06/14 1247	7630
BI-212	G,N	1.18E+00	3.40E-01	2.40E-01	4.99E-01	pCi/g	JRO	09/06/14 1247	7630
AC-228	G,N	1.05E+00	7.18E-02	7.10E-02	1.48E-01	pCi/g	JRO	09/06/14 1247	7630
CD-109	F,N	4.71E-01	5.24E-01	4.20E-01	8.55E-01	pCi/g	JRO	09/06/14 1247	7630
RA-224		6.38E-01	2.83E-01	2.99E-01	6.13E-01	pCi/g	JRO	09/06/14 1247	7630
Ra_D_214		7.71E-01	3.88E-02	3.78E-02	7.76E-02	pCi/g	JRO	09/06/14 1247	7630
K-40		1.48E+01	9.48E-01	2.29E-01	4.79E-01	pCi/g	JRO	09/06/14 1247	7630

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 7 of 17 Pages



Certificate of Analysis

mpany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140006

 Lab Sample ID:
 S1408005-06A

Matrix: Soil Client ID: 02060 C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Lc	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec I	'ull								
U-235	G,N	6.42E-02	3.09E-02	2.49E-02	5.06E-02	pCi/g	JRO	09/06/14 1848	7630
TL-208	G,N	3.40E-01	3.44E-02	2.07E-02	4.28E-02	pCi/g	JRO	09/06/14 1848	7630
TH-234	G,N	2.53E+00	2.89E+00	1.10E+00	2.24E+00	pCi/g	JRO	09/06/14 1848	7630
TH-228	G,N	2.06E+00	1.33E+00	1.16E+00	2.37E+00	pCi/g	JRO	09/06/14 1848	7630
PB-212	G,N	1.17E+00	5.86E-02	3.15E-02	6.44E-02	pCi/g	JRO	09/06/14 1848	7630
BI-212	G,N	1.21E+00	3.33E-01	2.28E-01	4.76E-01	pCi/g	JRO	09/06/14 1848	7630
AC-228	G,N	1.01E+00	7.83E-02	8.26E-02	1.72E-01	pCi/g	JRO	09/06/14 1848	7630
RA-224		1.06E+00	2.99E-01	3.01E-01	6.17E-01	pCi/g	JRO	09/06/14 1848	7630
Ra_D_214		8.24E-01	4.12E-02	4.12E-02	8.49E-02	pCi/g	JRO	09/06/14 1848	7630
K-40		1.69E+01	1.05E+00	2.14E-01	4.50E-01	pCi/g	JRO	09/06/14 1848	7630

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 8 of 17 Pages



Certificate of Analysis

ompany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140007

 Lab Sample ID:
 S1408005-07A

Matrix: Soil Client ID: 02060 C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Lc	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec F	'ull								
TL-208	G,N	1.79E-01	2.60E-02	1.93E-02	3.97E-02	pCi/g	JRO	09/07/14 049	7630
TH-232	G,N	1.96E+01	1.11E+01	8.86E+00	1.80E+01	pCi/g	JRO	09/07/14 049	7630
TH-228	G,N	1.07E+00	6.86E-01	9.63E-01	1.96E+00	pCi/g	JRO	09/07/14 049	7630
PB-212	G,N	7.70E-01	4.41E-02	2.32E-02	4.77E-02	pCi/g	JRO	09/07/14 049	7630
BI-212	G,N	8.38E-01	2.68E-01	2.03E-01	4.24E-01	pCi/g	JRO	09/07/14 049	7630
AC-228	G,N	7.05E-01	7.12E-02	6.72E-02	1.40E-01	pCi/g	JRO	09/07/14 049	7630
RA-224		6.25E-01	2.50E-01	2.39E-01	4.90E-01	pCi/g	JRO	09/07/14 049	7630
Ra_D_214		6.76E-01	3.68E-02	3.22E-02	6.65E-02	pCi/g	JRO	09/07/14 049	7630
K-40		1.49E+01	9.20E-01	2.02E-01	4.23E-01	pCi/g	JRO	09/07/14 049	7630

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 9 of 17 Pages



Certificate of Analysis

mpany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140008

 Lab Sample ID:
 S1408005-08A

Matrix: Soil Client ID: 02060 C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Lc	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec F	'ull								
J-235	G,N	1.09E-01	3.85E-02	3.08E-02	6.25E-02	pCi/g	TDR	09/05/14 1453	7631
rL-208	G,N	2.63E-01	3.18E-02	2.00E-02	4.13E-02	pCi/g	TDR	09/05/14 1453	7631
гн-234	G,N	2.89E+00	1.98E+00	1.05E+00	2.13E+00	pCi/g	TDR	09/05/14 1453	7631
ГН-228	G,N	9.80E-01	9.82E-01	1.25E+00	2.55E+00	pCi/g	TDR	09/05/14 1453	7631
PB-212	G,N	1.03E+00	5.83E-02	3.25E-02	6.64E-02	pCi/g	TDR	09/05/14 1453	7631
PA-234M	G,N	5.77E+00	6.11E+00	4.77E+00	1.01E+01	pCi/g	TDR	09/05/14 1453	7631
31-212	G,N	9.69E-01	3.45E-01	2.61E-01	5.42E-01	pCi/g	TDR	09/05/14 1453	7631
AC-228	G,N	9.59E-01	9.62E-02	7.40E-02	1.54E-01	pCi/g	TDR	09/05/14 1453	7631
RA-224		1.03E+00	3.01E-01	3.12E-01	6.39E-01	pCi/g	TDR	09/05/14 1453	7631
Ra_D_214		7.12E-01	4.40E-02	3.13E-02	6.52E-02	pCi/g	TDR	09/05/14 1453	7631
K-40		1.41E+01	9.48E-01	2.46E-01	5.15E-01	pCi/q	TDR	09/05/14 1453	7631

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory

AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 10 of 17 Pages



Certificate of Analysis

ompany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140009

 Lab Sample ID:
 S1408005-09A

Matrix: Soil Client ID: 02060 C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Lc	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec I	Full								
U-235	G,N	7.29E-02	3.05E-02	2.47E-02	5.01E-02	pCi/g	TDR	09/05/14 2055	7631
TL-208	G,N	2.44E-01	2.65E-02	1.64E-02	3.39E-02	pCi/g	TDR	09/05/14 2055	7631
TH-234	G,N	9.71E-01	7.32E-01	9.27E-01	1.88E+00	pCi/g	TDR	09/05/14 2055	7631
TH-228	G,N	8.23E-01	6.86E-01	1.03E+00	2.10E+00	pCi/g	TDR	09/05/14 2055	7631
PB-212	G,N	8.76E-01	4.84E-02	2.83E-02	5.77E-02	pCi/g	TDR	09/05/14 2055	7631
BI-212	G,N	9.65E-01	2.40E-01	1.67E-01	3.50E-01	pCi/g	TDR	09/05/14 2055	7631
AC-228	G,N	7.64E-01	7.29E-02	6.59E-02	1.37E-01	pCi/g	TDR	09/05/14 2055	7631
RA-224		7.52E-01	2.54E-01	2.58E-01	5.28E-01	pCi/g	TDR	09/05/14 2055	7631
Ra_D_214		7.18E-01	3.50E-02	3.43E-02	7.06E-02	pCi/g	TDR	09/05/14 2055	7631
K-40		1.43E+01	8.81E-01	1.72E-01	3.63E-01	pCi/g	TDR	09/05/14 2055	7631

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 11 of 17 Pages



Certificate of Analysis

ompany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140010

 Lab Sample ID:
 S1408005-10A

Matrix: Soil Client ID: 02060C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Lc	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec 1	Full								
TL-208	G,N	2.34E-01	2.81E-02	1.84E-02	3.79E-02	pCi/g	TDR	09/06/14 256	7631
TH-234	G,N	1.69E+00	7.12E-01	8.87E-01	1.80E+00	pCi/g	TDR	09/06/14 256	7631
TH-228	G,N	9.46E-01	6.89E-01	1.04E+00	2.11E+00	pCi/g	TDR	09/06/14 256	7631
PB-212	G,N	8.22E-01	4.73E-02	2.88E-02	5.87E-02	pCi/g	TDR	09/06/14 256	7631
PA-234M	G,N	4.37E+00	4.83E+00	3.77E+00	8.05E+00	pCi/g	TDR	09/06/14 256	7631
BI-212	G,N	9.22E-01	2.89E-01	2.16E-01	4.48E-01	pCi/g	TDR	09/06/14 256	7631
AC-228	G,N	8.28E-01	6.85E-02	6.15E-02	1.28E-01	pCi/g	TDR	09/06/14 256	7631
CD-109	F,N	5.21E-01	3.22E-01	3.64E-01	7.40E-01	pCi/g	TDR	09/06/14 256	7631
RA-224		5.60E-01	2.42E-01	2.61E-01	5.35E-01	pCi/g	TDR	09/06/14 256	7631
Ra_D_214		5.78E-01	3.60E-02	4.26E-02	8.70E-02	pCi/g	TDR	09/06/14 256	7631
K-40		1.69E+01	1.00E+00	2.00E-01	4.18E-01	pCi/g	TDR	09/06/14 256	7631

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 12 of 17 Pages



Certificate of Analysis

mpany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140011

 Lab Sample ID:
 S1408005-11A

Matrix: Soil Client ID: 02060 C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Le	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec I	Full								
U-235	G,N	1.59E-01	7.71E-02	6.58E-02	1.32E-01	pCi/g	TDR	09/06/14 858	7631
TL-208	G,N	3.51E-01	3.37E-02	1.93E-02	3.99E-02	pCi/g	TDR	09/06/14 858	7631
TH-234	G,N	2.01E+00	1.13E+00	1.09E+00	2.22E+00	pCi/g	TDR	09/06/14 858	7631
TH-228	G,N	3.78E+00	1.01E+00	1.30E+00	2.64E+00	pCi/g	TDR	09/06/14 858	7631
PB-212	G,N	1.19E+00	6.23E-02	3.45E-02	7.02E-02	pCi/g	TDR	09/06/14 858	7631
PB-210	G,N	4.54E+00	2.51E+00	2.01E+00	4.10E+00	pCi/g	TDR	09/06/14 858	7631
BI-212	G,N	1.14E+00	3.60E-01	2.70E-01	5.60E-01	pCi/g	TDR	09/06/14 858	7631
AC-228	G,N	1.11E+00	8.97E-02	7.51E-02	1.56E-01	pCi/g	TDR	09/06/14 858	7631
CO-60	E,N	1.27E-02	1.44E-02	1.09E-02	2.40E-02	pCi/g	TDR	09/06/14 858	7631
BA-133	E,N	7.03E-02	2.81E-02	4.68E-02	9.51E-02	pCi/g	TDR	09/06/14 858	7631
RA-224		8.41E-01	3.25E-01	3.23E-01	6.61E-01	pCi/g	TDR	09/06/14 858	7631
Ra_D_214		1.07E+00	5.17E-02	3.76E-02	7.77E-02	pCi/g	TDR	09/06/14 858	7631
K-40		1.81E+01	1.10E+00	2.24E-01	4.70E-01	pCi/g	TDR	09/06/14 858	7631

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 13 of 17 Pages



Certificate of Analysis

ompany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140012

 Lab Sample ID:
 S1408005-12A

Matrix: Soil Client ID: 02060C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Lc	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec H	'ull								
rL-208	G,N	3.06E-01	3.17E-02	1.86E-02	3.84E-02	pCi/g	TDR	09/06/14 1459	7631
гн-234	G,N	2.56E+00	8.59E-01	1.07E+00	2.18E+00	pCi/g	TDR	09/06/14 1459	7631
гн-228	G,N	2.30E+00	8.43E-01	1.08E+00	2.20E+00	pCi/g	TDR	09/06/14 1459	7631
PB-212	G,N	1.20E+00	8.51E-02	3.21E-02	6.55E-02	pCi/g	TDR	09/06/14 1459	7631
31-212	G,N	8.84E-01	2.62E-01	2.00E-01	4.20E-01	pCi/g	TDR	09/06/14 1459	7631
AC-228	G,N	1.09E+00	8.22E-02	6.10E-02	1.28E-01	pCi/g	TDR	09/06/14 1459	7631
r-131	E,N	2.71E-01	3.12E-01	2.49E-01	5.23E-01	pCi/g	TDR	09/06/14 1459	7631
CD-109	E,N	4.62E-01	3.52E-01	4.37E-01	8.87E-01	pCi/g	TDR	09/06/14 1459	7631
RA-224		7.30E-01	3.55E-01	3.03E-01	6.20E-01	pCi/g	TDR	09/06/14 1459	7631
a_D_214		7.64E-01	4.30E-02	3.48E-02	7.20E-02	pCi/g	TDR	09/06/14 1459	7631
<−40		1.63E+01	1.02E+00	2.43E-01	5.08E-01	pCi/g	TDR	09/06/14 1459	7631

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory

AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 14 of 17 Pages



Certificate of Analysis

ompany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140013

 Lab Sample ID:
 S1408005-13A

Matrix: Soil Client ID: 02060C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Le	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec I	Full								
U-235	G,N	7.21E-02	2.86E-02	2.31E-02	4.70E-02	pCi/g	TDR	09/06/14 2100	7631
TL-208	G,N	2.69E-01	2.77E-02	1.73E-02	3.57E-02	pCi/g	TDR	09/06/14 2100	7631
TH-228	G,N	1.48E+00	7.01E-01	1.17E+00	2.37E+00	pCi/g	TDR	09/06/14 2100	7631
PB-212	G,N	9.75E-01	1.01E-01	2.92E-02	5.95E-02	pCi/g	TDR	09/06/14 2100	7631
BI-212	G,N	1.19E+00	2.99E-01	2.05E-01	4.27E-01	pCi/g	TDR	09/06/14 2100	7631
AC-228	G,N	8.15E-01	7.26E-02	5.62E-02	1.17E-01	pCi/g	TDR	09/06/14 2100	7631
RA-224		8.24E-01	3.23E-01	2.59E-01	5.30E-01	pCi/g	TDR	09/06/14 2100	7631
Ra_D_214		7.49E-01	3.53E-02	3.02E-02	6.23E-02	pCi/g	TDR	09/06/14 2100	7631
K-40		1.79E+01	1.03E+00	1.93E-01	4.05E-01	pCi/g	TDR	09/06/14 2100	7631

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 15 of 17 Pages



Certificate of Analysis

ompany: USAFSAM/OEC

Address: 2510 Fifth St. Bldg. 840 Report Date: 9/10/2014 10:30:43 AM

WRIGHT-PATTERSON AFB OH, 45433

Contact: SSgt Michael Ames

 Client Sample ID:
 GS140014

 Lab Sample ID:
 S1408005-14A

Matrix: Soil Client ID: 02060C

 Collection Date:
 7/30/2014

 Receive Date:
 8/5/2014

 Collector:
 Client

Parameter	Qualifier	Activity	Uncertainty	Lc	MDA	Units	Analyst	Date/Time	Batch
Gamma Spec F	'ull								
U-235	G,N	6.63E-02	3.10E-02	2.50E-02	5.07E-02	pCi/g	TDR	09/07/14 302	7631
TL-208	G,N	2.58E-01	3.01E-02	1.87E-02	3.87E-02	pCi/g	TDR	09/07/14 302	7631
TH-234	G,N	9.72E-01	8.00E-01	1.02E+00	2.08E+00	pCi/g	TDR	09/07/14 302	7631
PB-212	G,N	9.66E-01	8.97E-02	3.10E-02	6.33E-02	pCi/g	TDR	09/07/14 302	7631
BI-212	G,N	8.84E-01	2.40E-01	1.74E-01	3.66E-01	pCi/g	TDR	09/07/14 302	7631
AC-228	G,N	8.28E-01	7.97E-02	6.93E-02	1.44E-01	pCi/g	TDR	09/07/14 302	7631
CS-137	D, F, N	3.69E-02	2.00E-02	1.53E-02	3.19E-02	pCi/g	TDR	09/07/14 302	7631
RA-224		6.22E-01	3.48E-01	2.82E-01	5.78E-01	pCi/g	TDR	09/07/14 302	7631
Ra_D_214		7.59E-01	4.34E-02	3.59E-02	7.40E-02	pCi/g	TDR	09/07/14 302	7631
K-40		1.46E+01	9.24E-01	2.07E-01	4.36E-01	pCi/q	TDR	09/07/14 302	7631

BRIAN J STROH, Capt, USAF Chief, Radioanalytical Laboratory

AURELIE M SOREFAN, DR-II, PhD, DAF Technical Director, Radioanalytical Laboratory

Page 16 of 17 Pages




Radiation Qualifier List

Qualifier	Qualifier Description
A	Identification Rejected
В	(ELAP) Blank Containination
C	Inconclusive
D	Misidentification
E	False Positve
F	Consistent with False Positive
G	Qualitatively consistent with detector background
Н	Quantitatively consistent with detector background
I	Inconsistent energy shift
J	(ELAP) The reported results is an esimated value
N	(ELAP) Non-target analyte
Q	(ELAP) One or more quality control criteria failed
U	(ELAP) Analyte was not detected and is reported as less than the LOD
K	Inconsistent Full Width Half Max (FWHM)
L	Hold time exceeded
M	Results invalid

Page 17 of 17 Pages

Attachment 2
Map of Gamma Walkover with Soil Sample Locations



Attachment 3 Instrument Calibration Sheets



DEPARTMENT OF THE AIR FORCE USAF SCHOOL OF AEROSPACE MEDICINE (AFMC)

OCCUPATIONAL ENVIRONMENTAL HEALTH WRIGHT-PATTERSON AFB OHIO CERTIFICATE OF CALIBRATION

NIST Traceable Check Sources Reference Instruments	NIST Traceable Check Sources Serial # Cert. Date DPM Mtg. Model Serial # Cal. Dute Cs. 137 MU851 01-Nov-04 1974084 Ludium 500-1 102952 4 DEC 201		TIO FOURTH	AND DIAGNOS	unesses.	FEET ME		
Sorial # Cert. Date DPM Mfg. Model Serial # Cal. Due Date	Serial # Cert. Date DPM Mitg. Model Serial # Cal. Due Cs 137 MU851 01-Nov-04 1974084 Ludium 500-1 102952 4 DEC 20	trumante		AND DIAGNOS				MIS
Cx 137 MU851 01-Nov-04 1974084 Ludium 500-1 102952 4 DEC 2014	Cs 137 MU851 01-Nov-04 1974084 Ludium 500-1 102952 4 DEC 20			DRM M				05775757
Measurement Siandards and lest equipment used are traceable to the National Institute of Standards and Technology, to the extent allowed by the last calibration facilities. ✓ Battery Ck ✓ Mechanical Ck ✓ Meter Zeroed ✓ Reset Ck. NA Alarm Ck. ✓ Audio Ck. ✓ Geotropism Ck. ✓ F/S Resp. Ck. ✓ Window Op. As Found HV 494 VDC Temperature 70.1 °F Relative Humidity 62.4 °F Final Volt, Set 700 VDC Threshold (LLD) 9 mV Window (ULD) 20 mV Window width 11 m HV Readout (2 points) Reference: 500 V Reference: 1000 V Inst. Readout: 497 V ± 2% Inst. Readout: 1.012 V ± 2% RANGE MULTIPLIER CAL POINT READING READING x 1000 400 CPM 380,000 CPM 400,000 CPM x 1000 400 CPM 40,000 CPM 40,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 1 400 CPM 400 CPM 400 CPM	Measurement Standards and test equipment used are traceable to the National Institute of Standards and Technology, to the extent allowed exhibitation facilities. Battery Ck Mechanical Ck Meter Zeroed Reset Ck. NA Alarm Ck. Audio Ck. Geotropiam Ck. F/S Resp. Ck. Window Op.	THE RESERVE OF THE PERSON NAMED IN COLUMN 1						_
Saltery Ck	Battery Ck Mechanical Ck Meter Zeroed Reset Ck NA Alarm Ck	102932 4 DEC 2014	aium 500-1	ration Luc	01-04		MUSSI	CN 137
As Found HV 494 VDC Temperature 70.1 °F Relative Humidity 62.4 °9 Final Volt. Set 700 VDC Threshold (LLD) 9 mV Window (ULD) 20 mV Window width 11 m HV Readout (2 points) Reference: 500 V Reference: 1000 V Inst. Readout: 497 V ± 2% Inst. Readout: 1.012 V ± 2% RANGE REFERENCE "AS FOUND" CORRECTED READING x 1000 400 CPM 380,000 CPM 400,000 CPM x 1000 100 CPM 110,000 CPM 110,000 CPM x 1000 400 CPM 40,000 CPM 40,000 CPM x 1000 400 CPM 40,000 CPM 40,000 CPM x 100 400 CPM 11,000 CPM 11,000 CPM x 100 400 CPM 11,000 CPM 11,000 CPM x 100 400 CPM 11,000 CPM 11,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 100 400 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 1,000 CPM 4,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,000 CPM 1,000 CPM x 11 400 CPM 4,000 CPM 4,000 CPM x 1 400 CPM 1,000 CPM 1,000 CPM x 1 400 CPM 1,000 CPM 1,000 CPM x 1 400 CPM 1,000 CPM 1,000 CPM x 1 100 CPM 1,000 CPM x 1 100 CPM 1,000 CPM x 1 100 CPM 1,000 CPM x 1	As Found HV	Technology, to the extent allowed by the Institut	stitute of Standards at	cubic to the National Inv	nt used are tra	and test equips		
As Found HV	As Found HV	NA Alarm Ck.	d Reset C	Ck. / Meter Zeroed	Mechanic	Battery Ck.	1	
Final Volt. Set	Final Volt. Set 700 VDC Threshold (LLD) 9 mV Window (ULD) 20 mV Window width HV Readout (2 points) Reference: 500 V Reference: 1000 V Inst. Readout: 497 V ± 2% Inst. Readout: 1.012 V ± RANGE MULTIPLIER REFERENCE CAL. POINT "AS FOUND" READING CORRECTED READING X 1000 400 CPM 380,000 CPM 400,000 CPM X 1000 100 CPM 110,000 CPM 110,000 CPM X 100 400 CPM 40,000 CPM 40,000 CPM X 100 100 CPM 11,000 CPM 11,000 CPM X 10 400 CPM 4,000 CPM 4,000 CPM X 10 100 CPM 1,100 CPM 1,100 CPM X 1 400 CPM 400 CPM 400 CPM X 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	. Ck. Window Op.	Ck. F/S Red	✓ Geatropism (Audio Ck.			
Final Volt. Set	Final Volt. Set 700 VDC Threshold (LLD) 9 mV Window (ULD) 20 mV Window width HV Readout (2 points) Reference: 500 V Reference: 1000 V Inst. Readout: 497 V ± 2% Inst. Readout: 1.012 V ± RANGE MULTIPLIER REFERENCE CAL. POINT "AS FOUND" READING CORRECTED READING X 1000 400 CPM 380,000 CPM 400,000 CPM X 1000 100 CPM 110,000 CPM 110,000 CPM X 100 400 CPM 40,000 CPM 40,000 CPM X 100 100 CPM 11,000 CPM 11,000 CPM X 10 400 CPM 4,000 CPM 4,000 CPM X 10 100 CPM 1,100 CPM 1,100 CPM X 1 400 CPM 400 CPM 400 CPM X 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	Salativa Musaidhi an 4 N	70.4 °E			404 V/F	numd LIV	An Fo
HV Readout (2 points) Reference: 500 V Reference: 1000 V Inst. Readout: 497 V ± 2% Inst. Readout: 1.012 V ± 2%	HV Readout (2 points) Reference: 500 V Reference: 1000 V	Relative Humidity 62.4 %	70.1	erature	rem	494 VL	ound HV _	AS FO
RANGE REFERENCE "AS FOUND" CORRECTED READING	RANGE REFERENCE CAL. POINT READING CORRECTED READING	20 mV Window width 11 mV	Window (ULD)	(LLD) 9 mV	Threshol	700 VC	Volt. Set	Final \
RANGE REFERENCE "AS FOUND" CORRECTED READING	RANGE REFERENCE CAL. POINT READING CORRECTED READING		De	500 V	eference:	points)	Deadout (2	HV
RANGE MULTIPLIER REFERENCE CAL. POINT "AS FOUND" READING CORRECTED READING x 1000 400 CPM 380,000 CPM 400,000 CPM x 1000 100 CPM 110,000 CPM 110,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	RANGE MULTIPLIER REFERENCE CAL. POINT "AS FOUND" READING CORRECTED READING x 1000 400 CPM 380,000 CPM 400,000 CPM x 1000 100 CPM 110,000 CPM 110,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM					A CONTRACTOR OF THE PARTY OF TH	readout (2	
MULTIPLIER CAL. POINT READING READING x 1000 400 CPM 380,000 CPM 400,000 CPM x 1000 100 CPM 110,000 CPM 110,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	MULTIPLIER CAL. POINT READING READING x 1000 400 CPM 380,000 CPM 400,000 CPM x 1000 100 CPM 110,000 CPM 110,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	1012 V ± 276	270 Hist. 19		residuat.	,,,,		
MULTIPLIER CAL. POINT READING READING x 1000 400 CPM 380,000 CPM 400,000 CPM x 1000 100 CPM 110,000 CPM 110,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	MULTIPLIER CAL. POINT READING READING x 1000 400 CPM 380,000 CPM 400,000 CPM x 1000 100 CPM 110,000 CPM 110,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	BRECTED	IND" CO	"AS FOU	ERENCE	P	RANGE	
x 1000 100 CPM 110,000 CPM 110,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	x 1000 100 CPM 110,000 CPM 110,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM							40000
x 1000 100 CPM 110,000 CPM 110,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	x 1000 100 CPM 110,000 CPM 110,000 CPM x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	400.000 CPM	.000 СРМ	380.	400 CPM		1000	X
x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	x 100 400 CPM 40,000 CPM 40,000 CPM x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM		The same of the sa					
x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	x 100 100 CPM 11,000 CPM 11,000 CPM x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM		And the same of th	-	400 CPM		100	×
x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM DIGITAL SCALER READOUT	x 10 400 CPM 4,000 CPM 4,000 CPM x 10 100 CPM 1,100 CPM 1,100 CPM x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM		,000 CPM	11,	100 CPM		100	×
x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM DIGITAL SCALER READOUT	x 1 400 CPM 400 CPM 400 CPM x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM		,000 CPM	4,	400 CPM		10	×
x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM DIGITAL SCALER READOUT	x 1 100 CPM 110 CPM 110 CPM Log Scale 200 CPM 200 CPM 200 CPM	1,100 CPM	,100 CPM	1,	100 CPM		10	×
Log Scale 200 CPM 200 CPM 200 CPM DIGITAL SCALER READOUT	Log Scale 200 CPM 200 CPM 200 CPM	400 CPM	400 CPM		400 CPM		(1	×
DIGITAL SCALER READOUT		110 CPM	110 CPM		100 CPM		(1	×
	DIGITAL SCALER READOUT	200 CPM	200 CPM		200 CPM		Scale	Log !
CAL REE POINT AS EQUIND READING CORRECTED READING	DIGITAL SCALER READOUT		EADOUT	ITAL SCALER R	DI			
CAL REF. FOINT AS FOUND READING CORRECTED READING	CAL. REF. POINT AS FOUND READING CORRECTED READING	CORRECTED READING	DING	AS FOUND REAL	7000	DINT	AL. REF. PO	CA
40,000 CPM 40,000 CPM 40,000 CPM	40,000 CPM 40,000 CPM 40,000 CPM	40,000 CPM	CPM	40,000		M	40,000 CP	
*UNCERTAINTY WITHIN + 10% CORRECTION FACTOR WITHIN + 20%	*UNCERTAINTY WITHIN ± 10% CORRECTION FACTOR WITHIN ± 20%	20%	CTOR WITHIN	ORRECTION FAC	N ± 10% C	INTY WITH	UNCERTA	*1
() 유리 () 구리 () 이 () () () () 이 () () () ()	OMMENTS: Calibration Interval = 1 year Use "Window OUT		W OUT	Use "Windo	val = 1 ves	bration Inte	ITS: Call	OMMEN
	, and the state of		001					
	Detector Parameters: Page2- 3 Procedural Authority - ICP#222100	1			ge2- 3	ters: I	ector Parame	Dete
OMMENTS: Calibration Interval = 1 year Use "Window OUT	Trocedural Additionary - 1CF #222100	uthority - ICP#22210000	Procedural /					

Jun 16 2014 03 13 PM Kurt Shorts HotSpot FIDLER Calibration Information

```
: Jun 16 2014 03:13 PM
Report Date
                                        16 June 2014
Calibration Date
                                     : Other Nuclide Check Source
Target Mix
                                        Cs-137
Radionuclide
Detector Barcode Number
                                     : N/A
                                    : 102209
: Scionix
: 76BRS76/3M-E1-X
: SAG420
: Ludlum
Meter Barcode Number
Detector Manufacturer
Detector Model Number
Detector Serial Number
Meter Manufacturer
Meter Model Number
                                     : 2221
: 287511
                                          2221
Meter Serial Number
                                     : RP 3067
: 16 June 2014
Check Source I.D.
Calibration Date
                                      : Kurt Shorts
Calibrated by
Check Source Activity (uCi): 8.890E-01
Sample Counting Time (minutes) : 1 Detector Height (cm) : 3.000E+01
                                            : 1.000E+00
Cs-137 Window Information:
Background (cpm) : 9,928
Areal Limit of Sensitivity (uCi/m2) : 6.9E-02
Point Limit of Sensitivity (uCi) : 6.4E-02
K-factor (m2) : 0.92
K-factor (m2)
Counting Data (counts):
0-cm: 16455
   20-cm: 14967
40-cm: 12820
   60-cm: 11625
 80-cm: 11051
100-cm: 10321
Instrument Type
                           :Other
Window Option:Only 60 keV
Units:Classic
This is an actual 3x3 calibration, and the values are typical of most 3x3
configurations.
Detector Calibration Results
Cs-137 Window Information:
Cs-137 Detector Efficiency (cpm/(uci/m2)): 6.8E+03
Cs-137 Detector Areal LOS (uCi/m2) : 6.9E-02
Cs-137 Detector Point LOS (uCi) : 6.4E-02
Cs-137 Detector Background Rate (cpm) : 9,928
Cs-137 Detector Check Source Rate (cpm) : 6,527
Cs-137 Detector K-Factor (m2) : 0.92
Cs-137 Detector K-Factor sdev (%) : 5.3
```



DEPARTMENT OF THE AIR FORCE USAF SCHOOL OF AEROSPACE MEDICINE (AFMC) OCCUPATIONAL ENVIRONMENTAL HEALTH

WRIGHT-PATTERSON AFB OHIO

Mfg. Sc		768R\$76/2ME1-X	Serial # SAC		index#	40300	Cal. Due Date: 24
		MEASUREME	NT AND DIA	GNOSTIC E	QUIPMEN	IT	
NIST	T Traceable Che	ck Sources		Ref	erence In	struments	
Isotope	Serial #	Cert. Date	DPM	Mfg.	Model	Serial #	Cal. Due Date
Cs 137	RP3067	01-Nov-04	1972997	Ludlum	500-1	102951	06 MAR 2015
Messurements of the contract o	nt Standards and test e facilities.	quipment used are to	rscesble to the No	ional Institute of	Standards ut	d Technology, to	the extent allowed by the i
	√ Battery	Ck. Mechanic			Reset C	ik. NA Alai ip. Ck. Win	
As Fou	and HV 647	VDC Ten	nperature _	0.88		Relative Hun	
Final V	olt. Set 675	VDC Thresho	old (LLD) 10	mV Wind	ow (ULD)	19 mV V	/indow width 9
HVR	eadout (2 points)			v	1000		
114.14	eadout (2 points)	Inst. Readout		v v ± 2%		_	000 V ,006 V ± 2%
		met, redución.		v ± 270	miet. N	eedout	,000 V ± 276
R	ANGE	REFERENCE	E "AS	S FOUND*	C	DRRECTED	
	TIPLIER	CAL. POINT		EADING		READING	
X 1	1000	400 CPN	Л	400,000	СРМ	400.0	00 CPM
x 1	1000	100 CPM	A	100,000			00 CPM
x 1	100	400 CPN	A	40,000	CPM	40,0	00 CPM
x 1	100	100 CPN	Λ	10,000	CPM	10,0	00 CPM
x 1	10	400 CPN	И	4,000	CPM		00 CPM
x 1	10	100 CPN	Λ	1,000			00 срм
x 1	1	400 CPN	Λ	400			00 CPM
x 1	1	100 CPN	A		CPM		00 CPM
Log S		200 CPN		200			00 CPM
		D	IGITAL SCAL				
CAL	REF. POINT		AS FOUND	READING		CORRECT	ED READING
- 4	10,000 CPM		40	.067 CPM		40	0,067 CPM
*U	NCERTAINTY W	/ITHIN ± 10%			WITHIN		
DMMENT		Interval = 1 ye	25 75 - 107 77 77	Window Ot	A CONTRACTOR		
2111112141	C. Cambration	minor van - 1 ye	a: 000	Million Of	01		
Detect	tor Parameters:	Page2- 3		Dece	downol d	orthodes 1	CD#22210000
	1	1	1	Proc	edural A	uthority -	CP#22210000
	10/11	1 7/1	//_				
2012	(VO STV)	1 41	11.				
rated By:	11 hou	////	X-			Date:_2	24-Jun-2014
wed By:	1/2	5/2					25- Jun-201



DEPARTMENT OF THE AIR FORCE USAF SCHOOL OF AEROSPACE MEDICINE (AFMC)

OCCUPATIONAL ENVIRONMENTAL HEALTH
WRIGHT-PATTERSON AFB OHIO

WRIGHT-PATTERSON AFB OHIO CERTIFICATE OF CALIBRATION

	Mod	iel 7688876/20061-)	Serial # SA	G 423	index#	08335	Cal. Due Date: 24-Jun-
	TEST	, MEASUREM	ENT AND DIA	AGNOSTIC E	QUIPMEN	т	
NIST TO		heck Sources				struments	
Isotope	Serial #	Cert. Date	DPM	Mfg.	Model	Serial #	Cal. Due Date
Cs 137 R	23067	01-Nov-04	1972997	Ludium	500-1	102951	06 MAR 2015
Probe #1 Mfg. Scionis Model 76BRS Serial # SAI Index # 08: Isotope: Cs High Voltage	NaI x 976/2ME1-X G 423 335 -137 @ 6*	DETECTOR P M M S			Pro Mfg Mo See Ind Iso	ATION bbe #3	o the extent allowed by the Institu
500	25427		1				
550	27564	_ L	1			1	
600	27676						
650	28510						
*675	28290					1	
700	29123		V				
750	28905						
800	29710					1	
850	33455						
	-						
*****	******					- 100	

Bkgd@ 675v	9,258	7 6		7			7
Final Volt.	Set (75 VDC	Final Volt. Set	VD	C F	inal Volt. Se	t VDC
		-	Efficiency	CPM/µci/m	² @12*	Efficiency	CPM/µci/m² @

31

Page 30f3

Jun 24 2014 01 40 PM Willis Hosley HotSpot FIDLER Calibration Information

```
Report Date
Calibration Date
                                              : Jun 24 2014 01:40 PM
                                                   June 24 2014
                                              : Other Nuclide Check Source
 Target Mix
 Radionuclide
                                                   Cs-137
 Detector Barcode Number
                                              : 08335
Meter Barcode Number : 102210
Detector Manufacturer : Scionix
Detector Model Number : 76BRS76/3M-E1-X
Detector Serial Number : SAG423
Meter Barcode Number
Detector Manufacturer
 Meter Manufacturer
                                             : Ludlum
Meter Model Number
                                           : 2221
: 287537
Meter Serial Number
Check Source I.D.
                                             : RP 3067
                                                  June 24 2014
Willis Hosley
Calibration Date
 Calibrated by
Check Source Activity (uCi): 8.890E-01
Sample Counting Time (minutes) : 1
Detector Height (cm) : 3.000E+01
                                                            : 1.000E+00
Cs-137 Window Information:
Background (cpm)
Areal Limit of Sensitivity (uCi/m2)
Point Limit of Sensitivity (uCi)
                                                              : 9,258
                                                              : 1.1E-02
: 5.5E-02
K-factor (m2)
                                                               : 4.83
Counting Data (counts):
0-cm: 16549
    20-cm:
                 14818
    40-cm: 12314
    60-cm:
                 11101
    80-cm: 10529
100-cm: 10378
Instrument Type :Other
Window Option:Only 60 keV
Units:Classic
This is an actual 3x3 calibration, and the values are typical of most 3x3
configurations.
Detector Calibration Results
Cs-137 Window Information:
Cs-137 Detector Efficiency (cpm/(uci/m2)): 4.0E+04
Cs-137 Detector Areal LOS (uci/m2) : 1.1E-02
Cs-137 Detector Point LOS (uci) : 5.5E-02
Cs-137 Detector Background Rate (cpm) : 9,258
CS-137 Detector Areal LOS (UCi/m2): 4.0E+04
CS-137 Detector Areal LOS (UCi/m2): 1.1E-02
CS-137 Detector Point LOS (UCi): 5.5E-02
CS-137 Detector Background Rate (cpm): 9,258
CS-137 Detector Check Source Rate (cpm): 7,291
CS-137 Detector K-Factor (m2): 4.83
CS-137 Detector K-Factor sdev (%): 5.2
```

Page 1

Page 3053



DEPARTMENT OF THE AIR FORCE USAF SCHOOL OF AEROSPACE MEDICINE (AFMC) OCCUPATIONAL ENVIRONMENTAL HEALTH

WRIGHT-PATTERSON AFB OHIO CERTIFICATE OF CALIBRATION

TEST	MEASUREMEN	NT AND DIAG	NOSTIC E	OLUPMEN	т	
NIST Traceable Ch		TI AND DIAG			struments	
Isotope Serial #	Cert. Date	DPM	Mfg.	Model	Serial #	Cal. Due Date
Cs 137 MU851		975850	Ludlum	500-1	102952	4 DEC 2014
culibration facilities.	ry Ck. Mechanic	cal Ck. Meter	Zeroed	√ Roset C		
7511277109100	-			_	_	
As Found HV 69	94 VDC Ten	perature _	64.8	F	Relative Hur	midity 69.7 %
Final Volt. Set 67	75 VDC Thresho	id (LLD) 10	mV Winds	ow (ULD)	20 mV V	Window width 10 mV
HV Readout (2 point		r wasti e	,		ference:	
HA Meadout (5 boils	Inst. Readout		/ ± 2%		_	1,010 V ± 2%
	mac Readout		_ ± 276	Iller i	eauoui	1,010 V ± 276
RANGE	REFERENCI	*AS	FOUND*	CC	PRECTED	
MULTIPLIER	CAL POINT		EADING		READING	
x 1000	400 CPI	1	420,000	CPM	400.0	000 CPM
x 1000	100 CPI	4	100,000	СРМ	- Company of the Comp	00 CPM
x 100	400 CP1	4	40,000	CPM	40,0	100 CPM
x 100	100 CPM	4	10,000	CPM	10,0	00 CPM
x 10	400 CPN	1	4,000	CPM	4,0	00 CPM
x 10	100 CPI	Λ	1,000	CPM	1,0	00 CPM
x 1	400 CPI	A	400	CPM	4	00 CPM
x 1	100 CPM	1	100	CPM	1	00 CPM
Log Scale	200 CPN	A	200	CPM	2	00 CPM
1371423155 CERTAIN		IGITAL SCAL				
CAL. REF. POINT			READING			TED READING
40,000 CPM			,304 СРМ			0,304 CPM
*UNCERTAINTY	WITHIN ± 10%	CORRECTIO	N FACTOR	WITHIN;	± 20%	
OMMENTS: Calibratio	on Interval = 1 ye	ar Use "	Window of	UT.		
	•					
Detector Parameters:	Page2- 3		Proc	edural A	uthority -	ICP#22210000
					-	
rated By: Luct	Shorts				Date:	03-Jun-2014
1	1 1-	- 1-				
rwed By: War	Vintati	113			Date	Stun 14



DEPARTMENT OF THE AIR FORCE USAF SCHOOL OF AEROSPACE MEDICINE (AFMC)

WRIGHT-PATTERSON AFB OHIO

Mfg. Ludi Mfg. Scion		768RS78/3M-E1-3	Serial # 2 Serial # 5		Index #		Date: 03-Ju Cal. Due Date: 03-Ju
	TEST, M	MEASUREM	ENT AND D	IAGNOSTIC E	QUIPMEN	т	
NIST T	raceable Che	ck Sources		Re	ference ins	struments	
Isotope	Serial #	Cert. Date	DPM	Mfg.	Model	Serial #	Cai. Due Date
Cs 137 N	4U851	01-Nov-04	1975850	Ludlum	500-1	102952	4 DEC 2014
		pripment med an	e traceable to the	National Institute	of Standards an	d Technology, to	the extent allowed by the Inst
calibration ficil		FTECTOR	men v	OLTAGE O	DTIME A	TION	
Probe #1	Nai D		robe #2	OLTAGE O		be#3	
Mfg. Scion	ioc		fg. N/A		Mfg		4
	S76/3M-E1-X		odel		Mo		
Serial # S			erial #			ial #	
Index# N	A	In	dex#		Ind	ex#	
Isotope: C	s-137 @ 6°	Is	otope:		Isot	tope:	
High		Г	High			High	
Voltage	CPM		Voltage	CPM	V	oltage	CPM
500	26601				7		
550	27584						
600	27895						
650	27826						
*675	27780						
700	28274						
750	28287						
800	28173						
850	30217						
900	41729						
*****	*****						

Bkgd@ 675v	8,781						
Final Volt.	Set 675	VDC F	Final Volt. Se	etVI	DC Fi	nal Volt, Set	VDC
Efficiency	9900 CPM/µ	ci/m² @12* E	Efficiency	CPM/µci/n	n ² @12*	Efficiency	CPM/µci/m²
OMMENTS	Calibratian	leterual = 1	unar II	ne Millindow ()	U IT		
OMMENTS	. Calibration	Interval = 1	year 0	se "Window ()	NC I		
				Pro	ocedural A	Authority -	- ICP#22210000
	41.4	C 1 1					
brated By:	17	- \ A	er K			200	03-Jun-2014

34

Jun 03 2014 11 07 AM Kurt Shorts HotSpot FIDLER Calibration Information

```
Report Date
                                       : Jun 03 2014 11:07 AM
Calibration Date
                                       : 03 June 2014
Target Mix
                                           Other Nuclide Check Source
Radionuclide
                                       : Cs-137
Detector Barcode Number
                                       : N/A
: 1027
                                           102204
Meter Barcode Number
Detector Manufacturer
                                       : Scionix
Detector Model Number
Detector Serial Number
                                       : 76BRS76/3M-E1-X
: SAG427
Meter Manufacturer
                                       : Ludlum
Meter Model Number
                                            2221
                                      : 2221
: 290803
Meter Serial Number
                                       : RP 3067
: 03 June 2014
Check Source I.D.
Calibration Date
Calibrated by
                                           Kurt Shorts
Check Source Activity (uCi): 8.900E-01
                                              : 1.000E+00
Sample Counting Time (minutes)
                                  : 3.000E+01
Detector Height (cm)
Cs-137 Window Information:
Background (cpm) : 8,781
Areal Limit of Sensitivity (uCi/m2) : 4.4E-02
Point Limit of Sensitivity (uCi) : 5.5E-02
: 1.24
                                                      : 1.24
K-factor (m2)
Counting Data (counts):
0-cm: 15887
    0-cm: 13996
   20-cm:
   40-cm: 12147
   60-cm:
               10389
   80-cm:
              10225
 100-cm: 9606
Instrument Type :Other window Option:Only 60 kev
Units:Classic
This is an actual 3x3 calibration, and the values are typical of most 3x3
configurations.
Detector Calibration Results
Cs-137 Window Information:
Cs-137 Detector Efficiency (cpm/(uCi/m2)): 9.9E+03
Cs-137 Detector Areal LOS (uCi/m2) : 4.4E-02
Cs-137 Detector Point LOS (uCi) : 5.5E-02
Cs-137 Detector Background Rate (cpm) : 8.781
CS-137 Detector Efficiency (cpm/(uc1/m2)) 3.52-03
CS-137 Detector Areal LOS (uci/m2) 4.4E-02
CS-137 Detector Point LOS (uci) 5.5E-02
CS-137 Detector Background Rate (cpm) 8,781
CS-137 Detector Check Source Rate (cpm) 7,106
Cs-137 Detector K-Factor (m2)
Cs-137 Detector K-Factor sdev (%)
                                                                1.24
```

Attachment 4
Radiation Meter Quality Control (QC) Log

Radiation Meter QC Log

Acceptable Range 1109%	19713 - 24095	P660 - 27030	19713-27093	16574- 2025 8	19713-24093	16574- 20258				
Source Check Reading	21.9 63 cpm	21845 Gan	21084cpm	18416 com	20438 Cpm	14532 COM				
HV/cables/Bat check	4.1 V /700V	6.00/6BV	6.00 /699V	6.14/674V	6.1/ 693V	V 669 / 6.8	1			
Date/Time	293417 0846	29.2hry/8146	30 JANY 0730	307414/1318	30JMy/618	5A9427 30JUN/1620				
N/S	28751 SAG420 76/3m-E1-X	287537 SAC423		15754 SAG-427	CHURS WAYSO	287511 SAG427				
Model	2221 768K576/3m-E1-X	Ladlum 2221			1222 mayon	Ludlum 2221				


Attachment 5 Statistical Summary of Instrument Data

Instrument Model Number: Ludlum 2221 / Scionix 76BRS76 3x3 NaI

Instrument Serial Number: 287511 / SAG420

Surface Type: Dirt/Soil

Area: Trench




Instrument Model Number: Ludlum 2221 / Scionix 76BRS76 3x3 NaI

Instrument Serial Number: 290803 / SAG 427

Surface Type: Dirt/Soil

Area: Trench



Instrument Model Number: Ludlum 2221 / Scionix 76BRS76 3x3 NaI

Instrument Serial Number: 284511 / SAG420

Surface Type: Asphalt

Area: Alongside trench to the Northeast

