# NONINVASIVE CARBON DIOXIDE MONITORING IN A PORCINE MODEL OF ACUTE LUNG INJURY DUE TO SMOKE INHALATION AND BURNS

# Slava Belenkiy,\*<sup>†</sup> Katherine M. Ivey,\*<sup>†</sup> Andriy I. Batchinsky,\* Thomas Langer,\*<sup>‡</sup> Corina Necsoiu,\* William Baker,\* José Salinas,\* and Leopoldo C. Cancio\*

\*US Army Institute of Surgical Research and <sup>†</sup>San Antonio Military Medical Center, Fort Sam Houston, Texas; and <sup>‡</sup>Dipartimento di Fisiopatologia medico-chirurgica e dei trapianti Universitá degli Studi di Milano, Milan, Italy

Received 28 Jan 2013; first review completed 13 Feb 2013; accepted in final form 18 Mar 2012

ABSTRACT In critically ill intubated patients, assessment of adequacy of ventilation relies on measuring partial pressure of arterial carbon dioxide (Paco2), which requires invasive arterial blood gas analysis. Alternative noninvasive technologies include transcutaneous CO<sub>2</sub> (tPco<sub>2</sub>) and end-tidal CO<sub>2</sub> (Etco<sub>2</sub>) monitoring. We evaluated accuracy of tPco<sub>2</sub> and Etco<sub>2</sub> monitoring in a porcine model of acute lung injury (ALI) due to smoke inhalation and burns. Eight anesthetized Yorkshire pigs underwent mechanical ventilation, wood-bark smoke inhalation injury, and 40% total body surface area thermal injury. tPco2 was measured with a SenTec system (SenTec AG, Therwil, Switzerland) and Etco2 with a Capnostream-20 (Oridion Medical, Jerusalem, Israel). These values were compared with Paco2 measurements from an arterial blood gas analyzer. Paired measurements of Etco<sub>2</sub>-Paco<sub>2</sub> (n 276) and tPco<sub>2</sub>-Paco<sub>2</sub> (n 250) were recorded in the Paco<sub>2</sub> range of 25 to 85 mmHg. Overlapping data sets were analyzed based on respiratory and hemodynamic status of animals. Acute lung injury was defined as  $Pao_2/Fio_2 \le 300$  mmHg; hemodynamic instability was defined as mean arterial pressure  $\le 60$  mmHg. Before ALI, Etco<sub>2</sub> demonstrated moderate correlation with Paco<sub>2</sub> ( $R^2$  0.45; P < 0.0001), which deteriorated after onset of ALI ( $R^2$  0.12; P < 0.001) 0.0001). Before ALI, tPco<sub>2</sub> demonstrated moderate correlation ( $R^2$  0.51, P < 0.0001), which was sustained after onset of ALI  $(R^2 = 0.78; P < 0.0001)$ . During hemodynamic stability, Etco<sub>2</sub> demonstrated moderate correlation with Paco<sub>2</sub> ( $R^2 = 0.44; P < 0.001$ ). 0.0001). During hemodynamic instability,  $Etco_2$  did not correlate with  $Paco_2$  ( $R^2$  0.03; P 0.29). tPco<sub>2</sub> monitoring demonstrated strong correlation with  $Paco_2$  during hemodynamic stability ( $R^2$  0.80, P < 0.0001), which deteriorated under hemodynamically unstable conditions ( $R^2 = 0.39$ ; P < 0.0001). Noninvasive carbon dioxide monitors are acceptable for monitoring trends in Paco<sub>2</sub> under conditions of hemodynamic and pulmonary stability. Under unstable conditions, reevaluation of patient status and increased caution in the interpretation of results are required.

KEYWORDS Transcutaneous carbon dioxide, end-tidal carbon dioxide, blood gas analysis, acute lung injury, swine, burns, inhalation injury

#### INTRODUCTION

Measurement of the partial pressure of carbon dioxide in arterial blood ( $Paco_2$ ) remains the criterion standard for evaluating the adequacy of ventilation. In certain populations, such as trauma patients with head injury, providing appropriate ventilation (avoiding either overventilation or underventilation) has been shown to save lives (1–2). Obtaining blood gas samples, however, is invasive and requires special equipment, which is

DOI: 10.1097/SHK.0b013e318292c331

Copyright © 2013 by the Shock Society

means of estimating  $Paco_2$  would be useful (3). Several such methods have been developed. End-tidal carbon dioxide (Etco<sub>2</sub>) monitoring is performed with increasing frequency in injured patients (4.5). However,

often impractical in prehospital settings. Thus, a noninvasive

with increasing frequency in injured patients (4,5). However, this method may become inaccurate in unstable trauma patients (6). The  $Paco_2$ -Etco\_2 gradient is related to the physiologic dead space as described by Enghoff modification of the Bohr equation (7).  $Paco_2$ -Etco\_2 gradient increases with increasing physiologic dead space, due to lung injuries or decreased pulmonary perfusion.

Another noninvasive technique that can be used as a surrogate for  $Paco_2$  is the measurement of the transcutaneous partial pressure of  $CO_2$  (tPco<sub>2</sub>). This technique was pioneered by Severinghaus (8) in the 1960s using a temperature-stabilized heated electrode. Transcutaneous gas tension is a function of dermal capillary blood and in turn of arterial blood flow. Previous studies by Nishiyama et al. (9, 10) demonstrated good correlations between transcutaneous  $CO_2$  measurements and  $Paco_2$  in adults undergoing general anesthesia. Several studies, performed in critically ill adults (11–14) as well as in adults undergoing noninvasive ventilation (15, 16), reported good correlation between tPco<sub>2</sub> measurements and Paco<sub>2</sub>. Most notably, a study by Hinkelbein et al. (13) demonstrated the feasibility of tPco<sub>2</sub> monitoring in critically ill adults during interhospital transport.

Address reprint requests to Leopoldo C. Cancio, MD, US Army Institute of Surgical Research, 3698 Chambers Pass, Ft Sam Houston, TX. E-mail: lee.cancio@us.army.mil.

All work has been performed at the US Army Institute of Surgical Research, Ft Sam Houston, Texas.

This study has been conducted in compliance with the Animal Welfare Act, the Implementing Animal Welfare Regulations and in accordance with the principles of the Guide for the Care and Use of Laboratory Animals.

This study has been approved by the US Army Institute of Surgical Research Institutional Animal Care and Use Committee.

This study was funded by the Combat Critical Care Engineering Task Area, Combat Casualty Care Research Area Directorate, US Army Medical Research, and Materiel Command.

This research was presented at the American Heart Association Resuscitation Science Symposium; November 4, 2012; Los Angeles, California.

The authors have not disclosed any potential conflicts of interest.

The opinions or assertions contained herein are the private views of the authors and are not to be construed as official or as reflecting the views of the Department of the Army or the Department of Defense.

	Form Approved OMB No. 0704-0188								
Public reporting burden for the collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to a penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number.									
1. REPORT DATE 01 JUL 2013			3. DATES COVERED -						
4. TITLE AND SUBTITLE			5a. CONTRACT NUMBER						
Noninvasive Carbon Dioxide Monitoring in a Porcine Model of Acute					5b. GRANT NUMBER				
Lung Injury Due u	o Smoke Innalation	5c. PROGRAM ELEMENT NUMBER							
6. AUTHOR(S) Belenkiy S., Ivey K. M., Batchinsky A. I., Langer T., Necsoiu C., Baker					5d. PROJECT NUMBER				
					5e. TASK NUMBER				
w., Salinas J., Can	c10 L. C.,		5f. WORK UNIT NUMBER						
7. PERFORMING ORGANI United States Arm Houston, TX	8. PERFORMING ORGANIZATION REPORT NUMBER								
9. SPONSORING/MONITO	RING AGENCY NAME(S) A		10. SPONSOR/MONITOR'S ACRONYM(S)						
		11. SPONSOR/MONITOR'S REPORT NUMBER(S)							
12. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release, distribution unlimited									
13. SUPPLEMENTARY NOTES									
14. ABSTRACT									
15. SUBJECT TERMS									
16. SECURITY CLASSIFIC	17. LIMITATION OF	18. NUMBER	19a. NAME OF						
a. REPORT <b>unclassified</b>	b. ABSTRACT unclassified	c. THIS PAGE unclassified	UU	6	RESPONSIBLE PERSON				

Standard Form 298 (Rev. 8-98) Prescribed by ANSI Std Z39-18

# 496 SHOCK Vol. 39, No. 6

TABLE 1. Animal data						
Weight, kg	$\textbf{45.0} \pm \textbf{2.8}$					
Total smoke volume, L	$\textbf{27.6} \pm \textbf{4.9}$					
Post smoke-inhalation arterial COHb, %	$\textbf{88.3} \pm \textbf{4.1}$					
Time to ALI, h	$\textbf{22.5} \pm \textbf{10.6}$					
Survival time, h	$56.8\pm40.9$					

COHb indicates carboxyhemoglobin.

Nevertheless, there is a paucity of studies comparing  $Etco_2$  and  $tPco_2$  monitoring in settings of evolving acute lung injury (ALI). Therefore, we carried out an animal study to evaluate the usability and accuracy of these methods in a clinically relevant porcine model of ALI secondary to smoke inhalation and burns.

#### MATERIALS AND METHODS

This study was approved by the US Army Institute of Surgical Research Animal Care and Use Committee. It was conducted in compliance with the Animal Welfare Act and the implementing Animal Welfare Regulations and in accordance with the principles of the *Guide for the Care and Use of Laboratory Animals*.

#### Animal preparation

For this study, we selected a subgroup of eight female, nonpregnant, Yorkshire pigs from an ongoing study that investigates treatment of acute respiratory distress syndrome (ARDS) due to smoke inhalation and burns. Total intrave nous anesthesia with ketamine HCL (20 30 mg/kg per hour), midazolam HCL (1.0 1.5 mg/kg per hour), and propofol (100 µg/kg per hour) was used during the experiment. All animals underwent tracheostomy and placement of central and arterial lines. After instrumentation, animals were allowed to recover for 2 to 3 h. After stabilization, they received inhalation injury with 22 breaths of pine bark smoke at room temperature (Table 1). This model of smoke inhalation injury has been described previously (17). Smoke injury was followed immedi ately by a 40% total body surface area, full thickness flame burn. Upon com pletion of smoke inhalation and burn, animals were transferred to the animal intensive care unit where they were continuously monitored for the duration of the experiment. Total intravenous anesthesia levels were adjusted to effect as needed to attain no response to painful stimuli. In addition, buprenorphine HCl (0.1 mg/kg) was administered intramuscularly every 6 h for analgesia for the duration of the study. Resuscitation with lactated Ringer's solution was performed by means of a computerized burn resuscitation decision support system for the first 24 h after burn. Subsequently, the fluid rate was adjusted manually to achieve a urinary output of at least 0.5 mL/kg per hour. Controlled mechanical ventilation (CMV) was initiated, with a tidal volume of 10 mL/kg and a respiratory rate adjusted to maintain Paco2 between 35 and 45 mmHg. Before onset of ARDS, positive end expiratory pressure was set to 5 cm H<sub>2</sub>O and remained constant. Fraction of inspired oxygen (Fio2) was titrated to achieve peripheral oxygen saturation (Spo2) at or above 90%. Fiberoptic bronchoscopy for pulmonary toilet was performed at 1, 2, 6, 12, and 24 h after injury, every 24 h thereafter, and anytime airway obstruction was clinically suspected. During suctioning and bronchoscopies, F102 was increased to 100%. After the onset of ARDS, animals were transitioned to low tidal volume CMV, and further venti lator changes, including adjustments of positive end expiratory pressure and Fio2, were made in accordance with the ARDSnet protocol (18). The study was designed to continue for up to 7 days after injury. Experiments were terminated earlier if ARDS did not develop in 72 h after injury or if an animal reached terminal cardiopulmonary failure before 7 days. Animals were killed with intra venous injection of 20 mL Fatal Plus (Vortech Pharmaceuticals, Dearborn, Mich).

Ratio of the partial pressure of oxygen in arterial blood (Pao<sub>2</sub>) to the frac tion of inspired oxygen (PFR) was measured at baseline, 2 h after injury, and every 6 h thereafter. A single set of baseline measurements was recorded after surgery but before the induction of smoke inhalation and burns. Baseline du ration generally was 2 to 3 h and was adequate to provide stabilization time for all vital signs after surgical preparation. We defined ALI as a sustained PFR less than 300 mmHg, and ARDS as a sustained PFR less than 200 mmHg. As an index of pulmonary instability, we recorded the frequency of changes in the respiratory rate (as set on the ventilator) and in the Fto<sub>2</sub>. We defined hemo dynamic instability as mean arterial pressure (MAP) of 60 mmHg or less (or a requirement for norepinephrine infusion to maintain MAP >60 mmHg).

A SenTec Digital Monitor and V Sign digital sensor (SenTec Ag, Therwil, Switzerland), described previously (12, 15, 19), were used for tPco2 monitor ing. One point tPco2 sensor calibration was performed in vitro according to the manufacturer's recommendations utilizing an 8%  $\pm$  0.05%  $CO_2$  standard gas before placement. Calibration was repeated every 6 to 8 h, as indicated by the device. The V Sign sensor was attached to the base of the pig's right auricle utilizing a single use attachment ring. This area was chosen because the skin is not thick and because it is well perfused by a branch of external carotid artery. Before sensor placement, the attachment site was cleansed with iso propyl alcohol, and one drop of SenTec viscous contact gel was applied to the center of the sensor membrane. The V Sign sensor was heated to 43.5°C. Use of a heating element within the electrode elevates local skin temperature and causes hyperemia. Application of heat to the skin surface, promotes "liquefaction" or changes in the matrix of the dermal lipid component in the stratum corneum layer, which increases gas diffusion through the skin (20, 21). It has been reported previously (22) that such changes begin to occur at 41°C and reach optimal levels for maximal gas diffusion at 44°C. Therefore, local hyperemia improves correlation between transcutaneous and arterial CO2 values. Previous publications (23, 24) indicated better Paco2 correlation with increased electrode temperature; therefore, we used maximal temperature (43.5°C) allowed by the device. The SenTec monitor displays tPco2 values in mmHg as measured at 37°C without applying a temperature correction.

For  $Etco_2$  monitoring, a FilterLine H Set  $CO_2$  sidestream sampling line and airway adapter were used, along with a Capnostream 20 system (Oridion Med ical, Jerusalem, Israel). The FilterLine was changed daily or sooner if it became contaminated with secretions.

An i STAT blood gas analysis system (Abbot Point of Care, Princeton, NJ) and EG7+ cartridges were utilized for obtaining  $Paco_2$  values. Paired mea surements of  $Etco_2$  and  $Paco_2$ , and of  $tPco_2$  and  $Paco_2$ , were recorded every 6 h (as well as intermittently when clinically indicated) to assess the adequacy of ventilation. All measurements were done while animals were mechanically ventilated in the CMV mode.

#### Statistical analysis

SigmaPlot for Windows version 12.0 (Systat Software, San Jose, Calif) was used. Overlapping sets of data were analyzed based on (i) respiratory status (before vs. after the onset of ALI) and on (ii) hemodynamic status (stable vs. unstable). All results are expressed as mean  $\pm$  SD. P < 0.05 was considered significant.

Linear regression was performed for pairs of  $Paco_2$  and  $tPco_2$  as well as for pairs of  $Paco_2$  and  $Etco_2$  data. Bias and 95% limits of agreement were cal culated using Bland Altman analysis (25). Student *t* tests and analyses of vari ance with post hoc Holm Sidak multiple comparison tests were used.

#### RESULTS

A total of 454 h of animal intensive care unit care were provided during this study. We recorded 276 paired measurements of Paco<sub>2</sub> and Etco<sub>2</sub>, and 250 paired measurements of Paco<sub>2</sub> and tPco<sub>2</sub>, from eight animals. The average duration of monitoring was 56.8  $\pm$  40.9 h. Etco<sub>2</sub> and tPco<sub>2</sub> measurements were made across a Paco<sub>2</sub> range of 25 to 85 mmHg. Animal data are presented in Table 1.

It is important to highlight that our injury model was fluid rather than static with rapidly declining respiratory status. The period between smoke/burn injury and onset of ALI was marked by pulmonary instability while animals progressed from healthy baseline conditions to ALI over  $22.5 \pm 10.6$  h. The mean number of ventilator changes per hour was greater before onset of ALI ( $3.07 \pm 1.53$ ) than during ALI ( $1.84 \pm 1.68$ , P = 0.02). Similarly, the average decrease in PFR in the first 24 h after injury was 199; in the second 24-h period, it was 100 (Fig. 1). There was a statistically significant decrease in PFR between baseline and 24 h, as well as between 12 and 24 h (P < 0.05), but not between 24 and 48 h (P > 0.05). During the study, two animals required norepinephrine infusion due to hypotension. During the study, we observed no thermal complications due to the placement of a heated tPco<sub>2</sub> sensor on the auricle.



Fig. 1. PFR during the first 48 h of study; at each time point, a single measurement was recorded from each animal (n = 8); error bars represent  $\pm$  SD.

Two partially overlapping sets of data were analyzed based on (i) pulmonary status (before vs. after the onset of ALI) and (ii) cardiovascular status (hemodynamically stable vs. unstable). The results of this analysis, describing the Etco<sub>2</sub>-Paco<sub>2</sub> and tPco<sub>2</sub>-Paco<sub>2</sub> relationships, are presented in Table 2 and are described in the following sections.

## Before versus after the onset of ALI

Linear regression analysis of 105 Etco<sub>2</sub>-Paco<sub>2</sub> pairs, recorded before the onset of ALI (Fig. 2A), demonstrated moderate correlation: Paco<sub>2</sub> = 13.45 + 0.68 \* Etco<sub>2</sub> ( $R^2$  = 0.45, P < 0.0001). There was systemic underreading by capnography (P = 0.02). Bland-Altman analysis revealed a mean bias of 0.91 ± 3.77 mmHg (Fig. 2B).

Analysis of 171 Etco<sub>2</sub>-Paco<sub>2</sub> pairs, recorded after ALI (Fig. 3A), demonstrated low correlation: Paco<sub>2</sub> = 41.51 + 0.34 \* Etco<sub>2</sub> ( $R^2 = 0.12$ , P < 0.0001). There was systemic underreading by capnography (P < 0.001). Mean bias increased to  $14.84 \pm 11.76$  mmHg (Fig. 3B).

Analysis of 88 tPco<sub>2</sub>-Paco<sub>2</sub> pairs, recorded before the onset of ALI (Fig. 2A), revealed moderate correlation:  $Paco_2 = 17.01 +$ 

 $0.56 * tPco_2 (R^2 = 0.51, P < 0.0001)$ . No systemic difference was observed (P = 0.85). Mean bias was  $0.09 \pm 4.56$  mmHg (Fig. 2C).

Analysis of 140 tPco<sub>2</sub>-Paco<sub>2</sub> pairs, recorded after the onset of ALI (Fig. 3A), revealed strong correlation: Paco<sub>2</sub> = 7.74 + 0.83 \* tPco<sub>2</sub> ( $R^2$  = 0.78, P < 0.0001). Slight systemic underreading was present (P = 0.026). Mean bias was 0.03 ± 5.44 mmHg (Fig. 3C).

#### Hemodynamically stable versus unstable

We repeated the analysis based on hemodynamic status. Linear regression of 233 Etco<sub>2</sub>-Paco<sub>2</sub> pairs under hemodynamically stable conditions (Fig. 4A) demonstrated moderate correlation: Paco<sub>2</sub> =  $5.32 + 1.03 * \text{Etco}_2$  ( $R^2 = 0.44$ , P < 0.0001). There was systemic underreading by capnography. The mean bias was  $6.74 \pm 8.26$  mmHg (Fig. 4B).

Analysis of 41 Etco<sub>2</sub>-Paco<sub>2</sub> pairs under hemodynamically unstable conditions (Fig. 5A) did not demonstrate a linear relationship: Paco<sub>2</sub> = 61.4 - 0.16 \* Etco<sub>2</sub> ( $R^2$  = 0.03, P = 0.29). Systemic underreading was present (P < 0.001). Mean bias was 25.59 ± 15.30 mmHg (Fig. 5B).

Analysis of 212 tPco<sub>2</sub>-Paco<sub>2</sub> pairs under hemodynamically stable conditions (Fig. 4A) demonstrated strong correlation: Paco<sub>2</sub> = 7.86 + 0.82 \* tPco<sub>2</sub> ( $R^2$  = 0.80, P < 0.0001). There was no systemic difference between the two techniques (P = 0.29). Mean bias was  $-0.35 \pm 4.75$  mmHg (Fig. 4C).

Finally, analysis of 38 tPco<sub>2</sub>-Paco<sub>2</sub> pairs under hemodynamically unstable conditions (Fig. 5A) demonstrated poor correlation: Paco<sub>2</sub> = 33.60 + 0.3 \* tPco<sub>2</sub> ( $R^2$  = 0.39, P < 0.0001). There was systemic overreading (P < 0.001). Mean bias was  $-11.32 \pm 14.87$  mmHg (Fig. 5C).

## DISCUSSION

In this study, we evaluated the utility of two commonly used noninvasive carbon dioxide monitoring technologies in a severely injured, mechanically ventilated porcine model of ALI due to smoke inhalation and burns. Our principal findings were (i) the period of time between smoke/burn injury and the onset of ALI was characterized by pulmonary instability, manifested by a steady decrease in the PFR and by frequent ventilator changes and other interventions; (ii) both Etco<sub>2</sub> and tPco<sub>2</sub> were moderately correlated with Paco<sub>2</sub> during this period; (iii) after

Variable compared with Paco <sub>2</sub>	Condition	R <sup>2</sup>	P	Bias, mean ± SD (95% limits of agreement), mmHg	Student t test, P
Etco <sub>2</sub>	Pre-ALI	0.45	<0.0001	0.91 ± 3.77 ( 6.48 to 8.31)	0.02
Etco <sub>2</sub>	ALI	0.12	<0.0001	14.84 $\pm$ 11.76 ( $$ 8.21 to 37.89)	<0.001
tPco <sub>2</sub>	Pre-ALI	0.51	<0.0001	$0.09 \pm 4.56$ ( $8.85$ to 9.03)	0.85
tPco <sub>2</sub>	ALI	0.78	<0.0001	$0.03 \pm 5.44$ ( 10.64 to 10.70)	0.026
Etco <sub>2</sub>	HD stable	0.44	<0.0001	6.74 $\pm$ 8.26 ( $$ 9.45 to 22.95)	<0.001
Etco <sub>2</sub>	HD unstable	0.03	0.29	25.59 $\pm$ 15.30 ( $$ 4.39 to 55.57)	<0.001
tPco <sub>2</sub>	HD stable	0.80	<0.0001	$0.35 \pm 4.75  ( 9.67 \text{ to } 8.97)$	0.29
tPco <sub>2</sub>	HD unstable	0.39	<0.0001	11.32 $\pm$ 14.87 ( $$ 40.46 to 17.82)	<0.001
	Variable compared with Paco <sub>2</sub> Etco <sub>2</sub> Etco <sub>2</sub> tPco <sub>2</sub> Etco <sub>2</sub> Etco <sub>2</sub> Etco <sub>2</sub> tPco <sub>2</sub> tPco <sub>2</sub>	Variable compared with Paco2ConditionEtco2Pre-ALIEtco2ALItPco2Pre-ALItPco2ALItPco2HD stableEtco2HD unstabletPco2HD stabletPco2HD stabletPco2HD unstabletPco2HD unstable	Variable compared with Paco2Condition $R^2$ Etco2Pre-ALI0.45Etco2ALI0.12tPco2Pre-ALI0.51tPco2ALI0.78Etco2HD stable0.44Etco2HD unstable0.03tPco2HD stable0.80tPco2HD unstable0.39	Variable compared with Paco <sub>2</sub> Condition $R^2$ $P$ Etco <sub>2</sub> Pre-ALI         0.45         <0.0001	$\begin{tabular}{ c c c c } \hline Wariable compared \\ \hline With Paco_2 & Condition & $R^2$ & $P$ & $mmHg$ \\ \hline (95\% limits of agreement), $mmHg$ \\ \hline (95\%$

TABLE 2. Etco<sub>2</sub>-Paco<sub>2</sub> and tPco<sub>2</sub>-Paco<sub>2</sub> relationships

HD indicates hemodynamically.



Fig. 2. A, Scatter plot and linear regression analysis comparing Etco<sub>2</sub> and Paco<sub>2</sub> as well as tPco<sub>2</sub> and Paco<sub>2</sub> before the onset of ALI. Solid line represents linear regression between tPco<sub>2</sub> and Paco<sub>2</sub>. Dashed line represents linear regression between Etco<sub>2</sub> and Paco<sub>2</sub>. B and C demonstrate Bland Altman analysis between Etco<sub>2</sub> Paco<sub>2</sub> and tPco<sub>2</sub> Paco<sub>2</sub> pairs; middle solid line indicates the mean difference (bias), and outer solid lines represent limits of agreement (mean ± 1.96 SD) between two methods.

the onset of ALI,  $Etco_2$  became relatively inaccurate, whereas  $tPco_2$  did not; (iv) hemodynamic instability caused  $Etco_2$  values to lose their linear relationship with Paco<sub>2</sub>. Based on these observations, we conclude that both methods can be useful in the monitoring of patients with developing severe lung injury, but that caution should be used in the interpretation of results when patients are changing rapidly.

Good correlation between  $Etco_2$  and  $Paco_2$  across all ranges of dead space was reported by McSwain et al. (7). Our previous study (26) demonstrated close correlation between  $Etco_2$ and  $Paco_2$  in a porcine model of severe chest injury during periods of hemodynamic stability and during settings of hypoventilation and hyperventilation induced by varying tidal volumes and different respiratory rates in healthy swine. At the same time, results were not as promising when accuracy of capnography was evaluated in trauma patients (6). In addition, several studies compared accuracy of end-tidal and  $tPco_2$ monitoring in human patients (13, 27, 28), favoring  $tPco_2$ . In our study,  $Etco_2$  systemically underestimated  $Paco_2$ . Presence of systemic bias was confirmed by paired Student *t* test results (Table 2). This is to be expected, particularly in patients with developing ALI, in whom dead space progressively increases, thereby leading to an increase in Paco<sub>2</sub>-Etco<sub>2</sub> gradient.

There was moderate correlation between tPco<sub>2</sub> and Paco<sub>2</sub> ( $R^2 = 0.51$ ) before ALI. This correlation became more linear ( $R^2 = 0.78$ ) after onset of ALI. We expected better correlation during pre-ALI stage, and such findings were surprising. We speculate that lower tPco<sub>2</sub>-Paco<sub>2</sub> correlation before onset of ALI was possibly due to more frequent ventilator changes: approximately  $3.07 \pm 1.53$  changes per hour in the initial phase of the experiment, compared with  $1.84 \pm 1.68$  changes per hour after the onset of ALI. Also, more frequent suctioning was required in the first 24 h after injury due to copious secretions. Because it can cause alveolar collapse and loss of recruitment, repeated suctioning may have had a destabilizing effect on Paco<sub>2</sub> correlation. In addition, the tPco<sub>2</sub> sensor was more frequently disconnected and removed from the animals during the first 24 h of the experiment to accommodate bronchoscopies.

When the tPco<sub>2</sub>-Paco<sub>2</sub> relationship was examined based on hemodynamic status, it was strongly linear ( $R^2 = 0.80$ ) with



Fig. 3. A, Scatter plot and linear regression analysis between Etco<sub>2</sub> and Paco<sub>2</sub> as well as tPco<sub>2</sub> and Paco<sub>2</sub> after the onset of ALI. Solid line represents linear regression between tPco<sub>2</sub> and Paco<sub>2</sub>. Dashed line represents linear regression between Etco<sub>2</sub> and Paco<sub>2</sub>. B and C demonstrate Bland Altman analysis between Etco<sub>2</sub> Paco<sub>2</sub> and tPco<sub>2</sub> Paco<sub>2</sub> pairs; middle solid line indicates the mean difference (bias), and outer solid lines represent limits of agreement (mean ± 1.96 SD) between two methods.



Fig. 4. A, Scatter plot and linear regression analysis between  $Etco_2$  and  $Paco_2$  as well as  $tPco_2$  and  $Paco_2$  during hemodynamic stability (HD stable). Solid line represents linear regression between  $tPco_2$  and  $Paco_2$ . Dashed line represents linear regression between  $Etco_2$  and  $Paco_2$ . B and C demonstrate Bland Altman analysis between  $Etco_2$   $Paco_2$  and  $tPco_2$   $Paco_2$  pairs; middle solid line indicates the mean difference (bias), and outer solid lines represent limits of agreement (mean  $\pm$  1.96 SD) between two methods.

minimal bias ( $-0.35 \pm 4.75$  mmHg) as long as MAP remained greater than 60 mmHg. Our results confirm those reported previously in human studies (11, 13, 20, 27, 28) that demonstrated similar correlation. tPco<sub>2</sub> had a less linear correlation to Paco<sub>2</sub> during hemodynamic instability ( $R^2 = 0.39$ ). In addition, transcutaneous monitoring tended to overestimate Paco<sub>2</sub> under these conditions. Previous reporting (29, 30) indicated that vasoactive medications did not have a significant effect on tPco<sub>2</sub> monitoring accuracy. However, our results were different, possibly because of the sensor location on the right auricle. Although the animals were hypotensive, their ears were cold to touch, and there was a visible area of vascular congestion. Preferred locations for tPco<sub>2</sub> monitoring have been reported in humans (9, 10), but no such location has been defined in pigs.

Because we used a subset of animals for this work from a larger study, which pursued treatment of ARDS due to smoke inhalation and burns, the samples we obtained were convenience samples. This limited our ability to further investigate effects of suctioning, ventilator changes, and sensor disconnections on stabilizations/steady state times of the measured variables. Future studies designed specifically to address these limitations may be warranted.

Based on our data, transcutaneous capnometry is an acceptable trend-monitoring the tool in settings of lung injury in hemodynamically stable patients and may be useful in a prehospital environment. However, the current generation of tPco2 sensors has some limitations. First, the tPco2 sensor requires stabilization time after placement. Several articles (31, 32) recommended a 20-min stabilization time while using the SenTec Monitor. Second, to maintain accuracy, the sensor has to be regularly recalibrated in vitro. Recalibration frequency appears to depend on sensor temperature. At 43.5°C, recalibrations were required every 6 to 8 h. On average, recalibration can be completed in approximately 3 min. Afterward, the sensor is repositioned on a patient and again requires stabilization time. In our experience, about 25 to 30 min was spent on moving, cleaning, recalibrating, replacing, and waiting for the sensor to stabilize. In this study, because of previously described limitations, we could not establish the precise stabilization timing necessary for improved accuracy; further study should be considered to



Fig. 5. A, Scatter plot and linear regression analysis between Etco<sub>2</sub> and Paco<sub>2</sub> as well as tPco<sub>2</sub> and Paco<sub>2</sub> during hemodynamic instability (HD unstable). Solid line represents linear regression between tPco<sub>2</sub> and Paco<sub>2</sub>. Dashed line represents linear regression between Etco<sub>2</sub> and Paco<sub>2</sub>. B and C demonstrate Bland Altman analysis between Etco<sub>2</sub> Paco<sub>2</sub> and tPco<sub>2</sub> Paco<sub>2</sub> pairs; middle solid line indicates the mean difference (bias), and outer solid lines represent limits of agreement (mean ± 1.96 SD) between two methods.

# 500 SHOCK Vol. 39, No. 6

address this issue. Also, we should point out that a prolonged stabilization time may limit utilization of  $tPco_2$  monitors in prehospital environments. Finally, the sensor membrane requires replacement every 42 days and between patients. Currently, several companies are developing next-generation solid-state  $tPco_2$  sensors that may overcome these shortcomings. Given promising results obtained from the use of a transcutaneous  $CO_2$  sensor, these new developments will be a welcomed addition to critical care monitoring armamentarium as well as potentially opening new possibilities for servo controlling mechanical ventilators and extracorporeal life support devices.

## CONCLUSIONS

In a porcine model of ALI due to smoke inhalation and burns, transcutaneous  $CO_2$  monitoring is an acceptable noninvasive surrogate for  $Paco_2$  under hemodynamically stable conditions and can be useful as a trend monitoring tool. However,  $tPco_2$  readings should be correlated with  $Paco_2$  with increased frequency when a patient's condition is dynamically changing, for example, during periods of hemodynamic instability. End-tidal  $CO_2$  monitoring offers an ease-of-use advantage over the current generation of transcutaneous  $CO_2$  monitors. However,  $Etco_2$  readings should be correlated with  $Paco_2$  with increased frequency during offers an ease-of-use advantage over the current generation of transcutaneous  $CO_2$  monitors. However,  $Etco_2$  readings should be correlated with  $Paco_2$  with increased frequency during evolution of lung injury (changes in dead space).

## ACKNOWLEDGMENTS

The authors thank Oridion, Inc, for providing support for this study, under a Cooperative Research and Development Agreement with the US Army Institute of Surgical Research.

#### REFERENCES

- 1. Davis DP, Idris AH, Sise MJ, Kennedy F, Eastman AB, Velky T, Vilke GM, Hoyt DB: Early ventilation and outcome in patients with traumatic brain injury. *Crit. Care Med* 34(4):1202 1208, 2006.
- Neumar RW, Otto CW, Link MS, Kronick SL, Shuster M, Callaway CW, Kudenchuk PJ, Omato JP, McNally B, Silvers SM, et al.: Part 8: Adult Advanced Cardiovascular Life Support: 2010 American Heart Association Guidelines for Cardiopulmonary Resuscitation and Emergency Cardiovascular Care. *Circulation* 122(18 suppl 3):S729 S767, 2010.
- Cancio LC, Chung KK: The role of normoventilation in improving traumatic brain injury outcomes. US Army Med Dep J Apr Jun:49 54, 2011.
- Committee on Trauma, American College of Surgeons: Advanced Trauma Life Support for Doctors (Student Course Manual). 9th ed. Chicago, IL: American College of Surgeons, 2012.
- Guidelines for the Field Management of Combat-Related Head Trauma. New York, NY: The Brain Trauma Foundation; 2005. Available at: https:// www.braintrauma.org/pdf/protected/btf field management guidelines.pdf. Accessed August 17, 2012.
- Warner KJ, Cuschieri J, Garland B, Carlbom D, Baker David, Copass MK, Jurkovich GJ, Bulger EM: The utility of early end-tidal capnography in monitoring ventilation status after severe injury. *J Trauma* 66(1):26 31, 2009.
- McSwain SD, Hamel DS, Smith PB, Gentile MA, Srinivasan S, Meliones JN, Cheifetz IM: End-tidal and arterial carbon dioxide measurements correlate across all levels of physiologic dead space. *Respir Care* 55(3):288–293, 2010.
- Severinghaus JW. Methods of measurement of blood and gas carbon dioxide during anesthesia. Anesthesiology 21:717 726, 1960.

- Nishiyama T, Nakamura S, Yamashita K: Comparison of the transcutaneous oxygen and carbon dioxide tension in different electrode locations during general anesthesia. *Eur J Anesthesiology* 23:1049 1055, 2006.
- Nishiyama T, Kohono Y, Koishi K: Comparison of ear and chest probes in transcutaneous carbon dioxide pressure measurements during general anesthesia in adults. *J Clin Monit Comput* 25:323 328, 2011.
- Bendejelid K, Schutz N, Stotz M, Gerard I, Suter PM, Romand JA: Transcutaneous Pco<sub>2</sub> monitoring in critically ill adults: clinical evaluation of a new sensor. *Crit Care Med* 33(10):2203 2206, 2005.
- Rodriguez P, Lellouche F, Aboab J, Buisson CB, Brochard L: Transcutaneous arterial carbon dioxide pressure monitoring in critically ill adult patients. *Intensive Care Med* 134:309 312, 2006.
- Hinkelbein J, Floss F, Denz C, Krieter H: Accuracy and precision of three different methods to determine Pco<sub>2</sub> (Paco<sub>2</sub> vs. PECO<sub>2</sub> vs. PECO<sub>2</sub>) during interhospital ground transport of critically ill and ventilated adults. *J Trauma* 65(1):10 18, 2008.
- Nicolini A, Ferrari M: Evaluation of transcutaneous carbon dioxide monitor in patients with acute respiratory failure. Ann Thorac Med 6(4):217 220, 2011.
- Storre J, Steurer B, Hans-Joachim K, Dreher M, Windisch W: Transcutaneous Pco<sub>2</sub> monitoring during initiation of noninvasive ventilation. *Chest* 132(6): 1810 1816, 2007.
- Senn O, Clarenbach F, Kaplan V, Maggiorini M, Bloch KE: Monitoring carbon dioxide tension and arterial oxygen saturation by a single earlobe sensor in patients with critical illness or sleep apnea. *Chest* 128(3):1291 1296, 2005.
- Batchinsky A, Burkett S, Zanders T, Chung KK, Regn DD, Jordan BS, Necsoiu C, Nguyen R, Hanson MA, Morris MJ, et al.: Comparison of airway pressure release ventilation to conventional mechanical ventilation in the early management of smoke inhalation injury in swine. *Crit Care Med* 39(10) 2314 2321, 2011.
- The Acute Respiratory Distress Syndrome Network: Ventilation with lower tidal volumes as compared with traditional tidal volumes for acute lung injury and the acute respiratory distress syndrome. N Engl J Med 342:1301 1308, 2000.
- Kocher S, Rohlin R, Tschupp A: Performance of a digital Pco<sub>2</sub>/Spo<sub>2</sub> ear sensor. J Clin Monit Comput 18(2):75 79, 2004.
- Eberhard P: The design, use, and results of transcutaneous carbon dioxide analysis: current and future directions. *Anesth Analg* 105(6):s48 s52, 2007.
- Franklin ML: Transcutaneous measurement of partial pressure of oxygen and carbon dioxide. *Respir Care Clin N Am* 1(1):119 131, 1995.
- 22. Lubbers DW: Theoretical basis of the transcutaneous blood gas measurements . *Crit Care Med* 9(10):721 733, 1981.
- Nishiyama T, Nakamura S, Yamashita K: Effects of electrode temperature of the new monitor, TCM4, on the measurement of transcutaneous oxygen and carbon dioxide tension. J Anesth 20(4):331–334, 2006.
- Sorensen L, Brage-Andersen L, Greisen G: Effects of the transcutaneous electrode temperature on the accuracy of transcutaneous carbon dioxide tension. *Scand J Clin Lab Invest* 71(7):548 552, 2011.
- Bland JM, Altman DJ: Statistical methods for assessing agreements between two methods of clinical measurements. *Lancet* 1(8476):307–310, 1986.
- Isbell CL, Batchinsky AI, Hetz KM, Baker WL, Cancio LC: Correlation between capnography and arterial carbon dioxide before, during and after severe chest injury in swine. *Shock* 37(1):103–109, 2012.
- Hirabayashi M, Fujiwara C, Ohtani N, Kagawa S, Kamide M: Transcutaneous Pco<sub>2</sub> monitors are more accurate than end-tidal Pco<sub>2</sub> monitors. J Anesth 23(2):198 202, 2009.
- Phan CQ, Tremper KK, Lee SE, Barker SJ: Noninvasive monitoring of carbon dioxide: a comparison of the partial pressure of transcutaneous and end-tidal carbon dioxide with the partial pressure of arterial carbon dioxide. *J Clin Monit* 3(3):149 154, 1987.
- Palmisano B, Severinghaus J: Transcutaneous Pco<sub>2</sub> and Po<sub>2</sub>: a multicenter study of accuracy. J Clin Monit 6(3):189–195, 1990.
- Janssens JP, Howarth-Frey C, Chevrolet JC: Transcutaneous Pco<sub>2</sub> to monitor noninvasive mechanical ventilation in adults: assessment of a new transcutaneous Pco<sub>2</sub> device. *Chest.* 113(3):768–773, 1998.
- Domingo C, Canturri E, Lujan M, Moreno O, Espuelas H, Marin A: Transcutaneous measurement of partial pressure of carbon dioxide and oxygen saturation: validation of the SenTec Monitor. *Arch Bronconeumol* 42(5):246 251, 2006.
- Janssens JP, Perrin E, Bennani I, DeMuralt B, Titelion V, Picaud C: Is continuous transcutaneous monitoring of Pco<sub>2</sub> (tcPco<sub>2</sub>) over 8 h reliable in adults? *Respir Med* 95(5) 331 335, 2001.