

AD755750

OFFICE OF NAVAL RESEARCH

Contract N00014-67-A-0467-0023

Project No. NR 056-515

TECHNICAL REPORT NO. 4

Quantitative Analysis of Aqueous  $\text{NO}_2^-$ - $\text{NO}_3^-$  Solutions  
by Infrared Internal Reflection Spectroscopy

by

R. T. Yang and M. J. D. Low

Prepared for Publication

in

Analytical Chemistry

January 30, 1973

Reproduced by  
NATIONAL TECHNICAL  
INFORMATION SERVICE  
U S Department of Commerce  
Springfield VA 22151

Reproduction in whole or in part is permitted for  
any purpose of the United States Government.

Approved for Public Release: Distribution Unlimited.

Unclassified  
Security Classification

| DOCUMENT CONTROL DATA - R&D                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 |                                                                                                                           |                                                                                                                    |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------|
| (Security classification of title, body of abstract and indexing annotation must be entered when the overall report is classified)                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                          |                                                                                                                           |                                                                                                                    |
| 1 ORIGINATING ACTIVITY (Corporate author)<br>Department of Chemistry<br>New York University<br>New York, N. Y. 10453                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        |                                                                                                                           | 2a REPORT SECURITY CLASSIFICATION<br><b>Unclassified</b><br>2b GROUP                                               |
| 3 REPORT TITLE<br>Quantitative Analysis of Aqueous $\text{NO}_2^-$ - $\text{NO}_3^-$ Solutions by Infrared Internal Reflection Spectroscopy                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 |                                                                                                                           |                                                                                                                    |
| 4 DESCRIPTIVE NOTES (Type of report and inclusive dates)<br>Technical Report                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |                                                                                                                           |                                                                                                                    |
| 5 AUTHOR(S) (Last name, first name, initial)<br>Yang, R. T. and Low, M. J. D.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |                                                                                                                           |                                                                                                                    |
| 6. REPORT DATE<br>1/30/73                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   | 7a TOTAL NO OF PAGES<br>23                                                                                                | 7b NO OF PAGES<br>13                                                                                               |
| 8a CONTRACT OR GRANT NO.<br>N00014-67-A-0467-0023<br>b. PROJECT NO<br>NR-056-515<br>c.<br>d.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                | 9a ORIGINATOR'S REPORT NUMBER(S)<br>ONR - 4<br>9b OTHER REPORT NO(S) (Any other numbers that may be assigned this report) |                                                                                                                    |
| 10. AVAILABILITY/LIMITATION NOTICES<br>Approved for Publication Release; Distribution Unlimited                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             |                                                                                                                           |                                                                                                                    |
| 11. SUPPLEMENTARY NOTES                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     |                                                                                                                           | 12 SPONSORING MILITARY ACTIVITY<br>Department of the Navy<br>Office of Naval Research<br>Arlington, Virginia 22217 |
| 13 ABSTRACT<br>Infrared spectra of aqueous $\text{NO}_2^-$ and $\text{NO}_3^-$ solutions, and of binary mixtures, were recorded using internal reflection techniques and a Fourier transform spectrometer. $\text{NO}_2^-$ could be detected at a concentration of 0.02 M, and $\text{NO}_3^-$ at 0.001 M. However, $\text{NO}_3^-$ also became adsorbed on the Ge prism. After making corrections for $\text{NO}_3^-$ adsorption and for overlapping of $\text{NO}_2^-$ and $\text{NO}_3^-$ bands, it was possible to analyze binary mixtures with concentrations as low as $[\text{NO}_3^-] = 0.03 \text{ M}$ and $[\text{NO}_2^-] = 0.05 \text{ M}$ . By adjusting the optics so that the usually totally absorbing water stretching band had a transmittance of about 50%, it was possible to observe N-H and C-H stretching bands of solutes. Dissolved materials cause the water structure and consequently the water absorption to change, causing distortions of the background of ratioed spectra of aqueous solutions. These distortions may influence qualitative and quantitative observations. |                                                                                                                           |                                                                                                                    |

DD FORM 1473  
JAN 64

I

Unclassified  
Security Classification

**Unclassified**  
Security Classification

| 14. KEY WORDS                              | LINK A |    | LINK B |    | LINK C |    |
|--------------------------------------------|--------|----|--------|----|--------|----|
|                                            | ROLE   | WT | ROLE   | WT | ROLE   | WT |
| Quantitative analysis of aqueous solutions |        |    |        |    |        |    |
| Infrared spectra of water solutions        |        |    |        |    |        |    |
| Internal reflection Spectroscopy           |        |    |        |    |        |    |
| Fourier transform Spectroscopy             |        |    |        |    |        |    |
| Aqueous solutions, infrared analysis of    |        |    |        |    |        |    |
| Nitrate solutions                          |        |    |        |    |        |    |
| Nitrite solutions                          |        |    |        |    |        |    |
| Water, infrared spectra of                 |        |    |        |    |        |    |
| Water solutions, infrared spectra of       |        |    |        |    |        |    |

**INSTRUCTIONS**

1. **ORIGINATING ACTIVITY:** Enter the name and address of the contractor, subcontractor, grantee, Department of Defense activity or other organization (*corporate author*) issuing the report.

2a. **REPORT SECURITY CLASSIFICATION:** Enter the overall security classification of the report. Indicate whether "Restricted Data" is included. Marking is to be in accordance with appropriate security regulations.

2b. **GROUP:** Automatic downgrading is specified in DoD Directive 5200.10 and Armed Forces Industrial Manual. Enter the group number. Also, when applicable, show that optional markings have been used for Group 3 and Group 4 as authorized.

3. **REPORT TITLE:** Enter the complete report title in all capital letters. Titles in all cases should be unclassified. If a meaningful title cannot be selected without classification, show title classification in all capitals in parentheses immediately following the title.

4. **DESCRIPTIVE NOTES:** If appropriate, enter the type of report, e.g., interim, progress, summary, annual, or final. Give the inclusive dates when a specific reporting period is covered.

5. **AUTHOR(S):** Enter the name(s) of author(s) as shown on or in the report. Enter last name, first name, middle initial. If military, show rank and branch of service. The name of the principal author is an absolute minimum requirement.

6. **REPORT DATE:** Enter the date of the report as day, month, year, or month, year. If more than one date appears on the report, use date of publication.

7a. **TOTAL NUMBER OF PAGES:** The total page count should follow normal pagination procedures, i.e., enter the number of pages containing information.

7b. **NUMBER OF REFERENCES:** Enter the total number of references cited in the report.

8a. **CONTRACT OR GRANT NUMBER:** If appropriate, enter the applicable number of the contract or grant under which the report was written.

8b, 8c, & 8d. **PROJECT NUMBER:** Enter the appropriate military department identification, such as project number, subproject number, system numbers, task number, etc.

9a. **ORIGINATOR'S REPORT NUMBER(S):** Enter the official report number by which the document will be identified and controlled by the originating activity. This number must be unique to this report.

9b. **OTHER REPORT NUMBER(S):** If the report has been assigned any other report numbers (*either by the originator or by the sponsor*), also enter this number(s).

10. **AVAILABILITY/LIMITATION NOTICES:** Enter any limitations on further dissemination of the report, other than those

imposed by security classification, using standard statements such as:

- (1) "Qualified requesters may obtain copies of this report from DDC."
- (2) "Foreign announcement and dissemination of this report by DDC is not authorized."
- (3) "U. S. Government agencies may obtain copies of this report directly from DDC. Other qualified DDC users shall request through \_\_\_\_\_."
- (4) "U. S. military agencies may obtain copies of this report directly from DDC. Other qualified users shall request through \_\_\_\_\_."
- (5) "All distribution of this report is controlled. Qualified DDC users shall request through \_\_\_\_\_."

If the report has been furnished to the Office of Technical Services, Department of Commerce, for sale to the public, indicate this fact and enter the price, if known.

11. **SUPPLEMENTARY NOTES:** Use for additional explanatory notes.

12. **SPONSORING MILITARY ACTIVITY:** Enter the name of the departmental project office or laboratory sponsoring (*paying for*) the research and development. Include address.

13. **ABSTRACT:** Enter an abstract giving a brief and factual summary of the document indicative of the report, even though it may also appear elsewhere in the body of the technical report. If additional space is required a continuation sheet shall be attached.

It is highly desirable that the abstract of classified reports be unclassified. Each paragraph of the abstract shall end with an indication of the military security classification of the information in the paragraph, represented as (TS), (S), (C), or (U).

There is no limitation on the length of the abstract. However, the suggested length is from 150 to 225 words.

14. **KEY WORDS:** Key words are technically meaningful terms or short phrases that characterize a report and may be used as index entries for cataloging the report. Key words must be selected so that no security classification is required. Identifiers, such as equipment model designation, trade name, military project code name, geographic location, may be used as key words but will be followed by an indication of technical context. The assignment of links, roles, and weights is optional.

II

**UNCLASSIFIED**  
Security Classification

QUANTITATIVE ANALYSIS OF AQUEOUS  $\text{NO}_2^-$ - $\text{NO}_3^-$  SOLUTIONS BY  
INFRARED INTERNAL REFLECTION SPECTROSCOPY

R. T. Yang and M. J. D. Low

Department of Chemistry  
New York University  
New York, N. Y. 10453

Abstract

Infrared spectra of aqueous  $\text{NO}_2^-$  and  $\text{NO}_3^-$  solutions, and of binary mixtures, were recorded using internal reflection techniques and a Fourier transform spectrometer.  $\text{NO}_2^-$  could be detected at a concentration of 0.02 M, and  $\text{NO}_3^-$  at 0.001 M. However,  $\text{NO}_3^-$  also became adsorbed on the Ge prism. After making corrections for  $\text{NO}_3^-$  adsorption and for overlapping of  $\text{NO}_2^-$  and  $\text{NO}_3^-$  bands, it was possible to analyze binary mixtures with concentrations as low as  $[\text{NO}_3^-] = 0.03 \text{ M}$  and  $[\text{NO}_2^-] = 0.05 \text{ M}$ . By adjusting the optics so that the usually totally absorbing water stretching band had a transmittance of about 50%, it was possible to observe N-H and C-H stretching bands of solutes. Dissolved materials cause the water structure and consequently the water absorption to change, causing distortions of the background of ratioed spectra of aqueous solutions. These distortions may influence qualitative and quantitative observations.

---

The utility of infrared internal reflection spectroscopy (IRS) in providing information about adsorption from organic solvent at the liquid-solid interface (1) led us to attempt to study adsorption from aqueous solution, using a Ge IRS prism as adsorbent as well as for optical purposes. It was found that  $\text{NO}_3^-$  became adsorbed on Ge, prompting us to examine the quantitative aspects of the infrared

analysis of aqueous solutions of  $\text{NO}_3^-$ ,  $\text{NO}_2^-$ , and of  $\text{NO}_3^-$ - $\text{NO}_2^-$  mixtures.

IRS analysis of aqueous  $\text{NaNO}_3$  solutions has been described by Katlafsky and Keller (2), who found a linear variation,  $[\text{NO}_3^-] = aA_{\text{NO}_3^-}$ , of the  $1335\text{-cm}^{-1}$   $\text{NO}_3^-$  absorbance over the 5-20% concentration range.  $A_{\text{NO}_3^-}$  is the  $\text{NO}_3^-$  absorbance. Wilhite and Ellis (3), using IRS, reported a similar linear variation for the 0.1-1 M  $\text{NO}_3^-$  range, but another linear relation,  $[\text{NO}_3^-] = b + cA_{\text{NO}_3^-}$ , for the 0.01-0.1 M  $\text{NO}_3^-$  range. Ahliah and Mooney (4) used IRS to analyze aqueous  $\text{NaNO}_3$ - $\text{NaNO}_2$  solutions; their calibration plots were linear but their graphs (5) show them to be  $[\text{NO}_3^-] = d + eA_{\text{NO}_3^-}$  and  $[\text{NO}_2^-] = f + gA_{\text{NO}_2^-}$ . Values of d and f are near 0.08 and 0.1 absorbance. As our observations differ somewhat from those made earlier, and the method is of potential utility, we describe our IRS study of aqueous  $\text{NO}_3^+$ ,  $\text{NO}_3^-$ , and  $\text{NO}_3^-$ - $\text{NO}_2^-$  solutions.

#### EXPERIMENTAL

A Ge single-pass plate, 1 mm thick, was used in conjunction with a suitable variable-angle attachment described elsewhere (1). Spectra having a constant resolution of  $8\text{ cm}^{-1}$  were recorded with a modified Digilab, Inc. Model FTS-14 Fourier transform spectrometer (6). All single-beam spectra were obtained by summing 400 interferograms. After the cell containing the Ge plate was installed, the sampling compartment was flushed with dry air. The cell remained in place throughout a series of measurements so that its transmission remained constant; small changes in cell position can cause transmission changes which can affect the results obtained when two single-beam spectra are ratioed. Pure water or solutions of various  $\text{NH}_4\text{NO}_2$ ,  $\text{NH}_4\text{NO}_3$ ,  $\text{NaNO}_3$ ,  $\text{NaNO}_2$ , or  $\text{NaNO}_3 + \text{NaNO}_2$  concentrations, prepared from distilled water and reagent-grade chemicals,

were placed into the cell or were withdrawn from it by means of syringes and Teflon tubing permanently attached to the cell.

### RESULTS AND DISCUSSION

Only single-beam spectra were measured. These were stored by the instrument's disc memory, retrieved, and ratioed by computation as required, using the instrument's digital computer. Some examples of single-beam spectra are shown in Fig. 1. The equivalent penetration depth (7), computed by taking the molar extinction coefficient of liquid water to be 81 at  $3450\text{ cm}^{-1}$  (8), was  $3.1\text{ }\mu$  at  $3450\text{ cm}^{-1}$  and about  $8\text{ }\mu$  at  $1600\text{ cm}^{-1}$ . Spectrum A is the "background" of the Ge plate in air. There were prominent  $\text{CO}_2$  bands, and some water vapor bands in the  $1900\text{--}1400\text{ cm}^{-1}$  region, caused by residual water vapor in the spectrometer. When water was in contact with the Ge plate, the prominent OH stretching and deformation bands were superimposed on the background (B, Fig. 1). The single-beam spectra of dilute solutions (C,D, Fig. 1) are similar to the spectrum of water; the bands of the solute are not prominent. The instrument function was removed when two such single-beam spectra were ratioed. If the empty cell was used as reference, the ratioed spectra had the appearance of spectra obtained by the usual double-beam operation, e.g. A, B, Fig. 2. Using water as the reference, the ratioed spectra had the appearance of double-beam, solvent-compensated spectra (C,D, Fig. 2). The slight  $\text{CO}_2$  bands which appeared in some spectra were brought about by changes in the efficiency of the air dryer. The noise above  $3000\text{ cm}^{-1}$  and below  $1000\text{ cm}^{-1}$  in spectra C and D of Fig. 2 was the result of ratioing spectra in regions where the transmission was very low. In spectrum C, the negative OH stretching band, the negative OH deformation band, and the increase in absorption below  $1000\text{ cm}^{-1}$ , which are barely detectable in

spectrum D, are not artifacts. These changes were caused by variations in the absorptivity of the water present in the solution, the water structure being affected by the solutes.

The absorptions of  $\text{NO}_3^-$  and  $\text{NO}_2^-$ , barely detectable in the single-beam spectra, became more recognizable in the ratioed spectra (A,B, Fig. 2); computer-produced scale expansion then made the bands usable. Examples are shown in Fig. 3. Trace A of that figure was obtained by recording two single-beam spectra of water and ratioing them against one another, in order to obtain an estimate of the noise level; it can be taken as a "100% line." Spectrum B shows the  $\text{NO}_3^-$  and  $\text{NO}_2^-$  bands. Spectrum C is a highly scale-expanded segment of spectrum D of Fig. 2, plotted to the same scale as spectrum C of Fig. 3; the noise level is consequently higher.

The ratioing and scale-expansion techniques outlined were used to examine single and binary solutions of various concentrations. Some segments of spectra of  $\text{NH}_4\text{NO}_2$  solutions are shown in Fig. 4. Similar results were obtained with  $\text{NaNO}_2$  solutions, the intensity of the  $1237\text{-cm}^{-1}$   $\text{NO}_2^-$  band decreasing progressively with decreasing  $\text{NO}_2^-$  concentration. A reasonably linear relation between the absorbance of the  $1237\text{-cm}^{-1}$  band and  $\text{NO}_2^-$  concentration was obtained (Fig. 5). The lowest concentration yielding a usable spectrum was 0.02 M.

Similar measurements were carried out with  $\text{NO}_3^-$  solutions. Some results obtained with  $\text{NaNO}_3$  are shown in Fig. 6. The intensity of the  $\text{NO}_3^-$  band decreased progressively as the  $\text{NO}_3^-$  decreased, but remained constant at concentrations less than 0.001 M (Fig. 6A).

With relatively concentrated solutions the  $\text{NO}_3^-$  band was broad and had a maximum at  $1352\text{ cm}^{-1}$  and a shoulder near  $1400\text{ cm}^{-1}$  (Fig. 6B). (The weak symmetric stretching band of  $\text{NO}_3^-$  could also be observed near  $1050\text{ cm}^{-1}$  with 1 M  $\text{NO}_3^-$  solutions.) When the  $\text{NO}_3^-$  concentration declined, the  $1400\text{-cm}^{-1}$  absorption diminished and the

maximum of the residual, sharp band shifted to  $1350\text{ cm}^{-1}$ . The  $\text{NO}_3^-$  band of the more concentrated solutions is more like that of solid  $\text{NaNO}_3$ , for which the asymmetric stretching band is at  $1374\text{ cm}^{-1}$  (9).

The sharp residual  $\text{NO}_3^-$  band, which could not be diminished by extensive flushing with water but which could be removed by treating the Ge plate with hot chromic acid cleaning solution, is attributed to an  $\text{NO}_3^-$  species fairly strongly bound to the Ge surface. Using the absorbance of the residual peak, and assuming that the species causing the band was  $\text{NO}_3^-$ , and that the area of the Ge plate was  $20\text{ cm}^2$ , leads to an estimate of one adsorbed unit per  $5\text{-}10\text{ \AA}^2$  of surface.

The total  $\text{NO}_3^-$  absorbance at  $1352\text{ cm}^{-1}$  could be corrected by subtracting the absorbance of the residual band of the adsorbed material. When this was done, the absorbance of  $\text{NO}_3^-$  in solution was found, and a reasonably linear relation between the corrected absorbance values and  $\text{NO}_3^-$  concentration was obtained (Fig. 5). From a  $\lambda$  plot, the absorbance of the residual band (0.0111) is equivalent to the absorbance of  $\sim 0.05\text{ M NO}_3^-$  solution. Using the extinction coefficient of  $\text{NO}_3^-$  for the adsorbed species and estimating an equivalent penetration depth of about  $7\text{ \AA}$  at  $1350\text{ cm}^{-1}$  leads to an estimate that the absorbance of the adsorbed species was equivalent to the absorbance of a solution of about  $0.06\text{ M NO}_3^-$ , in fair agreement with the value obtained from the  $\text{NO}_3^-$  calibration curve.

However, the calibration plots of Fig. 5 cannot be used directly for estimating the concentrations of  $\text{NO}_3^-$ - $\text{NO}_2^-$  mixtures. The  $1237\text{-cm}^{-1}\text{ NO}_2^-$  band (asymmetric stretching) has a shoulder at  $1340\text{ cm}^{-1}$  (symmetric stretching) which, as shown clearly by Fig. 7, overlapped the  $1352\text{-cm}^{-1}\text{ NO}_3^-$  band. The latter does not cause the

absorbance of the  $1237\text{-cm}^{-1}$   $\text{NO}_2^-$  band to change significantly. The ratio of the absorbance at the  $1340\text{-cm}^{-1}$  shoulder and that of the  $1237\text{-cm}^{-1}$  band is about 4:21. Taking the ratio of the absorbances of the  $\text{NO}_2^-$  asymmetric and symmetric stretching bands to be constant then gives

$$\left[ \begin{array}{c} \text{absorbance at } 1352 \text{ cm}^{-1} \\ \text{due to } \text{NO}_3^- \end{array} \right] = \left[ \begin{array}{c} \text{total absorbance} \\ \text{at } 1352 \text{ cm}^{-1} \end{array} \right] - \left[ \frac{\text{absorbance at } 1237 \text{ cm}^{-1}}{5.25} \right]$$

Some results obtained with dilute solutions are summarized in Table 1. The correction for  $\text{NO}_3^-$  has been made. The agreement between "actual" and "found" values is not unreasonable. If the correction for the band overlapping had not been made, the  $\text{NO}_3^-$  concentration values would have been too high by about  $1/5^{\text{th}}$  of the concentration of  $\text{NO}_2^-$ . It is pertinent to note that the lowest concentrations reported earlier (4) for a binary mixture were  $[\text{NO}_2^-] = 0.7 \text{ M}$ ,  $[\text{NO}_3^-] = 0.5 \text{ M}$ , 14 and 17 times as great as the lowest values in Table 1.

There are considerable differences in the frequencies of the  $\text{NO}_3^-$  and  $\text{NO}_2^-$  absorptions which have been reported; a summary is given in Table 2. The dispersion of the index of refraction in the infrared region is known (7); i.e., in the vicinity of an absorption band, the index of refraction first decreases, then increases sharply, and decreases again to approach the normal value after the absorption has been passed. Consequently, with IRS, the penetration depth reaches a maximum when the index of refraction reaches a maximum. When the sine of the incident angle is very close to the ratio of the refractive indices, the absorption band maximum is distorted toward lower frequency.

In the present experiments, Ge and an average angle of incidence more than  $60^\circ$  were employed. The sine of the incident angle and the

refractive index ratio of the Ge and liquid are far apart (0.87 and 0.33). The effect of the dispersion of the refractive index of the aqueous solution on the absorption band positions was therefore negligible. However, Katlafsky and Keller (2) used an Irtran-2 plate with an incident angle of  $40^\circ$ . The critical angle for the system ZnS(Irtran-2)-H<sub>2</sub>O in the infrared range is  $37.5^\circ$ . Consequently, the incident angle was quite close to the critical angle, and any dispersion of the refractive index would cause a shift of absorption band maxima. Similarly, Ahliah and Mooney (4, 10) used a KRS-5 plate and stated that the incident angle was  $30^\circ$ . However, the critical angle for the KRS-5-H<sub>2</sub>O system is  $33.7^\circ$ , the refractive index of KRS-5 and H<sub>2</sub>O being taken as 2.4 (7) and 1.33, respectively. Their measurements must have been made with the incident angle close to the critical angle, so that, again, band distortions occurred. The low values of band positions reported by Katlafsky and Keller and by Ahliah and Mooney can be attributed to band distortions caused by improper IRS techniques.

It was of interest to determine how well spectral features of dissolved materials could be discerned in the vicinity of the intense water stretching band. Consequently, the incident angle was adjusted to decrease the penetration depth to about  $0.8\lambda$  at  $3400\text{ cm}^{-1}$ . The spectra of Fig. 8 were then recorded.

The water stretching band (A, Fig. 8), obtained at about 50% maximum transmittance, has the same contour as the band observed by transmission measurements by Thompson et al. (13). Two broad absorptions due to  $\text{NH}_4^+$  (8) appear on the low-wavenumber side of the OH band in the  $3100\text{--}2600\text{ cm}^{-1}$  region when  $\text{NH}_4\text{NO}_3$  is dissolved in the water (B, Fig. 8), and these become more prominent when the spectra are solvent-compensated (C,D, Fig. 8) and then scale-expanded

(Fig. 9). There is, however, some distortion of the  $3050\text{-cm}^{-1}$  band because of the negative OH band. The distortion decreases as the solute concentration is decreased, but this also decreases the solute bands. Consequently, as the liquid layer is necessarily thin as well as dilute, the signal-to-noise ratio decreases. With the present technique the lowest concentrations at which N-H and C-H bands (similar experiments were made with aqueous methanol, Na citrate, etc.) were observed were near 0.1 M.

The results thus show that it is feasible to obtain spectral information near or even at the position of one of the strong water absorptions. It must be kept in mind, however, that the presence of solutes causes the water bands to change in absorption, so that distortions result. Such distortions may also affect quantitative results by causing base lines to fluctuate. The distortions decrease in intensity with decreasing solute concentration. However, because their intensities and frequencies are a function of the nature of the solute, they cannot be precisely compensated.

#### Acknowledgment

Support by Contract No. N00014-67-A-0467-0023 from ONR is gratefully acknowledged.

## Literature Cited

- (1) Yang, R. T., Low, M. J. D., Haller, G. L., Fenn, J., J. Colloid Interfac. Sci., in press.
- (2) Katlafsky, B., Keller, R. E., Anal. Chem. 35, 1665 (1963).
- (3) Wilhite, R. N., Ellis, R. F., Appl. Spectrosc. 17, 168 (1963).
- (4) Ahliah, G. E. B. Y., Mooney, E. F., Spectrochim. Acta 25A, 619 (1969).
- (5) Ref. 4, Fig. 5.
- (6) Low, M. J. D., Goodsel, A. J., Liark, H., in "Molecular Spectroscopy 1971," P. Hepple, ed., Proc. 5<sup>th</sup> Conf. Molecular Spectroscopy, Brighton, England, 1971; Institute of Petroleum, London, 1972, pp 383ff.
- (7) Harrick, N. J., "Internal Reflection Spectroscopy," Interscience Publishers, New York, 1967.
- (8) Thompson, W. K., Trans. Faraday Soc. 61, 2635 (1965).
- (9) Anbar, M., Halman, M., Pinchas, S., J. Chem. Soc. 1960, 1242.
- (10) Ahliah, G. E. B. Y., Mooney, E. F., Spectrochim. Acta 22A, 547 (1966).
- (11) Volod'ko, L. V., Huoah, L. T., Zh. Prikl. Spektrosk. 9[4], 644 (1968).
- (12) Weston, Jr., R. E., Brodasky, T. F., J. Chem. Phys. 27, 683 (1957).
- (13) Thompson, W. K., Senior, W. A., Pethica, B. A., Nature 211, 1086 (1966).

Table 1  
Analysis of  $\text{NaNO}_3$ - $\text{NaNO}_2$  Solutions

| $\text{M NO}_2^-$<br>actual | $\text{M NO}_3^-$<br>actual | $\text{M NO}_2^-$<br>found | $\text{M NO}_3^-$<br>found <sup>a</sup> |
|-----------------------------|-----------------------------|----------------------------|-----------------------------------------|
| 0.20                        | 0.20                        | 0.20                       | 0.21                                    |
| 0.050                       | 0.050                       | 0.059                      | 0.058                                   |
| 0.050                       | 0.150                       | 0.064                      | 0.18                                    |
| 0.15                        | 0.050                       | 0.14                       | 0.06                                    |
| 0.050                       | 0.030                       | 0.044                      | 0.033                                   |

<sup>a</sup>Corrected for adsorption and band overlapping.

Table 2  
Absorptions of  $\text{NaNO}_2$  and  $\text{NaNO}_3$

| This<br>Work | IRS                            |                              | Transmission |        | Assignment                                      |
|--------------|--------------------------------|------------------------------|--------------|--------|-------------------------------------------------|
|              | Katlafsky<br>and<br>Keller (2) | Ahlijah<br>and<br>Mooney (4) | (11)         | (9,12) |                                                 |
| 1352         | 1335                           | 1333                         | 1370         | 1385   | $\text{NO}_3^-$ } asym. stretch<br>sym. stretch |
| 1047         |                                | 1045                         | 1050         | 1049   |                                                 |
| 1340         |                                | 1316                         |              | 1335   | $\text{NO}_2^-$ } asym. stretch<br>sym. stretch |
| 1237         | 1225                           | 1220                         |              | 1240   |                                                 |

## Figures

## Fig. 1. Single-Beam Spectra

A: dry Ge plate in empty cell. B: water. C: aqueous solution 0.2 M in  $\text{NaNO}_3$  and 0.2 M in  $\text{NaNO}_2$ . D: aqueous solution 0.03 M in  $\text{NaNO}_2$  and 0.05 M in  $\text{NaNO}_3$ .

## Fig. 2. Ratioed Spectra

Each spectrum shown was obtained by ratioing two of the spectra of Fig. 1. A: water (B, Fig. 1/A, Fig. 1). B: 0.2 M  $\text{NaNO}_2$  + 0.2 M  $\text{NaNO}_3$  solution (C, Fig. 1/A, Fig. 1). C: 0.2 M  $\text{NaNO}_2$  + 0.2 M  $\text{NaNO}_3$  solution (C, Fig. 1/B, Fig. 1). D: 0.03 M  $\text{NaNO}_2$  + 0.05 M  $\text{NaNO}_3$  solution (D, Fig. 1/B, Fig. 1).

## Fig. 3. Scale-Expanded Spectra

A: "100% line"; see text. B: 0.2 M  $\text{NaNO}_2$  + 0.2 M  $\text{NaNO}_3$  solution; scale-expanded segment of spectrum B of Fig. 2. C: 0.03 M  $\text{NaNO}_2$  + 0.05 M  $\text{NaNO}_3$  solution; scale-expanded segment of spectrum D of Fig. 2.

Fig. 4. Spectra of  $\text{NH}_4\text{NO}_2$  Solutions

The numbers next to each trace indicate the molar concentrations.

## Fig. 5. Calibration Curves

Fig. 6. Spectra of  $\text{NaNO}_3$  Solutions

The numbers next to each trace indicate  $[\text{NO}_3^-]$  in M. All spectra are scale-expanded. Part A shows segments taken from a series of spectra, all spectra of the series being scale-expanded by the same factor; band intensity is consequently proportional to concentration. The same segments are shown again in Part B. In order to make changes in band contour more detectable, only the absorption in the  $1550\text{-}1000\text{ cm}^{-1}$  range was expanded. For each segment, the lowest absorption within that range was placed at the bottom of the chart and the highest absorption at the top of the

chart, so that the peak heights are the same. The expansion factor is different for each segment and is supplied by the computer.


Fig. 7. Spectra of  $\text{NO}_3^-$  and  $\text{NO}_2^-$


Fig. 8. Spectra of  $\text{H}_2\text{O}$  and  $\text{NH}_4\text{NO}_3$  Solutions

A: water. B: 2 M  $\text{NH}_4\text{NO}_3$  solution. The reference used for the ratioed spectra A and B was the single-beam spectrum of the dry Ge. C: 2 M  $\text{NH}_4\text{NO}_3$  solution, the ratio of spectra B and A. D: 1 M  $\text{NH}_4\text{NO}_3$  solution, the ratio of a spectrum (of 1 M  $\text{NH}_4\text{NO}_3$  solution similar to spectrum C) and spectrum A.

Fig. 9. Spectra of  $\text{NH}_4^+$

Scale-expanded segments of spectra C and D of Fig. 8.





2

