AFRL-AFOSR-JP-TR-2019-0050



Aptamer-Based Nano-Slit Platforms for Characterizing Human Performance Biomarkers

Nathan Swami UNIVERSITY OF VIRGINIA 1001 N EMMET ST CHARLOTTESVILLE, VA 22903-4833

08/14/2019 Final Report

DISTRIBUTION A: Distribution approved for public release.

Air Force Research Laboratory Air Force Office of Scientific Research Asian Office of Aerospace Research and Development Unit 45002, APO AP 96338-5002

https://livelink.ebs.afrl.af.mil/livelink/llisapi.dll

REPORT DOCUMENTATION PAGE						Form Approved OMB No. 0704-0188		
The public reportin data sources, gat any other aspect Respondents shou if it does not displic PLEASE DO NOT R	ng burden for this co hering and maintair of this collection of JId be aware that no ay a currently valid IETURN YOUR FORM	ollection of information ning the data needed information, includin ofwithstanding any o OMB control number I TO THE ABOVE ORC	on is estimated to average d, and completing and rev g suggestions for reducing ther provision of law, no pe r. SANIZATION.	1 hour per respons iewing the collection the burden, to Dep erson shall be subje	e, including th on of informati partment of De oct to any pend	e time for reviewing instructions, searching existing on. Send comments regarding this burden estimate or fense, Executive Services, Directorate (0704-0188). alty for failing to comply with a collection of information		
1. REPORT DA	TE (DD-MM-YY)	(Y) 2. R	EPORT TYPE			3. DATES COVERED (From - To)		
4. TITLE AND S Aptamer-Base Biomarkers	S UBTITLE ed Nano-Slit Pla	tforms for Charc	icterizing Human Performance		5a.	5a. CONTRACT NUMBER		
biomarkers	DOMUKEIS					5b. GRANT NUMBER FA2386-15-1-4105		
					5c.	PROGRAM ELEMENT NUMBER 61102F		
6. AUTHOR(S) Nathan Swami					5d.	I. PROJECT NUMBER		
					5e.	TASK NUMBER		
					5f.	WORK UNIT NUMBER		
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) UNIVERSITY OF VIRGINIA 1001 N EMMET ST CHARLOTTESVILLE, VA 22903-4833 US						8. PERFORMING ORGANIZATION REPORT NUMBER		
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) AOARD						10. SPONSOR/MONITOR'S ACRONYM(S) AFRL/AFOSR IOA		
APO AP 96338	3-5002					11. SPONSOR/MONITOR'S REPORT NUMBER(S) AFRL-AFOSR-JP-TR-2019-0050		
12. DISTRIBUTI A DISTRIBUTIO	ON/AVAILABILI N UNLIMITED: PE	TY STATEMENT 3 Public Release						
13. SUPPLEME	NTARY NOTES							
14. ABSTRACT In this mannel electrophores immunoglobu on graphene- to the PSA imm	t r, using DC-offse sis in a nanoslit c ulin antibodies t modified surfac munoassay. Tot	et AC fields, diffe channel to enak hat cause signa ces, we demons al of eight publi	erences in frequency ole the selective enri I interferences to imr trate the elimination cations.	v dispersion for chment of pro munoassays. Th of false positiv	negative c state speci nrough cou res caused	lielectrophoresis are balanced against fic antigen (PSA) versus anti-mouse pling enrichment to capture by receptors by anti-mouse immunoglobulin antibodies		
15. SUBJECT TERMS Biomarkers, Microfluidics, Sensing, Aptamers, Single-Molecule								
16. SECURITY	CLASSIFICATIO	N OF:	17. LIMITATION OF	18. NUMBER OF PAGES	19a. NAN	AE OF RESPONSIBLE PERSON		
a. REPORT	D. ABSTRACT	c. THIS PAGE	ABSTRACT		WINDER, S	HEENA		
Unclassified	classified Unclassified Unclassified SAR			19b. TELEPHONE NUMBER (Include area code) +81-42-511-2008				
	•	•			I	Standard Form 298 (Rev. 8/98		

Prescribed by ANSI Std. Z39.18

Final Report for AOARD Grant 15IOA105 Contract FA2386-15-1-4105

"Aptamer-based nano-slit platforms for characterizing human performance biomarkers"

Date: Annual Report submitted Oct 2016, Final Report - August 2019

	US PI	Taiwan PI
Name	Nathan S. Swami	Chia-Fu Chou
e-mail address	nswami@virginia.edu	cfchou@phys.sinica.edu.tw
Institution	University of Virginia	Academia Sinica, Taiwan
Mailing Address	351 McCormick Road, Charlotesville,	
	VA 22904	
Phone	(434) 924 1390	
Fax	(434) 924 8818	

Period of Performance: 09/24/2015 - 05/23/2018

Abstract: The need for methods to assess for human performance biomarkers that indicate fatigue, vigilance, and stress among DoD personnel during key missions has been identified as a major priority by the AFOSR. We seek to develop aptamer-based sensing platforms within or immobilized to nanoscale structures to enrich biomarkers of interest, for downstream coupling to multiplexed detection strategies from biofluids that contain physiological levels of interfering molecules. The chief scientific advancements include the development of electric field-based biomarker enrichment platforms, nanoconfined strategies for determining aptamer binding kinetics to target biomarkers and nanoparticle strategies for surface immobilization-free detection of biomarkers. The work was performed in collaboration with AFRL's 711th Human Performance Wing.

Introduction: Monitoring the expression profile of stress-related biomarkers can enable an improved understanding of the underlying signaling pathways that cause stress and lead to the development of therapies to mitigate stress. In this project, we seek to apply electrokinetic force fields for selective enrichment of stress-related neurological biomarkers within nano-slit devices, so that they could be coupled to aptamer-based detection strategies to assess human performance. The overall aims include:

(1) Enhance detection sensitivity by electrokinetic enrichment of biomarkers within nano-slit device platforms

- (2) Enhance detection selectivity using aptamers and frequency selective AC fields
- (3) Quantify biomarker binding interactions of aptamer libraries with target biomarkers

Experiment: The enabling device used in this work is a nanoslit device with nanoscale confinement in depth.

Per Fig. 1, a nanoslit to nanoscale depth (50-200 nm) is microfabricated by etching into quartz or imprinting into COC (cyclic copolymer). olefin This device was immobilized with aptamers of interest using gold-thiol chemistry and room temperature bonding under pressure to ensure biofunctionality. In some situations, the sensing performed without was surface immobilization.



Results and Discussion: In the following sections, we describe the highlights of results in each significant publication from this work.

1. "Aptamer-Functionalized Nanoparticles for Surface Immobilization-Free Electrochemical Detection of Cortisol in a Microfluidic Device", B. J. Sanghavi; J. H Moore; J. L Chavez; J. Hagen; N. Kelley-Loughnane; C. –F. Chou, N. S. Swami*. *Biosens. Bioelectron.* (2016), 78, pp. 244-252; DOI: 10.1016/j.bios.2015.11.044. Journal Impact Factor =7.8; <u>Times Cited</u> = 80.

Monitoring the periodic diurnal variations in cortisol from small volume samples of serum or saliva is of great interest, due to the regulatory role of cortisol within various physiological functions and stress symptoms. Current detection assays are immunologically based and require cumbersome antibody immobilization chemistries, thereby limiting the assay versatility, kinetics, and reproducibility. We present a quantitative aptamer-based detection methodology for cortisol that does not require target labeling, capture probe immobilization on the detection surface or wash steps prior to readout. Using a recognition system of aptamer functionalized gold nanoparticles pre-bound with electro-active triamcinolone, the cortisol level is detected based on its competitive binding to the aptamer by following signal from the displaced triamcinolone using square wave voltammetry at patterned graphene-modified electrodes in a microfluidic or nanoslit device. Due to the 3D analyte diffusion profile at the aptamer interface and the ability to enhance the surface area for cortisol capture, this assay shows signal linearity over a five-log analyte concentration range (10 μ g/mL – 30 pg/mL) and exhibits rapid binding kinetics with cortisol versus other glucocorticoids, as apparent from the absence of interferences from estradiol, testosterone and progesterone. The assay is carried out within the biologically relevant range for glucocorticoids in serum and saliva matrices, and benchmarked versus ELISA and radioimmunoassays. Based on absence of cumbersome surface immobilization and wash steps for carrying out this assay, its quantitative signal characteristics and its ability to resist interferences from other glucocorticoids, we envision its application towards routine monitoring of cortisol within bio-fluids.



Fig. 2: (a) Top view of device on glass coverslip within an insulator layer for fluidic and electrical isolation; (b) Displacement assay for aptamer-based cortisol detection, due to signal from electrochemical reduction of displace triamcinolone (TA) on graphene-modified electrodes.

The electrochemical aptamer assay for cortisol is based on close similarity in its size and structure to that of electro-active triamcinolone (TA). While this similarity causes the aptamer to bind with triamcinolone (Fig. 1b(i)), it is displaced from the aptamer in the presence of cortisol (Fog. 1b(ii)), since the aptamer was specifically selected for preferential binding to cortisol. The displaced triamcinolone can be electrochemically reduced (Fig. 1b(iii)) for detection at graphene-modified electrodes. The chief analytical figures of merit are summarized in Fig. 3. This aptamer-functionalized nanoparticle platform has the following features:

- (i) the high capture-probe level enables a wide dynamic range $(10 \ \mu g/mL 10 \ pg/mL)$;
- (ii) the 3D diffusion profile of cortisol towards aptamers for rapid binding;
- (iii) wash-free detection allows for real-time monitoring



Fig. 3: (c) Cortisol signal at its limit of detection (LOD) at 10 pg/mL versus negative control (0.1 μ g/mL of Neuropeptide Y); (d) Dilution plot close to LOD; and (e) Dilution plot over five-log signal linearity range (30 pg/mL to 10 μ g/mL). Error bars are mean ± standard deviation.

Figure 4 shows the quantification of samples using the voltammetry assay (left) and validation of voltammetry versus ELISA and radio-labeling assays for three samples from AFRL.



Serum	Voltammetry	ELISA	Radio-
Sample			labeling
1 (Sample 17)	0.09 ± 2.3	0.08 ± 3.9	0.11
2 (Sample 20)	0.15 ± 2.1	0.13 ± 3.4	0.21
3 (Sample 54)	0.11 ± 2.5	0.09 ± 3.6	0.16

Fig. 4 (left) Sample quantification using voltammetry; (right) validation of voltammetry versus ELISA & radio-labeling assays for three samples from AFRL (±%).

2."Nanoslit design for ion conductivity gradient enhanced dielectrophoresis for ultrafast biomarker enrichment in physiological media", A. Rohani, W. Varhue, K. -T. Liao, C. –F. Chou, N. Swami*. Biomicrofluidics (2016), 10 (3) 033109:1-14. Journal Impact Factor =3.1. Times Cited = 8.

Selective and rapid enrichment of biomolecules is of great interest for biomarker discovery, protein crystallization and in biosensing for speeding assay kinetics and reducing signal interferences. The current state of the art is based on DC electrokinetics, wherein localized ion depletion at the microchannel to nanochannel interface is used to enhance electric fields and the resulting biomarker electromigration is balanced against electro-osmosis in the microchannel to cause high degrees of biomarker enrichment. However, biomarker enrichment is not selective and the levels fall off within physiological media of high conductivity, due to a reduction in ion concentration polarization and electro-osmosis effects. Herein, we present a methodology for coupling AC electrokinetics with ion concentration



enrichment of biomarkers in nanoslits

polarization effects in nanoslits under DC fields, for enabling ultrafast biomarker enrichment in physiological media. Per Fig. 5, using AC fields at the critical frequency necessary for negative dielectrophoresis of the biomarker of interest, along with a critical offset DC field to create proximal ion accumulation and depletion regions along the perm-selective region inside a nanoslit, we enhance the localized field and field gradient to enable biomarker enrichment over a wide spatial extent along the nanoslit length. While enrichment under DC electrokinetics relies solely on ion depletion to enhance fields, this AC electrokinetic mechanism utilizes ion depletion as well as ion accumulation regions to enhance the field and its gradient. Hence, biomarker enrichment continues to be substantial in spite of the steady drop in nanostructure perm-selectivity within physiological media.

3. "Ultrafast immunoassays by coupling dielectrophoretic biomarker enrichment on nanoslit molecular dam with electrochemical detection on graphene", B.J. Sanghavi+, W. Varhue, A. Rohani, K.T. Liao, L. Bazydlo, C.F. Chou, N.S. Swami*; *Lab Chip* (2015), 15, 4563 – 4570. Journal Impact Factor =6.1; Times Cited = 64.

Heterogeneous immunoassays usually require long incubation times to promote specific target binding and several wash steps to eliminate non-specific binding. Hence, signal saturation is rarely achieved at detection limit levels of analyte, leading to significant errors in analyte quantification due to extreme sensitivity of the signals to incubation time and methodology. The poor binding kinetics of immunoassays at detection limit levels can be alleviated through creating an enriched analyte plug in the vicinity of immobilized capture probes to enable signal saturation at higher levels and at earlier times, due to higher analyte association and its faster replenishment at the binding surface. Herein, we achieve this by coupling frequency-selective dielectrophoretic molecular dam enrichment of the target biomarker in physiological media to capture probes immobilized on graphene-modified surfaces in a nanoslit to enable ultrafast immunoassays with near-instantaneous (< 2 minutes) signal saturation at dilute biomarker levels (picomolar) within ultra-low sample volumes (picoliters). This methodology is applied to the detection of Prostate Specific Antigen (PSA) diluted in serum samples, followed by validation against a standard two-step immunoassay using three de-identified patient samples. Based on the ability of dielectrophoretic molecular dam analyte enrichment methods to enable the detection of PSA at 1-5 pg/mL levels within a minute, and the relative

insensitivity of the signals to incubation time after the first two minutes, we envision its application for improving the sensitivity of immunoassays and their accuracy at detection limit levels.



4. "Aptamer-functionalized graphene-gold nanocomposites for label-free detection of dielectrophoretic-enriched neuropeptide Y"; R. E. Fernandez, B. J. Sanghavi, V. Farmehini, J. L. Chavez, J. A. Hagen, N. Kelley-Loughnane, C. F. Chou, N. S. Swami*. *Electrochemistry Communications* (2016), 72, 144-147. Journal Impact Factor =4.6. Times Cited = 19.

We present an electrokinetically enhanced aptamer sensing platform on a disposable plastic chip for label-free detection of neuropeptide Y (NPY), which is a key neurological biomarker. The sensor consists of aptamer-functionalized graphene-gold nanocomposites (Gr-AuNs) patterned inside a nanoslit that is embossed on cyclic olefin copolymer via nanoimprint lithography. Analyte molecules are dielectrophoretically focused through the nanoslit onto aptamer-immobilized Gr-AuNs for rapid and selective electrochemical detection of NPY at picomolar levels. Per Fig. 8, picomolar level detection limits can be achieved under electric fields to enrich NPY, which was then confirmed for selectivity for serum samples with cortisol spiked in at μ g/mL levels.





Fig. 8: (a) Dynamic range and (b) linear detection regions for NPY; and (c) detection kinetics.

5. "Frequency-selective electrokinetic enrichment of biomolecules in physiological media"



Fig. 9: Frequency-selective dielectrophoresis is used for selective enrichment of prostate specific antigen versus anti-mouse immunoglobulin to red immunoassay signal interference

A. Rohani, B. J. Sanghavi+, A. Salahi, K. –T. Liao, C.-F. Chou, N.S. Swami*, Nanoscale (2017), 9, 12124 – 12131. Impact Factor =7.4; Cited = 15.

Proteomic biomarkers of interest to the early diagnosis of diseases and infections are present at trace levels versus interfering species. Hence, their selective enrichment is needed within bio-assays for speeding binding kinetics with receptors and for reducing signal interferences. While DC fields can separate biomolecules based on their electrokinetic mobilities, they are unable to selectively enrich biomarkers versus interfering species, which may possess like-charges. We present the utilization of AC electrokinetics to enable

frequency-selective enrichment of nanocolloidal biomolecules, based on the characteristic time constant for polarization of their electrical double-layer, since surface conduction in their ion cloud depends on colloidal size, shape and surface charge. In this manner, using DC-offset AC fields, differences in frequency dispersion for negative dielectrophoresis are balanced against electrophoresis in a nanoslit channel to enable the selective enrichment of prostate specific antigen (PSA) versus anti-mouse immunoglobulin antibodies that cause signal interferences to immunoassays. Through coupling enrichment to capture by receptors on graphene-modified surfaces, we demonstrate the elimination of false positives caused by anti-mouse immunoglobulin antibodies to the PSA immunoassay.

List of Publications and Significant Collaborations that resulted from your AOARD supported project:

- A. Rohani, B. J. Sanghavi, A. Salahi, K. –T. Liao, C.-F. Chou, N.S. Swami*, "Frequency-selective electrokinetic enrichment of biomolecules in physiological media based on electrical double-layer polarization"; *Nanoscale* (2017), 9, 12124 – 12131. Impact Factor =7.4; Cited = 22.
- R. E. Fernandez, B. J. Sanghavi, V. Farmehini, J. L. Chávez, J. Hagen, N. Kelley-Loughnane, C.-F. Chou, N. S. Swami; "Aptamer-functionalized graphene-gold nanocomposites for label-free detection of dielectrophoretic-enriched neuropeptide Y", *Electrochem. Comm.* (2016), 72, 144-147. Times Cited = 19.
- dielectrophoretic-enriched neuropeptide Y", *Electrochem. Comm.* (2016), 72, 144-147. Times Cited = 19.
 B.J. Sanghavi, W. Varhue, A. Rohani, K.T. Liao, L. Bazydlo, C.F. Chou, N.S. Swami*; "Ultrafast immunoassays by coupling dielectrophoretic biomarker enrichment on nanoslit molecular dam with electrochemical detection on graphene", Lab Chip (2015), 15, 4563 4570. Times Cited = 64.
- B. J. Sanghavi; J. A Moore; J. L Chavez; J. Hagen; N. Kelley-Loughnane; C. –F. Chou, N. S. Swami*. ""Aptamer-Functionalized Nanoparticles for Surface Immobilization-Free Electrochemical Detection of Cortisol in a Microfluidic Device", *Biosens. Bioelectron.* (2016), 78, pp. 244-252; Times Cited = 80
- 5. A. Rohani, W. Varhue, K.-T. Liao, C.-F. Chou, N.S. Swami; "Nanoslit design for ion conductivity gradient enhanced dielectrophoresis; *Biomicrofluidics* (2016) 10, 033109; Times Cited = 8.
- 6. P Teerapanich, M Pugnière, C Henriquet, YL Lin, CF Chou, T Leïchlé, Nanofluidic Fluorescence Microscopy (NFM) for real-time monitoring of protein binding kinetics and affinity studies, *Biosensors & Bioelectronics*, 2017, 88, 25–33. Times Cited = 6.
- 7. Pattamon Teerapanich, Martine Pugniere, Corinne Henriquet, Yii-Lih Lin, Antoine Naillon, Pierre Joseph, Chia-Fu Chou, Thierry Leichle, Nanofluidic fluorescence microscopy with integrated concentration gradient generation for one-shot parallel kinetic assays, *Sensors and Actuators B: Chemical* (2018), 274, 338-342.
- 8. KK Sriram, S Nayak, S Pengel, CF Chou, A Erbe, 10 nm deep, sub-nanoliter fluidic nanochannels on germanium for attenuated total reflection infrared (ATR-IR) spectroscopy, *Analyst* (2017) 142 (2), 273-278

Collaborative interactions with industry or with Air Force Research Laboratory scientists:

- Joint papers with the following scientists from AFRL's 711-th Human Performance Wing: Jorge Chavez, Nancy Kelley-Loughnane, and Joshua Hagen

- Collaboration with UES, Inc. on multi-target sensor development

Attachments: Publications a), b) and c) listed above if possible.

DD882: As a separate document, please complete and sign the inventions disclosure form.