

AD-A056 264

MATERIALS RESEARCH LABS ASCOT VALE (AUSTRALIA)
A CRITICAL EVALUATION OF KOCOUR ELECTRONIC THICKNESS TESTERS - --ETC(U)
JAN 78 B L MOURANT, L H ESMORE

F/G 14/2


UNCLASSIFIED

MRL-TN-409

NL


| OF |

AD
A056264



END
DATE
FILMED
8-78

DDC



MICROCOPY RESOLUTION TEST CHART
NATIONAL BUREAU OF STANDARDS-1963-A

LEVEL

(12) 94 AR-000-881



DEPARTMENT OF DEFENCE
DEFENCE SCIENCE AND TECHNOLOGY ORGANISATION
MATERIALS RESEARCH LABORATORIES

MELBOURNE, VICTORIA

TECHNICAL NOTE

MRL-TN-409

A CRITICAL EVALUATION OF KOCOUR ELECTRONIC
THICKNESS TESTERS - MODELS 955 AND S77

B.L. Mourant and L.H. Esmore

Approved for Public Release



© COMMONWEALTH OF AUSTRALIA. 1978

JANUARY, 1978

78 07 17 034

AD A 056264

AD No. ~~1~~
DDC FILE COPY

APPROVED
FOR PUBLIC RELEASE

THE UNITED STATES NATIONAL
TECHNICAL INFORMATION SERVICE
IS AUTHORIZED TO
REPRODUCE AND SELL THIS REPORT

LEVEL II

DEPARTMENT OF DEFENCE
MATERIALS RESEARCH LABORATORIES

(12)

(9) TECHNICAL NOTE

(14) MRL-TN-409

(11) Jan 76

(12) 17p

(6)

A CRITICAL EVALUATION OF KOCOUR ELECTRONIC
THICKNESS TESTERS - MODELS 955 AND S77.

(10) B.L. Mourant and L.H. Esmore

ABSTRACT

The factors affecting the accuracy of two models of the Kocour Electronic Thickness Tester are evaluated including physical measurements of cells and gaskets, determination of the electrical characteristics and the evaluation of the standard electroplated discs. Constants determined allow the instruments to be used for coulometric methods described in ISO, SAA and BS specifications.

Approved for Public Release



© COMMONWEALTH OF AUSTRALIA . 1978

POSTAL ADDRESS: Chief Superintendent, Materials Research Laboratories
P.O. Box 50, Ascot Vale, Victoria 3032, Australia

409 014 78 07 13 034 503

DOCUMENT CONTROL DATA SHEET

Security classification of this page:

UNCLASSIFIED

1. DOCUMENT NUMBERS:

a. AR Number: AR-000-881

b. Series & Number: TECHNICAL NOTE

c. Report Number: MRL-TN-409 ✓

2. SECURITY CLASSIFICATION:

a. Complete document: UNCLASSIFIED

b. Title in isolation: UNCLASSIFIED

c. Abstract in isolation: UNCLASSIFIED

3. TITLE:

A CRITICAL EVALUATION OF KOCOUR ELECTRONIC
THICKNESS TESTERS - MODELS 955 AND S77 ✓

4. PERSONAL AUTHOR(S):

MOURANT, B.L. and
ESMORE, L.H.

5. DOCUMENT DATE:

JANUARY, 1978 ✓

6. TYPE OF REPORT & PERIOD COVERED:

7. CORPORATE AUTHOR(S):

Materials Research Laboratories

8. REFERENCE NUMBERS:

a. Task: DST 98/015

b. Sponsoring Agency:

9. COST CODE: 571600

10. IMPRINT (Publishing establishment)

Materials Research Laboratories,
P.O. Box 50, Ascot Vale, Vic.3032
JANUARY, 1978

11. COMPUTER PROGRAMME(S):

(Title(s) and language(s)):

12. RELEASE LIMITATIONS (of the document):

Approved for Public Release

12-0. OVERSEAS:

N.O. ☐P.R. ☒A ☐B ☐C ☐D ☐E ☐

13. ANNOUNCEMENT LIMITATIONS (of the information on this page):

No Limitation

14. DESCRIPTORS:

Coulometers; Electrodeposited Coatings; Metal Coatings;
Evaluation; Accuracy; Thickness Gages.

15. COSATI CODES: 1402, 1103

16. ABSTRACT (if this is security classified, the announcement of this report will be similarly classified):

✓ The factors affecting the accuracy of two models of the Kocour Electronic Thickness Tester are evaluated including physical measurements of cells and gaskets, determination of the electrical characteristics and the evaluation of the standard electroplated discs. Constants determined allow the instruments to be used for coulometric methods described in ISO, SAA and BS specifications.

- j -

C O N T E N T S

	<u>Page No.</u>
INTRODUCTION	1
EXPERIMENTAL	2
CONCLUSIONS	5
REFERENCE	5
TABLE 1: TYPICAL CELL CURRENTS	6
TABLE 2: REPRODUCIBILITY OF STRIPPED AREA - TYPE "A" CELL	7
TABLE 3: REPRODUCIBILITY OF STRIPPED AREA - TYPE "A" CELLS	8
TABLE 4: STANDARD THICKNESS DISCS	9
APPENDIX: CALCULATIONS OF SPOT DIAMETER FOR "A" CELL	10

* * *

ACCESSION for	
NTIS	White Section <input checked="" type="checkbox"/>
DDC	Off Section <input type="checkbox"/>
UNANNOUNCED	<input type="checkbox"/>
JUSTIFICATION	
BY	
DISTRIBUTION/AVAILABILITY NOTES	
Dist.	
A	

A CRITICAL EVALUATION OF KOCOUR ELECTRONIC
THICKNESS TESTERS - MODELS 955 AND S77

INTRODUCTION

The methods of measurement of the thickness of electroplated coatings may be classified into three groups :

1. Non-destructive
2. Destructive to both the coating and the component itself
3. Destructive to the coating only.

The methods in the first group are essential for many critical applications because every component in a batch may be examined without damage. These techniques include magnetic attraction, electromagnetic induction, eddy current, beta-ray back-scatter and X-ray fluorescence instruments. The accuracy of all of these may be affected by variations in coating purity or density and by underlayers of metals which differ from the coating being measured.

The second group includes sectioning of the components and, after suitable preparation, measurement of the coating thickness with a calibrated microscope. This method is widely adopted as the standard method in specifications but is limited to a minimum thickness of 2 μm and an absolute accuracy of $\pm 0.8 \mu\text{m}$ (1).

Methods in the third group employ dissolution of the coating either chemically or electrochemically without attacking the basis metal. The methods are commercially attractive because valuable items can be recovered by stripping and re-plating, the tests can be carried out in a works environment, and the accuracy is more than adequate for control purposes. Coulometric methods fall into this group and a number of suitable instruments have been designed and marketed. The Kocour* instruments were early in the

* Manufactured by the Kocour Co., 4800 S. St. Louis Ave., Chicago 32, Illinois, U.S.A.

field and have been widely employed throughout the industry. These instruments dissolve the coating electrochemically from a small area at constant current density, using electrolytes which do not attack the coating until current is applied. The area to be stripped is defined by a rubber gasket on a small Monel-metal cell. The d.c. current is controlled by a constant voltage source and precision series resistors. Time is recorded by a simple counting device driven by a synchronous clock and, by selecting the d.c. currents for particular coatings, the readings can be made directly in thickness units ($2s = 0.00001$ in). When the coating is perforated, the resultant rapid change in cell voltage triggers a relay system which terminates the test.

The Model 955 instrument is an early unit employing thermionic valves whereas the latest Model S77 (modified) uses solid state devices. The cells and gaskets for both models are the same and are used with accessory units as follows :

- (a) Type "A" cells for normal operation.
- (b) Accessory units "B" and "CB" with Type "B" gaskets and cells for smaller areas.
- (c) Accessory units "MT" and "CM" with Type "A" cells and gaskets for coatings thinner than about 0.0001 in ($2.5 \mu m$).

The following examination of the two instruments was aimed at assessing the factors affecting the reproducibility and accuracy of the thickness measurements. These are :

- 1. The stability of the d.c. supplies and cell currents.
- 2. The consistency of dimensions of cells and gaskets.
- 3. The consistency of the areas stripped.
- 4. The reliability of the electroplated standard discs supplied by the manufacturer for calibration checks.

The Model 955 instrument is designed for 110V power supplies and was supplied with a 230V/110V stepdown transformer. The Model S77 instrument has optional transformer taps for either 110 or 220V operation, and was set to the latter.

The composition of the electrolytes supplied by the manufacturer are not disclosed but are similar to some described in specifications.

EXPERIMENTAL

Constant Current Supply

The voltage sources of both instruments are stabilised. These were supplied from a "Stabilac" Model SP2500 unit through a "Variac" which allowed variation of the supply within the range of 200-240 V. In each instance, the stabilised d.c. voltage from each instrument was measured with a high-quality digital voltmeter. The instruments were first balanced and

calibrated in accordance with the manufacturer's instructions at 220 V mains setting and the supply then varied in steps of 10 V.

The instruments have provision for varying the calibration by $\pm 10\%$ from the normal settings and the accuracy of this offset was similarly examined holding the supply constant at 220 V.

The cell currents for each setting of the coating selector switch including those on the accessory units Type "B", "CB" and "CM" were also measured again with a precision digital meter. (An "MT" unit was not available for the Model 955 instrument).

Uniformity of Cells and Gaskets

The important cell dimension is the external diameter over which the gasket is fitted. If this is too small, electrolyte leakage could occur and, if too large, the gaskets could be unduly stretched. These dimensions on two type "A" and type "B" cells were measured at two positions at right angles.

Two dimensions on a number of new gaskets, shown in Fig. 1 as "P" and "Q", were measured by means of a travelling microscope on two diameters at right angles for both type "A" and type "B" gaskets.

In use, the gaskets are slightly stretched when fitted to the cell and they are compressed under normal cell loading. The area stripped is affected by these factors and, in addition, the gaskets may be distorted by relative movements of the sample and the cell during set-up of the tests.

The combined effects of these factors in actual measurements of coating thickness were examined by stripping the coating from tin plate. The coating on this was thin enough to allow numerous tests to be conducted rapidly and the contrast between the tin and the stripped steel aided in the precise measurement of the spots with a travelling microscope. The reproducibility of spot sizes, using a single "A" gasket on a particular cell, was determined to check the effect of re-positioning the cell ten times. For each spot, two diameters were measured at right angles to check possible ovality of the spots which may arise due to the off-centre cell mounting. These measurements were repeated with a "B" gasket and cell.

The consistency of eleven gaskets used with two cells was similarly determined for a single spot with each combination to check the variation in cells and gaskets as supplied. These measurements were also repeated with "B" gaskets.

Kocour Thickness Standards

The manufacturer supplies circular discs of various basis metals which have been electroplated with the coatings of copper, nickel, zinc, etc. It should be noted that these standards have been specifically prepared for coulometric measurement and should not be used for magnetic and other instruments because undercoats are sometimes plated under the standard coating.

These discs are plated in cells designed to achieve uniformity of thickness and with accurately controlled current and time to obtain specified thicknesses. The coatings are specified to be within $\pm 5\%$ of a marked thickness, and the reliability of this claim is of paramount importance if they are used to calibrate the instruments, as recommended by the manufacturer. It is also of some importance to establish whether the calibration of the instrument should be adjusted so that it reads the value marked on the standard or whether the instrument should be set up with zero correction and checked against the standard to ensure that the readings are within 5% of the labelled figure.

The thicknesses of the coatings on five standards were determined by cutting a sector from each disc, overplating this with copper, mounting, polishing and etching to permit measurement with a calibrated metallurgical microscope.

RESULTS AND DISCUSSION

Constant Current Supply

The effects of varying the a.c. mains supply ± 20 V from the nominal 220 V are negligible, the stabilised open-circuit voltage across the cell being 296 ± 1 V for the Model 955 instrument and 85.9 ± 0.3 V for the Model S77. When the calibration controls were set 10% high or low, the d.c. voltage agreed with the calculated figures for the Model 955 instrument, and were better than $\pm 1\%$ for the S77.

The actual cell currents at each setting of the coating selector switches for each instrument and accessory unit are recorded in Table 1.

Cells and Gaskets

In the unstressed condition, the dimension "P" for twelve "A" gaskets averaged 8.26 ± 0.07 mm compared a mean cell neck dimension of 8.35 mm. For eleven "B" gaskets, dimension "P" averaged 6.31 ± 0.04 mm compared with the cell dimension of 6.45 mm. No electrolyte leakage occurred when the largest gaskets were tested with any cell.

The effects of cell positioning and pressures on a single "A" gasket on one cell applied ten times to the tin-plate surface gave spot diameters shown in Table 2. Calculations of the areas of the spots as circles (using mean diameters) or as ellipses show that this effect is not significant. The spot diameters for eleven gaskets each fitted to two different cells with two spots produced with each combination are shown in Table 3. Similar measurements were made on a set of "B" gaskets. For ten applications of a single gasket on one cell, the mean spot diameter was 2.49 mm (s.d. 0.07 mm) and for eleven gaskets on two cells with each combination tested twice, the means and standard deviations were 2.50 mm (s.d. 0.10 mm) and 2.51 mm (s.d. 0.09 mm).

A calculation of the theoretical area of a spot, based on a zinc coating, shows that the diameter of the spot with an "A" cell should be 3.45 mm

and for a "B" cell, 2.44 mm. The calculations are in the Appendix. In the two tables, the largest spot was 3.66 mm, which corresponds to an area 11.3% greater than the calculated figure, and the smallest spot was 3.33 mm, which is 6.7% below the theoretical value. The mean of all 44 readings in Table 3 is 3.45 mm, which coincides with the calculated value.

Thickness Standard Discs

The microscope measurements of the thicknesses of the coatings on five discs are recorded in Table 4, together with mean values and a calculation of errors. Accepting the accuracy of the microscope to be $\pm 0.8 \mu\text{m}$ for a single reading, it is important to note that some readings show greater errors than the claimed figure of $\pm 5\%$, even when the mean is within the limit. It is therefore, not recommended that the instrument be re-calibrated if a test on a disc shows an error in respect to the marked value. If the reading is within 5% of the value, the instrument should be taken to be correctly calibrated. If the reading is not within the 5% value, the gasket, which may be damaged, should be replaced and a further test conducted.

A final observation is that the standard coatings may be damaged by cleaning with a pencil eraser as recommended by the manufacturer. It is preferable to remove passive films with a small swab of cotton wool mounted on a match stick and charged with freshly prepared paste of magnesia. The swab should be applied lightly through the gasket after the cell has been positioned. It is thus only applied to the test area thereby avoiding damage to adjacent areas of the standard coating.

CONCLUSIONS

Both Kocour instruments have been shown to be reliable for the coulometric determination of thickness of electroplated coatings. An accuracy of better than $\pm 10\%$ should be achieved provided that the rubber cell gaskets and electrolytes are in good condition. Some recommendations are made in respect to the standard-thickness discs and their use.

REFERENCE

1. Australian Standard, ASK 173-1971, para. 2.2.1.1.

T A B L E 1

TYPICAL CELL CURRENTS

Coating Setting	Model S77, S/N 5085			Model 955, S/N 1626	
	Normal mA	Accessory "CB", mA	Accessory "CM", mA	Normal mA	Accessory "B", mA
Cr ÷ 2 = 10 ⁻⁵	46.8	23.1	-	45.4	22.95
Ni	35.5	17.65	3.49	35.0	17.60
Cu	32.7	16.25	3.23	32.0	16.17
Brass	30.5	15.15	3.01	29.7	14.95
Zn	25.32	12.74	2.52	24.9	12.44
Cd	17.77	8.89	1.77	17.45	8.74
Sn	14.18	7.08	1.407	13.87	6.92
Ag	11.35	5.67	1.127	11.12	5.55
Cr x 10 ⁻⁶	9.29	4.63	0.927	9.07	4.55
Au	2.23	1.14	-	-	-

T A B L E 2

REPRODUCIBILITY OF STRIPPED AREA - TYPE "A" CELL

(One gasket/One cell)

Spot No.	Diameters of Spots, mm		Areas of Spots, mm ²	
	a	b	$\frac{\pi}{4} \left(\frac{a+b}{2} \right)^2$	$\frac{\pi}{4} (a \times b)$
1	3.62	3.57	10.15	10.15
2	3.64	3.60	10.29	10.29
3	3.56	3.60	10.07	10.07
4	3.62	3.60	10.24	10.24
5	3.54	3.64	10.12	10.12
6	3.62	3.53	10.04	10.04
7	3.66	3.56	10.24	10.23
8	3.66	3.57	10.26	10.26
9	3.49	3.55	9.73	9.73
10	3.51	3.62	9.98	9.98
Mean, 20 readings		3.588	10.11	10.11
Standard Deviation		0.049	(Circle)	(Ellipse)

T A B L E 3

REPRODUCIBILITY OF STRIPPED AREA - TYPE "A" CELLS

(11 gaskets/Two cells)

Gasket No.	Diameters of Spots, mm			
	Cell No. 1		Cell No. 2	
1	3.58	3.52	3.46	3.38
2	3.48	3.41	3.44	3.40
3	3.47	3.50	3.32	3.33
4	3.54	3.53	3.43	3.44
5	3.55	3.51	3.41	3.47
6	3.47	3.49	3.48	3.48
7	3.37	3.33	3.34	3.35
8	3.45	3.53	3.52	3.49
9	3.52	3.46	3.71	3.48
10	3.46	3.46	3.36	3.39
11	3.55	3.43	3.37	3.38
Mean, 22 spots = 3.48 mm			Mean, 22 spots = 3.43 mm	
Standard Deviation = 0.061 mm			Standard Deviation = 0.085 mm	

T A B L E 4

STANDARD THICKNESS DISCS

Coating	Zinc	Zinc	Cadmium	Nickel	Copper
Basis Metal	Steel	Steel	Brass	Steel	Steel
Marked Thickness, μm	11.2	11.7	12.2	12.5	12.7
Microscope Thickness, μm	11.8	12.1	10.7	12.2	12.8
	11.8	12.2	11.0	13.6	13.3
	11.8	13.0	10.7	13.0	12.8
	11.5	12.5	11.9	13.4	11.9
	10.5	12.1	11.3	12.2	11.9
	11.6	12.4	10.7	12.4	13.9
	10.4	12.7	12.2	13.0	12.8
	10.4	11.9	12.2	13.1	12.5
	10.7	12.7	12.4	12.2	12.5
	10.8	12.1	12.8	13.0	12.5
Mean Thickness, μm	11.1	12.4	11.6	12.8	12.7
Standard Deviation, μm	0.6	0.35	0.8	0.5	0.6
% Error, Mean v Nominal	-0.9	+6.0	+4.9	+3.2	0
% Error, Maximum v Nominal	+5.4	+11.1	+4.9	+8.8	+9.4
% Error, Minimum v Nominal	-7.1	-1.7	-12.2	-2.4	-6.3

APPENDIX

CALCULATION OF SPOT DIAMETER FOR "A" CELL

If we take the following constants,

Atomic Weight of Zinc	=	65.37
One Faraday	=	96,500 coulombs
Density of Zinc	=	7.14 grams/cm ³
Anodic Current Effic.	=	100%

and apply these to the following instrument and coating characteristics,

Zinc Coating Thickness	=	0.00100 inch = 25.40 micrometre
Deplating Current	=	25 mA (Model S77 25.32 mA, Model 955 24.9 mA)
Stripping Time	=	200 s

the following calculations may be made

$$200 \text{ s} \times 25 \text{ mA} = 5000 \text{ millicoulombs}$$

$$\begin{aligned} \text{Mass of Zinc} &= \frac{\text{millicoulombs}}{F} \times \frac{A.W.Zn}{z} \\ &= \frac{5000}{96,500} \times \frac{65.37}{2} \text{ milligram Zn} \\ &= 1.694 \text{ mg Zn} \end{aligned}$$

$$\begin{aligned} \text{Volume of Zinc} &= \frac{\text{Mass}}{\text{Density}} = \frac{1.694 \times 10^{-3}}{7.14} \\ &= 0.2373 \times 10^{-3} \text{ cm}^3 \end{aligned}$$

$$\begin{aligned} \text{Equivalent Area for 25.4 } \mu\text{m} &= \frac{0.2373 \times 10^{-3}}{25.4 \times 10^{-6}} = 9.3425 \text{ mm}^2 \\ \text{thickness} & \end{aligned}$$

$$\text{Diameter of Spot} = \sqrt{\frac{9.3425}{3.1416} \times 4} = 3.449 \text{ mm}$$

This figure should be compared with the diameters measured and shown in Tables 2 and 3.

For the Type "B" gasket, the area of the spot should be half that for the Type "A".

$$\text{Area} = 4.6712 \text{ mm}^2$$

$$\text{Diameter of Spot} = 2.439 \text{ mm}$$

This figure should be compared with the diameters quoted on Page 4.

Taking the average "A" gasket spot size to be 3.45 mm, the current required for each metal coating can similarly be calculated. A comparison between these values and the measured values for each instrument is as follows :

Coating	Valency Change	Density (g/cm ³)	Calculated Current, mA	Measured Currents, mA		Difference, %	
				955	S77	955	S77
Cr	3	7.14	47.1	45.4	46.8	-3.6	-0.6
Ag	1	10.50	11.15	11.12	11.35	-0.3	+1.8
Sn	2	7.29	14.07	13.87	14.18	-0.1	+0.8
Cd	2	8.64	17.60	17.45	17.77	-0.9	+1.0
Zn	2	7.14	25.00	24.9	25.32	-0.4	+1.3
Cu	2	8.93	32.17	32.0	32.7	-0.2	+1.7
Ni	2	8.90	34.72	35.0	35.5	+0.8	+2.2

It should be realised that the density figures for electrodeposits may vary from the values for wrought metals used in the above calculations.




Fig 1 TYPE 'A' GASKET

DISTRIBUTION LIST

MATERIALS RESEARCH LABORATORIES

Chief Superintendent
Superintendent, Metallurgy Division
Dr. J.J. Batten
Mr. B.L. Mourant
Mr. L.H. Esmore
Library
Librarian, N.S.W. Branch (Through Officer-in-Charge)
Officer-in-Charge, Joint Tropical Trials and Research
Establishment

DEPARTMENT OF DEFENCE

Chief Defence Scientist
Executive Controller, ADSS
Superintendent, Defence Science Administration, DSTO
Superintendent, Military Advisers Branch
Head, Laboratory Programs Branch
Army Scientific Adviser
Air Force Scientific Adviser
Naval Scientific Adviser
Chief Superintendent, Aeronautical Research Laboratories
Director, Weapons Research Laboratories
Senior Librarian, Weapons Research Laboratories
Librarian, R.A.N. Research Laboratory
Document Exchange Centre (16 copies)
Principal Librarian, Campbell Park Library ADSATIS Annex
Central Office, Directorate of Quality Assurance - Air Force
Director, Joint Intelligence Organisation
Head, Engineering Development Establishment
Mr. P. Ramsden, No.3 Army Quality Assurance Unit, Ordnance
Reserve, Maribyrnong
Mr. B.H.S. Day, Weapons Research Establishment
Mr. B. Wearne, Naval Dockyard, Garden Island Dockyard Laboratory
Mr. D. Glanvill, Aeronautical Research Laboratories

DEPARTMENT OF PRODUCTIVITY

NASA Canberra Office
Head, B.D.R.S.S. (Aust.)
Manager, Ammunition Factory, Footscray
Manager, Small Arms Factory, Lithgow
Manager, Government Aircraft Factory
Mr. M. Finnegan, Ammunition Factory, Footscray
Mr. O. Broughton, Government Aircraft Factory
Mr. A. Cottrell, Ordnance Factory, Maribyrnong
Mr. G. Sutherland, Ordnance Factory, Bendigo
Mr. P. Keown, Ordnance Factory, Bendigo
Mr. K. Coates, Small Arms Factory, Lithgow
Mr. D. Colbourne, Small Arms Factory, Lithgow
Mr. J. Porebski, Small Arms Factory, Lithgow

DISTRIBUTION LIST

(Continued)

OTHER FEDERAL AND STATE DEPARTMENTS AND INSTRUMENTALITIES

The Chief Librarian, Central Library, C.S.I.R.O.
Australian Atomic Energy Commission Research Establishment

MISCELLANEOUS - AUSTRALIA

Deputy Registrar, National Association of Testing Authorities
Assistant Technical Director, Standards Association of Australia
General Manager, Commonwealth Aircraft Corporation
Mr. J. Barbaro, Secretary Committee MT/9, Standards Association
of Australia
Mr. R.D. Slade, A.P.C. Research Laboratories
Mr. M.S. Pennisi, Department of Mining & Metallurgical
Engineering, University of Queensland
Mr. D. Smith, Electromasurements Pty.Ltd.
Mr. J. Isgar, H.B. Selby & Co. Pty.Ltd.
Mr. R. Johnson, M & T Products Pty.Ltd.
Mr. L. Ponsford, Avin Plating Pty.Ltd.
Mr. D. Stroud, Dept. of Metallurgy, Mining & Geology, Royal
Melbourne Institute of Technology (3 copies)
Mr. C.M. Whittington, International Nickel (Aust.) Pty.Ltd.
Mr. B.D. Reaburn, Nilsens Industries Pty.Ltd.
Mr. M. Taylor, Amalgamated Wireless (Aust.)
Mr. Clyde Naylor, Ford Motor Co. of Aust.
Mr. I. Rose, Robert Bryce & Co.
Mr. F. Budge, Ajax-Nettlefolds Ltd.
Mr. K. Stanton, National Association of Testing Authorities
(2 copies)
Mr. J. Huggard, T.A.A. Overhaul Section, Essendon Airport
Mr. W. Jenkins, Technisearch Ltd.
Mr. K.L.M. Staggard, Wilmot-Breeden (Aust.) Pty.Ltd.
Mr. J. McAuliffe, Australian Zinc Development Association
Mr. B.A. Footner, Chrysler (Aust.) Ltd.
Mr. R.F. Darnell, C/o Alcan Aust. Ltd.
Mr. I.L. Evans, C/o Lawrence Smith & Canning Pty.Ltd.
Mr. A. Goninan, C/o Goninans Industrial Platers Pty.Ltd.
Mr. B.A. Green, 4/7 Battery St., Clovelly, N.S.W.
Mr. S.G. Anderson, 3 Station Rd., North Williamstown
Mr. M. Barter, Electrolytic Zinc Co. of Aust.
Mr. F.J. Byrt, C/o McPhersons Research & Development Pty.Ltd.
Mr. K.L. Clark, C/o Applied Chemicals Pty.Ltd.
Mr. A.T. Corcoran, 20 Elevers St., Watsonia
Mr. H.P. Cotter, Ajax Nettlefolds Pty.Ltd.
Mr. D.D. Forbes, C/o Ogden Industries Pty.Ltd.
Mr. E.J. Layton, Ford Motor Co. of Aust.
Mr. G.K. Mills, C/o Gibson Chemicals Pty.Ltd.

(MRL-TN-409)

DISTRIBUTION LIST

(Continued)

MISCELLANEOUS - OVERSEAS

Defence Scientific and Technical Representative, Department
of Defence, England
Assistant Director/Armour and Materials, Military Vehicles and
Engineering Establishment, England
Reports Centre, Directorate of Materials Aviation, England
Library - Exchange Desk, National Bureau of Standards, U.S.A.
U.S. Army Standardization Group, Office of the Scientific
Standardization Representative, Canberra, A.C.T.
Senior Standardization Representative, U.S. Army Standardization
Group, Canberra, A.C.T.
Chief, Research and Development, Defence Scientific Information
Service, Department of National Defence, Canada (2 copies)
The Director, Defence Scientific Information and Documentation
Centre, India
Colonel B.C. Joshi, Military, Naval and Air Adviser, High
Commission of India, Red Hill, A.C.T.
Director, Defence Research Centre, Malaysia
Accessions Department, British Library, England
Official Publications Section, British Library, England
Librarian, Periodicals Recording Section, National Reference
Library of Science and Invention, England
INSPEC: Acquisition Section, Institution of Electrical Engineers,
England
Overseas Reports Section, Defence Research Information Centre,
England
Science Information Division, Department of Scientific and
Industrial Research, New Zealand
Librarian, Auckland Industrial Development Division,
Department of Scientific and Industrial Research, New Zealand
Mr. S.W. Armstrong, Armstrong & Farr Ltd., Christchurch,
New Zealand
Mr. D.J. Mitchell, Christchurch Electroplaters Co., Christchurch,
New Zealand
Mr. K.V. Lay, Universal Electroplaters Ltd., Auckland, New
Zealand.

IED
78