

MRL-TN-413

DC FILE COPY.



AR-000-900

DEPARTMENT OF DEFENCE

DEFENCE SCIENCE AND TECHNOLOGY ORGANISATION MATERIALS RESEARCH LABORATORIES

MELBOURNE, VICTORIA

TECHNICAL NOTE

MRL-TN-413

THE EFFECT OF MERCURY-VAPOUR PRESSURE IN A MERCURY MANOMETER

David B. Prowse* and David J. Hatt

Approved for Public Release

* Previously at Materials Research Laboratories. Now at the National Measurement Laboratory.



JULY, 1978

COMMONWEALTH OF AUSTRALIA 1978.

79 05 08 020

APPROVED

FOR PUBLIC RELEASE

THE UNITED STATES NATIONAL TECHNICAL INFORMATION SERVICE IS AUTHORISED TO ' REPRODUCE AND SELL THIS REPORT

Sector Second

MCEBIOD IN -.... MANO DEPARTME STIFICATION MATERIALS RESEARCH LABORATORIES 27 DISTON A/ATAILADILITY COD ATAL ON/ STEMAL TECHNICAL NOTE MRL-TN-413 THE EFFECT OF MERCURY YAPOUR PRESSURE IN A MERCURY MANOMETER 10 David B. David J. Prowse* Hatt ABSTRACT The effect of the pressure of mercury vapour acting on the free surfaces of the two arms of a manometer is discussed, particularly in relation to the MRL interferometric manometer. Approved for Public Release * Previously at Materials Research Laboratories. Now at the National Measurement Laboratory. \mathbf{C} 1979 3 C COMMONWEALTH OF AUSTRALIA 1978. POSTAL ADDRESS: Chief Superintendent, Meterials Research Laboratories P.O. Box 50, Accot Vale, Vistoria 3032, Australia 409 914

DOCUMENT CONTROL DATA SHEET

Security classification of this page:			UNCLASSIFIED						
1000	1. DOCUMENT NUMBERS:			2.	SECURITY CLASSIFICATION:				
		e. AR Number:	AR-000-900			Complete document:	UNCLASSIFIED		
		b. Series & Number:	TECHNICAL NOTE		b.	Title in isolation:	UNCLASSIFIED		
		c. Report Number:	MRL-TN-413		c.	Abstract in isolation:	UNCLASSIFIED		

3. TITLE:

THE EFFECT OF MERCURY-VAPOUR PRESSURE IN A MERCURY MANOMETER

4.	PERSONAL AUTHOR(S):	5.	DOCUMENT DATE: JULY, 1978		
	PROWSE, David B. and HATT, David J.	6.	TYPE OF REPORT & PERIOD COVERED:		
7.	CORPORATE AUTHOR(S): Materials Research Laboratories		REFERENCE NUMBERS: a. Tesk: DST 97/017 b. Sponsoring Agency:		
		· 9.	COST CODE: 451840		
10.	IMPRINT (Publishing establishment)	11.	COMPUTER PROGRAMME(S):		
	Materials Research Laboratories, P.O. Box 50, Ascot Vale, Vic.3039 JULY, 1978		(/fue(s) and language(s)):		
12.	RELEASE LIMITATIONS (of the document):				

Approved for Public Release

12-0. OVERSEAS:	N.O.	P.R. 1		8	c	0	E	
13. ANNOUNCEMEN	LIMITATIONS (of the information	n on this pa	ge):				2000

No Limitation

14. DESCRIPTORS:

Manometers; Vapor Pressure; Mercury (metal).

15. COSATI CODES: 1402

16. ABSTRACT (If this is security classified, the ennouncement of this report will be similarly classified):

The effect of the pressure of mercury vapour acting on the free surfaces of the two arms of a manometer is discussed, particularly in relation to the MRL interferometric manometer.

DEFENCE RELEVANCE STATEMENT

The measurement of pressure is of defence interest both for altimetry and for monitoring fluid pressures, as well as for scientific studies. The corrections to a mercury manometer described in this paper are relevant to measurements of the highest accuracy.

.

use opened the ion pump current indicated a pressure of 13 pPa chroughout the experiment.

THE EFFECT OF MERCURY-VAPOUR PRESSURE IN A MERCURY MANOMETER

Over the past few years there has been considerable improvement in both the accuracy and resolution of mercury manometers (see for example the list of references in [1]). Most of these manometers have accuracies approaching 1 part in 10^6 while the capacitance manometer described by Guildner et al. [2] has a resolution of 0.1 mPa and our interferometric manometer [1] has a resolution of 10 mPa [3].

The saturation vapour pressure of mercury (160 mPa at 20°C) is approximately 2 parts in 10^6 of atmospheric pressure (101 kPa) which is significant in terms of the accuracies claimed and especially at lower pressures where the uncertainty of temperature effects is less. Some authors (Guildner et al. [2], Kaneda et al. [4], and Bass and Green [5]) take into account the vapour pressure of mercury in determining the pressure measured by their manometers whereas other authors do not mention this effect. It was therefore important to decide if it was necessary to allow for the vapour pressure of mercury in absolute pressure measurement with our interferometric manometer.

In normal use both arms of the manometer are evacuated with a vacuum pump (either sorption or ion pump) and are connected by a valve which, when shut, isolates the measuring arm from both the pump and the reference arm. With the valve closed, gas is admitted to the measuring arm while the reference arm continues to be evacuated by the pump.

To measure the effect of the vapour pressure both arms of the manometer were evacuated by the ion pump through a liquid-nitrogen cold trap, and the output from the interferometer was obtained, via a digital-to-analog converter, on a chart recorder. The chart-recorder reading was set to zero and the valve connecting the two manometer arms was then closed.

When the valve was closed no initial change in the chart-recorder reading occurred. However, the reading increased linearly with time so that after 7 hours a reading of 8 fringes was recorded (equivalent to a pressure of 340 mPs). When the valve was opened at the end of this time the reading immediately returned to zero. The increase in pressure was due to small leaksge in the system, since, had it been caused by the vapour pressure of mercury, the chart-recorder would have reached a steady value between 0 and 4 fringes very quickly (approximately 0.1 s as discussed below). In fact, however, except for a small momentary increase in pressure after the valve

1

was opened the ion pump current indicated a pressure of 13 µPa throughout the experiment.

A simple calculation with Diels and Jaeckel's [6] value for the rate of evaporation of molecules $(1.42 \times 10^{21} \text{ molecule/m}^2.s)$ indicates that in our manometer (35 mm diameter tubing, 1 m long connected to each arm) the saturation vapour pressure of mercury should be established in a time of the order of 0.1 s. If this value is used to calculate the volume of mercury lost during evacuation a value of 30 mm³/h, much higher than actually occurs, is obtained; thus it appears that saturation vapour pressure is not reached near the pump.

These results show that under 'vacuum' the saturation vapour pressure of mercury, P, acts on the mercury surfaces in the two arms of our manometer (to an accuracy of 10 mPa - the resolution of the manometer) and that this pressure is not influenced by pumping with a vacuum pump. However when the measuring arm is connected to a comparison instrument, the vapour pressure at this instrument may, or may not, be the saturated vapour pressure of mercury, depending on the geometry of the connecting tubing. Kinetic theory states that the pressure in a closed system of constant and uniform temperature is proportional to the number of gas molecules, is independent of the nature of the gas and that equilibrium in pressure is established when there is no nett transfer of molecules to or from any region. If nitrogen is leaked into the measuring arm so that the pressure is in the viscous flow region (greater than approximately 1.3 Pa, where the transition from molecular to viscous flow occurs) then flow will occur until equilibrium as defined above is established and the pressure P at the comparison instrument will be equal to the pressure at the mercury surface. Thus, apart from altitude effects, the pressure P is balanced by the height of the mercury column h (measured by the interferometer) and P acting on the mercury surface in the reference arm, giving

Initially there may be concentration gradients of mercury vapour and nitrogen but mixing by diffusion will occur at uniform pressure and the total pressure in the system will slowly increase as the vapour pressure of mercury increases to the saturation vapour pressure throughout the system. (As shown in [7] this should take approximately 10 hours).

abust a site betacone at Pro-engent ogh i ter at the state of the stat

connected by a value which, when

2

For the molecular-flow region (pressures less than approximately 1.3 Pa), if a cold trap is placed between the manometer and the comparison instrument then, as shown in [8], the presence of the cold trap does not affect the pressure of any non-condensable gas at either instrument. In this case P cancels out and the absolute pressure is given directly by the number of counts on the manometer, at 340 afait. When the valve was opposed at

temodiately forunged to sero. The increase in prelative was one to small leakage in the system, since, had it. det not by the vapour presence of marcury, the chart-recorder would have reached a steady value between 0 and a fringes very quickly (sporestantely 0.1 a se discovered below). In fact, provided that equilibrium has been established.

Hence, as the saturation vapour pressure of mercury always acts on the mercury surface in the reference or 'vacuum' arm of our manometer, to obtain absolute pressure it is necessary to add the value of the mercury vapour pressure (160 mPa) to the pressure determined from the height difference between the two surfaces. This result applies for the region of viscous flow; but for molecular flow, if a cold trap is used, then the pressure is given by the height difference between the two surfaces. These results should apply to most other mercury manometers, but are significant only for measurements of very high precision.

REFERENCES

- Harrison, E.R., Hatt, D.J., Prowse, D.B., Wilbur-Ham, J. (1976). Metrologia, <u>12</u>, 115.
- Guildner, L.A., Stimson, H.F., Edsinger, R.E., Anderson, R.L. (1970). Metrologia, <u>6</u>, 1.
- 3. Hatt, D.J. and Prowse, D.B. (July 1977). Technical Note MRL-TN-403.
- Kaneda, R., Sudo, S., Nishibata, K. (1964). Bull. Nat. Res. Lab. Metrology, 9, 24.
- 5. Bass, A.H. and Green, E. (1972). ISA Transactions, <u>11</u>, 113.
- 6. Diels, K. and Jaeckel, R. (1966). Leybold Vacuum Handbook, p.55. Oxford: Pergamon Press.
- 7. Dushman, S. (1949). Scientific Foundations of Vacuum Technique, pp.77-78. London: Wiley.
- Guthrie, A. and Wakerling, R.K. (1949). Vacuum Equipment and Techniques, p.45. New York: McGraw-Hill.

3

(MRL-TN-413)

DISTRIBUTION LIST

MATERIALS RESEARCH LABORATORIES

Chief Superintendent Superintendent, Physics Division Mr. E.R. Harrison Dr. D.B. Prowse Mr. D.J. Hatt Library Librarian, N.S.W. Branch (Through Officer-in-Charge) Officer-in-Charge, Joint Tropical Trials and Research Establishment

DEPARTMENT OF DEFENCE

Chief Defence Scientist Executive Controller, ADSS Superintendent, Defence Science Administration, DSTO Superintendent, Military Advisers Branch Head, Laboratory Programs Branch Army Scientific Adviser Air Force Scientific Adviser Naval Scientific Adviser Chief Superintendent, Aeronautical Research Laboratories Senior Librarian, Defence Research Centre Director, Defence Research Centre Librarian, R.A.N. Research Laboratory Officer-in-Charge, Document Exchange Centre (16 copies) Principal Librarian, Campbell Park Library ADSATIS Annex Central Office, Directorate of Quality Assurance - Air Force Director, Joint Intelligence Organisation Head, Engineering Development Establishment

DEPARTMENT OF PRODUCTIVITY

NASA Canberra Office Head, B.D.R.S.S. (Aust.)

OTHER FEDERAL AND STATE DEPARTMENTS AND INSTRUMENTALITIES

The Chief Librarian, Central Library, C.S.I.R.O. Australian Atomic Energy Commission Research Establishment The Librarian, National Measurement Laboratory

(MRL-TN-413)

DISTRIBUTION LIST

(Continued)

MISCELLANEOUS - OVERSEAS

Defence Scientific and Technical Representative, Australian High Commission, London

Assistant Director/Armour and Materials, Military Vehicles and Engineering Establishment, England

Reports Centre, Directorate of Materials Aviation, England Library - Exchange Desk, National Bureau of Standards, U.S.A.

U.S. Army Standardization Group, Office of the Scientific Standardization Representative, Canberra, A.C.T.

Chief, Research and Development, Defence Scientific Information

Service, Canada (2 copies)

The Director, Defence Scientific Information and Documentation Centre, India.

Colonel B.C. Joshi, Military, Naval and Air Adviser, High Commission of India, Red Hill, A.C.T.

Director, Defence Research Centre, Malaysia

Exchange Section, British Library, England

Periodicals Recording Section, Science Reference Library. The British Library, England

INSPEC: Acquisition Section, Institution of Electrical Engineers, England

Overseas Reports Section, Defence Research Information Centre, England

Engineering Societies Library, U.S.A.

Library, Chemical Abstracts Service, U.S.A.

Director, National Engineering Laboratory, Glasgow

Librarian, Auckland Industrial Development Division, Department of Scientific and Industrial Research, Auckland.