



1. REPORT NUMBER	TION PAGE	READ INSTRUCTIONS
14 1574-01-1-2077	2. GOVT ACCESSION NO.	3. RECIPIENT'S CATALOG NUMBER
4. TITLE (and Sublide) (6) MILITARY (OTS ETE) SPECIFI AMMETER AC CLAMP-ON	CATION FOR THE	5. TYPE OF REPORT & PERIOD COVERED
ATTEIER, AC, CLAFF-ON 2	• 700- 1 · · · · · · · · · · · · · · · · · ·	6. PERFORMING ORG. REPORT NUMBER 1574-01-1-2077
Benjamin F. /Paiz Albert L./Simmons		B. CONTRACT OR GRANT NUMBER(+)
9. PERFORMING ORGANIZATION NAME AND A	DDRESS	10. PROGRAM ELEMENT, PROJECT, TASK
ARINC Research Corp. 2551 Riva Road		1315
Annapolis, MD 21401	e	12 070007 0475
CONTROLLING OFFICE NAME AND ADDRES	Ĩ	December 6079
·		12
14. MONITORING AGENCY NAME & ADDRESS(II	different from Controlling Office)	15. SECURITY CLASS. (of this report) Unclassified
		15. DECLASSIFICATION/DOWNGRADING SCHEDULE
Unlimited Thi	is document has been appropublic release and sale; it	oved
		· ·
18. SUPPLEMENTARY NOTES		
18. SUPPLEMENTARY NOTES 19. KEY WORDS (Continue on reverse side if neces	seary and identify by block number)	
18. SUPPLEMENTARY NOTES 19. KEY WORDS (Continue on reverse side if neces	every and identify by block number)	
<ol> <li>SUPPLEMENTARY NOTES</li> <li>KEY WORDS (Continue on reverse side if neces</li> <li>ABSTRACT (Continue on reverse side if neces</li> <li>This specification is appr Development Command, Depar all Departments and Agenci</li> </ol>	every and identify by block number) every and identify by block number) oved for use by the Co tment of the Army, and es of the Department of	ommunications Research and I is available for use by of Defense.
<ol> <li>SUPPLEMENTARY NOTES</li> <li>KEY WORDS (Continue on reverse elde if neces</li> <li>ABSTRACT (Continue on reverse elde if neces</li> <li>This specification is appr Development Command, Depar all Departments and Agenci</li> </ol>	eary and identify by block number) eary and identify by block number) oved for use by the Co tment of the Army, and es of the Department o	ommunications Research and d is available for use by of Defense.

р.,

H<sub>t</sub>T

.



MILITARY (OTS ETE) SPECIFICATION FOR THE AMMETER, AC, CLAMP-ON

December 1979

JAN 30 1980

.

. . . . .

Prepared for

U.S. Army Communications and Electronics Materiel Readiness Command Fort Monmouth, New Jersey under Contract DAAB07-78-A-6606

by Benjamin F. Paiz Albert L. Simmons

ARINC Research Corporation a Subsidiary of Aeronautical Radio, Inc. 2551 Riva Road Annapolis, Maryland 21401 Publication 1574-01-1-2077

# MILITARY (OTS ETE) SPECIFICATION

This specification is approved for use by the Communications Research and Development Command, Department of the Army, and is available for use by all Departments and Agencies of the Department of Defense.

# 1. SCOPE

1.1 Scope. This specification defines an Ammeter, AC, Clamp-On, hereinafter called the equipment.

1.2 <u>Classification</u>. The equipment defined by this specification shall be Type II, Class 5, Style E, Color R, per MIL-T-28800.

## 2. APPLICABLE DOCUMENTS

2.1 <u>Issues of documents</u>. The following documents of the issue in effect on the data of invitation for bids, or the request for proposal, form a part of this specification to the extent specified herein.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: U.S. Army Communications Research and Development Command, Fort Monmouth, New Jersey 07703 by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

1

Actual	a en Fer	
Nilo DOJ TA Unilat Justif	indel B icition	B
By Distri Avail	hution/	Codes
Dist	Avail and special	1/or 1

FSC 6625

#### SPECIFICATIONS

#### MILITARY

MIL-T-28800	Test Equipment for Use with Electrical and
	Electronic Equipment, General Specification for

# STANDARDS

MILITARY

MIL-STD-461	Electromagnetic Interference Characteristics, Requirements for Equipments
MIL-STD-462	Electromagnetic Interference Characteristics, Measurement of

MIL-STD-781B Reliability Tests, Exponential Distribution

(Copies of specifications, standards, drawings and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

# 3. REQUIREMENTS

3.1 <u>Classification of requirements</u>. The requirements for the equipment are classified as follows:

REQUIREMENT	PARAGRAPH
Safety	3.2
Parts, materials and processes	3.3
Design and construction	3.4
Dimensions and weight	3.5
Enclosure requirements	3.6
Marking and identification	3.7
Environmental requirements	3.8
Reliability requirements	3.9
Performance characteristics	3.10

3.2 <u>Safety</u>. Unless otherwise specified herein, the equipment shall comply with the Type II safety requirements specified in MIL-T-28800.

3.3 Parts, materials and processes. Unless otherwise specified herein, the equipment shall comply with the Type II parts, materials and processes requirements of MIL-T-28800.

3.3.1 <u>Restricted materials</u>. Equipment shall comply with the restricted material requirements of MIL-T-28800 including the requirement for mercury or radioactive materials and shall contain no combination of materials which cause deterioration of any material contained in the equipment due to effects of outgassing.

3.4 <u>Design and construction</u>. Unless otherwise specified herein, the equipment shall comply with the Type II, Class 5, Style E, Color R, design and construction requirements of MIL-T-28800.

3.4.1 First article. When specified, the contractor shall furnish 5 sample equipments for first article inspection and approval. (See 4.3 and 6.2)

3.4.2 <u>Controls</u>. Unless otherwise specified herein, built-in adjustments and compensating devices shall not be externally accessible.

3.4.2.1 Front panel controls. All controls which are required to operate the equipment throughout its specified performance characteristics shall be located on the front panel.

3.4.3 Accessibility. The equipment shall be constructed so that:

3.4.3.1 Subassemblies and chassis components can be removed without removing any other hard wired subassembly, printed circuit card or component.

3.4.3.2 Adjustments can be made without removing any component, printed circuit card or subassembly, except that the use of extender cards is permitted.

3.4.3.3 Printed circuit cards can be removed without the need to unsolder cables and interconnecting wiring (connections to all printed circuit cards shall be through pin and socket connectors). Printed circuit cards (mother boards) designed primarily to distribute power and signals to other printed cards (daughter boards) are excluded from this requirement. When such mother boards are used, they shall be accessible from both sides to allow maintenance testing.

3.4.3.4 Encapsulation and embedment. Encapsulation and embedment (potting) of subassemblies shall not be used.

3.4.4 <u>Solid-state construction</u>. Unless otherwise specified herein, the equipment shall be of solid-state, modular, miniaturized construction.

3.5 Dimensions and weight.

3.5.1 <u>Dimensions</u>. The overall dimensions shall be 229 mm/9" maximum height, 102 mm/4" maximum depth and 152 mm/6" maximum width.

3.5.2 Weight. The maximum weight of the equipment shall be 1.45 kg/3 lbs.

3.6 Enclosure requirements. Unless otherwise specified herein, the equipment shall comply with the Style E enclosure requirements of MIL-T-28800.

3.7 <u>Marking and identification</u>. Unless otherwise specified herein, the equipment shall comply with the Type II marking and identification requirements of MIL-T-28800.

3.7.1 <u>Supplemental identification plate</u>. The supplemental identification plate specified in MIL-T-28800 shall contain the following data only:

- a. Nomenclature
- b. Procurement instrument identification number (PIIN)
- c. Serial number
- d. National stock number
- e. US

3.8 <u>Environmental requirements</u>. Unless otherwise specified herein, the equipment shall comply with the Class 5 environmental requirements of MIL-T-28800.

3.8.1 <u>Electromagnetic interference</u>. The equipment shall comply with the following emission and susceptibility requirements of Notice 4, MIL-STD-461:

CE02 CS02 RE02.1 RS03 CE04 CS06 RE02

**RE02.1** and RS03 shall be performed from (To Be Determined, TBD\*) with RS03 at a susceptibility level of one volt per meter (1 V/m).

**3.8.2** <u>Humidity</u>. The equipment shall meet the humidity requirements in accordance with the test specified in paragraph 4.5.5.1.1.3 of MIL-T-28800.

3.8.3 <u>Vibration</u>. The equipment shall comply with the Class 5 vibration requirements of MIL-T-28800, except that the equipment need not be operating during vibration.

#### 3.9 Reliability requirements.

3.9.1 <u>Reliability burn-in</u>. Each equipment delivered against this specification shall be subjected to a minimum 96-hour on-time burn-in procedure as specified in 4.4.4. The last 24 hours of burn-in shall be failure free.

3.9.2 <u>Reliability</u>. Reliability shall comply with requirements as specified herein. The specified MTBF shall be 5,000 hours when tested as specified in 4.4.4. A failure shall be as defined in MIL-STD-781B, and as any departure from the required performance or operation of the required accuracies (not correctable by normal use of the operating controls) after the test is initiated. Test Level B, MIL-STD-781B shall be the required test level.

3.9.3 <u>Maintainability requirements</u>. The equipment shall comply with the Type II maintainability requirements of MIL-T-28800. (See 6.4)

# 3.10 Performance characteristics.

3.10.1 <u>Current, ac</u>. The equipment shall measure ac current from 0 to 300 amperes, in full-scale readings of 3 A, 30 A, 300 A. The measurement accuracy of the equipment shall be within  $\pm 3$  percent of the indicated value, at 50 or 60 Hz, and within  $\pm 5$  percent, at 400 Hz.

<sup>\*</sup>All TBD items in this specification must be completed by CERCOM/CORADCOM prior to official U.S. Army approval and publication.

3.10.2 <u>Clamp-on</u>. The equipment shall contain a clamp-on type device, trigger-operated, which permits current measurement and readings without loading or circuit interruption.

4. QUALITY ASSURANCE PROVISIONS

4.1 <u>Responsibility for inspection</u>. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.2 <u>Classification of inspections</u>. The inspections required herein are classified as in a and b.

- a. First article inspection (see 4.3).
- b. Quality conformance inspection (see 4.3).

4.3 First article and quality conformance inspection. Unless otherwise specified herein, the first article and quality conformance inspections shall be in accordance with MIL-T-28800.

4.3.1 Test plan. Unless otherwise stated in the contract, the contractor shall prepare a test plan in accordance with paragraphs 4.3(a) and 4.4(a) of MIL-T-28800, for use with both classes of inspection. The test plan shall include as a minimum the tests listed in TABLES I, II, & III, the tests noted in section 4.4, and a description of the satisfactory operation check and satisfactory operation test, as defined in para 4.5 of MIL-T-28800. Unless otherwise required, the tests and inspections to be performed shall be identical for both classes of inspection.

4.3.2 <u>Inspection sampling plan</u>. The inspection sampling to be performed during both classes of inspection shall be as specified in MIL-T-28800, except as indicated below:

4.3.2.1 Unless otherwise specified herein or in the contract, the Group C, D, E & F quality conformance inspections shall be performed on the first production lot only.

4.3.2.2 The Group D EMI tests and Group F Reliability test shall be performed during quality conformance inspection only if not performed during first article or bid sampling testing.

4.4 Quality assurance tests. Unless otherwise specified herein, the following tests shall be performed in accordance with MIL-T-28800.

#### TABLE I

DESCRIPTION	REQUIREMENT	TEST METHOD
Group A Preoperational Inspection Leakage Current Level A Performance	4.5.3.1* 3.2.1.3.1* TABLE II	4.3.1
Group B Level B Performance	TABLE III	4.3.1
Group C Electrical Power Environmental Requirements Humidity Vibration	3.5 3.9 3.9.2 3.9.3	4.3.1
Group D Electromagnetic Interference	3.9.1	4.4.1
Group E Dimensions Weight Front Panel Marking	3.6.1 3.6.2 3.8	4.3.1 4.3.1
Group F Reliability	3.10	4.3.1, 4.4.4
*Paragraphs of MIL-T-28800.		

#### Examination and Test Groups

4.4.1 Electromagnetic interference. One equipment shall be subjected to an EMI test for compliance with the requirements of 3.9.1. All test set-ups and procedures shall comply with the measurement techniques of Notice 3, MIL-STD-462. All emission and susceptibility tests shall be performed with the output cable connected to, and extended parallel with, the front of the equipment. The output cable shall be terminated to simulate a normal loading configuration.

**4.4.2** <u>Humidity</u>. The equipment shall be subjected to the humidity test specified in para 4.5.5.1.1.3 of MIL-T-28800.

#### TABLE II

#### Level A Performance Tests\*

DESCRIPTION	REQUIREMENTS	
Current, ac Clamp-on	3.11.1 3.11.2	
*Level A Testing is Abbreviated Testing (see 6.4.1).		

4.4.3 <u>Vibration</u>. The equipment shall be subjected to the vibration tests specified in MIL-T-28800, for Class 5 equipment, except that the equipment shall not be operated during the test. The satisfactory operation test shall be performed prior to and following vibration testing.

4.4.4 Reliability. The reliability tests shall be performed as follows:

4.4.4.1 <u>Burn-in</u>. Each deliverable equipment shall be subjected to a minimum 96 hours on-time burn-in period, prior to Group A testing. During the last 24 hours of burn-in, the equipment must operate failure-free. Up until this time, equipment will be allowed to accumulate failures. Each equipment which fails during the final 24-hour period shall be repaired and returned to test until it successfully survives a 24-hour period without failure. Failures which occur during the burn-in test shall be noted and reported, but shall not count toward the establishment of equipment MTBF. Prior to burn-in, the satisfactory operation test of 4.3.1 shall be conducted. Daily satisfactory operation checks (4.3.1) shall be conducted. For the last 24-hour failure-free period, a complete satisfactory operation test shall be conducted prior to and after the period.

# 4.4.4.2 Reliability sampling plan.

4.4.4.2.1 <u>First Article</u>. The Group F reliability demonstration shall be conducted on at least five (5) samples. The use of ten (10) samples is encouraged, but not more than ten (10) samples shall be used.

4.4.4.2.2 Quality conformance inspection. The Group F reliability tests shall be performed on the first production lot, only if not performed during First Article testing. From those units of the first lot that have passed the required Group A and B tests, a random sample of ten (10) units shall be selected for Group F testing.

4.4.4.2.3 <u>Reliability test plan</u>. The reliability tests shall be conducted in accordance with the following test plan:

 $\Theta_{O} = (TBD)$ 

Number of Failures	Accept** (Equal or More)	Total Test Time* Reject*** _(Equal or Less)
0		
1		
2		
3	(TBD)	
4		
5		
6		* * * *

\* Total test time is total unit hours of "equipment ON" time (in hours).
\*\* Accept if test time is greater than or equal to that listed.
\*\*\* Reject if test time is less than or equal to that listed.
\*\*\*\* Reject if test time is strictly less than (TBD) with 6 failures.

4.4.4.3 <u>Test length</u>. Testing shall continue until the total unit hours together with the total count of relevant equipment failures permit eithe, an accept or reject decision in accordance with the specified test plan. Only equipment "ON" time may be used in MTBF or longevity determinations. Testing shall be monitored in such a manner that the times to failure may be estimated with reasonable accuracy. No single equipment "ON" time shall be less than one-half the average operating time of all equipments "ON" test.

4.4.4.4 Test conditions.

4.4.4.1 Operating mode (duty cycle).

4.4.4.2 <u>Test level</u>. While under test, the equipment shall be subjected to the conditions specified by Test Level B, MIL-STD-781B.

4.4.4.3 <u>Input voltage cycling</u>. When so directed by the procuring activity, voltage cycling shall be accomplished as follows: The input voltage shall be maintained at one hundred ten percent (110%) nominal voltage for one-third of the equipment "ON" cycle, at the nominal value for the second one-third of the equipment "ON" cycle, and at ninety percent (90%) for the final one-third of the equipment "ON" cycle. This cycling procedure is to be repeated continuously throughout the reliability test.

4.4.4.5 <u>Reliability examination and test method</u>. The following inspections shall be used to verify equipment operation during reliability testing. The satisfactory operation test of 4.3.1 shall be used to confirm proper equipment operation prior to and following reliability testing. The satisfactory operation test shall also be performed weekly to verify equipment operation. The satisfactory operation check of 4.3.1 shall be used to monitor proper operation of the equipment daily.

4.4.4.6 <u>Reliability failure actions, Group F</u>. In the event of failure(s) during reliability testing, failure action shall be taken as required in MIL-STD-781B.

4.4.4.7 <u>Corrective action, Group F</u>. In the event the reliability test reaches a reject decision, corrective action shall be taken as required in MIL-STD-781B.

4.4.5 Optional quality assurance system. The procuring activity may substitute tests from a Department of Defense (DoD) approved supplier's quality assurance system for any or all tests in TABLE I.

5. PACKAGING

5.1 Preservation, packaging. Packaging for delivery shall be in accordance with MIL-T-28800 and as specified by the procuring activity.

6. NOTES

6.1 <u>Intended use</u>. The equipment is intended for use in maintaining electronic communications equipment by equipment repair units in intermediate and depot maintenance.

6.1.1 Equipment replaced. Equipment procured according to this specification is intended to replace all versions of various older equipments now fielded. These equipments to be replaced are listed below: (TBD)

6.2 Ordering data. Procurement documents should specify the following:

- a. Title, number and data of this specification and any amendment thereto
- b. Packaging requirements (see Section 5)
- c. When rough handling and functional tests are required
- d. Place of final inspection
- e. Technical literature required
- f. Quantity of tools and running spare parts required
- g. Marking for shipment and shipping containers
- h. Test plans and test reports
- i. Classification of inspection and number of samples required
- j. Rack mounting requirements
- k. Alternate power requirements
- 1. Maintainability rationale
- m. Reliability rationale

#### 6.3 Contract data requirements.

- a. Nomenclature assignment
- b. National stock number
- c. Equipment sample test plans
- d. Pretest performance records
- e. First production test data
- f. Identification plate drawing

# 6.4 Definitions.

6.4.1 Level A performance tests. Level A testing (TABLE II) is a reduced amount of testing which is performed on each equipment produced. Its purpose is to insure that all functions and modes of operation of the equipment are evaluated without extensively checking each parameter as required in a Level B test (TABLE III). The approved equipment test procedure shall specify the actual amount of testing to be performed.

6.4.2 <u>Reliability rationale</u>. Reliability rationale submitted with the bid should provide clear and concise rationale showing how the reliability of the equipment complies with minimum requirements of the solicitation. This may include data from previous reliability tests, reliability predictions and other data available to the offeror. In the absence of such data on the equipment being offered, such data on similar equipment of equal or greater complexity produced by the offeror may be submitted. However, such data must be clearly identified as comparative data, and accompanied with specifications and technical literature on the similar equipment. The acceptance of the submitted rationale does not relieve the successful bidder of performing and successfully completing the production reliability testing.

6.4.3 <u>Maintainability rationale</u>. Maintainability rationale submitted with the bid should provide clear and concise rationale showing how the maintainability of the equipment complies with requirements of the solicitation. This may include data from previous maintainability tests, records of repair and calibration data, and other data on the equipment available to the offeror. Data on selection of module size and other design features relevant to maintainability may also be submitted.

Custodian:

Army - CR

Preparing Activity: Army - CR

(Project 6625-

**Review** Activities:

Army -

User Activities:

Army -

STANDARDIZATION DOCUMENT IMPROV	EMENT PROPOSAL	OMB Approval No. 22-R255	
INSTRUCTIONS: The purpose of this form is to soli ment of suitable products at reasonable cost and min DoD contractors, government activities, or manufactu are invited to submit comments to the government. F preparing activity. Comments submitted on this form portion of the referenced document(s) or to amend con may be of use in improving this document. If there a envelope addressed to preparing activity.	cit beneficial comments w imum delay, or will otherw rers/vendors who are pro 'old on lines on reverse si do not constitute or imply htractual requirements. A re additional papers, state	hich will help achieve procure- rise enhance use of the document. spective suppliers of the product de, staple in comer, and send to r authorization to waive any ttach any pertinent data which th to form and place both in an	
DOCUMENT IDENTIFIER AND TITLE			
NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER		
	MATERIAL PROCURED UN	DER A	
	DIRECT GOVERNMEN	T CONTRACT USUBCONTRACT	
1. HAS ANY PART OF THE DOCUMENT CREATED PROB	LENS OR REQUIRED INTER	RPRETATION IN PROCUREMENT	
A. GIVE PARAGRAPH NUMBER AND WORDING.			
. NECOMMENCE INTO FOR CONNECTING THE DEF			
2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONS	IDERED TOO RIGID		
S. IS THE DOCUMENT RESTRICTIVET			
TES INO (Il "Yee", in what way?)			
4. REMARKS			
· .		•	
SUMMITTED BY (Printed or typed name and address - Option			
	DAT		
•			
		C16061	
UU 1 JAN 72 1420 REPLACES EDIT	TION OF 1 JAN 66 WHICH M	AY DE USED	

FOLD

Department of the Army Headquarters U.S.Army Electronics Command Fort Monmouth, New Jersey 07703

OFFICIAL BUSINESS

Commanding General U.S. Army Electronics Command ATTN: DRSEL-RD-TS Fort Monmouth, New Jersey 07703

FOLD

& U.S. COVERNMENT PRINTING OFFICE: 1976- - ADD-Pop/4149

- -

POSTAGE AND FEES PAID

To detach this form, cut along this line