

AD-A080 011

ARINC RESEARCH CORP ANNAPOLIS MD

F/G 14/2

MILITARY (OTS ETC) SPECIFICATION FOR THE TRANSMISSION TEST SET, (U)

DEC 79 B F PAIZ, A L SIMMONS

DAAB07-78-A-6606

UNCLASSIFIED


1574-01-1-2087

NL

(OF)
AD
A080011



END
DATE
FILMED
2-80
DDC



MICROCOPY RESOLUTION TEST CHART
NATIONAL BUREAU OF STANDARDS-1963-A

11

LEVEL II

ADA080011

**MILITARY (OTS ETE) SPECIFICATION
FOR THE
TRANSMISSION TEST SET**

December 1979


**DDC
RECEIVED
JAN 30 1980
E**

DDC FILE COPY

Prepared for
**U.S. ARMY COMMUNICATIONS AND ELECTRONICS
MATERIEL READINESS COMMAND
FORT MONMOUTH, NEW JERSEY
under Contract DAAB07-78-A-6606**

This document has been approved
for public release and sale; its
distribution is unlimited.

ARINC RESEARCH CORPORATION




REPORT DOCUMENTATION PAGE		READ INSTRUCTIONS BEFORE COMPLETING FORM
1. REPORT NUMBER 14 1574-01-1-2087	2. GOVT ACCESSION NO.	3. RECIPIENT'S CATALOG NUMBER
4. TITLE (and Subtitle) 6 MILITARY (OTS ETE) SPECIFICATION FOR THE TRANSMISSION TEST SET		5. TYPE OF REPORT & PERIOD COVERED
7. AUTHOR(s) 10 Benjamin F. Paiz Albert L. Simmons		6. PERFORMING ORG. REPORT NUMBER 1574-01-1-2087
9. PERFORMING ORGANIZATION NAME AND ADDRESS ARINC Research Corp. 2551 Riva Road Annapolis, MD 21401		8. CONTRACT OR GRANT NUMBER(s) 15 DAAB07-78-A-6606
11. CONTROLLING OFFICE NAME AND ADDRESS 12 16		10. PROGRAM ELEMENT, PROJECT, TASK AREA & WORK UNIT NUMBERS
14. MONITORING AGENCY NAME & ADDRESS (if different from Controlling Office)		12. REPORT DATE 17 December 1979
		13. NUMBER OF PAGES 14
		15. SECURITY CLASS. (of this report) Unclassified
		15a. DECLASSIFICATION/DOWNGRADING SCHEDULE
16. DISTRIBUTION STATEMENT (of this Report) Unlimited This document has been approved for public release and sale; its distribution is unlimited.		
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report)		
18. SUPPLEMENTARY NOTES		
19. KEY WORDS (Continue on reverse side if necessary and identify by block number)		
20. ABSTRACT (Continue on reverse side if necessary and identify by block number) This specification is approved for use by the Communications Research and Development Command, Department of the Army, and is available for use by all Departments and Agencies of the Department of Defense.		

400 247

JCB

MILITARY (OTS ETE) SPECIFICATION
FOR THE
TRANSMISSION TEST SET

December 1979



Prepared for
U.S. Army Communications and Electronics
Materiel Readiness Command
Fort Monmouth, New Jersey
under Contract DAAB07-78-A-6606

by
Benjamin F. Paiz
Albert L. Simmons

ARINC Research Corporation
a Subsidiary of Aeronautical Radio, Inc.
2551 Riva Road
Annapolis, Maryland 21401
Publication 1574-01-1-2087

This document has been approved
for public release and sale; its
distribution is unlimited.

MILITARY (OTS ETE) SPECIFICATION

This specification is approved for use by the Communications Research and Development Command, Department of the Army, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification defines a Transmission Test Set, hereinafter called the equipment.

1.2 Classification. The equipment defined by this specification shall be Type II, Class 5, Style E, Color R per MIL-T-28800 and, as herein stated, with the convertible/rack-mountable capability.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on the date of invitation for bids, or the request for proposal, form a part of this specification to the extent specified herein.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: U.S. Army Communications Research and Development Command, Fort Monmouth, New Jersey 07703 by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

ACQUISITION FOR	
THIS GRA&I	☒
DDC TAB	☐
Unannounced	☐
Justification	
By _____	
Distribution/ _____	
Availability Codes	
Dist.	Availend/or special
A	

FSC 6625

SPECIFICATIONS

MILITARY

MIL-T-28800 Test Equipment for Use with Electrical and
Electronic Equipment, General Specification for

STANDARDS

MILITARY

MIL-STD-461 Electromagnetic Interference Characteristics,
Requirements for Equipments

MIL-STD-462 Electromagnetic Interference Characteristics,
Measurement of

MIL-STD-781B Reliability Tests, Exponential Distribution

(Copies of specifications, standards, drawings and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Classification of requirements. The requirements for the equipment are classified as follows:

<u>REQUIREMENT</u>	<u>PARAGRAPH</u>
Safety	3.2
Parts, materials and processes	3.3
Design and construction	3.4
Electrical power sources and connections	3.5
Dimensions and weight	3.6
Enclosure requirements	3.7
Marking and identification	3.8
Environmental requirements	3.9
Reliability requirements	3.10
Performance characteristics	3.11

3.2 Safety. Unless otherwise specified herein, the equipment shall comply with the Type II safety requirements specified in MIL-T-28800.

3.3 Parts, materials and processes. Unless otherwise specified herein, the equipment shall comply with the Type II parts, materials and processes requirements of MIL-T-28800.

3.3.1 Restricted materials. Equipment shall comply with the restricted material requirements of MIL-T-28800 including the requirement for mercury or radioactive materials and shall contain no combination of materials which cause deterioration of any material contained in the equipment due to effects of outgassing.

3.4 Design and construction. Unless otherwise specified herein, the equipment shall comply with the Type II, Class 5, Style E, Color R, design and construction requirements of MIL-T-28800 with the convertible/rack-mountable capability.

3.4.1 First article. When specified, the contractor shall furnish five sample equipments for first article inspection and approval. (See 4.3 and 6.2)

3.4.2 Mainframe plug-in concept. Equipment utilizing externally accessible, externally removable subassemblies (mainframe plug-in concept) do not meet the requirements of this specification and therefore are unacceptable. Items using this concept may be offered, provided the plug-in(s) shall not be removable by access through the front panel or rear panel with the unaided hand. A system part number shall be assigned to any such mainframe plug-in combination required to meet the requirements of this specification. The system part number shall be marked on the mainframe identification plate and any plug-in(s).

3.4.3 Controls. Unless otherwise specified herein, built-in adjustments and compensating devices shall not be externally accessible.

3.4.3.1 Front panel controls. All controls which are required to operate the equipment throughout its specified performance characteristics shall be located on the front panel.

3.4.4 Accessibility. The equipment shall be constructed so that:

3.4.4.1 Subassemblies and chassis components can be removed without removing any other hard wired assembly, printed circuit card or component.

3.4.4.2 Adjustments can be made without removing any component, printed circuit card or subassembly, except the use of extender cards is permitted.

3.4.4.3 Printed circuit cards can be removed without the need to unsolder cables and interconnecting wiring (connections to all printed circuit cards shall be through pin and socket connectors). Printed circuit cards (mother boards) designed primarily to distribute power and signals to other printed cards (daughter boards) are excluded from this requirement. When such mother boards are used, they shall be accessible from both sides to allow maintenance testing.

3.4.4.4 Indicator lights. Unless approved by the procuring activity upon presentation of acceptable reliability data, indicator lights other than light emitting diodes (LEDs) shall be accessible from the operator's side of the front panel.

3.4.4.5 Encapsulation and embedment. Encapsulation and embedment (potting) of subassemblies shall not be used.

3.4.5 Solid-state construction. Unless otherwise specified herein, the equipment shall be of solid-state, modular, miniaturized construction.

3.5 Electrical power sources and connections. Unless otherwise specified herein, the equipment shall comply with the Type II electrical power sources and connections requirements of MIL-T-28800. The equipment shall operate from nominal 115/230 volts, single phase, 50, 60 and 400 Hz source. The equipment shall be capable of continuous operation from a rechargeable battery source at a minimum of eight hours before recharge.

3.5.1 Maximum power. The maximum power consumption of the equipment shall be 30 watts.

3.5.2 Input power selection device. An input power selection device shall be provided for selection of input power voltages of 115 Vac or 230 Vac. Provision shall be incorporated to prevent accidental switching. When the equipment is delivered, the power selection device shall be in the 115 Vac position.

3.5.3 Fuses and circuit breakers. Fuses and circuit breakers shall be in accordance with MIL-T-28800. (115 Vac/230 Vac) Either common or separate fuseholders may be provided. If only one fuseholder is used (common), the equipment shall be provided with the 115 Vac fuse installed and the 230 Vac fuse shall be stowed with the accessories.

3.5.4 Input power switch. A front panel-mounted power switch shall be provided. The ON position shall have panel identification lights for ac operation. The switch shall break both sides of the power source.

3.6 Dimensions and weight.

3.6.1 Dimensions. The overall dimensions shall be 229 mm/9" maximum height, 356 mm/14" maximum depth and maximum width of 483 mm/19" or as specified in MIL-T-28800 for rack-mounted equipment. A blank plate may be required to satisfy incremental height requirements.

3.6.2 Weight. The maximum weight of the equipment shall be 13.6 kg/30 lbs.

3.7 Enclosure requirements. Unless otherwise specified herein, the equipment shall comply with the Style E enclosure requirements of MIL-T-28800.

3.8 Marking and identification. Unless otherwise specified herein, the equipment shall comply with the Type II marking and identification requirements of MIL-T-28800.

3.8.1 Supplemental identification plate. The supplemental identification plate specified in MIL-T-28800 shall contain the following data only:

- a. Nomenclature
- b. Procurement instrument identification number (PIIN)
- c. Serial number
- d. National stock number
- e. US

3.9 Environmental requirements. Unless otherwise specified herein, the equipment shall comply with the Class 5 environmental requirements of MIL-T-28800.

3.9.1 Electromagnetic interference. The equipment shall comply with the following emission and susceptibility requirements of Notice 4, MIL-STD-461:

CE02	CS02	RE02.1	RS03
CE04	CS06	RE02	

RE02.1 and RS03 shall be performed from (To Be Determined, TBD*) with RS03 at a susceptibility level of one volt per meter (1 V/m).

3.9.2 Humidity. The equipment shall meet the humidity requirements in accordance with the test specified in paragraph 4.5.5.1.1.3 of MIL-T-28800.

3.9.3 Vibration. The equipment shall comply with the Class 5 vibration requirements of MIL-T-28800, except that the equipment need not be operating during vibration.

3.10 Reliability requirements.

3.10.1 Reliability burn-in. Each equipment delivered against this specification shall be subjected to a minimum 96-hour on-time burn-in procedure as specified in 4.4.4. The last 24 hours of burn-in shall be failure-free.

3.10.2 Reliability. Reliability shall comply with requirements as specified herein. The specified MTBF shall be 3,000 hours when tested as specified in 4.4.4. A failure shall be as defined in MIL-STD-781B and as any departure from the required performance or operation of the required accuracies (not correctable by normal use of the operating controls) after the test is initiated. Test Level B, MIL-STD-781B shall be the required test level.

3.10.3 Maintainability requirements. The equipment shall comply with the Type II maintainability requirements of MIL-T-28800. (See 6.4)

3.11 Performance Characteristics.

3.11.1 Frequency range. The equipment shall be capable of sending and receiving a selected oscillator signal level simultaneously in the frequency range that extends from 5 Hz to 560 Hz. The output (send) frequency shall be continuously variable and, at a minimum, be within ± 2 percent of the dial reading.

3.11.2 Output level. The output level (send side) shall be controlled by a precision step attenuator which allows the selection of output levels from -110 to +10 dBm in full-scale readings. The maximum number of output level selection shall be 12 range settings.

*All TBD items in this specification must be completed by CERCOM/CORADCOM prior to official U.S. Army approval and publication.

3.11.2.1 Output Harmonic Distortion. The oscillator output (send side) harmonic distortion shall, at a minimum, be 40 dB below the level of the fundamental frequency.

3.11.2.2 Spurious response. The spurious response of the oscillator send circuit shall, at a minimum, be 65 dB below the level of the fundamental frequency.

3.11.3 Input frequency response. The input frequency response (receive side) of the equipment shall be flat ± 0.2 dB between 5 Hz and 560 Hz.

3.11.3.1 Input level measurement. The input level measurement of the equipment shall be selectable with full-scale readings of -110 to +10 dBm or 0 to 300 volts rms with no more than 12 range settings.

3.11.4 Impedance matching. The equipment shall have input and output impedance selection of 135, 600 and 900 ohms. It shall also have provision for high impedance bridging on both send and receive sides.

3.11.4.1 Input/output balance. The equipment input and output terminals shall be balanced with respect to ground.

3.11.4.2 Longitudinal balance. The longitudinal balance of input and output circuits shall, at a minimum, be 40 dB.

3.11.5 Primary connectors. The equipment shall be capable of accepting Western Electric 309, 310 and 241 type plugs for both the input and output jacks.

3.11.6 Readout method. The equipment shall contain a meter, analog or digital, that is calibrated to read in dBm and volts rms.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.2 Classification of inspections. The inspections required herein are classified as in a and b.

- a. First article inspection (see 4.3).
- b. Quality conformance inspection (see 4.3).

4.3 First article and quality conformance inspection. Unless otherwise specified herein, the first article and quality conformance inspections shall be in accordance with MIL-T-28800.

4.3.1 Test plan. Unless otherwise stated in the contract, the contractor shall prepare a test plan in accordance with paragraphs 4.3(a) and 4.4(a) of MIL-T-28800, for use with both classes of inspection. The test plan shall include as a minimum the tests listed in TABLES I, II and III, the tests noted in section 4.4, and a description of the satisfactory operation check and satisfactory operation test, as defined in para 4.5 of MIL-T-28800. Unless otherwise required, the tests and inspections to be performed shall be identical for both classes of inspection.

TABLE I

Examination and Test Groups

DESCRIPTION	REQUIREMENT	TEST METHOD
Group A		
Preoperational Inspection	4.5.3.1*	
Leakage Current	3.2.1.3.1*	4.3.1
Level A Performance	TABLE II	
Group B		
Level B Performance	TABLE III	4.3.1
Group C		
Electrical Power	3.5	
Environmental Requirements	3.9	4.3.1
Humidity	3.9.2	
Vibration	3.9.3	
Group D		
Electromagnetic Interference	3.9.1	4.4.1
Group E		
Dimensions	3.6.1	
Weight	3.6.2	4.3.1
Front Panel Marking	3.8	4.3.1
Group F		
Reliability	3.10	4.3.1, 4.4.4
*Paragraphs of MIL-T-28800.		

TABLE II

Level A Performance Tests*

DESCRIPTION	REQUIREMENTS
Frequency Range	3.11.1
Output Level	3.11.2
Input Frequency Response	3.11.3
Impedance Matching	3.11.4
Primary Connectors	3.11.5
Readout Method	3.11.6
*Level A Testing is Abbreviated Testing (see 6.4.1).	

TABLE III

Level B Performance Test

DESCRIPTION	REQUIREMENTS
Output Harmonic Distortion	3.11.2.1
Spurious Response	3.11.2.2
Input Level Measurement	3.11.3.1
Input/Output Balance	3.11.4.1
Longitudinal Balance	3.11.4.2

4.3.2 Inspection sampling plan. The inspection sampling to be performed during both classes of inspection shall be as specified in MIL-T-28800, except as indicated below:

4.3.2.1 Unless otherwise specified herein or in the contract, the Group C, D, E and F quality conformance inspections shall be performed on the first production lot only.

4.3.2.2 The Group D EMI tests and Group F Reliability test shall be performed during quality conformance inspection only if not performed during first article or bid sampling testing.

4.4 Quality assurance tests. Unless otherwise specified herein, the following tests shall be performed in accordance with MIL-T-28800.

4.4.1 Electromagnetic interference. One equipment shall be subjected to an EMI test for compliance with the requirements of 3.9.1. All test set-ups and procedures shall comply with the measurement techniques of Notice 3, MIL-STD-462. All emission susceptibility tests shall be performed with the output cable connected to, and extended parallel with, the front of the equipment. The output cable shall be terminated to simulate a normal loading configuration.

4.4.2 Humidity. The equipment shall be subjected to the humidity test specified in para 4.5.5.1.1.3 of MIL-T-28800.

4.4.3 Vibration. The equipment shall be subjected to the vibration tests specified in MIL-T-28800, for Class 5 equipment, except that the equipment shall not be operated during the test. The satisfactory operation test shall be performed prior to and following vibration testing.

4.4.4 Reliability. The reliability tests shall be performed as follows:

4.4.4.1 Burn-in. Each deliverable equipment shall be subjected to a minimum 96 hours on-time burn-in period, prior to Group A testing. During the last 24 hours of burn-in, the equipment must operate failure-free. Up until this time, equipment will be allowed to accumulate failures. Each equipment which fails during the final 24-hour period shall be repaired and returned to test until it successfully survives a 24-hour period without failure. Failures which occur during the burn-in test shall be noted and reported, but shall not count toward the establishment of equipment MTBF. Prior to burn-in, the satisfactory operation test of 4.3.1 shall be conducted. Daily satisfactory operation checks (4.3.1) shall be conducted. For the last 24-hour failure-free period, a complete satisfactory operation test shall be conducted prior to and after the period.

4.4.4.2 Reliability sampling plan.

4.4.4.2.1 First article. The Group F reliability demonstration shall be conducted on at least five (5) samples. The use of ten (10) samples is encouraged, but not more than ten (10) samples shall be used.

4.4.4.2.2 Quality conformance inspection. The Group F reliability tests shall be performed on the first production lot, only if not performed during first article testing. From those units of the first lot that have passed the required Group A and B tests, a random sample of ten (10) units shall be selected for Group F testing.

4.4.4.2.3 Reliability test plan. The reliability tests shall be conducted in accordance with the following test plan:

$$\theta_0 = (\text{TBD})$$

<u>Number of Failures</u>	<u>Accept** (Equal or More)</u>	<u>Total Test Time* Reject*** (Equal or Less)</u>
0		
1		
2		
3	(TBD)	
4		
5		
6		****

* Total test time is total unit hours of "equipment ON" time (in hours).

** Accept if test time is greater than or equal to that listed.

*** Reject if test time is less than or equal to that listed.

**** Reject if test time is strictly less than (TBD) with six failures.

4.4.4.3 Test length. Testing shall continue until the total unit hours together with the total count of relevant equipment failures permit either an accept or reject decision in accordance with the specified test plan. Only equipment "ON" time may be used in MTBF or longevity determinations. Testing shall be monitored in such a manner that the times to failure may be estimated with reasonable accuracy. No single equipment "ON" time shall be less than one-half the average operating time of all equipments "ON" test.

4.4.4.4 Test conditions.

4.4.4.4.1 Operating mode (duty cycle).

4.4.4.4.2 Test level. While under test, the equipment shall be subjected to the conditions specified by Test Level B, MIL-STD-781B.

4.4.4.4.3 Input voltage cycling. When so directed by the procuring activity, voltage cycling shall be accomplished as follows: The input voltage shall be maintained at one hundred ten percent (110%) nominal voltage for one-third of the equipment "ON" cycle, at the nominal value for the second one-third of the equipment "ON" cycle, and at ninety percent (90%) for the final one-third of the equipment "ON" cycle. This cycling procedure is to be repeated continuously throughout the reliability test.

4.4.4.5 Reliability examination and test method. The following inspections shall be used to verify equipment operation during reliability testing. The satisfactory operation test of 4.3.1 shall be used to confirm proper equipment operation prior to and following reliability testing. The satisfactory operation test shall also be performed weekly to verify equipment operation. The satisfactory operation check of 4.3.1 shall be used to monitor proper operation of the equipment daily.

4.4.4.6 Reliability failure actions, Group F. In the event of failure(s) during reliability testing, failure action shall be taken as required in MIL-STD-781B.

4.4.4.7 Corrective action, Group F. In the event the reliability test reaches a reject decision, corrective action shall be taken as required in MIL-STD-781B.

4.4.5 Optional quality assurance system. The procuring activity may substitute tests from a Department of Defense (DoD) approved supplier's quality assurance system for any or all tests in TABLE I.

5. PACKAGING

5.1 Preservation, packaging. Packaging for delivery shall be in accordance with MIL-T-28800 and as specified by the procuring activity.

6. NOTES

6.1 Intended use. The equipment is intended for use in maintaining electronic communications equipment by equipment repair units in intermediate and depot maintenance.

6.1.1 Equipment replaced. Equipment procured according to this specification is intended to replace all versions of various older equipments now fielded. These equipments to be replaced are listed below: (TBD)

6.2 Ordering data. Procurement documents should specify the following:

- a. Title, number and date of this specification and any amendment thereto
- b. Packaging requirements (see Section 5)
- c. When rough handling and functional tests are required
- d. Place of final inspection
- e. Technical literature required
- f. Quantity of tools and running spare parts required
- g. Marking for shipment and shipping containers
- h. Test plans and test reports
- i. Classification of inspection and number of samples required
- j. Rack mounting requirements
- k. Alternate power requirements
- l. Maintainability rationale
- m. Reliability rationale

6.3 Contract data requirements.

- a. Nomenclature assignment
- b. National stock number
- c. Equipment sample test plans
- d. Pretest performance records
- e. First production test data
- f. Identification plate drawing

6.4 Definitions.

6.4.1 Level A performance tests. Level A testing (TABLE II) is a reduced amount of testing which is performed on each equipment produced. Its purpose is to insure that all functions and modes of operation of the equipment are evaluated without extensively checking each parameter as required in the Level B test (TABLE III). The approved equipment test procedure shall specify the actual amount of testing to be performed.

6.4.2 Reliability rationale. Reliability rationale submitted with the bid should provide clear and concise rationale showing how the reliability of the equipment complies with minimum requirements of the solicitation. This may include data from previous reliability tests, reliability predictions and other data available to the offeror. In the absence of such data on the equipment being offered, such data on similar equipment of equal or greater complexity produced by the offeror may be submitted. However, such data must be clearly identified as comparative data, and accompanied with specifications and technical literature on the similar equipment. The acceptance of the submitted rationale does not relieve the successful bidder of performing and successfully completing the production reliability testing.

6.4.3 Maintainability rationale. Maintainability rationale submitted with the bid should provide clear and concise rationale showing how the maintainability of the equipment complies with requirements of the solicitation. This may include data from previous maintainability tests, records of repair and calibration data, and other data on the equipment available to the offeror. Data on selection of module size and other design features relevant to maintainability may also be submitted.

Custodian:

Army - CR

Preparing Activity:

Army - CR

Review Activities:

Army -

(Project 6625-

User Activities:

Army -

**OMB Approval
No. 22-R255**

DOCUMENT IDENTIFIER AND TITLE**CONTRACT NUMBER**

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO (If "Yes", in what way?)

4. REMARKS

SUBMITTED BY (Printed or typed name and address - Optional)

TELEPHONE NO.

DATE _____

DD FORM 1426
1 JAN 73

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

C18081

FOLO

Department of the Army
Headquarters
U.S. Army Electronics Command
Fort Monmouth, New Jersey 07703

POSTAGE AND FEES PAID

OFFICIAL BUSINESS

Commanding General
U.S. Army Electronics Command
ATTN: DRSEL-RD-TS
Fort Monmouth, New Jersey 07703

FOLO

To detach this form, cut along this line