

AD-A085 130

INDIANA UNIV AT BLOOMINGTON DEPT OF CHEMISTRY

F/6 7/5

CORRECTION OF QUENCHING ERRORS IN ANALYTICAL FLUORIMETRY THROUG--ETC(U)

MAY 80 G M HIEFTJE, G R HAUGEN

N00014-76-C-0838

UNCLASSIFIED

TR-25

NL

1-1

2-1

3-1

4-1

5-1

6-1

7-1

8-1

9-1

10-1

11-1

12-1

13-1

14-1

15-1

16-1

17-1

18-1

19-1

20-1

21-1

22-1

23-1

24-1

25-1

26-1

27-1

28-1

29-1

30-1

31-1

32-1

33-1

34-1

35-1

36-1

37-1

38-1

39-1

40-1

41-1

42-1

43-1

44-1

45-1

46-1

47-1

48-1

49-1

50-1

51-1

52-1

53-1

54-1

55-1

56-1

57-1

58-1

59-1

60-1

61-1

62-1

63-1

64-1

65-1

66-1

67-1

68-1

69-1

70-1

71-1

72-1

73-1

74-1

75-1

76-1

77-1

78-1

79-1

80-1

81-1

82-1

83-1

84-1

85-1

86-1

87-1

88-1

89-1

90-1

91-1

92-1

93-1

94-1

95-1

96-1

97-1

98-1

99-1

100-1


END

DATE

FILED

7 80

DTIC



UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

REPORT DOCUMENTATION PAGE		READ INSTRUCTIONS BEFORE COMPLETING FORM
1. REPORT NUMBER TWENTY-FIVE	2. GOVT ACCESSION NO. ADA085130	3. RECIPIENT'S CATALOG NUMBER
4. TITLE (and Subtitle) Correction of Quenching Errors in Analytical Fluorimetry Through Use of Time Resolution.		5. TYPE OF REPORT & PERIOD COVERED Technical Report
7. AUTHOR(s) Gary M. Hieftje Gilbert R. Haugen		6. PERFORMING ORG. REPORT NUMBER 52
8. PERFORMING ORGANIZATION NAME AND ADDRESS Department of Chemistry Indiana University Bloomington, IN 47405		9. CONTRACT OR GRANT NUMBER(s) N-14-76-C-0838
11. CONTROLLING OFFICE NAME AND ADDRESS Office of Naval Research Washington, D.C.		10. PHS-GM-24473
14. MONITORING AGENCY NAME & ADDRESS (if different from Controlling Office) LEVEL III		NR 051-622
16. DISTRIBUTION STATEMENT (of this Report) Approved for public release; distribution unlimited		12. NUMBER OF PAGES 21
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report)		13. SECURITY CLASS. (of this Report) UNCLASSIFIED
18. SUPPLEMENTARY NOTES Prepared for publication in ANALYTICAL CHEMISTRY		14. DECLASSIFICATION/DOWNGRADING SCHEDULE
19. KEY WORDS (Continue on reverse side if necessary and identify by block number)		
20. ABSTRACT (Continue on reverse side if necessary and identify by block number) Varying concentrations of quenching agents can cause serious errors in analytical fluorimetry. The origin of these errors is an unexpected change in the quantum efficiency for the observed luminescence. In this paper, it is recalled that quantum efficiency can be expressed as the ratio of an observed luminescence decay time to the decay time which would be observed in the absence of quenchers. Because this latter quantity is a constant for any particular fluorophore, quantum efficiency variations can be compensated through measurement of the decay time. For such measurements, the		

ADA085130

DDC FILE COPY.

DD FORM 1 JAN 73 1473

EDITION OF 1 NOV 65 IS OBSOLETE
S/N 0102-014-6601

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

176685

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)

20. (Abstract - continued)

time-correlated single photon technique was employed and measured luminescence values were taken both from averaged photon count rates and from integrated fluorescence decay plots. Division of these values by measured luminescence lifetimes produced values which were independent of quencher concentration. Systems studied were quinine bisulfate quenched with chloride ion and 1-pyrenebutyric acid quenched by iodide.

Accession For	
NTIS GAKI	☒
DDC TAB	☐
Unpublished	☐
Justification	
By _____	
Date _____	
As _____	
Dist _____	
A	

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)

OFFICE OF NAVAL RESEARCH

Contract N14-76-C-0838

Task No. NF 051-622

TECHNICAL REPORT NO. ~~24~~ *25*

CORRECTION OF QUENCHING ERRORS IN ANALYTICAL
FLUORIMETRY THROUGH USE OF TIME RESOLUTION

by

Gary M. Hieftje and Gilbert R. Haugen

Prepared for Publication

in

ANALYTICAL CHEMISTRY

Indiana University

Department of Chemistry

Bloomington, Indiana 47405

May 27, 1980

Reproduction in whole or in part is permitted for
any purpose of the United States Government

Approved for Public Release; Distribution Unlimited

80 6 3 070

ABSTRACT

~~~~~

Varying concentrations of quenching agents can cause serious errors in analytical fluorimetry. The origin of these errors is an unexpected change in the quantum efficiency for the observed luminescence. In this paper, it is recalled that quantum efficiency can be expressed as the ratio of an observed luminescence decay time to the decay time which would be observed in the absence of quenchers. Because this latter quantity is a constant for any particular fluorophore, quantum efficiency variations can be compensated through measurement of the decay time. For such measurements, the time-correlated single photon technique was employed and measured luminescence values were taken both from averaged photon count rates and from integrated fluorescence decay plots. Division of these values by measured luminescence lifetimes produced values which were independent of quencher concentration. Systems studied were quinine bisulfate quenched with chloride ion and 1-pyrenebutyric acid quenched by iodide.

# INTRODUCTION

~~~~~

Fluorimetry provides one of the most sensitive means for the detection of many species of clinical, environmental, or forensic interest. However, fluorimetric analysis, unlike absorption methods, can suffer interferences from foreign species which quench excited analyte molecules. Although a number of mechanisms for such quenching action exist, they produce the same result--a reduced quantum efficiency and a consequent loss in fluorescence intensity. Because the concentration of the quenching agent is ordinarily unknown, correction for quenching is difficult or impossible; consequently, strong measures are taken to exclude quenchers from the analytical medium or to control carefully their concentration.

In the present paper, an alternative approach to overcoming quenching errors is explored. In this approach, quantum efficiency is indirectly measured by monitoring a sample's luminescence lifetime; correction for quenching then simply involves division of the measured fluorescence intensity by the observed lifetime.

Fundamentally, the new approach is quite straightforward. It is well known that the fluorescence intensity from the sample (F) is related to sample concentration (c) through the quantum efficiency (Φ) and a proportionality constant K.

$$F = K\Phi c. \quad (1)$$

In turn, quantum efficiency is simply the ratio of the intrinsic decay rate of an excited state (k_f) to the sum of the decay rates of all processes which depopulate the state (Σk):

$$\Phi = k_f / \Sigma k \quad (2)$$

Conveniently, the demoninator in equation 2 is simply the reciprocal of the decay time for the sample under examination (τ_m), whereas the numerator is a reciprocal of the intrinsic lifetime of the fluorophore (τ_F):

$$\Phi = \tau_m / \tau_F \quad (3)$$

Combining equations 1 and 3 yields

$$F / \tau_m = k'c \quad (4)$$

where k' contains both K from equation 1 and τ_F .

From equation 4, it should be possible to obtain a quantity which is independent of sample quantum efficiency simply by dividing a measured fluorescence signal (F) by the measured luminescence lifetime (τ_m). In the present investigation, these measurements were carried out for two different fluorophore/quencher combinations using a time-correlated single photon technique for measurement of both luminescence lifetimes and fluorescence intensities. Although the precision of the resulting corrected values was limited to approximately 4% by errors in the lifetime measurement procedure, it was found possible to correct for quenching effects in samples whose fluorescence had been quenched by 95%. Although the procedure is not able to overcome errors caused by all kinds of quenching processes, it is expected to be important in situations involving diffusional quenching and changes in intersystem crossing rates.

EXPERIMENTAL

~~~~~

Two different instrumental configurations were employed for both the measurement of fluorescence intensities and luminescence decay curves. These configurations will be described separately.

System 1. For the lifetime measurements, a commercial time-correlated single photon fluorimeter (ORTEC model 9200), modified for high data acquisition rates ( 1 ) was employed. The pulsed excitation source was run in air at 20 psig at a frequency of 25 KHz. For excitation wavelength selection, an interference filter peaked at 334 nm and having a bandwidth of 11 nm was used; emission was selected by a cutoff filter (Corning #0-52). The photomultiplier (RCA 8850) was uncooled and operated at a voltage of 2700 V.

Fluorescence intensities were measured using a computer-compatible system described previously (2). An excitation wavelength of 340 nm and emission wavelengths 375 and 395 nm were employed. Analog data collection was accomplished by chopping of the exciting radiation (4 KHz) and tuned, lock-in amplification. A 4-pole Bessel active filter with a bandwidth of 0.2 Hz smoothed the output of the lock-in amplifier. The S-20 photomultiplier (Hamamatsu 65TUV) was operated at 2060 volts.

System 1 was utilized for examining the correction of 1-pyrenebutyric acid quenched by potassium iodide (3).

System 2. A second instrumental array was employed in an attempt to extract all necessary data for corrected luminescence measurements from a single device. Unfortunately, such capability does not exist in conventional time-correlated single photon instruments because of large fluctuations in source intensity. Attempts were made to overcome this limitation through

counting of lamp pulses during the elapsed measurement interval and integrating the resulting luminescence decay curve. However, precision of these measurements was far poorer than could be obtained from a conventional luminescence spectrometer.

To overcome these difficulties, the high-pressure air-discharge lamp was replaced by a much stabler source consisting of a synchronously pumped, cavity-dumped, frequency-doubled dye laser. The laser was a commercial unit (models 171, 375, and 344, Spectra Physics, Mountain View, CA) operated with rhodamine 6-G as the dye. A pump wavelength of 514.5 nm and dye emission wavelength of 617 nm were utilized, resulting in an excitation pulse train at 308.5 nm. The frequency doubling crystal (Cleveland Crystal Co.) was KDP. This arrangement resulted in a primary beam having a power of approximately 17 mW and consisting of a train of 20 ps pulses at a frequency of 0.8 MHz. These primary pulses are calculated to have a peak power of approximately 1 KW and result in the production of frequency-doubled pulses of approximately 2 W peak power.

To serve as a trigger for the TAC of the correlation fluorimeter, a PIN photodiode registered the arrival of a fraction of the primary beam pulses. The primary and doubled beams were separated using a simple prism arrangement. The only other change from system 1, described above, is the use of a different cut-off filter (Schott #KV389) for separating the fluorescence radiation from background. A similar arrangement has been described previously (1).

With this stable pulsed source, it became possible to measure fluorescence intensity directly from recorded decay curves. Two approaches for this mea-

surement were examined. The first approach involved simply measuring the average anode pulse rate from the photomultiplier detector over a 10-second time period. The second approach utilized the integral of the processed luminescence decay curve as an indication of fluorescence intensity. This integral was readily obtained and could be directly read out from the commercial multichannel pulse-height analyzer employed in these studies.

The second system was employed to study the quenching of quinine bisulfate by chloride ion.

Reagents and chemicals. Both quenching systems have been discussed in some detail in the literature and solutions were prepared in accordance with instructions in that publication (3).

1-pyrenebutyric acid stock solutions were prepared in 0.01 M KOH and combined with varying amounts of iodide ion (as KI) in Tris buffer solutions (pH 8.0). KI was stabilized by the addition of  $\text{Na}_2\text{S}_2\text{O}_3$ . Solutions were made to have a range of lifetimes between 18 and 115 ns, the latter solution being totally unquenched. No attempt was made to exclude oxygen from the sample solution, since the technique should also correct for intersystem crossing rates induced by the presence of dissolved  $\text{O}_2$ .

Quinine bisulfate solutions were prepared in 0.1 N  $\text{H}_2\text{SO}_4$  and were combined with varying amounts of chloride ion (added as NaCl) to yield a group of solutions having expected lifetimes between 1.5 and 18.9 ns.

## RESULTS AND DISCUSSION

~~~~~

In general, results were similar for the studies on 1-pyrenebutyric acid and those on quinine bisulfate. However, instrumental system 2 seemed

to perform better than system 1 and was employed for all subsequent investigations. Consequently, we will concentrate on results obtained with this more elegant (but more complex) arrangement.

It was verified initially that quenching in both systems occurred by diffusion-controlled processes; Stern-Volmer plots for quinine bisulfate are shown in Figure 1. In Figure 1 are shown the results for both the anode pulse rate (curve a) and pulse-height spectrum integral (curve b) methods of monitoring fluorescence intensity. Examination of these curves indicates the slightly higher precision of the latter approach, an unsurprising result in view of the larger number of data points utilized by the technique.

The influence of quencher (chloride ion) concentration on quinine bisulfate luminescence intensity is shown in Figures 2a and 2b. Again, these two parts correspond to the different methods for monitoring fluorescence intensity. The lower curve in each figure reflects the drastic change in fluorescence which would be noted if no correction is made. In contrast, the upper, nearly flat line shows the improvement which is obtained by dividing the measured fluorescence intensity for each sample by its measured luminescence lifetime. All curves are least-squares fit to recorded data. For the lifetime-corrected values, relative standard deviations of 5.9% and 4.0% were obtained for, respectively, the anode pulse rate and pulse height integral techniques for monitoring fluorescence intensity. Essentially all this error can be ascribed to uncertainties in the measurement of luminescence decay times.

One of the limitations in the time-correlated single photon technique for the measurement of luminescence lifetimes is that data acquisition times

are long, being limited by the necessity to avoid pulse pileup in the detection system. Although there are ways to minimize this limitation (1,4), there is often a tendency to push data collection limits to improve acquisition time. Consequently, it was decided to examine the influence of unusually high anode pulse rates on the ability to correct for quantum efficiency changes.

The results of these investigations are shown in Figures 3a and 3b. Although the Stern-Volmer plots were linear for both methods of monitoring fluorescence intensity, even at high anode pulse rates, errors in measuring accurate luminescence lifetimes increased dramatically. At high anode pulse rates, fluorescence decay curves appear skewed toward shorter times; unless anode rates are matched exactly for samples having different degrees of quenching, decay curves yield erroneous lifetime values. Moreover, anode pulse overlap would result in a loss of integrated counts in the pulse height spectrum and in the measurement of anode pulse rate, thereby yielding incorrect values for fluorescence intensity. In the present experiment, anode pulse rates were measured with a high-frequency counter, so error in those values was minimized. In contrast, the pulse height spectrum integral was obtained directly from the time-correlated single photon counting apparatus, which had a relatively long dead time. Consequently, data scatter in the corrected values obtained with this latter technique was much greater (relative standard deviation = 7.9% in Figure 3b).

In summary, it has been shown that errors ordinarily caused by quantum efficiency variations in analytical fluorimetry can be overcome through use of time resolution. To perform the correction, one need only monitor simultaneously or consecutively the fluorescence intensity and luminescence

lifetime for each sample. Ratioing these two values then provides a measure of concentration which is independent of collisional quenching processes. Conveniently, both intensity and lifetime can be monitored using the time-correlated single photon technique, if a source having stable pulse amplitudes is employed. However, for conventional instruments, it will be necessary to utilize an independent method for monitoring fluorescence signals.

In this correction procedure, as in most others using the time correlated method, errors in measured lifetimes arise if fluorescence photons are detected at excessively high frequencies. These errors have greatest effect in the present measurements if fluorescence intensity is monitored through use of the pulse height spectrum integral; their effect is minimized if the anode pulse rate is measured directly.

ACKNOWLEDGEMENT

~~~~~

The authors wish to thank Esther E. Vogelstein for helpful comments and suggestions made during the preparation of this manuscript. One of the authors (GMH) wishes to acknowledge the support in part by the Office of Naval Research and by the National Institutes of Health through grant PHS GM 24473. The other author (GRH) wishes to acknowledge the support of the U.S. Energy Research and Development Administration under contract No. W-7405-ENG-118.

## LITERATURE CITED


~~~~~

1. G. R. Haugen, B. W. Wallin, and F. E. Lytle, *Rev. Sci. Instrum.* 50, 65 (1979).
2. G. R. Haugen, B. A. Rabi, and L. P. Rigdon, *Chem. Instrum.* 6(3), 205 (1975).
3. R. F. Chen, *Analyt. Biochem.* 57, 593-604 (1974).
4. D. W. Peterson and J. M. Hayes, in Contemporary Topics in Analytical and Clinical Chemistry, vol. 3, D. M. Hercules, G. M. Hieftje, L. R. Snyder, and M. A. Evenson, eds., Plenum Press, N.Y., 1978, ch. 5.


Figure Captions

- Figure 1.** Stern-Volmer plots for the quenching of quinine bisulfate by chloride ion. Curve A - fluorescence intensity monitored through average anode pulse rate. B - fluorescence intensity obtained by integration of luminescence decay curve (pulse-height spectrum).
- Figure 2.** Correction of measured fluorescence by ratioing with observed luminescence lifetime. Bottom curve in each figure is directly monitored fluorescence; upper plot reveals corrected values. Part A employs anode pulse rate to indicate fluorescence intensity; part B utilizes integral of pulse height spectrum.
- Figure 3.** Effect of high frequency of detected fluorescence photons on validity of quenching correction method. All parameters similar to those in Figure 2.


QUENCHING OF QUININE BISULFATE BY CHLORIDE ION




CORRECTION FOR DIFFUSIONAL QUENCHING OF QUININE




CORRECTION FOR DIFFUSIONAL QUENCHING OF QUININE



EFFECT OF HIGH ANODE RATE ON QUENCHING CORRECTION



EFFECT OF HIGH ANODE RATE ON QUENCHING CORRECTION



TECHNICAL REPORT DISTRIBUTION LIST, GEN

	<u>No. Copies</u>	
Office of Naval Research Attn: Code 472 800 North Quincy Street Arlington, Virginia 22217	2	U.S. Army Research Office Attn: CRD-AA-IP P.O. Box 1211 Research Triangle Park, N.C. 27709
ONR Branch Office Attn: Dr. George Sandoz 336 S. Clark Street Chicago, Illinois 60605	1	Naval Ocean Systems Center Attn: Mr. Joe McCartney San Diego, California 92152
ONR Branch Office Attn: Scientific Dept. 715 Broadway New York, New York 10003	1	Naval Weapons Center Attn: Dr. A. B. Amster, Chemistry Division China Lake, California 93555
ONR Branch Office 1030 East Green Street Pasadena, California 91106	1	Naval Civil Engineering Laboratory Attn: Dr. R. W. Drisko Port Hueneme, California 93401
ONR Branch Office Attn: Dr. L. H. Peebles Building 114, Section D 666 Summer Street Boston, Massachusetts 02210	1	Department of Physics & Chemistry Naval Postgraduate School Monterey, California 93940
Director, Naval Research Laboratory Attn: Code 6100 Washington, D.C. 20390	1	Dr. A. L. Slafkosky Scientific Advisor Commandant of the Marine Corps (Code RD-1) Washington, D.C. 20380
The Assistant Secretary of the Navy (R,E&S) Department of the Navy Room 4E736, Pentagon Washington, D.C. 20350	1	Office of Naval Research Attn: Dr. Richard S. Miller 800 N. Quincy Street Arlington, Virginia 22217
Commander, Naval Air Systems Command Attn: Code 310C (H. Rosenwasser) Department of the Navy Washington, D.C. 20360	1	Naval Ship Research and Development Center Attn: Dr. G. Bosmajian, Applied Chemistry Division Annapolis, Maryland 21401
Defense Documentation Center Building 5, Cameron Station Alexandria, Virginia 22314	12	Naval Ocean Systems Center Attn: Dr. S. Yamamoto, Marine Sciences Division San Diego, California 91232
Dr. Fred Saalfeld Chemistry Division Naval Research Laboratory Washington, D.C. 20375	1	Mr. John Boyle Materials Branch Naval Ship Engineering Center Philadelphia, Pennsylvania 19112

TECHNICAL REPORT DISTRIBUTION LIST, 051C

	<u>No. Copies</u>		<u>No. Copies</u>
Dr. M. B. Denton Department of Chemistry University of Arizona Tucson, Arizona 85721	1	Dr. John Duffin United States Naval Postgraduate School Monterey, California 93940	1
Dr. R. A. Osteryoung Department of Chemistry State University of New York at Buffalo Buffalo, New York 14214	1	Dr. G. M. Hieftje Department of Chemistry Indiana University Bloomington, Indiana 47401	1
Dr. B. R. Kowalski Department of Chemistry University of Washington Seattle, Washington 98105	1	Dr. Victor L. Rehn Naval Weapons Center Code 3813 China Lake, California 93555	1
Dr. S. P. Perone Department of Chemistry Purdue University Lafayette, Indiana 47907	1	Dr. Christie G. Enke Michigan State University Department of Chemistry East Lansing, Michigan 48824	1
Dr. D. L. Venezky Naval Research Laboratory Code 6130 Washington, D.C. 20375	1	Dr. Kent Eisentraut, MBT Air Force Materials Laboratory Wright-Patterson AFB, Ohio 45433	1
Dr. H. Freiser Department of Chemistry University of Arizona Tucson, Arizona 85721	1	Walter G. Cox, Code 3632 Naval Underwater Systems Center Building 148 Newport, Rhode Island 02840	1
Dr. Fred Saalfeld Naval Research Laboratory Code 6110 Washington, D.C. 20375	1	Dr. Rudolph J. Marcus Office of Naval Research Scientific Liaison Group American Embassy APO San Francisco 96503	
Dr. H. Chernoff Department of Mathematics Massachusetts Institute of Technology Cambridge, Massachusetts 02139	1	Mr. James Kelley DTNSRDC Code 2803 Annapolis, Maryland 21402	
Dr. K. Wilson Department of Chemistry University of California, San Diego La Jolla, California	1		
Dr. A. Zirino Naval Undersea Center San Diego, California 92132	1		