



NATIONAL BUREAU OF STANDARDS MICROCOPY RESOLUTION TEST CHART .

AD-A156 197



NITROGEN ANALYSIS BY CHEMILUMINESCENCE

Richard C. Striebich, 1Lt, USAF Fuels Branch Fuels and Lubrication Division

April 1985

Final Report for Period August 1982 - December 1983



Approved for public release, distribution unlimited.

310

AERO PROPULSION LABORATORY AIR FORCE WRIGHT AERONAUTICAL LABORATORIES AIR FORCE SYSTEMS COMMAND WRIGHT-PATTERSON AIR FORCE BASE, OHIO 45433-6563



1 K

NOTICE

When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not be be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

This report has been reviewed by the Office of Public Affairs (ASD/PA) and is releasable to the National Technical Information Service (NTIS). At NTIS, it will be available to the general public, including foreign nations.

This technical report has been reviewed and is approved for publication.

Richard C Strictich

RICHARD C. STRIEBICH, 1Lt, USAF Project Engineer Fuels Branch Fuels and Lubrication Division Aero Propulsion Laboratory

hur V. a.

ARTHUR V. CHURCHILL, Chief Fuels Branch Fuels and Lubrication Division Aero Propulsion Laboratory

FOR THE COMMANDER

ROBERT D. SHERRILL, Chief Fuels and Lubrication Division Aero Propulsion Laboratory

"If your address has changed, if you wish to be removed from our mailing list, or if the addressee is no longer employed by your organization, please notify AFWAL/POSF, Wright-Patterson AFB, Ohio 45433-6563 to help us maintain a current mailing list".

Copies of this report should not be returned unless return is required by security considerations, contractual obligations, or notice on a specific document.

REPORT DOCUMENTATIO	N PAGE	BE	READ INSTRUCTIONS FORE COMPLETING FORM
REPORT NUMBER	2. GOVT ACCESSION NO.	3. RECIPI	ENT'S CATALOG NUMBER
AFWAL-TR-84-2089	AD-41.56 19	7	
TITI E (and Subtitle)		5. TYPE	F REPORT & PERIOD COVERED
NITROCEN ANALYSIS BY CHEMILIMINE	SCENCE	Final August	1982 December 1983
NITRUGEN ANALISIS DI CHEMILOMINE.	JUENUE	August	1982 - December 1983
		6. PERFO	RMING ORG. REPORT NUMBER
		B. CONTR	ACT OR GRANT NUMBER(3)
AUTHOR(#)			
1Lt Richard C. Striebich			
PERFORMING ORGANIZATION NAME AND ADDRE	SS	10. PROG AREA	RAM ELEMENT, PROJECT, TASK & WORK UNIT NUMBERS
Aero Propulsion Laboratory (AFWA	L/POSF)	P.E.	62203F Project: 3048
AF Wright Aeronautical Laborator		lask	304805
Wright-Patterson AFB UH 45435 -0		WORK	BT DATE
I. CONTROLLING OFFICE NAME AND ADDRESS Apro Propulsion Laboratory (AFWA)	(POSE)	Apri	1 1985
AF Wright Aeronautical Laborator:	ies	13. NUME	ER OF PAGES
Wright-Patterson AFB OH 45433-6	563		26
. MONITORING AGENCY NAME & ADDRESS(If diffe	rent from Controlling Office)	15. SECU	RITY CLASS. (of this report)
		Uncla	assified
		150 DEC	LASSIFICATION DOWN GRADING
		SCH	ÊDULE
DISTRIBUTION STATEMENT (of this Report)			
			ACCESSION FOF
7. DISTRIBUTION STATEMENT (of the abstract enter	red in Block 20, if different fr	om Report)	NTIS GRAAI DTIC TAB
7. DISTRIBUTION STATEMENT (of the ebstract enter	red in Block 20, 11 different fr	rom Report)	NTIS GRAAI DTIC TAB Unannounced Justification
7. DISTRIBUTION STATEMENT (of the obstract enter	red in Block 20, if different fr	om Report)	NTIS GRAAI DTIC TAB
7. DISTRIBUTION STATEMENT (of the obstract enter 8. SUPPLEMENTARY NOTES	red in Block 20, if different fr	om Report)	NTIS GRAAI Image: Constraint of the second
7. DISTRIBUTION STATEMENT (of the obstract enter B. SUPPLEMENTARY NOTES	red in Block 20, il different fr	om Report)	NTIS GRA&I Image: Constraint of the second
7. DISTRIBUTION STATEMENT (of the obstract enter 8. SUPPLEMENTARY NOTES	red in Block 20, if different fr	om Report)	NTIS GRA&I DTIC TAB Unannounced Justification By Distribution/ Availability Codes
17. DISTRIBUTION STATEMENT (of the obstract enter 18. SUPPLEMENTARY NOTES	red in Block 20, il dillerent fr	om Report)	NTIS GRA&I DTIC TAB Unannounced Justification By Distribution/ Availability Codes [Avail and/or
7. DISTRIBUTION STATEMENT (of the obstract enter 18. SUPPLEMENTARY NOTES 9. KEY WORDS (Continue on reverse side if necessar	red in Block 20, if different in	r)	NTIS GRA&I DTIC TAB Unannounced Justification By Distribution/ Availability Codes Avail and/or Dist
 DISTRIBUTION STATEMENT (of the obstract enter SUPPLEMENTARY NOTES KEY WORDS (Continue on reverse side if necessar) Chemiluminescence Nitrogon analysis 	red in Block 20, if different fr	r) 0 T I.;	NTIS GRA&I DTIC TAB Unannounced Justification By
7. DISTRIBUTION STATEMENT (of the obstract enter 8. SUPPLEMENTARY NOTES 9. KEY WORDS (Continue on reverse side if necessary Chemiluminescence Nitrogen analysis Shale JP-4	red in Block 20, if different in	r) 0 T I .: C T Y	NTIS GRA&I DTIC TAB Unannounced Justification By Distribution/ Availability Codes Avail and/or Dist Special
DISTRIBUTION STATEMENT (of the obstract enter S. SUPPLEMENTARY NOTES Chemiluminescence Vitrogen analysis Shale JP-4 Model mixtures	red in Block 20, if different in	r) 0 T	NTIS GRA&I DTIC TAB Unannounced Justification By Distribution/ Availability Codes Avail and/or Dist Special Available
DISTRIBUTION STATEMENT (of the obstract enter B. SUPPLEMENTARY NOTES Chemiluminescence Vitrogen analysis Shale JP-4 Model mixtures	red in Block 20, if different in	r) 0 T 10 T 10 T 10 T	NTIS GRA&I DTIC TAB Unannounced Justification By
 DISTRIBUTION STATEMENT (of the obstract enter SUPPLEMENTARY NOTES KEY WORDS (Continue on reverse side if necessar) Chemiluminescence Vitrogen analysis Shale JP-4 Model mixtures ABSTRACT (Continue on reverse side if necessar) 	red in Block 20, if different in y and identify by block numbe , and identify by block numbe	r) 0 T .; 0 T .; 1 1	NTIS GRA&I DTIC TAB Unannounced Justification By
 DISTRIBUTION STATEMENT (of the obstract enter SUPPLEMENTARY NOTES Supplementary notes Key words (Continue on reverse side if necessary Chemiluminescence Nitrogen analysis Shale JP-4 Model mixtures ABSTRACT (Continue on reverse side if necessary With high overall accuracy, a che ppm nitrogen levels in model mixt Forty model mixtures were analyzed detector to demonstrate the absol A severely hydrotreated shale JP- mixtures to obtain six fuels rang parts-per-million added nitrogen. 	red in Block 20, if different in y and identify by block number emiluminescent det tures and shale-de ed using an Antek lute accuracy and -4 jet fuel was bl ging from less tha . Total nitrogen	r) ector c rived j Model 7 precisi ended w n l to analysi	NTIS GRA&I DTIC TAB Unannounced Justification By
 DISTRIBUTION STATEMENT (of the obstract enter SUPPLEMENTARY NOTES KEY WORDS (Continue on reverse side if necessary Chemiluminescence Nitrogen analysis Shale JP-4 Model mixtures ABSTRACT (Continue on reverse side if necessary With high overall accuracy, a cheppm nitrogen levels in model mixtures ABSTRACT (Continue on reverse side if necessary With high overall accuracy, a cheppm nitrogen levels in model mixtures ABSTRACT to demonstrate the absol A severely hydrotreated shale JP- nixtures to obtain six fuels rango parts-per-million added nitrogen. 	red in Block 20, if different in y and identify by block number emiluminescent det tures and shale-de ed using an Antek lute accuracy and -4 jet fuel was bl ging from less tha . Total nitrogen	ector corrived ju model 77 precisio ended wint to analysio	NTIS GRA&I DTIC TAB Unannounced Justification By
 DISTRIBUTION STATEMENT (of the obstract enter Distribution statement (of the obstract enter Supplementary notes Supplementary notes Supplementary notes Continue on reverse side if necessary Chemiluminescence Nitrogen analysis Shale JP-4 Model mixtures ABSTRACT (Continue on reverse side if necessary Nith high overall accuracy, a che opm nitrogen levels in model mixt Forty model mixtures were analyzed detector to demonstrate the absol A severely hydrotreated shale JP- nixtures to obtain six fuels rango parts-per-million added nitrogen FORM 1473 EDITION OF 1 NOV 65 IS OF 	red in Block 20, if different in y end identify by block number emiluminescent det tures and shale-de ed using an Antek lute accuracy and -4 jet fuel was bl ging from less tha . Total nitrogen BSOLETE UNCLA	ector C rived j Model 7 precisi ended w n 1 to analysi SSIFIED	NTIS GRA&I DTIC TAB Unannounced Justification By Distribution/ Availability Codes Avail and/or Dist Special Avail and Avail a

UNCLASSIFIED SECURITY CLASSIFICATION OF THIS PAGE(When Date Entered) 20. Abstract (Continued) by two different chemiluminescent detection systems, as well as by the microkjeldahl and microcombustion methods. Results clearly point out the ability of the chemiluminescent method to measure the nitrogen level of low ppm shale-derived aviation turbine fuel, UNCLASSIFIED SECURITY CLASSIFICATION OF THIS PAGE(When Date Entered)

5

FOREWORD

This technical report describes an experimental study of the possible limitations of a new instrumental technique for the determination of organically bound nitrogen in liquid samples by chemiluminescence. The work reported here was performed in-house under Work Unit 30480591, "Fuel Evaluation and Development" which is administered by the Fuels Branch (POSF), Fuels and Lubrication Division (POS), Aero Propulsion Laboratory, Air Force Wright Aeronautical Laboratories, Wright-Patterson Air Force Base, Chio.

The work described herein was performed by Lt Richard Striebich from September 1982 to December 1983. Special acknowledgement is extended to members of the Fuels Branch (POSF) as well as the Fuels Technical Support Branch (POFF) for their guidance and encouragement.

TABLE OF CONTENTS

SECTIO	N	PAGE
I	INTRODUCTION	1
II	CURRENT RESEARCH	3
	1. EXPERIMENTAL	3
	a. Theory	3
	b. Instrumentation	4
	c. Operating Conditions	10
	d. Operating Procedure	12
	e. Experimental Design	14
	2. DATA AND CALCULATIONS	17
	3. RESULTS	17
	a. Model Mixture Studies	17
	b. Shale Fuel Additions	17
III	DISCUSSION	23 .
	1. MODEL MIXTURE STUDY	23
	a. Linearity of Response with Concentration	23
	b. Non-dependence on Compound Type	23
	c. General Observations	23
	2. STANDARD ADDITIONS TO SHALE FUEL	25
	3. COMPARISON OF CHEMILUMINESCENCE TO OTHER METHODS OF NITROGEN ANALYSIS	25
IV	RECOMMENDATIONS	27
	REFERENCES	28

LIST OF ILLUSTRATIONS

	•	
FIGUR	Ε	PAGE
1	Activity for Deposit Formation vs. Nitrogen Concentrations in Fuel	2
2	Conversion of Nitrogen to Nitric Oxide	3
3	Chemiluminescent Detection System	5
4	Chemiluminescent Detector	6
5	Calibration Curve for Inject Rate Rotometer	7
6	Calibration Curve for Carrier Gas Rotometer	8
7	Calibration Curve for Oxygen Rotometer	9
8	Calibration Curve for Constant Rate Injector	11
9	Syringe Loading Technique	13
10	Mean Calibration Curve for ANTEK Nitrogen Analyzer	19
11	Mean Calibration Curve for ANTEK Nitrogen Analyzer (low ppm)	20
12	Known vs Measured Nitrogen Addition for Shale Fuel Samples	22
13	Repeatability Study for Pyrrole in Toluene	24
14	Repeatability Study: Low ppm Nitrogen	24

LIST OF TABLES

TABL	E	PAGE
1	Solutes, Solvents and Concentration used in Model Mixture Studies	15
2	Standard Additions to Shale JP-4	16
3	Results of Model Mixture Experiments	18
4	Known vs. Measured Values for Shale Fuel	21
5	Nitrogen Analyses of Shale Fuels	26

vii

LIST OF ABBREVIATIONS

ASTM	American Society for Testing and Materials
CRI	Constant Rate Injector
С	Degrees Celsius
DMA	Dimethylaniline
DMQ	2,7-Dimethylquinoline
1EP	1-Ethylpiperidine
F	Degrees Fahrenheit
MV-sec	Millivolts per second
N	Nitrogen
NO	Nitric Oxide
ng	nanogram (10 ⁻⁹ grams)
nm	nanometers (10 ⁻⁹ meters)
PM	Photomultiplier
ppm	parts-per-million
PYD	Pyridine
μ Έ	microliters

SECTION I

INTRODUCTION

Nitrogen compounds are detrimental to the thermal and storage stability of aviation turbine fuel. Nitrogen-bearing deposits are formed readily when a fuel which contains heteroatoms such as nitrogen is thermally stressed (Reference 1). In work completed for the Aero Propulsion Laboratory in 1967, Taylor describes increasing rates of deposit formation with increasing nitrogen concentration (Reference 2) (Figure 1). Since deposits tend to plug combustor nozzles and reduce heat flow in an aircraft engine fuel system, it would be advantageous to monitor the amount of organically bound nitrogen in a jet fuel.

Since total nitrogen content is greater in future non-petroleum fuel sources such as oil shale, tar sands, and coal liquids, monitoring the nitrogen level in distillate fuels will become more crucial. In addition to concerns of thermal and storage stability, nitrogen compounds poison catalysts, making hydrogenation processing more inefficient and expensive. Naturally, these alternative sources of fuel, when combusted, produce an increase in NOx emissions unless nitrogen levels are monitored and controlled during hydrotreating.

The purpose of this report is to validate the accuracy and precision of the ANTEK Model 720 Chemiluminescent Nitrogen Detector. This instrument is designed to detect organically bound nitrogen in concentrations ranging from one part-per-million (weight/weight) to approximately 10,000 ppm nitrogen (1.0 wt. % nitrogen).

Until just recently, elemental nitrogen has been measured almost exclusively by the Kjeldahl method. In the Kjeldahl method, boiling sulfuric acid is used to convert nitrogen to ammonia which is subsequently titrated. This method is hazardous and somewhat lengthy, and can give inaccurate results for low ppm nitrogen samples. In the past few years, a number of systems have been developed to measure nitrogen at the ppm level. Microcoulmetry (ASTM 3431) and gas chromatography with nitrogen selective detector are two widely used methods. Chemiluminescence, however, is becoming more and more popular because of its wide range of applicability (0 to 10,000 ppm N), analysis time (less than five minutes per sample), and excellent accuracy and precision.



SECTION II

CURRENT RESEARCH

1. EXPERIMENTAL

a. Theory

In the chemiluminescent process, nitrogen-bearing compounds are pyrolyzed at high temperatures to form nitric oxide:

 $N-R + O_2 \longrightarrow NO + combustion products$ (1)

At these temperature conditions, the formation of NO is greatly favored regardless of the physical or chemical state of the nitrogen. Temperatures of 1000 degrees C or higher tend to give the highest conversion of organically-bound nitrogen to nitric oxide (Figure 2).



3

DATA AND CALCULATIONS

The raw data (area counts vs. nanograms nitrogen injected,) were incorporated into ne RESULTS section.

Parts-per-million (ppm) nitrogen was determined on a weight-to-weight basis:

$$ppm nitrogen = \frac{weight solute (nitrogen)}{weight of solvent} \times 10^6 = \frac{ng N}{mg solvent}$$
(4)

The detector for the nitrogen analyzer generates integration counts which are raphed against nanograms nitrogen injected to construct a calibration curve. inograms nitrogen injected were calculated by:

$$\frac{ng N}{ng \text{ solvent } x \rho \text{ solvent } x \text{ sample size } (\mu 1) = ng \text{ Nitrogen}$$
(5)

where ρ = density: $\frac{\text{mg solvent}}{\mu \text{l solvent}}$

. RESULTS

a. Model Mixture Studies

• •

Table 3 lists each solute and solvent in the model mixture study. The average umber of nanograms nitrogen injected is plotted against the average area count for ach mixture. A least squares fit provides the best straight line through the data; he slopes, intercepts, and correlation coefficients are reported. Figures 10 and 11 raphically depict these data. These figures were used as calibration curves for the econd part of the experiment.

b. Standard Addition Studies

Table 4 lists the known vs. measured values for the shale fuel additions. igure 12 describes these data graphically.

~	
TABLE	

JP-4
SHALE
10
AD"1 TIONS
STANDARD

	Nitrogen from	Nitrogen from	Nitrogen from	Nitrogen from	Nitrogen from	lotal Nitrogen added (sum of
CODE	c,/-dimetriyiquinoiine	(bpm N)	(bpm N)	(bpm N)	(N mud)	(bpm N)
JP4-1	0	0	0	0	0	0
JP4-2	1.81	2.64	1.33	1.44	1.13	8.35
JP4-3	3.62	5.28	2.66	2.88	2.26	16.70
JP4-4	18.05	13.19	13.32	14.44	11.26	70.26
JP4-5	36.10	26.38	26 .64	28.88	22.52	140.52
JP4-6	6.37	4.66	5.89	6.38	4.98	28.28

. .

	TABLE 1					
SOLUT	ES, SOLVENTS AND CON MODEL MIXTURE STU	CENTRATIONS				
SOLUTES	SOLVENT		ppm	levels	(wt/wt)	_
2,7-Dimethylquinoline (a)	Toluene	3.2, 3	31.8,	159.3,	637.2,	1593
Indole (b)	Toluene	3.1, 3	31.2,	156.1,	624.4,	1561
Pyrrole (c)	Toluene	4.7, 4	¥6.6,	232.8,	698.4,	1164
Pyridine (d)	Isooctane	2.5, 2	24.9,	124.5,	497.8,	1244.5
i-e thylpiperidine(e)	Isooctane	3.0, 2	29.5,	147.3,	589.2,	1473
N-ethylcarbazole(f)	Toluene	2.9, 2	29.4,	147.1,	588.4,	1471
Pyrrole (c)	Toluene	4.7, 4	¥6.6,	232.8,	698.4,	1164
$N_{\rm N}$ -dimethylaniline (q)	Isooctane	3.2. 3	31.9.	159.6.	638.4.	1596





15

ecause the sample was already quite dilute. The larger injection volume provided ore nanograms of nitrogen which increased the detectability of the sample. In ither low or high ppm samples, five to ten measurements were taken for each sample ith an average deviation of plus or minus 3% (relative).

e. Experimental Design

This experiment consisted of two tasks: first, to show the linearity of etector response with respect to nitrogen concentration as well as to demonstrate a on-dependence of detector response on compound type; secondly, to determine letector response for organically-bound nitrogen in a complex shale JP-4 fuel natrix.

(1) Model Mixture Studies: Though many nitrogen compounds were available, only those classes of compounds found in jet fuels were used, i.e., pyrroles, amines indoles, pyridines, and quinolines (Reference 5). A stock solution of each solute vas prepared and diluted so that five different solutions were available with ppm nitrogen values ranging from 2 to 1500 ppm N (wt/wt). Solutes, solvents, and concentrations are given in Table 1. These solutions were then analyzed at the conditions specified earlier. A least squares curve fit of the data for the 40 different model mixtures was then constructed and used as a calibration curve for further experiments.

(2) Standard Addition Studies: Instead of using isooctane or toluene solvents, shale-derived JP-4 provided a more complex matrix in the second part of the experiment. Aliquots of five of the nitrogen-bearing standards used in the nodel mixture study were blended into the shale fuel to produce four mixtures that were 9 to 150 ppm added nitrogen. A standard solution containing several nitrogen compounds in isooctane was also prepared as well as a control sample of JP-4 (without additives) (Table 2).

By obtaining a calibration curve from the model mixture experiments, the nitrogen content of the shale fuel sample could then be measured.

•







(2) Furnace Temperature: A non-feedback, manual setting controlled the power supply to the furnace heating elements. A setting of 7.5 created a furnace temperature of 1000 degrees C.

(3) Range and Sensitivity: A low and high range were available. Since the experiment concentrated on a range of 0 to 1500 ppm nitrogen, the low range was appropriate for all measurements. The high setting could be utilized for high nitrogen crudes and only slightly refined products. Sensitivity levels of 1, 10, 20, and 50 were available: concentrations of 0-80 ppm nitrogen usually required a sensitivity of 1. Approximately 80 to 800 ppm required a sensitivity of 10 and above 800 generally needed a level of 20.

(4) Injection Rate: Injection rate was probably the most crucial parameter to control. Too high an injection rate caused coking of the sample within the combustion tube. A rate of 11μ /min was normally used because this rate maximized response and was suggested by other experimenters (Reference 4).

d. Operating Procedure

Gas flow rates and rotometer settings were first established as described in Section 2. c, "Operating Conditions." Injections were made with detector control in the "on" position. Syringe injection technique had a significant impact on the accuracy and precision of a nitrogen determination.

(1) Syringe Technique: To ensure that the entire sample was transferred from the syringe to the combustion furnace, a 2 - 3 μ l pocket of solvent (toluene) was placed behind a 1 - 2 μ l pocket of sample in the syringe, as shown in Figure 9. The solvent served to "rinse" the sample out of the syringe. Since this technique was employed for all samples, (with the exception of extremely low ppm nitrogen samples) the possible effect of solvent dilution was not observed in the calibration curve. Further, since the detector response was plotted versus the number of nanograms of nitrogen detected, no dilution effect was present.

(2) Syringe Technique (low ppm N samples): If the solution to be analyzed was suspected to be low in nitrogen content (less than 20 ppm), approximately 5.7 μ l was drawn into the syringe and injected. No solvent rinse was employed



11

AFWAL-TR-84-2089

(4) Injection System: A Houston-Atlas Model 1001 constant rate injector (CRI) forced the sample out of a 10 μ l Hamilton syringe. A digital rate setting on the CRI controlled the speed of sample introduction. The calibration curve relating digital setting to volumetric flow rate for a 10 μ l syringe is included in Figure 8.

(5) Detection System: One-sixteenth inch O.D. teflon tubing and quartz connectors allowed transfer of hot pyrolysis gases from the combustion tube to the reaction chamber. The ozone generator separated and recombined oxygen molecules to form ozone at high temperatures. Ozone and combustion gases combined in a reaction chamber to give off light which was detected by a photomultiplier (PM) tube. The resulting signal was electrically transferred to an integrator and strip chart recorder. The detection system also included zero baseline adjustment, diagnostic lights, as well as different sensitivity settings (Figure 4).

(6) Gas Scrubbers: Gases from the reaction chamber were transferred to a scrubber of activated charcoal and indicating drierite. The charcoal adsorbed any nitrogen dioxide or ozone present in the gas stream after leaving the detection system. When the charcoal was completely saturated with adsorbed gases, the ozone would move into the section of the scrubber with indicating drierite. Since ozone turns indicating drierite black, the scrubber would then be ready to be recharged with fresh adsorbents.

c. Operating Conditions

Many of the operating conditions were obtained by varying conditions until maximum response (integration counts) was obtained for each injection. Other conditions were suggested by the manufacturers or their technical representatives.

(1) Gas flow rates: At the suggestion of ANTEK Instruments Inc., the oxygen flow rate was increased well beyond that which was initially suggested in the operating manual. The greater rate gave an increased response for an 80 parts-per-million nitrogen standard. Normally, the 0_2 flow rate was 790 cc/min, (60 psi pressure), the argon flow rate was 47 cc/min (20 psi pressure), and the ozone flow rate was 120 cc/min. The pyrolysis system included a means of increasing the flow of oxygen to the inlet of the combustion tube; this means was not utilized in this experiment.



.



Figure 6. Calibration Curve for Carrier Gas Rotometer

AFWAL-TR-84-2089







· .

•

ł,







5

Nitric oxide then combines with ozone accordingly:

$$NO + O_3 ----- NO_2 + O_2$$
 (2)

Nitrogen dioxide (NO_2^*) is in an excited state. As its electrons fall to ground state energy, the molecule emits a photon in the 700 - 900 nm range:

$$NO_{2}^{*} ----- NO_{2} + h_{v}$$
 (3)

The amount of light emitted with this wavelength can be quantitated with a light sensitive detector (photomultiplier tube) and is directly proportional to the amount of nitrogen in the sample.

b. Instrumentation

Diagrams of the chemiluminescent detection system and detector are shown in Figures 3 and 4. The major components of the system are described as follows:

(1) Gases: Argon and oxygen (99.999% purity) were the only two gases required for this experiment. Gas filters of molecular sieve and indicating drierite removed water in both gas lines before introduction into the detector system. One branch of the oxygen line was connected to the ozone generator while another branch fed directly into the combustion tube. Rotometers just upstream of the combustion tube controlled gas flow rates (Figures 5, 6, and 7).

(2) Furnace: Although an ANTEK Model 771 pyroreactor was purchased along with the chemiluminescent detection system, this furnace could not be used. The heating elements burned out repeatedly due to the thermal shock of cycling the elements between low and high temperatures. A Lindberg Model 55035 tube furnace with a manual temperature controller (non-feedback) was a suitable substitution. This furnace, which had a single heating zone, required insulation to maintain constant temperature throughout the length of the pyrolysis tube.

(3) Pyrolysis Tube: The quartz pyrolysis tube could withstand temperatures exceeding 1100 degrees C without fracturing. Though some pyrolysis tubes are adaptable to a "boat" inlet technique (for high-boiling liquids and solids) this tube could only be used for syringe injection.

CORRELATION COEFFICIENT
COEFFICIENT
9
9
9
*
0.3344
9.4
6.6 0.9998
2
7.9 0.9999
0
0.7 0.9996
· · ·
3
6.5 0.9998
24
5.18 0.9999
08
4.U 0.9998
23
2.12 0.9996

TABLE 3

. .





ŀ,

۲

5

TABLE 4

KNOWN VS. MEASURED VALUES FOR ADDITIONS TO SHALE FUEL

Code	Known Addition ppm	Measured Addition ppm
JP4-1	O ppm ·	2 ppm
JP4-2	8.35 ppm	8.8 ppm
JP4-3	16.7 ppm	30.2 ppm
.1P4_4	70.3 ppm	71.9 ppm
10/ 5	140.6 ppm	135.0 ppm
JP4-5		37 9 nnm
JP4-6	28.1 ppm	37.3 ppm



SECTION III

DISCUSSION

1. MODEL MIXTURE STUDY

a. Linearity of Response with Concentration

The ANTEK Model 720 Chemiluminescent Nitrogen Detector shows excellent linearity of response according to concentration. Graphical analysis for five different concentrations of each nitrogen-bearing compound produces a linear fit with correlation coefficients of between 0.9996 and 0.9999.

b. Non-dependence on Compound Type

Figures 10 and 11 clearly indicate that detector response does not differ significantly with compound type. Thus, between 2 and 1500 ppm total nitrogen, the ANTEK analyzer is extremely effective in measuring total nitrogen content of a liquid, regardless of nitrogen type.

c. General Observations

The repeatability of detector response had to be determined at the very beginning of the project. Pyrrole in toluene standards were analyzed at two different times (three weeks apart) to demonstrate that there was very little drift in detector response. Figure 13 indicates a very slight change for the concentration range of 2 - 1500 ppm N. Figure 14, an expanded section of Figure 13, shows virtually no detector response change between 0 and 400 nanograms nitrogen during the three-week period.

One of the parameters that had to be closely monitored was the ozone flow rate. No repeatable results were obtained when the ozone rate was too low (approximately 60 cc/min instead of the preferred 120 cc/min). The ozone in the system tended to corrode lines and dissolve O-rings. Experimenters must ensure that needle valves for ozone control remain free of obstructions.



Figure 13. Repeatability Study for Pyrrole in Toluene



Figure 14. Repeatability Study: Low ppm Nitrogen

In order to analyze low parts-per-million solutions, i.e., less than 5 ppm nitrogen, a larger sample size is required. This larger sample size increases the amount of nitrogen injected, producing a greater detector response. Thus, it should be possible to analyze for total nitrogen in the parts-per-billion range merely by increasing the injection volume. Subsequent experiments, however, showed that this instrument cannot measure concentrations below 1 ppm nitrogen regardless of the sample size. Further investigations should be performed to document this preliminary work.

2. STANDARD ADDITIONS TO SHALE FUEL

Although the ANTEK analyzer accurately measured the nitrogen content of model mixtures, its ability to do so for samples with a complex hydrocarbon matrix had to be demonstrated. The fuel chosen for this experiment was a severely hydrotreated shale oil from the Caribou Refinery, Woods Cross, Utah. Since the fuel was so severely hydrotreated, it did not produce any area count when injected neatly into the ANTEK analyzer. This, of course, indicates little, if any, nitrogen in the fuel. Thus, the fuel was blended with five nitrogen-containing standard solutions to bring the concentration of N to a measurable level (Table 3). The results of the experiment show that the ANTEK chemiluminescent detection system can accurately measure nitrogen, even in a complex hydrocarbon matrix such as a shale derived turbine fuel.

3. COMPARISONS OF CHEMILUMINESCENCE TO OTHER METHODS OF NITROGEN ANALYSIS

Chemiluminescence seems to be the method of choice for many because of its short analysis time and its high accuracy. In order to see how accurately this instrument predicted the nitrogen level of the shale-derived JP-4 samples, another experiment was conducted. One pint from each sample (JP4-1 through JP4-6) was sent to four different commercial laboratories for nitrogen analysis with their best method. The Perkin-Elmer 240C was used by one laboratory but was found to be ineffective at the low concentrations used in this experiment (Table 5). The microkjeldahl method was accomplished by two separate laboratories and gave fair results. Note that the results were especially good from the analysis by Laboratory A, but not especially accurate by Laboratory C. The chemiluminescent method was the method used by Laboratory B; this instrument is a more recent model than the instrument described in this report. As can be seen, the results from this Monsanto instrument compare very favorably to those obtained on the Air Force unit.

TABLE 5

NITROGEN ANALYSES OF SHALE FUELS

	Known	Additions	0	70.3	16.7	28.1	140.6	8.35	
-	Antek ^d	bpm	2	71.9	30.2	37.9	135.0	8.8	
	Microkjeldahl, ^C	mqq	ۍ	17	5	8	155	21	
L		(Avg)	1	ı	ı	100	ı	ı	
n-Elm∉ 240,c	Antek, ^U 240, ^C ppm	(2)	400	0	0	00L	001	0	
dahl, ^a Antek, ^b Perkin 2		(E)	0	100	100	100	0	200	
		(Avg)	-	69	27	40	136	20	
		(2)	-	69	27	39	134	19	
		Ξ	-	68	27	40	137	20	
	ppm	(Avg)	32	88	27	44	121	59	
okjel		[2]	3	84	25	41	116	60	
Micro		(\mathbf{i})	22	92	28	47	125	57	
		Sample	(JP4-1) 83-POSF-1572	(JP4-4) 83-P0SF-1573	(JP4-3) 83-POSF-1574	(JP4-6) 83-POSF-1575*	(JP4-5) 83-POSF-1576	(JP4-2) 83-POSF-1577	

^aResults from Laboratory A.

^bResults from Laboratory B.

^CResults from Laboratory C.

d_{Results} from WPAFB.

*Iso-octane (2,2,4-trimethypentane) used as solvent.

SECTION IV

RECOMMENDATIONS

It is highly recommended that the Air Force use a chemiluminescent detector to measure the nitrogen level in aviation turbine fuels. If levels of nitrogen are below 1500 ppm, the chemiluminescent method is the best available in terms of accuracy, precision, and ease of operation. Greater concentrations of nitrogen can also be measured using this instrument by diluting a sample in solvent. One of the reasons for increased interest in this instrument has been the Air Force's Shale Fuel Validation Program. The purpose of this program has been to develop jet fuel resources in an operational environment. Since the shale crude is high in organically bound nitrogen and nitrogen is a suspected cause of thermal instability, it is certainly necessary to monitor this parameter during production as well as at the base level. Using the ANTEK nitrogen analyzer as a quality assurance instrument for this program is highly recommended. L-TR-84-2089

REFERENCES

Charles E. Baker, David A. Bittker, Stephen M. Cohen, and Gary T. Seng, "Research on Aviation Fuel Instability," National Aeronautics and Space Administration, Lewis Research Center, NASA Technical Memorandum 83420, Oct 83.

W. F. Taylor, "The Study of Hydrocarbon Fuel Vapor Phase Deposits," Exxon Research and Engineering Company, Linden, NJ, Quarterly Progress Report No. 7, November 1967 - February 1968.

Antek Instruments Operation/Service Manual, 1980.

Harry V. Drushel, "Determination of Nitrogen in Petroleum Fractions by Combustion with Chemiluminescent Detection of Nitric Oxide," Analytical Chemistry, 49, No. 7, 932-938, June 1977.

J. S. Warner, T. H. Danison, J. S. McNulty, <u>Analytical Techniques for Aromatic</u> <u>Components in Aircraft Fuels</u>, Battelle Laboratories for Aero Propulsion Laboratory, Wright-Patterson Air Force Base, Ohio 45433, AFWAL-TR-82-2015, 1982.

Monsanto Technical Operating Report, T2035-106T, 5 March 1984.



المعليمة بعارفان مراجة مراجة مراجع أحراجه أحراجه مراجع مراجع أرام