

# Power Measurement System for 1 mW at 1 GHz

Fred R. Clague

**NOTICE:**

This material may be protected  
by copyright law.  
(Title 17 U.S. Code)

Electromagnetic Fields Division  
Center for Electronics and Electrical Engineering  
National Engineering Laboratory  
National Institute of Standards and Technology  
Boulder, Colorado 80303-3328




U.S. DEPARTMENT OF COMMERCE, Robert A. Mosbacher, Secretary  
NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY, John W. Lyons, Director  
Issued November 1990

**DISTRIBUTION STATEMENT A**

Approved for public release;  
Distribution Unlimited

19960911 025

# DISCLAIMER NOTICE



**THIS DOCUMENT IS BEST QUALITY AVAILABLE. THE COPY FURNISHED TO DTIC CONTAINED A SIGNIFICANT NUMBER OF PAGES WHICH DO NOT REPRODUCE LEGIBLY.**

National Institute of Standards and Technology Technical Note 1345  
Natl. Inst. Stand. Technol., Tech. Note 1345, 32 pages (November 1990)  
CODEN:NTNOEF

U.S. GOVERNMENT PRINTING OFFICE  
WASHINGTON: 1990

---

For sale by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402-9325

## CONTENTS

	Page
ABSTRACT .....	v
1. INTRODUCTION .....	1
2. OPERATION .....	2
2.1 Initial Steps .....	2
2.2 Measurement .....	2
3. SYSTEM DESCRIPTION .....	6
3.1 Theory of Operation .....	6
3.2 Hardware .....	8
3.3 Software .....	8
4. ERROR ANALYSIS .....	9
4.1 Systematic Error Components .....	9
4.1.1 Voltmeter Uncertainty .....	9
4.1.2 Uncertainty in Thermistor Mount Effective Efficiency .....	11
4.1.3 Mismatch Uncertainty .....	11
4.1.4 Dual Element Error .....	12
4.2 Random Error .....	14
5. REFERENCES .....	15
APPENDIX A Instrument Specifications .....	16
APPENDIX B Software Listing .....	17

## ABSTRACT

An automated measurement system designed to measure power accurately at the level of 1 mW and at the frequency of 1 GHz is described. The system consists of commercial IEEE Std-488 bus-controlled instruments, a computer controller, and software. The results of a series of measurements are output to the computer display and, optionally, to a printer. The results are the mean of the measurement series and an estimate of the systematic and random uncertainty. The total estimated uncertainty for the average of six consecutive measurements of a nominal 1 mW, 1 GHz source is typically less than 1 percent. The system can measure any power from 0.1 to 10 mW at any microwave frequency by making appropriate changes to the software and possibly, the hardware.

Key words: automated measurement; microwave; microwave power measurement; power; power measurement; power measurement system.

# POWER MEASUREMENT SYSTEM FOR 1 mW at 1 GHz

Fred R. Clague

National Institute of Standards and Technology  
Boulder, Colorado 80303

## 1. INTRODUCTION

This system is especially designed to accurately measure microwave power at the level of 1 mW and the frequency of 1 GHz. Specifically, it supports the calibration of the Wavetek 8502A<sup>1</sup> pulse power meter, which has a 1 mW, 1 GHz calibrator output port. The manufacturer's specification on the power level of that output is  $\pm 1.5$  percent. Use of the system is not restricted to this specific application; relatively simple modifications to the software would make it possible to measure other power levels and frequencies.

The microwave power measurement method is based on the dc substitution technique. The system is implemented using a commercial version of the NIST-developed Type IV microwave power meter, a commercial coaxial thermistor mount, a digital voltmeter, and a dedicated computer controller. The Type IV power meter is not direct reading; the substituted dc power is calculated using readings obtained from the digital voltmeter. The computer controls the measurement process, calculates the results, and prints them out. The measurement results include an estimate of uncertainty for each data set. The automation also allows the implementation of a procedure that adequately corrects for thermistor mount drift caused by external temperature changes. The system is packaged in a combination operating/shipping case.

---

<sup>1</sup> Certain commercial instruments and software products are identified in this document in order to adequately specify the instrument supported and the measurement system. Such identification does not imply recommendation or endorsement by NIST nor does it imply that the identified items are necessarily the best available for the purpose.

## 2. OPERATION

### 2.1 Initial Steps

Before turning on the Type IV power meter be certain that the thermistor mount is connected to it. The output of the Wavetek 8502A calibrator is found to be more stable after a 2 hour warmup, rather than the 30 minutes specified by the manual. If possible, the 2 hour warmup period is recommended for both the 8502A and the power measurement system. It is also recommended that the thermistor mount be attached to the calibrator output for at least 30 minutes before making the measurement. This will minimize the temperature drift of the mount, improving the measurement accuracy.

Before turning on the computer, load the disk marked "System and Program" in the drive, then turn on the power. The operating system will be automatically loaded. The computer screen will display the time and the several soft-key options: SET CLOCK, LOAD PROGRAM, and EXIT. (The soft keys, or function keys, are the set of eight dark grey keys along the top of the keyboard labeled F1 through F8.) Set the time if needed, and then press the LOAD PROGRAM soft key. The measurement program will be loaded and run.

### 2.2 Measurement

The first screen displayed by the program is shown in figure 2.1. To see instructions on how to operate the 8502A (to turn the calibrator output on and off), press F1. To enter the serial number of the 8502A being measured, press F2; the serial number will then be printed with the measurement result. To change the number of repeated measurements to be averaged in a set (at least 6 to 10 is recommended), press F3. To begin the measurement set, press F4. To exit the program, press F5.




Figure 2.1. Screen display of the measurement menu.

Figure 2.2 shows the screen that appears when the first item is selected from the Measurement Menu. It gives brief instructions for manually controlling the 8502A calibrator output based on information given in the instrument's operating manual. The four numbered steps shown on the screen should be carried out before proceeding with the measurement.




Figure 2.2. Screen display of operating instructions for the calibrator output.

Figure 2.3 shows the screen that appears when F4 is pressed to start the measurement. Just before the message TURN RF ON (PRESS 8502A KEY '7') is displayed, the computer will beep once. At that point press key 7 on the 8502A to turn the rf on and wait for a pair of beeps from the computer. The message will change to TURN RF OFF (PRESS 8502A 'CLEAR'). After pressing the CLEAR key, wait until a single beep sounds again, before pressing key 7 to begin the next measurement in the set. This sequence will be automatically repeated until all the measurements making up the set have been made.




Figure 2.3. Screen display while the measurement is made.



When the desired number of measurements is complete, the final screen that is displayed is shown in figure 2.4.

PWR_MTR1		MEASUREMENT COMPLETE				
MOUNT: HP 8478B, S/N 2106A20054		16:59:46 16 Aug 1990				
POWER METER: WAVETEK MODEL 8502A, S/N 1813641						
NO.	POWER (mW)	PWR - 1.0mW (%)	V1 (VOLT)	Delta V (mV)	V1 Drift (uV/s)	Ref. Offset (mV)
1	1.004607	+00.461	2.296978	43.701	-2.1	-7.348
2	1.005146	+00.515	2.296942	43.726	-2.1	-7.348
3	1.005618	+00.562	2.296909	43.747	-2.2	-7.347
4	1.005635	+00.564	2.296878	43.748	-2.5	-7.347
5	1.005678	+00.568	2.296842	43.751	-2.7	-7.345
6	1.005731	+00.573	2.296804	43.754	-3.1	-7.345
RESULTS:						
	AVG PWR (mW)	AVG - 1.0mW (%)	MAX DEV (%)	STD DEV (%)	SYS UNC (%)	TOT UNC (%)
	1.005493	+00.540	+0.033, -0.079	0.044	0.843	0.897
		User:1 Caps Command				
		MENU PS-UNIT [ ] [ ] [ ] [ ] [ ] [ ] [ ] [ ] *				

Figure 2.4. Screen display of the measurement results.

The upper part of the display summarizes each measurement in the set as explained in table 2.1 below.

Table 2.1. Explanation of the upper part of the measurement screen	
Column Heading	Explanation
NO.	Number of the power measurement.
POWER	Result of the power measurement in milliwatts.
PWR - 1 mW	Percent deviation of the measured power from 1 milliwatt.
V1	Power meter voltage with the rf off (see section 3.1).
Delta V	Change that occurs in the power meter voltage when the rf is turned on.
V1 Drift	Drift of $V_1$ in $\mu\text{V/s}$ that occurred from the beginning of the measurement until it was complete. Note that if the drift is greater than $10 \mu\text{V/s}$ the measurement should be repeated after waiting a period of time for the mount temperature to further stabilize.
Ref. Offset	The compensation element channel is used as the voltage reference; this column shows the voltage difference between the measurement thermistor channel and the compensation thermistor channel when the rf is off.

The final results are displayed on the screen below the horizontal dashed line. The explanation of each column is given in the following table.

Table 2.2. Explanation of the results section of the measurement screen	
Column Heading	Explanation
AVG PWR	Average power in milliwatts computed from the measured data set.
AVG - 1mW	Percent deviation of the average power level from 1 milliwatt.
MAX DEV	The maximum positive and negative deviations from 1 milliwatt.
STD DEV	The standard deviation of the individual measurements.
SYS UNC	The total calculated systematic uncertainty in the measurement.
TOT UNC	Total uncertainty; the systematic uncertainty plus three times the standard deviation of the mean.

### 3. SYSTEM DESCRIPTION

#### 3.1 Theory of Operation

The NIST Type IV power meter is not a direct reading instrument. An external precision dc voltmeter must be connected to the power meter, and the power is calculated from the voltmeter readings. The power,  $P$ , is given by

$$P = \frac{1}{R_0} (V_1^2 - V_2^2), \quad (3.1)$$

where  $V_1$  is the output voltage without rf power,  $V_2$  is the voltage with rf power, and  $R_0$  is the operating resistance of the mount. Note that the so-called "bolometric power" is simply the change of the mount dc bias power as rf power is applied and removed.

It can be seen from eq (3.1) above that, as the rf power becomes small,  $V_2$  approaches  $V_1$ . Because of the uncertainty "magnification" that occurs in the computed difference of two nearly equal numbers, the power measurement uncertainty becomes very large as the power decreases. The solution to this problem is to measure the difference between  $V_1$  and  $V_2$  directly. This requires a reference voltage generator (RVG) which is set nominally equal to  $V_1$  and, in effect, stores  $V_1$ .

When an RVG is used, the expression for calculating power from measured voltages becomes,

$$P = \frac{1}{R_0} (2V_1 - \Delta V) \Delta V, \quad (3.2)$$

where  $R_0$  and  $V_1$  were previously defined, and  $\Delta V$  is the change in the power meter voltage when rf is applied. In providing for a first-order correction of mount drift, the value of  $V_1$  and  $\Delta V$  are estimated by assuming linear drift and measuring several other voltages while the rf is off, as shown in figure 3.1.

The diagram in figure 3.1 depicts the outputs of the power meter and RVG as a function of time while the rf is cycled on and off. The measurement sequence of five voltage and time readings used to calculate the power and correct for the mount drift is also shown. Note that the reference voltage generator is not set equal to  $V_{1i}$ , nor is it constant with time. This is because it is convenient to use the compensation element of the mount, biased by




Figure 3.1. Measured power meter voltages vs time.

the second power meter channel, as the reference voltage generator. Thus the RVG does drift during the measurement, but this change is also corrected, to first order, by the measurement series.

In terms of the measured voltages, the values to be used in eq (3.2) are given by,

$$V_1 = V_{1i} + \left( \frac{t_3 - t_1}{t_5 - t_1} \right) (V_{1f} - V_{1i}) \quad (3.3)$$

and,

$$\Delta V = V_{2X} - \left[ V_{1Xi} + \left( \frac{t_3 - t_2}{t_4 - t_2} \right) (V_{1Xf} - V_{1Xi}) \right] \quad (3.4)$$

### 3.2 Hardware

The system block diagram is shown in figure 3.2. The input switching to the digital voltmeter (DVM) is done with the multiplexer internal to the DVM. The dual power meter also has an IEEE Std-488 bus interface with controlled output switching, but it is not used in this application. The specifications for the instruments are given in appendix A.




Figure 3.2. System block diagram.

### 3.3 Software

A software listing is included as appendix B. Comments at the beginning of the code define the variables (and their location) that one might want to change for other applications such as a different power level or a new mount calibration factor.

## 4. ERROR ANALYSIS

### 4.1 Systematic Error Components

The factors contributing to the total systematic uncertainty are:

1. Uncertainty in the dc voltage measurements.
2. Uncertainty in the thermistor mount effective efficiency calibration.
3. Mismatch uncertainty due to the source (8502A calibrator output) reflection coefficient and the thermistor mount reflection coefficient.
4. The "dual element substitution error" associated with the coaxial thermistor mount.
5. Type IV power meter uncertainty. There are four sources of possible error internal to the power meter. They are, the reference resistors, the operational amplifier open loop gain, input offset voltage, and input bias current. The Type IV error analysis [1] indicates that all of them are negligible compared to the four factors listed above.

The first four of these items will be considered individually in the following sections.

#### 4.1.1 Voltmeter Uncertainty

The effect of uncertainty in the individual voltmeter readings can be determined by taking the total differential of the expression for power, eq (3.2),

$$dP = \frac{2}{R_0} [\Delta V dV_1 + (V_1 - \Delta V) d\Delta V], \quad (4.1)$$

where, in terms of the measured parameters,

$$dV_1 = (1 + T_{1f}) \delta V_{1i} + T_{1f} \delta V_{1f}, \quad (4.2)$$

$$d\Delta V = \delta V_{2x} + (1 + T_{2f}) \delta V_{1xi} + T_{2f} \delta V_{1xf}, \quad (4.3)$$

$$T_{1f} = \frac{t_3 - t_1}{t_5 - t_1}, \quad (4.4)$$

and,

$$T_{2f} = \frac{t_3 - t_2}{t_4 - t_2} \quad (4.5)$$

The quantities  $\delta V_{1i}$ ,  $\delta V_{1f}$ ,  $\delta V_{1Xi}$ ,  $\delta V_{1Xf}$ , and  $\delta V_{2X}$ , are the uncertainties in the measured values of  $V_{1i}$ ,  $V_{1f}$ ,  $V_{1Xi}$ ,  $V_{1Xf}$ , and  $V_{2X}$ . These uncertainties in the measured voltages are based on the voltmeter specifications, which are usually given in two parts as a fraction of reading term,  $\alpha$ , and a fraction of full scale term,  $\beta$ . The general expression for the voltmeter uncertainty is given by,

$$\delta V = \alpha V_{\text{reading}} + \beta V_{\text{fullscale}} \quad (4.6)$$

Figure 4.1 shows the uncertainty in power measurement as a function of power level near 1 mW, as calculated using the above procedure (in the calculations, the sign of the independent terms are chosen to give the maximum contribution to the total uncertainty) for the voltmeter, power meter, and measurement configuration used in this system.




Figure 4.1. Power measurement uncertainty from the DVM.

#### 4.1.2 Uncertainty in Thermistor Mount Effective Efficiency

This is the uncertainty of the NIST thermistor mount calibration. The NIST calibration also gives a value for the mount calibration factor  $C_f$ , which is the factor used in this measurement rather than effective efficiency alone, and is defined later in this section. The values listed on the report of calibration will, of course, be constant for any given mount, until the unit is recalibrated.

The thermistor mount should be recalibrated periodically.

#### 4.1.3 Mismatch Uncertainty

The net power delivered to a termination by a source is given by,

$$P_t = P_0 \frac{1 - |\Gamma_t|^2}{|1 - \Gamma_g \Gamma_t|^2}, \quad (4.7)$$

where  $P_0$  is the power the source would deliver to a nonreflecting termination,  $\Gamma_g$  is the generator reflection coefficient, and  $\Gamma_t$  is the termination reflection coefficient. Ideally, the calibrator should deliver a net power of 1 mW to the power detector being calibrated, but that can only be accomplished if the complex reflection coefficients of the power detector, generator, and calibrating thermistor mount are known, which is generally not the case. Assuming, then, that the calibrator output specification is the power delivered to a nonreflecting load,  $P_0$ , the measured output is given by,

$$P_0 = \frac{P_m}{\eta_m} \frac{|1 - \Gamma_g \Gamma_m|^2}{1 - |\Gamma_m|^2}, \quad (4.8)$$

where  $P_m$  is the bolometrically measured power,  $\eta_m$  is the effective efficiency of the thermistor mount,  $\Gamma_g$  is the generator reflection coefficient, and  $\Gamma_m$  is the thermistor mount reflection coefficient. The denominator of eq (4.8) is the mount calibration factor,

$$C_f = \eta_m (1 - |\Gamma_m|^2), \quad (4.9)$$



so that eq (4.8) becomes,

$$P_0 = \frac{P_m}{C_f} |1 - \Gamma_g \Gamma_m|^2. \quad (4.10)$$

The value of  $\Gamma_m$  has been measured during the NIST calibration, but only an upper limit to the magnitude of  $\Gamma_g$  is known (from the source return loss specification). Thus, only the limits to the term involving the reflection coefficients are known,

$$(1 - |\Gamma_g| |\Gamma_m|)^2 \leq |1 - \Gamma_g \Gamma_m|^2 \leq (1 + |\Gamma_g| |\Gamma_m|)^2, \quad (4.11)$$

so that  $P_0$  is also only known within the limits,

$$\frac{P_m}{C_f} (1 - |\Gamma_g| |\Gamma_m|)^2 \leq P_0 \leq \frac{P_m}{C_f} (1 + |\Gamma_g| |\Gamma_m|)^2. \quad (4.12)$$

This uncertainty in  $P_0$  is the mismatch uncertainty and its relative value is given to first order by,

$$\pm 2 |\Gamma_g| |\Gamma_m|. \quad (4.13)$$

The return loss specification on the calibrator output is greater than 25 dB, which results in a value for  $|\Gamma_g|$  of  $\leq 0.056$ . The value of  $|\Gamma_m|$  for the thermistor mount provided is 0.019; together these give a mismatch uncertainty in  $P_0$  of  $\pm 0.21$  percent.

#### 4.1.4 Dual Element Error

The power detector is a dual-element coaxial thermistor mount. Dual-element bolometer units are nonlinear with power level as a result of a dc-rf substitution error that arises because the two elements are not identical [2]. The error is of concern in this measurement because it is being made at 1 mW,

while the NIST calibration of mount efficiency is done at 10 mW. The only way to determine the error magnitude is by direct measurement.

In this case, the method used was to connect the coax mount to one arm of a nominally equal power splitter (for this measurement, a waveguide "magic tee" in WR 90), and a single-element waveguide mount to the other arm. The ratio of the two bolometric powers was determined at 10 mW and again at a randomly selected level between 10 mW and 0.1 mW. The change in the ratios as determined at the two power levels was a measure of the dual-element error.

Figure 4.2 shows results for two identical model waveguide mounts at 9.1 GHz. The increased spread of the data as the power level decreases is typical of bolometric measurements because of the small change in dc power that occurs at low microwave power levels. The -10 dB point on the plot is approximately equal to 1 mW.




Figure 4.2. Change in the power ratio of 2 waveguide mounts vs power level.

Figure 4.3 is the result for a coax mount compared with one of the waveguide mounts. The change in ratio at the 1 mW level (-10 dB point) is about 0.035 percent. This is the uncertainty that can be expected in the effective efficiency and thus the power measurement at 1 mW, given the calibration is done at 10 mW.




Figure 4.3. Change in the power ratio of a coax mount to a waveguide mount vs power level.

## 4.2 Random Error

In section 2.2, figure 2.4 shows the measurement screen. The last three columns under the Results section show the standard deviation, the systematic uncertainty, and the total uncertainty of that measurement set. The random contribution to the total uncertainty is chosen to be three times the standard deviation of the mean.

## 5. REFERENCES

- [1] Larsen, N.T. A new self-balancing dc-substitution rf power meter. IEEE Trans. Instrum. Meas. IM-25: 343-347; 1976 December.
- [2] Engen, G.F. A dc-rf substitution error in dual-element bolometer mounts. IEEE Trans. Instrum. Meas. IM-13: 58-64; 1964 June-Sept.

## APPENDIX A

### Instrument Specifications

1. Digital voltmeter: 6½ digit resolution; 3 volt dc range with 0.0025% of reading and 0.0002% of full scale accuracy; 300 mV dc range with 0.0035% of reading and 0.0013% of full scale accuracy; IEEE Std-488 bus; optional integrated reed relay multiplexer.  
Note: meters with other dc ranges such as 100 mV, 1 volt, and 10 volts are also usable. For instance, a 6½ digit meter with 0.00034% of reading and 0.002% of full scale accuracy on the 100 mV range, 0.00024% of reading and 0.00033% of full scale accuracy on the 1 volt range, and 0.00023% of reading and 0.00016% of full scale accuracy on the 10 volt range, gives results comparable to the 3 volt-300 mV meter.
2. Multiplexer: integrated with the DVM (or separate unit); minimum 6 single-pole, single-throw contacts; maximum thermal offset of 3μV; IEEE Std-488 bus.
3. Dual NIST Type IV power meter (or two single units).
4. Coaxial thermistor mount: type N male connector; temperature compensation thermistors; dc bias power  $\approx 30\text{mW}$ ; maximum  $|\Gamma| < 0.025$ ; NIST calibration at 1 GHz.
5. Computer controller: programmable in Hewlett Packard Work Station Basic version 5.13 ("Rocky Mountain Basic"), or TransEra "HT Basic" with IEEE Std-488 capability; IEEE Std-488 bus.

# APPENDIX B

## Software Listing

```

100 File$="PWRM1"           I Started:9001111632/FRC
105 Rev$="9011210805" I FRC I NTL author of the subprograms
110 I Errors, Select_v, and Hp_3457
115 I
120 I This program application is the measurement of the 1 mW
125 I calibrator output of the Wavetek 8501A peak power meter.
130 I
135 I
140 I NOTES:
145 I
150 I This version measures V1 and delta V with the compensation element
155 I used as an RVG. It also calculates the measurement uncertainty.
160 I
165 I Total measurement uncertainty includes:
170 I Mount calibration factor uncertainty of 0.5973% (For #20054
175 I with Cal Factor of 0.9897)
180 I and calculated mismatch uncertainty for the source ( $|\Gamma| \leq 0.056$ )
185 I and the mount ( $|\Gamma| \leq 0.019$ ) of 0.21%.
190 I The total is 0.8073% plus the DVM and Type IV contribution.
195 I
200 I
205 I INSTRUMENTS CONTROLLED: ADDRESS
210 I 1. HP3457A DVM 722
215 I 2. HP2225A PRINTER 701
220 I
225 I
230 I DESCRIPTION OF THE MAIN INITIAL VALUE VARIABLES:
235 I
240 I -----
245 I The following are in the labeled common named "/Dvm/":
250 I
255 I ** "Dvm_name$" - the DVM identifier (ie, HP3457A)
260 I
265 I * "P0" - power level in milliwatts. The measurement results are
270 I compared with this value. Default setting is 1 mW.
275 I
280 I * "R0" - mount operating resistance in ohms. Normally 200 ohms
285 I for a coax mount and may be either 100 or 200 ohms for
290 I a waveguide mount. Default setting is 200 ohms.
295 I
300 I -----
305 I The following are in the labeled common named "/Mount/":
310 I
315 I * "Mount$" - bolometer mount identifier (manufacturer,
320 I model, and serial number).
325 I
330 I * "Cf" - NIST measured mount calibration factor. Default setting
335 I is 0.9897 for the supplied mount. Value must be changed
340 I after mount replacement or recalibration.
345 I
350 I -----
355 I The following are in the labeled common named "/Errs/":
360 I
365 I * "Cfu" - total quoted uncertainty of the NIST measured mount
370 I calibration factor. Default setting is 0.5973% for the
375 I supplied mount.
380 I
385 I * "Mmu" - calculated mismatch uncertainty. Default setting is
390 I 0.21% as indicated in the notes above.
395 I
400 I -----
405 I The following is in the labeled common named "/Wavetek/":
410 I
415 I "Sn$" - records the serial number of the Wavetek meter
420 I being measured. It can be input before the measure-
425 I ment from an item on the initial menu.
430 I
435 I -----
440 I
445 I CHANGING INITIAL VALUE OF VARIABLES
450 I
455 I * These variables are initially defined in the subprogram "Set_up".
460 I To change them, move to the subprogram by executing, "EDIT S".
465 I Change the values as needed and "Re-store" the program if the
470 I changes are to be permanent.
475 I
480 I
485 I ** This variable is initially defined in the subprogram "Hp_3457".
490 I If a different DVM is used, along with the name, the percent

```

```

495 I |of reading and the percent of full scale specifications must |
500 I |also be changed in that subprogram. Execute "EDIT hp3457" to |
505 I |move to the subprogram. |
510 I
515 I * * * * * MAIN PROGRAM * * * * *
520 I
525 OPTION BASE 1
530 COM /Dvm/ P0,R0,A1,A2,A3,A4,A5,B1,B2,B3,B4,B5,R1,R2,R3,R4,R5
535 COM /Dvm/ Dvm_name$(40) IDVM ID
540 COM /Errs/ Dp,V1c,V1i,V1f,V1xi,V1xf,V1x,V2x,T1fac,T2fac,Cfu,Mmu
545 COM /Mount/ Mount$(40),Cf IMount ID
550 COM /Wavetek/ Sn$(7) IFor the serial number
555 REAL P(100,1) IFor the power measurements
560 CONTROL 2,1;0 ITurn PRT ALL off
565 KEY LABELS OFF ITurn off key labels
570 I
575 CALL Set_up IFor mount & measurement parameters
580 CALL Hp_3457 IGet DVM parameters
585 CALL Init IHardware initialization
590 Nt=6 IDefault No. of meas
595 LOOP ITo repeat measurement sets
600 CALL Menu1(Nt,Quit)
605 IF Quit THEN Quit ITerminate
610 CALL Hdr IScreen header
615 REDIM P(Nt,1) IRedimension
620 FOR N=1 TO Nt IMeasurement loop
625 DISP N
630 CALL Meas(N,P1) IDo the measurement
635 P(N,1)=P1 IFill array for statistics
640 WAIT 1 IWait before measuring again
645 NEXT N
650 CALL Stats(P(*)) ICalculate the statistics of the run
655 OUTPUT 722;"TRIG AUTO" ILet DVM continue reading
660 PRINT TABXY(30,1),CHR$(128);CHR$(136);" M E A S U R E M E N T C O M P L E T E "
665 CALL Menu2 IPost measurement soft keys
670 END LOOP
675 Quit: ITerminate program
680 CLEAR SCREEN
685 END
690 I
695 I * * * * * SUB PROGRAMS * * * * *
700 I
705 M: SUB Meas(N,P1)
710 OPTION BASE 1
715 Sys_prt=VAL(SYSTEM$( "SYSTEM PRIORITY" )) IDetermine system priority
720 Lcl_prt=Sys_prt+1 ISet local priority 1 higher for ON KEY
725 ON KEY 0 LABEL " ",Lcl_prt GOTO Bail_out
730 I
735 COM /Dvm/ P0,R0,A1,A2,A3,A4,A5,B1,B2,B3,B4,B5,R1,R2,R3,R4,R5
740 COM /Dvm/ Dvm_name$(40) IDVM ID
745 COM /Errs/ Dp,V1c,V1i,V1f,V1xi,V1xf,V1x,V2x,T1fac,T2fac,Cfu,Mmu
750 COM /Mount/ Mount$(40),Cf IMount ID
755 I
760 CALL Dvm(V1i,T1i) IV1 before rf turn_on
765 OUTPUT 722;"CHAN 0" IConnect for delta V
770 WAIT .2 I
775 CALL Dvm(V1xi,T1xi) IInitial delta V1 (V1xi) with rf off
780 Vt=V1xi+V1i-SQR(V1i^2-9.E-4*R0) ICalculate threshold for Rf sub
785 CALL Rf(1,Vt) ICalls for rf ON and determines when
790 WAIT 1 IFor source to settle
795 CALL Dvm(V2x,T2x) IRead delta V2 (V2x) with rf on
800 CALL Rf(0,Vt) ICalls for rf OFF and determines when
805 WAIT 1 IWait again
810 CALL Dvm(V1xf,T1xf) IFinal delta V1 (V1xi) with rf off
815 OUTPUT 722;"CHAN 1" IReconnect for V1
820 WAIT .2
825 CALL Dvm(V1f,T1f) IFinal V1 with rf off
830 I
835 T1fac=(T2x-T1i)/(T1f-T1i) IFirst timing factor
840 V1c=V1i+T1fac*(V1f-V1i) IV1 corrections
845 I
850 T2fac=(T2x-T1xi)/(T1xf-T1xi) ISecond timing factor
855 V1x=V1xi+T2fac*(V1xf-V1xi) IDelta V corrections
860 Dv1=(V1f-V1i)*1.E+6 IChange in V1
865 Dv1_dt=Dv1/(T1f-T1i) IDrift rate of V1 in mV/sec
870 Dv2=V2x-V1x IChange in V2 - (delta V)
875 I
880 CALL Errors ICalculate errors
885 I

```

```

890 P1=1000/R0*(2*V1c-(Dv2))*(Dv2) !Power in mW
895 P1=P1/Cf !Cal factor correction
900 !
905 GOSUB Printout !Print results
910 SUBEXIT !Normal exit
915 P: !Printout
920 Printout: !
925 IMAGE 3D,5X,Z.6D,5X,S2Z.3D,8X,Z.6D,2X,3D.3D,5X,S2D.0,8X,2D.3D
930 PRINT USING 925;N,P1,100*(P1-P0)/P0,V1c,Dv2*1.E+3,Dv1_dt,V1x*1.E+3
935 RETURN !
940 !
945 Bail_out: !As it says
950 OUTPUT 722;"TRIG AUTO" !DVM continue reading
955 PRINT
960 PRINT TABXY(30,1),CHR$(128);CHR$(136);" M E A S U R E M E N T S T O P P E D "
965 PAUSE
970 !
975 Exit: !Finished
980 SUBEND ! SUB Meas
985 ! * * * * *
990 !
995 Rf:SUB Rf(On,Vt) !Turn rf ON/OFF
1000 IF On THEN
1005 DISP CHR$(129);" TURN RF ON (PRESS 8502A KEY '7') ";CHR$(128) !Tell operator
1010 BEEP 1000,.01 !Get his attention
1015 LOOP !Wait for rf to be turned on/off
1020 CALL Dvm(V,T) !Read DVM
1025 WAIT 1 !
1030 EXIT IF V>Vt !If rf is turned ON
1035 END LOOP
1040 ELSE
1045 DISP CHR$(129);" TURN RF OFF (PRESS 8502A 'CLEAR') ";CHR$(128) !Tell operator
1050 BEEP 1000,.01 !Get his attention
1055 WAIT .2
1060 BEEP 1000,.01
1065 LOOP !Wait for rf to be turned on/off
1070 CALL Dvm(V,T) !Read DVM
1075 WAIT 1 !
1080 EXIT IF V<Vt !If rf is turned OFF
1085 END LOOP
1090 END IF
1095 DISP ""
1100 SUBEND
1105 ! * * * * *
1110 !
1115 Dvm:SUB Dvm(V,T) !DVM reading
1120 SEND 7;UNL LISTEN 22 !Get dvm's attention
1125 TRIGGER 7 !trig to read
1130 ENTER 722;V !Read DVM
1135 T=TIMEDATE !Get the time
1140 SUBEND
1145 ! * * * * *
1150 !
1155 Init:SUB Init !Initialize instruments
1160 !
1165 CLEAR 722 !Clear 3457
1170 OUTPUT 722;"TERM SCANNER" !Connect input to scanner
1175 OUTPUT 722;"NPLC 10" !10 PLC
1180 OUTPUT 722;"DCV -1" !Auto Range
1185 OUTPUT 722;"TRIG AUTO" !Set up for single readings
1190 !
1195 OUTPUT 722;"CHAN 1" !Connect for V1, floating DVM
1200 WAIT 1 !Make sure everything is settled
1205 SUBEND
1210 ! * * * * *
1215 !
1220 H:SUB Hdr
1225 !
1230 OPTION BASE 1
1235 CLEAR SCREEN
1240 !
1245 COM /Dvm/ P0,R0,A1,A2,A3,A4,A5,B1,B2,B3,B4,B5,R1,R2,R3,R4,R5
1250 COM /Dvm/ Dvm_name$[40] !DVM ID
1255 COM /Mount/ Mount$[40],Cf !Mount ID
1260 COM /Wavetek/ Sn$[7] !For the serial number
1265 !
1270 PRINT TABXY(1,1),CHR$(137)&"P W R _ M T R 1"&CHR$(136)
1275 PRINT TABXY(30,1),CHR$(136);CHR$(129);" M E A S U R E M E N T I N P R O G R E S S ";CHR$(128)
1280 !

```



```

1285 PRINT TABXY(1,3),CHR$(140);"MOUNT: ";Mount$;CHR$(136)
1290 PRINT TABXY(59,3),CHR$(140);TIME$(TIMEDATE);" ";DATE$(TIMEDATE);CHR$(136)
1295 PRINT TABXY(1,4),CHR$(140);"POWER METER: WAVETEK MODEL 8502A, S/N ";Sn$;CHR$(136)
1300 I
1305 DIM A$(80),B$(80),C$(80),D$(80),Scr$(80) !String variables to build IMAGE statement
1310 Ima:DATA "#," NO,"",4X,"" POWER "",4X,""PWR-" ""
1315 Imc:DATA "#," "mw",6X,"" V1 "",3X,""Delta V"",3X,""V1 Drift"",3X,""Ref. Offset""
1320 RESTORE Ima
1325 READ Scr$ !Read as IMAGE statement
1330 OUTPUT A$ USING Scr$
1335 OUTPUT B$ USING "#,2D.0";PO
1340 RESTORE Imc
1345 READ Scr$
1350 OUTPUT C$ USING Scr$
1355 D$=A$&B$&C$
1360 PRINT D$
1365 I
1370 IMAGE "ä C",4X,"ä (mW) C",4X,"ä (% C",7X,"ä (VOLT) C",3X,"ä (mV) C",3X,"ä (uV/s) C",3X,"ä (mV) C"
1375 PRINT USING 1370
1380 I
1385 SUBEND I Hdr
1390 I*****
1395 E:SUB Errors
1400 OPTION BASE 1
1405 COM /Dvm/ PO,R0,A1,A2,A3,A4,A5,B1,B2,B3,B4,B5,R1,R2,R3,R4,R5
1410 COM /Dvm/ Dvm_name$(40) !DVM ID
1415 COM /Errs/ Dp,V1c,V1i,V1f,V1xi,V1xf,V1x,V2x,T1fac,T2fac,Cfu,Mmu
1420 I
1425 CALL Select_v(V1i,Aa1i,Bb1i,Ss1i) !Aa_ - fraction of reading error
1430 CALL Select_v(V1f,Aa1f,Bb1f,Ss1f) !
1435 CALL Select_v(V1xi,Aa1xi,Bb1xi,Ss1xi) !
1440 CALL Select_v(V1xf,Aa1xf,Bb1xf,Ss1xf) !Bb_ - fraction of FS error
1445 CALL Select_v(V2x,Aa2x,Bb2x,Ss2x) !Ss_ - fullscale reading
1450 I
1455 GOSUB With_rvg !
1460 I GOSUB Servo_errors !Very small error - not used for this application
1465 !sub routine removed
1470 Total_error:=! Without RVG.
1475 Total=Without+Err+Ierr
1480 I
1485 SUBEXIT
1490 I
1495 With_rvg: ! Eq's derived 900111/FRC
1500 Dv1i=Aa1i*V1i+Bb1i*Ss1i ! Delta-V due to initial V1 measmnt
1505 Dv1f=Aa1f*V1f+Bb1f*Ss1f ! Delta-V due to final V1 measmnt
1510 I
1515 Dv1xi=ABS(Aa1xi*V1xi)+Bb1xi*Ss1xi ! Delta due to initial V1x measmnt
1520 Dv1xf=ABS(Aa1xf*V1xf)+Bb1xf*Ss1xf ! Delta-V due to final V1x measmnt
1525 I
1530 Dv2x=ABS(Aa2x*V2x)+Bb2x*Ss2x ! Delta-V due to V2x measmnt
1535 I
1540 Dv1c=(1+T1fac)*Dv1i+T1fac*Dv1f !Error in corrected V1
1545 Dv1x=(1+T2fac)*Dv1xi+T2fac*Dv1xf !Error in delta V correction
1550 I
1555 Dpv1=ABS((V2x-V1x)*Dv1c) ! Delta-power due to V1 measmnt errors
1560 Dpv1x=ABS((V1c-V2x-V1x)*Dv1x) ! Delta-power due to V1x " "
1565 Dpv2x=ABS((V1c-V2x-V1x)*Dv2x) ! Delta-power due to V2x " "
1570 Dp=2*(Dpv1+Dpv1x+Dpv2x)/R0 ! Sum (2 & R0 left out above)
1575 Dp=Dp*1.E+3 ! Dp in mW
1580 RETURN
1585 I
1590 SUBEND
1595 I*****
1600 I
1605 Hp3457:SUB Hp_3457
1610 OPTION BASE 1
1615 COM /Dvm/ PO,R0,A1,A2,A3,A4,A5,B1,B2,B3,B4,B5,R1,R2,R3,R4,R5
1620 COM /Dvm/ Dvm_name$(40) !DVM ID
1625 Dvm_name$="HP 3457"
1630 !FOR DVM: VALUE QUANTITY (HP 3457, 1 yr, 167 ms, 6-1/2 dig)
1635 Nc: DATA 3.03E6 ! number of counts, full scale
1640 A1: DATA 4.5E-5 ! fraction-of-rdg error, range R1, 1 yr
1645 A2: DATA 3.5E-5 ! fraction-of-rdg error, range R2, etc.
1650 A3: DATA 2.5E-5 ! fraction-of-rdg error, range R3
1655 A4: DATA 4.0E-5 ! fraction-of-rdg error, range R4
1660 A5: DATA 5.5E-5 ! fraction-of-rdg error, range R5
1665 B1: DATA 385. ! fraction-of-FS error, counts, range R1, 10 PLC
1670 B2: DATA 40. ! fraction-of-FS error, counts, range R2
1675 B3: DATA 7. ! fraction-of-FS error, counts, range R3

```

```

1680 B4: DATA      20.          ! fraction-of-FS error, counts, range R4
1685 B5: DATA      7.          ! fraction-of-FS error, counts, range R5
1690 R1: DATA      0.0303      ! lowest range (including overrange), volts
1695 R2: DATA      0.303       ! next range up
1700 R3: DATA      3.03        ! next range up
1705 R4: DATA      30.3        ! next range up
1710 R5: DATA      300.        ! next range up
1715 READ Nc,A1,A2,A3,A4,A5,B1,B2,B3,B4,B5,R1,R2,R3,R4,R5
1720 !
1725 Convert_fs_errs: ! Normalize FS count errors to fractional errors
1730 B1=B1/Nc
1735 B2=B2/Nc
1740 B3=B3/Nc
1745 B4=B4/Nc
1750 B5=B5/Nc
1755 SUBEND
1760 !
1765 Select: SUB Select_v(V,Aa,Bb,Ss)
1770 OPTION BASE 1
1775 COM /Dvm/ P0,R0,A1,A2,A3,A4,A5,B1,B2,B3,B4,B5,R1,R2,R3,R4,R5
1780 COM /Dvm/ Dvm_name$[40] IDVM ID
1785 SELECT ABS(V) ! V may be of either polarity
1790 CASE <=R1 ! Start at lowest range
1795 Aa=A1 ! Fraction of rdg error for V on range R1
1800 Bb=B1 ! Fraction of FS error for V on range R1
1805 Ss=R1 ! Fullscale reading for V, range R1
1810 Range=1 ! Range_no number for plot
1815 CASE <=R2 ! Uprange if necessary
1820 Aa=A2
1825 Bb=B2
1830 Ss=R2 ! Etc. for range R2
1835 Range=2
1840 CASE <=R3 ! And again
1845 Aa=A3
1850 Bb=B3
1855 Ss=R3
1860 Range=3
1865 CASE <=R4
1870 Aa=A4
1875 Bb=B4
1880 Ss=R4
1885 Range=4
1890 CASE <=R5
1895 Aa=A5
1900 Bb=B5
1905 Ss=R5
1910 Range=5
1915 CASE ELSE
1920 BEEP
1925 PRINT "Voltage is in excess of 300 volts. Don't be ridiculous."
1930 PAUSE
1935 END SELECT
1940 SUBEND
1945 !
1950 S: SUB Set_up ! Initialize mount parameters
1955 OPTION BASE 1
1960 COM /Dvm/ P0,R0,A1,A2,A3,A4,A5,B1,B2,B3,B4,B5,R1,R2,R3,R4,R5
1965 COM /Dvm/ Dvm_name$[40] IDVM ID
1970 COM /Errs/ Op,V1c,V1i,V1f,V1xi,V1xf,V1x,V2x,T1fac,T2fac,Cfu,Mmu
1975 COM /Mount/ Mount$[40],Cf
1980 Mount$="HP 8478B, S/N 2106A20054"
1985 Cf=.9897 ! Mount calibration factor
1990 Cfu=.5973 ! Calibration factor uncertainty in %
1995 Mmu=.21 ! Mismatch factor uncertainty in %
2000 R0=200 ! Mount operating resistance in ohms
2005 P0=1.0 ! Comparison power in mW. Note that
2010 ! the following line limits this setting
2015 ! to a 0.1 mW resolution.
2020 P0=DROUND(P0,2) ! Limit P0 to 1 place beyond decimal
2025 SUBEND
2030 !
2035 Stats: SUB Stats(REAL P(*))
2040 OPTION BASE 1
2045 COM /Dvm/ P0,R0,A1,A2,A3,A4,A5,B1,B2,B3,B4,B5,R1,R2,R3,R4,R5
2050 COM /Dvm/ Dvm_name$[40] IDVM ID
2055 COM /Errs/ Op,V1c,V1i,V1f,V1xi,V1xf,V1x,V2x,T1fac,T2fac,Cfu,Mmu
2060 ALLOCATE Dum(SIZE(P,1),1) ! Use Dum(*) to preserve P(*)
2065 GOSUB Sd ! Standard dev. of original set
2070 !

```

```

2075 Sys_err=Cfu+Mmu+100*Dp/Mean !Systematic error % (See header notes)
2080 Sdm=Sd/SQR(SIZE(P,1)) !Standard Deviation of the mean
2085 Tot_unc=Sys_err+300*(Sdm/Mean) !Total uncertainty % with 3*SD mean
2090 !
2095 GOSUB Prt !Print results
2100 DEALLOCATE Dum(*)
2105 SUBEXIT
2110 !
2115 Prt:PRINT "-----"
2120 PRINT "RESULTS:"
2125 !
2130 DIM A$(128),B$(128),C$(128),D$(128),Scr$(128) !String variables to build IMAGE statement
2135 Imd:DATA "#,8X,""AVG PWR """,4X,""AVG-""
2140 Ime:DATA "#,""mw""",6X,"" MAX DEV """,3X,""STD DEV""",3X,""SYS UNC""",3X,""TOT UNC""
2145 RESTORE Imd
2150 READ Scr$ !Read as IMAGE statement
2155 OUTPUT A$ USING Scr$
2160 OUTPUT B$ USING "#,20.0";P0
2165 RESTORE Ime
2170 READ Scr$
2175 OUTPUT C$ USING Scr$
2180 D$=A$&B$&C$
2185 PRINT D$
2190 !
2195 IMAGE 8X,"ä (mw) Ç",4X,"ä (%) Ç",7X,"ä (%) Ç",3X,"ä (%) Ç",3X,"ä (%) Ç",3X,"ä (%) Ç"
2200 PRINT USING 2195
2205 IMAGE 8X,Z.60,5X,SZ.30,8X,SZ.30,K,SZ.30,4X,Z.30,5X,Z.30,5X,Z.30
2210 PRINT USING 2205;Mean,100*(Mean-P0)/P0,100*Maxpdv/Mean,"",100*Maxndv/Mean,100*Sd/Mean,Sys_err,Tot_unc
2215 RETURN
2220 !
2225 Sd:I
2230 MAT Dum= P
2235 Sum=SUM(Dum) ! Sum of the elements in P(*)
2240 Mean=Sum/SIZE(P,1) ! Mean of P(*)
2245 MAT Dum= P-(Mean) ! Dum(*) contains deviations from mean
2250 Maxpdv=MAX(Dum(*)) ! Largest positive deviation
2255 Maxndv=MIN(Dum(*)) ! Largest negative deviation
2260 Maxdv=MAX(ABS(Maxpdv),ABS(Maxndv)) ! Largest largest deviation
2265 MAT Dum= Dum . Dum ! Dum holds squares of deviations
2270 IF SIZE(P,1)>1.1 THEN ! Check for single measurement
2275 Var=SUM(Dum)/(SIZE(P,1)-1) ! Variance
2280 ELSE
2285 Var=SUM(Dum)
2290 END IF
2295 Sd=SQR(Var) ! Standard deviation
2300 Max_al=3*Sd ! Maximum allowable standard deviation
2305 RETURN
2310 !
2315 SUBEND
2320 !
2325 Menu2:SUB Menu2 !Post measurement soft keys
2330 OPTION BASE 1
2335 Sys_prt=VAL(SYSTEM$("SYSTEM PRIORITY")) !Determine system priority
2340 Lcl_prt=Sys_prt+1 !Set local priority 1 higher for ON KEY
2345 USER 1 KEYS !1st set of soft keys
2350 KEY LABELS ON !Turn on soft keys
2355 FOR N=0 TO 19 !Clear keys
2360 ON KEY N LABEL "" GOTO Top !Default destination
2365 NEXT N
2370 ON KEY 1 LABEL " MENU ",Lcl_prt GOTO Exit
2375 ON KEY 2 LABEL " PRINT ",Lcl_prt GOSUB Print
2380 !
2385 Top:LOOP !Wait for input
2390 END LOOP
2395 Print: !Alpha dump
2400 KEY LABELS OFF !Turn off soft keys
2405 DUMP ALPHA !As it says
2410 KEY LABELS ON !Turn keys back on
2415 RETURN
2420 Exit: !
2425 KEY LABELS OFF
2430 SUBEND
2435 !
2440 Menu1:SUB Menu1(Nt,Quit) !PRE measurement set up & soft keys
2445 OPTION BASE 1
2450 Sys_prt=VAL(SYSTEM$("SYSTEM PRIORITY")) !Determine system priority
2455 Lcl_prt=Sys_prt+1 !Set local priority 1 higher for ON KEY
2460 !
2465 COM /wavetek/ Sn$(7) !For the serial number

```

```

2470 M_flag=1                                !To write menu
2475 USER 1 KEYS                             !1st set of soft keys
2480 KEY LABELS ON                           !Turn on soft keys
2485 FOR N=0 TO 19                           !Clear keys
2490   ON KEY N LABEL "" GOTO Top             !Default destination
2495 NEXT N
2500 !
2505 ON KEY 1 LABEL " SELECT (1)",Lcl_prty GOSUB Help
2510 ON KEY 2 LABEL " SELECT (2)",Lcl_prty GOSUB Sn
2515 ON KEY 3 LABEL " SELECT (3)",Lcl_prty GOSUB Change
2520 ON KEY 4 LABEL " SELECT (4)",Lcl_prty GOTO Exit
2525 ON KEY 5 LABEL " SELECT (5)",Lcl_prty GOTO Quit
2530 !
2535 Top:LOOP                               !Wait for input
2540   IF M_flag=1 THEN GOSUB Menu
2545   END LOOP
2550 !
2555 Menu: CLEAR SCREEN
2560 PRINT TABXY(5,2),CHR$(129);" P W R _ M T R 1 ";CHR$(128)
2565 CLIP 10,110,24,88                       !To draw a box
2570 FRAME
2575 PRINT TABXY(24,5)," - - MEASUREMENT MENU - - "
2580 PRINT TABXY(20,8),CHR$(129);" (1)";CHR$(128);" WAVETEK OPERATING INSTRUCTIONS"
2585 PRINT TABXY(20,10),CHR$(129);" (2)";CHR$(128);" INPUT WAVETEK SERIAL NUMBER"
2590 IF Sn="" THEN
2595   PRINT TABXY(25,11),"(No S/N in memory)"
2600 ELSE
2605   PRINT TABXY(25,11),"(S/N ";Sn$;" in memory)"
2610 END IF
2615 PRINT TABXY(20,13),CHR$(129);" (3)";CHR$(128);" CHANGE # OF MEASUREMENT POINTS"
2620 PRINT TABXY(25,14),"(Present setting =";Nt;" ) "
2625 PRINT TABXY(20,16),CHR$(129);" (4)";CHR$(128);" BEGIN MEASUREMENT"
2630 PRINT TABXY(20,18),CHR$(129);" (5)";CHR$(128);" EXIT PROGRAM"
2635 M_flag=0
2640 RETURN
2645 Sn:                                     !Input the WAVETEK serial number
2650 KEY LABELS OFF                           !Turn off soft keys
2655 INPUT "WAVETEK SERIAL NUMBER ?";Sn$[1,7]
2660 Sn$=TRIM$(Sn$)
2665 PRINT TABXY(25,11),"(S/N ";Sn$;" in memory)"
2670 KEY LABELS ON                           !Turn keys back on
2675 RETURN
2680 Change:                                !Change # of meas points
2685 KEY LABELS OFF                           !Turn off soft keys
2690 INPUT "NUMBER OF MEASUREMENT POINTS ?";Nt
2695 Nt=MIN(Nt,100)
2700 Nt=MAX(Nt,1)
2705 PRINT TABXY(25,14),"(Present setting =";Nt;" ) "
2710 KEY LABELS ON                           !Turn keys back on
2715 RETURN
2720 Help:                                  !With operation of Wavetek
2725 CALL Help
2730 M_flag=1
2735 RETURN
2740 Quit:                                  !Terminate program
2745 Quit=1
2750 Exit:                                  !
2755 KEY LABELS OFF
2760 SUBEND
2765 !
2770 Help: SUB Help
2775 CLEAR SCREEN
2780 OPTION BASE 1
2785 Sys_prty=VAL(SYSTEM$("SYSTEM PRIORITY")) !Determine system priority
2790 Lcl_prty=Sys_prty+1                     !Set local priority 1 higher for ON KEY
2795 USER 1 KEYS                             !1st set of soft keys
2800 KEY LABELS ON                           !Turn on soft keys
2805 FOR N=0 TO 19                           !Clear keys
2810   ON KEY N LABEL "" GOTO Top             !Default destination
2815 NEXT N
2820 ON KEY 1 LABEL "CONTINUE",Lcl_prty GOTO Exit
2825 GOSUB Text                               !Print info
2830 !
2835 Top:LOOP                               !Wait for input
2840   END LOOP
2845 !
2850 Text: PRINT TABXY(22,2),"CONTROLLING 8502A CALIBRATOR OUTPUT"
2855 PRINT TABXY(12,4),"Press the following 8502A front panel control keys in the"
2860 PRINT TABXY(12,5),"sequence indicated:"

```

```

2865 PRINT TABXY(14,7),"(1)-'CW',"
2870 PRINT TABXY(25,7),"(2)-'Menu',"
2875 PRINT TABXY(38,7),"(3)-'F3',"
2880 PRINT TABXY(49,7),"(4)-'F1'."
2885 PRINT TABXY(12,9),"Then, pressing the 8502A key '7' will turn the calibrator"
2890 PRINT TABXY(12,10),"output ON, and pressing the 8502A key 'CLEAR' will turn the"
2895 PRINT TABXY(12,11),"calibrator output OFF."
2900 PRINT TABXY(12,13),"For more detail see 'Calibrator Output Level Test' on page 6-2"
2905 PRINT TABXY(12,14),"of the 8502A manual."
2910 PRINT TABXY(12,17),"CAUTION: Do not press any UNITS key when the mount is"
2915 PRINT TABXY(12,18),"connected to the calibrator. This will cause the"
2920 PRINT TABXY(12,19),"calibrator to output 100 mW which might damage the mount."
2925 RETURN
2930 !
2935 Exit:
2940 SUBEND

```

## BIBLIOGRAPHIC DATA SHEET

## 4. TITLE AND SUBTITLE

Power Measurement System for 1 mW at 1 GHz

## 5. AUTHOR(S)

Fred R. Clague

## 6. PERFORMING ORGANIZATION (IF JOINT OR OTHER THAN NIST, SEE INSTRUCTIONS)

U.S. DEPARTMENT OF COMMERCE  
NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY  
BOULDER, COLORADO 80303-3328

## 7. CONTRACT/GRANT NUMBER

## 8. TYPE OF REPORT AND PERIOD COVERED

## 9. SPONSORING ORGANIZATION NAME AND COMPLETE ADDRESS (STREET, CITY, STATE, ZIP)

U. S. Air Force  
Logistics Command  
Directorate of Metrology  
Newark AFB, Ohio 43057

## 10. SUPPLEMENTARY NOTES

## 11. ABSTRACT (A 200-WORD OR LESS FACTUAL SUMMARY OF MOST SIGNIFICANT INFORMATION. IF DOCUMENT INCLUDES A SIGNIFICANT BIBLIOGRAPHY OR LITERATURE SURVEY, MENTION IT HERE.)

An automated measurement system designed to measure power accurately at the level of 1 mW and at the frequency of 1 GHz is described. The system consists of commercial IEEE Std-488 bus-controlled instruments, a computer controller, and software. The results of a series of measurements are output to the computer display and, optionally, to a printer. The results are the mean of a measurement series and estimate of the systematic and random uncertainty. The total estimated uncertainty for the average of six consecutive measurements of a nominal 1 mW, 1 GHz source is typically less than 1 percent. The system can measure any power from 0.1 to 10 mW at any microwave frequency by making appropriate changes to the software and possibly, the hardware.

## 12. KEY WORDS (6 TO 12 ENTRIES; ALPHABETICAL ORDER; CAPITALIZE ONLY PROPER NAMES; AND SEPARATE KEY WORDS BY SEMICOLONS)

automated measurement; microwave; microwave power measurement; power; power measurement; power measurement system.

## 13. AVAILABILITY

☒

UNLIMITED

☐

FOR OFFICIAL DISTRIBUTION. DO NOT RELEASE TO NATIONAL TECHNICAL INFORMATION SERVICE (NTIS).

☒ORDER FROM SUPERINTENDENT OF DOCUMENTS, U.S. GOVERNMENT PRINTING OFFICE,  
WASHINGTON, DC 20402.☐

ORDER FROM NATIONAL TECHNICAL INFORMATION SERVICE (NTIS), SPRINGFIELD, VA 22161.

## 14. NUMBER OF PRINTED PAGES

32

## 15. PRICE