

## Detection of *Staphylococcus aureus* enterotoxin A and B genes with PCR-EIA and a hand-held electrochemical sensor

Mohamed Aitichou<sup>a,\*</sup>, Robert Henkens<sup>b</sup>, Afroz M. Sultana<sup>c</sup>, Robert G. Ulrich<sup>c</sup>, M. Sofi Ibrahim<sup>c</sup>

<sup>a</sup>Clinical Research Management, 1425 Porter Street, Frederick, MD 21702, USA

<sup>b</sup>Alderon Biosciences, Inc., 2810 Meridian Parkway, Suite 152, Durham, NC 27713, USA

<sup>c</sup>US Army Medical Research Institute of Infectious Diseases, 1425 Porter Street, Frederick, MD 21702, USA

Received 22 March 2004; accepted for publication 11 June 2004

### Abstract

Two electrochemical assays for detecting *Staphylococcus aureus* enterotoxin A and B genes were developed. The assays are based on PCR amplification with biotinylated primers, hybridization to a fluorescein-labeled probe, and detection with horseradish peroxidase-conjugated anti-fluorescein antibody using a hand-held electrochemical detector. The limit of detection (LOD) for both assays was approximately 16 copies of the *sea* and *seb* genes. The assays were evaluated in blinded studies, each with 81 samples that included genomic and cloned *S. aureus* DNA, and genomic DNA from *Alcaligenes*, *Bacillus*, *Bacteroides*, *Bordetella*, *Borkholderia*, *Clostridium*, *Comamonas*, *Enterobacter*, *Enterococcus*, *Escherichia*, *Francisella*, *Haemophilus*, *Klebsiella*, *Listeria*, *Moraxella*, *Neisseria*, *Proteus*, *Pseudomonas*, *Salmonella*, *Serratia*, *Shigella*, *Streptococcus*, *Vibrio* and *Yersinia* species. Both assays showed 100% sensitivity. The specificity was 96% for the SEA assay and 98% for the SEB assay. These results demonstrate the feasibility of performing probe-based detection of PCR products with a low-cost, hand-held, electrochemical detection device as a viable alternative to colorimetric enzyme-linked assays of PCR products. © 2004 Elsevier Ltd. All rights reserved.

**Keywords:** *Staphylococcus aureus*; Enterotoxins; SEA; SEB; Electrochemical PCR-EIA

### 1. Introduction

*Staphylococcus aureus* is a gram-positive bacterium that colonizes and persists in diverse host habitats. It produces 15 enterotoxins, SEA-SEE, SEG-SEO, and SEU. Based on their antigenic properties, these enterotoxins were divided into three subfamilies: the *sea* sub-family includes *sea*, *sed*, *see*, *seh* and *sej*; the *seb* sub-family includes *seb*, *sec* and *seg*; and the *sei* subfamily includes *sei*, *sek* and *se*. [1]. These and other *S. aureus* toxins are known to cause human food poisoning and toxic shock [2–4]. *S. aureus* is also considered to be a potential biological threat agent because it can be produced in large quantities, its toxins are relatively stable to chemical and physical inactivation, and small amounts of toxins (< 1 µg) can trigger symptoms of toxic shock.

Various immunoassays based on antigen capture have been developed to identify and assay enterotoxins produced by *S. aureus*. Among these assays are the reversed passive latex agglutination (RPLA), manual and automated enzyme-linked immunosorbent assays (ELISAs), and immuno-magnetic electrochemiluminescence assays [5–7]. With these assays, it was possible to detect 1 pg/ml of *S. aureus* B toxin [5] or 0.2–1.25 ng/ml of *S. aureus* A, B, C, D, E toxins [5]. The issue of cross-reactivity and limited specificity of these assays continue to be a major disadvantage and explain the need for DNA based toxin identification methods.

PCR-based assays with gel electrophoresis, dot blots, or Southern blots were developed to specifically identify genes of *S. aureus* enterotoxins A, B, C, D, E, exfoliative toxins (ETA, ETB) and TSST-1 [8–13]. While these assays are useful as research tools, they do not lend themselves for use in rapid diagnosis. PCR-based colorimetric enzyme-immuno assays (PCR-EIA) for detecting *sea* and *seb* genes have been developed. Becker et al. developed

\* Corresponding author. Tel.: +1-301-619-2415; fax: +1-301-619-2290.  
E-mail address: mohamed.aitichou@det.amedd.army.mil (M. Aitichou).

## Report Documentation Page

*Form Approved*  
OMB No. 0704-0188

Public reporting burden for the collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to a penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number.

1. REPORT DATE <b>11 JUN 2004</b>			2. REPORT TYPE <b>N/A</b>			3. DATES COVERED <b>-</b>			
4. TITLE AND SUBTITLE <b>Detection of Staphylococcus aureus enterotoxin A and B genes with PCR-EIA and a hand-held electrochemical sensor. Molecular and Cellular Probes 18:373-377</b>						5a. CONTRACT NUMBER			
						5b. GRANT NUMBER			
						5c. PROGRAM ELEMENT NUMBER			
						5d. PROJECT NUMBER			
6. AUTHOR(S) <b>Ichou, MA Henkins, R Sultana, A Ulrich, RG Ibrahim, MS</b>						5e. TASK NUMBER			
						5f. WORK UNIT NUMBER			
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) <b>United States Army Medical Research Institute of Infectious Diseases</b>						8. PERFORMING ORGANIZATION REPORT NUMBER <b>RPP-03-147</b>			
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES)						10. SPONSOR/MONITOR'S ACRONYM(S)			
						11. SPONSOR/MONITOR'S REPORT NUMBER(S)			
12. DISTRIBUTION/AVAILABILITY STATEMENT <b>Approved for public release, distribution unlimited</b>									
13. SUPPLEMENTARY NOTES									
14. ABSTRACT <b>Two electrochemical assays for detecting Staphylococcus aureus enterotoxin A and B genes were developed. The assays are based on PCR amplification with biotinylated primers, hybridization to a fluorescein-labeled probe, and detection with horseradish peroxidase (HRP)-conjugated anti-fluorescein antibody using a hand-held electrochemical detector. The limit of detection (LOD) for both assays was approximately 16 copies of the sea and seb genes. The assays were evaluated in blinded studies, each with 81 samples that included genomic and cloned S. aureus DNA, and genomic DNA from Alcaligenes, Bacillus, Bacteroides, Bordetella, Burkholderia, Clostridium, Comamonas, Enterobacter, Enterococcus, Escherichia, Francisella, Haemophilus, Klebsiella, Listeria, Moraxella, Neisseria, Proteus, Pseudomonas, Salmonella, Serratia, Shigella, Streptococcus, Vibrio and Yersinia species. Both assays showed 100% sensitivity. The specificity was 96% for the SEA assay and 98% for the SEB assay. These results demonstrate the feasibility of performing probe-based detection of PCR products with a low-cost, hand-held, electrochemical detection device as a viable alternative to colorimetric enzymelinked assays of PCR products.</b>									
15. SUBJECT TERMS <b>Staphylococcal enterotoxin, SEB, SEA, genes, methods, PCR-EIA, hand-held sensor</b>									
16. SECURITY CLASSIFICATION OF:					17. LIMITATION OF ABSTRACT		18. NUMBER OF PAGES	19a. NAME OF RESPONSIBLE PERSON	
a. REPORT <b>unclassified</b>		b. ABSTRACT <b>unclassified</b>		c. THIS PAGE <b>unclassified</b>		<b>SAR</b>		<b>5</b>	

PCR-EIA assays for the *sea* and *seb* genes and detected as little as 50 and 100 pg of genomic DNA for single and multiplex PCR, respectively [8]. Gilligan et al. developed PCR-EIA assays specific for *sea* and *seb* [14]. Both assays detected about 250 gene copies of the *sea* or *seb*. More recently, quantitative Real-Time PCR assays for the detection of enterotoxins genes *sea* to *sej* and *seu* were reported [15,16a,16b]. The detection limits varied from 100 to 400 copies for the *sea* gene [15] and 250 to  $>10^5$  copies per PCR reaction for the *sea* to *sej* genes [16b].

The aim of this study was to develop an electrochemical PCR-EIA as an alternative to a colorimetric PCR-EIA to detect the *sea* and *seb* genes of *S. aureus*.

## 2. Materials and methods

### 2.1. Bacterial strains and DNA samples

The bacterial strains and cloned DNA used in this study are listed in Table 1. The DNA from cultures and clinical specimens was isolated by standard phenol-chloroform extraction followed by ethanol precipitation [17] or by using the QIAamp DNA purification kit (Qiagen, Hilden, Germany). The final DNA pellets were suspended in 100  $\mu$ l of TE buffer (1 mM EDTA; 10 mM Tris-Cl pH 7.5), the concentration was determined with a spectrophotometer and the DNA was stored at  $-20^\circ\text{C}$  until use.

### 2.2. PCR amplification

The primers used to amplify *sea* and *seb* genes are listed in Table 2. PCR amplifications were carried out in 50  $\mu$ l volumes. Each reaction contained PCR buffer (10 mM Tris-HCl pH 9.0, 50 mM KCl, 0.1% Triton X 100), 0.5  $\mu$ M of each primer, 0.250 mM dNTPs (Roche, Indianapolis, IN), 3.125 mM MgCl<sub>2</sub> (Promega, Madison, WI), 0.0125 U/ $\mu$ l uracil N-glycosylase (Roche), 0.25 U of Taq DNA polymerase (Promega) and 5  $\mu$ l of template DNA. Thermal cycling consisted of one cycle at 95  $^\circ\text{C}$  for 2 min; 45 cycles of 95  $^\circ\text{C}$  for 15 s and 60  $^\circ\text{C}$  for 60 s; followed by one final extension cycle at 72  $^\circ\text{C}$  for 5 min on a PTC 100 thermal cycler (MJ Research, Waltham, MA).

### 2.3. Electrochemical detection

The fluorescein-labeled probes used for detection were purchased from Invitrogen (Carlsbad, CA) and are listed in Table 2. Five microlitre of PCR product were added to individual wells of a microtiter plate containing 40.5  $\mu$ l of lambda-exonuclease (0.06 U/ $\mu$ l) (Roche) and incubated at room temperature for 5 minutes to generate single-stranded PCR products. Sixty microlitre of fluorescein labeled probe (40 nM) were added, mixed, and incubated for 5 min at

Table 1  
Bacterial species and strains used for electrochemical PCR-EIA

Organism	Strain/ isolate	Organism	Strain/ isolate
<i>Alcaligenes xylosoxidans</i>	19606	<i>Pseudomonas aeruginosa</i>	17933D
<i>Bacillus anthracis</i>	Ames	<i>Salmonella choleraesuis</i>	9150D
<i>Bacillus anthracis</i>	4728	<i>Serratia odorifera</i>	33077
<i>Bacillus cereus</i>	13061	<i>Shigella flexneri</i>	12022
<i>Bacillus cereus</i>	10867	<i>Shigella sonnei</i>	9290
<i>Bacillus coagulans</i>	7050	<i>Staphylococcus aureus</i>	25923
<i>Bacillus licheniformis</i>	2759	<i>S. aureus</i>	29247
<i>Bacillus megaterium</i>	8244	Staph-SEA	RIID clinic
<i>Bacillus polymexa</i>	4525	Staph-SEB	RIID clinic
<i>Bacillus</i>	7953	Staph-SEC	RIID clinic
<i>stearothermophilus</i>			
<i>Bacillus subtilis var niger</i>	6633	Staph-SED	RIID clinic
<i>Bacillus thuringiensis</i>	35646	Staph-SEE	RIID clinic
<i>Bacteroides distasonis</i>	8503	Staph-TSST-1	RIID clinic
<i>Bordetella bronchiseptica</i>	10580	<i>S. hominis</i>	27844
<i>Burkholderia cepacia</i>	25416	<i>S. epidermidis</i>	49134
<i>Clostridium perfringens</i>	13124	<i>Stenotrophomonas maltophilia</i>	13637
<i>Comamonas acidovorans</i>	15668	<i>Streptococcus pneumoniae</i>	33400
<i>Enterobacter cloacae</i>	49141	<i>S. pyogenes</i>	19615
<i>E. aerogenes</i>	m10822	<i>Vibrio cholerae</i>	N16961
<i>Enterococcus durans</i>	6056	<i>Yersinia enterocolitica</i>	NA
<i>E. faecalis</i>	29212	<i>Y. kristensii</i>	33639
<i>Escherichia coli</i>	25922	<i>Y. pestis</i>	CO92
<i>Francisella tularensis</i>	NA	<i>Y. pestis</i>	K25
<i>Haemophilus influenzae</i>	51	<i>Y. pestis</i>	Nairobi
	907D		
<i>Klebsiella pneumoniae</i>	13883	<i>Y. pestis</i>	Java9
<i>Listeria monocytogenes</i>	15313	<i>Y. pseudotuberculosis</i>	6904
<i>Moraxella cattaharalis</i>	25240	<i>Yersinia ruckeri</i>	29473
<i>Neisseria lactamica</i>	23970	SEAWT	clone
<i>Proteus mirabilis</i>	7002	SEBWT	clone
<i>Proteus vulgaris</i>	49132		

80  $^\circ\text{C}$ , then for 5 min at room temperature. Fifty microlitre were transferred to wells of neutravidin-coated microtiter plate (Roche) and incubated at room temperature for 10 min. The plate was then washed five times with washing buffer that contained 24 mM Tris, 1 mM EDTA, 75 mM NaCl and 0.05% Tween 20. Fifty microlitre of horseradish peroxidase (HRP)-conjugated anti-fluorescence antibody (0.75 U/ $\mu$ l) were added to each well and incubated for 10 min at room temperature. The plate was washed five times. The electrochemical reaction was initiated by adding 50  $\mu$ l of 3,3',5,5'-tetramethylbenzidine (0.4 g/ $\mu$ l) and HRP peroxide (0.02%) (Pierce, Rockford, IL). The current resulting from the enzyme-substrate reactions was measured with an electrochemical reader using intermittent pulse amperometry. For comparison, additional readings at 410 nm were recorded by using an ELISA reader (Dynatech, Chantilly, VA).

Table 2  
Primers and probes sequences

Primers/ probes	Nucleotide sequence	Reference
SEAU	Biotin-5'-ATGGTAGCGAGAAAAGCGAA-3'	Gilligan et al. [14]
SEAL	PO4-5'-GCCATAAATTGATCGGCACT-3'	Gilligan et al. [14]
SEAP	Fluorescein-5'-CTAAAGCTGTCCCTG-CAATTCA-3'	Gilligan et al. [14]
SEBU	Biotin-5'-TGTATGTATGGTGGTGAAC-3'	Sharma et al. [13]
SEBL	PO4-5'-ACAAATCGTTAAAAACGGCG-3'	Gilligan et al. [14]
SEBP	Fluorescein-5'-ATAGTGACGAGTTAGGTA-3'	This work

#### 2.4. Data analysis

Student's *t*-test and regression analysis were used to evaluate statistical significance and assay linearity. To determine positivity and negativity, three negative controls were run in each assay and a threshold value was calculated from the mean electrochemical signal of the negatives controls plus two standard deviations. Sensitivity was defined as  $(TP/(TP+FN) \times 100)$  and specificity as  $(TN/(TN+FP) \times 100)$ , where TP, TN, FP and FN represents true positive, true negative, false positive, and false negative, respectively.




Fig. 1. Detection limits of SEA (A) and SEB (B) assays. Serial 10-fold dilutions of plasmid DNA ranging from 1,000,000 to 0.1 fg/ $\mu\text{l}$  were used. Each bar represents the mean of three replicates. Plasmid DNA was detected at 0.1 fg/ $\mu\text{l}$  which represents 0.5 fg per PCR reaction (about 12 copies). Regression analysis and the coefficient of correlation are indicated. Agarose gel analysis of PCR product obtained with 10-fold serial dilutions of plasmid DNA is shown below each graph.

### 3. Results

To determine the assay's limit of detection (LOD), 10-fold serial dilutions of cloned DNA ranging from 100 pg/ $\mu$ l to 0.1 fg/ $\mu$ l were made and the assays were carried out according to optimized protocols. The LOD of the SEA and SEB assay was 0.5 fg of plasmid DNA (12 copies) per PCR reaction (Fig. 1). The electrochemical signals obtained with 0.5 fg of plasmid DNA (23 replicates) were significantly higher ( $p < 0.001$ ) than those obtained with the negative controls (23 replicates) for both assays. Regression analysis of electrochemical signals obtained with the SEA and SEB assays showed a linear correlation with DNA concentrations spanning seven orders of magnitude ( $r^2 \geq 93\%$ ) indicating the utility of the assays as quantitative tests (Fig. 1).

When a 10-fold dilution series of *S. aureus* genomic DNA was used, the LOD of each assay was 10 fg/ $\mu$ l (50 fg per PCR reaction) or 16 copies of template. Each assay was tested by colorimetric EIA detection and compared with the electrochemical detection assay. Similar sensitivity levels were obtained with the colorimetric assays (data not shown). Tests with spiked DNA showed that addition of non-enterotoxigenic staphylococcal DNA or human DNA to a cloned DNA had no effect on the amplification and detection limits. This indicates the absence of PCR inhibition and interference with ELISA detection (data not shown) and that the performance of our hand held electrochemical detection system was comparable to the colorimetric PCR-EIA methods.

To determine the sensitivity and specificity of the assays, we tested DNA samples from 44 different bacterial species in a blinded study. A total of 81 samples plus 11 positive controls and four negative controls were tested. Included in this panel were DNA from *Streptococcus pyogenes* strains containing *spea* or *spec* genes that exhibit a high amount of sequence similarity to *seb* and *sea* genes, respectively. All eight samples positive for *sea* and 18 samples positive for *seb* genes were correctly identified (Table 3). Thus the sensitivity

was 100% for both assays. Of 56 DNA samples that did not contain *sea* or *seb* genes, two were considered false positive by SEA assay and one was considered false positive by the SEB assay (Table 3). Thus the specificity of the SEA and SEB assays were 96 and 98%, respectively. However, upon repetition, these samples were negative, and agarose gel analysis did not show any PCR products. These results indicate that the assays are specific for *sea* and *seb* genes.

### 4. Discussion

The objective of this study was to develop an assay that determines the presence or absence of *S. aureus* harboring enterotoxin A and B genes. We have documented in the forgoing results a highly sensitive and specific electrochemical assay that quantitates PCR products using microtitre strips with wells equipped with immunosensors. The strips were applied for testing a large number of sample DNAs with both increased sensitivity and specificity. The results were automatically acquired and no gel-based detection was required.

Identifying infectious disease and biological threat agents relies on a variety of microbiological, serological, and molecular techniques, each has advantages and disadvantages. Microbiological techniques provide reliable information on viability but they are not necessarily specific. Serological techniques can rapidly detect antigens or antibodies, but they have limited sensitivity and specificity. Molecular techniques are highly sensitive and specific, but do not provide information on viability. In a naturally occurring epidemic or a bioterrorism attack scenario, it may be necessary to use a battery of assays and platforms that when used in tandem can provide reliable information on the quantity, identity, and viability of the agent. In this report we described PCR-EIA methods for detecting *S. aureus sea* and *seb* genes using a hand-held electrochemical detector. Although the detector is used here for DNA detection, it can also be used for antigen or antibody detection using ELISA techniques.

When the assays were tested with a reference panel of 81 samples, all eight samples positive for *sea* and 18 samples positive for *seb* were correctly identified. Thus the sensitivity for both assays was 100%. Of the 56 samples that lacked *sea* or *seb* genes, two false positives (*Enterobacter aerogenes* and *Yersinia ruckeria*) were observed with the SEA assay and one false positive (*Yersinia ruckeria*) was observed with the SEB assay. These false positive results were not due to non-specific amplification or non-specific probe hybridization. Upon repetition, these samples were negative, and there was no visible PCR products by agarose gel analysis. There was 100% correlation between phenotypic and genotypic identification and the results were reproducible and statistically significant. Furthermore, there was no cross-reactivity with streptococcal enterotoxins, which are genetically related to staphylococcal enterotoxins. However, we elected to state the specificity as observed in the first run to illustrate the fact

Table 3  
Cross reactivity results of SEA and SEB electrochemical PCR-EIA

Organism	Toxin	Number of positive/total	
		PCR-EIA <i>sea</i>	PCR-EIA <i>seb</i>
<i>Staphylococcus aureus</i>	–	0/1	0/1
<i>S. aureus</i>	SEA	7/7	0/7
<i>S. aureus</i>	SEB	0/17	17/17
<i>S. aureus</i>	SEA/SEB	1/1	1/1
Staph-SEC	SEC	0/2	0/2
Staph-SED	SED	0/2	0/2
Staph-SEE	SEE	0/1	0/1
Staph-TSST-1	TSST-1	0/1	0/1
Other bacterial species <sup>a</sup>	–	2/49	1/49
Sensitivity	–	100%	100%
Specificity	–	96%	98%

<sup>a</sup> See Table 1.

that carry-over contamination is always a risk in all PCR-based assays. Therefore, it was important to repeat questionable results at least three times to verify the results.

The same underlying enzymatic reactions are used in electrochemical and colorimetric PCR-EIA assays, and the two approaches show similar sensitivity and specificity. The results reported here show that electrochemical PCR-EIA assays can reliably carry out assays previously done with colorimetric PCR-EIA. In addition, the electrochemical detection procedure can be carried out at ambient temperature with a hand-held device.

The utility of electrochemical PCR-EIA as a screening tool for analyzing organic molecules such as steroids, protein, and antibiotics was previously reported [18,19a,19b]. Although electrochemical detection scores are based on peroxidase activity, phosphatase and catalase activities measured by electrochemical sensors could also be used as an alternative. There is an amperometric immunosensor reported to detect and assay levels of *S. aureus* electrically in pure cultures and in foods [19]. The assay measures amperometrically the catalase activity that was conjugated to an anti-protein A antibody. The use of phosphatase substrate in this format has yet to be tested.

Although TaqMan<sup>®</sup> assays offer Real-Time detection, PCR-EIA assays are more affordable in some laboratories. Moreover, the electrochemical detection platform can be used for antigen or antibody detection as well as for DNA assays.

## Acknowledgements

This work was supported by DOD Biological Defense Research Program funds managed by the United States Army Medical Research and Materiel Command, Fort Detrick, Maryland. We thank Alan Schmaljohn, Brian Moore and Kathy Kenyon for reviewing the manuscript. The mention of materials or products in this article does not constitute endorsement by the Department of Defense or the United States government. The DNA samples used in this work were kindly provided by the Diagnostic Systems Division and Toxinology Division of the US Army Medical Research Institute of Infectious Diseases, Fort Detrick, Maryland.

## References

- [1] Dinges MM, Orwin PM, Schlievert PM. Enterotoxins of *Staphylococcus aureus*. Clin Microbiol Rev 2000;13:16–34.
- [2] Barg NL, Harris T. Toxin-mediated syndromes. In: Crossley KB, Archer GL, editors. The staphylococci in human disease. New York, NY: Churchill Livingstone; 1997, p. p527–544.
- [3] Fraser JD, Arcus V, Kong P, Baker E, Profit T. Superantigens-powerful modifiers of the immune system. Mol Med Today 2000;6:125–32.
- [4] Ulrich RG. Evolving superantigens of *Staphylococcus aureus*. FEMS Immunol Med Microbiol 2000;27:1–7.
- [5] Kijek TM, Rossi CA, Moss D, Parker RW, Henchal EA. Rapid and sensitive immunomagnetic-electrochemiluminescent detection of staphylococcal enterotoxin B. J Immunol Methods 2000;236:9–17.
- [6] Park CE, Akhtar M, Rayman MK. Nonspecific of a commercial enzyme-linked immunosorbent assays kit (TECRA) for detection of staphylococcal enterotoxins in foods. Appl Environ Microbiol 1992;58:2509–12.
- [7] Thomson NE, Razdan M, Kuntsmann G, Aschenbach JM, Evenson ML, Bergdoll MS. Detection of *Staphylococcus* enterotoxins by enzyme-linked immunosorbent assays and radio-immunoassays: comparison of monoclonal and polyclonal antibody systems. Appl Environ Microbiol 1986;51:885–90.
- [8] Becker K, Roth R, Peters G. Rapid and specific detection of toxigenic *Staphylococcus aureus*: use of two multiplex PCR enzyme immunoassays for amplification and hybridization of staphylococcal enterotoxin genes, exfoliative toxin genes, and toxic shock syndrome toxin-1 gene. J Clin Microbiol 1998;36:2548–53.
- [9] Jauldhac B, Bes M, Bornstein N, Piemont Y, Brun Y, Fleurette J. Synthetic DNA probes for detection of genes for enterotoxins A, B, C, D, E, and for TSST-1 in staphylococcal strains. J Appl Bacteriol 1992;72:386–92.
- [10] Johnson WM, Tyler SD, Ewan EP, Ashton FE, Pollard DR, Rozee KR. Detection of genes for enterotoxins, exfoliative toxin, and toxic shock syndrome toxin 1 in *Staphylococcus aureus* by the polymerase chain reaction. J Clin Microbiol 1991;29:426–30.
- [11] Monday SR, Bohach GA. Use of multiplex PCR to detect classical and newly described pyrogenic toxin genes in staphylococcus isolates. J Clin Microbiol 1999;37:3411–4.
- [12] Neil RJ, Fanning RG, Delahoz F, Wolff R, Gemski P. Oligonucleotide probes for detection and differentiation of *Staphylococcus aureus* strains containing genes for enterotoxin A, B, C, and toxic shock syndrome toxin-1. J Clin Microbiol 1990;28:1514–8.
- [13] Sharma NK, Rees CED, Dodd CER. Development of a single-reaction multiplex PCR toxin typing assay for *Staphylococcus aureus* strains. Appl Environ Microbiol 2000;66:1347–53.
- [14] Gilligan K, Shipley M, Stiles B, Hadfield TL, Ibrahim MS. Identification of *Staphylococcus aureus* enterotoxins A and B genes by PCR-ELISA. Mol Cell Probes 2000;14:71–8.
- [15] Klotz M, Opper S, Heeg K, Zimmermann S. Detection of *Staphylococcus aureus* enterotoxins A to D by real-time fluorescence PCR assay. J Clin Microbiol 2003;41(10):4683–7.
- [16] (a) Letertre C, Perelle S, Dilasser F, Fach P. Identification of a new putative enterotoxin SEU encoded by the egc cluster of *Staphylococcus aureus*. J Appl Microbiol 2003;95(1):38–43.  
(b) Letertre C, Perelle S, Dilasser F, Fach P. Detection and genotyping by real-time PCR of the staphylococcal enterotoxin genes sea to sej. Mol Cell Probe 2003;17(4):139–47.
- [17] Sambrook JE, Fritsh F, Maniatis T. Molecular cloning: a laboratory manual. NY, USA: Cold Spring Harbor Laboratory Press; 1989.
- [18] (a) Draisci R, Delli F, Quadri L, Achene G, Volpe L, Paleschi G. A new electrochemical enzyme-linked immunosorbent assay for the screening of macrolide antibiotic residue in bovine meat. Analysts 2001;126:1942–6.  
(b) Draisci R, Volpe G, Compagne D, Purificato I, Delli Quadri F, Paleschi G. Development of an electrochemical ELISA for the screening of 17B-estradiol and application to bovine serum. Analyst 2000;125:1419–23.
- [19] Mirhabibollahi B, Brooks JL, Kroll RG. Development and performance of an enzyme-linked amperometric immunosensor for the detection of *Staphylococcus aureus* in foods. J Appl Bacteriol 1990;68:577–85.