REPORT DOCUMENTATION PAGE				Form Approved OMB No. 0704-0188			
The public reporting burden for this collection or gathering and maintaining the data needed, and of information, including suggestions for redu (0704-0188), 1215 Jefferson Davis Highway, S subject to any penalty for failing to comply with PLEASE DO NOT RETURN YOUR FO	f information completing an cing the burd uite 1204, Ar a collection of RM TO TH	is estimated to average 1 hour d reviewing the collection of inf den, to Department of Defens lington, VA 22202-4302. Resg f information if it does not displa IE ABOVE ADDRESS.	per response, incl prmation. Send cor e, Washington Hea pondents should be y a currently valid (uding the tim mments regar adquarters S aware that i OMB control	ne for reviewing instructions, searching existing data sources, rding this burden estimate or any other aspect of this collection Services, Directorate for Information Operations and Reports notwithstanding any other provision of law, no person shall be number.		
1. REPORT DATE (DD-MM-YYYY)	REPORT DATE (DD-MM-YYYY) 2. REPORT TYPE			3. DATES COVERED (From - To)			
4. TITLE AND SUBTITLE				5a. CONTRACT NUMBER			
				5b. GRANT NUMBER			
				5c. PRC	OGRAM ELEMENT NUMBER		
6. AUTHOR(S)				5d. PRC	DJECT NUMBER		
				5e. TAS	SK NUMBER		
				5f. WORK UNIT NUMBER			
7. PERFORMING ORGANIZATION N	ame(s) ar	ND ADDRESS(ES)		L	8. PERFORMING ORGANIZATION REPORT NUMBER		
9. SPONSORING/MONITORING AGE	NCY NAM	IE(S) AND ADDRESS(ES)			10. SPONSOR/MONITOR'S ACRONYM(S)		
					11. SPONSOR/MONITOR'S REPORT NUMBER(S)		
12. DISTRIBUTION/AVAILABILITY S	FATEMEN'	Т					
13. SUPPLEMENTARY NOTES							
14. ABSTRACT							
15. SUBJECT TERMS							
16. SECURITY CLASSIFICATION OF		17. LIMITATION OF ABSTRACT	18. NUMBER OF	19a. NAI	ME OF RESPONSIBLE PERSON		
	a. REPORT b. ABSTRACT C. THIS PAGE PAGE PAGES 19b. TEL				PHONE NUMBER (Include area code)		



59th Medical Wing Institutional Animal Care and Use Committee (IACUC) 59 Clinical Research Division/SGVUS 1100 Wilford Hall Loop, Bldg 4430 Lackland AFB, TX 78236-5300

NOTICE OF ACTION REGARDING IACUC REVIEW

Date: 12 Dec 17

TO: Maj Joseph Maddry/59EMDS/SGO3D

Your **Final Report** was reviewed by the WHASC IACUC during the 14 Nov 17 meeting. The Committee's decision is provided below:

FWH20140100A (Research) "The Accuracy and Precision of the CONMED ECOMTM in the Porcine (Sus scrofa - Yorkshire cross) Model" PI: Maj Joseph Maddry/59 EMDS/SGO3D

Introduced at the 10 Oct 17 IACUC meeting: Due to conflicts of interest and early departures, the committee no longer met quorum requirements and therefore this item will be reviewed at the next convened meeting in November.

14 Nov 17: The committee voted that this item be approved as written. FOLLOW-UP: CLOSED

Name of Official Elizabeth L. Prince	Title/Office Symbol/Phone Office of Research Protocol Support /SGVUS/292-4210		
Signature		Info Copy To	

1. Protocol Number: FWH20140100A

2. Type of Research: Animal Research

3. Title: The Accuracy and Precision of the CONMED ECMTM in the Porcine Model (sus scrofa – Yorkshire cross)

4. Principal Investigator (PI):

Name	Rank	Date of IACUC Training	Branch of Service/ Corps	Staff Resident Fellow Civilian	Department / Office Symbol	Email (if other than WHASC Outlook)	Phone	Pager
Joseph Maddry	O-4	May 2014	USAF	Staff	59th EMDS/SG OED	<u>Joseph.madd</u> ry@gmail.co <u>m</u>	210-916- 3693	

5. Purpose: The purpose of this study was to explore the accuracy and precision of a recently FDA approved device, the CONMED endotracheal cardiac output monitor (ECOM) [™] (device information in appendix) apparatus, by comparing it to the PAC in under hemorrhagic and hypothermic conditions.

6. Results: Using GraphPad Prism® to conduct non-linear fit analyses comparing the slopes of the curves for ECOM versus PAC, we found that the curves from the ECOM data were significantly different from the PAC data curves under both conditions, but more pronounced differences were found under hemorrhagic conditions. Future studies will compare the ECOM stroke volume variability data to the PAC catheter data. Stroke volume variation has been shown to be a reliable predictor of fluid responsiveness. These data may compare favorably.

7. How may your findings benefit the Air Force? If the CONMED ECOM proves to be accurate and precise, management of hypovolemic shock can be more precisely targeted to the need for fluid rather than vasoactive medications. Findings from studies may suggest improved short term and long-term survival for our injured service members with targeted care. Because the injured service member will very commonly have need for placement of an endotracheal tube for management of a patent airway. This device can potentially simplify acquisition of data while protecting the service member from the added potential complications associated with placement and maintenance of the PAC.

8. Number of Animals

Projected Enrollment of Animals at the Beginning of Study: 20

Actual Number of Animals Enrolled: 20

9. Status of Animals Entered into the Protocol: The animals were in good general health, and were euthanized at the end of the study per protocol.

10. Number of animals since last status report:

	Number enrolled since last report	Total enrollment to date
Number of animals entered into the Study	0	20

11. Status of Funds: There are no remaining funds. All funds were used accordingly.

12. Reason for Closure: Study is complete

13. Specific Problems: At the beginning of the study we identified a few issues with the ECOM monitor and tube. The tube was very positional and sensitive to movement of the animal. The ECOM monitor was sent back to the company for updates to the software that would allow the monitor to record information pertinent to the study.

14. Publications and Presentations:

Presentations:

"Comparison of an endotracheal cardiac output monitor to a pulmonary artery catheter" Tri Service Nursing Research Program; Elliot City, Maryland; April 25, 2017 "Comparison of an endotracheal cardiac output monitor to a pulmonary artery catheter" American Association of Nurse Anesthetists; Seattle, Washington; Sept 8 - 12, 2017

These Presentations and Publications have been cleared by 59 CRD and Public Affairs

Publications: None

- 15. Exceptional Achievements: None
- 16. Signature of Principal Investigator:

Joseph Maddry, MD Maj, USAF, MC, FS Emergency Physician/Medical Toxicologist Director USAF En route Care Research Center Director, Clinical Research, Emergency Sciences and Toxicology 59th MDW/ST Chief Scientist Office U.S. Army Institute of Surgical Research